



# SAR Test exclusion documentation according to FCC KDB 447498, RSS-102 and EN 62479

Report identification number: 1-5229/17-02-19

Certification numbers and labeling requirements				
FCC ID	U28AUCIC			
IC number	1350B-AUCIC			
HVIN (Hardware Version Identification Number)	Aurora CIC			
PMN (Product Marketing Name)	Aurora CIC			
FVIN (Firmware Version Identification Number)	-/-			
HMN (Host Marketing Name)	-/-			

This test report is electronically signed and valid without handwriting signature. For verification of the electronic signatures, the public keys can be requested at the testing laboratory.

# **Document authorized:**

Thomas Vogler Lab Manager Radio Communications & EMC



#### EUT technologies:

Technologies:	Max. power: (AVG)	Max. gain:	
3.84 MHz module )*		Fieldstrength 50.8 dBµV/m @ 1 m	

)\* exempted from routine evaluation for FCC, for RSS-102 see additional test report of the frequency range 3 kHz – 10 MHz

#### SAR test exclusion according to KDB447498 (General RF Exposure Guidance)

Equations from Chapter 4.3.1: Standalone SAR test exclusion considerations page 11 and ff. and tables in Annex C

The table below gives the calculated maximal power that could be used for source based time averaged conducted or radiated power, adjusted for tune up tolerance. If this is at or below the calculated value the DUT is exempted from SAR evaluation.

f in [MHz]	d <sub>separation</sub> [mm]	Threshold <sub>1-g</sub>	Powerlimit [mW]	P <sub>max-declared</sub> [mW]	Exclusion
0.10	< 50	3	948.00	< 1 mW	yes

## SAR test exclusion according to RSS-102 Issue 5 Section 2.5.1/Table 1

The table below gives the calculated maximal power that could be used for source based time averaged conducted or radiated power, adjusted for tune up tolerance. If this is at or below the calculated value the DUT is exempted from SAR evaluation.

f in [MHz]	d <sub>separation</sub> [mm]	tissue volume	Powerlimit [mW]	P <sub>max-declared</sub> [mW]	Exclusion
< 300	5	1 g	71.00	< 1 mW	yes

### SAR test exclusion according to EN 62479

Compliance is given according to EN 62479 because the output power of the DUT is smaller than 20 mW.