

**Product Equality Declaration**

We,

Oticon A/S  
 Kongebakken 9  
 DK-2765 Smørum  
 Denmark

Declare on our sole responsibility that the radio part of DUAL Connect with FCC ID U28FUDS01 and IC: 1350B-FUDS01,

is identical with respect to wireless performance with style RITE series, certified under FCC ID: U28FUBTE01 and IC: 1350B-FUBTE01. Therefore Test Report No. 5-5766-04-03/07 continues to apply.

The electrical behaviour of all styles and models are identical because each of the audio amplifier modules includes the same radio module for wireless functionality.

Technical description: Oticon hearing aid family

Operation frequency : 3.84MHz  
 RF-output Power : ≤-42dBμA/m @10m  
 Type of Modulation : AM  
 ITU-Designation : 388KA1D  
 Number of Channels : 1  
 Antenna : Integrated coil antenna  
 Operating Temperature range : 0°C to +50°C

Modules and the hearing aid modules have been tested to prove evidence for compliance to the standards as stated in the table below.

Standard	Test report reference
EN 300 330-1, V1.5.1	Test Report No. 2-4458-02-02 BTE/06 Test Report No. 2-4458-02-02c/06 Test Report No. 2-4458-02-02 RITE/06 Test Report No. 5-5766-04-02/07 Testing Laboratory: CETECOM ICT Services GmbH
FCC, CFR Part 15.209, Part 15.223 IC and RSS 210, Issue 6	Test Report No. 2-4458-02-06b/06 Test Report No. 2-4458-02-06c/06 Test Report No. 5-5766-04-03/07 Testing Laboratory: CETECOM ICT Services GmbH
EN 301 489-1, V1.6.1 EN 301 489-3, V1.4.1	Test Report No. 2-4458-02-03/06 Test Report No. 5-5766-04-05/07 Testing Laboratory: CETECOM ICT Services GmbH
IEC 60065:2001 + A1:2005	Test Report No. NC9631-A2-CB-1

EN 60601-1 (3rd edition)	Testing Laboratory: UL International Demko A/S Test Report No. 5-5766-4-6/07 Testing Laboratory: CETECOM ICT Services GmbH
--------------------------	--

Name: Kristine Klitgaard Pedersen  
Title: Wireless Regulatory Manager

Company: Oticon A/S  
Address: Kongebakken 9  
DK-2765 Smørum  
Denmark

Email: [kkp@oticon.dk](mailto:kkp@oticon.dk)  
Phone: +45 3913 8583

22/10-2008   
(Date and Signature)