

## Class II Permissive Change (C2PC)

2021-03-15

### Federal Communication Commission

Equipment Authorization Division, Application Processing Branch  
7435 Oakland Mills Road  
Columbia, MD 21048

### Certification and Engineering Bureau

Industry Canada  
Spectrum Engineering Branch  
3701 Carling Avenue, Building 94  
Ottawa, Ontario K2H 8S2

### TO WHOM IT MAY CONCERN

Pursuant to the applicable paragraphs in ISED's RSP-100 §7 and CFR 47 FCC part §2.1043(b)(2) we hereby declare the following changes under Class II Permissive Change to the product:

Hearing Instrument amplifier module: Aurora miniRITE

**FCC ID: U6XAUMRIT**

**IC: 7031A-AUMRIT**

Original grant: May 29<sup>th</sup>-2017 and last C2PC amended grant: January 12<sup>th</sup> -2018.

We, Bernafon AG hereby declare the following Class 2 Permissive Change to the amplifier module Aurora miniRITE related to the FCC part 15.247 / RSS-247 DTS radio:

- The change of Bluetooth BAW filter and ESD protection diode due to "End-of-Life" components changes has been performed with a minor PCB layout change handled in correlation with KDB 178919 D01, section III, subclause D + Page 3, Note 5, bullet 2.
- Only passive components have been changed.
- The Limited Modular Approval (LMA) still apply after the C2PC product update.
- No mechanical changes has been applied to the product for this C2PC.
- The radio module performance is equal to the previously certified module performance.
- The device capability of operating the transmitter in 2 Mbps mode has never been implemented in any application since first product launch. The functionality has thus been disabled in the SW.

Bernafon AG  
Morgenstrasse 131 Phone +41 31 998 15 15  
P.O. Box 550 Fax +41 31 998 15 90  
CH-3018 Bern info@bernafon.ch  
Switzerland www.bernafon.com

**bernafon**<sup>®</sup>  
Your hearing · Our passion

We, Bernafon AG declare that the hearing aid amplifier module changes as documented in the test report will have a similar module performance as established in the FCC/ISED original certification.

If you have any questions, please feel free to contact me at the address shown below.

Signed on behalf of Bernafon AG,



Per Klaus Nielsen, Certification Manager

Demant A/S  
Kongebakken 9  
DK-2765 Smørum, Denmark

Phone : +45 3917 7100  
Email : [pkni@bernafon.com](mailto:pkni@bernafon.com)