Bernafon AG Switzerland

Morgenstrasse 131 Phone +41 31 998 15 15 Phone +41 31 998 15 Phone +41 31 998 15 Phone +41 31 998 15 Phone Phon www.bernafon.com



Federal Communication Commission

Equipment Authorization Division, Application Processing Branch 7435 Oakland Mills Road Columbia, MD 21048

Certification and Engineering Bureau

Industry Canada Spectrum Engineering Branch 3701 Carling Avenue, Building 94 Ottawa, Ontario K2H 8S2, Canada

December 20th, 2018.

Declaration for IC and FCC Label Exemption for Bernafon AG hearing aid module

With reference to Industry Canada, RSP-100 issue 11 chapter 3 Certification of Radio Apparatus and KDB 784748 D01v08 GUIDELINES FOR LABELLING AND USER INFORMATION FOR DEVICES SUBJECT TO PART 15 AND PART 18, Bernafon hereby declare an exemption regarding the labelling of the IC and FCC certification number on the Aurora BTE power wireless hearing aid families. This exception applied also to Bernafon affiliated OEMs and private label customers using the Aurora miniRITE 312 T Rc radio module in their hearing aids.

The granted exemption for the identical Oticon Aurora miniRITE 312 T Rc radio module under IC and FCC ID 1350B-AUMRTRC and U28AMRTRC respectively is used for the Bernafon exemption declaration.

The Bernafon IC and FCC certification number: 7031A-AUMRTRC and U6XAUMRTRC is located next to the device model number in the user manual and/or on the product packaging material. The exemption declaration is based on the fact that the hearing aids will be too small (Final product size 31.0 x 12.7 x 8.2 mm) for the IC and FCC certification number to be legible without the aid of magnification while printed on a surface of the hearing aid not exposed to wear and user functions.

For your information, our TCB is CTC Advanced GmbH in Germany.

Sincerely,

muhammad Asif

Asif Muhammad (On behalf of Bernafon AG)

Director, Regulatory Affairs, QARA Oticon A/S Kongebakken 9 2765 Smoerum **DK-Denmark**

Phone: +45 3913 7100