

Berne, 15 March 2013

Federal Communication Commission  
Equipment Authorization Division, Application Processing Branch  
7435 Oakland Mills Road  
Columbia, MD 21048

Certification and Engineering Bureau  
Industry Canada  
Spectrum Engineering Branch  
3701 Carling Avenue, Building 94  
Ottawa, Ontario K2H 8S2

## **Declaration on Electrically Identical Models**

TO WHOM IT MAY CONCERN

We, Bernafon AG, Morgenstrasse 131, 3018 Bern, Switzerland, declare on our sole responsibility that the new models

**AR9 ITCPD, AR7 ITCPD, PC9 ITCPD, PC7 ITCPD**  
**AR9 ITED, AR7 ITED, PC9 ITED, PC7 ITED**  
**AR9 ITCD, AR7 ITCD, PC9 ITCD and PC7 ITCD**  
with FCC ID: **U6XF2ITE01** and IC: **7031A-F2ITE01**

are electrically identical to the certified model

**ITC/ITE Fusion 2** under FCC ID: **U28FU2ITE** and IC: **1350B-FU2ITE**

The products are hearing instruments. The only differences between the new models and the original model are:

- Models xx9 yyy and xx7 yyy offer different audiological signal processing. This is implemented by different parameter sets for the digital signal processing of the audio path.
- Models xxnn ITCPD, xxnn ITED and xxnn ITCD offer different electro-acoustical performance (maximum possible sound pressure level). This is implemented by the use of different electro-acoustic transducers (loudspeaker).
- The prefixes AR and PC indicate different product families.

We attest that above changes are not relevant for any RF behavior subject to regulatory items. All products use the same electronic module as tested and described in Cetecom Test Report 1-4852/12-01-02.

Therefore Cetecom Test Report 1-4852/12-01-02 remains applicable.

continued

If you have any questions, please feel free to contact us at the address shown below

Sincerely,



---

Name:	Thomas Leuzinger
Function:	Team Leader Electro-Acoustic Design
Company:	Bernaфон AG
Address:	Morgenstrasse 131 3018 Bern Switzerland
Phone:	+41 (31) 998 16 92
Fax:	+41 (31) 998 15 90
Email:	<a href="mailto:tle@bernaфон.ch">tle@bernaфон.ch</a>