

Carestream

Carestream Health Inc.
150 Verona Street
Rochester, NY 14608

Oct. 4, 2010

Federal Communications Commission
Authorization and Evaluation Division
7435 Oakland Mills Road
Columbia, MD 21046

Attn: Equipment Authorization Branch

Ref: Class II Permissive Change for FCC ID: U72DRX1 and for FCC ID: U72DRX1-4

The purpose of this letter is to request a Class II Permissive change for FCC ID: U72DRX1 and FCC ID: U72DRX1-4 the changes made to these applications is as follows:

1. The single band antenna used for frequencies 5180.0- 5240.0 MHz in the DRX1 and the single band antenna for 5745 – 5805 MHz in the DRX1-4 are being replaced by a single antenna that operates in both frequency ranges.
2. The radio continues to transmit in frequencies 5180 – 5230 MHz with output power of 0.05 watts for FCC ID: U72DRX1 and in frequencies 5745 – 5805 MHz at an output power of 0.0389 watts for the FCC ID: U72DRX1-4 radio.
3. The Transmitters remains identical as described in all documentation for the initial certification of FCC ID:U72DRX1 and FCC ID: U72DRX1-4
4. No changes were made in the transmitter section. The firmware allowed the radio to operate on the additional channels using the same modulation and power output.

Respectfully Submitted,



Ronald L Cain
CarestreamHealth Inc.
150 Verona Street
Rochester NY, 14608