



● FPO SFDA CERTIFICATION NUMBER

● PLACE MANUFACTURER'S ID MARK HERE

4-R1.70

NOTES:

- 1) MATERIAL, .010" THICK POLYCARBONATE, MATTE FINISH, WITH 3M 467 PRESSURE SENSITIVE ADHESIVE, FARSIDE.
- 2) PARTS SHALL BE SUPPLIED DIE-CUT TO SHAPE WITH PEEL-OFF LINER.
- 3) COLORS: BACKGROUND TO BE SILVER LINING (PANTONE 428), TEXT AND SYMBOLS TO BE ECLIPSE EXCEPT WHERE INDICATED.
- 4) SHALL MEET THE DURABILITY REQUIREMENTS OF UL 60950-1.
- 5) PLACE WRITE-ON OVERLAMINATE OVER ENTIRE SURFACE.
- 6) LABEL TO BE SERIALIZED STARTING WITH 0101 AND RUNNING SEQUENTIALLY,VENDOR TO TRACK NUMBERS.
- 7) SEE PACKAGING GROUP FOR ARTWORK.

NO. **2G5555**
SHEET 1 NEXT SHEET 0

CHG. NO. DATE REL	REVISIONS	DR BY APPS.
REV A 04/16/09	INITIAL RELEASE	DR

UNLESS OTHERWISE SPECIFIED
LINEAR TOLERANCE ± 1.6
ANGULAR TOLERANCE \pm —
DIMENSIONS ARE IN MILLIMETERS

MD

CARESTREAM HEALTH, INC.
ROCHESTER, NY 14608

DR. WESTON	DATE 04-16-09	CK
DFTG. JEM	DES. ENG. L. CHEN	MFG. ENG. L. CHEN

NAME
LABEL-DOCKING STATION, WIRELESS

SCALE 1X SIZE DWG SIZE **A**

NO. **2G5555**
SHEET 1 NEXT SHEET 0

Attachment 4. Reference photo

Docking station Left view



Docking station Bottom view

