

# RF Exposure Evaluation Report

**Product** : Digital Blood Pressure Monitor  
**Trade mark** : **microlife**  
**Model/Type reference** : BP3GT1-6F  
**Serial Number** : N/A  
**Report Number** : EED32N00021902  
**FCC ID** : U7I-BP3GT1-6F  
**Date of Issue** : Apr. 19, 2021  
**Test Standards** : 47 CFR Part 1.1307  
47 CFR Part 2.1093  
KDB447498D01 General RF Exposure Guidance v06  
**Test result** : PASS

Prepared for:

**Microlife Corporation**  
9F, 431, RuiGuang Road, NeiHu Taipei 11492, Taiwan

Prepared by:

**Centre Testing International Group Co., Ltd.**  
Hongwei Industrial Zone, Bao'an 70 District,  
Shenzhen, Guangdong, China  
**TEL: +86-755-3368 3668**  
**FAX: +86-755-3368 3385**



Compiled by:

*Martin Lee*

Martin Lee

Approved by:

*David Wang*

David Wang

Reviewed by:

*Aaron Ma*

Aaron Ma

Date:

Apr. 19, 2021

Check No.:4538028527

## 2 Version

Version No.	Date	Description
00	Apr. 19, 2021	Original

### 3 Contents

	Page
<b>1 COVER PAGE.....</b>	<b>1</b>
<b>2 VERSION.....</b>	<b>2</b>
<b>3 CONTENTS.....</b>	<b>3</b>
<b>4 GENERAL INFORMATION.....</b>	<b>4</b>
4.1 CLIENT INFORMATION.....	4
4.2 GENERAL DESCRIPTION OF EUT.....	4
4.3 TEST LOCATION.....	5
4.4 DEVIATION FROM STANDARDS.....	5
4.5 ABNORMALITIES FROM STANDARD CONDITIONS.....	5
4.6 OTHER INFORMATION REQUESTED BY THE CUSTOMER.....	5
<b>5 SAR EVALUATION.....</b>	<b>6</b>
5.1 RF EXPOSURE COMPLIANCE REQUIREMENT.....	6
5.1.1 Standard Requirement.....	6
5.1.2 EUT RF Exposure.....	7
<b>PHOTOGRAPHS OF EUT CONSTRUCTIONAL DETAILS.....</b>	<b>8</b>

## 4 General Information

### 4.1 Client Information

Applicant:	Microlife Corporation
Address of Applicant:	9F, 431, RuiGuang Road, NeiHu Taipei 11492, Taiwan
Manufacturer:	ONBO Electronic (Shenzhen) Co., Ltd.
Address of Manufacturer:	No.138, Huasheng Road, Langkou Community, Dalang Street, Longhua District, Shenzhen, China
Factory:	ONBO Electronic (Shenzhen) Co., Ltd.
Address of Factory:	No.138, Huasheng Road, Langkou Community, Dalang Street, Longhua District, Shenzhen, China

### 4.2 General Description of EUT

Product Name:	Digital Blood Pressure Monitor
Mode No.:	BP3GT1-6F
Trade Mark:	<b>microlife</b>
Product Type:	<input type="checkbox"/> Mobile <input checked="" type="checkbox"/> Portable <input type="checkbox"/> Fix Location
Bluetooth Version:	V4.0
Operation Frequency:	2402MHz~2480MHz
Modulation Type:	GFSK
Transfer Rate:	<input checked="" type="checkbox"/> 1Mbps <input type="checkbox"/> 2Mbps
Number of Channel:	40
Antenna Type:	PCB antenna
Antenna Gain:	0dBi
Power Supply:	4*AAA Battery or Switching Adapter MODEL: DSA-6E-05 US INPUT: 100-240V~50/60Hz 0.3A Output: 6V 0.6A
Sample Received Date:	Apr. 8, 2021
Sample tested Date:	Apr. 8, 2021 to Apr. 13, 2021
Company Name and Address shown on Report, the sample(s) and sample Information was/ were provided by the applicant who should be responsible for the authenticity which CTI hasn't verified.	

### 4.3 Test Location

All tests were performed at:

Centre Testing International Group Co., Ltd

Building C, Hongwei Industrial Park Block 70, Bao'an District, Shenzhen, China

Telephone: +86 (0) 755 33683668 Fax:+86 (0) 755 33683385

No tests were sub-contracted.

FCC Designation No.: CN1164

### 4.4 Deviation from Standards

None.

### 4.5 Abnormalities from Standard Conditions

None.

### 4.6 Other Information Requested by the Customer

None.

## 5 SAR Evaluation

### 5.1 RF Exposure Compliance Requirement

#### 5.1.1 Standard Requirement

According to KDB447498D01 General RF Exposure Guidance v06  
Standalone SAR test exclusion considerations

Unless specifically required by the published RF exposure KDB procedures, standalone 1-g head or body and 10-g extremity SAR evaluation for general population exposure conditions, by measurement or numerical simulation, is not required when the corresponding SAR Exclusion Threshold condition, listed below, is satisfied.

#### Limits

The 1-g and 10-g SAR test exclusion thresholds for 100 MHz to 6 GHz at test separation distances  $\leq 50$  mm are determined by:

$$\left[ \frac{\text{max. power of channel, including tune-up tolerance, mW}}{\text{min. test separation distance, mm}} \right] \cdot \sqrt{f(\text{GHz})} \leq 3.0$$
 for 1-g SAR and  $\leq 7.5$  for 10-g extremity SAR, where  $f(\text{GHz})$  is the RF channel transmit frequency in GHz

Power and distance are rounded to the nearest mW and mm before calculation<sup>17</sup>

The result is rounded to one decimal place for comparison

The test exclusions are applicable only when the minimum test separation distance is  $\leq 50$  mm and for transmission frequencies between 100 MHz and 6 GHz. When the minimum test separation distance is  $< 5$  mm, a distance of 5 mm is applied to determine SAR test exclusion

### 5.1.2 EUT RF Exposure

#### Measurement Data

GFSK mode				
Test channel	Peak Output Power (dBm)	Tune up tolerance (dBm)	Maximum tune-up Power	
			(dBm)	(mW)
Lowest(2402MHz)	-0.25	-0.5±0.5	0	1.000
Middle(2440MHz)	-1.95	-2.0±0.5	-1.5	0.708
Highest(2480MHz)	-3.65	-4.0±0.5	-3.5	0.447

Worst case: GFSK						
Channel	Maximum Peak Conducted Output Power (dBm)	Tune up tolerance (dBm)	Maximum tune- up Power		Calculated value	Exclusion threshold
			(dBm)	(mW)		
Lowest (2402MHz)	-0.25	-0.5±0.5	0	1.000	0.31	3.0
Middle (2440MHz)	-1.95	-2.0±0.5	-1.5	0.708	0.22	
Highest (2480MHz)	-3.65	-4.0±0.5	-3.5	0.447	0.14	

Conclusion: the calculated value  $\leq 3.0$ , SAR is exempted.

Remark: The Max Conducted Peak Output Power data refer to report Report No.: EED32N00021901

## PHOTOGRAPHS OF EUT Constructional Details

Refer to Report No. EED32N00021901 for EUT external and internal photos.

The test report is effective only with both signature and specialized stamp, The result(s) shown in this report refer only to the sample(s) tested. Without written approval of CTI, this report can't be reproduced except in full.

\*\*\* End of Report \*\*\*