


DRAFT

SpO₂/Aux

IPX7



FCC ID: URA-MS24895
IC: 6895B-MS24895



Dräger



Caution: Federal Law in the United States restricts this device to sale by, or on the order of a physician.

This equipment utilizes RF energy. Maximum radiated power is 100mW EIRP in the frequency range of 2.400-2.4835 GHz. To minimize in-band interference, do not exceed more than eight similar devices operating within a 1 meter area.

Dräger Medical Systems, Inc.
3135 Quarry Road
Telford, PA 18969 USA

ID:



3.00"

type
drawing

language
en

title
Label Regulatory Valve M366

replacement for

number
MS17848

creating system
coreldraw

release status
released

revision index
02

occurs

page/of
2 / 2

Drägermedical
A Dräger and Siemens Company