



Date: 20.01.2016

Federal Communications Commission
Authorization and Evaluation Division
7435 Oakland Mills Road
Columbia, MD 21046

Attn: OET Dept.

Ref: FCC Class II Permissive change for FCC ID: URA-MS32018
Applicant: Draeger Medical Systems, Inc.

Dear Examiner,

This is to request a Class II permissive change for FCC ID: URA-MS32018,

Reason for Class II permissive change:

1) Changing of antenna:

New antenna type:

CU5006-02, Antenova (Omni), gain: 2.4 dBi

Antenna details are described in technical descriptions "CU5006 Rev02_Draeger.pdf" and "CU5006-2_Draeger_PS_Rev02.pdf".

2) Technical decrease of output power for body-worn devices

For details see our OET test report

If you have any questions regarding this application, please feel free to contact me.

Sincerely,

A handwritten signature in black ink, appearing to read 'Bettina Stangl'.

Bettina Stangl
Team Lead Global Market Access

CSA Group Bayern GmbH
Ohmstrasse 1-4
94342 Strasskirchen
Germany
Phone: +49 9424 9481 250
Fax: +49 9424 9481 440
Email: Bettina.Stangl@csagroup.org