

Antw Question on calibration interval.txt
From: wolfgang klos [Wolfgang.Klos@vde.com]
Sent: Thursday, April 14, 2011 3:55 AM
To: Clark, Randy
Cc: Stephan Kloska
Subject: Antw: Question on calibration interval
Attachments: 91571_Calibration info.XLS

Hello Randy,

I took the time to determine the time period between the day of test and the date of last calibration.
As you can see from the attached table, the date of the last calibration of all items was less than 2 years ago at the time of testing for all equipment except for the horn antenna used for measurements above 18 GHz. For this antenna the calibration intervall was extended to five years, with respect to Note 2 in section 4.4.2 of ANSI C63.4:2003.

Therefore we think that you can trust our test data.
When evaluating the test results for the current project, you may also consider that we did not see any emissions above the noise floor at frequencies above 10 GHz.

Mit freundlichen Grüßen/
with best regards
wolfgang klos

VDE Prüf- und Zertifizierungsinstitut GmbH/ VDE Testing and Certification
Institute

EMV und Akustik, Fachgebiet FG43/
EMC and Acoustics, Section FG43

Tel.: +49 69 8306 - 218
Fax: +49 69 8306 - 620

e-mail: wolfgang.klos@vde.com
www.vde.com