

RF Exposure Evaluation

Product : Blood Pressure Monitor

Model Number : BPM18R


Applicant : AVITA Corporation


Address : 9F, No.78, Sec. 1, Kwang-Fu Road, San-Chung,
New Taipei City, Taiwan

Issued By : Interocean EMC Technology Corp.

LAB Location : No. 5-2, Lin 1, Tin-Fu, Lin-Kou Dist., New Taipei City,
Taiwan 244, R.O.C.

Report Issued : 2017/11/06

Project Engineer : 
Elli Chang

Approved : 
Jerry Liu

RF Exposure Evaluation

Portable Device

According to §15.247(i) and §1.1307(b)(1), systems operating under the provisions of this section shall be operated in a manner that ensures that the public is not exposed to radio frequency energy level in excess of the Commission's guidelines. According to KDB 447498_D01_V06 4.3.1(a)

SAR exclusion thresholds by:

[max. power of channel, including tune-up tolerance, mW]/(min, test separation distances, mm)]* $[\sqrt{f(\text{GHz})}] \leq 3.0$ for 1-g SAR and ≤ 7.5 for 10-g extremity SAR.

Maximum measured transmitter power

Frequency Range (MHz)	Maximum measured transmitter power (dBm)	Tune-up power tolerance (dB)	Total Maximum power (dBm)	Total Maximum power (mW)
2402	-8.30	(±)2	-6.30	0.234
2440	-9.44	(±)2	-7.44	0.180
2480	-10.39	(±)2	-8.39	0.145

$$(0.234/5)*(\sqrt{2.402}) = 0.073 \leq 3$$

Conclusion: No SAR is required.

SIMULTANEOUS TRANSMISSION EVALUATION

N/A