



FCC ID : UWDA00044109
Model : A00044109
Applicant : Beckman Coulter, Inc.

Date: January 19, 2007
Page: 1 of 2

Request for Limited Modular Approval

The modular transmitter (Model: A00044109) is designed by Beckman Coulter, Inc. This modular transmitter will be equipped only on Pre-analytical Automation System (PAS)[Controlling ID of patient's blood samples for hematological examination.]. The modular transmitter complies with the module basic requirements (Item 1 to 8) in FCC Public Notice: Part 15 Unlicensed Modular Transmitter Approval (DA00-1407) as indicated below:

[Module basic requirement]

(1)Have its own RF shielding.: This modular transmitter has it's own RF shielding. Please refer to the A1 Photo.

This part has been submitted as separate exhibit.

A1 Photo (Modular transmitter)

- (2) The modular transmitter must have buffered modulation/data inputs.:
Data input to Modular transmitter is controlled as not generating excessive data speed and abnormal modulation.
- (3)The modular transmitter must have its own power supply regulation.:
The modular transmitter does not have its own power supply, so that the regulated power of 5V is supplied from the end product. Therefore the use of this device is limited to the end product which has 5V stabilized power source.
- (4)The modular transmitter must comply with the antenna requirements of Section 15.203 and 15.204(c):
The antenna connector of this product is a PH Connector produced by J.S.T.Mfg Co.,Ltd. It deems to be special. Moreover, this product is installed in the PAS by professional installation. Therefore, the equipment complies with the antenna requirement of 15.203 and 15.204(c).
- (5)The modular transmitter must be tested in a stand alone.: The modular transmitter has been tested as a stand alone and then confirmed the compliance. Please refer to the Test report.
- (6)The modular transmitter must be labeled with its own FCC ID number, and, if the FCC ID number is not visible when the module is installed inside another device, then the outside of the device into which the module is installed must also display a label referring to the enclosed module. This modular transmitter complies with this requirement. Contains FCC ID: UWDA00044109 will be indicated on outside of this product.


2007-900-0011

Beckman Coulter, Inc.
Corporate Headquarters
4300 N. Harbor Blvd.
P.O. Box 3100
Fullerton, CA 92834-3100

Telephone: (714) 871-4848
Facsimile: (714) 773-8283
Internet: www.beckmancoulter.com

(7) The modular transmitter must comply with any specific rule or operating requirements applicable to the transmitter and the manufacturer must provide adequate instructions along with the module to explain any such requirements.: The necessary explanations to comply with this requirement are contained in the manual.

(8) The modular transmitter must comply with any applicable RF exposure requirements.: This equipment complies with FCC radiation exposure limits set forth for uncontrolled equipment and meets with the FCC radio frequency (RF) Exposure Guidelines in Supplement C to OET 65. This equipment has very low levels of RF energy that it is deemed to comply without testing of specific absorption ratio (SAR).



Jeffrey Gonzalez

Manager, Product Compliance Engineering