#### Calibration Laboratory of Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland



S Schweizerischer Kalibrierdienst

- C Service suisse d'étalonnage
- Servizio svizzero di taratura
- S Swiss Calibration Service

Accreditation No.: SCS 0108

Accredited by the Swiss Accreditation Service (SAS) The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

#### Client Sporton

Certificate No: D900V2-190\_Sep20

# CALIBRATION CERTIFICATE

	D900V2 - SN:190	0			
Calibration procedure(s)	QA CAL-05.v11 Calibration Procedure for SAR Validation Sources between 0.7-3 GHz				
Calibration date:	September 04, 20	020			
		onal standards, which realize the physical uni robability are given on the following pages an			
All calibrations have been conducte	ed in the closed laborator	ry facility: environment temperature $(22 \pm 3)^{\circ}$ C	C and humidity < 70%.		
Calibration Equipment used (M&TE	critical for calibration)				
Primary Standards	ID #	Cal Date (Certificate No.)	Scheduled Calibration		
Power meter NRP	SN: 104778	01-Apr-20 (No. 217-03100/03101)	Apr-21		
Power sensor NRP-Z91	SN: 103244	01-Apr-20 (No. 217-03100)	Apr-21		
	SN: 103244 SN: 103245	01-Apr-20 (No. 217-03100) 01-Apr-20 (No. 217-03101)	Apr-21 Apr-21		
Power sensor NRP-Z91					
Power sensor NRP-Z91 Reference 20 dB Attenuator	SN: 103245	01-Apr-20 (No. 217-03101)	Apr-21		
Power sensor NRP-Z91 Reference 20 dB Attenuator Type-N mismatch combination	SN: 103245 SN: BH9394 (20k)	01-Apr-20 (No. 217-03101) 31-Mar-20 (No. 217-03106)	Apr-21 Apr-21		
Power sensor NRP-Z91 Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4	SN: 103245 SN: BH9394 (20k) SN: 310982 / 06327	01-Apr-20 (No. 217-03101) 31-Mar-20 (No. 217-03106) 31-Mar-20 (No. 217-03104)	Apr-21 Apr-21 Apr-21		
Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4	SN: 103245 SN: BH9394 (20k) SN: 310982 / 06327 SN: 7349	01-Apr-20 (No. 217-03101) 31-Mar-20 (No. 217-03106) 31-Mar-20 (No. 217-03104) 29-Jun-20 (No. EX3-7349_Jun20)	Apr-21 Apr-21 Apr-21 Jun-21		
Power sensor NRP-Z91 Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards	SN: 103245 SN: BH9394 (20k) SN: 310982 / 06327 SN: 7349 SN: 601	01-Apr-20 (No. 217-03101) 31-Mar-20 (No. 217-03106) 31-Mar-20 (No. 217-03104) 29-Jun-20 (No. EX3-7349_Jun20) 27-Dec-19 (No. DAE4-601_Dec19)	Apr-21 Apr-21 Apr-21 Jun-21 Dec-20		
Power sensor NRP-Z91 Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards	SN: 103245 SN: BH9394 (20k) SN: 310982 / 06327 SN: 7349 SN: 601	01-Apr-20 (No. 217-03101) 31-Mar-20 (No. 217-03106) 31-Mar-20 (No. 217-03104) 29-Jun-20 (No. EX3-7349_Jun20) 27-Dec-19 (No. DAE4-601_Dec19) Check Date (in house)	Apr-21 Apr-21 Apr-21 Jun-21 Dec-20 Scheduled Check		
Power sensor NRP-Z91 Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards Power meter E4419B	SN: 103245 SN: BH9394 (20k) SN: 310982 / 06327 SN: 7349 SN: 601 ID # SN: GB39512475	01-Apr-20 (No. 217-03101) 31-Mar-20 (No. 217-03106) 31-Mar-20 (No. 217-03104) 29-Jun-20 (No. EX3-7349_Jun20) 27-Dec-19 (No. DAE4-601_Dec19) Check Date (in house) 30-Oct-14 (in house check Feb-19)	Apr-21 Apr-21 Apr-21 Jun-21 Dec-20 Scheduled Check In house check: Oct-20		
Power sensor NRP-Z91 Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards Power meter E4419B Power sensor HP 8481A	SN: 103245 SN: BH9394 (20k) SN: 310982 / 06327 SN: 7349 SN: 601 ID # SN: GB39512475 SN: US37292783	01-Apr-20 (No. 217-03101) 31-Mar-20 (No. 217-03106) 31-Mar-20 (No. 217-03104) 29-Jun-20 (No. EX3-7349_Jun20) 27-Dec-19 (No. DAE4-601_Dec19) Check Date (in house) 30-Oct-14 (in house check Feb-19) 07-Oct-15 (in house check Oct-18)	Apr-21 Apr-21 Apr-21 Jun-21 Dec-20 Scheduled Check In house check: Oct-20 In house check: Oct-20		
Power sensor NRP-Z91 Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards Power meter E4419B Power sensor HP 8481A Power sensor HP 8481A RF generator R&S SMT-06	SN: 103245 SN: BH9394 (20k) SN: 310982 / 06327 SN: 7349 SN: 601 ID # SN: GB39512475 SN: US37292783 SN: MY41092317	01-Apr-20 (No. 217-03101) 31-Mar-20 (No. 217-03106) 31-Mar-20 (No. 217-03104) 29-Jun-20 (No. EX3-7349_Jun20) 27-Dec-19 (No. DAE4-601_Dec19) Check Date (in house) 30-Oct-14 (in house check Feb-19) 07-Oct-15 (in house check Oct-18) 07-Oct-15 (in house check Oct-18)	Apr-21 Apr-21 Apr-21 Jun-21 Dec-20 Scheduled Check In house check: Oct-20 In house check: Oct-20 In house check: Oct-20		
Power sensor NRP-Z91 Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards Power meter E4419B Power sensor HP 8481A Power sensor HP 8481A	SN: 103245 SN: BH9394 (20k) SN: 310982 / 06327 SN: 7349 SN: 601 ID # SN: GB39512475 SN: US37292783 SN: MY41092317 SN: 100972	01-Apr-20 (No. 217-03101) 31-Mar-20 (No. 217-03106) 31-Mar-20 (No. 217-03104) 29-Jun-20 (No. EX3-7349_Jun20) 27-Dec-19 (No. DAE4-601_Dec19) Check Date (in house) 30-Oct-14 (in house check Feb-19) 07-Oct-15 (in house check Oct-18) 07-Oct-15 (in house check Oct-18) 15-Jun-15 (in house check Oct-18)	Apr-21 Apr-21 Apr-21 Jun-21 Dec-20 Scheduled Check In house check: Oct-20 In house check: Oct-20 In house check: Oct-20 In house check: Oct-20 In house check: Oct-20		
Power sensor NRP-Z91 Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards Power meter E4419B Power sensor HP 8481A Power sensor HP 8481A RF generator R&S SMT-06	SN: 103245 SN: BH9394 (20k) SN: 310982 / 06327 SN: 7349 SN: 601 ID # SN: GB39512475 SN: US37292783 SN: MY41092317 SN: 100972 SN: US41080477	01-Apr-20 (No. 217-03101) 31-Mar-20 (No. 217-03106) 31-Mar-20 (No. 217-03104) 29-Jun-20 (No. 217-03104) 29-Jun-20 (No. EX3-7349_Jun20) 27-Dec-19 (No. DAE4-601_Dec19) Check Date (in house) 30-Oct-19 (No. DAE4-601_Dec19) 07-Oct-15 (in house check Feb-19) 07-Oct-15 (in house check Oct-18) 15-Jun-15 (in house check Oct-18) 31-Mar-14 (in house check Oct-19)	Apr-21 Apr-21 Apr-21 Jun-21 Dec-20 Scheduled Check In house check: Oct-20 In house check: Oct-20 In house check: Oct-20 In house check: Oct-20 In house check: Oct-20		

Calibration Laboratory of

Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland



Schweizerischer Kalibrierdienst

S Service suisse d'étalonnage

С Servizio svizzero di taratura

S **Swiss Calibration Service** 

Accreditation No.: SCS 0108

Accredited by the Swiss Accreditation Service (SAS) The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

## Glossarv:

tissue simulating liquid
sensitivity in TSL / NORM x,y,z
not applicable or not measured

# Calibration is Performed According to the Following Standards:

- a) IEEE Std 1528-2013, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", June 2013
- b) IEC 62209-1, "Measurement procedure for the assessment of Specific Absorption Rate (SAR) from hand-held and body-mounted devices used next to the ear (frequency range of 300 MHz to 6 GHz)", July 2016
- c) IEC 62209-2, "Procedure to determine the Specific Absorption Rate (SAR) for wireless communication devices used in close proximity to the human body (frequency range of 30 MHz to 6 GHz)", March 2010
- d) KDB 865664, "SAR Measurement Requirements for 100 MHz to 6 GHz"

# Additional Documentation:

e) DASY4/5 System Handbook

# Methods Applied and Interpretation of Parameters:

- Measurement Conditions: Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL: The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.
- Feed Point Impedance and Return Loss: These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.
- Electrical Delay: One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
- SAR measured: SAR measured at the stated antenna input power.
- SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

# **Measurement Conditions**

DASY system configuration, as far as not given on page 1.

DASY Version	DASY5	V52.10.4
Extrapolation	Advanced Extrapolation	and the second
Phantom	Modular Flat Phantom	
Distance Dipole Center - TSL	15 mm	with Spacer
Zoom Scan Resolution	dx, dy, dz = 5 mm	
Frequency	900 MHz ± 1 MHz	

# Head TSL parameters

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	41.5	0.97 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	42.0 ± 6 %	0.96 mho/m ± 6 %
Head TSL temperature change during test	< 0.5 °C		

## SAR result with Head TSL

SAR averaged over 1 cm <sup>3</sup> (1 g) of Head TSL	Condition	
SAR measured	250 mW input power	2.70 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	10.9 W/kg ± 17.0 % (k=2)

SAR averaged over 10 cm <sup>3</sup> (10 g) of Head TSL	condition	
SAR measured	250 mW input power	1.72 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	6.93 W/kg ± 16.5 % (k=2)

# Appendix (Additional assessments outside the scope of SCS 0108)

## **Antenna Parameters with Head TSL**

Impedance, transformed to feed point	49.7 Ω - 6.5 jΩ
Return Loss	- 23.7 dB

# **General Antenna Parameters and Design**

Electrical Delay (one direction)	1.418 ns

After long term use with 100W radiated power, only a slight warming of the dipole near the feedpoint can be measured.

The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard.

No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feedpoint may be damaged.

# Additional EUT Data

Manufactured by	SPEAG

# **DASY5 Validation Report for Head TSL**

Date: 04.09.2020

Test Laboratory: SPEAG, Zurich, Switzerland

## DUT: Dipole 900 MHz; Type: D900V2; Serial: D900V2 - SN:190


Communication System: UID 0 - CW; Frequency: 900 MHz Medium parameters used: f = 900 MHz;  $\sigma$  = 0.96 S/m;  $\epsilon_r$  = 42;  $\rho$  = 1000 kg/m<sup>3</sup> Phantom section: Flat Section Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2011)

DASY52 Configuration:

- Probe: EX3DV4 SN7349; ConvF(9.4, 9.4, 9.4) @ 900 MHz; Calibrated: 29.06.2020
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn601; Calibrated: 27.12.2019
- Phantom: Flat Phantom 4.9 (front); Type: QD 00L P49 AA; Serial: 1001
- DASY52 52.10.4(1527); SEMCAD X 14.6.14(7483)


# Dipole Calibration for Head Tissue/Pin=250 mW, d=15mm/Zoom Scan (7x7x7)/Cube 0:

Measurement grid: dx=5mm, dy=5mm, dz=5mmReference Value = 64.94 V/m; Power Drift = -0.02 dB Peak SAR (extrapolated) = 4.12 W/kg SAR(1 g) = 2.7 W/kg; SAR(10 g) = 1.72 W/kg Smallest distance from peaks to all points 3 dB below = 16 mm Ratio of SAR at M2 to SAR at M1 = 65.8% Maximum value of SAR (measured) = 3.59 W/kg



0 dB = 3.59 W/kg = 5.55 dBW/kg

# Impedance Measurement Plot for Head TSL





# D900V2, serial no. 190 Extended Dipole Calibrations

Referring to KDB 450824, if dipoles are verified in return loss (<-20dB, within 20% of prior calibration), and in impedance (within 5 ohm of prior calibration), the annual calibration is not necessary and the calibration interval can be extended.


#### <Justification of the extended calibration>

D <b>900</b> V2 – serial no. <b>190</b>						
		900MHZ				
Date of Measurement	Return-Loss (dB)	Delta (%)	Real Impedance (ohm)	Delta (ohm)	Imaginary Impedance (ohm)	Delta (ohm)
09.04.2020 (Cal. Report)	-23.712		49.676		-6.5067	
09.03.2021 (extended)	-21.323	-10.08	50.402	-0.726	-9.0583	2.5516

The return loss is < -20dB, within 20% of prior calibration; the impedance is within 5 ohm of prior calibration. Therefore the verification result should support extended calibration.



<Dipole Verification Data> - D900 V2, serial no. 190 (Data of Measurement : 09.03.2021) 900 MHz - Head



# **Calibration Laboratory of** Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland



Schweizerischer Kalibrierdienst

- Service suisse d'étalonnage
- С Servizio svizzero di taratura S
  - Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS) The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

Client Sporton Certificate No: DAE4-1311\_Aug21

Accreditation No.: SCS 0108

S

Dbject	DAE4 - SD 000 D04 BM - SN: 1311				
Calibration procedure(s)	QA CAL-06.v30 Calibration procee	dure for the data acquisition elec	tronics (DAE)		
Calibration date:	August 20, 2021				
		•			
		nal standards, which realize the physical uni bbability are given on the following pages an			
All calibrations have been conduc	cted in the closed laboratory	facility: environment temperature (22 ± 3)°C	and humidity - 70%		
		adding. contracting of a tor (22 1 0) C	rand number < r0.		
			s and furniting < 70%.		
Calibration Equipment used (M&			and the second of Marian Arts second		
Calibration Equipment used (M&T Primary Standards	TE critical for calibration)	Cal Date (Certificate No.) 07-Sep-20 (No:28647)	Scheduled Calibration Sep-21		
Calibration Equipment used (M&T Primary Standards Keithley Multimeter Type 2001	TE critical for calibration)	Cal Date (Certificate No.) 07-Sep-20 (No:28647)	Scheduled Calibration Sep-21		
Calibration Equipment used (M&T Primary Standards Keithley Multimeter Type 2001 Secondary Standards Auto DAE Calibration Unit	TE critical for calibration) ID # SN: 0810278 ID # SE UWS 053 AA 1001	Cal Date (Certificate No.) 07-Sep-20 (No:28647) Check Date (in house)	Scheduled Calibration		
Calibration Equipment used (M&T Primary Standards Keithley Multimeter Type 2001 Secondary Standards Auto DAE Calibration Unit Calibrator Box V2.1	TE critical for calibration) ID # SN: 0810278 ID # SE UWS 053 AA 1001	Cal Date (Certificate No.) 07-Sep-20 (No:28647) Check Date (in house) 07-Jan-21 (in house check)	Scheduled Calibration Sep-21 Scheduled Check In house check: Jan-22		
Calibration Equipment used (M&T Primary Standards Keithley Multimeter Type 2001 Secondary Standards Auto DAE Calibration Unit Calibrator Box V2.1	TE critical for calibration) ID # SN: 0810278 ID # SE UWS 053 AA 1001	Cal Date (Certificate No.) 07-Sep-20 (No:28647) Check Date (in house) 07-Jan-21 (in house check)	Scheduled Calibration Sep-21 Scheduled Check In house check: Jan-22		
Calibration Equipment used (M&T Primary Standards Keithley Multimeter Type 2001 Secondary Standards Auto DAE Calibration Unit	TE critical for calibration)           ID #           SN: 0810278           ID #           SE UWS 053 AA 1001           SE UMS 006 AA 1002	Cal Date (Certificate No.) 07-Sep-20 (No:28647) Check Date (in house) 07-Jan-21 (in house check) 07-Jan-21 (in house check)	Scheduled Calibration Sep-21 Scheduled Check In house check: Jan-22 In house check: Jan-22		

#### Calibration Laboratory of Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland





S Schweizerischer Kalibrierdienst

C Service suisse d'étalonnage

Servizio svizzero di taratura

S Swiss Calibration Service

Accreditation No.: SCS 0108

Accredited by the Swiss Accreditation Service (SAS) The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

# Glossary

DAE Connector angle data acquisition electronics information used in DASY system to align probe sensor X to the robot coordinate system.

# Methods Applied and Interpretation of Parameters

- DC Voltage Measurement: Calibration Factor assessed for use in DASY system by comparison with a calibrated instrument traceable to national standards. The figure given corresponds to the full scale range of the voltmeter in the respective range.
- Connector angle: The angle of the connector is assessed measuring the angle mechanically by a tool inserted. Uncertainty is not required.
- The following parameters as documented in the Appendix contain technical information as a
  result from the performance test and require no uncertainty.
  - DC Voltage Measurement Linearity: Verification of the Linearity at +10% and -10% of the nominal calibration voltage. Influence of offset voltage is included in this measurement.
  - Common mode sensitivity: Influence of a positive or negative common mode voltage on the differential measurement.
  - Channel separation: Influence of a voltage on the neighbor channels not subject to an input voltage.
  - AD Converter Values with inputs shorted: Values on the internal AD converter corresponding to zero input voltage
  - Input Offset Measurement. Output voltage and statistical results over a large number of zero voltage measurements.
  - Input Offset Current: Typical value for information; Maximum channel input offset current, not considering the input resistance.
  - Input resistance: Typical value for information: DAE input resistance at the connector, during internal auto-zeroing and during measurement.
  - Low Battery Alarm Voltage: Typical value for information. Below this voltage, a battery alarm signal is generated.
  - Power consumption: Typical value for information. Supply currents in various operating modes.

# DC Voltage Measurement A/D - Converter Resolution nominal

High Range:	1LSB =	6.1µV,	full range =	-100+300 mV
Low Range:	1LSB =	61nV,	full range =	-1+3mV
DASY measurement	parameters: Aut	to Zero Time: 3	sec; Measuring	time: 3 sec

ŝą.

1.4

<b>Calibration Factors</b>	x	Y <sup>1</sup>	z
High Range	405.510 ± 0.02% (k=2)	405.047 ± 0.02% (k=2)	404.821 ± 0.02% (k=2)
Low Range	3.96328 ± 1.50% (k=2)	3.99400 ± 1.50% (k=2)	3.97320 ± 1.50% (k=2)

# **Connector Angle**

Connector Angle to be used in DASY system	222.5 ° ± 1 °

# Appendix (Additional assessments outside the scope of SCS0108)

# 1. DC Voltage Linearity

High Range	Reading (µV)	Difference (µV)	Error (%)
Channel X + Input	200031.77	-5.20	-0.00
Channel X + Input	20006.58	0.39	0.00
Channel X - Input	-20002.34	3.46	-0.02
Channel Y + Input	200032.86	-4.26	-0.00
Channel Y + Input	20001.39	-4.67	-0.02
Channel Y - Input	-20005.28	0.77	-0.00
Channel Z + Input	200032.31	-5.12	-0.00
Channel Z + Input	20004.31	-1.66	-0.01
Channel Z - Input	-20004.31	1.82	-0.01

Low Range	Reading (µV)	Difference (µV)	Error (%)
Channel X + Input	2001.11	-0.37	-0.02
Channel X + Input	201.74	0.40	0.20
Channel X - Input	-197.72	0.81	-0.41
Channel Y + Input	2001.85	0.48	0.02
Channel Y + Input	200.73	-0.57	-0.28
Channel Y - Input	-200.26	-1.56	0.79
Channel Z + Input	2001.67	0.41	0.02
Channel Z + Input	201.03	-0.17	-0.09
Channel Z - Input	-199.06	-0.31	0.15

2. Common mode sensitivity DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec

	Common mode Input Voltage (mV)	High Range Average Reading (μV)	Low Range Average Reading (µV)
Channel X	200	13.39	11.44
	- 200	-10.26	-12.53
Channel Y	200	-13.63	-13.74
	- 200	12.59	12.05
Channel Z	200	-18.60	-18.48
	- 200	17.68	17.19

# 3. Channel separation

DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec

	Input Voltage (mV)	Channel X (µV)	Channel Y (µV)	Channel Z (µV)
Channel X	200	(46)	3.58	-2.54
Channel Y	200	8.76	3.e.	5.69
Channel Z	200	9.62	6.67	

Certificate No: DAE4-1311\_Aug21

# 4. AD-Converter Values with inputs shorted

DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec

	High Range (LSB)	Low Range (LSB)
Channel X	15446	16713
Channel Y	16320	15746
Channel Z	16580	17710

# 5. Input Offset Measurement

DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec Input 10M $\Omega$ 

	Average (μV)	min. Offset (μV)	max. Offset (μV)	Std. Deviation (µV)
Channel X	0.50	-0.98	1.81	0.67
Channel Y	-0.01	-1.13	1.26	0.57
Channel Z	0.08	-1.25	1.61	0.57

# 6. Input Offset Current

Nominal Input circuitry offset current on all channels: <25fA

# 7. Input Resistance (Typical values for information)

	Zeroing (kOhm)	Measuring (MOhm)
Channel X	200	200
Channel Y	200	200
Channel Z	200	200

# 8. Low Battery Alarm Voltage (Typical values for information)

Typical values	Alarm Level (VDC)	
Supply (+ Vcc)	+7.9	
Supply (- Vcc)	-7.6	

# 9. Power Consumption (Typical values for information)

Typical values	Switched off (mA)	Stand by (mA)	Transmitting (mA)
Supply (+ Vcc)	+0.01	+6	+14
Supply (- Vcc)	-0.01	-8	-9

Calibration Laboratory of Schmid & Partner Engineering AG

Zeughausstrasse 43, 8004 Zurich, Switzerland

Iac MRA

S Schweizerischer Kalibrierdienst

C Service suisse d'étalonnage Servizio svizzero di taratura

S Swiss Calibration Service

Accreditation No.: SCS 0108

Accredited by the Swiss Accreditation Service (SAS) The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

Client

Auden

Certificate No

EX-7628\_Jun22

# CALIBRATION CERTIFICATE

Object	EX3DV4 - SN:7628
Calibration procedure(s)	QA CAL-01.v9, QA CAL-12.v9, QA CAL-14.v6, QA CAL-23.v5, QA CAL-25.v7 Calibration procedure for dosimetric E-field probes
Calibration date	June 22, 2022
This calibration certificate do The measurements and the u	cuments the traceability to national standards, which realize the physical units of measurements (SI). incertainties with confidence probability are given on the following pages and are part of the certificate.
All calibrations have been cor	nducted in the closed laboratory facility: environment temperature (22 $\pm$ 3) °C and humidity < 70%.
Calibration Equipment used (	M&TE critical for calibration)

Primary Standards	ID	Cal Date (Certificate No.)	Scheduled Calibration
Power meter NRP	SN: 104778	04-Apr-22 (No. 217-03525/03524)	Apr-23
Power sensor NRP-Z91	SN: 103244	04-Apr-22 (No. 217-03524)	Apr-23
OCP DAK-3.5 (weighted)	SN: 1249	20-Oct-21 (OCP-DAK3.5-1249 Oct21)	Oct-22
OCP DAK-12	SN: 1016	20-Oct-21 (OCP-DAK12-1016 Oct21)	Oct-22
Reference 20 dB Attenuator	SN: CC2552 (20x)	04-Apr-22 (No. 217-03527)	Apr-23
DAE4	SN: 660	13-Oct-21 (No. DAE4-660_Oct21)	Oct-22
Reference Probe ES3DV2	SN: 3013	27-Dec-21 (No. ES3-3013 Dec21)	Dec-22
	and the second se		and the second se

ID	Check Date (in house)	Scheduled Check
SN: GB41293874	06-Apr-16 (in house check Jun-20)	In house check: Jun-22
SN: MY41498087	06-Apr-16 (in house check Jun-20)	In house check: Jun-22
SN: 000110210	06-Apr-16 (in house check Jun-20)	In house check: Jun-22
SN: US3642U01700		In house check: Jun-22
SN: US41080477		In house check: Oct-22
	SN: GB41293874 SN: MY41498087 SN: 000110210 SN: US3642U01700	SN: GB41293874         06-Apr-16 (in house check Jun-20)           SN: MY41498087         06-Apr-16 (in house check Jun-20)           SN: 000110210         06-Apr-16 (in house check Jun-20)           SN: US3642U01700         04-Aug-99 (in house check Jun-20)

	Name	Function	Signature
Calibrated by	Jeton Kastrati	Laboratory Technician	fell
Approved by	Sven Kühn	Technical Manager	Sh
This calibration certifica	te shall not be reproduced except in	n full without written approval of the lab	Issued: June 28, 2022 oratory.

Calibration Laboratory of Schmid & Partner Engineering AG





Schweizerischer Kalibrierdienst S

Service suisse d'étalonnage С Servizio svizzero di taratura

S Swiss Calibration Service

Accreditation No.: SCS 0108

Accredited by the Swiss Accreditation Service (SAS) The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

#### Glossary

TSL	tissue simulating liquid
NORMx,y,z	sensitivity in free space
ConvF	sensitivity in TSL / NORMx,y,z
DCP	diode compression point
CF	crest factor (1/duty_cycle) of the RF signal
A, B, C, D	modulation dependent linearization parameters
Polarization $\varphi$	$\varphi$ rotation around probe axis
Polarization $\vartheta$	$\vartheta$ rotation around an axis that is in the plane normal to probe axis (at measurement center), i.e., $\vartheta = 0$ is normal to probe axis
Connector Angle	information used in DASY system to align probe sensor X to the robot coordinate system

# Calibration is Performed According to the Following Standards:

- a) IEC/IEEE 62209-1528, "Measurement Procedure For The Assessment Of Specific Absorption Rate Of Human Exposure To Radio Frequency Fields From Hand-Held And Body-Worn Wireless Communication Devices - Part 1528: Human Models, Instrumentation And Procedures (Frequency Range of 4 MHz to 10 GHz)", October 2020.
- b) KDB 865664, "SAR Measurement Requirements for 100 MHz to 6 GHz"

#### Methods Applied and Interpretation of Parameters:

- NORMx, y,z: Assessed for E-field polarization θ = 0 (f ≤ 900 MHz in TEM-cell; f > 1800 MHz; R22 waveguide). NORMx, y,z are only intermediate values, i.e., the uncertainties of NORMx,y,z does not affect the E2-field uncertainty inside TSL (see below ConvF).
- NORM(f)x,y,z = NORMx,y,z \* frequency\_response (see Frequency Response Chart). This linearization is implemented in DASY4 software versions later than 4.2. The uncertainty of the frequency response is included in the stated uncertainty of ConvF.
- · DCPx, y, z: DCP are numerical linearization parameters assessed based on the data of power sweep with CW signal. DCP does not depend on frequency nor media.
- · PAR: PAR is the Peak to Average Ratio that is not calibrated but determined based on the signal characteristics
- · Ax,y,z; Bx,y,z; Cx,y,z; Dx,y,z; VRx,y,z; A, B, C, D are numerical linearization parameters assessed based on the data of power sweep for specific modulation signal. The parameters do not depend on frequency nor media. VR is the maximum calibration range expressed in RMS voltage across the diode.
- · ConvF and Boundary Effect Parameters: Assessed in flat phantom using E-field (or Temperature Transfer Standard for f ≤ 800 MHz) and inside waveguide using analytical field distributions based on power measurements for f > 800 MHz. The same setups are used for assessment of the parameters applied for boundary compensation (alpha, depth) of which typical uncertainty values are given. These parameters are used in DASY4 software to improve probe accuracy close to the boundary. The sensitivity in TSL corresponds to NORMx, y,z \* ConvF whereby the uncertainty corresponds to that given for ConvF. A frequency dependent ConvF is used in DASY version 4.4 and higher which allows extending the validity from ±50 MHz to ±100 MHz.
- · Spherical isotropy (3D deviation from isotropy): in a field of low gradients realized using a flat phantom exposed by a patch antenna.
- · Sensor Offset: The sensor offset corresponds to the offset of virtual measurement center from the probe tip (on probe axis). No tolerance required.
- · Connector Angle: The angle is assessed using the information gained by determining the NORMx (no uncertainty required).

#### **Basic Calibration Parameters**

	Sensor X	Sensor Y	Sensor Z	Unc $(k = 2)$
Norm (µV/(V/m) <sup>2</sup> ) A	0.62	0.57	0.62	±10.1%
DCP (mV) B	111.0	106.0	106.0	±4.7%

#### **Calibration Results for Modulation Response**

UID	Communication System Name	Î	A dB	$^{B}_{dB\sqrt{\mu V}}$	с	D dB	VR mV	Max dev.	Max Unc <sup>E</sup> k = 2
0	CW	X	0.00	0.00	1.00	0.00	149.7	±2.5%	±4.7%
		Y	0.00	0.00	1.00		132.0		
		Z	0.00	0.00	1.00	-	143.0		
10352	Pulse Waveform (200Hz, 10%)	X	1.62	60.95	6.32	10.00	60.0	±3.2%	±9.6%
	1	Y	1.58	60.90	6.47		60.0	i i contra contra de la contra de	0.051950.051
		Z	1.49	60.28	5.91	-	60.0		_
10353	Pulse Waveform (200Hz, 20%)	X	0.86	60.00	4.84	6.99	80.0	±2.6%	±9.6%
	1	Y	0.85	60.00	4.91	-201332	80.0		~~~~~~~~~~~~~~~~~~~~~~~~~~~~~~~~~~~~~~~
		Z	0.82	60.00	4.56		80.0		_
10354	Pulse Waveform (200Hz, 40%)	X	0.51	60.00	3.65	3.98	95.0	±2.5%	±9.6%
0.00.000000		Y	24.00	72.00	7.00	1.00000	95.0	5977424-255 -	-30-9637N
		Z	0.07	139.09	0.19		95.0		
10355	Pulse Waveform (200Hz, 60%)	X	0.60	154.87	21.18	2.22	120.0	±1.5%	±9.6%
052220		Y	5.90	159.97	7.34		120.0		
		Z	3.18	159.30	17.99		120.0		
10387	QPSK Waveform, 1 MHz	X	0.53	66.90	14.75		150.0	±3.7%	±9.6%
	BOTES POTSTATION IN AND ITS	Y	0.40	62.66	11.45		150.0		0000000
		Z	0.41	62.09	11.37		150.0		
10388	QPSK Waveform, 10 MHz	X	1.42	68.83	15.02	0.00	150.0	±0.9%	±9.6%
		Y	1.17	65.68	13.20	1 824.6	150.0		
		Z	1.17	65.30	13.17		150.0		
10396	64-QAM Waveform, 100 kHz	X	1.85	66.44	16.84	3.01	150.0	±1.1%	±9.6%
	2 - 2 AN - ANA	Y	1.79	66.01	16.59		150.0	1	
		Z	1.63	64.49	16.12		150.0		
10399	64-QAM Waveform, 40 MHz	X	2.78	67.22	15.55	0.00	150.0	±2.7%	±9.6%
		Y	2.66	66.30	14.94	-	150.0		
		Z	2.78	66.75	15.21		150.0		
10414	WLAN CCDF, 64-QAM, 40 MHz	X	3.76	67.22	15.72	0.00	150.0	±4.3%	±9.6%
		Y	3.71	66.68	15.38		150.0		
		Z	3.70	66.38	15.30	2	150.0		

Note: For details on UID parameters see Appendix

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

<sup>A</sup> The uncertainties of Norm X,Y,Z do not affect the E<sup>2</sup>-field uncertainty inside TSL (see Pages 5 and 6). <sup>B</sup> Linearization parameter uncertainty for maximum specified field strength.

E Uncertainty is determined using the max, deviation from linear response applying rectangular distribution and is expressed for the square of the field value.

# Sensor Model Parameters

	C1 fF	C2 fF	а V <sup>-1</sup>	T1 msV <sup>-2</sup>	T2 ms V <sup>-1</sup>	T3 ms	T4 V <sup>-2</sup>	T5 V <sup>-1</sup>	Т6
x	7.4	51.85	31.86	5.33	0.00	4.90	0.73	0.00	1.00
y	7.8	55.92	32.57	4.93	0.00	4.94	0.68	0.00	1.00
z	8.1	58.19	33.16	2.75	0.00	4.90	0.34	0.00	1.00

# **Other Probe Parameters**

Sensor Arrangement	Triangular
Connector Angle	-144.3°
Mechanical Surface Detection Mode	enabled
Optical Surface Detection Mode	disabled
Probe Overall Length	337 mm
Probe Body Diameter	10 mm
Tip Length	9 mm
Tip Diameter	2.5 mm
Probe Tip to Sensor X Calibration Point	1 mm
Probe Tip to Sensor Y Calibration Point	1 mm
Probe Tip to Sensor Z Calibration Point	1 mm
Recommended Measurement Distance from Surface	1.4 mm

Note: Measurement distance from surface can be increased to 3-4 mm for an Area Scan job.

#### Calibration Parameter Determined in Head Tissue Simulating Media

f (MHz) <sup>C</sup>	Relative Permittivity <sup>F</sup>	Conductivity <sup>F</sup> (S/m)	ConvF X	ConvF Y	ConvF Z	Alpha <sup>G</sup>	Depth <sup>G</sup> (mm)	Unc (k = 2)
750	41.9	0.89	10.22	10.22	10.22	0.52	0.80	±12.0%
835	41.5	0.90	9.96	9.96	9.96	0.45	0.85	±12.0%
900	41.5	0.97	9.68	9.68	9.68	0.46	0.80	±12.0%
1450	40.5	1.20	8.79	8.79	8.79	0.50	0.80	±12.0%
1640	40.2	1.31	8.67	8.67	8.67	0.32	0.86	±12.0%
1750	40.1	1.37	8.62	8.62	8.62	0.34	0.86	±12.0%
1900	40.0	1.40	8.23	8.23	8.23	0.36	0.86	±12.0%
2000	40.0	1.40	8.19	8.19	8.19	0.30	0.86	±12.0%
2300	39.5	1.67	8.12	8.12	8.12	0.28	0.90	±12.0%
2450	39.2	1.80	7.82	7.82	7.82	0.35	0.90	±12.0%
2600	39.0	1.96	7.74	7.74	7.74	0.27	0.90	±12.0%
3300	38.2	2.71	7.21	7.21	7.21	0.30	1.35	±13.1%
3500	37.9	2.91	7.00	7.00	7.00	0.30	1.35	±13.1%
3700	37.7	3.12	6.95	6.95	6.95	0.30	1.35	±13.1%
3900	37.5	3.32	6.74	6.74	6.74	0.35	1.50	±13.1%
4100	37.2	3.53	6.64	6.64	6.64	0.35	1.50	±13.19
4200	37.1	3.63	6.38	6.38	6.38	0.35	1.60	±13.19
4400	36.9	3.84	6.32	6.32	6.32	0.35	1.60	±13.1%
4600	36.7	4.04	6.04	6.04	6.04	0.35	1.70	±13.1%
4800	36.4	4.25	6.08	6.08	6.08	0.40	1.80	±13.1%
4950	36.3	4.40	5.87	5.87	5.87	0.40	1.80	±13.19
5250	35.9	4.71	5.49	5.49	5.49	0.40	1.80	±13.1%
5600	35.5	5.07	4.93	4.93	4.93	0.40	1.80	±13.1%
5750	35.4	5.22	4.90	4.90	4.90	0.40	1.80	±13.1%
5850	35.2	5.32	4.78	4.78	4.78	0.40	1.80	±13.1%

<sup>C</sup> Frequency validity above 300 MHz of ±100 MHz only applies for DASY v4.4 and higher (see Page 2), else it is restricted to ±50 MHz. The uncertainty is the RSS of the ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band. Frequency validity below 300 MHz is ±10, 25, 40, 50 and 70 MHz for ConvF assessments at 30, 64, 128, 150 and 220 MHz respectively. Validity of ConvF assessed at 6 MHz is 4–9 MHz, and ConvF assessed at 13 MHz is 9–19 MHz. Above 5 GHz frequency validity can be extended to ±110 MHz.

assessed at 13 MHz is 9–19 MHz. Above 5 GHz frequency validity can be extended to ±110 MHz. <sup>F</sup> At frequencies below 3 GHz, the validity of tissue parameters (ε and σ) can be relaxed to ±10% if liquid compensation formula is applied to measured SAR values. At frequencies above 3 GHz, the validity of tissue parameters (ε and σ) is restricted to ±5%. The uncertainty is the RSS of the ConvF uncertainty for indicated target tissue parameters.

<sup>G</sup> Alpha/Depth are determined during calibration. SPEAG warrants that the remaining deviation due to the boundary effect after compensation is always less than ±1% for frequencies below 3 GHz and below ±2% for frequencies between 3–6 GHz at any distance larger than half the probe tip diameter from the boundary.


## Calibration Parameter Determined in Head Tissue Simulating Media

f (MHz) <sup>C</sup>	Relative Permittivity <sup>F</sup>	Conductivity <sup>F</sup> (S/m)	ConvF X	ConvF Y	ConvF Z	Alpha <sup>G</sup>	Depth <sup>G</sup> (mm)	Unc (k = 2)
6500	34.5	6.07	5.45	5.45	5.45	0.20	2.50	±18.6%
7000	33.9	6.65	5.55	5.55	5.55	0,25	2.50	±18.6%
8000	32.7	7.84	5.35	5.35	5.35	0.50	1.50	±18.6%
9000	31.6	9.08	5.40	5.40	5.40	0.50	1.80	±18.6%

C Frequency validity at 6.5 GHz is -600/+700 MHz, and ±700 MHz at or above 7 GHz. The uncertainty is the RSS of the ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band. <sup>F</sup> At frequencies 6–10 GHz, the validity of tissue parameters ( $\varepsilon$  and  $\sigma$ ) can be relaxed to ±10% if liquid compensation formula is applied to measured SAR

values. The uncertainty is the RSS of the ConvF uncertainty for indicated target tissue parameters.


G Alpha/Depth are determined during calibration. SPEAG warrants that the remaining deviation due to the boundary effect after compensation is always less than ±1% for frequencies below 3 GHz; below ±2% for frequencies between 3-6 GHz; and below ±4% for frequencies between 6-10 GHz at any distance larger than half the probe tip diameter from the boundary.




# Frequency Response of E-Field

(TEM-Cell:ifi110 EXX, Waveguide:R22)

Uncertainty of Frequency Response of E-field: ±6.3% (k=2)




# Receiving Pattern ( $\phi$ ), $\vartheta = 0^{\circ}$



Dynamic Range f(SAR<sub>head</sub>)

Uncertainty of Linearity Assessment: ±0.6% (k=2)



# **Conversion Factor Assessment**

0 Uncertainty of Spherical Isotropy Assessment: ±2.6% (k=2)

0.2

0.4

0.6

0.8

Certificate No: EX-7628\_Jun22

-0.8

-1

-0.6 -0.4 -0.2

# Appendix: Modulation Calibration Parameters

UID	Rev	Communication System Name	Group	PAR (dB)	Unc <sup>E</sup> k =
0	-	CW	CW	0.00	±4.7
0010	CAA	SAR Validation (Square, 100 ms, 10 ms)	Test	10.00	±9.6
10011	CAB	UMTS-FDD (WCDMA)	WCDMA	2.91	±9.6
10012	CAB	IEEE 802.11b WiFi 2.4 GHz (DSSS, 1 Mbps)	WLAN	1.87	±9.6
10013	CAB	IEEE 802.11g WiFi 2.4 GHz (DSSS-OFDM, 6 Mbps)	WLAN	9.46	±9.6
10021	DAC	GSM-FDD (TDMA, GMSK)	GSM	9.39	±9.6
10023	DAC	GPRS-FDD (TDMA, GMSK, TN 0)	GSM	9.57	±9.6
10024	DAC	GPRS-FDD (TDMA, GMSK, TN 0-1)	GSM	6.56	±9.6
10025	DAC	EDGE-FDD (TDMA, 8PSK, TN 0)	GSM	12.62	±9.6
10026	DAC	EDGE-FDD (TDMA, 8PSK, TN 0-1)	GSM	9.55	±9.6
10027	DAC	GPRS-FDD (TDMA, GMSK, TN 0-1-2)	GSM	4.80	±9.6
10028	DAC	GPRS-FDD (TDMA, GMSK, TN 0-1-2-3)	GSM	3.55	±9.6
10029	DAC	EDGE-FDD (TDMA, 8PSK, TN 0-1-2)	GSM	7.78	±9.6
10030	CAA	IEEE 802.15.1 Bluetooth (GFSK, DH1)	Bluetooth	5.30	±9.6
10031	CAA	IEEE 802.15.1 Bluetooth (GFSK, DH3)	Bluetooth	1.87	±9.6
10032	CAA	IEEE 802.15.1 Bluetooth (GFSK, DH5)	Bluetooth	1,16	±9.6
10033	CAA	IEEE 802.15.1 Bluetooth (PI/4-DQPSK, DH1)	Bluetooth	7.74	±9.6
10034	CAA	IEEE 802.15.1 Bluetooth (PI/4-DQPSK, DH3)	Bluetooth	4.53	±9.6
10035	CAA	IEEE 802.15.1 Bluetooth (PI/4-DQPSK, DH5)	Bluetooth	3.83	±9.6 ±9.6
10036	CAA	IEEE 802.15.1 Bluetooth (8-DPSK, DH1)			and the state of t
10037	CAA	IEEE 802.15.1 Bluetooth (8-DPSK, DH3)	Bluetooth	8.01	±9.6
10038	CAA	IEEE 602.15.1 Bluetooth (8-DPSK, DH3)	Bluetooth	4.77	±9.6
10039	CAB		Bluetooth	4.10	±9.6
10039	CAB	CDMA2000 (1xRTT, RC1)	CDMA2000	4.57	±9.6
10042		IS-54 / IS-136 FDD (TDMA/FDM, PI/4-DQPSK, Halfrate)	AMPS	7.78	±9.6
10044	CAA	IS-91/EIA/TIA-553 FDD (FDMA, FM)	AMPS	0.00	±9.6
1.	CAA	DECT (TDD, TDMA/FDM, GFSK, Full Slot, 24)	DECT	13.80	±9.6
0049	CAA	DECT (TDD, TDMA/FDM, GFSK, Double Slot, 12)	DECT	10.79	±9.6
10056	CAA	UMTS-TDD (TD-SCDMA, 1.28 Mcps)	TD-SCDMA	11.01	±9.6
10058	DAC	EDGE-FDD (TDMA, 8PSK, TN 0-1-2-3)	GSM	6.52	±9.6
10059	CAB	IEEE 802.11b WiFi 2.4 GHz (DSSS, 2 Mbps)	WLAN	2.12	±9.6
10060	CAB	IEEE 802,11b WiFi 2.4 GHz (DSSS; 5.5 Mbps)	WLAN	2.83	±9.6
10061	CAB	IEEE 802.11b WiFi 2.4 GHz (DSSS, 11 Mbps)	WLAN	3.60	±9.6
10062	CAD	IEEE 802.11a/h WiFi 5 GHz (OFDM, 6 Mbps)	WLAN	8.68	±9.6
10063	CAD	IEEE 802.11a/h WiFi 5 GHz (OFDM, 9 Mbps)	WLAN	8.63	±9.6
10064	CAD	IEEE 802.11a/h WiFi 5 GHz (OFDM, 12 Mbps)	WLAN	9.09	±9.6
10065	CAD	IEEE 802.11a/h WiFi 5 GHz (OFDM, 18 Mbps)	WLAN	9.00	±9.6
10066	CAD	IEEE 802.11a/h WiFi 5 GHz (OFDM, 24 Mbps)	WLAN	9.38	±9.6
0067	CAD	IEEE 802.11a/h WiFi 5 GHz (OFDM, 36 Mbps)	WLAN	10.12	±9.6
10068	CAD	IEEE 802.11a/h WiFi 5 GHz (OFDM, 48 Mbps)	WLAN	10.24	±9.6
10069	CAD	IEEE 802.11a/h WiFi 5 GHz (OFDM, 54 Mbps)	WLAN	10.56	±9.6
10071	CAB	IEEE 802.11g WiFi 2.4 GHz (DSSS/OFDM, 9 Mbps)	WLAN	9.83	±9.6
10072	CAB	IEEE 802.11g WiFi 2.4 GHz (DSSS/OFDM, 12 Mbps)	WLAN	9.62	±9.6
10073	CAB	IEEE 802.11g WiFi 2.4 GHz (DSSS/OFDM, 18 Mbps)	WLAN	9.94	±9.6
10074	CAB	IEEE 802.11g WiFi 2.4 GHz (DSSS/OFDM, 24 Mbps)	WLAN	10.30	±9.6
10075	CAB	IEEE 802.11g WiFi 2.4 GHz (DSSS/OFDM, 36 Mbps)	WLAN	10.30	±9.6
10076	CAB	IEEE 802.11g WiFi 2.4 GHz (DSSS/OFDM, 48 Mbps)	WLAN	10.77	±9.6
0077	CAB	IEEE 802.11g WiFi 2.4 GHz (DSSS/OFDM, 48 Mbps)	WLAN		
0081	CAB	CDMA2000 (1xRTT, RC3)		11.00	±9.6
0082	CAB	IS-54 / IS-136 FDD (TDMA/FDM, Pl/4-DQPSK, Fullrate)	CDMA2000	3.97	±9.6
0090	DAC	GPRS-FDD (TDMA, GMSK, TN 0-4)	AMPS	4.77	±9.6
0090	CAC	UMTS-FDD (HSDPA)	GSM	6.56	±9.6
0097	DAC		WCDMA	3.98	±9.6
0098	CAC	UMTS-FDD (HSUPA, Subtest 2)	WCDMA	3.98	±9.6
		EDGE-FDD (TDMA, 8PSK, TN 0-4)	GSM	9.55	±9.6
0100	CAC	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	5.67	±9.6
0101	CAB	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, 16-QAM)	LTE-FDD	6.42	±9.6
0102	CAB	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, 64-QAM)	LTE-FDD	6.60	±9.6
0103	DAC	LTE-TDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-TOD	9.29	±9.6
0104	CAE	LTE-TDD (SC-FDMA, 100% RB, 20 MHz, 16-QAM)	LTE-TOD	9.97	±9.6
0105	CAE	LTE-TDD (SC-FDMA, 100% RB, 20 MHz, 64-QAM)	LTE-TDD	10.01	±9.6
0108	CAE	LTE-FDD (SC-FDMA, 100% RB, 10 MHz, QPSK)	LTE-FDD	5.80	±9.6
0109	CAG	LTE-FDD (SC-FDMA, 100% RB, 10 MHz, 16-QAM)	LTE-FDD	6.43	±9.6
0110	CAG	LTE-FDD (SC-FDMA, 100% RB, 5 MHz, QPSK)	LTE-FDD	5.75	±9.6
0111	CAG	LTE-FDD (SC-FDMA, 100% RB, 5MHz, 16-QAM)	LTE-FDD	6.44	±9.6

UID	Rev	Communication System Name	Group	PAR (dB)	Unc <sup>E</sup> k =
10112	CAG	LTE-FDD (SC-FDMA, 100% RB, 10 MHz, 64-QAM)	LTE-FDD	6.59	±9.6
10113	CAG	LTE-FDD (SC-FDMA, 100% RB, 5 MHz, 64-QAM)	LTE-FDD	6.62	±9.6
0114	CAG	IEEE 802.11n (HT Greenfield, 13.5 Mbps, BPSK)	WLAN	8.10	±9.6
0115	CAG	IEEE 802.11n (HT Greenfield, 81 Mbps, 16-QAM)	WLAN	8.46	±9.6
0116	CAG	IEEE 802.11n (HT Greenfield, 135 Mbps, 64-QAM)	WLAN	8.15	±9.6
0117	CAG	IEEE 802.11n (HT Mixed, 13.5 Mbps, BPSK)	WLAN	8.07	±9.6
0118	CAD	IEEE 802.11n (HT Mixed, 81 Mbps, 16-QAM)	WLAN	8.59	±9.6
0119	CAD	IEEE 802.11n (HT Mixed, 135 Mbps, 64-QAM)	WLAN	8.13	±9.6
0140	CAD	LTE-FDD (SC-FDMA, 100% RB, 15 MHz, 16-QAM)	LTE-FDD	6.49	±9.6
0141	CAD	LTE-FDD (SC-FDMA, 100% RB, 15 MHz, 64-QAM)	LTE-FDD	6.53	±9.6
0142	CAD	LTE-FDD (SC-FDMA, 100% RB, 3 MHz, QPSK)	LTE-FDD	5.73	±9.6
0143	CAD	LTE-FDD (SC-FDMA, 100% RB, 3 MHz, 16-QAM)	LTE-FDD	6.35	±9.6
0144	CAC	LTE-FDD (SC-FDMA, 100% RB, 3 MHz, 64-QAM)	LTE-FDD	6.65	±9.6
0145	CAC	LTE-FDD (SC-FDMA, 100% RB, 1.4 MHz, QPSK)	LTE-FDD	5.76	±9.6
0146	CAC	LTE-FDD (SC-FDMA, 100% RB, 1.4 MHz, 16-QAM)	LTE-FDD	6.41	±9.6
0147	CAC	LTE-FDD (SC-FDMA, 100% RB, 1.4 MHz, 64-QAM)	LTE-FDD	6.72	±9.6
0149	CAE	LTE-FDD (SC-FDMA, 50% RB, 20 MHz, 16-QAM)	LTE-FDD	6.42	±9.6
0150	CAE	LTE-FDD (SC-FDMA, 50% RB, 20 MHz, 64-QAM)	LTE-FDD	6.60	±9.6
0151	CAE	LTE-TDD (SC-FDMA, 50% RB, 20 MHz, QPSK)	LTE-TDD	9.28	±9.6
0152	CAE	LTE-TDD (SC-FDMA, 50% RB, 20 MHz, 16-QAM)	LTE-TDD	9.92	±9.6
0153	CAE	LTE-TDD (SC-FDMA, 50% RB, 20 MHz, 64-QAM)	LTE-TDD	10.05	±9.6
0154	CAF	LTE-FDD (SC-FDMA, 50% RB, 10 MHz, QPSK)	LTE-FDD	5.75	±9.6
0155	CAF	LTE-FDD (SC-FDMA, 50% RB, 10 MHz, 16-QAM)	LTE-FDD	6.43	±9.6
0156	CAF	LTE-FDD (SC-FDMA, 50% RB, 5MHz, QPSK)	LTE-FDD	5.79	±9.6
0157	CAE	LTE-FDD (SC-FDMA, 50% RB, 5 MHz, 16-QAM)	LTE-FDD	6.49	±9.6
0158	CAE	LTE-FDD (SC-FDMA, 50% RB, 10 MHz, 64-QAM)	LTE-FDD	6.62	±9.6
0159	CAG	LTE-FDD (SC-FDMA, 50% RB, 5 MHz, 64-QAM)	LTE-FDD	6.56	±9.6
0160	CAG	LTE-FDD (SC-FDMA, 50% RB, 15 MHz, QPSK)	LTE-FDD	5.82	±9,6
0161	CAG	LTE-FDD (SC-FDMA, 50% RB, 15 MHz, 16-QAM)	LTE-FDD	6.43	±9.6
0162	CAG	LTE-FDD (SC-FDMA, 50% RB, 15 MHz, 64-QAM)	LTE-FDD	6.58	±9.6
0166	CAG	LTE-FDD (SC-FDMA, 50% RB, 1.4 MHz, QPSK)	LTE-FDD	5.46	±9.6
0167	CAG	LTE-FDD (SC-FDMA, 50% RB, 1.4 MHz, 16-QAM)	LTE-FDD	6.21	±9.6
0168	CAG	LTE-FDD (SC-FDMA, 50% RB, 1.4 MHz, 64-QAM)	LTE-FDD	6.79	±9.6
0169	CAG	LTE-FDD (SC-FDMA, 1 RB, 20 MHz, QPSK)	LTE-FDD	5.73	±9.6
0170	CAG	LTE-FDD (SC-FDMA, 1 RB, 20 MHz, 16-QAM)	LTE-FDD	6.52	±9.6
0171	CAE	LTE-FDD (SC-FDMA, 1 RB, 20 MHz, 64-QAM)	LTE-FDD	6.49	±9.6
0172	CAE	LTE-TDD (SC-FDMA, 1 RB, 20 MHz, QPSK)	LTE-TDD	9.21	±9.6
0173	CAE	LTE-TDD (SC-FDMA, 1 RB, 20 MHz, 16-QAM)	LTE-TDD	9.48	±9.6
0174	CAF	LTE-TDD (SC-FDMA, 1 RB, 20 MHz, 64-QAM)	LTE-TDD	10.25	±9.6
0175	CAF	LTE-FDD (SC-FDMA, 1 RB, 10 MHz, QPSK)	LTE-FDD	5.72	±9.6
0176	CAF	LTE-FDD (SC-FDMA, 1 RB, 10 MHz, 16-QAM)	LTE-FDD	6.52	±9.6
0177	CAE	LTE-FDD (SC-FDMA, 1 RB, 5 MHz, QPSK)	LTE-FDD	5.73	±9.6
0178	CAE	LTE-FDD (SC-FDMA, 1 RB, 5 MHz, 16-QAM)	LTE-FDD	6.52	±9.6
0179	AAE	LTE-FDD (SC-FDMA, 1 RB, 10 MHz, 64-QAM)	LTE-FDD	6.50	±9.6
0180	CAG	LTE-FDD (SC-FDMA, 1 RB, 5 MHz, 64-QAM)	LTE-FDD	6.50	±9.6
0181	CAG	LTE-FDD (SC-FDMA, 1 RB, 15 MHz, QPSK)	LTE-FDD	5.72	±9.6
0182	CAG	LTE-FDD (SC-FDMA, 1 RB, 15 MHz, 16-QAM)	LTE-FDD	6.52	±9.6
0183	CAG	LTE-FDD (SC-FDMA, 1 RB, 15 MHz, 64-QAM)	LTE-FDD	6.50	±9.6
0184	CAG	LTE-FDD (SC-FDMA, 1 RB, 3 MHz, QPSK)	LTE-FDD	5.73	±9.6
0185	CAI	LTE-FDD (SC-FDMA, 1 RB, 3 MHz, 16-QAM)	LTE-FDD	6.51	±9.6
0186	CAG	LTE-FDD (SC-FDMA, 1 RB, 3 MHz, 64-QAM)	LTE-FDD	6.50	±9.6
0187	CAG	LTE-FDD (SC-FDMA, 1 RB, 1.4 MHz, QPSK)	LTE-FDD	5.73	±9.6
0188	CAG	LTE-FDD (SC-FDMA, 1 RB, 1.4 MHz, 16-QAM)	LTE-FDD	6.52	±9.6
0189	CAE	LTE-FDD (SC-FDMA, 1 RB, 1.4 MHz, 64-QAM)	LTE-FDD	6.50	±9.6
0193	CAE	IEEE 802.11n (HT Greentield, 6.5 Mbps, BPSK)	WLAN	8,09	±9.6
0194	AAD	IEEE 802.11n (HT Greenfield, 39 Mbps, 16-QAM)	WLAN	8.12	±9.6
0195	CAE	IEEE 802.11n (HT Greenfield, 65 Mbps, 64-QAM)	WLAN	8.21	±9.6
0196	CAE	IEEE 802.11n (HT Mixed, 6.5 Mbps, BPSK)	WLAN	8.10	±9.6
0197	AAE	IEEE 802.11n (HT Mixed, 39 Mbps, 16-QAM)	WLAN	8,13	±9.6
0198	CAF	IEEE 802.11n (HT Mixed, 65 Mbps, 64-QAM)	WLAN	8.27	±9.6
0219	CAF	IEEE 802.11n (HT Mixed, 7.2 Mbps, BPSK)	WLAN	8.03	±9.6
0220	AAF	IEEE 802.11n (HT Mixed, 43.3 Mbps, 16-QAM)	WLAN	8.13	±9.6
0.004	CAC	IEEE 802.11n (HT Mixed, 72.2 Mbps, 64-QAM)	WLAN	8.27	±9.6
	0.1-	THE ODD HAN THE READ HERBARY DOOLD	WLAN	0.00	±9.6
0221	CAC	IEEE 802.11n (HT Mixed, 15 Mbps, BPSK) IEEE 802.11n (HT Mixed, 90 Mbps, 16-QAM)	WLAN	8.06	±9.6

UID	Rev	Communication System Name	Group	PAR (dB)	Unc <sup>E</sup> k =
10225	CAD	UMTS-FDD (HSPA+)	WCDMA	5.97	±9.6
10226	CAD	LTE-TDD (SC-FDMA, 1 RB, 1.4 MHz, 16-QAM)	LTE-TDD	9.49	±9.6
10227	CAD	LTE-TDD (SC-FDMA, 1 RB, 1.4 MHz, 64-QAM)	LTE-TDD	10,26	±9.6
10228	CAD	LTE-TDD (SC-FDMA, 1 RB, 1.4 MHz, QPSK)	LTE-TDD	9.22	±9.6
10229	DAC	LTE-TDD (SC-FDMA, 1 RB, 3 MHz, 16-QAM)	LTE-TDD	9.48	±9.6
10230	CAC	LTE-TDD (SC-FDMA, 1 RB, 3 MHz, 64-QAM)	LTE-TDD	10.25	±9.6
10231	CAC	LTE-TDD (SC-FDMA, 1 RB, 3 MHz, QPSK)	LTE-TDD	9,19	±9.6
10232	CAD	LTE-TDD (SC-FDMA, 1 RB, 5 MHz, 16-QAM)	LTE-TDD	9.48	±9,6
10233	CAD	LTE-TDD (SC-FDMA, 1 RB, 5 MHz, 64-QAM)	LTE-TDD	10.25	±9.6
10234	CAD	LTE-TDD (SC-FDMA, 1 RB, 5 MHz, QPSK)	LTE-TDD	9.21	±9.6
10235	CAD	LTE-TDD (SC-FDMA, 1 RB, 10 MHz, 16-QAM)	LTE-TDD	9,48	±9.6
10236	CAD	LTE-TDD (SC-FDMA, 1 RB, 10 MHz, 64-QAM)	LTE-TDD	10.25	±9.6
10237	CAD	LTE-TDD (SC-FDMA, 1 RB, 10 MHz, QPSK)	LTE-TDD	9.21	±9.6
10238	CAB	LTE-TDD (SC-FDMA, 1 RB, 15 MHz, 16-QAM)	LTE-TDD	9.48	±9.6
10239	CAB	LTE-TDD (SC-FDMA, 1 RB, 15 MHz, 64-QAM)	LTE-TDD	10.25	±9.6
10240	CAB	LTE-TDD (SC-FDMA, 1 RB, 15MHz, QPSK)	LTE-TDD	9.21	±9.6
10241	CAB	LTE-TDD (SC-FDMA, 50% RB, 1.4 MHz, 16-QAM)	LTE-TDD	9.82	±9.6
10242	CAD	LTE-TDD (SC-FDMA, 50% RB, 1.4 MHz, 64-QAM)	LTE-TDD	9.86	±9.6
10243	CAD	LTE-TDD (SC-FDMA, 50% RB, 1.4 MHz, QPSK)	LTE-TDD	9.46	±9.6
10244	CAD	LTE-TDD (SC-FDMA, 50% RB, 3 MHz, 16-QAM)	LTE-TDD	10.06	±9.6
10245	CAG	LTE-TDD (SC-FDMA, 50% RB, 3 MHz, 64-QAM)	LTE-TDD	10.06	±9.6
10246	CAG	LTE-TDD (SC-FDMA, 50% RB, 3 MHz, QPSK)	LTE-TDD	9.30	±9.6
0247	CAG	LTE-TDD (SC-FDMA, 50% RB, 5MHz, 16-QAM)	LTE-TDD	9.91	±9.6
10248	CAG	LTE-TDD (SC-FDMA, 50% RB, 5MHz, 64-QAM)	LTE-TDD	10.09	±9.6
10249	CAG	LTE-TDD (SC-FDMA, 50% RB, 5 MHz, QPSK)	LTE-TDD	9.29	±9.6
10,250	CAG	LTE-TDD (SC-FDMA, 50% RB, 10 MHz, 16-QAM)	LTE-TDD	9.81	±9.6
10251	CAF	LTE-TDD (SC-FDMA, 50% RB, 10 MHz, 64-QAM)	LTE-TDD	10.17	±9.6
10252	CAF	LTE-TDD (SC-FDMA, 50% RB, 10 MHz, QPSK)	LTE-TDD	9.24	±9.6
10253	CAF	LTE-TDD (SC-FDMA, 50% RB, 15 MHz, 16-QAM)	LTE-TDD	9.90	±9.6
0254	CAB	LTE-TDD (SC-FDMA, 50% RB, 15 MHz, 64-QAM)	LTE-TDD	10.14	±9.6
0255	CAB	LTE-TDD (SC-FDMA, 50% RB, 15 MHz, QPSK)	LTE-TDD	9.20	±9.6
0256	CAB	LTE-TDD (SC-FDMA, 100% RB, 1.4 MHz, 16-QAM)	LTE-TDD	9.96	±9.6
0257	CAD	LTE-TDD (SC-FDMA, 100% RB, 1.4 MHz, 64-QAM)	LTE-TDD	10.08	±9.6
0258	CAD	LTE-TDD (SC-FDMA, 100% RB, 1.4 MHz, QPSK)	LTE-TDD	9.34	±9.6
0259	CAD	LTE-TDD (SC-FDMA, 100% RB, 3 MHz, 16-QAM)	LTE-TDD	9.98	±9.6
0260	CAG	LTE-TDD (SC-FDMA, 100% RB, 3 MHz, 64-QAM)	LTE-TDD	9.97	±9.6
0261	CAG	LTE-TDD (SC-FDMA, 100% RB, 3 MHz, QPSK)	LTE-TDD	9.24	±9.6
0262	CAG	LTE-TDD (SC-FDMA, 100% RB, 5 MHz, 16-QAM)	LTE-TDD	9.83	±9.6
0263	CAG	LTE-TDD (SC-FDMA, 100% RB, 5 MHz, 64-QAM)	LTE-TDD	10.16	±9.6
0264	CAG	LTE-TDD (SC-FDMA, 100% RB, 5 MHz, QPSK)	LTE-TDD	9.23	±9.6
0265	CAG	LTE-TDD (SC-FDMA, 100% RB, 10 MHz, 16-QAM)	LTE-TDD	9.92	±9.6
0266	CAF	LTE-TDD (SC-FDMA, 100% RB, 10 MHz, 64-QAM)	LTE-TDD	10.07	±9.6
0267	CAF	LTE-TDD (SC-FDMA, 100% RB, 10 MHz, QPSK)	LTE-TDD	9.30	±9.6
0268	CAF	LTE-TDD (SC-FDMA, 100% RB, 15 MHz, 16-QAM)	LTE-TDD	10.06	±9.6
0269	CAB	LTE-TDD (SC-FDMA, 100% RB, 15 MHz, 64-QAM)	LTE-TDD	10.13	±9.6
0270	CAB	LTE-TDD (SC-FDMA, 100% RB, 15 MHz, QPSK)	LTE-TDD	9.58	±9.6
0274	CAB	UMTS-FDD (HSUPA, Subtest 5, 3GPP Rel8.10)	WCDMA	4.87	±9.6
0275	CAD	UMTS-FDD (HSUPA, Subtest 5, 3GPP Rel8.4)	WCDMA	3.96	±9.6
0277	CAD	PHS (QPSK)	PHS	11.81	±9.6
0278	CAD	PHS (QPSK, BW 884 MHz, Rolloff 0.5)	PHS	11.81	±9.6
0279	CAG	PHS (QPSK, BW 884 MHz, Rolloff 0.38)	PHS	12.18	±9.6
0290	CAG	CDMA2000, RC1, SO55, Full Rate	CDMA2000	3.91	±9.6
0291	CAG	CDMA2000, RC3, SO55, Full Rate	CDMA2000	3.46	±9.6
0292	CAG	CDMA2000, RC3, SO32, Full Rate	CDMA2000	3.39	±9.6
0293	CAG	CDMA2000, RC3, SO3, Full Rate	CDMA2000	3.50	±9.6
0295	CAG	CDMA2000, RC1, SO3, 1/8th Rate 25 fr.	CDMA2000	12.49	±9.6
0297	CAF	LTE-FDD (SC-FDMA, 50% RB, 20 MHz, QPSK)	LTE-FDD	5.81	±9.6
0298	CAF	LTE-FDD (SC-FDMA, 50% RB, 3 MHz, QPSK)	LTE-FDD	5.72	±9.6
0299	CAF	LTE-FDD (SC-FDMA, 50% RB, 3 MHz, 16-QAM)	LTE-FDD	6.39	±9.6
0300	CAC	LTE-FDD (SC-FDMA, 50% RB, 3 MHz, 64-QAM)	LTE-FDD	6.60	±9.6
0301	CAC	IEEE 802.16e WIMAX (29:18, 5 ms, 10 MHz, QPSK, PUSC)	WIMAX	12.03	±9.6
0302	CAB	IEEE 802.16e WiMAX (29:18, 5 ms, 10 MHz, QPSK, PUSC, 3CTRL)	WIMAX	12.57	±9.6
0303	CAB	IEEE 802.16e WiMAX (31:15, 5 ms, 10 MHz, 64QAM, PUSC)	WIMAX	12.52	±9.6
0304	CAA	IEEE 802.16e WIMAX (29:18, 5 ms, 10 MHz, 64QAM, PUSC)	WEMAX	11.86	±9.6
0305	CAA	IEEE 802.16e WIMAX (31:15, 10 ms, 10 MHz, 64QAM, PUSC)	WIMAX	15.24	±9.6
0000			A REAL WORLD	220010000	

UID	Rev	Communication System Name	Group	PAR (dB)	Unc <sup>E</sup> k =
10307	AAB	IEEE 802.16e WIMAX (29:18, 10 ms, 10 MHz, QPSK, PUSC)	WIMAX	14.49	±9.6
10308	AAB	IEEE 802.16e WIMAX (29:18, 10 ms, 10 MHz, 16QAM, PUSC)	WIMAX	14,46	±9.6
10309	AAB	IEEE 802.16e WIMAX (29:18, 10 ms, 10 MHz, 16QAM, AMC 2x3)	WIMAX	14.58	±9.6
10310	AAB	IEEE 802.16e WIMAX (29:18, 10 ms, 10 MHz, QPSK, AMC 2x3	WiMAX	14.57	±9.6
10311	AAB	LTE-FDD (SC-FDMA, 100% RB, 15MHz, QPSK)	LTE-FDD	6.06	±9.6
10313	AAD	IDEN 1:3	IDEN	10.51	±9.6
10314	AAD	IDEN 1:6	IDEN	13.48	±9.6
10315	AAD	IEEE 802.11b WiFi 2.4 GHz (DSSS, 1 Mbps, 96pc dc)	WLAN	1.71	±9.6
10316	AAD	IEEE 802.11g WiFi 2.4 GHz (ERP-OFDM, 6 Mbps, 96pc dc)	WLAN	8.36	±9.6
10317	AAA	IEEE 802.11a WiFi 5 GHz (OFDM, 6 Mbps, 96pc dc)	WLAN		
10352	AAA	Pulse Waveform (200 Hz, 10%)		8.36	±9.6
10353	AAA	Pulse Waveform (200 Hz, 20%)	Generic	10.00	±9.6
10354	AAA	Pulse Waveform (200 Hz, 40%)	Generic	6.99	±9.6
10355	AAA	Pulse Waveform (200 Hz, 60%)	Generic	3.98	±9.6
10356	AAA	Pulse Waveform (200 Hz, 80%)	Generic	2.22	±9.6
10387	AAA	OPSK Waveform (200 Hz, 80%)	Generic	0.97	±9.6
10388			Generic	5.10	±9.6
THE REAL PROPERTY AND	AAA	QPSK Waveform, 10 MHz	Generic	5.22	±9.6
10396	AAA	64-QAM Waveform, 100 kHz	Generic	6.27	±9.6
10399	AAA	64-QAM Waveform, 40 MHz	Generic	6.27	±9.6
0400	AAD	IEEE 802.11 ac WiFi (20 MHz, 64-QAM, 99pc dc)	WLAN	8.37	±9.6
10401	AAA	IEEE 802.11ac WiFi (40 MHz, 64-QAM, 99pc dc)	WLAN	8.60	±9.6
0402	AAA	IEEE 802.11 ac WiFi (80 MHz, 64-QAM, 99pc dc)	WLAN	8.53	±9.6
10403	AAB	CDMA2000 (1xEV-DO, Rev. 0)	CDMA2000	3.76	±9.6
0404	AAB	CDMA2000 (1xEV-DO, Rev. A)	CDMA2000	3.77	±9.6
0406	AAD	CDMA2000, RC3, SO32, SCH0, Full Rate	CDMA2000	5.22	±9.6
0410	AAA	LTE-TDD (SC-FDMA, 1 RB, 10 MHz, QPSK, UL Sub=2.3,4,7,8,9)	LTE-TDD	7.82	±9.6
0414	AAA	WLAN CCDF, 64-QAM, 40 MHz	Generic		
0415	AAA	IEEE 802.11b WiFi 2.4 GHz (DSSS, 1 Mbps, 99pc dc)	WLAN	8.54	±9.6
0416	AAA	IEEE 802.11g WiFi 2.4 GHz (ERP-OFDM, 6 Mbps, 99pc dc)	1.1077-018	1.54	±9,6
0417	AAA	IEEE 802.11a/h WiFi 5 GHz (OFDM, 6 Mbps, 99pc dc)	WLAN	8.23	±9.6
0418	AAA		WLAN	8.23	±9.6
0419	AAA	IEEE 802.11g WiFi 2.4 GHz (DSSS-OFDM, 6 Mbps, 99pc, Long)	WLAN	8,14	±9.6
0419	AAA	IEEE 802.11g WiFi 2.4 GHz (DSSS-OFDM, 6 Mbps, 99pc, Short)	WLAN	8.19	±9.6
the second s		IEEE 802.11ri (HT Greenfield, 7.2 Mbps, BPSK)	WLAN	8.32	±9.6
0423	AAA	IEEE 802.11n (HT Greenfield, 43.3 Mbps, 16-QAM)	WLAN	8.47	±9.6
0424	AAE	IEEE 802.11n (HT Greenfield, 72.2 Mbps, 64-QAM)	WLAN	8.40	±9.6
0425	AAE	IEEE 802.11n (HT Greenfield, 15 Mbps, BPSK)	WLAN	8.41	±9.6
0426	AAE	IEEE 802.11n (HT Greenfield, 90 Mbps, 16-QAM)	WLAN	8.45	±9.6
0427	AAB	IEEE 802.11n (HT Greenfield, 150 Mbps, 64-QAM)	WLAN	8.41	±9.6
0430	AAB	LTE-FDD (OFDMA, 5MHz, E-TM 3.1)	LTE-FDD	8.28	±9.6
0431	AAC	LTE-FDD (OFDMA, 10 MHz, E-TM 3.1)	LTE-FDD	8.38	±9.6
0432	AAB	LTE-FDD (OFDMA, 15 MHz, E-TM 3.1)	LTE-FDD	8.34	±9.6
0433	AAC	LTE-FDD (OFDMA, 20 MHz, E-TM 3.1)	LTE-FDD	8.34	±9.6
0434	AAG	W-CDMA (BS Test Model 1, 64 DPCH)	WCDMA	8.60	±9.6
0435	AAA	LTE-TDD (SC-FDMA, 1 RB, 20 MHz, QPSK, UL Sub)			
0447	AAA	LTE-FDD (OFDMA, 5 MHz, E-TM 3.1, Clipping 44%)	LTE-TDD	7.82	±9.6
0448	AAA	LTE-FDD (OFDMA, 10 MHz, E-TM 3.1, Clippin 44%)	LTE-FDD	7.56	±9.6
0449	AAC	LTE-FDD (OFDMA, 15 MHz, E-TM 3.1, Cliping 44%)	LTE-FDD	7.53	±9.6
0450	AAA	LTE-FDD (OFDMA, 15MHz, E-1M 3.1, Cliping 44%) LTE-FDD (OFDMA, 20 MHz, E-TM 3.1, Clipping 44%)	LTE-FDD	7.51	±9.6
0451	AAA	W-CDMA (BS Tast Model 1, 64 ODOLL OF	LTE-FDD	7.48	±9.6
0453	AAC	W-CDMA (BS Test Model 1, 64 DPCH, Clipping 44%)	WCDMA	7.59	±9.6
0456	at a set of second	Validation (Square, 10 ms, 1 ms)	Test	10.00	±9.6
al come	AAC	IEEE 802.11ac WiFi (160 MHz, 64-QAM, 99pc dc)	WLAN	8.63	±9.6
0457	AAC	UMTS-FDD (DC-HSDPA)	WCDMA	6.62	±9,6
0458	AAC	CDMA2000 (1xEV-DO, Rev. B, 2 carriers)	CDMA2000	6.55	±9.6
0459	AAC	CDMA2000 (1xEV-DO, Rev. B, 3 carriers)	CDMA2000	8.25	±9.6
0460	AAC	UMTS-FDD (WCDMA, AMR)	WCDMA	2.39	±9.6
)461	AAC	LTE-TDD (SC-FDMA, 1 RB, 1.4 MHz, QPSK, UL Sub)	LTE-TDD	7.82	±9.6
0462	AAC	LTE-TDD (SC-FDMA, 1 RB, 1.4 MHz, 16-QAM, UL Sub)	LTE-TDD	8.30	±9.6
0463	AAD	LTE-TDD (SC-FDMA, 1 RB, 1.4 MHz, 64-QAM, UL Sub)	LTE-TDD	8.56	±9.6
1464	AAD	LTE-TDD (SC-FDMA, 1 RB, 3 MHz, QPSK, UL Sub)	LTE-TDD	7.82	±9.6
0465	AAC	LTE-TDD (SC-FDMA, 1 RB, 3 MHz, 16-QAM, UL Sub)	LTE-TDD	8.32	
0466	AAC	LTE-TDD (SC-FDMA, 1 RB, 3 MHz, 64-QAM, UL Sub)	LTE-TDD		±9.6
0467	AAA	LTE-TDD (SC-FDMA, 1 RB, 5 MHz, QPSK, UL Sub)	LTE-TDD	8.57	±9.6
0468	AAF	LTE-TDD (SC-FDMA, 1 RB, 5MHz, 16-QAM, UL Sub)	and the second se	7.82	±9.6
0469	AAD	LTE-TDD (SC-FDMA, 1 RB, 5MHz, 64-QAM, UL Sub)	LTE-TDD	8.32	±9.6
0470	AAD	LTE-TDD (SC-FDMA, 1 RB, 10 MHz, QPSK, UL Sub)	LTE-TDD	8.56	±9.6
	1.0.10		LTE-TDD	7.82	±9.6
471	AAC	LTE-TDD (SC-FDMA, 1 RB, 10 MHz, 16-QAM, UL Sub)	LTE-TDD	8.32	±9.6

UID 10472	Rev	Communication System Name	Group	PAR (dB)	Unc <sup>E</sup> k = 2
10472	-	LTE-TDD (SC-FDMA, 1 RB, 10 MHz, 64-QAM, UL Sub)	LTE-TDD	8.57	±9.6
		The rest of the rest of the start of the sta	LTE-TDD	7.82	±9.6
10474		THE TOTAL THE TOTAL TOTAL OF SUD	LTE-TDD	8.32	±9.6
10475	- Andrews	LTE-TDD (SC-FDMA, 1 RB, 15 MHz, 64-QAM, UL Sub)	LTE-TDD	8.57	±9.6
10477		LTE-TDD (SC-FDMA, 1 RB, 20 MHz, 16-QAM, UL Sub)	LTE-TDD	8.32	±9.6
10478	AAC	LTE-TDD (SC-FDMA, 1 RB, 20 MHz, 64-QAM, UL Sub)	LTE-TOD	8.57	±9.6
10479	AAC	LTE-TDD (SC-FDMA, 50% RB, 1.4 MHz, QPSK, UL Sub)	LTE-TDD	7,74	±9.6
10480	AAA	LTE-TDD (SC-FDMA, 50% RB, 1.4 MHz, 16-QAM, UL Sub)	LTE-TDD	8.18	±9.6
10481	AAA	LTE-TDD (SC-FDMA, 50% RB, 1.4 MHz, 64-QAM, UL Sub)	LTE-TDD	8.45	100.041
10482	AAA	LTE-TDD (SC-FDMA, 50% RB, 3 MHz, QPSK, UL Sub)	LTE-TDD		±9,6
10483	AAA	LTE-TDD (SC-FDMA, 50% RB, 3 MHz, 16-QAM, Sub)	LTE-TDD	7.71	±9.6
10484	AAB	LTE-TDD (SC-FDMA, 50% RB, 3 MHz, 64-QAM, UL Sub)	LTE-TDD	8.39	±9.6
10485	AAB	LTE-TDD (SC-FDMA, 50% RB, 5 MHz, QPSK, UL Sub)	LTE-TDD	8.47	±9.6
10486	AAB	LTE-TDD (SC-FDMA, 50% RB, 5 MHz, 16-QAM, UL Sub)	and the second se	7.59	±9.6
10487	AAC	LTE-TDD (SC-FDMA, 50% RB, 5MHz, 64-QAM, UL Sub)	LTE-TOD	8.38	±9.6
10488	AAC	LTE-TDD (SC-FDMA, 50% RB, 10 MHz, QPSK, UL Sub)	LTE-TDD	8.60	±9.6
10489	AAC	LTE-TDD (SC-FDMA, 50% RB, 10 MHz, 16-QAM, UL Sub)	LTE-TOD	7.70	±9.6
10490	AAF	LTE-TDD (SC-FDMA, 50% RB, 10 MHz, 64-QAM, UL Sub)	LTE-TOD	8.31	±9.6
10491	AAF	LTE-TDD (SC-FDMA, 50% RB, 15 MHz, QPSK, UL Sub)	LTE-TDD	8.54	±9.6
10492	AAF	LTE-TDD (SC-FDMA, 50% RB, 15 MHz, 16-QAM, UL Sub)	LTE-TDD	7.74	±9.6
10493	AAF	TE-TOD (SC-FDMA, 50% PD (SMHZ, 16-QAM, UL Sub)	LTE-TDD	8.41	±9.6
10494	AAF	LTE-TDD (SC-FDMA, 50% RB, 15 MHz, 64-QAM, UL Sub) LTE-TDD (SC-FDMA, 50% RB, 20 MHz, QPSK, UL Sub)	LTE-TDD	8.55	±9.6
10495	AAF	LTE-TDD (SC-FDMA, 50% RB, 20 MHz, QPSK, UL Sub)	LTE-TDD	7.74	±9.6
10496	AAE	LTE-TDD (SC-FDMA, 50% RB, 20 MHz, 16-QAM, UL Sub)	LTE-TDD	8.37	±9.6
10497	AAE	LTE-TDD (SC-FDMA, 50% RB, 20 MHz, 64-QAM, UL Sub)	LTE-TDD	8.54	±9.6
10498		LTE-TDD (SC-FDMA, 100% RB, 1.4 MHz, QPSK, UL Sub)	LTE-TDD	7.67	±9.6
10498	AAE	LTE-TDD (SC-FDMA, 100% RB, 1.4 MHz, 16-QAM, UL Sub)	LTE-TDD	8.40	±9.6
	AAC	LTE-TDD (SC-FDMA, 100% RB, 1.4 MHz, 64-QAM, UL Sub)	LTE-TDD	8.68	±9.6
10500	AAF	LTE-TDD (SC-FDMA, 100% RB, 3 MHz, QPSK, UL Sub)	LTE-TDD	7.67	±9.6
0501	AAF	LTE-TDD (SC-FDMA, 100% RB, 3 MHz, 16-QAM, UL Sub)	LTE-TDD	8.44	±9.6
0502	AAB	LTE-TDD (SC-FDMA, 100% RB, 3 MHz, 64-OAM, UL Sub)	LTE-TDD	8.52	±9.6
0503	AAB	LTE-TDD (SC-FDMA, 100% RB, 5 MHz, QPSK, UL Sub)	LTE-TDD	7.72	±9.6
0504	AAB	LTE-TDD (SC-FDMA, 100% RB, 5 MHz, 16-QAM, UL Sub)	LTE-TDD	8.31	±9.6
10505	AAC	LTE-TDD (SC-FDMA, 100% RB, 5MHz, 64-QAM, UL Sub)	LTE-TDD	8.54	
0506	AAC	LTE-TDD (SC-FDMA, 100% RB, 10 MHz, QPSK, UL Sub)	LTE-TDD	7.74	±9.6
0507	AAC	LTE-TDD (SC-FDMA, 100% RB, 10 MHz, 16-QAM, UL Sub)	LTE-TDD		±9.6
0508	AAF	LTE-TDD (SC-FDMA, 100% RB, 10 MHz, 64-QAM, UL Sub)	LTE-TDD	8.36	±9.6
0509	AAF	LTE-TDD (SC-FDMA, 100% RB, 15MHz, QPSK, UL Sub)	LTE-TOD	8.55	±9.6
0510	AAF	LTE-TDD (SC-FDMA, 100% RB, 15 MHz, 16-QAM, UL Sub)	the second se	7.99	±9.6
0511	AAF	LTE-TDD (SC-FDMA, 100% RB, 15 MHz, 64-QAM, UL Sub)	LTE-TDD	8.49	±9.6
0512	AAF	LTE-TOD (SC-FDMA, 100% RB, 20 MHz, QPSK, UL Sub)	the second se	8.51	±9.6
0513	AAF	LTE-TDD (SC-FDMA, 100% RB, 20 MHz, 16-QAM, UL Sub)	LTE-TDD	7,74	±9.6
0514	AAE	LTE-TDD (SC-FDMA, 100% RB, 20 MHz, 64-QAM, UL Sub)	LTE-TDD	8.42	±9.6
0515	AAE	IEEE 802.11b WiFi 2.4 GHz (DSSS, 2 Mbps, 99pc dc)	LTE-TDD	8.45	±9.6
0516	AAE	IEEE 802.11b WiFi 2.4 GHz (DSSS, 5.5 Mbps, 99pc dc)	WLAN	1.58	±9.6
0517	AAF	IEEE 802.11b WiFi 2.4 GHz (DSSS, 5.5 Mbps, 99pc dc)	WLAN	1.57	±9.6
0518	AAF	IEEE 802.11a/h WIFI 5 GHz (OFDM, 9 Mbps, 99pc dc)	WLAN	1.58	±9.6
0519	100 C 10 C 10 C	IEEE 802.11a/h WiFI 5 GHz (OFDM, 9 Mbps, 99pc dc)	WLAN	8.23	±9.6
Contraction of the	AAB	IEEE 802 11a/b WEE 5 CH2 (OF DM, 12 Mbps, 99pc dc)	WLAN	8.39	±9.6
0521	AAB	IEEE 802.11a/h WiFi 5 GHz (OFDM, 18 Mbps, 99pc dc)	WLAN	8.12	±9.6
-	AAB	IEEE 802 11a/h WiFi 5 GHz (OFDM, 24 Mbps, 99pc dc)	WLAN	7.97	±9.6
	and the second se	IEEE 802.11a/h WIFI 5 GHz (OFDM, 36 Mbps, 99pc dc)	WLAN	8.45	±9.6
	AAC	IEEE 802.11a/h WiFi 5 GHz (OFDM, 48 Mbps, 99pc dc)	WLAN	8.08	±9.6
The second s	AAC	IEEE 802.11a/h WiFi 5 GHz (OFDM, 54 Mbps, 99pc dc)	WLAN	8.27	±9.6
No.	AAE	IEEE 802.11ac WiFi (20 MHz, MCS0, 99pc dc)	WLAN	8.36	±9.6
and the second se	AAF	IEEE 802.11ac WiFI (20 MHz, MCS1, 99pc dc)	WLAN	8.42	±9.6
Contractor of the	AAF	IEEE 802.11ac WiFi (20 MHz, MCS2, 99pc dc)	WLAN	8.21	±9.6
Contractory on the local division of the loc	AAF	IEEE 802.11ac WiFi (20 MHz, MCS3, 99pc dc)	WLAN	8.36	±9.6
	AAF	IEEE 802.11ac WiFi (20 MHz, MCS4, 99pc dc)	WLAN	8.36	±9.6
	AAF	IEEE 802.11ac WiFi (20 MHz, MCS6, 99pc dc)	WLAN	8.43	±9.6
the second s	AAF	IEEE 802.11ac WiFi (20 MHz, MCS7, 99pc dc)	WLAN	8.29	
the second second	AAE	IEEE 802.11ac WiFi (20 MHz, MCS8, 99pc dc)	WLAN	8.38	±9.6
	AAE	IEEE 802.11ac WiFi (40 MHz, MCS0, 99pc dc)	WLAN		±9.6
the second s	AAE	IEEE 802.11ac WiFi (40 MHz, MCS1, 99pc dc)	WLAN	8.45	±9.6
536	AAF	IEEE 802.11ac WiFi (40 MHz, MCS2, 99pc dc)	WLAN	8.45	±9.6
537	AAF	IEEE 802.11ac WiFi (40 MHz, MCS3, 99pc dc)	and the second s	8.32	±9.6
538	AAF	IEEE 802.11ac WiFi (40 MHz, MCS4, 99pc dc)	WLAN	8.44	±9.6
540 /	AAA	EEE 802.11ac WiFi (40 MHz, MCS6, 99pc dc)	WLAN	8.54	±9.6
		( a mine, model, seperoc)	WLAN	8.39	±9.6

1054	1 AAA	a standarion System Name	Group	PAR (dB)	Unc <sup>E</sup> k =
1054		I HELL OVER THE I (OUMPLE, MUS7, 990C dc)	WLAN	8.46	±9.6
1054		The sould will (+0 MHz, MUS8, 99Dc dc)	WLAN	8.65	±9.6
1054	200	40 MHZ, MCS9, 990C dc)	WLAN	8.65	
		Contract the contract mice with a second second	WLAN	8.47	±9.6
1054		1 Minute over 1 40 Minute Mic 51, 9900 Mc1	WLAN	8.55	±9.6
1054	-	Contract of the traction of traction of the tr	WLAN		±9.6
1054		IEEE 802.11ac WiFi (80 MHz, MCS3, 99pc dc)	WLAN	8.35	±9.6
1054	-	IEEE 802.11ac WiFi (80 MHz, MCS4, 99pc dc)	WLAN	8.49	±9.6
10550		IEEE 802.11ac WiFi (80 MHz, MCS6, 99pc dc)	and the second se	8.37	±9.6
1055	1 AAC	IEEE 802.11ac WiFi (80 MHz, MCS7, 99pc dc)	WLAN	8.38	±9.6
10552	2 AAC	IEEE 802.11ac WiFi (80 MHz, MCS8, 99pc dc)	WLAN	8.50	±9.6
10553	AAC	IEEE 802.11ac WiFi (80 MHz, MCS9, 99pc dc)	WLAN	8.42	±9.6
10554	AAC	IEEE 802.11ac WiFi (160 MHz, MCS0, 99pc dc)	WLAN	8.45	±9.6
10555	5 AAC	IEEE 802.11ac WiFi (160 MHz, MCS1, 99pc dc)	WLAN	8.48	±9.6
10556	AAC	IEEE 802.11ac WiFi (160 MHz, MCS2, 99pc dc)	WLAN	8.47	±9.6
10557	AAC	IEEE 802.11ac WiFi (160 MHz, MCS2, 99pc dc)	WLAN	8.50	±9.6
10558	AAC	IEEE 802.11ac WiFi (160 MHz, MCS3, 99pc dc)	WLAN	8.52	±9.6
10560	and the second second	IEEE 802 11ac WIFI (160 MHz, MCS4, 99pc dc)	WLAN	8.61	±9.6
10561		IEEE 802.11ac WiFi (160 MHz, MCS6, 99pc dc)	WLAN	8.73	±9.6
10562		IEEE 802.11ac WiFi (160 MHz, MCS7, 99pc dc)	WLAN	8.56	±9.6
10563	1.0.100	IEEE 802.11ac WiFi (160 MHz, MCS8, 99pc dc)	WLAN	8.69	±9.6
10564		IEEE 802.11ac WiFi (160 MHz, MCS9, 99pc dc)	WLAN	8.77	±9.6
10565	-	IEEE 802.11g WiFi 2.4 GHz (DSSS-OFDM, 9 Mbps, 99pc dc)	WLAN	8.25	
0565	-	IEEE 802.11g WIFI 2.4 GHz (DSSS-OFDM_12 Mbns_99nc do)	WLAN	8.45	±9.6
	AAC	TEEE 802.11g WiFi 2.4 GHz (DSSS-OFDM 18 Mbps, 99pc da)	WLAN		±9.6
0567	AAC	IEEE 802.11g WiFi 2.4 GHz (DSSS-OFDM, 24 Mbos, 99nc do)	WLAN	8.13	±9.6
0568	AAC	TEEE 802.11g WiFi 2.4 GHz (DSSS-OFDM, 36 Mbns, 99pc do)	WLAN	8.00	±9.6
0569	AAC	TEEE 802.11g WIFI 2.4 GHz (DSSS-OFDM, 48 Mbps, 99pp de)		8.37	±9.6
0570	AAC	TEEE 802.11g WiFi 2.4 GHz (DSSS-OFDM, 54 Mbps, 99pc de)	WLAN	8.10	±9.6
0571	AAC	IEEE 802.11b WiFI 2.4 GHz (DSSS, 1 Mbps, 90pc dc)	WLAN	8.30	±9.6
0572	AAC	IEEE 802.11b WiFI 2.4 GHz (DSSS, 2 Mbps, 90pc dc)	WLAN	1.99	±9.6
0573	AAC	IEEE 802.11b WiFi 2.4 GHz (DSSS, 5.5 Mbps, 90pc dc)	WLAN	1.99	±9.6
0574	AAC	IEEE 802.11b WiFi 2.4 GHz (DSSS, 11 Mbps, 90pc dc)	WLAN	1.98	±9.6
0575	AAC	IEEE 802.11g WiFi 2.4 GHz (DSSS-OFDM, 6 Mbps, 90pc dc)	WLAN	1.98	±9.6
0576	AAC	IEEE 802.11g WiFi 2.4 GHz (DSSS-OFDM, 9 Mbps, 90pc dc)	WLAN	8.59	±9.6
0577	AAC	IEEE 802 11g WIE 24 GHz (DSSS-OFDM, 9 Mpps, 90pc dc)	WLAN	8.60	±9.6
0578	AAD	IEEE 802.11g WiFi 2.4 GHz (DSSS-OFDM, 12 Mbps, 90pc dc)	WLAN	8.70	±9.6
0579	AAD	IEEE 802.11g WiFi 2.4 GHz (DSSS-OFDM, 18 Mbps, 90pc dc)	WLAN	8.49	±9.6
0580	AAD	IEEE 802.11g WiFi 2.4 GHz (DSSS-OFDM, 24 Mbps, 90pc dc)	WLAN	8.36	±9.6
0581	AAD	IEEE 802.11g WiFi 2.4 GHz (DSSS-OFDM, 36 Mbps, 90pc dc)	WLAN	8.76	±9.6
0582	AAD	IEEE 802.11g WiFi 2.4 GHz (DSSS-OFDM, 48 Mbps, 90pc dc)	WLAN	8.35	±9.6
0583	AAD	IEEE 802.11g WiFi 2.4 GHz (DSSS-OFDM, 54 Mbps, 90pc dc)	WLAN	8.67	±9.6
0584	AAD	IEEE 802.11a/h WiFi 5 GHz (OFDM, 6 Mbps; 90pc dc)	WLAN	8.59	the loss of the lo
585	and the second se	IEEE 802.11a/h WiFi 5 GHz (OEDM 9 Mbor, 90pg do)	WLAN	8.60	±9.6
	AAD	IEEE 802.11a/h WiFi 5 GHz (OFDM, 12 Mbps, 90nc do)	WLAN		±9.6
586	AAD	IEEE 802.11a/h WiFi 5 GHz (OFDM, 18 Mbns, 90oc de)	WLAN	8,70	±9.6
587	AAA	IEEE 802.11a/h WiFi 5 GHz (OFDM, 24 Mbns, 90nc do)	WLAN	8.49	±9.6
588	AAA	IEEE 802.11a/h WiFi 5 GHz (OFDM, 36 Mbps, 90nc do)	and the second se	8.36	±9.6
589	AAA	IEEE 802.11a/h WIFI 5 GHz (OFDM, 48 Mbos, 90nc do)	WLAN	8.76	±9.6
590	AMA	IEEE 802.11a/h WiFi 5 GHz (OFDM, 54 Mbps, 90nc da)	WLAN MI	8.35	±9.6
591	AAA	IEEE 802.11n (HT Mixed, 20 MHz, MCS0, 90pc dc)	WLAN	8.67	±9.6
592	AAA	IEEE 802.11n (HT Mixed, 20 MHz, MCS0, 90pc dc)	WLAN	8.63	±9.6
593	AAA	IEEE 802.11n (HT Mixed, 20 MHz, MCS2, 90pc dc)	WLAN	8.79	±9.6
594	AAA	IEEE 802.11n (HT Mixed, 20 MHz, MCS3, 90pc dc)	WLAN	8.64	±9.6
595	AAA	IEEE 802.11n (HT Mixed, 20 MHz, MCS3, 90pc dc)	WLAN	8.74	±9.6
596	AAA	IEEE 802.11n (HT Mixed, 20 MHz, MCS4, 90pc dc)	WLAN	8.74	±9.6
	AAA	IEEE 802 11n (HT Mixed 201012, MCS5, 90pc dc)	WLAN	8.71	±9.6
	AAA	IEEE 802.11n (HT Mixed, 20 MHz, MCS6, 90pc dc)	WLAN	8.72	±9.6
and the second second	AAA	EEE 802.11n (HT Mixed, 20 MHz, MCS7, 90pc dc)	WLAN	8.50	±9.6
	AAA	EEE 802.11n (HT Mixed, 40 MHz, MCS0, 90pc dc)	WLAN	8.79	±9.6
	AAA	EEE 802 11n (HT Mixed, 40 MHz, MCS1, 90pc dc)	WLAN	8.88	
	AAA	EEE 802.11n (HT Mixed, 40 MHz, MCS2, 90pc dc)	WLAN	8.82	±9.6
	AAA	EEE 802.11n (HT Mixed, 40 MHz, MCS3, 90nc do)	WLAN		±9.6
	MAA	EEE 802.11n (HT Mixed, 40 MHz, MCS4, 90oc dc)	WLAN	8.94	±9.6
	AAA	EEE 802.11n (HT Mixed, 40 MHz, MCS5, 90nc do)	WLAN	9.03	±9.6
-	AAA II	EEE 802.11n (HT Mixed, 40 MHz, MCS6, 90nc do)	WLAN	8.76	±9.6
the second second	AAC   I	EEE 802.11n (HT Mixed, 40 MHz, MCS7, 90nc do)		8.97	±9.6
and the second division of	AAC I	EEE 802.11ac WiFi (20 MHz, MCS0, 90oc do)	WLAN	8.82	±9.6
08 /	AAC II	EEE 802.11ac WiFi (20 MHz, MCS1, 90pc dc)	WLAN	8.64	±9.6
	and the state of t	(of a start ways of a start of a	WLAN	8.77	±9.6

1060	9 Re	System Name	Group	PAR (dB)	Unc <sup>E</sup> k =
1061	and the second se	C IFFE 802 11ac WIFI (20 MHz, MCS2, 90pc dc)	WLAN	8.57	±9.6
1061		The source of the rest with rest wit	WLAN	8.78	±9.6
1061		(20 MHZ, MCS4, 90pc dc)	WLAN	8.70	±9.6
1061		C IEEE 802.11ac WiFI (20 MHz, MCS5, 90pc dc)	WLAN	8.77	±9.6
1061	1. A.	C IEEE 802.11ac WiFI (20 MHz, MCS6, 90pc dc)	WLAN	8.94	±9.6
1061	5 AA	C IEEE 802.11ac WiFi (20 MHz, MCS7, 90pc dc)	WLAN	8.59	±9.6
1061	6 AA	C IEEE 802.11ac WiFi (40 MHz, MCS8, 90pc dc)	WLAN	8.82	±9.6
1061	and the second se	C IEEE 802.11ac WiFi (40 MHz, MCS0, 90pc dc)	WLAN	8.82	±9.6
1061	8 AA	C IEEE 802.11ac WiFi (40 MHz, MCS1, 90pc dc)	WLAN	8.81	±9.6
1061	9 AAG	C IEEE 802.11ac WiFi (40 MHz, MCS2, 90pc dc)	WLAN	8.58	±9.6
1062	0 AAG	C IEEE 802.11ac WiFi (40 MHz, MCS3, 90pc dc)	WLAN	8.86	±9.6
1062	1 AAG	C IEEE 802.11ac WiFi (40 MHz, MCS4, 90pc dc)	WLAN	8.87	±9.6
1062	2 AAG	C IEEE 802.11ac WiFi (40 MHz, MCS6, 90pc dc)	WLAN	8.77	±9.6
10623	3 AAC	C IEEE 802.11ac WiFi (40 MHz, MCS7, 90pc dc)	WLAN	8.68	±9.6
10624	AAC	C IEEE 802.11ac WiFi (40 MHz, MCS8, 90pc dc)	WLAN	8.82	±9.6
10625	S AAC	C IEEE 802.11ac WiFi (40 MHz, MCS9, 90pc dc)	WLAN	8.96	±9.6
10626	AAC	IEEE 802.11ac WiFi (80 MHz, MCS0, 90pc dc)	WLAN	8.96	±9.6
10627	AAC	IEEE 802.11ac WiFi (80 MHz, MCS1, 90pc dc)	WLAN	8.83	±9.6
10628	AAC	IEEE 802.11ac WiFi (80 MHz, MCS2, 90pc dc)	WLAN	8.88	±9.6
10629	AAC	IEEE 802.11ac WiFi (80 MHz, MCS3, 90pc dc)	WLAN	8.71	±9.6
10630	AAC	IEEE 802.11ac WiFi (80 MHz, MCS3, 90pc dc)	WLAN	8.85	±9.6
10631	AAC	IEEE 802.11ac WiFi (80 MHz, MCS4, 90pc dc)	WLAN	8.72	±9.6
10632	AAC	IEEE 802.11ac WiFi (80 MHz, MCS6, 90pc dc)	WLAN	8.81	±9.6
10633	AAC	IEEE 802.11ac WiFI (80 MHz, MCS8, 90pc dc)	WLAN	8.74	±9.6
10634	AAC	IEEE 802.11ac WiFi (80 MHz, MCS8, 90pc dc)	WLAN	8.83	±9.6
10635	AAC	IEEE 802.11ac WiFi (80 MHz, MCS9, 90pc dc)	WLAN	8.80	±9.6
10636	AAC	IEEE 802.11ac WiFi (160 MHz, MCS9, 90pc dc)	WLAN	8.81	±9.6
10637	AAC	IEEE 802.11ac WiFi (160 MHz, MCS0, 90pc dc)	WLAN	8.83	±9.6
10538	AAC	IEEE 802.11 ac WiFi (160 MHz, MCS1, 90pc dc)	WLAN	8.79	±9.6
10639	AAC	IEEE 802.11ac WiFi (160 MHz, MCS2, 90pc dc)	WLAN	8.86	±9.6
10640	AAC	IEEE 802.11ac WiFi (160 MHz, MCS3, 90pc dc)	WLAN	8.85	±9.6
10641	AAC	IEEE 802.11ac WiFi (160 MHz, MCS4, 90pc dc)	WLAN	8.98	±9.6
10642	AAC	IEEE 802.11ac WiFi (160 MHz, MCS5, 90pc dc)	WLAN	9.06	±9.6
10643	AAC	IEEE 802.11ac WiFi (160 MHz, MCS6, 90pc dc)	WLAN	9.06	±9.6
10644	AAC	IEEE 802.11ac WiFi (160 MHz, MCS7, 90pc dc)	WLAN	8.89	±9.6
10645	AAC	IEEE 802.11ac WiFi (160 MHz, MCS8, 90pc dc)	WLAN	9.05	±9.6
10646	AAC	LTE-TDD (SC-FDMA, 1 RB, 5 MHz, QPSK, UL Sub=2,7)	WLAN	9.11	±9.6
10647	AAC	LTE-TDD (SC-FDMA, 1 RB, 20 MHz, QPSK, UL Sub=2,7)	LTE-TDD	11.96	±9.6
10648	AAC	CDMA2000 (1x Advanced)	LTE-TDD	11.96	±9.6
10652	AAC	LTE-TDD (OFDMA, 5MHz, E-TM 3.1, Clipping 44%)	CDMA2000	3.45	±9.6
10653	AAC	LIE-IDD (OFDMA 10 MH2 E TM 2.1 Official And	LTE-TDD	6.91	±9.6
10654	AAC	LTE-TDD (OFDMA, 15 MHz, E-TM 3.1, Clipping 44%)	LTE-TDD	7.42	±9.6
10655	AAC	LTE-TDD (OFDMA, 20 MHz, E-TM 3.1, Clipping 44%)	LTE-TDD	6.96	±9.6
10658	AAC	Pulse Waveform (200 Hz, 10%)	LTE-TDD	7.21	±9.6
10659	AAC	Pulse Waveform (200 Hz, 20%)	Test	10.00	±9.6
10660	AAC	Pulse Waveform (200 Hz, 40%)	Test	6,99	±9.6
10661	AAC	Pulse Waveform (200 Hz, 60%)	Test	3.98	±9.6
0662	AAC	Pulse Waveform (200 Hz, 80%)	Test	2.22	±9.6
0670	AAC	Bluetooth Low Energy	Test	0.97	±9.6
0671	AAD	IEEE 802.11ax (20 MHz, MCS0, 90pc dc)	Bluetooth	2.19	±9.6
0672	AAD	IEEE 802.11ax (20 MHz, MCS0, 90pc dc)	WLAN	9.09	±9.6
0673	AAD	IEEE 802.11ax (20 MHz, MCS2, 90pc dc)	WLAN	8.57	±9.6
0674	AAD	IEEE 802.11ax (20 MHz, MCS3, 90pc dc)	WLAN	8.78	±9.6
and the second se	AAD	IEEE 802.11ax (20 MHz, MCS4, 90pc dc)	WLAN	8.74	±9.6
0676	AAD	IEEE 802.11ax (20 MHz, MCS5, 90pc dc)	WLAN	8.90	±9.6
the set of	AAD	IEEE 802.11ax (20 MHz, MCS6, 90pc dc)	WLAN	8.77	±9.6
and the second second second	AAD	IEEE 802.11ax (20 MHz, MCS7, 90pc dc)	WLAN	8.73	±9.6
the second s	AAD	IEEE 802.11ax (20 MHz, MCS8, 90pc dc)	WLAN	8.78	±9.6
and the second se	AAD	IEEE 802.11ax (20 MHz, MCS9, 90pc dc)	WLAN	8.89	±9.6
0681	AAG	IEEE 802.11ax (20 MHz, MCS10, 90pc dc)	WLAN	8.80	±9.6
0682	AAF	IEEE 802.11ax (20 MHz, MCS11, 90pc dc)	WLAN	8.62	±9.6
0683 /	AAA	IEEE 802.11ax (20 MHz, MCS0, 99pc dc)	WLAN	8.83	±9.6
0684	AAC	IEEE 802.11ax (20 MHz, MCS1, 99pc dc)	WLAN	8.42	±9.6
0685 /	AAC	EEE 802.11ax (20 MHz, MCS2, 99pc dc)	WLAN	8.26	±9.6
686 4	AAC I	EEE 802.11ax (20 MHz, MCS2, 99pc dc)	WLAN	8.33	±9.6
		(and the moust sale dc)	WLAN	8.28	±9.6

10687	AAE	Communication System Name IEEE 802.11ax (20 MHz, MCS4, 99pc dc)	Group	PAR (dB)	Unc <sup>E</sup> k =
10688	AAE	IEEE 802.11ax (20 MHz, MCS4, 99pc dc)	WLAN	8.45	±9.6
10689	AAD	IEEE 802.11ax (20 MHz, MCS5, 99pc dc)	WLAN	8.29	±9.6
10690	-	IEEE 802.11ax (20 MHz, MCS6, 99pc dc)	WLAN	8.55	±9.6
10691	AAB	IEEE 802.11ax (20 MHz, MCS7, 99pc dc)	WLAN	8.29	±9.6
10692	AAA	IEEE 802.11ax (20 MHz, MCS9, 99pc dc)	WLAN	8.25	±9.6
10693	AAA	IEEE 802.11ax (20 MHz, MCS10, 99pc dc)	WLAN	8.29	±9.6
10694	AAA	IEEE 802.11ax (20 MHz, MCS10, 99pc dc)	WLAN	8.25	±9.6
10695	AAA	IEEE 802.11ax (40 MHz, MCS), 90pc dc)	WLAN	8.57	±9.6
10696	AAA	IEEE 802.11ax (40 MHz, MCS1, 90pc dc)	WLAN	8.78	±9.6
10697	AAA	IEEE 802.11ax (40 MHz, MCS2, 90pc dc)	WLAN	8.91	±9.6
10698	AAA	IEEE 802.11ax (40 MHz, MCS3, 90pc dc)	WLAN	8.61	±9.6
10699	AAA	IEEE 802.11ax (40 MHz, MCS4, 90pc dc)	WLAN	8.89	±9.6
10700	AAA	IEEE 802.11ax (40 MHz, MCS5, 90pc dc)	WLAN	8.82	±9.6
10701	AAA	IEEE 802.11ax (40 MHz, MCS6, 90pc dc)	WLAN	8.73	±9.6
10702	AAA	IEEE 802.11ax (40 MHz, MCS7, 90pc dc)	WLAN	8.86	±9.6
10703	AAA	IEEE 802.11ax (40 MHz, MCS8, 90pc dc)	WLAN	8.70	±9.6
10704	AAA	IEEE 802.11ax (40 MHz, MCS9, 90pc dc)	WLAN	8.82	±9.6
10705	AAA	IEEE 802.11ax (40 MHz, MCS10, 90pc dc)	WLAN	8.56	±9.6
10706	AAC	IEEE 802.11ax (40 MHz, MCS11, 90pc dc)	WLAN	8.69	±9.6
10707	AAC	IEEE 802.11ax (40 MHz, MCS0, 99pc dc)	WLAN	8.66	±9.6
0708	AAC	IEEE 802.11ax (40 MHz, MCS1, 99pc dc)	WLAN	8.32	±9.6
10709	AAC	IEEE 802.11ax (40 MHz, MCS2, 99pc dc)	WLAN	8.55	±9.6
0710	AAC	IEEE 802.11ax (40 MHz, MCS3, 99pc dc)	WLAN	8.33	±9.6
0711	AAC	JEEE 802.11ax (40 MHz, MCS4, 99pc dc)	WLAN	8.29	±9.6
0712	AAC	IEEE 802.11ax (40 MHz, MCS5, 99pc dc)	WLAN	8.39	±9.6
0713	AAC	IEEE 802.11ax (40 MHz, MCS6, 99pc dc)	WLAN	8.67	±9.6
0714	AAC	IEEE 802.11ax (40 MHz, MCS7, 99pc dc)	WLAN	8.33	±9.6
0715	AAC	IEEE 802.11ax (40 MHz, MCS8, 99pc dc)	WLAN	8.26	±9.6
0716	AAC	IEEE 802.11ax (40 MHz, MCS9, 99pc dc)	WLAN	8.45	±9.6
0717	AAC	IEEE 802.11ax (40 MHz, MCS10, 99pc dc)	WLAN	8.30	±9.6
0718	AAC	IEEE 802.11ax (40 MHz, MCS11, 99pc dc)	WLAN	8.48	±9.6
0719	AAC	IEEE 802.11ax (80 MHz, MCS0, 90pc dc)	WLAN	8.24	±9.6
and the second	AAC	IEEE 802.11ax (80 MHz, MCS1, 90pc dc)	WLAN	8.81	±9.6
termine of the local division of the	AAC	IEEE 802.11ax (80 MHz, MCS2, 90pc dc)	WLAN	8.87	±9.6
a sub-	AAC	IEEE 802.11ax (80 MHz, MCS3, 90pc dc)	WLAN	8.76	±9.6
	AAC	IEEE 802.11ax (80 MHz, MCS4, 90pc dc)	WLAN	8.55	±9.6
and the second se	AAC	IEEE 802.11ax (80 MHz, MCS5, 90pc dc)	WLAN	8.70	±9.6
A CONTRACTOR OF A CONTRACT	AAC	IEEE 802.11ax (80 MHz, MCS6, 90pc dc)	WLAN	8.90	±9.6
and the second se	AAC	IEEE 802.11ax (80 MHz, MCS7, 90pc dc)	WLAN	8.74	±9.6
	AAC	IEEE 802.11ax (80 MHz, MCS8, 90pc dc)	WLAN	8.72	±9.6
	AAC	IEEE 802.11ax (80 MHz, MCS9, 90nn de)	WLAN	8.66	±9.6
encertain in the	AAC	IEEE 802.11ax (80 MHz, MCS10, 90pc dc)	WLAN	8.65	±9.6
the second second second	AAG	IEEE 802.11ax (80 MHz, MCS11, 90pc dc)	WLAN	8.64	±9.6
the second se	AAC	IEEE 802.11ax (80 MHz, MCS0, 99pc dc)	WLAN	8.67	±9.6
California and State	AAC	IEEE 802.11ax (80 MHz, MCS1, 99pc dc)	WLAN	8.42	±9.6
the second s	AAC	IEEE 802.11ax (80 MHz, MCS2, 99pc dc)	WLAN	8.46	±9.6
and the second second	AAC	EEE 802.11ax (80 MHz, MCS3, 99pc dc)	WLAN	8.40	±9.6
	AAC I	EEE 802.11ax (80 MHz, MCS4, 99pc dc)	WLAN	8.25	±9.6
	AC 1	EEE 802.11ax (80 MHz, MCS5, 99pc dc)	WLAN	8.33	±9.6
	AC 1	EEE 802.11ax (80 MHz, MCS6, 99pc dc)	WLAN	8.27	±9.6
The second se	VAC 1	EEE 802.11ax (80 MHz, MCS7, 99pc dc)	WLAN	8.36	±9.6
the second second	VAC	EEE 802.11ax (80 MHz, MCS8, 99pc dc)	WLAN	8.42	±9.6
	AC 1	EEE 802.11ax (80 MHz, MCS9, 99pc dc)	WLAN	8.29	±9.6
and the second second	ACI	EEE 802.11ax (80 MHz, MCS10, 99pc dc)	WLAN WLAN	8.48	±9.6
and the owner of the owner	AC	EEE 802.11ax (80 MHz, MCS11, 99pc dc)	WLAN	8.40	±9.6
the state of the s	AC II	EEE 802.11ax (160 MHz, MCS0, 90pc dc)	WLAN	8.43	±9.6
the second se	AC I	EEE 802.11ax (160 MHz, MCS1, 90pc dc)	WLAN	8.94	±9.6
and the second	AC   IE	EE 802.11ax (160 MHz, MCS2, 90pc dc)	WLAN	9.16	±9.6
	AC IE	EE 802.11ax (160 MHz, MCS3, 90pc dc)	WLAN	8.93	±9.6
	AC TE	EE 802.11ax (160 MHz, MCS4, 90pc dc)	the second se	9.11	±9.6
the second second	AC IE	EE 802.11ax (160 MHz, MCS5, 90pc dc)	WLAN	9.04	±9.6
Colored Statements	AC   IE	EE 802.11ax (160 MHz, MCS6, 90pc dc)	WLAN	8.93	±9.6
the second se	AC   IE	EE 802.11ax (160 MHz, MCS7, 90pc dc)	WLAN	8.90	±9.6
and the second division of	AC IE	EE 802.11ax (160 MHz, MCS8, 90pc dc)	WLAN	8.79	±9.6
52 A/	AC IE	EE 802.11ax (160 MHz, MCS9, 90pc dc)	WLAN	8.82	±9.6
			WLAN	8.81	±9.6

10753	Rev AAC	System Name	Group	PAR (dB)	Unc <sup>E</sup> k =
10754		IEEE 802.11ax (160 MHz, MCS10, 90pc dc) IEEE 802.11ax (160 MHz, MCS11, 90pc dc)	WLAN	9.00	±9.6
10755		IEEE 802.11ax (160 MHz, MCS0, 99pc dc)	WLAN	8.94	±9.6
10756		IEEE 802.11ax (160 MHz, MCS0, 99pc dc)	WLAN	8.64	±9.6
10757		100 MHZ, MUS1, 990C (C)	WLAN	8.77	±9.6
10758	AAC	100 MHZ, MCSZ, 990c dc)	WEAN	8.77	±9.6
10759	AAC	IEEE 802.11ax (160 MHz, MCS3, 99pc dc)	WLAN	8.69	±9.6
10760	AAC	IEEE 802.11ax (160 MHz, MCS4, 99pc dc)	WLAN	8.58	±9.6
10761	AAC	IEEE 802.11ax (160 MHz, MCS5, 99pc dc)	WLAN	8.49	±9.6
10762	AAC	IEEE 802.11ax (160 MHz, MCS6, 99pc dc)	WLAN	8.58	±9.6
10763	AAC	IEEE 802.11ax (160 MHz, MCS7, 99pc dc)	WLAN	8.49	±9.6
10764	AAC	IEEE 802.11ax (160 MHz, MCS8, 99pc dc)	WLAN	8.53	±9.6
10765	AAC	IEEE 802.11ax (160 MHz, MCS9, 99pc dc)	WLAN	8.54	±9.6
10766	AAC	IEEE 802.11ax (160 MHz, MCS10, 99pc dc)	WLAN	8.54	±9.6
10767	AAC	5G NR (CP-OFDM, 1 RB, 5 MHz, QPSK, 15 kHz)	WLAN	8.51	±9.6
10768	AAC	5G NR (CP-OFDM, 1 RB, 10 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	7.99	±9.6
10769	AAC	5G NR (CP-OFDM, 1 RB, 15 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.01	±9.6
10770	AAC	5G NR (CP-OFDM, 1 RB, 20 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.01	±9.6
10771	AAC	5G NR (CP-OFDM, 1 RB, 25 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.02	±9.6
10772	AAC	5G NR (CP-OFDM, 1 RB, 30 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.02	±9.6
10773	AAC	5G NR (CP-OFDM, 1 RB, 40 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.23	±9.6
10774	AAC	5G NR (CP-OFDM, 1 RB, 50 MHz, OPSK, 15 kHz)	5G NR FR1 TDD	8.03	±9.6
10775	AAC	5G NR (CP-OFDM, 50% RB, 5 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.02	±9.6
10776	AAC	5G NR (CP-OFDM, 50% RB, 10 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.31	±9.6
10777	AAC	5G NR (CP-OFDM, 50% RB, 15 MHz, OPSK, 15 kHz)	5G NR FR1 TDD	8.30	±9.6
10778	AAC	5G NR (CP-OFDM, 50% RB, 20 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.30	±9.6
10779	AAC	5G NR (CP-OFDM, 50% RB, 25 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.34	±9.6
10780	AAC	5G NR (CP-OFDM, 50% RB, 30 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.42	±9.6
10781	AAC	5G NR (CP-OFDM, 50% RB, 40 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.38	±9.6
10782	AAC	5G NR (CP-OFDM, 50% RB, 50 MHz, QPSK, 15 KHz)	5G NR FR1 TDD	8.38	±9.6
10783	AAC	5G NR (CP-OFDM, 100% RB, 5 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.43	±9.6
10784	AAC	5G NR (CP-OFDM, 100% RB, 10 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.31	±9.6
10785	AAC	5G NR (CP-OFDM, 100% RB, 15 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.29	±9.6
10786	AAC	DG NH (CP-OFDM, 100% RB, 20 MHz OPSK 15 LUS)	5G NR FR1 TDD	8.40	±9.6
and the second se	AAC	5G NH (CP-OFDM, 100% RB, 25 MHz OPSK 15 HUN)	5G NR FR1 TDD	8.35	±9.6
	AAC	5G NH (CP-OFDM, 100% RB, 30 MHz OPSK 15 KHz)	5G NR FR1 TDD	8.44	±9.6
and an and a second	AAC	5G NH (CP-OFDM, 100% RB, 40 MHz OPSK 15 KHz)	5G NR FR1 TDD	8.39	±9.6
the local division in which the local division in the local divisi	AAC	5G NR (CP-OFDM, 100% RB, 50 MHz OPSK 15 KHz)	5G NR FR1 TDD	8.37	±9.6
of the local division in which the local division in the local div	AAC	DG NH (CP-OFDM, 1 RB, 5 MHz, OPSK 30 kHz)	5G NR FR1 TDD	8.39	±9.6
	AAC	5G NR (CP-OFDM, 1 RB, 10 MHz, OPSK 30 kHz)	5G NR FR1 TDD	7.83	±9.6
there are a second	AAC	5G NR (CP-OFDM, 1 RB, 15MHz, OPSK 30 kHz)	5G NR FR1 TDD	7.92	±9.6
and the second se	AAC	5G NR (CP-OFDM, 1 RB, 20 MHz, OPSK, 20 HUM	5G NR FR1 TDD	7.95	±9.6
	440	5G NR (CP-OFDM, 1 RB, 25 MHz, OPSK, 30 kHz)	5G NR FR1 TDD	7.82	±9.6
and the second second second	NAG	SGINH (CP-OFDM, 1 RB. 30 MHz, OPSK, 30 kHz)	5G NR FR1 TDD	7.84	±9.6
-	ANG .	5G NH (CP-OFDM, 1 RB, 40 MHz, OPSK 30 kHz)	5G NR FR1 TDD	7.82	±9.6
and the second se	140	5G NR (CP-OFDM, 1 RB, 50 MHz, OPSK, 30 kHz)	5G NR FR1 TDD 5G NR FR1 TDD	8.01	±9.6
and the second second second	140	SGINH (CP-OFDM, 1 RB, 60 MHz, OPSK 30 kHz)	5G NR FR1 TDD	7.89	±9.6
	MU	DG NH (CP-OFDM, 1 RB, 80 MHz, OPSK 30 kHz)	5G NR FR1 TDD	7.93	±9.6
	146	5G NH (CP-OFDM, 1 RB, 90 MHz, OPSK 30 kHz)	5G NR FR1 TDD	7.89	±9.6
Contraction in contraction of the local division of the local divi	WE :	G NR (CP-OFDM, 1 RB, 100 MHz, OPSK 30 kHz)	5G NR FR1 TDD	7.87	±9.6
and the state of t	040 13	DG NH (CP-OFDM, 50% RB, 10 MHz OPSK 30 KHz)	5G NR FR1 TDD	7.93	±9.6
	AD 3	OG NH (CP-OFDM, 50% RB, 15 MHz, OPSK 30 kHz)	5G NR FR1 TDD	8.34	±9.6
	nu :	GINH (CP-OFDM, 50% RB, 30 MHz, OPSK 30 kHz)	5G NR FR1 TDD	8.37	±9.6
	AD 1	SO NH (CP-OFDM, 50% RB, 40 MHz, OPSK, 20 KHz)	5G NR FR1 TDD	8.34	±9.6
	AU 12	G NH (CP-OFDM, 50% RB, 60 MHz, OPSK, 30 kHz)	5G NR FR1 TDD	8.34	±9.6
	10 3	GINH (CP-OFDM, 100% RB, 5 MHz, OPSK 30 kHz)	5G NR FR1 TDD	8.35	±9.6
	AU 5	G NR (CP-OFDM, 100% RB, 10 MHz, OPSK, 30 kHz)	5G NR FR1 TDD	8.35	±9.6
	10 3	GINH (CP-OFDM, 100% RB, 15MHz OPSK 20 Hust	5G NR FR1 TDD	8.33	±9.6
	AD D	GINH (CP-OFDM, 100% RB, 20 MHz, OPSK, 30 kHz)	5G NR FR1 TDD	8.30	±9.6
the second se	10 0	GINH (CP-OFDM, 100% RB, 25 MHz, OPSK 30 kHz)	5G NR FR1 TDD	8.41	±9.6
	10 million (1997)	G NR (CP-OFDM, 100% RB, 30 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	8.41	±9.6 ±9.6
	AD 5	G NR (CP-OFDM, 100% RB, 40 MHz, OPSK, 30 kHz)	5G NR FR1 TDD	8.36	
	1. C.	G NR (CP-OFDM, 100% RB, 50 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	8.39	±9.6
	D 50	G NR (CP-OFDM, 100% RB, 60 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	8.41	±9.6 ±9.6
828 A4	EE	3 NR (CP-OFDM, 100% RB, 80 MHz, QPSK, 30 kHz) 3 NR (CP-OFDM, 100% RB, 90 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	8.42	±9.6
M		THE OFUM, 100% HE SOME OPEN SOLUTION	5G NR FR1 TDD		A.0.0

1082	9 AA	section System Name	Group	PAR (dB)	Unc <sup>E</sup> k =
1083	1. C.	1 00 01 01 01 01 01 01 01 00 MB. 100 MHZ OPSK 20 LU-1	5G NR FR1 TDD	8.40	±9.6
1083	C	CONTROLOGICUM, THE, TUMHZ, OPSK 60 KH2)	5G NR FR1 TDD		±9.6
1083	1.5.7	CONTRACTOR OF DWILL THE TOTAL TOTAL CONSIGNATION OF DWILL THE TOTAL TO	5G NR FR1 TDD	7.73	
1083		(OF OF OW, FIRD, 20 MHZ, OPSK, 60 kHz)	5G NR FR1 TDD	7,74	±9.6
	7	CONTRACT OF DIVIL, I ND, 25 MMZ, QPSK 60 kHz)	5G NR FR1 TDD	7.70	±9.6
10834		D 5G NR (CP-OFDM, 1 RB, 30 MHz, OPSK, 60 kHz)	5G NR FR1 TDD		±9.6
10835	-	D 5G NR (CP-OFDM, 1 RB, 40 MHz, OPSK, 60 kHz)	5G NR FR1 TDD	7.75	±9.6
10836	-	E 5G NR (CP-OFDM, 1 RB, 50 MHz, OPSK, 60 kHz)		7.70	±9.6
10837		D 5G NR (CP-OFDM, 1 RB, 60 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	7.66	±9.6
10839	-	D 5G NR (CP-OFDM, 1 RB, 80 MHz, OPSK 60 kHz)	5G NR FR1 TDD	7.68	±9.6
10840	AAL	5G NR (CP-OFDM, 1 RB, 90 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	7.70	±9.6
10841	AAE	5G NR (CP-OFDM, 1 RB, 100 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	7.67	±9.6
10843	AAE	5G NR (CP-OFDM, 50% RB, 15MHz, QPSK, 60 kHz)	5G NR FR1 TDD	7.71	±9.6
10844	AAD	5G NR (CP-OFDM, 50% RB, 20 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	8.49	±9.6
10846	AAC	5G NR (CP-OFDM, 50% RB, 30 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	8.34	±9.6
10854	AAD	5G NB (CP.OEDM 100% PD 401111 005K, 60 kHz)	5G NR FR1 TDD	8.41	±9.6
10855	_	THE REPORT OF DWI, TOUGHE, TO MHZ, OPSK ROLLING	5G NR FR1 TDD	8.34	±9.6
10856		TOWN, TOUGO BB. TOMHZ OPSK ED (Ha)	5G NR FR1 TDD	8.36	±9.6
10857		TOTAL TOTAL TOTAL TOTAL TOTAL TOTAL TOTAL TOTAL	5G NR FR1 TDD	8.37	±9.6
10858		U PSK ED LAND	5G NR FR1 TDD	8.35	
	-	SG NR (CP-OFDM, 100% RB, 30 MHz, OPSK, 60 kHz)	5G NR FR1 TDD		±9.6
10859	AAD	5G NH (CP-OFDM, 100% RB, 40 MHz, OPSK, 60 kHz)	5G NR FR1 TDD	8.36	±9.6
10860	AAD	DO NH (CP-OFDM, 100% RB, 50 MHz, OPSK, 60 KHz)		8.34	±9.6
10861	AAD	5G NR (CP-OFDM, 100% RB, 60 MHz, OPSK, 60 HHz)	5G NR FR1 TDD	8.41	±9.6
10863	AAD	DG NR (CP-OFDM, 100% RB, 80 MHz, OPSK, 60 kHz)	5G NR FR1 TDD	8,40	±9.6
10864	AAE	5G NR (CP-OFDM, 100% RB, 90 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	8.41	±9.6
10865	AAD	5G NR (CP-OFDM, 100% RB, 100 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	8.37	±9.6
10866	AAD	5G NR (DFT-s-OFDM, 1 RB, 100 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	8.41	±9.6
10868	AAD	5G NB (DET & OEDM, 1988, 100 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.68	±9.6
10869	AAD	5G NR (DFT-s-OFDM, 100% RB, 100 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.89	±9.6
10870	AAD	5G NR (DFT-s-OFDM, 1 RB, 100 MHz, QPSK, 120 kHz)	5G NR FR2 TDD	5.75	±9.6
10871	AAD	5G NR (DFT-s-OFDM, 100% RB, 100 MHz, QPSK, 120 kHz)	5G NR FR2 TDD	5.86	
		5G NH (DFI-S-OFDM, 1 RB, 100 MHz 160AM 120 kHz)	5G NR FR2 TDD	and a second	±9.6
10872	AAD	3G NH (DFI-S-OFDM, 100% RB, 100 MHz, 160 AM, 120 HUL)	5G NR FR2 TDD	5.75	±9.6
10873	AAD	3G NH (DE1-S-OFDM, 1 RB, 100 MHz 640 AM, 120 LU-	5G NR FR2 TDD	6.52	±9.6
10874	AAD	5G NR (DFI-S-OFDM, 100% RB, 100 MHz 640 AM, 120 KU-S)	SG NR FR2 TOD	6.61	±9.6
10875	AAD	30 NA (CP-OFDM, 1 RB, 100 MHz OPSK 120 LUS)	5G NR FR2 TDD	6.65	±9.6
10876	AAD	5G NR (CP-OFDM, 100% RB, 100 MHz OPSK 120 kHz)	5G NR FR2 TDD	7.78	±9.6
10877	AAD	5G NR (CP-OFDM, 1 RB, 100 MHz, 16QAM, 120 kHz)	5G NR FR2 TDD	8.39	±9.6
10878	AAD	5G NR (CP-OFDM, 100% RB, 100 MHz, 16QAM, 120 kHz)	5G NR FR2 TDD	7.95	±9.6
10879	AAD	5G NR (CP-OFDM, 1 RB, 100 MHz, 64QAM, 120 kHz)	5G NR FR2 TDD	8.41	±9.6
10880	AAD	5G NR (CP-OFDM, 100% RB, 100 MHz, 64QAM, 120 kHz)	5G NR FR2 TDD	8.12	±9.6
10881	AAD	5G NR (DFT-s-OFDM, 1 RB, 50 MHz, QPSK, 120 kHz)	5G NR FR2 TDD	8.38	±9.6
10882	AAD	5G NB (DET & OFDM, 1989, 50 MHZ, QPSK, 120 kHz)	5G NR FR2 TDD	5.75	±9.6
10883	AAD	5G NR (DFT-s-OFDM, 100% RB, 50 MHz, QPSK, 120 kHz)	5G NR FR2 TDD	5.96	±9.6
10884	AAD	5G NR (DFT-s-OFDM, 1 RB, 50 MHz, 160AM, 120 kHz) 5G NR (DFT-s-OFDM, 1 RB, 50 MHz, 160AM, 120 kHz)	5G NR FR2 TDD	6.57	±9.6
	AAD	00 NH (0FT-S-OFDM, 100% RB, 50 MHz 160 AM 120 KH-1	5G NR FR2 TDD	6.53	
	and the second second second	5G NH (DEI-S-OFDM, 1 RB, 50 MHz, 640AM, 120 KHz)	5G NR FR2 TDD		±9.6
	AAD	5G NH (DFT-S-OFDM, 100% RB 50 MHz 640 AM 100 MLC)	5G NR FR2 TDD	6.61	±9.6
	AAD	3G NH (CP-OFDM, 1 RB, 50 MHz, OPSK 120 kHz)	5G NR FR2 TDD	6.65	±9.6
the second s	AAD	3G NH (CP-OFDM, 100% RB, 50 MHz, OPSK 120 LHz)	5G NR FR2 TDD	7.78	±9.6
the second se	AAD	3G NH (CP-OFDM, 1 RB, 50 MHz, 180AM, 120 kHz)	50 NR FR2 100	8.35	±9.6
and the second se	AAD	3G NH (GP-OFDM, 100% RB, 50 MHz, 160AM, 120 KHz)	5G NR FR2 TDD	8.02	±9.6
The second	AAD	3G NR (CP-OFDM, 1 RB, 50 MHz, 840AM, 120 kHz)	5G NR FR2 TDD	8.40	±9.6
The second second second second	AAD	5G NR (CP-OFDM, 100% RB, 50 MHz 640 AM 120 KHz)	5G NR FR2 TDD	8.13	±9.6
0897	AAD	5G NR (DFT-s-OFDM, 1 RB, 5MHz, QPSK, 30 kHz)	5G NR FR2 TDD	8.41	±9.6
	AAD	5G NR (DFT-s-OFDM, 1 RB, 10 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.66	±9.6
0899	AAD	5G NR (DFT-s-OFDM, 1 RB, 15 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.67	±9.6
and the second second	AAD	5G NR (DET-s-OEDM 1 PP ON HIL OPSK, 30 kHz)	5G NR FR1 TDD	5.67	±9.6
the second second	AAD	5G NR (DFT-s-OFDM, 1 RB, 20 MHz, QPSK, 30 kHz) 5G NB (DFT-s-OFDM, 1 RB, 20 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.68	±9.6
	AAD	5G NR (DFT-s-OFDM, 1 RB, 25 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.68	±9.6
	AAD	5G NR (DFT-s-OFDM, 1 RB, 30 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.68	and the second s
	AAD	5G NR (DFT-s-OFDM, 1 RB, 40 MHz, QPSK, 30 kHz)	5G NR FR1 TDD		±9.6
	MD	DG NH (DFT-S-OFDM, 1 RB, 50 MHz, OPSK, 30 kHz)	5G NR FR1 TDD	5.68	±9.6
and the local division of the local division	MU	SGINR (DET-S-OEDM, 1 RB, 60 MHz, OPSK, 30 kHz)	5G NR FR1 TDD	5.68	±9.6
	WD 3	DG NH (DF1-S-OFDM, 1 RB, 80 MHz, OPSK, 20 kHz)		5.68	±9.6
within the same	MU I	DG NH (DE I-S-OEDM, 50% RB, 5 MHz, OPSK, 30 kHz)	5G NR FR1 TDD	5.68	±9.6
Sector Se	AD !	5G NR (DFT-s-OFDM, 50% RB, 10 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.78	±9.6
0909 A	AD	5G NR (DFT-s-OFDM, 50% RB, 15 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.93	±9.6
0910 A	AD S	5G NR (DFT-s-OFDM, 50% RB, 20 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.96	±9.6
		COMPANY COMPANY CONTRACT OF STATES	5G NR FR1 TDD	5.83	±9.6

1091	Re AA	source and a system warne	Group	PAR (dB)	Unc <sup>E</sup> k =
10912		THE REPORT OF DWI, JUYO HE, 25 MHZ CIPSK 30 LUNI	5G NR FR1 TDD	5.93	±9.6
10913	- 6.5.7	THE REPORT OF DWI, SUM AB, SUMHZ OPSK SALUT	5G NR FR1 TDD	5.84	±9.6
10914		0 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1	5G NR FR1 TDD	5.84	±9.6
10915		101 13 01 0M, 50% HB, 50 MHZ, OPSK 30 kHz1	5G NR FR1 TDD	5.85	±9.6
10916		1 2 2 2 2 2 2 2 2 2 2 2 2 2 2 2 2 2 2 2	5G NR FR1 TDD	5.83	±9.6
10917		THE SUMPTION OF THE SUMPT OPSK 30 LUN	5G NR FR1 TDD	5.87	±9.6
10918			5G NR FR1 TDD	5.94	
		5G NR (DFT-S-OFDM, 100% RB, 5MHz, OPSK, 20 Hur)	5G NR FR1 TDD	5.86	±9.6
10919	100000	5G NR (DFT-S-OFDM, 100% RB, 10 MHz, OPSK, 30 kHz)	5G NR FR1 TDD	5.86	±9.6
10920		56 NR (DET-S-OEDM, 100% RB, 15 MHz, OPSK 30 KHz)	5G NR FR1 TDD	5.87	±9.6
10921	AAD	5G NR (DFT-S-OFDM, 100% RB, 20 MHz, OPSK, 30 kHz)	5G NR FR1 TDD	5.84	±9.6
10922	AAD	5G NR (DF1-S-OFDM, 100% RB, 25MHz, OPSK 30 kHz)	5G NR FR1 TDD		±9.6
10923	AAD	50 NR (DET-S-OFDM, 100% RB, 30 MHz, OPSK, 30 KHz)	5G NR FR1 TDD	5.82	±9.6
10924	AAD	5G NH (DFT-S-OFDM, 100% RB, 40 MHz, OPSK, 30 KHz)	5G NR FR1 TDD	5.84	±9.6
10925	AAD	DG NR (DFT-S-OFDM, 100% RB, 50 MHz, OPSK, 30 kHz)	and a second sec	5.84	±9.6
10926	AAD	5G NR (DFT-s-OFDM, 100% RB, 60 MHz, OPSK, 30 kHz)	5G NR FR1 TDD	5.95	±9.6
10927	AAD	5G NH (DFT-s-OFDM, 100% RB, 80 MHz, OPSK, 20 kHz)	5G NR FR1 TDD	5.84	±9.6
10928	AAD	5G NR (DFT-s-OFDM, 1 RB, 5MHz, OPSK 15 kHz)	5G NR FR1 TDD	5.94	±9.6
10929	AAD	5G NR (DFT-s-OFDM, 1 RB, 10 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.52	±9.6
10930	AAD	5G NR (DFT-s-OFDM, 1 RB, 15 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.52	±9.6
10931	AAD	5G NR (DFT-s-OFDM, 1 RB, 20 MHz, OPSK, 15 kHz)	5G NR FR1 FDD	5.52	±9.6
10932	AAB	5G NR (DFT-s-OFDM, 1 RB, 25 MHz, OPSK, 15 kHz)	5G NR FR1 FDD	5.51	±9.6
10933	AAA	5G NR (DFT-s-OFDM, 1 RB, 30 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.51	±9.6
10934	AAA	5G NR (DFT-s-OFDM, 1 RB, 40 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.51	±9.6
10935	AAA	5G NR (DFT-s-OFDM, 1 RB, 50 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.51	±9.6
10936	AAC	5G NR (DFT-s-OFDM, 50% RB, 5 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.51	±9.6
10937	AAB	5G NR (DFT-s-OFDM, 50% RB, 10 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.90	±9.6
10938	AAB	5G NB (DET & OFOM, 50% PB, 10 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.77	±9.6
10939	AAB	5G NR (DFT-s-OFDM, 50% RB, 15 MHz, OPSK, 15 kHz)	5G NR FR1 FDD	5.90	±9.6
10940	AAB	5G NR (DFT-s-OFDM, 50% RB, 20 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.82	±9.6
10941	AAB	5G NR (DFT-s-OFDM, 50% RB, 25 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.89	±9.6
10942	AAB	5G NR (DFT-s-OFDM, 50% RB, 30 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.83	±9.6
10943	AAB	5G NR (DFT-s-OFDM, 50% RB, 40 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.85	±9.6
10944	AAB	5G NR (DFT-s-OFDM, 50% RB, 50 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.95	±9.6
10945	AAB	5G NR (DFT-s-OFDM, 100% RB, 5MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.81	±9.6
10946	AAC	5G NR (DFT-s-OFDM, 100% RB, 10 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.85	±9.6
10947	AAB	5G NR (DFT-s-OFDM, 100% RB, 15 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.83	±9.6
10948	AAB	5G NR (DFT-s-OFDM, 100% RB, 20 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.87	±9.6
10949	AAB	5G NR (DFT-s-OFDM, 100% RB, 25 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.94	
10950	AAB	5G NR (DFT-s-OFDM, 100% RB, 30 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.87	±9.6
10951	AAB	5G NR (DFT-s-OFDM, 100% RB, 40 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.94	±9.6
		5G NH (DET-S-OFDM, 100% RB, 50 MHz, OPSK, 15 kHz)	5G NR FR1 FDD	5.92	±9.6
Gentle Contraction	AAB	SGINH UL (CP-OFDM, TM 3.1. 5MHz 64-OAM 15 HILL)	5G NR FR1 FDD	8.25	±9.6
	AAB	5G NR DL (CP-OFDM, TM 3.1, 10 MHz, 64-OAM, 15 kHz)	5G NR FR1 FDD		±9.6
	AAB	5G NH DL (CP-OFDM, TM 3.1, 15 MHz 64-OAM 15 HH-)	5G NR FR1 FDD	8.15	±9.6
	AAB	SG NH DL (CP-OFDM, TM 3.1, 20 MHz 64-OAM 15 LHz)	5G NR FR1 FDD	8.23	±9.6
	AAB	SGINH DL (CP-OFDM, TM 3.1, 5MHz 64-OAM 30 KHz)	5G NR FR1 FDD	8.42	±9.6
the second se	AAC	3G NH DL (CP-OFDM, TM 3.1, 10 MHz, 64-OAM 30 KHz)	5G NR FR1 FDD	8.14	±9.6
Territoria da la compañía de la comp	AAB	3G NH DL (CP-OFDM, TM 3.1, 15MHz 64-OAM 30 HU-)		8.31	±9.6
and a state of the	AAB	3G NH DL (CP-OFDM, TM 3.1, 20 MHz 64-OAM 30 HHz)	5G NR FR1 FDD	8.61	±9.6
the support	AAB	3G NH DL (CP-OFDM, TM 3.1, 5MHz, 64-OAM 15 +H+)	5G NR FR1 FDD	8.33	±9.6
and the second se	AAB	3G NH DL (CP-OFDM, TM 3.1, 10 MHz, 64-OAM, 15 kHz)	5G NR FR1 TDD	9.32	±9.6
the second se	AAB	3G NR DL (CP-OFDM, TM 3.1, 15 MHz 64-OAM 15 KHz)	5G NR FR1 TDD	9.36	±9.6
	040	3G NH DL (CP-OFDM, TM 3.1, 20 MHz, 64-OAM, 15 kHz)	5G NR FR1 TDD	9.40	±9.6
	200	3G NH DL (CP-OFDM, TM 3.1, 5MHz, 64-OAM, 30 KHz)	5G NR FR1 TDD	9.55	±9.6
and the second se	0.0	5G NR DL (CP-OFDM, TM 3.1, 10 MHz, 64-OAM, 30 kHz)	5G NR FR1 TDD	9.29	±9.6
and the second se	D	SG NH DL (CP-OFDM, TM 3.1, 15MHz B4-OAM 20 KHz)	5G NR FR1 TDD	9.37	±9.6
	AB	5G NR DL (CP-OFDM, TM 3.1, 20 MHz, 64-QAM, 30 KHz)	5G NR FR1 TDD	9.55	±9.6
0968 4	AB	5G NR DL (CP-OFDM, TM 3.1, 100 MHz, 64-QAM, 30 kHz)	5G NR FR1 TDD	9.42	±9.6
0972 A	AB	5G NR (CP-OFDM, 1 RB, 20 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	9.49	±9.6
973 A	AB	5G NR (DFT-s-OFDM, 1 RB, 100 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	11.59	±9.6
974 A	AB	5G NR (CP-OFDM, 100% RB, 100 MHz, 256-QAM, 30 kHz)	5G NR FR1 TDD	9.06	±9.6
and the second second	AA	ULLA BDR	5G NR FR1 TDD 1	0.28	±9.6
Address of the owner of		ULLA HDR4	1011.0	2.23	±9.6
		ULLA HDR8	X10.4 A	7.02	±9.6
Conception in which the		JLLA HDRp4	1111.4	8.82	±9.6
		JLLA HDRp8	1111.4	1.50	±9.6
	- er s   1		ULLA		10.0

UID	Rev	Communication System Name			
10983	AAA	5G NR DL (CP-OFDM, TM 3.1, 40 MHz, 64-QAM, 15 kHz)	Group	PAR (dB)	UncE k=2
10984	AAA	5G NR DL (CP-OFDM, TM 3.1, 40 MHz, 64-QAM, 15 kHz)	5G NR FR1 TDD	9.31	±9.6
10985	AAA	5G NR DL (CP-OFDM, TM 3.1, 50 MHz, 64-QAM, 15 kHz) 5G NR DL (CP-OFDM, TM 3.1, 40 MHz, 64-QAM, 30 kHz)	5G NR FR1 TDD	9.42	±9.6
10986	AAA	5G NR DL (CP-OFDM, TM 3.1, 40 MHz, 64-QAM, 30 kHz)	5G NR FR1 TDD	9.54	±9.6
10987	AAA	5G NR DL (CP-OFDM, TM 3.1, 50 MHz, 64-QAM, 30 kHz)	5G NR FR1 TDD	9.50	±9.6
10988	AAA	5G NR DL (CP-OFDM, TM 3.1, 50 MHz, 64-QAM, 30 kHz)	5G NR FR1 TDD	9.53	±9.6
10989	AAA	5G NR DL (CP-OFDM, TM 3.1, 70 MHz, 64-QAM, 30 kHz)	5G NR FR1 TDD	9.38	±9.6
10990	AAA	5G NR DL (CP-OFDM, TM 3.1, 90 MHz, 64-QAM, 30 kHz)	5G NR FR1 TDD	9.33	±9.6
		100 (01 01 01 01 01 01 01 01 01 01 01 01 01 0	5G NR FR1 TDD	9.52	±9.6

<sup>E</sup> Uncertainty is determined using the max. deviation from linear response applying rectangular distribution and is expressed for the square of the field value.