

Met Laboratories
914 West Patapsco Ave.
Baltimore, MD 21230

April 30, 2008

RE: RF exposure information FCC ID: V5AASDB1

<u>FCC ID Number</u>	<u>Product</u>	<u>Title / Model</u>
V5AASDB1		Medical Device – A Patient Worn Digital Sleep Recording Device / Brand name: Nox T3, Model name: ASDB1

TO WHOM IT MAY CONCERN

The product stated above complies with the FCC limits for Human exposure to RF emissions.

This device is categorically excluded from routine environmental, because it operates at very low power level. The equipment is deemed to comply with the SAR or MPE limits without testing due to its very low power level (EIRP < 2.5 mW).

Sincerely,

P.A.



Kolbrun Ottosdottir
Quality Manager
Nox Medical ehf
Keldnaholti, IS-112 Reykjavik, Iceland
Phone: +354-5229000
E-mail: keo@noxmedical.com