

Tissue Liquid:	HEAD							
Dipole Antenna:	835		1800		1900		2450	
S/N:	4d019		220		5d016		776	
Dipole Cal Date:	April 20, 2009		September 15, 2009		September 7, 2010		August 19, 2010	
Imp. verification Date:	March 22, 2010		August 27, 2010		NA		NA	
Impedance (ohm):	Real	Imag	Real	Imag	Real	Imag	Real	Imag
Target:	52.3	-3.7	47.2	-4.1				
Measured:	48.2	-7.6	44.1	-2.2				
Delta:	-4.1	-3.9	-3.1	1.9	0.0	0.0	0.0	0.0

Calibration Equipment Used:	Manuf.	Model	S/N #	Cal Due Date	Remarks
Network Analyzer	HP	8653E	US39172946	November 7, 2010	None