

---

**From:** oetech@fcc.gov  
**Sent:** Thursday, July 25, 2019 9:41 AM  
**To:** Rod Peloquin  
**Subject:** Response to Inquiry to FCC (Tracking Number 345183) (TCB)

**Importance:** High

**Inquiry on 07/25/2019 :**

**Inquiry:**

An applicant is seeking approval for a radio module that will be integrated into medical equipment, like patient monitors and Pulse CO-Oximeters that contain displays.

These types of products are not specifically listed in KDB 784748 Section 2.2 so we would like to know if they qualify to use e-labeling for the module.

Thanks

**FCC response on 07/25/2019**

Yes, if they have a display they qualify.

**Attachment Details:**

Do not reply to this message. Please select the [Reply to an Inquiry Response](#) link from the OET Inquiry System to add any additional information pertaining to this inquiry.