

Alcon Research Ltd.
20511 Lake Forest Dr., Lake Forest, CA 92630
T: 949.753.1393
www.alcon.com



Attestation Letter

May 01, 2014

Northwest EMC Inc.
22975 NW Evergreen Pkwy, Suite 400
Hillsboro, OR 97124

Ref: FCC ID: VMCNGPFSW1; Centurion® Vision System Footswitch

To Whom It May Concern:

Alcon Research Ltd., hereby attest that the design, functionality and intended use of the LuxOR Microscope Foot Controller charging circuitries and the Centurion System Footswitch charging circuitries are identical electrically and mechanically, with only minor changes to an enclosure (shape).

The LuxOR Microscope Foot Controller charger contains two air-core coils and circuitry performs three basic roles: it detects (the arrival and departure) of footswitch on the console hook, it provides energy transfer from the console to the footswitch, and it generates near-field communication of battery charging status between footswitch and console.

The technical reports and exhibits (FCC ID: VMCNGPFSW1) provided demonstrate compliance with FCC rules CFR 15.109; 15.209; 15.107; 15.207 and 18.305. The radio transmitter is not subject to routine environmental testing or evaluation.

Sincerely,

A handwritten signature in blue ink, appearing to read "Ahmad Salehi".

Ahmad Salehi
Dir. Development