


Certification And Engineering Bureau
Industry Canada
Spectrum Engineering Branch
3701 Carling Avenue, Building 94
Ottawa, Ontario K2H 8S2

15-Feb-2018

To Whom It May Concern

ID Label/ Label Location – T Moxi All and Vista T Rd 13 families

The hearing instruments Unitron T Moxi All and Vista T Rd 13 families, which are subject of this application, shall bear a permanently affixed, readily visible label. Due to the specific use of the hearing instrument (as shown photo below) its external surface is in contact with the user's skin in the ear.



Therefore, it is not appropriate to permanently affix a label on the outer surface of the device to prevent skin irritations.


Due to the small size of the devices the ID and the statement as required by RSS-210 of Industry Canada has been placed in the user manual as shown below.

The hearing aid described in this user guide is certified under:
Standard hearing system

	USA - FCC ID:	Canada - IC:
T Moxi All	VMY-UBTD1	2756A- UBTD1

The hearing aid described in this user guide is certified under:
Standard hearing system

	USA - FCC ID:	Canada - IC:
Vista T Rd 13	VMY-UBTD1	2756A- UBTD1



Amir Panah
Manager
Quality Assurance