

Federal Communication Commission  
Equipment Authorization Division,  
Application Processing Branch  
7435 Oakland Mills Road  
Columbia, MD 21048

06-Oct-2015

To Whom It May Concern

ID Label/ Label Location – Stride ITE Hearing Instruments & Vista N ITE Hearing Instruments

Pursuant to paragraph §2.925 of 47 C.F.R. the hearing instruments Unitron ITE (custom), which is subject of this filing, shall bear a permanently affixed, readily visible label listing the information as specified in §2.926 and §15.19(a) of the 47 C.F.R. Due to the specific use of the hearing instrument (as shown in figure 1) its external surface is in contact with the user's skin behind the ear.



Figure 1: Position of the Hearing Instrument on the User

Therefore, it is not appropriate to permanently affix a label on the outer surface of the device to prevent skin irritations. The only alternative place for the FCC ID and serial number of the device would be in the battery compartment.

In this case the size of the FCC ID will be smaller than 2 points, which does not satisfy the requirements for "readily visible" label as defined in §2.925 (d)(2) and §2.925(g).

Due to the small size of the devices and pursuant to paragraph § 15.19(a)(5) of 47 C.F.R., the FCC ID and the statement specified in §15.19(a)(3) are placed in the user manual as shown below.

	Stride ITE 13, 312	Stride ITE 10A	Stride ITE 10A omni
USA - FCC ID:	VMY-UWITE2	VMY-UWITE3	VMY-UWITE4
Canada - IC:	2756A-UWITE2	2756A-UWITE3	2756A-UWITE4

	Vista ITE 13, 312	Vista ITE 10A	Vista ITE 10A omni
USA - FCC ID:	VMY-UWITE2	VMY-UWITE3	VMY-UWITE4
Canada - IC:	2756A-UWITE2	2756A-UWITE3	2756A-UWITE4

*B. Matleski*

Brian Matleski  
Manager-Quality Assurance  
Unitron Hearing  
20 Beasley Drive Kitchener, ON  
519-895-0100 x2110