

Electromagnetic Compatibility Test Report

Tests Performed on a Delsys, Inc.

Trigno Centro Base Station, Model SP-W10

Radiometrics Document RP-9728B2



Product Detail:

FCC ID: W4P-SP-W10 IC: 8138A-SPW10

Equipment type: Biofeedback System with Transceiver

Test Standards:

US CFR Title 47, Chapter I, FCC Part 15 Subpart C

FCC Part 15 CFR Title 47: 2023

Canada ISED; RSS-210, Issue 10: 2019 as required for Category I Equipment

IC RSS-GEN Issue 5: 2018

This report concerns: Original Grant for Certification

FCC Part 15.249

Tests Performed For:	Test Facility:
Delsys, Inc.	Radiometrics Midwest Corporation
23 Strathmore Rd.	12 Devonwood Avenue
Natick, MA 01760	Romeoville, IL 60446-1349
·	(815) 293-0772
Test Date(s):	

September 5-15, 2023

Document RP-9728B2 Revisions:

Rev. Issue Date		Revised By			
0	December 22, 2023				
1	December 29, 2023	Joseph Strzelecki			

Radiomet.com Page 1 of 19



Table of Contents

1.0 ADMINISTRATIVE DATA	3
2.0 TEST SUMMARY AND RESULTS	3
2.1 RF Exposure Compliance Requirements	4
3.0 EQUIPMENT UNDER TEST (EUT) DETAILS	4
3.1 EUT Description	4
3.1.1 FCC Section 15.203 & RSS-GEN Antenna Requirements	4
4.0 TESTED SYSTEM DETAILS	
4.1 Tested System Configuration	4
4.2 EUT Operating Modes	4
4.3 Special Accessories	5
4.4 Equipment Modifications	5
5.0 TEST SPECIFICATIONS	5
6.0 TEST PROCEDURE DOCUMENTS	5
7.0 RADIOMETRICS' TEST FACILITIES	
8.0 DEVIATIONS AND EXCLUSIONS FROM THE TEST SPECIFICATIONS	6
9.0 CERTIFICATION	
10.0 TEST EQUIPMENT TABLE	6
11.0 TEST SECTIONS	
11.1 AC Conducted Emissions	
11.2 Radiated RF Emissions	
11.2.1 Field Strength Calculation	
11.2.2 Duty Cycle	
11.2.3 Radiated Emissions Test Results	
11.3 Occupied Bandwidth Data	
11.3.1 Measurement Instrumentation Uncertainty	
12.0 REVISION HISTORY	.19

Notice: This report must not be reproduced (except in full) without the written approval of Radiometrics Midwest Corporation.

B

Radiometrics Midwest Corporation

Testing of: Delsys, Inc., Model SP-W10, Trigno Centro Base Station

1.0 ADMINISTRATIVE DATA

Equipment Under Test:	
A Delsys, Inc., Trigno Centro Base Station	
Model: SP-W10 Serial Number: RMC-001	
This will be referred to as the EUT in this Report	
Date EUT Received at Radiometrics:	Test Date(s):
May 11, 2023	September 5-15, 2023
Test Report Written and Authorized By:	Test Witnessed By:
Joseph Strzelecki	The tests were not witnessed by personnel from
Senior EMC Engineer	Delsys, Inc.
Radiometrics' Personnel Responsible for Test:	EUT Checked By:
Joseph Strzelecki 12/22/2023	Joseph Strzelecki Chris D'Alessio Radiometrics
Date	
Joseph Strzelecki	
Senior EMC Engineer	
NARTE EMC-000877-NE	
Chris D'Alessio	
EMC Technician	

2.0 TEST SUMMARY AND RESULTS

The EUT (Equipment Under Test) is a Trigno Centro Base Station, Model SP-W10, manufactured by Delsys, Inc. The detailed test results are presented in a separate section. The following is a summary of the test results.

Emissions Tests Results

Environmental Phenomena	Frequency Range	Basic Standard	Test Result
RF Radiated Emissions	30-25,000 MHz	RSS-210 & FCC Part 15.249	Pass
Conducted Emissions, AC Mains	0.15 - 30 MHz	RSS-210 & FCC Part 15.207	Pass
Occupied Bandwidth Test	Fundamental Freq.	RSS-GEN & FCC Part 15	Pass

IEC 17025 Decision Rule:

The declaration of pass or fail is based on the specifications listed above. The declaration of pass or fail did not consider measurement uncertainty.

RP-9728B2 Rev. 1 Radiomet.com Page 3 of 19

R

Radiometrics Midwest Corporation

Testing of: Delsys, Inc., Model SP-W10, Trigno Centro Base Station

2.1 RF Exposure Compliance Requirements

Since the average power output is less than 10 mW, the EUT meets the FCC requirement for RF exposure and it is exempt from RSS-102 SAR and RF exposure evaluations. There are no power level adjustments available to the end user. The antenna is permanently attached. The detailed calculations for RF Exposure are presented in a separate document.

3.0 EQUIPMENT UNDER TEST (EUT) DETAILS

3.1 EUT Description

The EUT is a Trigno Centro Base Station, Model SP-W10, manufactured by Delsys, Inc. The EUT was in good working condition during the tests, with no known defects.

3.1.1 FCC Section 15.203 & RSS-GEN Antenna Requirements

The antenna has a reverse polarity SMA connector. RF-SMA is not readily available to the general public, therefore, it meets the 15.203 Requirements.

4.0 TESTED SYSTEM DETAILS

4.1 Tested System Configuration

The system was configured for testing in a typical fashion. The EUT was placed on an 80-cm or 150 cm high, nonconductive test stand. The testing was performed in conditions as close as possible to installed conditions.

The identification for all equipment, are:

Tested System Configuration List

Item	Description Type	e*	Manufacturer	Model Number	Serial Number
1	Trigno Centro Base Station	Е	Delsys, Inc.	SP-W10	RMC-001
2	Power Supply	Е	CUI	SDM24-12-UD	211600207
3	Personal Computer	Н	Dell	XPS 15 9570	9DDCC5E1D1CD
4	PC Power Supply	Н	Dell	DA130PM130	052-1A9Q-A02

^{*} Type: E = EUT, P = Peripheral, S = Support Equipment; H = Host Computer

Type of modulation including the bit rate and symbol rate	FSK, 1MBPS
Name and version of the test software used to exercise the device	Torrent App; Version: V1.58
Power settings used for the purpose of exercising the device	Tx setting -8
Firmware number of the EUT	N/A

4.2 EUT Operating Modes

The transmit mode for all tests was continuous. The continuous mode produces a Duty Cycle of at least 99%. The EUT was in its normal GFSK modulation during the tests. It was tested powered from its external power supply since that is the configuration in the final installation.



Testing of: Delsys, Inc., Model SP-W10, Trigno Centro Base Station

4.3 Special Accessories

No special accessories were used during the tests in order to achieve compliance.

4.4 Equipment Modifications

No modifications were made to the EUT at Radiometrics' test facility in order to comply with the standards listed in this report.

5.0 TEST SPECIFICATIONS

Document	Date	Title	
FCC CFR Title 47	2023	Code of Federal Regulations Title 47, Chapter 1, Federal Communications Commission, Part 15 - Radio Frequency Devices	
IC RSS-210 Issue 10	2019	Low Power Licence-Exempt Radiocommunication Devices (All Frequency Bands) Category I Equipment	
IC RSS-Gen Issue 5	2019	General Requirements and Information for the Certification of Radiocommunication Equipment (RSS-Gen)	

6.0 TEST PROCEDURE DOCUMENTS

The tests were performed using the procedures from the following specifications:

Document	Date	Title
ANSI C63.4-2014	2014	Methods of Measurement of Radio Noise Emissions from Low-Voltage Electrical and Electronic Equipment in the Range of 9 kHz to 40 GHz
ANSI C63.10-2013	2013	American National Standard for Testing Unlicensed Wireless Devices

7.0 RADIOMETRICS' TEST FACILITIES

The results of these tests were obtained at Radiometrics Midwest Corp. in Romeoville, Illinois, USA. Radiometrics is accredited by A2LA (American Association for Laboratory Accreditation) to conform to ISO/IEC 17025: 2017 "General Requirements for the Competence of Calibration and Testing Laboratories". Radiometrics' Lab Code is 121191 and Certification Number is 1495.01. Radiometrics' scope of accreditation includes all of the test methods listed herein. A copy of the accreditation can be accessed on our web site (www.radiomet.com). Radiometrics accreditation status can be verified at A2LA's web site (www.a2la2.org).

The following is a list of shielded enclosures located in Romeoville, Illinois used during the tests:

Chamber E: Is a custom-made anechoic chamber that measures 52' L X 30' W X 18' H. The walls and ceiling are fully lined with RF absorbers. Pro-shield of Collinsville, Oklahoma manufactured the chamber. The floor has a 9' x 9' section of microwave absorber for testing above 1 GHz.

Test Station F: Is an area that measures 10' D X 12' W X 10' H. The floor and back wall are metal shielded. This area is used for conducted emissions measurements.

A separate ten-foot long, brass plated, steel ground rod attached via a 6-inch copper braid grounds each of the above chambers. Each enclosure is also equipped with low-pass power line filters.

RP-9728B2 Rev. 1 Radiomet.com Page 5 of 19

R

Radiometrics Midwest Corporation

Testing of: Delsys, Inc., Model SP-W10, Trigno Centro Base Station

The FCC has accepted these sites as test site number US1065. The FCC test site Registration Number is 732175. Details of the site characteristics are on file with the Industry Canada as site number IC 3124A with a CAB ID of US0224.

A complete list of the test equipment is provided herein. The calibration due dates are indicated on the equipment list. The equipment is calibrated in accordance with ANSI/NCSL Z540-1 with traceability to the National Institute of Standards and Technology (NIST).

8.0 DEVIATIONS AND EXCLUSIONS FROM THE TEST SPECIFICATIONS

There were no deviations or exclusions from the test specifications.

9.0 CERTIFICATION

Radiometrics Midwest Corporation certifies that the data contained herein was taken under conditions that meet or exceed the requirements of the test specification and the data contained herein was taken with calibrated test equipment. The results relate only to the EUT listed herein.

10.0 TEST EQUIPMENT TABLE

					Frequency	Cal	
RMC ID	Manufacturer	Description	Model No.	Serial No.	Range	Period	Cal Date
AMP-05	RMC/Celeritek	Pre-amplifier	MW110G	1001	1.0-12GHz	12 Mo.	01/14/23
AMP-20	Avantek	Pre-amplifier	SF8-0652	15221	8-18GHz	12 Mo	06/02/23
AMP-59	Amplitech	Pre-amplifier	APTMP44	AMP-59	18-26 GHz	12 Mo.	01/10/23
ANT-48	RMC	Std. Gain Horn	HW2020	1001	18-26.5 GHz	24 Mo.	11/23/22
ANT-66	ETS-Lindgren	Horn Antenna	3115	62580	1.0-18GHz	24 Mo.	03/16/23
ANT-68	EMCO	Log-Periodic Ant.	93146	9604-4456	200-1000MHz	24 Mo.	02/07/22
ANT-80	AH Systems	Bicon Antenna	SAS-540	294	20-330MHz	24 Mo.	01/26/23
HPF-06	Mini-Circuits	High Pass Filter	VHF-3800+	31035	3-11 GHz	24 Mo.	05/23/22
LSN-01	Electrometrics	50 uH LISN	FCC/VDE 50/2	1001	0.01-30MHz	24 Mo.	09/21/23
REC-21	Agilent	Spectrum Analyzer	E7405A	MY45118341	9Hz-26.5 GHz	24 Mo.	02/24/22
REC-31	Agilent	Spectrum Analyzer	E7402A	US41160415	9kHz-3GHz	24 Mo.	06/23/23
REC-44	Agilent	Spectrum Analyzer	E4440A	US40420673	3Hz-26.5GHz	24 Mo.	03/31/22
THM-02	Fluke	Temp/Humid Meter	971	93490471	N/A	24 Mo.	11/22/22

Note: All calibrated equipment is subject to periodic checks.

Software Company	Test Software Name	Version	Applicable Tests
Radiometrics	EN550XX0	07.22.22	RF Conducted Emissions (FCC Part 15 & EN 55032)
Radiometrics	REREC11D	07.21.22	RF Radiated Emissions (FCC Part 15 & EN 55032)
Agilent	PSA/ESA-E/L/EMC	2.4.0.42	Bandwidth and screen shots

11.0 TEST SECTIONS

11.1 AC Conducted Emissions

The tests and limits are in accordance with FCC section 15.207 and RSS Gen section 8.8.



Testing of: Delsys, Inc., Model SP-W10, Trigno Centro Base Station

A computer-controlled analyzer was used to perform the conducted emissions measurements. The frequency range was divided into 500 subranges equally spaced on a logarithmic scale. The computer recorded the peak of each subrange. This data was then plotted on a semi-log graph generated by the computer. Adjusting the positions of the cables and orientation of the test system then maximizes the highest emissions.

Mains Conducted emission measurements were performed using a 50 Ohm/50 uH Line Impedance Stabilization Network (LISN) as the pick-up device. Measurements were repeated on both leads within the power cord. If the EUT power cord exceeded 80 cm in length, the excess length of the power cord was made into a 30 to 40 cm bundle near the center of the cord. The LISN was placed on the floor at the base of the test platform and electrically bonded to the ground plane.

FCC Limits of Conducted Emissions at the AC Mains Ports

Frequency Range	Class B Limits (dBuV)				
(MHz)	Quasi-Peak	Average			
0.150 - 0.50*	66 - 56	56 - 46			
0.5 - 5.0	56	46			
5.0 - 30	60	50			
* The limit decreases linearly with the logarithm of the frequency in this range					


* The limit decreases linearly with the logarithm of the frequency in this range.

The initial step in collecting conducted data is a peak detector scan and the plotting of the measurement range. Significant peaks are then marked as shown on the following table, and these signals are then measured with the quasi-peak detector. The following represents the worst case emissions from the EUT power supply, after testing all modes of operation. The Amplitude is the final corrected value with cable and LISN Loss.

Model	SP-W09	Specification	FCC part 15.207 & RSS-210
Serial Number	RMC-01	Test Date	09/05/2023
Test Personnel	Jeffery E Tomes	Test Location	Chamber F
Test Equipment	EMI Receiver (REC-20) (LSN-01)		


Judgement: Pass by at least 11 dB.

Testing of: Delsys, Inc., Model SP-W10, Trigno Centro Base Station



Frequency	QP	QP	Average	Average	Margin Under
(MHz)	Amp. (dBuV)	Limit (dBuV)	Amp. (dBuV)	Limit (dBuV)	Limit (dB)
0.158	55.1	65.6	37.8	55.6	10.5
0.189	50.2	64.1	32.7	54.1	13.9
0.206	47.9	63.4	31.1	53.4	15.5
0.419	38.9	57.5	27.7	47.5	18.6

Testing of: Delsys, Inc., Model SP-W10, Trigno Centro Base Station



Frequency (MHz)	QP Amp. (dBuV)	QP Limit (dBuV)	Average Amp. (dBuV)	Average Limit (dBuV)	Margin Under Limit (dB)
0.159	54.4	65.5	35.7	55.5	11.1
0.167	53.7	65.1	35.4	55.1	11.4
0.200	48.5	63.6	31.8	53.6	15.1
0.428	40.0	57.3	29.8	47.3	17.3

11.2 Radiated RF Emissions

Radiated emission measurements were performed with linearly polarized broadband antennas. The results obtained with these antennas can be correlated with results obtained with a tuned dipole antenna. The radiated emission measurements were performed with a spectrum analyzer. The bandwidth used from 150 kHz to 30 MHz is 9 or 10 kHz and the bandwidth from 30 MHz to 1000 MHz is 100 or 120 kHz. Above 1 GHz, a 1 MHz bandwidth is used. A 10 dB linearity check is performed prior to start of testing in order to determine if an overload condition exists. Figure 4 herein lists the details of the test equipment used during radiated emissions tests.

In addition, a high pass filter was used to reduce the fundamental emission. High pass filters were not needed above 10 GHz, since the preamplifiers attenuated the fundamental emission.

The EUT was rotated through three orthogonal axis as per 5.10.1 of ANSI C63.10 during the radiated tests.

RP-9728B2 Rev. 1 Radiomet.com Page 9 of 19



Testing of: Delsys, Inc., Model SP-W10, Trigno Centro Base Station

Final radiated emissions measurements were performed inside of an anechoic chamber at a test distance of 3 meters. The anechoic chamber is designated as Chamber E. This Chamber meets the Site Attenuation requirements of ANSI C63.4. Chamber E is located at 12 Devonwood Ave. Romeoville, Illinois EMI test lab.

The entire frequency range from 30 to 25000 MHz was slowly scanned. Measurements were performed using two antenna polarizations, (vertical and horizontal). The worst case emissions were recorded. All measurements may be performed using either the peak, average or quasi-peak detector functions. If the peak detector data exceeds or is marginally close to the limits, the measurements are repeated using a quasi-peak detector or average function as required by the specification for final determination of compliance. The QP and average detectors have a linear response.

The detected emission levels were maximized by rotating the EUT, adjusting the positions of all cables, and by scanning the measurement antenna from 1 to 4 meters above the ground.

Radiated Emissions Field Strength Limits

Frequency	Test Distance	Non-Fun	damental Limits
Range (MHz)	(meters)	uV/m	dB(uV/m)
30 - 88	3	100	40.0
88 - 216	3	150	43.5
216 - 960	3	200	46.0
Above 960	3	500	54.0

The emission limits shown in the above table are based on measurements using a CISPR quasi-peak detector below 1 GHz. Above 1 GHz, the radiated emission limits are based on measurements employing an average detector. Above 1 GHz peak emissions shall not be more than 20 dB above the average limits.

The fundamental limit is 94 dBuV/m for Average and 114 dBuV/m for peak detection.

11.2.1 Field Strength Calculation

The field strength is calculated by adding the Antenna Factor and Cable Loss, and by subtracting the Amplifier Gain from the measured reading. The basic equation is as follows:

Where: FS = Field Strength

RA = Receiver Amplitude

AF = Antenna Factor

CF = Cable Attenuation Factor

AG = Amplifier Gain

HPF = High pass Filter Loss

PKA = Peak to Average Factor (This is only used for average measurements above 1 GHz)

The Peak to average factor is used when average measurements are required. It is calculated by the highest duty cycle in percent over any 100mS transmission.

11.2.2 Duty Cycle

In accordance to 7.5 of ANSI C63.10 the following procedures were used.

a) The EUT was set to the "worst-case" pulse ON time.

W

Radiometrics Midwest Corporation

Testing of: Delsys, Inc., Model SP-W10, Trigno Centro Base Station

- b) The RF output was Coupled to the input of a spectrum analyzer by a "near-field" coupling method. The signal received shall be of sufficient level to trigger adequately the spectrum analyzer sweep display.
- c) The center frequency of the spectrum analyzer was set to the center of the RF signal.
- d) The spectrum analyzer was set for ZERO SPAN.
- e) The sweep time was of the analyzer was set to 100 ms and other times to show the duty cycle.
- f) Since the pulse train has a period that exceeds 100 ms, then:
 - 1) The trigger on the spectrum analyzer was set to capture the greatest amount of pulse "ON time" over 100 ms.
 - 2) The 100 ms period that contains the maximum "on time" was found.
 - 3) The duty cycle was determined by dividing the total maximum "ON time" by 100 ms (tON/100 ms).
- h) The duty cycle correction factor was used applying Equation (10) of ANSI C63.10 to the duty cycle determined in the preceding steps.

The width of each pulse is 0.18 mSeconds. There are, at most, 9 pulses per 100mS. This yields an effective duty cycle of 1.62%. The Peak to average factor is calculated by the highest duty cycle in percent over any 100mS transmission. The transmitter operates for a maximum duration of 1.62 ms in any 100 ms interval. 20 Log*(1.62mSec/100mSec) = -35.8 dB Peak to Average correction factor.

Since the difference between the peak and the average limits are 20 dB, there is no need to use a correction factor more than 20 dB. Therefore, a 20 dB factor was used.

11.2.3 Radiated Emissions Test Results

Test Date	September 14 and 15, 2023
EUT	SP-W10; Serial #RMC-001
Test Distance	3 Meters
Specification	FCC Part 15 Subpart C & RSS-210 Section B.10
Notes	Corr. Factors = cable loss - preamp gain
Abbreviations	Pol = Antenna Polarization; V = Vertical; H = Horizontal; P = peak; Q = QP
Configuration	The EUT is in the transmit mode with the receiver on

This table includes all emissions except Fundamental, Band edge and harmonics emissions.

	Meter			Ant	,	Dist			Margin	
Freq.	Reading		Ant.	Factor	Corr.	Fact	EUT	Limit	Under	
MHz	dBuV	Dect.	Pol.	dB/m	Factors	dB	dBuV/m	dBuV/m	Limit dB	Note
34.4	14.6	Р	Н	12.1	0.6	0.0	27.3	40.0	12.7	
46.8	12.4	Ρ	Н	9.7	0.7	0.0	22.8	40.0	17.2	
60.7	19.2	Ρ	Ι	8.8	0.8	0.0	28.8	40.0	11.2	
72.0	24.6	Р	I	8.9	0.9	0.0	34.4	40.0	5.6	
72.0	21.5	Q	Н	8.9	0.9	0.0	31.3	40.0	8.7	
84.0	21.0	Р	Н	9.2	1.0	0.0	31.2	40.0	8.8	
96.1	21.0	Р	Н	9.9	1.1	0.0	32.0	43.5	11.5	
108.1	20.1	Р	Н	10.7	1.1	0.0	31.9	43.5	11.6	
120.0	20.8	Р	Н	11.5	1.2	0.0	33.5	43.5	10.0	
143.9	29.6	Р	Н	12.5	1.3	0.0	43.4	43.5	0.1	
143.9	24.1	Q	I	12.5	1.3	0.0	37.9	43.5	5.6	
151.7	28.8	Р	I	12.6	1.3	0.0	42.7	43.5	8.0	
151.7	24.6	Q	Н	12.6	1.3	0.0	38.5	43.5	5.0	
168.0	18.6	Р	Н	13.0	1.4	0.0	33.0	43.5	10.5	
180.0	20.4	Р	Н	13.4	1.5	0.0	35.3	43.5	8.2	
180.0	17.3	Q	Н	13.4	1.5	0.0	32.2	43.5	11.3	
191.9	24.1	Р	Н	13.9	1.5	0.0	39.5	43.5	4.0	
191.9	18.7	Q	Н	13.9	1.5	0.0	34.1	43.5	9.4	
204.1	27.1	Р	Н	14.4	1.6	0.0	43.1	43.5	0.4	

RP-9728B2 Rev. 1 Radiomet.com Page 11 of 19



Testing of: Delsys, Inc., Model SP-W10, Trigno Centro Base Station

	Matan			Λ 4		D:-4			Manain	
Frog	Meter Reading		Ant.	Ant Factor	Corr.	Dist Fact	EUT	Limit	Margin Under	
Freq. MHz	dBuV	Dect.	Pol.	dB/m	Factors	dB	dBuV/m	dBuV/m	Limit dB	Note
204.1	24.5	Q	H	14.4	1.6	0.0	40.5	43.5	3.0	INOLE
216.0	20.7	P	H	14.7	1.6	0.0	37.0	43.5	6.5	
239.9	19.0	P	H	15.1	1.7	0.0	35.8	46.0	10.2	
252.0	21.2	P	H	12.0	1.7	0.0	34.9	46.0	11.1	
263.9	25.5	P	H	12.4	1.8	0.0	39.7	46.0	6.3	
276.0	16.2	P	H	12.9	1.8	0.0	30.9	46.0	15.1	
315.6	18.7	P	H	14.8	2.0	0.0	35.5	46.0	10.5	
331.7	16.8	P	H	14.2	2.0	0.0	33.0	46.0	13.0	
406.1	14.0	P	H	15.5	2.2	0.0	31.7	46.0	14.3	
547.0	9.8	P	H	18.0	2.6	0.0	30.4	46.0	15.6	
635.6	9.0	P	H	19.3	2.8	0.0	31.1	46.0	14.9	
770.8	10.9	P	H	21.0	3.1	0.0	35.0	46.0	11.0	
940.9	7.6	P	H	23.2	3.5	0.0	34.3	46.0	11.7	
1188.2	46.0	P	H	24.9	-33.7	0.0	37.2	74.0	36.8	1
1328.3	47.3	P	H	25.0	-33.7	0.0	38.6	74.0	35.4	1
1527.5	43.8	P	H	25.3	-33.6	0.0	35.5	74.0	38.5	1
1920.9	48.5	P	H	27.1	-33.4	0.0	42.2	74.0	31.8	1
2100.1	42.9	P	H	27.6	-33.3	0.0	37.2	74.0	36.8	1
2636.6	47.2	P	H	28.8	-32.7	0.0	43.3	74.0	30.7	1
2918.9	44.1	P	H	29.6	-32.4	0.0	41.3	74.0	32.7	1
3694.7	41.8	P	Н	32.0	-31.4	0.0	42.4	74.0	31.6	1
3896.9	41.1	P	Н	32.8	-31.3	0.0	42.6	74.0	31.4	1
4308.3	40.7	P	Н	32.5	-30.8	0.0	42.4	74.0	31.6	1
4500.5	41.0	P	Н	32.9	-30.7	0.0	43.2	74.0	30.8	1
30.2	22.0	P	V	13.5	0.6	0.0	36.1	40.0	3.9	
30.2	9.2	Q	V	13.5	0.6	0.0	23.3	40.0	16.7	
40.0	19.0	P	V	10.6	0.7	0.0	30.3	40.0	9.7	
43.5	28.5	Р	V	10.1	0.7	0.0	39.3	40.0	0.7	
43.9	15.5	Q	V	9.9	0.7	0.0	26.1	40.0	13.9	
46.1	22.5	P	V	9.9	0.7	0.0	33.1	40.0	6.9	
61.4	22.2	Р	V	8.9	0.8	0.0	31.9	40.0	8.1	
68.9	28.1	Р	V	8.8	0.9	0.0	37.8	40.0	2.2	
68.9	15.7	Q	V	8.8	0.9	0.0	25.4	40.0	14.6	
72.0	29.1	Р	V	8.9	0.9	0.0	38.9	40.0	1.1	
72.0	26.1	Q	V	8.9	0.9	0.0	35.9	40.0	4.1	
84.0	29.0	Р	V	9.2	1.0	0.0	39.2	40.0	0.8	
84.0	20.2	Q	V	9.2	1.0	0.0	30.4	40.0	9.6	
94.2	26.9	Р	V	9.8	1.0	0.0	37.7	43.5	5.8	
95.9	21.2	Q	V	9.9	1.1	0.0	32.2	43.5	11.3	
110.0	28.5	Р	V	10.9	1.1	0.0	40.5	43.5	3.0	
111.0	10.9	Q	V	10.9	1.1	0.0	22.9	43.5	20.6	
120.0	20.9	Р	V	11.5	1.2	0.0	33.6	43.5	9.9	
144.4	28.4	Р	V	12.5	1.3	0.0	42.2	43.5	1.3	
144.4	20.1	Q	V	12.5	1.3	0.0	33.9	43.5	9.6	
180.0	20.4	Р	V	13.4	1.5	0.0	35.3	43.5	8.2	
180.0	15.1	Q	V	13.4	1.5	0.0	30.0	43.5	13.5	
191.9	20.9	Р	V	13.9	1.5	0.0	36.3	43.5	7.2	
191.9	19.7	Q	V	13.9	1.5	0.0	35.1	43.5	8.4	
216.0	17.4	Р	V	14.7	1.6	0.0	33.7	43.5	9.8	
238.2	14.7	Р	V	15.0	1.7	0.0	31.4	46.0	14.6	
252.0	14.8	Р	V	12.0	1.7	0.0	28.5	46.0	17.5	
263.9	15.2	Р	V	12.4	1.8	0.0	29.4	46.0	16.6	
300.0	15.0	Р	V	14.0	1.9	0.0	30.9	46.0	15.1	
315.1	15.1	Р	V	14.9	2.0	0.0	32.0	46.0	14.0	



Testing of: Delsys, Inc., Model SP-W10, Trigno Centro Base Station

	Meter			Ant		Dist			Margin	
Freq.	Reading		Ant.	Factor	Corr.	Fact	EUT	Limit	Under	
MHz	dBuV	Dect.	Pol.	dB/m	Factors	dB	dBuV/m	dBuV/m	Limit dB	Note
337.5	12.8	Р	V	14.3	2.0	0.0	29.1	46.0	16.9	
348.6	13.1	Р	V	14.4	2.0	0.0	29.5	46.0	16.5	
369.3	14.2	Р	V	14.5	2.1	0.0	30.8	46.0	15.2	
396.1	15.6	Р	V	15.3	2.2	0.0	33.1	46.0	12.9	
470.2	13.7	Р	V	17.1	2.4	0.0	33.2	46.0	12.8	
580.6	9.2	Р	V	18.6	2.7	0.0	30.5	46.0	15.5	
665.2	7.6	Р	V	20.5	2.9	0.0	31.0	46.0	15.0	
753.3	9.1	Р	V	20.9	3.1	0.0	33.1	46.0	12.9	
843.8	8.6	Р	V	22.1	3.3	0.0	34.0	46.0	12.0	
939.9	8.0	Р	V	23.2	3.5	0.0	34.7	46.0	11.3	
1175.2	46.4	Р	V	24.8	-33.7	0.0	37.5	74.0	36.5	1
1328.3	48.6	Р	V	25.0	-33.7	0.0	39.9	74.0	34.1	1
1692.7	48.9	Р	V	26.2	-33.5	0.0	41.6	74.0	32.4	1
1719.7	50.1	Р	V	26.4	-33.5	0.0	43.0	74.0	31.0	1
2111.1	44.8	Р	V	27.6	-33.3	0.0	39.1	74.0	34.9	1
2807.8	45.0	Р	V	29.2	-32.5	0.0	41.7	74.0	32.3	1
2904.9	44.0	Р	V	29.6	-32.4	0.0	41.2	74.0	32.8	1
3695.7	40.7	Р	V	32.0	-31.4	0.0	41.3	74.0	32.7	1
4145.1	41.4	Р	V	32.4	-30.9	0.0	42.9	74.0	31.1	1
4652.7	40.7	Р	V	33.2	-30.5	0.0	43.4	74.0	30.6	1
4877.9	41.0	Р	V	33.5	-29.9	0.0	44.6	74.0	29.4	1

Judgment: Passed by 2.4 dB

Testing of: Delsys, Inc., Model SP-W10, Trigno Centro Base Station

Fundammental and Harmonic Emissions FCC 15.249

		Sp		Analyz dings	er		EUT	Peak	Ave	Peak	Ave	Margin
hrm	Tx					Corr.	Emission	Tot.	FS	Lin	nit	Under
		Vert	ical	Horiz	ontal		Freq					
#	Freq	Peak	Ave	Peak	Ave	Fact.	MHz	dBu	V/m	dBu\	//m	Limit
1	2402	78.7	58.7	70.1	50.1	30.3	2402.0	109.0	89.0	114	94	5.0
BE	2402	64.5	46.0	58.4	39.8	-5.3	2390.0	59.2	40.7	74	54	13.3
2	2402	60.9	40.9	55.5	35.5	2.4	4804.0	63.3	43.3	74	54	10.7
3	2402	40.8	20.8	40.1	20.1	7.8	7206.0	48.6	28.6	74	54	25.4
4	2402	39.0	19.0	41.9	21.9	12.0	9608.0	53.9	33.9	74	54	20.1
1	2440	79.8	59.8	70.2	50.2	30.5	2440.0	110.3	90.3	114	94	3.7
2	2440	56.1	36.1	54.7	34.7	2.7	4880.0	58.8	38.8	74	54	15.2
3	2440	49.9	29.9	48.5	28.5	8.2	7320.0	58.1	38.1	74	54	15.9
4	2440	47.8	27.8	48.0	28.0	12.7	9760.0	60.7	40.7	74	54	13.3
1	2480	80.4	60.4	72.8	52.8	30.6	2480.0	111.0	91.0	114	94	3.0
BE	2480	63.1	44.6	53.6	35.4	-5.0	2483.5	58.1	39.6	74	54	14.4
2	2480	55.7	35.7	58.1	38.1	2.7	4960.0	60.8	40.8	74	54	13.2
3	2480	48.8	28.8	49.2	29.2	8.3	7440.0	57.5	37.5	74	54	16.5
4	2480	47.9	27.9	47.5	27.5	13.3	9920.0	61.2	41.2	74	54	12.8
			(Column	numbe	rs (see	below for ex	kplanatic	ns)			
1	2	3	4	5	6	7	8	9	10	11	12	13

Column #1. hrm = Harmonic; BE = Band Edge emissions

Column #2. Frequency of Transmitter.

Column #3. Uncorrected readings from the spectrum analyzer with First Axis Rotation.

Column #4. Average Reading based on peak reading reduced by the Duty cycle correction

Column #5. Uncorrected readings from the spectrum analyzer with First Axis Rotation.

Column #6. Average Reading based on peak reading reduced by the Duty cycle correction

Column #7. Corr. Factors = Cable Loss - Preamp Gain + Antenna Factor

Column #8. Frequency of Tested Emission

Column #9. Highest peak field strength at listed frequency.

Column #10. Highest Average field strength at listed frequency.

Column #11. Peak Limit.

Column #12. Average Limit.

Column #13. The margin (last column) is the worst case margin under the peak or average limits for that row.


Overall Judgment: Passed by at least 10 dB

No other Emissions were detected from 30 to 25,000 MHz within 10 dB of the limits.

Testing of: Delsys, Inc., Model SP-W10, Trigno Centro Base Station

Figure 1. Drawings of Radiated Emissions Setup

Chamber E, anechoic



Frequency Range	Receive Antenna	Pre- Amplifier	Spectrum Analyzer	High Pass Filter
30 to 200 MHz	ANT-80	None	REC-21	None*
200 to 1000 MHz	ANT-68	None	REC-21	None*
1 to 10 GHz	ANT-66	AMP-05	REC-21	HPF-06
10 to 18 GHz	ANT-66	AMP-20	REC-21	None*
18-26 GHz	ANT-48	AMP-59	REC-21	None*

• AC cords not shown. They are connected to AC


outlet with low-pass filter on turntable

RP-9728B2 Rev. 1 Radiomet.com Page 15 of 19


^{*} A high pass filter was not needed since the fundamental frequency was outside of the amplifiers pass band.

Testing of: Delsys, Inc., Model SP-W10, Trigno Centro Base Station

Radiated Emissions Test Setup for Frequencies from 30MHz to 1000MHz (Side View)



Radiated Emissions Test Setup for Frequencies over 1000MHz (Side View)



RP-9728B2 Rev. 1 Radiomet.com Page 16 of 19

Testing of: Delsys, Inc., Model SP-W10, Trigno Centro Base Station

11.3 Occupied Bandwidth Data


The 99% OBW and the 20 dB bandwidth of the RF output were measured automatically using an Agilent Model: E4440A Spectrum Analyzer. The spectrum analyzer was set to MAX HOLD mode to record the worst case of the modulation. The plots of the occupied bandwidth for the EUT are supplied on the following pages.

The 20 dB bandwith and the 99% OBW is fully within the allowed 2400 to 2483.5 MHz authourized band.

Channel	99% EBW MHz	20 dB OBW MHz
2402	0.9253	1.0280
2440	0.9404	0.9864
2480	0.9287	0.9824


Test Date	September 13, 2023
EUT	SP-W10; Serial #RMC001


Figure 2. Occupied Bandwidth Plot



RP-9728B2 Rev. 1 Radiomet.com Page 17 of 19

Testing of: Delsys, Inc., Model SP-W10, Trigno Centro Base Station







Testing of: Delsys, Inc., Model SP-W10, Trigno Centro Base Station

11.3.1 Measurement Instrumentation Uncertainty

Measurement	Uncertainty
Radiated Emissions, E-field, 3 meters, 30 to 200 MHz	4.8 dB
Radiated Emissions, E-field, 3 meters, 200 to 1000 MHz	4.6 dB
Radiated Emissions, E-field, 3 meters, 1 to 6 GHz	5.0 dB
Radiated Emissions, E-field, 3 meters, 6 to 18 GHz	5.5 dB
Radiated Emissions, E-field, 3 meters, 18 to 26 GHz	5.9 dB
Bandwidth using marker delta method at a span of 10 MHz	4 kHz
99% Occupied Bandwidth using REC-44	1% of frequency span
Temperature THM-02	0.6 Deg C

The uncertainties represent expanded uncertainties expressed at approximately the 95% confidence level using a coverage factor of k=2 in accordance with CISPR 16-4-2.

12.0 REVISION HISTORY

Document RP-9728B2 Revisions:			
Rev.	Affected Sections	Description	Rationale
1	Cover page	Corrected FCC & IC numbers	Typographical error

RP-9728B2 Rev. 1 Radiomet.com Page 19 of 19