



CARL ZEISS MEDITEC

Federal Communication Commission
Equipment Authorization Division, Application Processing Branch
7435 Oakland Mills Road
Columbia, MD 21048

01/07/09

TO WHOM IT MAY CONCERN

Modular Approval Request

The wireless modules (FCP WL, FCP Gateway WL, FCP Interface WL) consist of a specific basic board carrying a general transmission module based on Mitsumi BT module WML-C46.

1. The WML-C46 has its own RF shielding.
2. All modules have a digital input buffer.
3. All modules have their own power supply regulation.
4. The modules "FCP WL" and "FCP Interface WL" have been tested with an integrated on board antenna module "FCP Gateway WL" has been tested with an external antenna connected via a unique antenna plug on the transmission module.
5. The modules have been tested in a stand-alone configuration.
6. The modules will be labelled with its own FCC ID number.
7. The modules comply with the with all applicable FCC rules. Instructions for maintaining compliance are given in the user instructions.
8. The modules comply with any applicable RF exposure requirements.

If you have any questions, please feel free to contact us at the address shown below

Sincerely,

Name: Jürgen Kasprowitsch
Company: Carl Zeiss Surgical GmbH
Address: Carl-Zeiss-Straße 22
D-73447 Oberkochen, Germany

Phone: +49 7364 20 2555
Fax: +49 7364 20 2035
Email: Kasprowitsch@meditec.zeiss.com