

TCB Comments:

- 1) In the SAR test report, it appears that the calibration certificates may be out of date.
 - a) Annex E, Probe Calibration Certificate, states that the certificate was issued on 7 September 2007. The test was performed on 1 September 2008, therefore this certificate appears to be valid. However, the part of the certificate on page 106 of the test report states that further calibration should be scheduled for January, March, April and August of 2008. Please can you confirm that the calibration of the probe is valid?
 - b) Annex F, Dipole Calibration Certificate, states that calibration was performed on 20 February 2008. Please confirm the re-calibration date.

Agent responses:

- a) We think the calibration of the probe is valid. Because the probe was calibrated on 7 September 2007 and the valid period is one year, then the expired date is 6 September 2008. The test was performed on 1 September 2008, and the report was issued on 3 September 2008, so the probe is still in the valid period. For the question of "the part of the certificate on page 106 of the test report states that further calibration should be scheduled for January, March, April and August of 2008", that is the recalibration schedule of SPEAG, not for us, and all the instrument of SPEAG in the probe calibration certificate are all in the valid period, because the calibration was performed on 7 September 2007 and the SPEAG's instrument were all calibrated on January, March, April and August of 2007, and the re-calibration date of those instrument is January, March, April and August of 2008.
- b) In Annex F the Dipole Calibration Certificate, states that calibration was performed on 20 February 2007, and on page 17 in the table 12, it states the valid period of the dipole is two years, so the re-calibration date will be 19 February 2009.