

Declaration of equivalence

Issue date: 20 April 2012

Product name: COCHLEAR NUCLEUS 6 SOUND PROCESSOR

Product ID/ models numbers: CP910, DP910, CP920 and DP920

Responsible party
/ trade name: Cochlear Limited

Address: 1 University Avenue
Macquarie University NSW 2109
Australia

Contact person: Bronwyn Evans

Email: bevans@cochlear.com

Phone number: +61 294286515

Test standards: 47 CFR Part 15, RSS-210 Issue 8

Registration numbers: FCC ID: WTOP900
IC: 8039A-P900

We, Cochlear Limited, declare that

The devices bearing the trade name and models specified above have an identical electrical bill of materials except for the CP910 and DP910 devices which do not include the accessory connector. The implant stimulation firmware may differ between CP910, CP920, DP910 and DP920 devices.



Bronwyn Evans,
Senior Vice President Global Quality, Clinical and Regulatory