



⚡ WARNING ⚡
RISK OF ELECTRIC SHOCK

- TO REDUCE THE RISK OF ELECTRIC SHOCK, DO NOT REMOVE FRONT OR BACK COVER.
- READ THE MANUAL AND SAFETY PRECAUTIONS BEFORE PERFORMING INSTALLATION.
- MAKE SURE THE POWER CORD IS PLUGGED INTO THE SYSTEM BEFORE CONNECTING THE POWER CORD TO AN AC POWER OUTLET.
- DISCONNECT POWER FROM THE SYSTEM BY UNPLUGGING THE POWER CORD FROM AC OUTLET.
- THE AC POWER OUTLET SHOULD BE COMPATIBLE WITH THE 3-PIN POWER CORD CONNECTOR.

1. Plug the power cord into your system.
2. Connect AC power cord to the power adapter.
3. Plug the power cord into an easily accessible AC power outlet.

Model Name: AID-240SBW
Input Rating: 100-240Vac, 2.0-1.0A, 50-60Hz

For use only with power adapter:
 Mfr.: EDAC POWER ELECTRONICS CO LTD
 Model: EM11013CWWWW
 Input: 100-240Vac, 50-60 Hz, 2.0-1.0A
 Output: 19Vdc, 4.73A, Max. 90W

Manufacturer : Avalue Technology Inc. TEL: +886-2-8226-2345
 7F, No.228, Lian Cheng Road, Zhonghe District, New Taipei City, 235, Taiwan

CLASSIFIED
UL US
 6344074

MEDICAL - GENERAL MEDICAL EQUIPMENT AS TO ELECTRICAL SHOCK, FIRE AND MECHANICAL HAZARDS ONLY IN ACCORDANCE WITH ANSI/AAMI ES 60601-1 (2005) + AMD (2012) and CAN/CSA-C22.2 No. 6060-1 (2008) + (2014)

This device complies with Part 15 of the FCC rules. Operation is subject to the following conditions: (1) this device may not cause harmful interference and (2) this device must accept any interference received, including interference that may cause undesired operation

AID-240SBW Medical Computers
FCC ID: XBG-AID240SBWCT IC: 21285-CTNFC

Contains:
FCC ID: TX2-RTL8821AE IC: 6317A-RTL8821AE

MADE IN TAIWAN

Manufacture Date: 2016-05
Rev.1.30 30ACH
P/N: AID-240SBW01-D02R

S/N: 0000AF00278