



CERTIFICATE OF COMPLIANCE



Report No.: NEI-FCCE-1-1210045


Issued Date: Dec. 05, 2012

Authorized under Declaration of Conformity according to 47 CFR, Part 2 and Part 15 of the FCC Rules. The product listed in follows was (were) tested in the Neutron EMC Laboratory to comply with the criteria limits Class B of conducted and radiated emissions of the Technical Standards FCC Part 15, Subpart B: 2010, established by the FCC, USA.

Equipment MEDIVista 18.5" Bedside Terminal
Model Name LS1852
Brand Name Lincor
Applicant Lincor Solutions Ltd.
Address Cork Business and Technology Park, Model Farm Road, Cork, Ireland

Measurement ANSI C63.4: 2009
Method ICES-003 Issue 5: 2012
CISPR 22: 2008
CAN/CSA-CISPR 22-10

The test data, data evaluation, and equipment configuration contained in our test report (Report No. NEI-FCCE-1-1210045) were obtained utilizing the test procedures, test instruments, test sites that has been accredited by the Authority of NVLAP and TAF according to the ISO-17025 quality assessment standard and technical standard(s). The test data contained in the referenced test report relate only to the EUT sample and item(s) tested.


Andy Chiu
Authorized Signatory

Neutron Engineering Inc.
B1, No. 37, Lane 365, YangGuang St.,
NeiHu District 114, Taipei, Taiwan.
TEL: +886-2-2657-3299
FAX: +886-2-2657-3331

