



## EMC Test Data

Client:	Nevro Corporation	Job Number:	JD103637
Model:	Senza Implantable Pulse Generator, Model IPG2000	T-Log Number:	T103654
		Project Manager:	Christine Krebill
Contact:	Jon Parker	Project Coordinator:	-
Standard:	FCC Part 95, EN 301 839 V2.1.1	Class:	2

### SAR Exclusion

#### Test Specific Details

Objective: Evaluate the RF Exposure requirements per FCC 1.1310, 2.1091 and RSS-102.

Date: 9/27/2017

Engineer: David W. Bare

#### Deviations From The Standard

No deviations were made from the requirements of the standard.

#### FCC SAR Exclusion

The EIRP power of the device is 122uW which is less than 1 mW EIRP required for SAR exclusion.