

NEUROMetrix

February 9, 2010

Federal Communications Commission
Authorization and Evaluation Division
Washington, DC 20554

RE: FCC ID: XUL-RF-SYNC

To whom it may concern:

This is a submittal for Limited Modular Approval of the module NeuroMetrix RF-SYNC, Model NeuroMetrix Bioamplifier.

This module incorporates the same radio technology and implementation as that used in NeuroMetrix Stimulator. This module differs in that it is modular, intended to be incorporated into NeuroMetrix Bioamplifier, thereby providing wireless communication to existing NeuroMetrix Stimulator. The module complies with FCC Rules as a self-contained entity; it does not rely on external mounting or shielding or other input/output mechanisms to achieve compliance. The device was designed in accordance with FCC Public Notice DA 00-1407, 26-Jun-2000, "Unlicensed Modular Transmitter Approval". The module is not being offered for sale as an OEM component. It is for incorporation into NeuroMetrix products only.

The module has been tested with onboard chip antennas which are permanently attached. It has no antenna plugs.

The module will be labeled with its own FCC ID number.

The module complies with all applicable FCC rules. Instructions for maintaining compliance are provided in documents accompanying the host devices in which the module will be incorporated (Instructions for Use or Service Guide).

The module complies with applicable SAR and RF exposure requirements.

Thank you in advance for your consideration.

Sincerely,



Michael Williams
Senior Vice President of Engineering
NeuroMetrix, Inc.