

## **Submit 3 Dipole Calibration Report**



SIMT

CNAS L0134

SHANGHAI INSTITUTE OF MEASUREMENT AND TESTING TECHNOLOGY  
NATIONAL CENTER OF MEASUREMENT AND TEST FOR EAST CHINA校准证书编号: 2008J10-10-812002  
Calibrated certificate series No.

## CALIBRATION CERTIFICATE

上海市计量测试技术研究院  
华东国家计量测试中心

## 校 准 证 书

委托者 Customer	程智科技股份(昆山)有限公司 Compliance Certification Services Inc.
委托者地址 Address of customer	江苏省昆山市(留学创业园)伟业路10号 No. 10, Wei-Ye Rd., Innovation park, Eco & Tec, Development Zone, Kun Shan City, Jiang Su, P. R. O. C.
器具名称 Name of instrument	偶极子天线 DIPOLE ANTENNA
制造厂 Manufacturer	ANTENNESSA 公司
型号/规格 Model/Specification	DIPOLE 900MHz
器具编号 No. of instrument	SN 48/05 DIPD33
器具准确度 Instrument accuracy	/

(机构校准专用章)

证书批准人 Approved by	
核 验 员 Checked by	
校 准 员 Calibrated by	

校准日期 2008 年 12 月 10 日  
Date for calibrated Year Month Day地址: 上海市张衡路1500号(总部) 电话: 021-38839800 传真: 021-50798390 邮编: 201203  
Address: No.1500 Zhangheng Road, Shanghai(headquarters) Tel.: Fax.: Post Code: 201203 Tel. for complaint上海市宜山路716号(分部) 电话: 021-64701390 传真: 021-64701810 邮编: 200233  
Address: No. 716 Yishan Road, Shanghai(branch) Tel.: Fax.: Post Code: 200233

投诉电话: 021-50798262

**SIMT**

SHANGHAI INSTITUTE OF MEASUREMENT AND TESTING TECHNOLOGY  
NATIONAL CENTER OF MEASUREMENT AND TEST FOR EAST CHINA

校准证书编号: 2008J10-10-812002  
Calibrated certificate series No.

国家法定计量检定机构计量授权证书号(中心/院): (国)法计(2002)01039号/ (2002)01019号  
The number of the Certificate of Metrological Authorization to The Legal Metrological Verification Institution is No. (2002) 01039 / No. (2002) 01019

中国合格评定国家认可委员会实验室认可证书号: No. CNAS L0134  
The number of the certificate accredited by CNAS is No. L0134

本次校准所依据的技术规范(代号、名称):  
Reference documents for the calibration (code, name)

JCJ/J101002.1/0-2007 SAR偶极子天线校准规范

IEEE Std 1528-2003 "IEEE Recommended Practice for Determining the Peak

Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head  
from Wireless Communications Devices: Measure Techniques"

IEC 62209-1: 2005 Procedure to measure the Specific Absorption Rate (SAR) in the  
frequency range of 300 MHz to 3 GHz Part 1: hand-held mobile wireless  
communication devices

本次校准所使用的主要计量标准器具:

Main measurement standards used in this calibration

名称/型号 Name/Model	编号 Number	证书编号/有效期限 Certificate No./Due date	测量范围/准确度 Measuring range/accuracy
VECTOR NETWORK ANALYZER ZVB 8	容-027-27	2008F31-10-001907 2009.06.26	300 kHz~8 GHz, Frequency resolution: 100 μHz, Measurement time: < 8 ms, Measurement bandwidths: 1 Hz~500 kHz

以上计量标准器具的量值溯源至国家基准。

Quantity values of above measurement standards used in this calibration are traced to those of the national primary standards in the P.R. China.

校准地点及环境条件:

Location and environmental condition for the calibration

地点: 宜山路 716 号 (No. 716 Yishan Road)  
Location:

温度: 23 °C; 湿度: 49 %RH; 其它: /  
Ambient temperature: 23 °C; Relative humidity: 49 %RH; Others: /

本次校准结果的扩展不确定度:

Expanded uncertainty

+3dB 至 -15dB:  $U = 0.8 \text{ dB}$  ( $k=2$ )

-15dB 至 -25dB:  $U = 1.2 \text{ dB}$  ( $k=2$ )

-25dB 至 -35dB:  $U = 3.1 \text{ dB}$  ( $k=2$ )

校准结果/说明:

Results of calibration and additional explanation

Pass

The requirements of the calibration criterion: return Loss must be less than -20dB

本证书提供的结果仅对本次被校的器具有效。

The data are valid only for the instrument(s).

校准证书续页专用

Continued page of calibration certificate

第 2 页 共 5 页  
Page of total pages

**SIMT**

CNAS L0134

SHANGHAI INSTITUTE OF MEASUREMENT AND TESTING TECHNOLOGY  
NATIONAL CENTER OF MEASUREMENT AND TEST FOR EAST CHINA

校准证书编号: 2008J10-10-812002  
Calibrated certificate series No.

### 校准结果/说明 (续页) :

Results of calibration and additional explanation (continued page)

#### 1. Calibration procedure:

Return Loss is measured with the dipole mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis. During calibration, the flat phantom is filled with the liquid whose parameters are calibrated relative to different frequency.

#### 2. Calibration Conditions:

##### A. The spacer from Dipole center to TSL

Distance Dipole Center - TSL	Frequency
15mm±0.2mm with spacer	900MHz

##### B. Head TSL parameters

The following parameters and calculation were applied.

Head TSL temperature change is well controlled to be within  $22\pm0.2^{\circ}\text{C}$  during test.

Frequency	Nominal Head TSL Parameters (Permittivity/ Conductivity)	Measurement Head TSL parameters (Permittivity/ Conductivity)
900 MHz	41.50/0.97	41.71/1.00

##### C. Body TSL parameters

The following parameters and calculation were applied.

Body TSL temperature change is well controlled to be within  $22\pm0.2^{\circ}\text{C}$  during test.

Frequency	Nominal Body TSL Parameters (Permittivity/ Conductivity)	Measurement Body TSL parameters (Permittivity/ Conductivity)
900 MHz	55.00/1.05	54.62/1.04


#### 3. Measurement Results

Frequency	Return Loss with Head TSL	Return Loss with Body TSL
900 MHz	-25.06 dB	-24.23 dB


## 校准结果/说明 (续页) :

Results of calibration and additional explanation (continued page)

## Return Loss Measurement Plot for head TSL



## Return Loss Measurement Plot for Body TSL



Remark: Attachment 1:SAR validation &amp; Test equipment

End

**SIMT**

CNAS L0134

SHANGHAI INSTITUTE OF MEASUREMENT AND TESTING TECHNOLOGY  
NATIONAL CENTER OF MEASUREMENT AND TEST FOR EAST CHINA校准证书编号: 2008J10-10-812002  
Calibrated certificate series No.

## Attachment 1: SAR validation &amp; Test equipment

Validation	Condition	SAR Value (W/kg)	
		1g	10g
SAR measured with Head TSL	1W (input power)	11.11	7.27
SAR measured with Body TSL	1W (input power)	10.98	7.29

名称/型号 Name/Model	编号 Number	证书编号/有效期限 Certificate No./Due date	测量范围/准确度 Measuring range/accuracy
6 axis Robot KR3	容-027-01	/	6 axes, Repeatability: $\pm$ 0.05 mm, Nominal payload: 3 kg
Vector Network Analyzer ZVB 8	容-027-27	2008F31-10-001907 2009.06.26	300 kHz to 8 GHz, Frequency resolution: 100 $\mu$ Hz, Measurement time: < 8 ms, Measurement bandwidths: 1 Hz to 500 kHz
Signal Generator SMT 06	容-027-15	2008F33-10-001469 2009.06.26	5 kHz - 6 GHz, Resolution: 0.1Hz, -144 to + 13 dBm, Max. RF power: 1W, Max. DC voltage: 0V / Level > -127 dBm: f < 1.5 GHz: < 1dB; F > 1.5 GHz: < 1.5dB; f > 3GHz: < 2dB
Power Meter NRVD	容-027-16	2008F31-10-001906 2009.06.24	100 kHz to 6 GHz, 10nW to 500mW
Millivoltmeter 2000	容-027-26	2008F11-10-001004 2009.06.19	Measurement range: 100.0000 mV ~ 1000.000V Sensitivity: 0.1 $\mu$ V ~ 1 m V.
Power Amplifier BLMA 0820-6	容-027-18	2008F33-10-001467 2009.06.26	0.8 - 2 GHz; Output: 6W; Gain: min 37.8 / typ 40, $\pm$ 2 dB; Harmonics: 2nd: 20dBc, 3rd: 20dBc; Line power: 125 W.
Isotropic E-Field Probe E-FIELD PROBE	容-027-54	2008J10-10-801001 2008.12.25	Dipole resistance (in the connector plane): 1M $\Omega$ to 2M Axial isotropy in human-equivalent liquids: < 0.25 dB Hemispherical Isotropy in human-equivalent liquids < 0.5 dB, Linearity < 0.5 dB, Lower SAR detection threshold: 0.0015 Watts/kg
SAM Phantom	容-027-22	/	/



SIMT

SHANGHAI INSTITUTE OF MEASUREMENT AND TESTING TECHNOLOGY  
NATIONAL CENTER OF MEASUREMENT AND TEST FOR EAST CHINA

校准证书编号: 2008J10-10-812003  
Calibrated certificate series No.

## CALIBRATION CERTIFICATE

上海市计量测试技术研究院  
华东国家计量测试中心

## 校 准 证 书

委托者 程智科技股份(昆山)有限公司  
Customer Compliance Certification Services Inc.

委托者地址 江苏省昆山市(留学创业园)伟业路10号  
Address of customer No. 10, Wei-Ye Rd., Innovation park, Eco & Tec, Development Zone, Kun Shan City, Jiang Su, P. R. O. C.

器具名称 偶极子天线  
Name of instrument DIPOLE ANTENNA

制造厂 ANTENNESSA 公司  
Manufacturer

型号/规格 DIPOLE 1800MHz  
Model/Specification

器具编号 SN 48/05 DIPF34  
No. of instrument

器具准确度 /  
Instrument accuracy

证书批准人  
Approved by

(机构校准专用章)

核 验 员 刘 鹏  
Checked by

校 准 员 高 龙  
Calibrated by

校准日期 2008 年 12 月 10 日  
Date for calibrated Year Month Day

投诉电话: 021-50798262

地址: 上海市张衡路1500号(总部) 电话: 021-38839800 传真: 021-50798390 邮编: 201203  
Address: No.1500 Zhangheng Road, Shanghai(headquarters) Tel.: Fax.: Post Code: 201203 Tel. for complaint

上海市宜山路716号(分部) 电话: 021-64701390 传真: 021-64701810 邮编: 200233  
No. 716 Yishan Road, Shanghai(branch) Tel.: Fax.: Post Code: 200233

未经本院批准, 部分采用本证书内容无效。

Partly using this certificate will not be admitted unless allowed by SIMT.

国家法定计量检定机构计量授权证书号 (中心/院) : (国) 法计 (2002) 01039 号 / (2002) 01019 号  
 The number of the Certificate of Metrological Authorization to The Legal Metrological Verification Institution is No. (2002) 01039 / No. (2002) 01019

中国合格评定国家认可委员会实验室认可证书号: No. CNAS L0134  
 The number of the certificate accredited by CNAS is No. L0134

本次校准所依据的技术规范 (代号、名称) :

Reference documents for the calibration (code, name)

JCJ/J101002.1/0-2007 SAR 偶极子天线校准规范

IEEE Std 1528-2003 "IEEE Recommended Practice for Determining the Peak

Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head  
 from Wireless Communications Devices: Measure Techniques"

IEC 62209-1: 2005 Procedure to measure the Specific Absorption Rate (SAR) in the  
 frequency range of 300 MHz to 3 GHz Part 1: hand-held mobile wireless  
 communication devices

本次校准所使用的主要计量标准器具:

Main measurement standards used in this calibration

名称/型号 Name/Model	编号 Number	证书编号/有效期限 Certificate No./Due date	测量范围/准确度 Measuring range/accuracy
VECTOR NETWORK ANALYZER ZVB 8	容-027-27	2008F31-10-001907 2009.06.26	300 kHz~8 GHz, Frequency resolution: 100 μHz, Measurement time: < 8 ms, Measurement bandwidths: 1 Hz~500 kHz

以上计量标准器具的量值溯源至国家基准。

Quantity values of above measurement standards used in this calibration are traced to those of the national primary standards in the P.R. China.

校准地点及环境条件:

Location and environmental condition for the calibration

地点: 宜山路 716 号 (No. 716 Yishan Road)  
 Location:

温度: 23 °C; 湿度: 49 %RH; 其它: /  
 Ambient temperature; Relative humidity; Others

本次校准结果的扩展不确定度:

Expanded uncertainty

+3dB 至 -15dB:  $U = 0.8 \text{ dB}$  ( $k=2$ )  
 -15dB 至 -25dB:  $U = 1.2 \text{ dB}$  ( $k=2$ )  
 -25dB 至 -35dB:  $U = 3.1 \text{ dB}$  ( $k=2$ )

校准结果/说明:

Results of calibration and additional explanation

Pass

The requirements of the calibration criterion: return Loss must be less than -20dB

本证书提供的结果仅对本次被校的器具有效。

The data are valid only for the instrument(s).

## 校准结果/说明 (续页) :

Results of calibration and additional explanation (continued page)

## 1. Calibration procedure:

Return Loss is measured with the dipole mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis. During calibration, the flat phantom is filled with the liquid whose parameters are calibrated relative to different frequency.

## 2. Calibration Conditions:

## A. The spacer from Dipole center to TSL

Distance Dipole Center - TSL	Frequency
10mm±0.2mm with spacer	1800MHz

## B. Head TSL parameters

The following parameters and calculation were applied.

Head TSL temperature change is well controlled to be within  $22\pm0.2^{\circ}\text{C}$  during test.

Frequency	Nominal Head TSL Parameters (Permittivity/ Conductivity)	Measurement Head TSL parameters (Permittivity/ Conductivity)
1800 MHz	40.00/1.40	39.40/1.37

## C. Body TSL parameters

The following parameters and calculation were applied.

Body TSL temperature change is well controlled to be within  $22\pm0.2^{\circ}\text{C}$  during test.

Frequency	Nominal Body TSL Parameters (Permittivity/ Conductivity)	Measurement Body TSL parameters (Permittivity/ Conductivity)
1800 MHz	53.30/1.52	51.86/1.52


## 3. Measurement Results

Frequency	Return Loss with Head TSL	Return Loss with Body TSL
1800 MHz	-20.82 dB	-22.01 dB


## 校准结果/说明 (续页) :

Results of calibration and additional explanation (continued page)

## Return Loss Measurement Plot for head TSL



## Return Loss Measurement Plot for Body TSL



Remark: Attachment 1:SAR validation &amp; Test equipment

End

## Attachment 1: SAR validation &amp; Test equipment

Validation	Condition	SAR Value (W/kg)	
		1g	10g
SAR measured with Head TSL	1W (input power)	38.49	20.39
SAR measured with Body TSL	1W (input power)	37.78	20.06

名称/型号 Name/Model	编号 Number	证书编号/有效期限 Certificate No./Due date	测量范围/准确度 Measuring range/accuracy
6 axis Robot KR3	容-027-01	/	6 axes, Repeatability: $\pm$ 0.05 mm, Nominal payload: 3 kg
Vector Network Analyzer ZVB 8	容-027-27	2008F31-10-001907 2009.06.26	300 kHz to 8 GHz, Frequency resolution: 100 $\mu$ Hz, Measurement time: < 8 ms, Measurement bandwidths: 1 Hz to 500 kHz
Signal Generator SMT 06	容-027-15	2008F33-10-001469 2009.06.26	5 kHz - 6 GHz, Resolution: 0.1Hz, -144 to + 13 dBm, Max. RF power: 1W, Max. DC voltage: 0V / Level > -127 dBm; f < 1.5 GHz: < 1dB; f > 1.5 GHz: < 1.5dB; f > 3GHz: < 2dB
Power Meter NRVD	容-027-16	2008F31-10-001906 2009.06.24	100 kHz to 6 GHz, 10nW to 500mW
Millivoltmeter 2000	容-027-26	2008F11-10-001004 2009.06.19	Measurement range: 100.0000 mV ~ 1000.000V Sensitivity: 0.1 $\mu$ V ~ 1 m V.
Power Amplifier BLMA 0820-6	容-027-18	2008F33-10-001467 2009.06.26	0.8 - 2 GHz; Output: 6W; Gain: min 37.8 / typ 40, $\pm$ 2 dB; Harmonics: 2nd: 20dBc, 3rd: 20dBc; Line power: 125 W.
Isotropic E-Field Probe E-FIELD PROBE	容-027-54	2008J10-10-801001 2008.12.25	Dipole resistance (in the connector plane): 1M $\Omega$ to 2M Axial isotropy in human-equivalent liquids: < 0.25 dB Hemispherical Isotropy in human-equivalent liquids < 0.5 dB, Linearity < 0.5 dB, Lower SAR detection threshold: 0.0015 Watts/kg
SAM Phantom	容-027-22	/	/

## **Submit 4 E-field Calibration Report**

# COMOSAR E-Field probe Calibration Report



Ref: CR-131-1-09-SATB-A

Page: 1/17 Issue: A Date: 2009/05/11

## **COMOSAR E-FIELD PROBE CALIBRATION REPORT**

Prepared By: BUTET Romain, SATIMO

Project Description: COMOSAR E-FIELD PROBE

Prepared For (End User): CCS

This document is issued by SATIMO, in confidence and is not to be reproduced in whole or in part without the prior written permission. The information contained herein is to be used only for the purpose for which it is submitted and is not to be released in whole or in part without the prior written permission of SATIMO.

# COMOSAR E-Field probe Calibration Report



Ref: CR-131-1-09-SATB-A

Page: 2/17

Issue: A

Date: 2009/05/11

## COMOSAR SEPT ISOTROPIC E-FIELD PROBE CALIBRATION REPORT

**DATE:** 6/8/2009

**OFFER REFERENCE:** PF.127.1.09.SATB.A

**OBJECT:** COMOSAR SEPT ISOTROPIC E-FIELD PROBE

**MANUFACTURER:** SATIMO

**SERIAL NUMBER:** SN 11/09 EP100

**CUSTOMER:** CCS

**CONTRACT:** B01351

**DATE OF CALIBRATION:** 16/04/2009

### **WARRANTY:**

This Calibration certificate may not be reproduced other than in full. Calibration certificates without signature and seal are not valid. This documentation contains property information which is protected by copyright. All rights are reserved. No part of this document may be photocopied, reproduced without the prior written agreement of SATIMO. SATIMO shall not be liable for errors contained herein or for incidental or consequential in connection with the furnishing, performance or use of this material. Warranty doesn't apply to Normal wear, Normal tear, Improper use, Improper maintain, Improper installation.

---

Date

11/05/2009

SAR TEAM MANAGER

A handwritten signature in black ink, appearing to read "R. A. S." or a similar initials.


# COMOSAR E-Field probe Calibration Report



Ref: CR-131-1-09-SATB-A

Page: 3/17 Issue: A Date: 2009/05/11

## PRODUCT DESCRIPTION



Frequency Range	100 MHz - 30 GHz
Probe length	330 mm
Length of one dipole	4.5 mm
Maximum external diameter	8 mm
Probe extremity diameter	6.5 mm
Distance between dipoles/probe extremity	< 2.7 mm
Resistance of the three dipole (at the connector)	Dipole 1: $R1=2.5307 \text{ M}\Omega$ Dipole 2: $R2=2.6353 \text{ M}\Omega$ Dipole 3: $R3=2.5471 \text{ M}\Omega$
Connector (HIROSE series SR30)	6 wire male (Hirose SR30series)

The probe could be checked by measuring the resistance of the three dipoles.

## CALIBRATION TEST EQUIPMENT

TYPE	IDENTIFICATION	DATE OF CALIBRATION
Calibration bench	CALISAR CALIBRATION SYSTEM V2.0	
Multimeter	Keithley (2000, SN: 1000572)	Date of calibration: 01-07-2008

# COMOSAR E-Field probe Calibration Report




Ref: CR-131-1-09-SATB-A

Page: 4/17 Issue: A Date: 2009/05/11

## MEASUREMENT PROCEDURE

Probe calibration is realized, in compliance with CENELEC EN 50361 and IEEE 1528 std, with CALISAR, SATIMO proprietary calibration system. The calibration is performed with the EN 50361 annexe technique using reference guide at the five frequencies.



$$SAR = \frac{4(P_{fw} - P_{bw})}{ab\delta} \cos^2\left(\pi \frac{y}{a}\right) e^{-(2z/\delta)}$$

Where :

- $P_{fw}$  = Forward Power
- $P_{bw}$  = Backward Power
- a and b = Waveguide dimensions
- d = Skin depth

Keithley configuration:

Rate = Medium; Filter =ON; RDGS=10; FILTER TYPE =MOVING AVERAGE; RANGE AUTO

After each calibration, a SAR measurement is performed on a validation dipole and compared with a NPL calibrated probe, to verify it.

**COMOSAR E-Field probe  
Calibration Report**



Ref: CR-131-1-09-SATB-A

Page: 5/17 Issue: A Date: 2009/05/11

**PROBE UNCERTAINTIES**

**Calibration report of dosimetric  
SATIMO probe**

**Uncertainty on calibration system**

ERROR SOURCES	Uncertainty value (%)	Probability Distribution	Divisor	ci	Standard Uncertainty (%)
Incident or forward power	3,00%	Rectangular	$\sqrt{3}$	1	1,732%
Reflected power	3,00%	Rectangular	$\sqrt{3}$	1	1,732%
Liquid conductivity	5,00%	Rectangular	$\sqrt{3}$	1	2,887%
Liquid permittivity	4,00%	Rectangular	$\sqrt{3}$	1	2,309%
Field homogeneity	3,00%	Rectangular	$\sqrt{3}$	1	1,732%
Field probe positioning	5,00%	Rectangular	$\sqrt{3}$	1	2,887%
Field probe linearity	3,00%	Rectangular	$\sqrt{3}$	1	1,732%
<b>Combined standard uncertainty</b>					4,761%
<b>Expanded uncertainty (confidence interval of 95%)</b>					9,331%

# COMOSAR E-Field probe Calibration Report



Ref: CR-131-1-09-SATB-A

Page: 6/17 Issue: A Date: 2009/05/11


## 1. Calibration at 835.00 MHz

### A. Calibration parameters.

Label	850
Epsilon	41.82
Sigma	0.89 S/m
Temperature	21°C
Cable loss	0.11 dB
Coupler loss	20.50 dB
Waveguide S11	-11.20 dB
Low limit detection	0.824 V/m (0.604 mW/kg)

Calibration curves  $ei=f(V)$  ( $i=1,2,3$ ) allow to obtain E-field value using the formula:

$$E = \sqrt{E_1^2 + E_2^2 + E_3^2}$$



The following tables represent the calibration curves linearization by curve segment in CW signal.

# COMOSAR E-Field probe Calibration Report



Ref: CR-131-1-09-SATB-A

Page: 7/17 Issue: A Date: 2009/05/11


Calibration coefficients for the three dipoles in CW:

Sensitivity in liquid:

Liquid	Epsilon	Sigma (S/m)	CF dipole 1 (W.kg <sup>-1</sup> (mV) <sup>-1</sup> )	CF dipole 2 (W.kg <sup>-1</sup> (mV) <sup>-1</sup> )	CF dipole 3 (W.kg <sup>-1</sup> (mV) <sup>-1</sup> )
Head	41.82	0.89	20.63	20.50	28.35
Body	55.09	0.94	20.01	19.89	27.76

## B. Isotropy.

- Axial isotropy: 0.029 dB
- Hemispherical isotropy: 0.030 dB



## C. Linearity.

- Linearity: 0.04 dB

# COMOSAR E-Field probe Calibration Report



Ref: CR-131-1-09-SATB-A

Page: 8/17 Issue: A Date: 2009/05/11


## 2. Calibration at 897.00 MHz

### A. Calibration parameters.

Label	900
Epsilon	41.24
Sigma	0.94 S/m
Temperature	21°C
Cable loss	0.10 dB
Coupler loss	20.27 dB
Waveguide S11	-16.70 dB
Low limit detection	0.795 V/m (0.59 mW/kg)

Calibration curves  $e_i=f(V)$  ( $i=1,2,3$ ) allow to obtain E-field value using the formula:

$$E = \sqrt{E_1^2 + E_2^2 + E_3^2}$$



The following tables represent the calibration curves linearization by curve segment in CW signal.

# COMOSAR E-Field probe Calibration Report



Ref: CR-131-1-09-SATB-A

Page: 9/17 Issue: A Date: 2009/05/11


Calibration coefficients for the three dipoles in CW:

Sensitivity in liquid:

Liquid	Epsilon	Sigma (S/m)	CF dipole 1 (W.kg <sup>-1</sup> (mV) <sup>-1</sup> )	CF dipole 2 (W.kg <sup>-1</sup> (mV) <sup>-1</sup> )	CF dipole 3 (W.kg <sup>-1</sup> (mV) <sup>-1</sup> )
Head	41.24	0.94	22.07	22.01	30.17
Body	55.99	1.02	21.56	21.33	29.11

## B. Isotropy.

- Axial isotropy: 0.029 dB
- Hemispherical isotropy: 0.030 dB



## C. Linearity.

- Linearity: 0.04 dB

**COMOSAR E-Field probe  
Calibration Report**



Ref: CR-131-1-09-SATB-A

Page: 10/17 Issue: A Date: 2009/05/11


**3. Calibration at 1747.00 MHz**

**A. Calibration parameters.**

Label	1800
Epsilon	38.57
Sigma	1.34 S/m
Temperature	21°C
Cable loss	0.18 dB
Coupler loss	20.20 dB
Waveguide S11	-13.15 dB
Low limit detection	0.832 V/m (0.93 mW/kg)

Calibration curves  $e_i=f(V)$  ( $i=1,2,3$ ) allow to obtain E-field value using the formula:

$$E = \sqrt{E_1^2 + E_2^2 + E_3^2}$$



The following tables represent the calibration curves linearization by curve segment in CW signal.

# COMOSAR E-Field probe Calibration Report



Ref: CR-131-1-09-SATB-A

Page: 11/17 Issue: A Date: 2009/05/11


Calibration coefficients for the three dipoles in CW:

Sensitivity in liquid:

Liquid	Epsilon	Sigma (S/m)	CF dipole 1 (W.kg-1 (mV)-1)	CF dipole 2 (W.kg-1 (mV)-1)	CF dipole 3 (W.kg-1 (mV)-1)
Head	38.57	1.34	37.12	38.57	50.40
Body	51.99	1.49	36.65	37.99	49.65

## B. Isotropy.

- Axial isotropy: 0.050 dB
- Hemispherical isotropy: 0.076 dB



## C. Linearity.

- Linearity: 0.03 dB

# COMOSAR E-Field probe Calibration Report



Ref: CR-131-1-09-SATB-A

Page: 12/17 Issue: A Date: 2009/05/11


## 4. Calibration at 1880.00 MHz

### A. Calibration parameters.

Label	1900
Epsilon	38.34
Sigma	1.45 S/m
Temperature	21°C
Cable loss	0.18 dB
Coupler loss	21.15 dB
Waveguide S11	-26.90 dB
Low limit detection	0.796 V/m (0.92 mW/kg)

Calibration curves  $e_i=f(V)$  ( $i=1,2,3$ ) allow to obtain E-field value using the formula:

$$E = \sqrt{E_1^2 + E_2^2 + E_3^2}$$



The following tables represent the calibration curves linearization by curve segment in CW signal.

# COMOSAR E-Field probe Calibration Report



Ref: CR-131-1-09-SATB-A

Page: 13/17 Issue: A Date: 2009/05/11


Calibration coefficients for the three dipoles in CW:

Sensitivity in liquid:

Liquid	Epsilon	Sigma (S/m)	CF dipole 1 (W.kg <sup>-1</sup> (mV) <sup>-1</sup> )	CF dipole 2 (W.kg <sup>-1</sup> (mV) <sup>-1</sup> )	CF dipole 3 (W.kg <sup>-1</sup> (mV) <sup>-1</sup> )
Head	38.34	1.45	41.07	42.36	55.46
Body	52.13	1.50	40.41	41.11	54.77

## B. Isotropy.

- Axial isotropy: 0.050 dB
- Hemispherical isotropy: 0.076 dB



## C. Linearity.

- Linearity: 0.03 dB

# COMOSAR E-Field probe Calibration Report



Ref: CR-131-1-09-SATB-A

Page: 14/17 Issue: A Date: 2009/05/11


## 5. Calibration at 1950.00 MHz

### A. Calibration parameters.

Label	2000
Epsilon	38.19
Sigma	1.47 S/m
Temperature	21°C
Cable loss	0.19 dB
Coupler loss	20.10 dB
Waveguide S11	-30.10 dB
Low limit detection	0.787 V/m (0.94 mW/kg)

Calibration curves  $e_i=f(V)$  ( $i=1,2,3$ ) allow to obtain E-field value using the formula:

$$E = \sqrt{E_1^2 + E_2^2 + E_3^2}$$



The following tables represent the calibration curves linearization by curve segment in CW signal.

# COMOSAR E-Field probe Calibration Report



Ref: CR-131-1-09-SATB-A

Page: 15/17 Issue: A Date: 2009/05/11


Calibration coefficients for the three dipoles in CW:

Sensitivity in liquid:

Liquid	Epsilon	Sigma (S/m)	CF dipole 1 (W.kg-1 (mV)-1)	CF dipole 2 (W.kg-1 (mV)-1)	CF dipole 3 (W.kg-1 (mV)-1)
Head	38.19	1.46	41.92	43.16	56.44
Body	54.05	1.52	41.01	42.41	55.66

## B. Isotropy.

- Axial isotropy: 0.050 dB
- Hemispherical isotropy: 0.076 dB



## C. Linearity.

- Linearity: 0.03 dB

# COMOSAR E-Field probe Calibration Report



Ref: CR-131-1-09-SATB-A

Page: 16/17 Issue: A Date: 2009/05/11


## 6. Calibration at 2450.00 MHz

### A. Calibration parameters.

Label	2450
Epsilon	37.44
Sigma	1.75 S/m
Temperature	21°C
Cable loss	0.20 dB
Coupler loss	21.50 dB
Waveguide S11	-13.65 dB
Low limit detection	0.793 V/m (1.09 mW/kg)

Calibration curves  $e_i=f(V)$  ( $i=1,2,3$ ) allow to obtain E-field value using the formula:

$$E = \sqrt{E_1^2 + E_2^2 + E_3^2}$$



The following tables represent the calibration curves linearization by curve segment in CW signal.

# COMOSAR E-Field probe Calibration Report



Ref: CR-131-1-09-SATB-A

Page: 17/17 Issue: A Date: 2009/05/11


Calibration coefficients for the three dipoles in CW:

Sensitivity in liquid:

Liquid	Epsilon	Sigma (S/m)	CF dipole 1 (W.kg <sup>-1</sup> (mV) <sup>-1</sup> )	CF dipole 2 (W.kg <sup>-1</sup> (mV) <sup>-1</sup> )	CF dipole 3 (W.kg <sup>-1</sup> (mV) <sup>-1</sup> )
Head	37.44	1.75	51.19	53.87	70.49
Body	53.70	1.96	50.36	52.99	69.77

## B. Isotropy.

- Axial isotropy: 0.050 dB
- Hemispherical isotropy: 0.076 dB



## C. Linearity.

- Linearity: 0.03 dB