PROBE CALIBRATION CERTIFICATES

chmid & Partner Engineering AG ughausstrasse 43, 8004 Zuri	ry of		Schweizerlacher Kalibrierdienst Service suisse d'étalonnage Servizio svizzero di taratura Swiss Calibration Service
credited by the Swiss Accredit e Swiss Accreditation Servic Itilateral Agreement for the r	ce is one of the signatories t	o the EA	editation No.: SCS 0108
ient BACL-SZ (Aud	Sector production		ES3-3019_Aug18
ALIBRATION	CERTIFICATE		
bject	ES3DV2 - SN:3019	9	
Calibration procedure(s)	QA CAL-01.v9, QA Calibration procedu	CAL-12:v9, QA CAL-23.v5, QA ure for dosimetric E-field probes	CAL-25.v6
Calibration date	August 20, 2018		
The measurements and the unc	pertainties with confidence prol	al standards, which realize the physical units bability are given on the following pages and a facility: environment temperature (22 ± 3) ² C a	are part of the certificato.
he measurements and the unc Il calibrations have been cond	pertainties with confidence prot ucted in the closed laboratory	bability are given in the following pages and a facility: environment temperature (22 \pm 3)°C a	are part of the certificato.
The measurements and the unc All calibrations have been cond Calibration Equipment used (Mi	pertainties with confidence prot ucted in the closed laboratory	bability are given on the following pages and a	are part of the certificato.
he measurements and the unc Il calibrations have been cond Calibration Equipment used (Mi Primary Standards	entainties with confidence prof ucted in the closed laboratory &TE critical for calibration)	bability are given on the following pages and a facility: environment temperature (22 \pm 3) $^{\circ}\mathrm{C}$ a	nd humidity < 70%.
he measurements and the unc Il calibrations have been cond alibration Equipment used (Mi Primary Standards Power meter NRP	entainties with confidence prof ucted in the closed laboratory &TE critical for calibration)	facility are given on the following pages and a facility: environment temperature (22 ± 3)°C a Cal Date (Centificate No.)	nd humidity < 70%.
he measurements and the unc Il calibrations have been cond alibration Equipment used (Mi Primary Standards Power meter NRP Power sensor NRP-Z91	entainties with confidence prof ucted in the closed laboratory &TE critical for calibration) ID SN: 104778	facility are given on the following pages and a facility: environment temperature (22 ± 3)°C a Cal Date (Certificate No.) 04-Apr-18 (No. 217-02672/02673)	scheduled Calibration
The measurements and the unc all calibrations have been cond Calibration Equipment used (Mi Primary Standards Power meter NRP Power sensor NRP-Z91 Power sensor NRP-Z91	entainties with confidence prof ucted in the closed laboratory &TE critical for calibration) ID SN: 104778 SN: 104244	bability are given on the following pages and a facility: environment temperature (22 ± 3)°C a Cal Date (Certificate No.) 04-Apr-18 (No. 217-02672/02673) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02672)	Scheduled Calibration Apr-19 Apr-19
The measurements and the unc all calibrations have been cond Calibration Equipment used (Mi Primary Standards Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator	entainties with confidence prot ucted in the closed laboratory &TE critical for calibration) ID SN: 104778 SN: 10474 SN: 103244 SN: 103245	bability are given on the following pages and a facility: environment temperature (22 ± 3)°C a Cal Date (Centificate No.) 04-Apr-18 (No. 217-02672/02673) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02672)	Scheduled Calibration Apr-19 Apr-19
he measurements and the unc al calibrations have been cond calibration Equipment used (Mi Primary Standards Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator Reference Probe ES3DV2	entainties with confidence prof ucted in the closed laboratory &TE critical for calibration) ID SN: 104778 SN: 104778 SN: 103244 SN: 103245 SN: SS277 (20x)	bability are given on the following pages and a facility: environment temperature (22 ± 3)°C a Cal Date (Certificate No.) 04-Apr-18 (No. 217-02672/02673) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02672)	Scheduled Calibration Apr-19 Apr-19 Apr-19 Apr-19
he measurements and the unc all calibrations have been cond calibration Equipment used (Mi Primary Standards Power meter NRP Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator Reference Probe ES3DV2 DAE4	entainties with confidence prof ucted in the closed laboratory &TE critical for calibration) ID SN: 104778 SN: 103244 SN: 103245 SN: 55277 (20x) SN: 3013	bability are given on the following pages and a facility: environment temperature (22 ± 3)°C a Cal Date (Certificate No.) 04-Apr-18 (No. 217-02672/02673) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02673) 04-Apr-17 (No. ES3-3013_Dec17)	Scheduled Calibration Apr-19 Apr-19 Apr-19 Dec-18 Dec-18 Scheduled Check
he measurements and the unc al calibrations have been cond calibration Equipment used (Mi Primary Standards Power meter NRP Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator Reference Probe ES3DV2 DAE4 Secondary Standards	entainties with confidence prof ucted in the closed laboratory &TE critical for calibration) ID SN: 104778 SN: 103244 SN: 103245 SN: 55277 (20x) SN: 5013 SN: 660	Cal Date (Centificate No.) Cal Date (Centificate No.) 04-Apr-18 (No. 217-02672/02673) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02682) 30-Dec-17 (No. 283-3013_Dec17) 21-Dec-17 (No. DAE4-660_Dec17) Check Date (in house) 06-Apr-16 (in house Apr-18)	Scheduled Calibration Apr-19 Apr-19 Apr-19 Apr-19 Dec-18 Dec-18 Dec-18 Dec-18 Dec-18 In house check: Jun-20
he measurements and the unc Il calibrations have been cond calibration Equipment used (Mi Primary Standards Power meter NRP Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator Reference Probe ES30V2 DAE4 Secondary Standards Power meter E44198	entainties with confidence prof ucted in the closed laboratory &TE critical for calibration) ID SN: 104778 SN: 103244 SN: 103245 SN: 3013 SN: 3013 SN: 680 ID	Dability are given on the following pages and a facility: environment temperature (22 ± 3)°C a Cal Date (Certificate No.) 04-Apr-18 (No. 217-02572/02573) 04-Apr-18 (No. 217-02572/02573) 04-Apr-18 (No. 217-02572) 04-Apr-18 (No. 217-02572) 04-Apr-18 (No. 217-02573) 04-Apr-18 (No. 217-02562) 30-Dec-17 (No. 217-02682) 30-Dec-17 (No. 216-02662) 30-Dec-17 (No. DAE4-660_Dec17) 21-Dec-17 (No. DAE4-660_Dec17) Check Date (in house) 05-Apr-16 (in house check Jum-18) 05-Apr-16 (in house check Jum-18)	Scheduled Calibration Apr-19 Apr-19 Apr-19 Apr-19 Apr-19 Dec-18 Dec-18 Dec-18 Scheduled Check In house check; Jun-20 In house check; Jun-20
he measurements and the unc alibration Equipment used (Mi Primary Standards Power meter NRP Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator Reference 20 dB Attenuator Reference Probe ES3DV2 DAE4 Secondary Standards Power sensor E44198 Power sensor E4412A	entainties with confidence prof ucted in the closed laboratory &TE critical for calibration) ID SN: 104778 SN: 103244 SN: 103245 SN: 55277 (20x) SN: 5013 SN: 660 ID SN: 680 SN: 6841293874 SN: MY41496087 SN: 000110210	Cal Date (Certificate No.) Cal Date (Certificate No.) 04-Apr-18 (No. 217-02672/02673) 04-Apr-18 (No. 217-02672/02673) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02682) 30-Dec-17 (No. DAE4-660_Dec17) 21-Dec-17 (No. DAE4-660_Dec17) Check Date (in house) 05-Apr-16 (in house check Jum-18) 06-Apr-16 (in house check Jum-18) 08-Apr-16 (in house check Jum-18)	Scheduled Calibration Apr-19 Apr-19 Apr-19 Apr-19 Apr-19 Dec-18 Dec-18 Dec-18 Scheduled Check In house check: Jun-20 In house check: Jun-20
he measurements and the unc all calibrations have been cond calibration Equipment used (Mi Primary Standards Power meter NRP Power sensor NRP-291 Reference Probe ES30V2 DAE4 Secondary Standards Power sensor E44198 Power sensor E4412A RF generator HP 8848C	entainties with confidence prof ucted in the closed laboratory &TE critical for calibration) ID SN: 104778 SN: 103244 SN: 103245 SN: 3013 SN: 660 ID SN: GB41293674 SN: 08110210 SN: 036110210 SN: US3642U01700	Cal Date (Certificate No.) Q4-Apr-18 (No. 217-02672/02673) Q4-Apr-18 (No. 217-02672/02673) Q4-Apr-18 (No. 217-02672/02673) Q4-Apr-18 (No. 217-02672) Q4-Apr-18 (No. 217-02672) Q4-Apr-18 (No. 217-02672) Q4-Apr-18 (No. 217-02672) Q4-Apr-18 (No. 217-02682) 30-Dec-17 (No. 217-0	Scheduled Calibration Apr-19 Apr-19 Apr-19 Apr-19 Dec-18 Dec-18 Dec-18 Scheduled Check In house check: Jun-20 In house check: Jun-20 In house check: Jun-20 In house check: Jun-20
he measurements and the unc all calibrations have been cond calibration Equipment used (Mi Primary Standards Power meter NRP Power sensor NRP-291 Reference Probe ES30V2 DAE4 Secondary Standards Power sensor E44198 Power sensor E4412A RF generator HP 8848C	entainties with confidence prof ucted in the closed laboratory &TE critical for calibration) ID SN: 104778 SN: 103244 SN: 103245 SN: 55277 (20x) SN: 5013 SN: 660 ID SN: 680 SN: 6841293874 SN: MY41496087 SN: 000110210	Cal Date (Certificate No.) Cal Date (Certificate No.) 04-Apr-18 (No. 217-02672/02673) 04-Apr-18 (No. 217-02672/02673) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02682) 30-Dec-17 (No. DAE4-660_Dec17) 21-Dec-17 (No. DAE4-660_Dec17) Check Date (in house) 05-Apr-16 (in house check Jum-18) 06-Apr-16 (in house check Jum-18) 08-Apr-16 (in house check Jum-18)	Scheduled Calibration Apr-19 Apr-19 Apr-19 Apr-19 Apr-19 Dec-18 Dec-18 Dec-18 Scheduled Check In house check: Jun-20 In house check: Jun-20
he measurements and the unc Il calibrations have been cond calibration Equipment used (Mi Primary Standards Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator Reference 20 dB Attenuator Reference Probe ES3DV2 DAE4 Secondary Standards Power sensor E44198 Power sensor E4412A Power sensor E4412A RF génerator HP B848C	entainties with confidence prof ucted in the closed laboratory &TE critical for calibration) ID SN: 104778 SN: 103244 SN: 103245 SN: 3013 SN: 660 ID SN: GB41293674 SN: 08110210 SN: 036110210 SN: US3642U01700	Cal Date (Certificate No.) Q4-Apr-18 (No. 217-02672/02673) Q4-Apr-18 (No. 217-02672/02673) Q4-Apr-18 (No. 217-02672/02673) Q4-Apr-18 (No. 217-02672) Q4-Apr-18 (No. 217-02672) Q4-Apr-18 (No. 217-02672) Q4-Apr-18 (No. 217-02672) Q4-Apr-18 (No. 217-02682) 30-Dec-17 (No. 217-0	Scheduled Calibration Apr-19 Apr-19 Apr-19 Apr-19 Dec-18 Dec-18 Dec-18 Scheduled Check In house check: Jun-20 In house check: Jun-20 In house check: Jun-20 In house check: Jun-20
The measurements and the uncertain and the uncer	entainties with confidence prof ucted in the closed laboratory 8TE critical for calibration) 8TE critical for calibration 8TE cr	bability are given on the following pages and a facility: environment temperature (22 ± 3)°C a Cal Date (Certificate No.) 04-Apr-18 (No. 217-02672/02673) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02662) 30-Dec-17 (No. DAE4-660_Dec17) 21-Dec-17 (No. DAE4-660_Dec17) Check Date (in house check Jun-18) 05-Apr-16 (in house check Jun-18) 06-Apr-16 (in house check Jun-18) 04-Aug-99 (in house check Jun-18) 31-Mar-14 (in house check Jun-18)	Scheduled Calibration Aar-19 Agr-19 Agr-19 Agr-19 Apr-19 Dec-18 Dec-18 Dec-18 Scheduled Check In house check: Jun-20 In house check: Jun-20
he measurements and the unc	entainties with confidence prof ucted in the closed laboratory &TE critical for calibration) ID SN: 104778 SN: 103244 SN: 103245 SN: 3013 SN: 680 ID SN: 680 ID SN: 680 ID SN: 680 SN: 00110210 SN: 00110210 SN: US3642U01700 SN: US3642U01700 SN: US41080477 Name	Cal Date (Centificate No.) Cal Date (Centificate No.) 04-Apr-18 (No. 217-02672/002673) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02682) 30-Dec-17 (No. 253-3013, Dec17) 21-Dec-17 (No. DAE4-660_Dec17) Check Date (in house) 06-Apr-16 (in house check Jun-18) 05-Apr-16 (in house check Jun-18) 06-Apr-16 (in house check Jun-18) 07-Apr-16 (in house check Jun-18) 08-Apr-16 (in house check Jun-18) 09-Apr-16 (in house check Oct-17)	Scheduled Calibration Apr-19 Apr-19 Apr-19 Apr-19 Apr-19 Dec-18 Dec-18 Dec-18 Scheduled Check In house check: Jun-20 In house check: Jun-20

Certificate No: ES3-3019_Aug18

Page 1 of 39

Calibration Laboratory of Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzenland



Schweizerischer Kalibriordienst Service suisse d'étalonnage C Servizio svizzero di taratura Swiss Calibration Service

S

S

Accreditation No.: SCS 0108

Accredited by the Swiss Accreditation Service (SAS) The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

Glossarv

orosoury.	
TSL	tissue simulating liquid
NORMx,y,z	sensitivity in free space
ConvF	sensitivity in TSL / NORMX,y.z.
DCP	diade compression point
CF	crest factor (1/duty_cycle) of the RF signal
A. B. C. D	modulation dependent linearization parameters
Polarization ()	in rotation around probe axis
Polanzation &	3 rotation around an axis that is in the plane normal to probe axis (at measurement center),
	i.e. 3 = 0 is normal to probe axis
Connector Angle	information used in DASY system to align probe sensor X to the robot coordinate system

Calibration is Performed According to the Following Standards:

- IEEE Std 1528-2013, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement
- Techniques", June 2013 IEC 62209-1, ", "Measurement procedure for the assessment of Specific Absorption Rate (SAR) from hard-held and body-mounted devices used next to the ear (frequency range of 300 MHz to 6 GHz)". July 2016 IEC 62209-2, "Procedure to determine the Specific Absorption Rate (SAR) for wireless communication devices. b)
- C) used in close proximity to the human body (frequency range of 30 MHz to 6 GHz)*, March 2010 KDB 865664, "SAR Measurement Requirements for 100 MHz to 6 GHz"
- d)

Methods Applied and Interpretation of Parameters:

- NORMx, y, z; Assessed for E-field polarization & = 0 (f < 900 MHz in TEM-cell; 1 > 1800 MHz; R22 waveguide). NORMx, y,z are only intermediate values, i.e., the uncertainties of NORMx, y,z does not affect the E2-field uncertainty inside TSL (see below ConvF).
- NORM(f)x,y,z = NORMx,y,z * frequency_response (see Frequency Response Chart). This linearization is implemented in DASY4 software versions later than 4.2. The uncertainty of the frequency response is included in the stated uncertainty of ConvF.
- DCPx, y, z: DCP are numerical linearization parameters assessed based on the data of power sweep with CW signal (no uncertainty required). DCP does not depend on frequency nor media.
- PAR: PAR is the Peak to Average Ratio that is not calibrated but determined based on the signal characteristics
- Ax,y,z; Bx,y,z; Cx,y,z; Dx,y,z; VRx,y,z; A, B, C, D are numerical linearization parameters assessed based on the data of power sweep for specific modulation signal. The parameters do not depend on frequency nor mode. We is the average of the second secon media. VR is the maximum calibration range expressed in RMS voltage across the diode.
- ConvF and Boundary Effect Parameters: Assessed in flat phantom using E-field (or Temperature Transfer Standard for f < 800 MHz) and inside waveguide using analytical field distributions based on power measurements for f > 800 MHz. The same setups are used for assessment of the parameters applied for boundary compensation (alpha, depth) of which typical uncertainty values are given. These parameters are used in DASY4 software to improve probe accuracy close to the boundary. The sensitivity in TSL corresponds to NORMx.y.z * ConvF whereby the uncertainty corresponds to that given for ConvF. A frequency dependent ConvF is used in DASY version 4.4 and higher which allows extending the validity from ± 50 MHz to ± 100 MHz.
- Spherical isotropy (3D deviation from isotropy): in a field of low gradients realized using a flat phantom exposed by a patch antenna.
- Sensor Offset: The sensor offset corresponds to the offset of virtual measurement center from the probe tip (on probe axis). No tolerance required.
- Connector Angle: The angle is assessed using the information gained by determining the NORMx (no uncertainty required).

Certificate No: ES3-3019_Aug18

Page 2 of 39

August 20, 2018

Probe ES3DV2

SN:3019

Manufactured: Calibrated: December 5, 2002 August 20, 2018

Calibrated for DASY/EASY Systems (Note: non-compatible with DASY2 system!)

Certificate No: ES3-3019_Aug18

Page 3 of 39

August 20, 2018

DASY/EASY - Parameters of Probe: ES3DV2 - SN:3019

Basic Calibration Parameters

	Sensor X	Sensor Y	Sensor Z	Unc (k=2)
Norm (uV/(V/m)2)A	1.01	1.13	0.93	± 10.1 %
DCP (mV) ^B	104.8	103.8	106.3	

Modulation Calibration Parameters

UID	Communication System Name		A dB	B dBõV	c	D dB	VR mV	Unc" (k=2)
0	CW	X	0.0	0.0	1.0	0.00	189.8	±3.0 %
-	1970	Y	0.0	0.0	1.0		205.7	
		Z	0.0	0.0	1.0		205.8	

Note: For details on UID parameters see Appendix.

Sensor Model Parameters

	C1 fF	C2 fF	a V−1	T1 ms.V-2	T2 ms.V ⁻¹	T3 ms	T4 V-=	T5 V-1	T6
X	27.76	200.6	35.82	18,61	0.506	5.10	0.000	0.276	1.005
Y	28,17	203.3	35.77	18.85	0.706	5.10	0.000	0.137	1.010
Z	26.29	187.2	34.88	16.18	0.325	5.10	1.142	0.072	1.007

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

⁶ The uncertainties of Norm X,Y,Z do not affect the E²-field uncertainty inside TSL (see Pages 5 and 6).
⁹ Numerical insanzation parameter: uncertainty not required.
⁴ Uncertainty is determined using the max, deviation from linear response applying rectangular distribution and is expressed for the square of the field value.

Certificate No: ES3-3019_Aug18

Page 4 of 39

August 20, 2018

DASY/EASY - Parameters of Probe: ES3DV2 - SN:3019

Calibration Parameter Determined in Head Tissue Simulating Media

f (MHz)	Relative Permittivity ^F	Conductivity (S/m) ^F	ConvF X	ConvF Y	ConvF Z	Alpha ⁰	Depth ^o (mm)	Unc (k=2)
150	52.3	0.76	7.67	7.67	7.67	0.05	1.50	± 13.3 %
450	43.5	0.87	7.18	7.18	7.18	0.15	1.60	± 13.3 %

^C Frequency validity above 300 MHz of ± 100 MHz only applies for DASY v4.4 and higher (see Page 2), else it is restricted to ± 50 MHz. The uncertainty is the RSS of the ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band. Frequency validity below 300 MHz is ± 10, 25, 40, 50 and 70 MHz for ConvF assessments at 30, 64, 128, 150 and 220 MHz respectively. Above 5 GHz frequency validity validity can be extended to ± 110 MHz.
^F At frequencies below 3 GHz, the validity of tissue parameters (c and e) can be released to ± 10% if liquid compensation formula is applied to measured SAR values. At frequencies above 3 GHz, the validity of tissue parameters (c and e) can be released to ± 10%. The uncertainty is the RSS of the ConvF uncertainty for indicated target tissue parameters.
^C AlphaDepth are detormined during calibration. SPEAC warrants that the remaining deviation due to the boundary effect after compensation is always less than ± 1% for frequencies below 3 GHz and below ± 2% for frequencies between 3-6 GHz at any distance larger than half the probe tip diameter from the boundary.

Certificate No: ES3-3019_Aug18

Page 5 of 39

August 20, 2018

DASY/EASY - Parameters of Probe: ES3DV2 - SN:3019

1	f (MHz) ^c	Relative Permittivity ^F	Conductivity (S/m) ^F	ConvF X	ConvF Y	ConvF Z	Alpha ^G	Depth ^d (mm)	Unc (k=2)
	150	61.9	0.80	7.30	7.30	7.30	0.07	1.50	± 13.3 %
	450	56.7	0.94	7.10	7.10	7.10	0.10	1.50	± 13.3 %

Calibration Parameter Determined in Body Tissue Simulating Media


^c Frequency validity above 300 MHz of ± 100 MHz only applies for DASY v4.4 and higher (see Page 2), else it is restricted to ± 50 MHz. The uncertainty is the RSS of the ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band. Frequency validity below 300 MHz is ± 10, 25, 40, 50 and 70 MHz for ConvF assessments at 30, 64, 128, 150 and 220 MHz respectively. Above 5 GHz frequency validity can be extended to ± 110 MHz.
^c At frequencies below 3 GHz, the validity of tissue parameters (c and d) can be relaxed to ± 10% if liquid compensation formula is applied to measured SAR values. At frequencies above 3 GHz, the validity of tissue parameters (c and d) is restricted to ± 5%. The uncertainty is the RSS of the ConvF uncertainty for indicated target lissue parameters.
^c Alpha/Depth are determined during calibration. SPEAC warrants that the remaining deviation due to the boundary effect after compensation is always less than ± 1% for frequencies below 3 GHz and below ± 2% for frequencies between 3-6 GHz at any distance larger than half the probe tip diameter from the boundary.

Certificate No: ES3-3019_Aug18

Page 6 of 39

August 20, 2018

Frequency Response of E-Field (TEM-Cell:ifi110 EXX, Waveguide: R22)




Uncertainty of Frequency Response of E-field: ± 6.3% (k=2)

Certificate No: ES3-3019_Aug18

Page 7 of 39

August 20, 2018



Receiving Pattern (ϕ), $\vartheta = 0^{\circ}$


Uncertainty of Axial Isotropy Assessment: ± 0.5% (k=2)

Certificate No: ES3-3019_Aug18

Page 8 of 39

August 20, 2018






Certificate No: ES3-3019_Aug18

Page 9 of 39

August 20, 2018



August 20, 2018

DASY/EASY - Parameters of Probe: ES3DV2 - SN:3019

Sensor Arrangement	Triangular
Connector Angle (*)	-18.7
Mechanical Surface Detection Mode	enabled
Optical Surface Detection Mode	disabled
Probe Overall Length	337 mm
Probe Body Diameter	10 mm
Tip Length	10 mm
Tip Diameter	4 mm
Probe Tip to Sensor X Calibration Point	2 mm
Probe Tip to Sensor Y Calibration Point	2 mm
Probe Tip to Sensor Z Calibration Point	2 mm
Recommended Measurement Distance from Surface	3 mm

Certificate No: ES3-3019_Aug18

Page 11 of 39

August 20, 2018

Appendix: Modulation	Calibration	Parameters
----------------------	-------------	------------

	Communication System Name		A dB	B dBõV	c	D dB	VR mV	Max Unc ^E (k=2)
0	CW	X	0.00	0.00	1,00	0.00	189.8	± 3.0 %
_		Y	0.00	0.00	1.00		205.7	
10010-	01011111111	Z	0.00	0.00	1.00		205.8	1
CAA	SAR Validation (Square, 100ms, 10ms)	x	3.29	69.40	11.94	10.00	25.0	± 9.6 %
		Y	6.86	78.15	16.03	-	25.0	
		Z	3.55	70.52	12.29		25.0	
10011- CAB	UMTS-FDD (WCDMA)	x	0.88	67.55	14,41	0.00	150.0	±9.6 %
		Y	1.00	69.15	15.56		150.0	
		Z	0.82	66.15	13.59		150.0	
10012- CAB	IEEE 802.11b WIFI 2.4 GHz (DSSS, 1 Mbps)	×	1.15	64.93	15.50	0.41	150.0	± 9.6 %
		Y	1.20	65.35	15.90	-	150.0	
		Z	1.13	64.32	14.98		150.0	
10013- CAB	IEEE 802.11g WiFi 2.4 GHz (DSSS- OFDM, 6 Mbps)	X	4.63	67.61	17,39	1.46	150.0	± 9.6 %
		Ŷ	4.69	67.72	17.52	1	150.0	
		Z	4.58	67.52	17.24		150.0	
10021- DAC	GSM-FDD (TDMA, GMSK)	x	100.00	114,64	27.74	9.39	50.0	± 9.6 %
-		Y	100.00	117.82	29.54		50.0	
		Z	100.00	114.90	27.69	1	50.0	
10023- DAC	GPRS-FOD (TDMA, GMSK, TN 0)	×	100.00	113.96	27.45	9.57	50.0	± 9.6 %
		Y	100.00	117.23	29.30		50.0	
		Z	100.00	114,00	27.30		50.0	
10024- DAC	GPRS-FDD (TDMA, GMSK, TN 0-1)	×	100.00	112,40	25.80	6.56	60.0	± 9.6 %
		Y	100.00	116.08	27.75		60.0	
		Z	100.00	113.76	26.24		60.0	
10025- DAC	EDGE-FDD (TDMA, 8PSK_ TN 0)	x	9.98	97.93	39.31	12.57	50.0	±9.6 %
		Y	32.24	137.81	54.11	-	50,0	
20240		Z	7,11	87.72	35.09		50.0	
10026- DAC	EDGE-FDD (TDMA, 8PSK, TN 0-1)	×	14.40	104.70	37.60	9.56	60.0	± 9.6 %
		Y	21.43	115.81	41.75		60.0	
40007	ODDO COD COMMA ONDER THE 2 YO	Z	10.88	98.23	35.49	1.0.0	60.0	
10027- DAC	GPRS-FDD (TDMA, GMSK, TN 0-1-2)	x	100.00	112.28	25.05	4.80	80.0	± 9.6 %
		Y	100.00	116.70	27.28		80.0	-
10000	0000 CDD (TOLL) 2000 TOLS	Z	100.00	114.64	25.92		80.0	
10028- DAC	GPRS-FDD (TDMA, GMSK, TN 0-1-2-3)	×	100.00	112.94	24.67	3.55	100.0	±9.6 %
		Y	100.00	118.58	27.39	-	100.0	_
10000	FROM FROM PROVIDENCE AND	Z	100.00	116.41	26.01		100.0	
10029- DAC	EDGE-FDD (TDMA, 8PSK, TN 0-1-2)	X	8.11	90.12	31.09	7.80	80.0	±9.6 %
		Y	9.67	94.63	33.13		80.0	-
40.030	IFFE ODS OF A DUNIE OF ADDRESS OF ADDRESS	Z	6.66	85.92	29,56		80.0	
10030- CAA	IEEE 802.15.1 Bluetooth (GFSK, DH1)	X	100.00	109.77	24,16	5.30	70.0	±9.6 %
	-	Y	100.00	113.94	26.30		70.0	
10031-	IEEE 802.15.1 Bluetooth (GFSK, DH3)	Z X	100.00	111.16 105.00	24.61 19.97	1.88	70.0	±9.6 %
CAA		Y	100.00	115.59	24.65		100.0	

Certificate No: ES3-3019_Aug18

Page 12 of 39

August 20, 2018

10032- CAA	IEEE 802.15.1 Bluetooth (GFSK, DH5)	х	100.00	100.25	17,19	1.17	100.0	±9.6 %
um		Y	100.00	119.00	25,00		100.0	
		Z	100.00	106.97	19.82		100.0	-
10033- CAA	IEEE 802.15.1 Bluetooth (Pt/4-DQPSK. DH1)	x	100.00	117.31	28.80	5.30	70.0	± 9.6 %
di si i	No. 10	Y	100.00	119.34	29.96		70.0	
-		Z	100.00	117.71	28.83		70.0	
10034- CAA	IEEE 802.15.1 Bluetooth (PI/4-DQPSK, DH3)	X	3.39	74.56	14.02	1.8B	100.0	±9.6 %
	191107	Y	7.50	83.72	17.60		100.0	
		Z	2.76	72.78	13.27		100.0	
10035- CAA	IEEE 802.15.1 Bluetooth (PI/4-DOPSK, DH5)	X	1.52	67,61	10.81	1.17	100.0	±9.6 %
		Y	2.55	73.11	13.53		100.0	-
		Z	1.33	66.63	10.29		100.0	
10036- CAA	IEEE 802.15.1 Bluetooth (8-DPSK, DH1)	Х	100.00	117.74	29.00	5.30	70.0	±9.6 %
		Y	100.00	119.74	30.14		70.0	
		Z	100.00	118.21	29.05	1	70.0	
10037- CAA	IEEE 802.15.1 Bluelooth (8-DPSK, DH3)	X	2.81	72.76	13.37	1.88	100.0	± 9.6 %
		Y	5.49	80.47	16.60		100.0	
		Z	2.33	71.15	12.66		100.0	
10038- CAA	IEEE 802.15.1 Bluetooth (8-DPSK, DH5)	X	1.57	68.12	11.15	1,17	100.0	± 9.6 %
		Y	2.68	73.90	13,96		100.0	
		Z	1.36	66.99	10.57		100.0	
10039- CAB	CDMA2000 (1xRTT, RC1)	x	0.51	60.81	6.58	0.00	150.0	± 9.6 %
		Y	0.63	62.47	B.01		150.0	
		Z	0.48	60.49	6.29	1.000	150.0	1.1.1
10042- CAB	IS-54 / IS-136 FDD (TDMA/FDM, PI/4- DQPSK, Halfrate)	х	100.00	109.34	24.59	7.78	50.0	± 9.6 %
		Y	100.00	113.01	26.55		50.0	
		Z	100.00	110.01	24.73		50.0	
10044- CAA	IS-91/EIA/TIA-553 FDD (FDMA, FM)	X	0.06	123.26	6.55	0.00	150.0	± 9.6 %
		Y	0.00	117.23	6.07		150.0	
		Z	0.02	120.00	0,96		150.0	1
10048- CAA	DECT (TDD, TDMA/FDM, GFSK, Full Slot, 24)	x	100.00	113.79	28.54	13.80	25.0	± 9.6 %
		Y	100.00	117.44	30.66		25.0	
		Z	100.00	112.50	27.83		25.0	
10049- CAA	DECT (TDD, TDMA/FDM, GFSK, Double Slot, 12)	×	100.00	113.10	27.30	10.79	40.0	± 9.6 %
oloc -		Y	100.00	116.50	29.24		40.0	
		Z	100.00	112.54	26.89		40.0	1.0
10056- CAA	UMTS-TDD (TD-SCDMA, 1.26 Mcps)	X	100.00	118.95	30.75	9.03	50.0	±9.6 %
		Y	100.00	120.98	31.98		50.0	
-		Z	100.00	119.02	30.62		50.0	
10058- DAC	EDGE-FDD (TDMA, 8PSK, TN 0-1-2-3)	×	5.91	83.19	27.57	6.55	100.0	± 9.6 %
		Y	6.56	85.60	28.82		100.0	
		Z	5.08	80.09	26.35	1.00	100.0	1.1
10059- CAB	IEEE 802.11b WiFi 2.4 GHz (DSSS, 2 Mbps)	X	1.25	66.85	16.51	0.61	110.0	± 9.6 %
		Y	1.32	67.32	16.93		110.0	
-		Z	1.20	65.88	15.84		110.0	
10060- CAB	IEEE 802.11b WiFi 2.4 GHz (DSSS, 5.5 Mbps)	X	100.00	134.06	33.85	1.30	110.0	± 9,6 %
	- istal	Y	100.00	136.81	35.28		110.0	
							110.0	

Certificate No: ES3-3019_Aug18

Page 13 of 39

August 20, 2018

10061- CAB	IEEE 802.11b WiFi 2.4 GHz (DSSS, 11 Mbps)	X	15.36	108.05	30.50	2.04	110.0	±9.6 %
		Y	15.08	108.15	30.91		110.0	-
10062-	IEEE 802.11a/h WIFI 5 GHz (OFDM, 6	Z	6.66	94.63	26.70	-	110.0	-
CAC	Mbps)	×	4.35	67.26	16.60	0.49	100.0	±9.6 %
_		Y	4.41	67.37	16.72		100.0	
10000		Z	4.30	67.18	16,44		100.0	land of the land
10063- CAC	IEEE 802.11a/h WiFi 5 GHz (OFDM, 9 Mbps)	×	4.38	67.45	16.75	0.72	100.0	±9.6 %
		Y	4.44	67.56	16.88		100.0	
Incol		Z	4.34	67.37	16.60		100.0	-
10064- CAC	IEEE 802.11a/h WiFi 5 GHz (OFDM, 12 Mops)	x	4.60	67,64	16.95	0.86	100.0	± 9.6 %
_		Y	4.66	67.75	17.07		100.0	
		Z	4.55	67.56	16.80		100.0	
10065- CAC	IEEE 802.11a/h WiFi 5 GHz (OFDM, 18 Mbps)	X	4.51	67.57	17.11	1,21	100.0	± 9.6 %
		Y	4.57	67.68	17.23	-	100.0	
		Z	4.46	67.48	16.95	-	100.0	
10066- CAC	IEEE 802.11a/h WiFi 5 GHz (OFDM, 24 Mbps)	X	4.54	67.62	17.29	1.46	100.0	± 9.6 %
		Y	4.60	67.74	17.42		100.0	
		Z	4.48	67.51	17.12		100.0	
10067- CAC	IEEE 802.11a/h WiFi 5 GHz (OFDM, 36 Mbps)	×	4,84	68.00	17.84	2.04	100.0	± 9.6 %
		Y	4.91	68.15	17.99		100.0	
		Z	4.77	67.85	17.66		100.0	
10068- CAC	IEEE 802.11a/h WiFi 5 GHz (OFDM, 48 Mbps)	X	4.93	68.11	18.14	2.55	100.0	± 9,6 %
		Y	5.00	68.27	18.29		100.0	
		Z	4.86	68.00	17.98		100.0	
10069- CAC	IEEE 802.11a/h WiFi 5 GHz (OFDM, 54 Mbps)	X	4.97	68.10	18.30	2.67	100.0	± 9.6 %
	and the second se	Y	5.05	68.27	18.47		100.0	-
		Z	4.90	67.95	18.12		100.0	
10071- CAB	IEEE 802 11g WiFi 2.4 GHz (DSSS/OFDM, 9 Mbps)	X	4.76	67.80	17.78	1.99	100.0	±9.6 %
	(Participation and Participation)	Y	4.82	67.92	17.91		100.0	-
		Z	4.71	67.71	17.63		100.0	
10072- CAB	IEEE 802.11g WiFi 2.4 GHz (DSSS/OFDM, 12 Mbps)	X	4.74	68.13	18.04	2.30	100.0	±9.6 %
UND	(bassior bm, 12 mops)	Y	4.81	68.27	18.18		100.0	-
-		Z	4.67	67.99	17.87		100.0	1.1.1.1
10073- CAB	IEEE 802.11g WiFi 2.4 GHz (DSSS/OFDM, 18 Mbps)	X	4.86	68.55	18.52	2.83	100.0	± 9.6 %
		Y	4.94	68.72	18.68		100.0	
		Z	4.79	68.40	18.35		100.0	
10074- CAB	IEEE 802.11g WiFi 2.4 GHz (DSSS/OFDM, 24 Mbps)	×	4,92	68.71	18,79	3.30	100.0	±9.6 %
	Contraction of the second seco	Y	5,01	68.89	18,96	-	100.0	
		Z	4.85	68.56	18.63		100.0	
10075- CAB	IEEE 802.11g WiFi 2.4 GHz (DSSS/OFDM, 36 Mbps)	X	4.99	68.87	19.13	3.82	90.0	±9.6 %
	The second secon	Y	5.08	69.07	19.31		90.0	
		Z	4.91	68.70	18.96	-	90.0	-
10076- CAB	IEEE 802.11g WiFi 2.4 GHz (DSSS/OFDM, 48 Mbps)	x	5.04	68.76	19.33	4.15	90.0	±9.6 %
	The same strength in the part of	Y	5.14	69.00	19.53		90.0	
		z	4.97	68.59	19.16		90.0	
10077-	IEEE 802.11g WiFi 2.4 GHz	X	5.09	68.91	19.48	4.30	90.0	±9.6 %
CAR								
CAB	(DSSS/OFDM, 54 Mbps)	Y	5.19	69.15	19.69		90.0	

Certificate No: ES3-3019_Aug18

Page 14 of 39

August 20, 2018

10081- CAB	CDMA2000 (1xRTT, RC3)	X	0.33	60.00	5.42	0.00	150.0	± 9.6 %
or us		Y	0.37	60.39	6.25		150.0	
		Z	0.32	60.00	5.39		150.0	
10082- CAB	IS-54 / IS-136 FDD (TDMA/FDM, PI/4- DQPSK, Fullrate)	×	0.87	60.00	4.66	4.77	80.0	±9.6 %
		Y	0.92	60.00	5.02		80.0	
_		Z	0.79	60.00	4.55		80.0	
10090- DAC	GPRS-FDD (TDMA, GMSK, TN 0-4)	х	100.00	112.50	25.87	6.56	60.0	± 9.6 %
		Y.	100.00	116.14	27.80		60.0	
		Z	100.00	113.84	26.29		60.0	
10097- CAB	UMTS-FDD (HSDPA)	X	1.68	68.74	14.96	0.00	150.0	±9.6 %
		Y	1.80	69.62	15.63		150.0	
		Z	1.59	67.84	14.37		160,0	
10098- CAB	UMTS-FDD (HSUPA, Subtest 2)	x	1.64	68.69	14.95	0.00	150.0	± 9.6 %
		Y	1.77	69.59	15,63		150.0	
		Z	1.56	67.78	14.35		150.0	
10099- DAC	EDGE-FDD (TDMA, 8PSK, TN 0-4)	x	14.55	104.92	37.67	9.56	60.0	± 9.6 %
		Y	21.65	116.02	41.81		60.0	
	111	Z	10.99	98.46	35.57	1000	60.0	
10100- CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	x	2.75	69.89	16.53	0.00	150.0	±9.6 %
		Y	2.86	70.39	16.89		150.0	
		Z	2.65	69.26	16.15		150.0	
10101- CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, 16-QAM)	×	2.92	67.35	15,74	0.00	150.0	± 9.6 %
		Y	2.99	67.63	15.98		150.0	
		Z	2.87	67.07	15.50		150.0	
10102+ CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, 64-QAM)	х	3.03	67.40	15.86	0.00	150.0	± 9.6 %
		Y	3.09	67.64	16.07		150.0	
-		Z	2.98	67.15	15.62		150.0	1
10103- CAF	LTE-TDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	X	7.90	81.11	22.78	3.98	65.0	± 9.6 %
		Y	7.81	80.69	22.73		65.0	
		Z	7.06	79.54	22.21		65.0	
10104- CAF	LTE-TDD (SC-FDMA, 100% RB, 20 MHz, 16-QAM)	X	6.89	76.73	21.71	3.98	65.0	±9.6 %
67.0	7777.04	Y	7.12	77.18	22.01		65.0	
		Z	6.50	75.95	21.35	A	65.0	
10105- CAF	LTE-TDD (SC-FDMA, 100% RB, 20 MHz, 64-QAM)	X	6.51	75.50	21.48	3.98	65.0	± 9.6 %
		Y	6.52	75.32	21.50		65.0	
-		Z	6.50	75.81	21.59	1000	65.0	1
10108- CAF	LTE-FDD (SC-FDMA, 100% RB, 10 MHz, QPSK)	x	2.35	69.45	16.36	0.00	150.0	± 9,6 %
		Y	2.45	69.99	16.76		150.0	
		Z	2.25	68.74	15.91		150.0	
10109- CAF	LTE-FDD (SC-FDMA, 100% RB, 10 MHz, 16-QAM)	X	2.55	67.44	15.50	0.00	150.0	± 9.6 %
	Contraction of the state	Y	2.63	67.76	15.78		150.0	
		Z	2.50	67.09	15.19		150.0	-
10110- CAF	LTE-FDD (SC-FDMA, 100% RB, 5 MHz, QPSK)	X	1.83	68.67	15.52	0.00	150.0	± 9.6 %
		Y	1.94	69.40	16.07		150.0	1.00
		Z	1.74	67.81	14,96	-	150.0	
10111- CAF	LTE-FDD (SC-FDMA, 100% RB, 5 MHz, 16-QAM)	X	2.27	68.67	15.34	0.00	150.0	±9.6.%
		Y	2.35	69.06	15.70		150.0	-
		2	2.18	68.08	14.88		150.0	

Certificate No: ES3-3019_Aug18

Page 15 of 39

August 20, 2018

10112- CAF	LTE-FDD (SC-FDMA, 100% RB, 10 MHz, 64-QAM)	×	2.68	67.57	15.60	0.00	150.0	± 9.6 %
		Y	2.75	67.85	15.85		150.0	-
		Z	2.62	67.25	15.31		150.0	
10113- CAF	LTE-FDD (SC-FDMA, 100% RB, 5 MHz, 64-QAM)	×	2.40	68.84	15,47	0.00	150.0	± 9.6.%
		Y	2.48	69.17	15,79		150.0	-
		Z	2.31	68.27	15.02		150.0	
10114- CAC	IEEE 802.11n (HT Greenfield, 13.5 Mbps, BPSK)	×	4.78	67.32	16.48	0.00	150.0	±9.6%
		Y	4.84	67.44	16.60		150.0	
		Z	4.74	67.21	16.33		150.0	
10115- CAC	IEEE 802.11n (HT Greenfield, 81 Mbps, 16-QAM)	×	5.03	67.47	16,54	0.00	150.0	± 9,6 %
_		Y	5.08	67.57	16.65		150.0	
		Z	4.98	67.38	16.40		150.0	
10116- CAC	IEEE 802.11n (HT Greenfield, 135 Mbps, 64-QAM)	×	4.86	67,52	16.51	0.00	150.0	± 9.6 %
		Y	4.91	67.64	16.62		150.0	
_		Z	4.81	67,41	16.35		150.0	
10117- CAC	IEEE 802.11n (HT Mixed, 13.5 Mbps, BPSK)	x	4.76	67.19	16.44	0.00	150.0	± 9.6 %
		Y	4.81	67.30	16.55	1	150.0	
		Z	4.72	67.12	16.30		150.0	
10118- CAC	IEEE 802.11n (HT Mixed, 81 Mbps, 16- QAM)	×	5.08	67.58	16.61	0.00	150.0	± 9.6 %
-		Y	5.14	67.72	16.74		150.0	-
		Z	5.01	67.43	16.43	1	150.0	
10119- CAC	IEEE 802.11n (HT Mixed, 135 Mbps, 64- QAM)	X	4.86	67.54	16.52	0.00	150.0	± 9.6 %
		Y	4.92	67.66	16.64		150.0	
		Z	4.82	67.43	16.37		150.0	
10140- CAE	LTE-FDD (SC-FDMA, 100% RB, 15 MHz, 16-QAM)	x	3.03	67.44	15.76	0.00	150.0	± 9.6 %
-		Y	3.10	67.70	15.99		150.0	
		Z	2.98	67.18	15.52	-	150.0	
10141- CAE	LTE-FDD (SC-FDMA, 100% RB, 15 MHz, 64-QAM)	x	3.16	67.67	15.99	0.00	150.0	± 9.6 %
		Y	3.23	67.88	16.18	-	150.0	
-		Z	3.11	67.44	15.76		150.0	
10142- CAE	LTE-FDD (SC-FDMA, 100% RB, 3 MHz, QPSK)	x	1.50	67.65	13.86	0.00	150.0	±9.6 %
		Y	1.64	68.71	14.65	L	150.0	
		Z	1.40	66.66	13.21		150.0	
10143- CAE	LTE-FDD (SC-FDMA, 100% RB, 3 MHz, 16-QAM)	x	1.79	66.97	12.89	0.00	150.0	±9.6 %
		Y	1.94	67.90	13.60		150.0	
		Z	1.68	66.15	12.29	-	150.0	
10144- CAE	LTE-FDD (SC-FDMA, 100% RB, 3 MHz, 64-QAM)	×	1.50	64.00	10.77	0.00	150.0	± 9.6 %
		Y	1.60	64.68	11.38		150.0	
		Z	1.43	63.53	10.33	-	150.0	1
10145- CAF	LTE-FDD (SC-FDMA, 100% RB, 1.4 MHz, QPSK)	×	0.50	60.00	5.08	0.00	150.0	±9.6 %
_		Y	0.53	60.00	5.45	-	150.0	
		Z	0.49	60.00	4.93		150,0	
10146- CAF	LTE-FDD (SC-FDMA, 160% RB, 1.4 MHz, 16-QAM)	×	0,66	60,00	4,66	0.00	150.0	±9.6 %
		Y	0.67	60.00	5.08		150.0	
		Z	0.67	60.00	4,44		150.0	
10147- CAF	LTE-FDD (SC-FDMA, 100% RB, 1.4 MHz, 64-QAM)	x	0.67	60.00	4.71	0.00	150.0	± 9.6 %
		Y	0.60	58.92	4.43		150.0	
		Z		60.00				

Certificate No: ES3-3019_Aug18

Page 16 of 39

August 20, 2018

10149- CAE	LTE-FDD (SC-FDMA, 50% RB, 20 MHz, 16-QAM)	X	2.56	67.52	15.55	0.00	150.0	±9.6 %
Lary yes	to writing	Y	2.64	67.83	15.83	-	150.0	
		Z	2.51	67.16	15.25		150.0	1000
10150- CAE	LTE-FDD (SC-FDMA, 50% RB, 20 MHz, 64-0AM)	X	2.69	67.64	15.65	0.00	150,0	±9.6 %
		Y	2.76	67.91	15.90		150.0	-
		Z	2.63	67.32	15.36	1.00	150.0	
10151- CAF	LTE-TDD (SC-FDMA, 50% RB, 20 MHz, QPSK)	X	9.73	86.59	24.67	3.98	65.0	± 9.6 %
		Y	9.80	86.53	24.79		65.0	
		Z	8.83	85.34	24.25		65.0	
10152- CAF	LTE-TDD (SC-FDMA, 50% RB, 20 MHz, 16-QAM)	×	6.54	77.11	21.21	3.98	65.0	±9.6 %
		Y	6.78	77.59	21.55		65.0	
		Z	6.11	76.24	20.80	-	65.0	-
10153- CAF	LTE-TDD (SC-FDMA, 50% RB, 20 MHz, 64-QAM)	×	7.17	78.82	22.30	3.98	65.0	±9.6 %
		Y	7.34	79.00	22.49		65.0	
		Z	6.72	77.97	21.92		65.0	100
10154- CAF	LTE-FDD (SC-FDMA, 50% RB, 10 MHz, QPSK)	x	1.87	69.05	15.75	0.00	150.0	± 9.6 %
		Y	1.98	69.74	16.28		150.0	
		Z	1.77	68.14	15.17		150.0	
10155- CAF	LTE-FDD (SC-FDMA, 50% RB, 10 MHz, 16-QAM)	x	2.28	68.74	15.39	0.00	150.0	± 9.6 %
		Y	2.36	69.12	15.74	-	150.0	
		Z	2.19	68,15	14.93	-	150.0	1
10156- CAF	LTE-FDD (SC-FDMA, 50% RB, 5 MHz, OPSK)	X	1.22	66.13	12.34	0.00	150.0	± 9,6 %
1411 C		Y	1.36	67.41	13.29		150.0	
		Z	1,14	65.20	11.69	1.02	150.0	1.
10157- CAF	LTE-FDD (SC-FDMA, 50% RB, 5 MHz. 16-QAM)	x	1,22	63.09	9.61	0.00	150.0	± 9.6 %
		Y	1.33	63.90	10.34		150.0	
		Z	1.15	62.61	9.16		150.0	
10158- CAF	LTE-FDD (SC-FDMA, 50% RB, 10 MHz, 64-QAM)	x	2.41	68.95	15.54	0.00	150.0	±9.6 %
		Y	2.49	69,27	15.85		150.0	
		Z	2.32	68.37	15.09		150.0	
10159- CAF	LTE-FDD (SC-FDMA, 50% RB, 5 MHz, 64-QAM)	X	1.25	63.14	9.66	0.00	150.0	± 9.6 %
		Y	1.36	63.97	10.39		150.0	
-		Z	1.18	62.64	9.19		150.0	
10160- CAE	LTE-FDD (SC-FDMA, 50% RB, 15 MHz, QPSK)	×	2.40	68.97	16.11	0.00	150.0	± 9.6 %
		Y	2.50	69.46	16.49		150.0	
A		Z	2.30	68.28	15.65		150.0	1
10161- CAE	LTE-FDD (SC-FDMA, 50% RB, 15 MHz, 16-QAM)	×	2.56	67.57	15.39	0.00	150.0	± 9.6 %
		Y	2.63	67.87	15.67		150.0	
	A CONTRACTOR OF THE OWNER	Z	2.50	67.21	15.07	1.1.1.1	150.0	
10162- CAE	LTE-FDD (SC-FDMA, 50% RB, 15 MHz, 64-QAM)	×	2.66	67.86	15,56	0.00	150.0	± 9.6 %
		Y.	2.74	68.13	15.82		150.0	
	The second se	Z	2.60	67.51	15.24		150.0	
10166- CAF	LTE-FDD (SC-FDMA, 50% RB, 1.4 MHz, QPSK)	×	2.73	67.85	18,46	3.01	150.0	± 9,6 %
-		Y	2.78	68,27	19.13		150.0	
		Z	2.81	68,48	18.76		150.0	
10167- CAF	LTE-FDD (SC-FDMA, 50% RB, 1.4 MHz, 16-QAM)	x	3.00	69.93	18.59	3.01	150.0	± 9.6 %
		Y	2.96	70.40	19.40		150.0	
			3.25	71.67	19.34		150.0	

Certificate No: ES3-3019_Aug18

Page 17 of 39

August 20, 2018

10168- CAF	LTE-FDD (SC-FDMA, 50% RB, 1.4 MHz, 64-QAM)	x	3.36	72.56	20.25	3.01	150.0	± 9.6 %
		Y	3.26	72.65	20.86		150.0	
_		Z	3.76	74.98	21.27		150.0	-
10169- CAE	LTE-FDD (SC-FDMA, 1 RB, 20 MHz, QPSK)	x	2.24	65.89	17.49	3.01	150.0	± 9.6 %
		Y	2.17	65.69	17.95	-	150.0	
		Z	2.39	67.30	18.17		150.0	
10170- CAE	LTE-FDD (SC-FDMA, 1 RB, 20 MHz, 16-QAM)	x	2.58	69.82	19.29	3.01	150.0	± 9.6 %
_		Ŷ	2.31	68.87	19.55	-	150.0	
		Z	3.10	73.74	20.99		150.0	
10171- AAE	LTE-FDD (SC-FDMA, 1 RB, 20 MHz, 64-QAM)	×	2.20	66.63	16.66	3.01	150.0	= 9.6 %
		Y	2.04	66.46	17.32		150.0	
	and the second sec	Z	2.49	69,12	17.74	1.000	150.0	
10172- CAF	LTE-TDD (SC-FDMA, 1 RB, 20 MHz, QPSK)	x	6.57	90.42	29.01	6.02	65.0	± 9.6 %
-		Y	6.70	92.44	30.69		65.0	
-		Z	5.30	87.29	28.20		65.0	
10173- CAF	LTE-TDD (SC-FDMA, 1 RB, 20 MHz, 16-QAM)	x	16.42	104.67	31.45	6.02	65.0	± 9.6 %
		Y	22.05	113,86	35.25		65.0	1
-		Z	42.83	124.16	36.91	1	65.0	
10174- CAF	LTE-TDD (SC-FDMA, 1 RB, 20 MHz, 64-QAM)	×	10,59	95.52	28.04	6.02	65.0	± 9.6 %
1		Y	12.26	101.14	30.88		65.0	
		Z	27.69	114.1B	33.55		65.0	1
10175- CAF	LTE-FDD (SC-FDMA, 1 RB, 10 MHz, QPSK)	x	2.22	65.67	17.27	3.01	150.0	± 9.6 %
		Y	2.15	65.53	17.77		150.0	
		Z	2.36	67.03	17.92	-	150.0	
10176- CAF	LTE-FDD (SC-FDMA, 1 RB, 10 MHz, 16-QAM)	x	2.58	69.84	19.30	3.01	150.0	± 9.6 %
		Y	2.31	68.89	19.56	-	150.0	
		Z	3.11	73.77	21.00		150.0	
10177- CAH	LTE-FDD (SC-FDMA, 1 RB, 5 MHz, QPSK)	x	2.23	65.75	17,33	3.01	150.0	± 9.6 %
		Y	2.16	65.60	17.82		150.0	
		Z	2.37	67.12	17.98	-	150.0	1
10178- CAF	LTE-FDD (SC-FDMA, 1 RB, 5 MHz, 16- QAM)	x	2.57	69.75	19.24	3.01	150.0	± 9.6 %
		Y	2.31	68.83	19.52		150.0	
		Z	3.09	73.64	20.92	1.0	150.0	1
10179- CAF	LTE-FDD (SC-FDMA, 1 RB, 10 MHz, 64-QAM)	x	2.36	68.14	17.85	3.01	150.0	±9.6 %
		Y	2.16	67.70	18.38		150.0	
		Z	2.76	71.27	19.21		150.0	
10180- CAF	LTE-FDD (SC-FDMA, 1 RB, 5 MHz, 64- QAM)	×	2.20	66.62	16.64	3.01	150.0	±9.6 %
		Y	2.04	66.46	17.31		150.0	
		Z	2.49	69.10	17.72		150.0	
10181- CAE	LTE-FDD (SC-FDMA, 1 RB, 15 MHz, QPSK)	x	2.22	65.74	17.33	3.01	150.0	± 9.6 %
		Y	2.16	65.59	17.81		150.0	
		Z	2.37	67.11	17.98	1.27	150.0	-
10182- CAE	LTE-FDD (SC-FDMA, 1 RB, 15 MHz, 16-QAM)	x	2.57	69.73	19.23	3.01	150.0	±9.6 %
		Y	2.30	68.81	19.50		150.0	
		Z	3.09	73.61	20.91		150.0	-
10183- AAD	LTE-FDD (SC-FDMA, 1 RB, 15 MHz, 64-QAM)	x	2.20	66.60	16.63	3,01	150.0	±9.6 %
		Y	2.04	66.44	17.30		150.0	
		Z	2.48	69.08	17.71		150.0	

Certificate No: ES3-3019_Aug18

Page 18 of 39

August 20, 2018

10184- CAE	LTE-FDD (SC-FDMA, 1 RB, 3 MHz, OPSK)	X	2.23	65.77	17.34	3.01	150.0	±9.6 %
CAN OF	arony	Y	2.16	65.62	17.83		150.0	
		Z	2.38	67.14	18.00		150.0	1.1.1.1
10185- CAE	LTE-FDD (SC-FDMA, 1 RB, 3 MHz, 16- QAM)	X	2.58	69.79	19.27	3,01	150.0	± 9,6 %
		Y	2.31	68.86	19.54		150.0	
		Z	3.10	73.69	20.95	-	150.0	
10186- AAE	LTE-FDD (SC-FDMA, 1 RB, 3 MHz, 64- QAM)	×	2.20	66.64	16.66	3.01	150.0	± 9.6.%
		Y	2.04	66.4B	17.33	-	150.0	
		Z	2.50	69.14	17.74		150.0	1
10187- CAF	LTE-FDD (SC-FDMA, 1 RB, 1.4 MHz, QPSK)	x	2.24	65.85	17.43	3.01	150.0	± 9.6 %
		Y	2.17	65.68	17.91	-	150.0	
		Z	2.39	67.25	18.10		150.0	
10188- CAF	LTE-FDD (SC-FDMA, 1 RB, 1.4 MHz, 16-QAM)	X	2.63	70.23	19.57	3.01	150.0	± 9.6 %
	in a my	Y	2.35	69.19	19.79		150.0	
		Z	3.20	74.34	21.34		150.0	
10189- AAF	LTE-FDD (SC-FDMA, 1 RB, 1.4 MHz, 64-QAM)	x	2.24	66.93	16.89	3,01	150.0	± 9.6 %
		Y	2.07	66.73	17.54		150.0	
		Z	2,55	69.53	18.02		150.0	
10193- CAC	IEEE 802.11n (HT Greenfield, 6.5 Mbps, BPSK)	x	4.17	67.12	16.09	0.00	150.0	± 9.6 %
40.14		Y	4.22	67.24	16.23	-	150.0	
		Z	4.13	67.06	15.95		150.0	
10194- CAC	IEEE 802.11n (HT Greenfield, 39 Mbps, 16-QAM)	х	4.28	67.28	16.23	0.00	150.0	± 9.6.%
		Y	4.34	67.40	16.36	1	150.0	
		Z	4.24	67.20	16.08		150.0	-
10195- CAC	IEEE 802.11n (HT Greenfield, 65 Mbps, 64-QAM)	×	4.31	67.25	16.23	0.00	150.0	±9.6 %
		Y	4.37	67.38	16.36		150.0	
		Z	4.26	67.17	16.08		150.0	
10196- CAC	IEEE 802.11n (HT Mixed, 6.5 Mbps, BPSK)	×	4.14	67.05	16.05	0.00	150.0	± 9.6 %
		Y	4.20	67.18	16.19		150.0	_
		Z	4.10	67.00	15.90		150.0	
10197+ CAC	IEEE 802.11n (HT Mixed, 39 Mbps, 16- QAM)	Х	4.29	67.27	16.23	0.00	150.0	± 9.6 %
		Y	4.35	67.39	16.37		150.0	
-		Z	4.25	67.20	16.08		150.0	
10198- CAC	IEEE 802.11n (HT Mixed, 65 Mbps, 64- OAM)	×	4.30	67.24	16:22	0.00	150.0	± 9.6 %
		Y	4.35	67.37	16.36		150.0	
	100000000000000000000000000000000000000	Z	4.25	67.16	16.07		150.0	
10219- CAC	IEEE 802.11n (HT Mixed, 7.2 Mbps, BPSK)	x	4.10	67,13	16.04	0.00	150.0	± 9.6 %
		Y	4.16	67.25	16.18		150.0	
		Z	4.06	67.07	15.89		150.0	-
10220- CAC	IEEE 802.11n (HT Mixed, 43.3 Mbps, 16- QAM)	×	4.28	67.23	16.22	0.00	150.0	± 9.6 %
		Y	4.34	67.35	16.35		150.0	-
		Z	4.24	67.16	16.07		150.0	-
10221- CAC	IEEE 802.11n (HT Mixed, 72.2 Mbps, 64- QAM)	X	4.32	67.21	16.22	0.00	150.0	± 9,6 %
-		Y	4.38	67.33	16.35	-	150.0	
		Z	4.27	67.14	16.07		150.0	1
10222- CAC	IEEE 802.11n (HT Mixed, 15 Mbps, BPSK)	X	4.74	67.21	16.43	0.00	150,0	±9.6 %
		Y	4.79	67.33	16.55		150.0	
		Z	4.70	67.12	16.29		150,0	

Certificate No: ES3-3019_Aug18

Page 19 of 39

August 20, 2018

10223- CAC	IEEE 802.11n (HT Mixed, 90 Mbps, 16- QAM)	×	4.94	67.26	16.45	0.00	150.0	± 9.6 %
-		Y	5.00	67.37	16.57		150.0	
		Z	4.90	67.18	16.32	-	150.0	
10224- CAC	IEEE 802.11n (HT Mixed, 150 Mbps, 64- QAM)	X	4.78	67.35	16.43	0.00	150.0	± 9.6 %
		Y	4.83	67.47	16.55		150.0	
		Z	4.74	67.26	16.29		150.0	
10225- CAB	UMTS-FDD (HSPA+)	×	2.38	66.08	14.04	0.00	150.0	± 9.6 %
		Y	2.46	66.40	14,36	1	150.0	-
	the second s	Z	2.33	65.77	13.68		150.0	
10226- CAA	LTE-TDD (SC-FDMA, 1 RB, 1.4 MHz, 16-QAM)	×	18.80	107.39	32.33	6.02	65.0	± 9.6 %
-		Y	25.18	116.67	36.12		65.0	-
		2	56.86	129.79	38.42		65.0	
10227- CAA	LTE-TOD (SC-FDMA, 1 RB, 1.4 MHz, 64-QAM)	x	17.62	104,50	30.74	6.02	65.0	± 9.6 %
	1.5 8 1 2.2	Y	24.98	114.46	34,69		65.0	
	and the second se	Z	53.65	126.02	36.59		65.0	
10228- CAA	LTE-TDD (SC-FDMA, 1 RB, 1.4 MHz, QPSK)	×	B.84	96.76	31.22	6.02	65.0	± 9.6 %
		Y	10.52	102,26	33.97		65.0	
		Z	8.68	97.67	31.80		65.0	
10229- CAC	LTE-TDD (SC-FDMA, 1 RB, 3 MHz, 16- OAM)	×	16.55	104.80	31.49	6.02	65.0	± 9.6 %
		Y	22.16	113.92	35.27		65.0	
in a		Z	43.51	124.44	36.98		65.0	
10230- CAC	LTE-TDD (SC-FDMA, 1 RB, 3 MHz, 64- QAM)	X	15.38	101.96	29.94	6.02	65.0	±9.6 %
		Y	21.58	111.59	33.84		65.0	
		Z	40.32	120.70	35.19	1	65.0	
10231- CAC	LTE-TDD (SC-FDMA, 1 RB, 3 MHz, QPSK)	x	8.23	95.17	30.61	6.02	65.0	±9.6 %
		Y	9.84	100.68	33.39		65.0	
		Z	8.03	95.88	31.12		65.0	
10232- CAE	LTE-TDD (SC-FDMA, 1 RB, 5 MHz, 16- QAM)	×	16.51	104.77	31.48	6.02	65.0	±9.6 %
		Y	22.12	113.91	35.27		65.0	
	and the second s	Z	43.30	124.37	36.97		65.0	
10233- CAE	LTE-TDD (SC-FDMA, 1 RB, 5 MHz, 64- QAM)	×	15.29	101.88	29.92	6.02	65.0	± 9.6 %
		Y	21.43	111.49	33.81		65.0	_
		Z	39.84	120.51	35.15		65.0	
10234- CAÉ	LTE-TDD (SC-FDMA, 1 RB, 5 MHz, QPSK)	×	7.86	94.06	30.11	6.02	65.0	±9.6 %
		Y	9.46	99.67	32.94		65.0	
Gun	Statement and a statement of the stateme	Z	7.64	94.69	30.59		65.0	
10235- CAE	LTE-TDD (SC-FDMA, 1 RB, 10 MHz, 16-QAM)	×	16.57	104,86	31.51	6.02	65.0	± 9.6 %
		Y	22.25	114.05	35.31		65.0	
		Z	43.57	124.50	37.01		65.0	
10236- CAE	LTE-TOD (SC-FDMA, 1 RB, 10 MHz, 64-QAM)	x	15.56	102.14	29.98	6.02	65.0	± 9.6 %
		Y	22.01	111.92	33.92		65.0	
		Z	41.07	120.98	35.26		65.0	
10237- CAE	LTE-TDD (SC-FDMA, 1 RB, 10 MHz, QPSK)	x	8.24	95.23	30.63	6,02	65.0	± 9.6 %
		Y	9.87	100.80	33.43		65.0	
		Z	8.02	95.91	31.14		65.0	
10238- CAE	LTE-TDD (SC-FDMA, 1 RB, 15 MHz, 16-QAM)	x	16.49	104.76	31.48	6,02	65.0	± 9.6 %
		Y	22.11	113.92	35.28		65.0	

Certificate No: ES3-3019_Aug18

Page 20 of 39

August 20, 2018

10239- CAE	LTE-TDD (SC-FDMA, 1 RB, 15 MHz, 64-QAM)	X	15.23	101.82	29.90	6.02	65.0	±9.6%
		Y	21.33	111.42	33.80		65.0	
		Z	39.55	120.41	35.13		65.0	
10240- CAE	LTE-TDD (SC-FDMA, 1 RB, 15 MHz, QPSK)	х	8.24	95.24	30.63	6.02	65.0	± 9.6 %
Lov Gas	di ony	Y	9.87	100.81	33.44		65.0	
		Z	8.02	95.92	31.14		65.0	
10241- CAA	LTE-TDD (SC-FDMA, 50% RB, 1.4 MHz, 16-QAM)	X	8.99	87.82	28.29	6.98	65.0	± 9.6 %
unn.	Tu-saran)	Y	9.75	90.67	30.00		65.0	1
		Z	9.34	89.67	29.13		65.0	
10242- CAA	LTE-TDD (SC-FDMA, 50% RB, 1,4 MHz, 64-0AM)	X	7.89	85.15	27.21	6.98	65.0	± 9.6 %
Gret	dif dif di	Y	8.15	86.79	28.48		65.D	
		Z	8.92	88.81	28.75	1000	65.0	
10243- CAA	LTE-TDD (SC-FDMA, 50% RB, 1.4 MHz, OPSK)	X	6.23	80.61	26.36	6.98	65.0	± 9.6 %
unn	Grony	Y	6.55	82.23	27.61	-	65.0	
		Z	5.44	78.21	25.48		65.0	
10244-	LTE-TDD (SC-FDMA, 50% RB, 3 MHz.	X	3.42	67.77	12.33	3.98	65.0	± 9.6 %
CAC	16-QAM)	Y	4.29	71.02	14.35		65.0	
		2	3.28	67.61	12.05		65.0	-
10245-	LTE-TDD (SC-FDMA, 50% RB, 3 MHz.	X	3.32	67.21	12.00	3.98	65.0	±9.6 %
CAC	64-QAM)	Y	4.07	70.09	13.87	-	65.0	
		Z	3.17	66.99	11.70		65.0	
10010	LTT TOD ICO FOLL FOR OR ALL.	X	4.43	73.96	15.64	3.98	65.0	±9.6 %
10246- CAC	LTE-TDD (SC-FDMA, 50% RB, 3 MHz, OPSK)	1		1.000	0.000	3.98	1.000	19.0 %
		Y	5.08	75.78	16.72		65.0	
		Z	3.85	72.56	14,95		65.0	
10247- CAE	LTE-TDD (SC-FDMA, 50% RB, 5 MHz, 16-QAM)	×	4.59	72.28	15.77	3.96	65.0	± 9.6 %
		Y	4.91	73.13	16,39		65.0	
		Z	4,19	71.34	15.21		65,0	
10248- CAE	LTE-TDD (SC-FDMA, 50% RB, 5 MHz, 64-QAM)	X	4.34	71.14	15.26	3.98	65.0	± 9.6 %
		1.8	4.65	72.00	15,89		65.0	
		Z	3.96	70.19	14.69		65.0	
10249- CAE	LTE-TDD (SC-FDMA, 50% RB, 5 MHz. OPSK)	X	10.92	88.36	22.53	3.98	65.0	± 9.6 %
Con the	a org	Y	10.64	88.09	22.75		65.0	
		Z	9.25	86.32	21.79		65.0	
10250- CAE	LTE-TDD (SC-FDMA, 50% RB, 10 MHz, 16-QAM)	×	7.53	81.41	22.33	3.98	65.0	± 9.6 %
	and the second	Y	7.46	80.99	22.30		65.0	
-		Z	6.98	80.48	21.90		65.0	
10251- CAE	LTE-TDD (SC-FDMA, 50% RB, 10 MHz, 64-QAM)	X	6.21	76.79	20.02	3.98	65.0	± 9.6 %
Sec. 100		Y	6.42	77.16	20.35		65.0	
		Z	5.76	75.88	19.56		65.0	
10252-	LTE-TDD (SC-FDMA, 50% RB, 10 MHz,	X	13.32	93.66	26.46	3.98	65.0	± 9.6 %
CAE	OPSK)	Y	12.73	92.76	26.36		65.0	
		Z	11.46	91.61	25.80		65.0	
10253-	LTE-TDD (SC-FDMA, 50% RB, 15 MHz.	X	6.41	76.60	20.76	3.98	65.0	± 9.6 %
CAE	16-QAM)	Y	6.65	77.06	21.12	-	65.0	-
-		Z	6.00	75.76	20.37	-	65.0	-
10254-	LTE-TDD (SC-FDMA, 50% RB, 15 MHz,	X	6.93	77.96	21.66	3.98	65.0	±9.6 %
10254- CAE	64-QAM)	100	7177	0.076		0.00		1.0.01
		Y	7,09	78.16	21.87		65.0 65.0	-
		Z	6.50	77.15	21.27	1	0.00	1

Certificate No: ES3-3019_Aug18

Page 21 of 39

August 20, 2018

10255- CAE	LTE-TDD (SC-FDMA, 50% RB, 15 MHz, OPSK)	X	9.17	85.70	24.40	3.98	65.0	±9.6.%
_		Y	9.28	85.72	24.56	-	65.0	
10256-	LTE-TDD (SC-FDMA, 100% RB, 1.4	Z	8.31	84,42	23.94	-	65.0	
CAA	MHz, 16-QAM)	×	2.33	63.37	8.84	3.98	65.0	±9.6 %
		Y	2.70	65.07	10.21		65.0	
		Z	2.17	63.01	8.43	1	65.0	
10257- CAA	LTE-TDD (SC-FDMA, 100% RB, 1.4 MHz, 64-QAM)	x	2.30	63.01	8.53	3.98	65.0	± 9.6 %
		Y	2.62	64.47	9.78	1	65.0	
_		Z	2.14	62.65	8.12		65.0	
10258- CAA	LTE-TDD (SC-FDMA, 100% RB, 1.4 MHz, QPSK)	x	2.37	65.48	10.54	3.98	65.0	± 9.6 %
		Y	2.73	66.98	11.62		65.0	
		Z	2.13	64.74	10.00		65.0	1
10259- CAC	LTE-TDD (SC-FDMA, 100% RB, 3 MHz, 16-QAM)	x	5.73	75.91	18.25	3.98	65.0	± 9.6 %
		Y	5.92	76.29	18.63		65.0	
		Z	5.25	74.96	17.74	-	65.0	
10260- CAC	LTE-TDD (SC-FDMA, 100% RB, 3 MHz. 64-QAM)	X	5.61	75.26	17.97	3,98	65.0	± 9.6 %
1		Y	5.82	75.68	18.36	-	65.0	
		Z	5.16	74.34	17.46		65.0	
10261- CAC	LTE-TDD (SC-FDMA, 100% RB, 3 MHz, QPSK)	X	11.22	89.48	23.72	3,98	65.0	±9.6 %
		Y	10.83	88.93	23.80		65.0	
	-	Z	9.64	87.55	23.03	-	65.0	
10262- CAE	LTE-TDD (SC-FDMA, 100% RB, 5 MHz, 16-QAM)	X	7,47	81.24	22.23	3.98	65.0	= 9.6 %
		Y	7.42	80.85	22.22		65.0	
		Z	6.92	80.30	21.81		65.0	
10263- CAE	LTE-TDD (SC-FDMA, 100% RB, 5 MHz, 64-QAM)	x	6.20	76.77	20.02	3.98	65.0	± 9.6 %
		Y	6.41	77.14	20.34		65.0	-
		Z	5.75	75.86	19.56		65.0	
10264- CAE	LTE-TDD (SC-FDMA, 100% RB, 5 MHz, QPSK)	×	12.97	93.14	26.27	3.98	65.0	±9.6 %
		Y.	12.47	92.36	26.20		65.0	
		Z	11.18	91.12	25.61		65.0	
10265- CAE	LTE-TDD (SC-FDMA, 100% RB, 10 MHz, 16-QAM)	X	6.54	77.12	21.22	3.98	65.0	± 9.6 %
		Y	6.78	77.60	21.56		65.0	
		Z	6,11	76.24	20.81		65.0	
10266- CAE	LTE-TDD (SC-FDMA, 100% RB, 10 MHz, 64-QAM)	X	7.17	78.80	22.29	3.98	65.0	±9.6 %
		Y	7.33	78.99	22.4B		65.0	
-		Z	6,72	77.95	21.91		65.0	
10267- CAE	LTE-TDD (SC-FDMA, 100% RB, 10 MHz, QPSK)	×	9.68	86.48	24,63	3.98	65.0	± 9.6 %
1.11	12.00	Ŷ	9.76	86.44	24.75	_	65.0	
		Z	8.79	85.24	24.21		65.0	
10268- CAE	LTE-TDD (SC-FDMA, 100% RB, 15 MHz, 16-QAM)	x	7.07	76.76	21.77	3.98	65.0	±9.6 %
		Y	7.29	77.15	22.04		65.0	
-		Z	6.70	76.06	21.43	1.000	65.0	
10269- CAE	LTE-TDD (SC-FDMA, 100% RB, 15 MHz, 64-QAM)	x	7.03	76.29	21.59	3.98	65.0	± 9.6 %
		Y	7.24	76.69	21.87		65.0	
	No como o Service Service de	Z	6.68	75.64	21.27	1	65.0	
10270- CAE	LTE-TDD (SC-FDMA, 100% RB, 15 MHz, QPSK)	x	8.04	80.94	22.90	3.98	65.0	± 9.6 %
		Y	8.18	81.04	23.04		65.0	

Certificate No: ES3-3019_Aug18

Page 22 of 39

August 20, 2018

10274- CAB	UMTS-FDD (HSUPA, Subtest 5, 3GPP Rel8.10)	X	2.26	66.83	14.19	0.00	150.0	±9.6%
		Y	2.35	67.29	14.60	-	150.0	
_		Z	2.19	66.41	13.80		150,0	
10275- CAB	UMTS-FDD (HSUPA, Subtest 5, 3GPP Rel8,4)	×	1.38	68.12	14.71	0.00	150.0	±9.6 %
Ser yes	(NUMBER)	Y	1.51	69.22	15.51	-	150.0	
		Z	1.30	67.09	14.07	-	150.0	
10277-	PHS (QPSK)	X	2.05	61.14	6.20	9.03	50.0	± 9.6 %
CAA	Fh3 (dF3R)	Y	2.35	62.20	7.24		50.0	
						_	50.0	-
		Z	1.85	60.65	5.69	0.00		
10278- CAA	PHS (QPSK: BW 884MHz, Rolloff 0.5)	X	3.29	66.09	10.86	9.03	50.0	±9.6%
		Y	3,79	67.79	12.18		50.0	1
		Z	3.05	65.55	10.36		50.0	
10279- CAA	PHS (QPSK, BW 884MHz, Rolloff 0.38)	x	3.33	66.19	10.97	9.03	50.0	± 9.6 %
GAM		Y	3.82	67.86	12.26		50.0	-
		Z	3.02	65.64	10.46		50.0	
10000	CONTRACTOR DOL COST T. A Date	X	0.45	60.00	5.81	0.00	150.0	± 9.6 %
10290- AAB	CDMA2000, RC1, SO55, Full Rate				and the	0.00		1 0.0 %
		Y	0.54	61.10	6.94		150.0	
-	The state of the s	Z	0.44	60.00	5.71		150.0	
10291- AAB	CDMA2000, RC3, SO55, Full Rate	x	0.32	60.00	5.40	0.00	150,0	± 9.6 %
		Y	0.36	60.31	6.19		150.0	
		Z	0.32	60.00	5.37		150.0	
10292- AAB	CDMA2000, RC3, SO32, Full Rate	X	0.31	60.00	5.66	0.00	150.0	± 9.6 %
AND		Y	0.42	62.09	7.53		150.0	
_		Z	0.31	60.00	5.64		150.0	
10000		X	0.40	61.64	7.04	0.00	150.0	± 9.6 %
10293- AAB	CDMA2000, RC3, SO3, Full Rate					0.00	10.00	1 3.0 7
		Y	D.72	66.90	10.33		150.0	
	A CONTRACTOR OF A CONTRACTOR OFTA CONTRACTOR O	Z	0.39	61.43	6.90		150.0	
10295- AAB	CDMA2000, RC1, SO3, 1/8th Rate 25 fr.	x	100.00	115.65	29.78	9.03	50.0	± 9.6 %
1.12.162		Y	100.00	117,45	30.86		50.0	
		Z	100.00	115.39	29.51		50.0	
10297- AAD	LTE-FDD (SC-FDMA, 50% RB, 20 MHz, QPSK)	x	2.36	69.57	16.44	0.00	150.0	± 9.6 %
MU	MP DNI	Y	2.47	70.10	16.83		150.0	-
		Z	2.27	68.85	15.99		150.0	
10298-	LTE-FDD (SC-FDMA, 50% RB, 3 MHz,	X	0.67	61.00	7.40	0.00	150.0	± 9.6 %
AAD	QPSK)	Y	0.76	61.94	8.33	-	150.0	
	-					-		-
	A DES DESCRIPTION OF DESCRIPTION OF DESCRIPTION	Z	0.64	60.66	7.04	0.00	150.0	1000
10299- AAD	LTE-FDD (SC-FDMA, 50% RB, 3 MHz, 16-QAM)	×	0.84	60.11	6.12	0.00	150.0	± 9.6 %
		Y.	0.94	61.36	7.42		150.0	
		Z	0.84	60.07	5.83		150.0	1 march
10300- AAD	LTE-FDD (SC-FDMA, 50% RB, 3 MHz, 64-QAM)	X	0.73	59.08	4.84	0.00	150.0	± 9.6 %
	a de sul	Y	0.79	59.77	5.75		150.0	
		Z	0.72	58.92	4.50		150.0	
10301-	IEEE 802.16e WiMAX (29:18, 5ms, 10MHz, QPSK, PUSC)	X	4.80	68.05	17.95	4.17	80.0	±9.6 °
AAA	IUMINZ, UPSN, PUSU	Y	5.02	68.71	18.39	-	80.0	-
						-	and the second division of the second divisio	-
		Z	4.55	67.08	17.32	1.00	80.0	10.00
10302- AAA	IEEE 802 16e WiMAX (29:18, 5ms, 10MHz, QPSK, PUSC, 3 CTRL symbols)	×	5.05	67.52	18.12	4.96	0.08	± 9.6 °
		Y	5.33	68.60	18.81	_	80.0	
							80.0	

Certificate No: ES3-3019_Aug18

Page 23 of 39

August 20, 2016

10303- AAA	IEEE 802.16e WIMAX (31:15, 5ms, 10MHz, 64QAM, PUSC)	×	4.87	67.42	17.96	4.96	0.08	± 9.6 %
		Y	5.17	68.58	18.70		80.0	
		Z	4.71	66.92	17.58		80.0	
10304- AAA	IEEE 602.16e WIMAX (29:18, 5ms, 10MHz, 64QAM, PUSC)	×	4.66	67.21	17.45	4.17	80.0	± 9.6 %
		Y	4.92	68.23	18.08		80.0	-
	and the second second second	Z	4.51	66.75	17.09		80.0	-
10305- AAA	IEEE 802.16e WiMAX (31:15, 10ms, 10MHz, 64QAM, PUSC, 15 symbols)	X	5.30	72.82	20.07	6.02	50.0	±9.6 %
		Y	6.32	76.30	21.85		50.0	
		Z	4.76	70.90	19.01		50.0	
10306- AAA	IEEE 802.16e WIMAX (29:18, 10ms, 10MHz, 64QAM, PUSC, 18 symbols)	x	5.08	70,37	19.67	6.02	50.0	±9.6 %
_		Y	5.60	72.42	20.88		50.0	
	and the second se	Z	4.78	69.24	18.95		50.0	
10307- AAA	IEEE 802.16e WIMAX (29:18, 10ms, 10MHz, QPSK, PUSC, 18 symbols)	×	5.03	70.60	19.63	6.02	50.0	±9.6 %
		Y	5.59	72.82	20.90		50.0	
		Z	4.70	69.37	18.87		50.0	
10308- AAA	IEEE 802.16e WIMAX (29:18, 10ms, 10MHz, 16QAM, PUSC)	×	5.07	71.06	19.88	6.02	50.0	± 9.6 %
		Y	5.69	73.45	21.23		50.0	
		Z	4.72	69.74	19.09		50.0	
10309- AAA	IEEE 802.16e WiMAX (29:18, 10ms, 10MHz, 16QAM, AMC 2x3, 18 symbols)	x	5.08	70,41	19.77	6,02	50.0	± 9.6 %
		Y	5.60	72.49	20.99		50.0	
		Z	4.78	69.29	19.05		50.0	
10310- AAA	IEEE 802.16e WiMAX (29:18, 10ms, 10MHz, QPSK, AMC 2x3, 18 symbols)	x	5.08	70.63	19.76	6.02	50.0	± 9.6 %
		Y	5.64	72.78	21.01		50.0	-
		Z	4.77	69.45	19.03		50.0	
10311- AAD	LTE-FDD (SC-FDMA, 100% RB, 15 MHz, QPSK)	X	2.71	68.65	16.12	0.00	150.0	± 9.6 %
		Y	2.82	69.11	16.45		150.0	
-		Z	2.61	68.03	15.74	1	150.0	
10313- AAA	IDEN 1:3	X	10,35	86.29	20.64	6.99	70.0	± 9.6 %
		Y	11.44	88.28	21.75	-	70.0	
-		Z	11.02	88.24	21.45		70.0	
10314- AAA	IDEN 1:6	X	62.09	120.10	33.12	10.00	30.0	± 9.6 %
		Y	24.40	106.12	30.15		30.0	-
		Z	80.12	126.18	34.93		30.0	-
10315- AAB	IEEE 802.11b WiFi 2.4 GHz (DSSS, 1 Mbps, 96pc duty cycle)	X	1.03	64.62	15.2B	0.17	150.0	±9.6 %
		Y	1.09	65.04	15.69		150.0	
		Z	1.02	64.04	14.75		150.0	
10316- AAB	IEEE 802.11g WIFI 2.4 GHz (ERP- OFDM, 6 Mbps, 96pc duty cycle)	x	4.22	67.15	16.29	0,17	150.0	± 9,6 %
		Y	4.28	67.28	16.43		150.0	
		Z	4.18	67.07	16.13		150.0	
10317- AAC	IEEE 802.11a WIFI 5 GHz (OFDM, 6 Mbps, 96pc duty cycle)	×	4.22	67.15	16.29	0.17	150.0	±9.6 %
		Y	4.28	67.28	16.43		150.0	
		Z	4.18	67.07	16.13		150.0	1
10400- AAD	IEEE 802.11ac WiFi (20MHz, 64-QAM, 99pc duty cycle)	X	4.21	67,17	16.16	0.00	150.0	± 9.6 %
		Y	4.27	67.33	16.31		150.0	
	1	Z	4.16	67.06	15.98		150.0	
						0.00		1.00.00.000
	IEEE 802.11ac WiFi (40MHz, 64-QAM, 99pc duty cycle)	X	4.99	67.22	16.39	0.00	150.0	± 9.6 %
10401- AAD	IEEE 802,11ac WiFi (40MHz, 64-QAM, 99pc duty cycle)	X	4.99	67.22	16.39	0.00	150.0	±9.6 %

Certificate No: ES3-3019_Aug18

Page 24 of 39

August 20, 2018

10402- AAD	IEEE 802.11ac WiFi (80MHz, 64-QAM, 99pc duty cycle)	X	5.31	67.53	16.47	0.00	150.0	± 9.6 %
		Y	5.36	67.64	16.59		150.0	
		Z	6.27	67.47	16.35		150.0	
10403- AAB	CDMA2000 (1xEV-DO, Rev. 0)	X	0.45	60.00	5.81	0.00	115.0	± 9.6 %
10.464		Y	0.54	61.10	6.94		115.0	
		Z	0.44	60.00	5.71		115.0	
10404- AAB	CDMA2000 (1xEV-DO, Rev. A)	×	0.45	60.00	5.81	0.00	115.0	± 9.6 %
nnb		Y	0.54	61.10	6.94		115.0	
		z	D.44	60.00	5.71		115.0	
10406- AAB	CDMA2000, RC3, SO32, SCH0, Full Rate	x	100.00	115.64	25.77	0.00	100.0	± 9.6 %
CY1L)		Y	100.00	128.79	31.14		100.0	
		Z	100.00	106.07	21.60	-	100.0	
10410- AAE	LTE-TDD (SC-FDMA, 1 RB, 10 MHz, OPSK, UL Subframe=2,3,4,7,8,9, Subframe Cont=4)	×	100.00	125.78	31.11	3.23	80.0	± 9.6 %
	and the second s	Y	100.00	134.61	35.21		80.0	
		Z	100.00	126.61	31.37		80.0	
10415- AAA	IEEE 802.11b WiFi 2.4 GHz (DSSS, 1 Mbps, 99pc duty cycle)	×	0.93	63.35	14,45	0.00	150.0	±9.6 %
		Y	0.98	63.79	14.89		150.0	
		Z	0.93	62.96	14.01		150.0	
10416- AAA	IEEE 802.11g WiFi 2.4 GHz (ERP- OFDM, 6 Mbps, 99pc duty cycle)	x	4.15	67.03	16.15	0.00	150.0	± 9.6 %
	of and a major sale and stant	Y	4.21	67.15	16.28		150.0	
		Z	4.11	66.96	15.99		150.0	-
10417- AAB	IEEE 802.11a/h WiFi 5 GHz (OFDM, 6 Mbps, 99pc duty cycle)	X	4.15	67.03	16.15	0.00	150.0	±9.6 %
	unabol cabo and aloual	Y	4.21	67.15	16.28		150.0	
-		Z	4.11	66.96	15.99		150.0	
10418- AAA	IEEE 802.11g WIFi 2.4 GHz (DSSS- OFDM, 6 Mbps, 99pc duty cycle, Long preambule)	X	4.14	67.26	16.23	0.00	150.0	± 9.6 %
		Y	4.20	67.39	16.37		150.0	
	the second s	Z	4.10	67.18	16.07		150.0	
10419- AAA	IEEE 802.11g WiFi 2.4 GHz (DSSS- OFDM, 6 Mbps, 99pc duty cycle, Short preambule)	X	4.16	67.18	16.21	0.00	150.0	±9.6 %
-		Y	4.22	67.31	16.34		150.0	
		2	4.12	67.11	16.05		150.0	
10422- AAB	IEEE 802.11n (HT Greenfield, 7.2 Mbps, BPSK)	X	4.26	67.14	16.22	0.00	150.0	± 9.6 %
		Ý	4.32	67.27	16.35		150.0	
		Z	4.22	67.07	16.07		150.0	
10423- AAB	IEEE 802.11n (HT Greenfield, 43.3 Mbps, 16-QAM)	x	4.36	67.37	16.29	0.00	150.0	± 9.6 %
C.F.G.W.	- A distance and	Y	4,42	67.49	16.43	1	150.0	
		Z	4.31	67.29	16.14	1	150.0	
10424- AAB	IEEE 802.11n (HT Greenfield, 72.2 Mbps, 64-QAM)	×	4.29	67.30	16.26	0.00	150.0	± 9.6 %
		Y	4.35	67.43	16.40		150.0	
		Z	4.25	67.21	16.11		150.0	-
10425- AAB	IEEE 802,11n (HT Greenfield, 15 Mbps, BPSK)	x	4.96	67.44	16.53	0.00	150.0	± 9.6 %
		Y	5.01	67.55	16.65		150.0	
		Z	4.91	67.33	16,37		150.0	
10426- AAB	IEEE 802.11n (HT Greenfield, 90 Mbps, 16-QAM)	×	5.01	67,66	16.64	0.00	150.0	± 9,6 %
		Y	5.07	67.78	16.75		150.0	
		Z	4.96	67.54	16,48		150.0	

Certificate No: ES3-3019_Aug18

Page 25 of 39

August 20, 2018

10427- AAB	IEEE 802.11n (HT Greenfield, 150 Mbps, 64-QAM)	X	4.95	67.36	16.48	0.00	150.0	± 9.6 %
1		Y	5.01	67.47	16.60		150.0	
		Z	4.91	67.28	16.35		150.0	
10430- AAC	LTE-FDD (OFDMA, 5 MHz, E-TM 3.1)	x	4,06	73.07	17.88	0.00	150.0	± 9.6 %
		Y	3.94	72.28	17.62		150.0	
		Z	3.97	72.74	17.53		150.0	
10431- AAC	LTE-FDD (OFDMA, 10 MHz, E-TM 3.1)	X	3.70	67.58	15.79	0.00	150.0	± 9.6 %
Contraction in the local division of the loc		Y	3.77	67.74	15.97		150.0	
		Z	3.64	67.41	15.56		150.0	
10432- AAC	LTE-FDD (OFDMA, 15 MHz, E-TM 3,1)	x	4.05	67,44	16,13	0.00	150.0	± 9.6 %
		Y	4.11	67.58	16.28		150.0	-
	Station of the second second	Z	4.00	67.33	15,96		150.0	
10433- AAC	LTE-FDD (OFDMA, 20 MHz, E-TM 3.1)	х	4.31	67.34	16.29	0.00	150.0	± 9.6 %
		Y	4.37	67.47	16.42	-	150.0	
	and the second sec	Z	4.27	67.26	16.14	-	150.0	-
10434- AAA	W-CDMA (BS Test Model 1, 64 DPCH)	x	3.89	72.67	16.88	0.00	150.0	± 9.6 %
		Y	3.81	72.16	16.78		150.0	
	and the second sec	Z	3.70	71.97	16.35		150.0	
10435- AAE	LTE-TDD (SC-FDMA, 1 RB, 20 MHz, QPSK, UL Subframe=2,3,4,7,8,9)	X	100.00	125.45	30,96	3.23	0,08	±9.6 %
		Y	100.00	134.26	35.05	_	80,0	
		Z	100.00	126.27	31.22		80.0	
10447- AAC	LTE-FDD (OFDMA, 5 MHz, E-TM 3.1, Clipping 44%)	X	2.79	66.54	13.72	0.00	150.0	± 9.6 %
		Y	2.89	66.90	14.06		150.0	
		Z	2,70	66.17	13.34		150.0	
10448- AAC	LTE-FDD (OFDMA, 10 MHz, E-TM 3.1, Clippin 44%)	x	3,58	67.39	15.68	0.00	150.0	± 9.6 %
		Y	3.65	67.57	15.87		150.0	÷
		Z	3.53	67.23	15.46		150.0	
10449- AAC	LTE-FDD (OFDMA, 15 MHz, E-TM 3.1. Cliping 44%)	x	3.91	67.27	16.04	0.00	150.0	±9.6 %
		Y	3.97	67.41	16.19		150.0	1
	A service of the second s	Z	3.86	67.16	15.86		150.0	
10450- AAC	LTE-FDD (OFDMA, 20 MHz, E-TM 3.1, Clipping 44%)	x	4.14	67.12	16.15	0.00	150.0	±9.6 %
		Y	4.20	67.25	16.28		150.0	
		Z	4.10	67.04	15.99		150.0	1
10451- AAA	W-CDMA (BS Test Model 1, 64 DPCH, Clipping 44%)	x	2.45	65.46	12.35	0.00	150.0	±9.6 %
	2500	Y	2.56	65.93	12.76		150.0	
		Z	2.35	65.03	11.91		150.0	
10456- AAB	IEEE 802.11ac WiFi (160MHz, 64-QAM, 99pc duty cycle)	х	6.45	69.52	17.52	0.00	150.0	± 9.6 %
		Y	6.40	69.35	17.50		150.0	
	and the second se	Z	6.49	69.67	17.50	_	150.0	
10457- AAA	UMTS-FDD (DC-HSDPA)	x	3.58	65.94	15.92	0.00	150.0	± 9.6 %
		Y	3.63	66.05	16.05		150.0	
		Z	3.57	65.91	15.77		150.0	
10458- AAA	CDMA2000 (1xEV-DO, Rev. B, 2 carriers)	×	2.65	67.20	13.29	0.00	150.0	± 9.6 %
		Y	2.77	67.71	13.74		150.0	_
		Z	2.45	66.19	12.51		150.0	
10459- AAA	CDMA2000 (1xEV-DO, Rev. B, 3 carriers)	x	4.33	68.48	16.53	0.00	150.0	±9.6 %
		Y	4.22	67.81	16.25		150.0	
		Z	4.21	68,13	16.10		150.0	

Certificate No: ES3-3019_Aug18

Page 26 of 39

August 20, 2018

10460- AAA	UMTS-FDD (WCDMA, AMR)	х	0.82	69.31	15.66	0.00	150.0	± 9.6 %
1001		Y	0.95	71.24	16.99		150.0	
_		Z	0.74	67.22	14.49		150.0	
10461- AAA	LTE-TDD (SC-FDMA, 1 RB, 1.4 MHz, QPSK, UL Subframe=2,3,4,7,8,9)	х	100.00	131.81	33.89	3.29	60.0	± 9.6 %
		Y	100.00	142.15	38.65		80.0	
		Z	100.00	134.92	35.12		80.0	
10462- AAA	LTE-TDD (SC-FDMA, 1 RB, 1.4 MHz, 16-QAM, UL Subframe=2,3,4,7,8,9)	X	0.90	62.21	8.81	3.23	80.0	± 9.6 %
		Y	100.00	110.30	23.82		0.08	
		Z	0.87	62.36	8.52		80.0	
10463- AAA	LTE-TDD (SC-FDMA, 1 RB, 1.4 MHz 64-QAM, UL Subframe=2.3,4,7,8,9)	x	0,71	60.00	7.04	3.23	80.0	±9.6 %
		Y	1.91	69.61	11.91	-	80.0	
1100	And the second se	Z	0.66	60.00	6.66	1	80.0	
10464- AAB	LTE-TDD (SC-FDMA, 1 RB, 3 MHz. QPSK, UL Subframe=2,3,4,7,8,9)	×	100.00	127.95	31,95	3.23	80.0	± 9.6 %
		Y	100.00	139.33	37.13		80.0	-
		Z	100.00	131.02	33,15		80.0	
10465- AAB	LTE-TDD (SC-FDMA, 1 RB, 3 MHz, 16- QAM, UL Subframe=2,3,4,7,8,9)	x	0.82	61.38	8.35	3.23	80.0	± 9.6 %
		Y.	100.00	109.29	23.38		0.08	
		Z	0.77	61.39	8.02		0.08	
10466- AAB	LTE-TDD (SC-FDMA, 1 RB, 3 MHz, 64- OAM, UL Subframe=2,3,4,7,8,9)	x	0.71	60.00	7.00	3.23	80.0	± 9,6 %
		Y	1.22	65.67	10.37		80.0	-
		Z	0.67	60.00	6.62		80.0	-
10467- AAD	LTE-TDD (SC-FDMA, 1 RB, 5 MHz, QPSK, UL Subframe=2,3,4,7,8,9)	X	100.00	128.47	32.18	3.23	80.0	± 9.6 %
		Ý	100.00	139.85	37.36		80.0	-
		Z	100.00	131.59	33,40		80.0	
10468- AAD	LTE-TDD (SC-FDMA, 1 RB, 5 MHz, 16- QAM, UL Subframe=2,3,4,7,8,9)	x	0.85	61.72	8.54	3.23	80.0	± 9,6 %
		Ŷ	100.00	109.79	23.59		80.0	1
		Z	0.81	61.80	8.24		80.0	
10469- AAD	LTE-TOD (SC-FDMA, 1 RB, 5 MHz, 64- QAM, UL Subframe=2.3.4,7,8.9)	x	0.71	60.00	7.00	3.23	80.0	± 9.6 %
		Y	1.26	65.97	10.50	-	80.0	-
		Z	0.66	60.00	6.63		80.0	
10470- AAD	LTE-TDD (SC-FDMA, 1 RB, 10 MHz, OPSK, UL Subframe=2.3,4,7,8,9)	X	100,00	128,50	32,19	3.23	80.0	± 9.6 %
		Y	100.00	139,93	37,39		0.08	
		Z	100:00	131.64	33,41	-	80.0	
10471- AAD	LTE-TDD (SC-FDMA, 1 RB, 10 MHz, 15- QAM, UL Subframe=2,3,4,7,8,9)	x	0.84	61.66	8.50	3.23	80.0	± 9.6 %
		Y	100.00	109.69	23.54	-	80.0	
	and the second sec	Z	0.80	61.73	8,19		0.08	1
10472- AAD	LTE-TDD (SC-FDMA, 1 RB, 10 MHz, 64- QAM, UL Subframe=2.3,4,7,8,9)	×	0,71	60.00	6.99	3.23	0.08	± 9.6 %
		Y	1,24	65.80	10.41	-	80.0	-
		Z	0.66	60.00	6.61	0.05	80.0	1000
10473- AAD	LTE-TDD (SC-FDMA, 1 RB, 15 MHz, QPSK, UL Subframe=2,3,4,7,8,9)	X	100.00	128.47	32.17	3,23	80.0	± 9.6 %
		Y	100.00	139.91	37.37		80.0	-
10474-	LTE-TDD (SC-FDMA, 1 RB, 15 MHz, 16-	X	100.00	131.60 61.63	33.40 8.49	3.23	80.0 80.0	± 9.6 %
AAD	QAM. UL Subframe=2,3,4,7,8,9}	-				-	000	-
		Y	100.00	109.69	23.54	-	0.08	-
		2	0.80	61.70	8.18	0.00	0.08	1000
10475- AAD	LTE-TDD (SC-FDMA, 1 RB, 15 MHz, 64- QAM, UL Subframe=2,3,4,7,8,9)	x	0.71	60.00	6.99	3.23	80.0	±9.6 %
_		Ŷ	1.23	65,75	10.39	-	80.0	-
		Z	0.66	60.00	6.61		80.0	1

Certificate No: ES3-3019_Aug18

Page 27 of 39

August 20, 2018

10477- AAE	LTE-TDD (SC-FDMA, 1 RB, 20 MHz, 16- QAM, UL Subframe=2,3,4,7,8,9)	x	0.82	61.39	8.34	3.23	B0.0	± 9.6 %
_		YZ	100.00	109.29	23.36		80.0 80.0	
10478- AAE	LTE-TDD (SC-FDMA, 1 RB, 20 MHz, 64- QAM, UL Subframe=2,3,4,7,8,9)	X	0.71	60.00	6.98	3.23	80.0	±9.6 %
		Y	1.20	65.55	10.30		80.0	
		Z	0.66	60.00	6.60		80.0	
10479- AAA	LTE-TOD (SC-FDMA, 50% RB, 1.4 MHz, QPSK, UL Subframe=2,3,4,7,8,9)	×	100.00	125.63	32.26	3.23	80.0	± 9.6 %
		Y	100.00	131.75	35.27		80.0	
		Z	100.00	127,84	33.13		80.0	
10480- AAA	LTE-TDD (SC-FDMA, 50% RB, 1.4 MHz, 16-QAM, UL Subframe=2,3,4,7,8,9)	X	19.63	91.54	20.27	3.23	80.0	± 9.6 %
		Y	100.00	115.20	27.34		80.0	1
	and a second	2	100.00	108.55	24.07	-	80.0	
10481- AAA	LTE-TDD (SC-FDMA, 50% RB, 1.4 MHz, 64-QAM, UL Subframe=2,3,4,7,8,9)	×	3.49	72.42	13.95	3.23	80.0	± 9.6 %
		Y	100.00	111.30	25.48		80.0	
_	and the second se	Z	5.91	77,84	15.53	-	80.0	
10482- AAB	LTE-TDD (SC-FDMA. 50% RB, 3 MHz, OPSK. UL Subframe=2,3,4,7,8,9)	×	1.67	65,33	11.27	2.23	80.0	± 9.6 %
		Y	2.30	68.66	13.07		80.0	1
		Z	1.38	63.63	10.34	1.00	80.0	
10483- AAB	LTE-TDD (SC-FDMA, 50% RB, 3 MHz, 16-QAM, UL Subframe=2,3,4,7,8,9)	x	1.25	60.06	7.80	2.23	80.0	±9.6 %
_		Y	1.97	64.67	10.70		80.0	
_		Z	1.19	60.00	7.59		80.0	P
10484- AAB	LTE-TDD (SC-FDMA, 50% RB, 3 MHz. 64-QAM, UL Subframe=2.3,4,7,8,9)	x	1.27	60.00	7.74	2.23	80.0	± 9.6 %
		Y.	1.80	63.53	10.14		80.0	
	and the second s	Z	1.22	60.00	7.56		80.0	-
10485- AAD	LTE-TDD (SC-FDMA, 50% RB, 5 MHz, QPSK, UL Subframe=2,3,4,7,8,9)	X	5.81	81.43	19.42	2.23	60.0	± 9.6 %
_		Y	6.11	82.25	20.03		80.0	
1. The second		Z	4.06	77.04	17.81		80.0	
10486- AAD	LTE-TDD (SC-FDMA, 50% RB, 5 MHz, 16-QAM, UL Subframe=2,3,4,7,8,9)	×	2.34	66.39	12.56	2.23	80.0	±9.6 %
_		Y	2.71	68.04	13.57		60.0	
		Z	2.02	64.93	11.73		80.0	1.7.7
10487- AAD	LTE-TDD (SC-FDMA, 50% RB, 5 MHz, 64-QAM, UL Subframe=2,3,4,7,8,9)	×	2.25	65.64	12.18	2.23	80.0	±9.6 %
		Y	2.60	67,18	13.15		80.0	
	and the second second second	Z	1.96	64,31	11.39		80.0	1
10488- AAD	LTE-TDD (SC-FDMA, 50% RB, 10 MHz, QPSK, UL Subframe=2,3,4,7,8,9)	x	5.09	80.09	21.16	2.23	80.0	± 9.6 %
		Y	5,16	80.10	21.33		80.0	1
		Z	4,19	77.34	20.09		80.0	
10489- AAD	LTE-TDD (SC-FDMA, 50% RB, 10 MHz, 16-QAM, UL Subframe=2,3,4,7,8,9)	×	4.01	73.05	18.07	2.23	80.0	±9.6 %
-		Y	4.07	73.07	18.23		80.0	
		Z	3.64	71.81	17.45		0.08	
10490- AAD	LTE-TDD (SC-FDMA, 50% RB, 10 MHz, 64-QAM, UL Subframe=2,3,4,7,8,9)	X	3.98	72.48	17,82	2.23	80.0	±9.6 %
_		Y	4.06	72.53	17.99		0.08	
-0.15		Z	3.63	71.32	17.22		80.0	
10491- AAD	LTE-TDD (SC-FDMA, 50% RB, 15 MHz, QPSK, UL Subframe=2,3,4,7,8,9)	x	4.41	75.53	19.94	2.23	80.0	± 9.6 %
		Y	4.53	75.73	20.14	-	80.0	_
		Z	3.93	73.90	19.24		80.0	
10492- AAD	LTE-TDD (SC-FDMA, 50% RB, 15 MHz, 16-QAM, UL Subframe=2,3,4,7,6,9)	x	4.05	71.20	18.08	2.23	0.08	± 9.6 %
-		Y	4.14	71.29	18.22		80.0	-
		Z	3.79	70,37	17.63		0.08	

Certificate No: ES3-3019_Aug18

Page 28 of 39

August 20, 2018

10493- AAD	LTE-TDD (SC-FDMA, 50% RB, 15 MHz, 64-QAM, UL Subframg=2,3,4,7,8,9)	x	4.07	70.90	17.94	2.23	80.0	±9.6 %
		YZ	4.15	71.00	18.07 17.49	_	0.08	
10494- AAE	LTE-TDD (SC-FDMA, 50% RB, 20 MHz, QPSK, UL Subframe=2,3,4,7,8,9)	X	4.91	77,32	20,61	2.23	80.0	± 9,6 %
AME	QF3R, 0L 300name=2,3,4,7,0,3)	Y	5.04	77.50	20.80		80.0	
		Z	4.30	75.42	19.85		80.0	
10495-	LTE-TDD (SC-FDMA, 50% RB, 20 MHz,	X	4.09	71.47	18,41	2.23	80.0	± 9.6 %
AAE	16-QAM, UL Subframe=2,3,4,7,8,9)	-	4.17	71.55	18,53	6169	80.0	20.0 10
		Y		70.60	17.95	-	80.0	
10495- AAE	LTE-TDD (SC-FDMA, 50% RB, 20 MHz, 64-QAM, UL Subframe=2.3,4,7,8,9)	X	3.83 4.13	71.08	18.26	2.23	80.0	# 9.6 %
MAE	04-07-14, OL Subiranie-2,0,4,7,0.31	Y	4.21	71.17	18.38		60.0	
		Z	3.88	70.29	17.84	-	B0.0	1
10107	175 700 400 50444 4000 00 44		0.96	60.00	6.99	2.23	80.0	# 9.6 %
10497- AAA	LTE-TDD (SC-FDMA, 100% RB, 1.4 MHz, QPSK, UL Subframe=2,3,4,7,8,9)	×	2450	100.01	2.2.2.4	6.63		1 9.0 %
		Y	0.98	60.00	7.40	_	80.0	
		Z	0.92	60.00	6.79		80.0	1000
10498- AAA	LTE-TDD (SC-FDMA, 100% RB, 1.4 MHz, 16-QAM, UL Subframe=2,3,4,7,8,9)	x	1,14	60.00	5.73	2.23	80.0	19.6 %
		Y	1.15	60.00	6.09		80.0	
		Z	1.11	60.00	5.47	1	80.0	
10499- AAA	LTE-TDD (SC-FDMA, 100% RB, 1.4 MHz, 64-QAM, UL	×	1.17	60.00	5.57	2.23	80.0	± 9.6 %
	Subframe=2.3,4,7,8,9)	1				_		
_	a state of the sta	Y	1.18	60.00	5.92		80.0	
		Z	1.15	60.00	5.29		80.0	
10500- AAB	LTE-TDD (SC-FDMA, 100% RB, 3 MHz, QPSK, UL Subframe=2.3,4,7,8.9)	X	5.67	81,49	20.32	2.23	B0.0	± 9.6 %
		Y	5.74	81.66	20,65		80.0	
		Z	4.28	77.81	18.94		80.0	
10501- AAB	LTE-TDD (SC-FDMA, 100% RB, 3 MHz, 16-QAM, UL Subframe=2,3,4,7,8,9)	X	3.25	70,19	15.15	2.23	80.0	± 9.6 %
		Y	3.52	71,13	15.78		0.08	
		Z	2.82	68,60	14,33		80.0	
10502- AAB	LTE-TDD (SC-FDMA_ 100% RB, 3 MHz, 64-QAM, UL Subframe=2,3,4,7,8,9)	x	3.16	69,49	14.76	2.23	0.08	± 9.6 %
		Y	3.44	70.48	15.41		0.08	
		Z	2.76	67.99	13.96		80.0	-
10503- AAD	LTE-TDD (SC-FDMA, 100% RB, 5 MHz, OPSK, UL Subframe=2.3,4,7,8,9)	x	4.96	79.69	20.99	2.23	80.0	± 9.6 %
		Y	5.06	79.77	21,19	-	80.0	
		Z	4.10	76.98	19.93		80.0	
10504- AAD	LTE-TDD (SC-FDMA, 100% RB, 5 MHz, 16-QAM, UL Subframe=2,3,4,7,8,9)	X	3.96	72.85	17.97	2.23	0.08	±9.6 %
		Y	4.04	72.91	18.14		0.08	
		Z	3.60	71.62	17,35		80.0	
10505- AAD	LTE-TDD (SC-FDMA, 100% RB, 5 MHz, 64-QAM, UL Subframe=2.3,4,7,8,9)	x	3.94	72.31	17.73	2.23	80.0	± 9.6 %
		Y	4.03	72.40	17.91		60.0	
		Z	3.60	71.15	17.13	1100	80.0	
10506- AAD	LTE-TDD (SC-FDMA, 100% RB, 10 MHz, QPSK, UL Subframo=2,3,4,7,8,9)	X	4.85	77.09	20.51	2.23	80.0	± 9.6 %
		Y	4.98	77.30	20.71		80.0	
		Z	4.25	75.22	19.75	-	80.0	
10507- AAD	LTE-TDD (SC-FDMA, 100% RB, 10 MHz, 16-QAM, UL Subframe=2,3,4,7,8,9)	×	4.07	71.38	18.36	2.23	80.0	± 9,6 %
		Y	4.16	71.48	18.49		80.0	

Certificate No: ES3-3019_Aug18

Page 29 of 39

August 20, 2018

10508- AAD	LTE-TDD (SC-FDMA, 100% RB, 10 MHz, 64-QAM, UL Subframe=2,3,4,7,8,9)	X	4.11	70.96	18.20	2.23	80.0	± 9,6 %
		Y	4.19	71.08	18.33	-	80.0	-
		Z	3.86	70.19	17.78	-	0.08	
10509- AAD	LTE-TDD (SC-FDMA, 100% RB, 15 MHz, QPSK, UL Subframe=2,3,4,7,8,9)	X	4.80	74.18	19.50	2.23	80.0	±9.6 %
		Y	4.92	74.37	19.67		80.0	
		Z	4.40	73.00	18.99		BD.0	
10510- AAD	LTE-TDD (SC-FDMA, 100% RB, 15 MHz, 16-QAM, UL Subframe=2,3,4,7,8,9)	x	4.38	70,21	18.17	2.23	80.0	± 9.6 %
		Y	4.48	70.36	18.31	_	80.0	-
	A second and second second second	Z	4.17	69.55	17.82		80.0	
10511- AAD	LTE-TDD (SC-FDMA, 100% RB, 15 MHz, 64-QAM, UL Subframe=2,3,4,7,8,9)	x	4,44	69.97	18.09	2.23	80.0	± 9.6 %
		Y	4.54	70.12	18.23		80.0	-
		Z	4.23	69.36	17.75		80.0	
10512- AAE	LTE-TDD (SC-FDMA, 100% RB, 20 MHz, QPSK, UL Subframe=2,3,4,7,8,9)	×	5.12	75.85	20.03	2.23	80.0	± 9.6 %
		Y	5.25	76.05	20.20		80.0	-
_		Z	4.61	74,40	19.43		80.0	
10513- AAE	LTE-TDD (SC-FDMA, 100% RB, 20 MHz, 16-QAM, UL Subframe=2,3,4,7,8,9)	x	4.29	70.38	18,28	2.23	80.0	±9.6%
		Y	4.39	70.55	18,43		80.0	
		Z	4.07	69.67	17.91		0.08	
10514- AAE	LTE-TDD (SC-FDMA, 100% RB, 20 MH2, 64-QAM, UL Subframe=2,3,4,7,8,9)	x	4,31	69.92	18,12	2.23	0.08	±9.6 %
		Y	4.40	70.08	18.26		80.0	
		Z	4.10	69.28	17.77		80.0	
10515- AAA	IEEE 802.11b WiFi 2.4 GHz (DSSS, 2 Mbps, 99pc duty cycle)	×	0.89	63.54	14.50	0.00	150.0	± 9.6 %
		Y	0.94	64.02	14.97		150.0	
		Z	0.89	63,10	14.03		150.0	
10516- AAA	IEEE 802.11b WIFI 2.4 GHz (DSSS, 5.5 Mbps, 99pc duty cycle)	x	0.63	73,64	17.73	0.00	150.0	±9.6 %
_		Y	0,85	77.87	20.24	-	150.0	
		Z	0.50	68.63	15.41		150.0	-
10517- AAA	IEEE 802.11b WiFi 2.4 GHz (DSSS, 11 Mbps, 99pc duty cycle)	x	0.73	65.48	15.07	0.00	150.0	± 9.6 %
_		Y	0.80	66.38	15.85		150.0	
10518- AAB	IEEE 802.11a/h WiFi 5 GHz (OFDM, 9 Mbps, 99pc duty cycle)	Z X	0.72	64.47 67.16	14.31 16.15	0.00	150.0 150.0	±9.6 %
3.16	aspot solic data chapi	Y	4.20	67.29	16.29		150.0	
		z	4.10	67.09	16.00	-	150.0	
10519- AAB	IEEE 802.11a/h WiFi 5 GHz (OFDM, 12 Mbps, 99pc duty cycle)	X	4.10	67.30	16.23	0.00	150.0	± 9,6 %
-		Y	4.32	67.42	16.36		150.0	
		Z	4.22	67.22	16.07		150.0	
10520- AAB	IEEE 802.11a/h WiFi 5 GHz (OFDM, 18 Mbps. 99pc duty cycle)	×	4.12	67.21	16.14	0.00	150.0	€ 9.6 %
		Y	4.18	67.34	16.28	-	150.0	
1000		Z	4.08	67.13	15,98		150.0	
10521- AAB	IEEE 802.11a/h WiFL5 GHz (OFDM, 24 Mbps, 99pc duty cycle)	×	4.05	67.13	16.10	0.00	150.0	± 9.6 %
		Y	4,11	67.26	16.24	-	150.0	
LOBOX		Z	4.01	67.03	15.93		150.0	
10522- AAB	IEEE 802.11a/h WiFr 5 GHz (OFDM, 36 Mbps, 99pc duty cycle)	×	4.08	67.17	16.14	0.00	150,0	± 9.6 %
-		Y	4.14	67.31	16.28		150.0	
		Z	4.03	67.05	15.96		150.0	

Certificate No: ES3-3019_Aug18

Page 30 of 39

August 20, 2018

10523- AAB	IEEE 802.11a/h WiFi 5 GHz (OFDM, 48 Mbps, 99pc duty cycle)	X	4.05	67.38	16.19	0,00	150.0	±9.6 %
	training a share of stand	Y	4.12	67.52	16.33		150.0	_
		Z	4.01	67.29	16.03		150.0	
10524- AAB	IEEE 802.11a/h WiFi 5 GHz (OFDM, 54 Mbps, 99pc daty cycle)	X	4.05	67.25	16.21	0.00	150.0	±9.6 %
	maps, superand every	Y	4.11	67.39	16.35		150.0	
		Z	4.00	67.15	16.04		150.0	
10525- AAB	IEEE 802.11ac WiFi (20MHz, MCS0, 99pc duty cycle)	X	4.12	66.42	15.87	0.00	150.0	± 9.6 %
		Y	4.18	66.55	16.00	-	150.0	
		Z	4.08	66.34	15.71		150.0	-
10526- AAB	IEEE 802.11ac WiFi (20MHz, MCS1, 99pc duty cycle)	X	4.20	66.62	15.96	0.00	150.0	± 9.6 %
1 il nut	and a start	Y	4.27	66.76	16.09		150.0	
-		Z	4.15	66.52	15.79		150.0	-
10527- AAB	IEEE 802.11ac WiFi (20MHz, MCS2, 99pc duty cycle)	x	4.15	66.61	15.90	0.00	150.0	± 9.6 %
	and stand	Y	4.21	66.75	16.04		150.0	
		Z	4.10	66.51	15.74		150.0	
10528- AAB	IEEE 802.11ac WiFi (20MHz, MCS3, 99pc duty cycle)	x	4.16	66.61	15.93	0.00	150.0	± 9.6 %
	and a state of the	Y	4.22	66.75	16.07		150.0	
		Z	4.11	66.51	15.76		150.0	
10529- AAB	IEEE 802.11ac WIFI (20MHz, MCS4, 99pc duty cycle)	x	4.16	66.61	15.93	0.00	150.0	±9.6 %
10.02	cope and of an	Y	4.22	66.75	16.07		150.0	
		Z	4.11	66.51	15.76		150.0	
10531- AAB	IEEE 802.11ac WiFi (20MHz, MCS6, 99pc duty cycle)	x	4.11	66.58	15.88	0.00	150.0	± 9.6 %
	sole and sheet	Y	4.17	66.72	16.02		150.0	-
		Z	4.06	66.47	15.71		150.0	
10532- AAB	IEEE 802.11ac WiFi (20MHz, MCS7, 99pc duty cycle)	X	4.01	66.45	15.81	0.00	150.0	± 9,6 %
10.00	aske and stand	Y	4.07	66.59	15.95		150.0	
		Z	3.96	66.35	15.65		150.0	
10533- AAB	IEEE 802,11ac WiFi (20MHz, MCSB, 99pc duty cycle)	X	4.16	66.72	15.94	0.00	150.0	± 9.6 %
		Y	4.22	66.86	16.08	-	150.0	
		Z	4.11	66.61	15,77	-	150.0	
10534- AAB	IEEE 802-11ac WiFi (40MHz, MCS0, 99pc duty cycle)	X	4.75	66.52	16.05	0.00	150.0	± 9.6 %
	and the second states of the s	Y	4.81	66.64	16,16		150.0	
		Z	4.71	66.44	15,91		150.0	
10535- AAB	IEEE 802,11ac WiFi (40MHz, MCS1, 99pc duty cycle)	x	4.78	66.63	16.11	0.00	150.0	± 9.6 %
		Y	4.84	66.75	16.22	1	150.0	
		Z	4.74	66.54	15.96	1	150.0	
10536- AAB	IEEE 802.11ac WIFI (40MHz, MCS2, 99pc duty cycle)	x	4.68	66.60	16.07	0.00	150.0	± 9.6 %
		Y	4.73	66.73	16.19		150.0	
1		Z	4.63	66.52	15.93	1000	150.0	
10537- AAB	IEEE 802.11ac WiFi (40MHz, MCS3, 99pc duty cycle)	X	4.77	66.73	16.14	0.00	150.0	± 9.6 %
-		Y	4.83	66.85	16.25		150.0	-
-	A Real Property of the second second	Z	4.72	66.63	15.99		150.0	
10538- AAB	IEEE 802.11ac WiFI (40MHz, MCS4, 99pc duty cycle)	×	4.79	66.55	16.08	0.00	150.0	± 9.6 %
	-	Y	4.85	66.67	16.20		150.0	
-	I Day of the state of the state of the	Z	4.75	66.46	15,94		150.0	
10540- AAB	IEEE 802.11ac WiFi (40MHz, MCS6, 99pc duty cycle)	X	4.73	66.52	16.09	0.00	150.0	±9.6 %
a la de la companya d		Y	4.79	66.64	16,21		150.0	

Certificate No: ES3-3019_Aug18

Page 31 of 39

August 20, 2018

10541- AAB	IEEE 802.11ac WiFi (40MHz, MCS7, 99pc duty cycle)	×	4.73	66.49	16.05	0.00	150.0	± 9,6 %
		Ϋ́	4.79	66.60	16.16		150.0	
		Z	4.69	66.41	15.91		150.0	
10542- AAB	IEEE 802.11ac WiFi (40MHz, MCS8, 99pc duty cycle)	X	4.87	66.56	16.10	0.00	150.0	±9.6 %
		Y	4.93	66.68	16.22		150.0	
		Z	4,83	66.49	15.97		150.0	
10543- AAB	IEEE 802.11ac WIFI (40MHz, MCS9, 99pc duty cycle)	×	4.95	66.68	16.20	0.00	150.0	± 9.6 9
		Y	5.00	66.80	16.31		150.0	
		Z	4.89	66.58	16.05		150.0	
10544- AAB	IEEE 802.11ac WIFI (80MHz, MCS0, 99pc duty cycle)	×	5.13	66.53	16.04	0.00	150.0	± 9.6 %
_	V A Y Y Y	Y	5,19	66,64	16.15		150.0	
		Z	5,10	66,46	15.91		150.0	
10545- AAB	IEEE 802.11ac WiFi (80MHz, MCS1, 99pc duty cycle)	×	5.32	67.06	16.27	0,00	150.0	± 9.6 %
		Y	5.37	67.18	16.39		150.0	
		2	5.26	66.93	16.12		150.0	
10546- AAB	IEEE 802.11ac WiFi (80MHz, MCS2, 99pc duty cycle)	×	5,16	66.63	16.06	0.00	150.0	±9.6 %
		Y	5.21	66.74	16.17		150.0	
		Z	5.12	66.56	15.93		150.0	
10547- AAB	IEEE 802.11ac WiFi (80MHz, MCS3, 99pc duty cycle)	×	5.32	67.06	16.27	0.00	150.0	±9.6 %
		Y	5.37	67.16	16.38		150.0	
		Z	5.27	66.95	16.13		150.0	
10548- AAB	IEEE 802.11ac WiFi (80MHz, MCS4, 99pc duty cycle)	×	5.33	67.26	16.36	0.00	150.0	± 9.6 %
		Y	5.39	67.41	16.48		150.0	
		Z	5.26	67.08	16.18		150.0	
10550- AAB	IEEE 802.11ac WiFi (80MHz, MCS6, 99pc duty cycle)	x	5.31	67,19	16.36	0.00	150.0	±9.6 %
		Y	5.36	67.30	16.47	-	150.0	
		Z	5.26	67.08	16.21		150.0	
10551- AAB	IEEE 802.11ac WiFI (80MHz, MCS7, 99pc duty cycle)	x	5,13	66.55	16.00	0.00	150.0	± 9.6 %
		Y	5.18	66.66	16.11	1	150.0	
		Z	5.09	66.49	15.88		150.0	-
10552- AAB	IEEE 802.11ac WiFr (80MHz, MCS8, 99pc duty cycle)	×	5.13	66.67	16.05	0.00	150.0	±9.6 %
		Y	5.19	66.78	16.16		150.0	
		Z	5.09	66.61	15.93	-	150.0	
10553- AAB	IEEE 802.11ac WiFi (80MHz, MCS9, 99pc duty cycle)	x	5.17	66.57	16,03	0.00	150.0	± 9.6 %
		Y	5.23	66.69	16.14		150.0	
		Z	5.14	66.52	15.91		150.0	
10554- AAC	IEEE 802.11ac WiFi (160MHz, MCS0, 99pc duty cycle)	×	5.59	66,85	16.12	0.00	150,0	± 9.6 %
		Y	5.64	66.96	16.23	-	150.0	
		Z	5.55	66.79	16.00		150.0	
10555- AAC	IEEE 802.11ac WiFi (160MHz, MCS1, 99pc duty cycle)	×	5.66	67.06	16.21	0.00	150,0	± 9,6 %
		Y	5,71	67.17	16.32		150.0	
	and the second se	Z	5.62	66.97	16.09		150.0	
10556- AAC	IEEE 802.11ac WiFi (160MHz, MCS2, 99pc duty cycle)	×	5.73	67.28	16.32	0.00	150.0	± 9.6 %
		Y	5.79	67.40	16.43		150.0	
	Contraction of the second	Z	5.68	67.17	16.18		150.0	
10557- AAC	IEEE 802.11ac WiFi (160MHz, MCS3, 99pc duty cycle)	x	5.65	67.02	16.20	0.00	150.0	± 9.6 %
		Y	5.70	67.14	16.31		150.0	

Certificate No: ES3-3019_Aug18

Page 32 of 39

August 20, 2018

10558- AAC	IEEE 802.11ac WIFI (160MHz; MCS4, 99pc duty cycle)	X	5.62	66.96	16,19	0.00	150.0	± 9.6 %
nymu.	appendig epone	Y	5.67	67.08	16.30		150.0	
		Z	5.57	66.8B	16.06		150.0	
10560- AAC	IEEE 802.11ac WiFi (160MHz, MCS6, 99pc duty cycle)	x	5,65	66.95	16.22	0.00	150.0	± 9.6 %
		Y	5.71	67.07	16.33		150.0	
		Z	5.61	66.87	16.10	1.1	150.0	-
10561- AAC	IEEE 802.11ac WIFI (160MHz, MCS7, 99pc duty cycle)	x	5.60	66.94	16.24	0.00	150.0	± 9.6 %
		Y	5.65	67.07	16.36	_	150.0	
		Z	5.55	66.86	16.11		150.0	
10562- AAC	IEEE 802.11ac WiFi (160MHz, MCS8, 99pc duty cycle)	x	5.63	67.07	16.31	0.00	150.0	±9.6 %
		Y	5.69	67.19	16.42	1	150.0	
		Z	5.59	66.99	16.18	-	150.0	
10563- AAC	IEEE 802.11ac WiFi (160MHz, MCS9, 99pc duty cycle)	x	5.78	67.23	16.36	0.00	150.0	± 9.6 %
		Y	5.83	67.32	16.46		150.0	
		Z	5,76	67,23	16.28		150.0	
10564- AAA	IEEE 802.11g WiFi 2.4 GHz (DSSS- OFDM, 9 Mbps, 99pc duty cycle)	х	4.46	67.18	16.31	0,46	150.0	± 9.6 %
		Y	4.52	67.32	16.46		150.0	_
		Z	4.42	67.11	16.17	1	150.0	
10565- AAA	IEEE 802.11g WiFi 2.4 GHz (DSSS- OFDM, 12 Mbps, 99pc duty cycle)	×	4.63	67.58	16.63	0.46	150.0	±9.6 %
		Y	4.69	67.69	16.75		150.0	
		Z	4,59	67.52	16.49		150,0	
10566- AAA	IEEE 802.11g WiFi 2.4 GHz (DSSS- OFDM, 18 Mbps, 99pc duty cycle)	×	4,47	67.37	16.42	0.46	150.0	± 9.6 %
		Y	4.54	67.50	16.57		150.0	
11000		Z	4,43	67.29	16.28		150.0	
10567- AAA	IEEE 802.11g WIFI 2.4 GHz (DSSS- OFDM, 24 Mbps, 99pc duty cycle)	x	4.52	67.79	16.83	0.46	150.0	± 9.6 %
		Y	4.57	67.86	16.92		150.0	
		Z	4.48	67,72	16.69		150.0	
10568- AAA	IEEE 802.11g WiFi 2.4 GHz (DSSS- OFDM, 36 Mbps, 99pc duty cycle)	X	4.34	66,96	16.07	0.46	150.0	± 9.6 %
		Y	4.41	67.16	16.26		150.0	1
		Z	4.29	66.85	15.90		150.0	
10569- AAA	IEEE 802.11g WiFi 2.4 GHz (DSSS- OFDM, 48 Mbps, 99pc duty cycle)	×	4.53	68.16	17:05	0.46	150.0	± 9.6 %
-		Y	4.58	68.21	17.13		150.0	
	and the second se	Z	4.49	68.10	16.91		150.0	
10570- AAA	IEEE 802.11g WiFi 2.4 GHz (DSSS- OFDM, 54 Mbps, 99pc duty cycle)	X	4,49	67.82	16,87	0,46	150.0	± 9,6 %
_		Y	4.55	67.90	16.97	-	150.0	
		Z	4,45	67.74	16.72		150.0	
10571- AAA	IEEE 802.11b WiFi 2.4 GHz (DSSS: 1 Mbps: 90pc duty cycle)	X	1.16	65.70	15,86	0.46	130.0	± 9.6 %
		Y	1.22	66.16	16.29		130.0	-
		Z	1.13	64,93	15.27		130.0	1000
10572- AAA	IEEE 802.11b WiFi 2.4 GHz (DSSS, 2 Mbps, 90pc duty cycle)	x	1.1B	66.44	16.31	0.46	130.0	± 9.6 %
		Y	1.24	66.88	16.72	-	130.0	-
(maning)	The second se	Z	1.14	65,55	15.67	-	130.0	
10573- AAA	IEEE 802.11b WiFi 2.4 GHz (DSSS, 5.5 Mbps, 90pc duty cycle)	x	15.28	116.08	30.90	0.46	130.0	19,6%
		Y	37.29	132.13	35.54	1	130.0	
		Z	2.39	87.75	23.05		130.0	1000
10574- AAA	IEEE 802.11b WiFi 2.4 GHz (DSSS, 11 Mbps, 90pc duty cycle)	×	1.45	74.49	20.17	0.46	130.0	± 9.6 %
		Y	1.51	74.69	20.49		130.0	-
		Z	1.27	71.67	18.75		130.0	

Certilicate No: ES3-3019_Aug18

Page 33 of 39

August 20, 2018

10575- AAA	IEEE 802.11g WiFi 2.4 GHz (DSSS- OFDM, 6 Mbps, 90pc duty cycle)	×	4.27	67.06	16.39	0.46	130.0	± 9.6 %
	-	Y	4,33	67.20	16.53		130.0	
		Z	4.23	66.99	16.23	1	130.0	
10576- AAA	IEEE 802.11g WiFi 2.4 GHz (DSSS- OFDM, 9 Mbps, 90pc duty cycle)	x	4.31	67.33	16.51	0.46	130.0	± 9.6 %
£		Y	4.37	67.44	16.64		130.0	
		Z	4.26	67.25	16.36		130.0	
10577- AAA	IEEE 802.11g WiFI 2.4 GHz (DSSS- OFDM, 12 Mbps, 90pc duty cycle)	x	4.44	67.51	16.64	0.46	130.0	± 9.6 %
		Y	4.50	67.62	16.76		130.0	
		Z	4.39	67.44	16.49		130.0	
10578- AAA	IEEE 802.11g WiFi 2.4 GHz (DSSS- OFDM, 18 Mbps, 90pc duty cycle)	x	4.36	67.69	16.78	0.46	130.0	± 9.6 %
		Y	4.41	67,76	16.87	-	130.0	
		Z	4.32	67.61	16.63		130.0	
10579- AAA	IEEE 802.11g WiFi 2.4 GHz (DSSS- OFDM, 24 Mbps, 90pc duty cycle)	x	4.10	66.72	15.92	0.46	130.0	± 9.6 %
		Y	4.17	66,93	16.12		130.0	-
		Z	4.05	66.61	15.75		130.0	
10580- AAA	IEEE 802.11g WiFi 2.4 GHz (DSSS- OFDM, 36 Mbps, 90pc duty cycle)	x	4,11	66.69	15.89	0.46	130.0	± 9.6 %
		Y	4.18	66.91	16.09		130.0	
-		Z	4.05	66.55	15.69		130.0	
10581- AAA	IEEE 802.11g WiFi 2.4 GHz (DSSS- OFDM, 48 Mbps, 90pc duty cycle)	×	4.30	67.88	16.81	0.46	130.0	± 9.6 %
		Y	4.35	67.96	16.92		130.0	
		Z	4.25	67.78	16.65		130.0	-
10582- AAA	IEEE 802.11g WiFi 2.4 GHz (DSSS- OFDM, 54 Mbps, 90pc duty cycle)	×	4,01	66.47	15.69	0.46	130.0	± 9.6 %
		Y	4.09	66.72	15.92		130.0	
		Z	3.95	66.34	15.50		130.0	
10583- AAB	IEEE 802.11a/h WiFi 5 GHz (OFDM, 6 Mbps, 90pc duty cycle)	X	4.27	67.06	16.39	0.46	130.0	±9.6 %
		Y	4.33	67.20	16.53		130.0	
	and the second state of th	Z	4.23	66.99	16.23	-	130.0	
10584- AAB	IEEE 802.11a/h WiFi 5 GHz (OFDM, 9 Mbps, 90pc duty cycle)	X	4.31	67.33	16.51	0,46	130.0	± 9.6 %
		Y	4.37	67.44	16.64		130.0	
	Contraction of the second second	Z	4.26	67.25	16.36		130.0	
10585- AAB	IEEE 802.11a/h WiFi 5 GHz (OFDM, 12 Mbps, 90pc duty cycle)	X	4.44	67:51	16.64	0.46	130.0	± 9.6 %
		Y	4.50	67.62	16.76		130.0	
		Z	4.39	67.44	16.49		130.0	-
10586- AAB	IEEE 802.11a/h WiFi 5 GHz (OFDM, 18 Mbps. 90pc duty cycle)	x	4.36	67.69	16.78	0.46	130.0	±9.6 %
		Y	4.41	67.76	16.87		130,0	
		Z	4.32	67.61	16.63		130.0	
10587- AAB	IEEE 802.11a/h WiFi 5 GHz (OFDM, 24 Mbps. 90pc duty cycle)	x	4,10	66,72	15.92	0.46	130.0	± 9.6 %
		Y	4.17	66.93	16.12		130.0	1.0
		Z	4.05	66.61	15.75	1000	130.0	12.04
10588- AAB	IEEE 802.11a/h WiFi 5 GHz (OFDM, 36 Mbps. 90pc duty cycle)	x	4.11	66.69	15.89	0.46	130.0	± 9.6 %
-		Y	4.18	66.91	16.09		130.0	
		Z	4.05	66.55	15.69		130.0	
10589- AAB	IEEE 802.11a/h WiFi 5 GHz (OFDM, 48 Mbps, 90pc duty cycle)	x	4.30	67.88	16.81	0.45	130.0	±9.6 %
		Y	4.35	67.96	16.92		130.0	
		Z	4.25	67.78	16.65		130.0	
10590- AAB	IEEE 802.11a/h WiFi 5 GHz (OFDM, 54 Mbps, 90pc duty cycle)	X	4.01	66.47	15.69	0.46	130.0	±9.6 %
		Y	4.09	66.72	15.92	1	130.0	
		Z	3.96	66.34	15.50		130.0	

Certificate No: ES3-3019_Aug18

Page 34 of 39

August 20, 2018

10591- AAB	IEEE 802.11n (HT Mixed, 20MHz, MCS0, 90pc duty cycle)	X	4.43	67.18	16.55	0.46	130.0	± 9.6 %
5.02	and an and a mark all and	Y	4,49	67.29	16.68		130.0	
		Z	4,39	67.12	16.41		130.0	
10592- AAB	IEEE 802.11n (HT Mixed, 20MHz, MCS1, 90pc duty cycle)	×	4.52	67.42	16.68	0,46	130.0	± 9.6 %
040	West, superany syster	Y	4.58	67.53	16.78		130.0	
_		Z	4,48	67.35	16.51		130.0	
10593-	IEEE 802.11n (HT Mixed, 20MHz,	X	4,40	67.30	16.51	0.46	130.0	± 9.6 %
10593- AAB	MCS2, 90pc duty cycle)			67.43	16.65	0.40	130.0	2 9.9 19
		Y	4.51					-
		Z	4.40	67.23	16.36		130.0	
10594- AAB	IEEE 802.11n (HT Mixed, 20MHz, MCS3, 90pc duty cycle)	×	4.50	67.49	16.69	0.46	130.0	±9.6 %
		Y	4,56	67.59	16.81		130.0	
_		Z	4.46	67.41	16.54		130.0	
10595- AAB	IEEE 802.11n (HT Mixed, 20MHz, MCS4, 90pc duty cycle)	X	4.47	67.48	16.61	0.46	130.0	± 9.6 %
COLUMN .	nices, supe dail chart	Y	4.53	67.60	16.74		130.0	
		Z	4.42	67.40	16.45		130.0	
10596-	IEEE 802.11n (HT Mixed, 20MHz.	X	4.39	67.39	16.57	0.46	130.0	± 9.6 %
AAB	MCS5, 90pc duty cycle)	100	196961	19011025	10214	0.40	VALLA	
		Y	4.45	67.52	16.71	_	130.0	-
		Z	4.33	67.28	16.41		130.0	
10597- AAB	IEEE 802.11n (HT Mixed, 20MHz, MCS6, 90pc duty cycle)	x	4.35	67.24	16.40	0.46	130.0	± 9.6 %
		Y	4.41	67.39	16.55	-	130.0	
		Z	4.30	67.15	16.24		130.0	
1059B- AAB	IEEE 802.11n (HT Mixed, 20MHz, MCS7, 90pc duty cycle)	X	4.36	67.56	16.72	0.46	130.0	19.6 9
	interest color and of col	Y	4.42	67.64	16.83	_	130.0	-
		Z	4.32	67.48	16.57	-	130.0	-
10599- AAB	IEEE 802.11n (HT Mixed, 40MHz, MCS0, 90pc duty cycle)	X	5.28	68.07	17.13	0.46	130.0	± 9.6 %
nna.	mood, appendity eyeley	Y	5.31	68.10	17.21		130.0	
		Z	5.25	68.02	17.00		130.0	-
10600-	IEEE 802.11n (HT Mixed, 40MHz,	X	5.26	68.02	17.07	0.46	130.0	± 9.6 %
AAB	MCS1, 90pc duty cycle)	Y	5.32	68.16	17.21		130.0	
		Z	5.18	67.83	16.88		130.0	
10601- AAB	IEEE 802.11n (HT Mixed, 40MHz, MCS2, 90pc duty cycle)	×	5,19	67.88	17.02	0.46	130.0	± 9.6 3
		Y	5.24	68.00	17.15		130.0	
		Z	5.13	67.77	16.87		130.0	
10602- AAB	IEEE 802.11n (HT Mixed, 40MHz, MCS3, 90pc duty cycle)	X	5.23	67.74	16.87	0.46	130.0	± 9.6 *
		Y	5.29	67.89	17.01		130.0	
		Z	5.16	67.59	16.70		130.0	
10603- AAB	IEEE 802.11n (HT Mixed, 40MHz, MCS4, 90pc duty cycle)	X	5.22	67.79	17.04	0.46	130.0	± 9.6 %
1010	moon, sope duty eyeley	Y	5.28	67.91	17.16		130.0	
		Z	5.16	67.66	16.88		130.0	
10001	INCOMENTAL AND A DESCRIPTION OF A DESCRI	X				0.46	130.0	± 9.6 9
10604- AAB	tEEE 602.11n (HT Mixed, 40MHz, MCS5, 90pc duty cycle)		5.14	67.45	16.84	0.46		29.63
-		Y	5.19	67.55	16.95	-	130.0	_
		Z	5.09	67.38	16.70		130.0	
10605- AAB	JEEE 802.11n (HT Mixed, 40MHz, MCS6, 90pc duty cycle)	×	5.18	67.64	16.94	0.46	130.0	± 9.6 %
		Y	5.24	67.79	17.08		130.0	-
		Z	5.12	67.50	16.77		130.0	
10606-	IEEE 802.11n (HT Mixed, 40MHz,	X	5.10	67.52	16.72	0.46	130.0	19.6
BAA	MCS7, 90pc duty cycle)							-
		Y	5.15	67.65	16.86		130.0	-
		Z	5.05	67.41	16.57		130.0	

Certificate No: ES3-3019_Aug18

Page 35 of 39

August 20, 2018

10607- AAB	IEEE 802.11ac WiFi (20MHz, MCS0, 90pc duty cycle)	x	4.29	66.54	16.21	0.46	130,0	± 9.6 %
		Y	4,34	66.66	16.33	-	130.0	
		Z	4.24	66.47	16.06		130.0	
10608- AAB	IEEE 802.11ac WiFi (20MHz, MCS1, 90pc duty cycle)	×	4.39	66.80	16.33	0.46	130.0	± 9,6 %
		Y	4,45	66.92	16.46		130.0	
		Z	4.34	66.71	16.1B		130.0	
10609- AAB	IEEE 802.11ac WiFi (20MHz, MCS2, 90pc duty cycle)	×	4,29	66.63	16.15	0.46	130.0	± 9.6 9
		Y	4.36	66.77	16.28		130.0	
	and a state of the	Z	4.24	66.54	15.98	-	130.0	
10610- AAB	IEEE 802.11ac WiFi (20MHz, MCS3, 90pc duty cycle)	×	4.35	66.83	16.34	0.46	130.0	± 9.6 %
		Y	4.41	66.94	16.46		130.0	
	and a second sec	Z	4.30	66.73	16.18		130.0	
10611- AAB	IEEE 802.11ac WIFI (20MHz, MCS4, 90pc duty cycle)	×	4.25	66.59	16.16	0.46	130.0	± 9.6 %
	and the second second	Y	4.32	66.72	16.29	-	130.0	
		Z	4.20	66.49	15.99		130.0	
10612- AAB	IEEE 802.11ac WIFI (20MHz, MCS5, 90pc duty cycle)	×	4.22	66.65	16.17	0.46	130.0	±9.6%
-		Y	4.29	66.81	16.32		130.0	
	and the second sec	Z	4.17	66.52	15.99		130.0	1000
10613- AAB	IEEE 802.11ac WiFi (20MHz, MCS6, 90pc duty cycle)	x	4,22	65.46	15.99	0.46	130.0	±9.6 %
		Y	4.29	66.63	16.15	-	130.0	
_		Z	4.17	66.34	15.82		130.0	
10614- AAB	IEEE 802.11ac WiFi (20MHz, MCS7, 90pc duty cycle)	×	4.22	66.76	16.30	0.46	130.0	± 9.6 %
		Y	4.28	66.86	16.41		130.0	
		Z	4,17	66.66	16.13		130.0	
10615- AAB	IEEE 802.11ac WiFi (20MHz, MCS8, 90pc duty cycle)	×	4.24	66.41	15.90	0.46	130.0	± 9.6 %
		Y	4.31	66.60	16.08		130.0	
		Z	4,19	66.31	15.73		130.0	
10616- AAB	IEEE 802.11ac WiFi (40MHz, MCS0, 90pc duty cycle)	×	4,03	66.65	16.40	0.46	130.0	± 9.6 %
		Y	4.98	66.76	16.51		130.0	
		Z	4.88	66.57	16.26	-	130.0	
10617- AAB	IEEE 802.11ac WiFi (40MHz, MCS1, 90pc duty cycle)	X	4.95	66.75	16.43	0.46	130.0	19.6 %
		Y	5.01	66.87	16.55		130.0	
		Z	4.90	66.65	16.28		130.0	
10618- AAB	IEEE 802.11ac WiFI (40MHz, MCS2, 90pc duty cycle)	X	4.86	66.79	16.47	0,46	130.0	± 9.6 %
		Y	4.92	66.89	16.57		130.0	
		Z	4.82	66.72	16.33		130.0	
10619- AAB	IEEE 802.11ac WIFI (40MHz, MCS3, 90pc duty cycle)	×	4,94	66.81	16.41	0.46	130.0	±9.6 %
-		Y	5.00	66.95	16.54		130.0	
ani at		Z	4.88	66.69	16.25		130.0	C
10620+ AAB	IEEE 802.11ac WiFi (40MHz, MCS4, 90pc duty cycle)	×	4.94	66.58	16.33	0.46	130.0	± 9,6 %
		Y	5.00	66.71	16.46		130.0	_
		Z	4,89	66,47	16.18	1.2.2	130.0	
10621- AAB	IEEE 802.11ac WIFI (40MHz, MCS5, 90pc duty cycle)	×	4,96	66.73	16.54	0.46	130.0	19.6 %
		Y	5.02	66.81	16.63		130.0	
		Z	4.92	66.66	16.41		130,0	-
10622- AAB	IEEE 802.11ac WiFi (40MHz, MCS6, 90pc duty cycle)	×	4.95	66.82	16.59	0.46	130,0	±9,6 %
		Y	5.00	66.92	16.68		130.0	
		Z	4.90	66.73	16.45		130.0	

Certificate No: ES3-3019_Aug18

Page 36 of 39

August 20, 2018

10523- AAB	IEEE 802.11ac WiFi (40MHz, MCS7, 90pc duty cycle)	×	4.87	66.45	16.24	0.46	130.0	± 9.6 %
		I Y I	4.92	66.58	16.37		130.0	
		Z	4.82	66.36	16.10		130.0	
10624- AAB	IEEE 802.11ac WiFi (40MHz, MCS8, 90pc duty cycle)	X	5.04	66.65	16.42	0.46	130.0	±9.6 %
		Y	5.09	66.77	16.53		130.0	
		Z	4.99	66.56	16.27		130.0	
10625- VAB	IEEE 802.11ac WiFi (40MHz, MCS9, 90pc duty cycle)	×	5.15	66.91	16.62	0.46	130.0	± 9.6 %
		Y	5.20	67.00	16,72		130.0	
		Z	5,10	66.84	16.48		130.0	
10626- AAB	IEEE 802.11ac WiFi (80MHz, MCS0, 90pc duty cycle)	x	5.29	66.59	16.34	0.46	130.0	± 9.6 %
		Y	5.34	66.70	16.45		130.0	
		Z	5.25	66.53	16.22		130.0	
10627- AAB	IEEE 802.11ac WiFi (80MHz, MCS1, 90pc duty cycle)	x	5.52	67.29	16.67	0.46	130.0	±9.6%
		Y	5.58	67.41	16.78		130.0	
		Z	5.47	67.18	16.52		130.0	
10628- AAB	IEEE 802.11ac WiFi (80MHz, MCS2, 90pc duty cycle)	×	5.27	66.55	16.22	0.46	130.0	± 9.6 %
		Y	5.33	66.68	16.34		130.0	
		Z	5.23	66.47	16.09		130.0	
10629- AAB	IEEE 802.11ac WiFi (80MHz, MCS3, 90pc duty cycle)	x	5.51	67.21	16.55	0.46	130.0	± 9.6 %
		Y	5.56	67.34	16.68		130.0	
		Z	5.45	67.08	16.40		130.0	
10630- AAB	IEEE 802.11ac WIFI (80MHz, MCS4, 90pc duty cycle)	x	5.53	67.47	16.69	0.46	130.0	± 9.6 %
		Y	5.60	67.63	16.83		130.0	
		Z	5.44	67.24	16.49		130.0	
10631- AAB	IEEE 802.11ac WiFi (80MHz, MCS5, 90pc duty cycle)	X	5.53	67.60	16.95	0.46	130.0	± 9.6 %
		Y	5.58	67.67	17.02		130.0	
		Ż	5.48	67.48	16.80		130.0	
10632- AAB	IEEE 802.11ac WiFi (80MHz, MCS6, 90pc duty cycle)	X	5.65	67.90	17.12	0.46	130.0	± 9.6 %
		Y	5.69	67.95	17.19		130.0	
		Z	5,59	67,77	16,96		130.0	
10633- AAB	IEEE 802.11ac WiFi (80MHz, MCS7, 90pc duty cycle)	x	5.28	66.62	16.30	0.46	130.0	± 9.6 %
		Y	5.34	66.72	16.40		130.0	
		z	5.25	66.56	16,18		130.0	
10634- AAB	IEEE 802.11ac WiFi (80MHz, MCS8, 90pc duty cycle)	x	5.32	66.86	16.47	0.46	130.0	± 9.6 %
		Y	5.37	66.95	16.56		130.0	
		Z	5.29	66.80	16.35		130.0	
10635- AAB	IEEE 802.11ac WiFi (80MHz, MCS9, 90pc duty cycle)	x	5.16	66.03	15.76	0.46	130.0	± 9.6 %
		Y	5.23	66.21	15.93		130.0	
		Z	5.12	65.96	15.63		130.0	
10536- AAC	IEEE 802.11ac WiFi (160MHz, MCS0, 90pc duty cycle)	×	5.75	66.94	16.43	0.46	130.0	± 9.6 %
		Y.	5.81	67.05	16.54		130.0	
		Z	5.72	66.87	16.31		130.0	
10637- AAC	IEEE 802.11ac WIFI (160MHz, MCS1, 90pc duty cycle)	×	5.87	67.25	16.58	0.46	130.0	± 9.6 %
		Y	5.92	67.36	16.69		130.0	
		Z	5.82	67.15	16.45		130.0	
10638- AAC	IEEE 802.11ac WiFi (160MHz, MCS2, 90pc duty cycle)	x	5.94	67.47	16.67	0.46	130.0	± 9.6 %
		Y	5.99	67.59	16.78		130.0	
		Z	5.89	67.37	16.53		130.0	

Certificate No: ES3-3019_Aug18

Page 37 of 39

August 20, 2018

10639- AAC	IEEE 802.11ac WiFi (160MHz, MCS3,	X	5.83	67.14	16.54	0.46	130.0	± 9.6 %
AAG	90pc duty cycle)	Y	5.88	C7 00	10.05		100.0	
		Z	5.78	67.25 67.06	16.65		130.0	
10640-	IEEE 802.11ac WiFi (160MHz, MCS4.	X	5.74	66.89	16.42	0.46	130.0	
AAC	90pc duty cycle)	1	3,14	00.09	10.30	0.40	130.0	± 9.6 %
		Y	5.80	67.03	16.49	-	130.0	
		Z	5.70	66.81	16.23		130.0	-
10641-	IEEE 802.11ac WiFi (160MHz, MCS5,	X	5.90	67.16	16.52	0.46	130.0	± 9.6 %
AAC	90pc duty cycle)	100			10.04	0.40	100.0	+ 0.0 10
		Y	5.96	67.30	16.65		130.0	
	and the second se	Z	5.84	67.04	16.37		130.0	-
10642-	IEEE 802.11ac WiFi (160MHz, MCS6,	X	5.88	67.23	16.72	0.46	130.0	± 9.6 %
AAC	90pc duty cycle)	1						1.0
_		Y	5.93	67.31	16.81		130.0	
		Z	5,84	67.16	16.60	1.20	130.0	
10643-	IEEE 802.11ac WiFi (160MHz, MCS7,	X	5.73	66.90	16.44	0.46	130.0	± 9.6 %
AAC .	90pc duty cycle)			1.1	1	1911		1.111
_		Y	5.79	67.04	16.57		130.0	
		Z	5.68	66.81	16.31	1	130.0	
10644-	IEEE 802.11ac WiFi (160MHz, MCS8,	X	5.78	67.07	16.55	0.46	130.0	±9.6 %
AAC	90pc duty cycle)	-				1	100 C	-
		Y.	5.84	67.20	16.67		130.0	-
10011	An and the second se	Z	5.74	66.99	16.42		130.0	1.11
10645-	IEEE 802.11ac WiFi (160MHz, MCS9,	X	5.92	67.19	16,58	0.46	130.0	± 9.6 %
AAC	90pc duty cycle)	-						
		Y	5.97	67.31	16.69		130.0	
10010		Z	5.89	67.17	16.48		130.0	-
10646- AAE	LTE-TDD (SC-FDMA, 1 RB, 5 MHz. QPSK, UL Subframe=2,7)	x	13.61	107.81	37.87	9.30	60.0	±9.6 %
		Y	25.75	125,86	44.42		60.0	
	Summer and the second second	Z	9.90	101.38	36.12		60.0	
10647-	LTE-TDD (SC-FDMA, 1 RB, 20 MHz,	X	11.23	103.89	36.78	9.30	60.0	±9.6 %
AAE	QPSK, UL Subframe=2.7)							
		Y	19.74	119.98	42.91		60.0	
		Z	8.22	97.43	34.92	-	60.0	
10648- AAA	CDMA2000 (1x Advanced)	x	0.30	60.00	4.87	0.00	150.0	± 9.6 %
		Y	0.33	60.00	5.44		150.0	
		Z	0.30	60.00	4.85		150.0	
10652- AAC	LTE-TDD (OFDMA, 5 MHz, E-TM 3.1, Clipping 44%)	X	3.70	69.20	16.78	2.23	B0.0	± 9.6 %
110	authority and	Y	3.79	69.35	16.97		80.0	
	2.0	Z	3.51	68.59	16.38		80.0	
10653- AAC	LTE-TDD (OFDMA, 10 MHz, E-TM 3.1, Clipping 44%)	X	4,13	67,83	17.08	2.23	80.0	± 9.6 %
	and the second second	Y	4.22	67.99	17.23		80.0	
		Z	4.00	67.46	16.80		80.0	
10654- AAC	LTE-TDD (OFDMA, 15 MHz, E-TM 3.1, Clipping 44%)	X	4.14	67.26	17.12	2.23	80.0	± 9.6 %
10.414	anistent da int	Y	4.22	67.42	17.27		80.0	
		Z	4.03	66.92	16.87		80.0	
10655-	LTE-TOD (OFDMA, 20 MHz, E-TM 3.1,	X	4.03	67.03	10.07	2.23	80.0	±9.6 %
AAD	Clipping 44%)			-40, 157,	1	6.20	1000	T 8/0 %
		Y	4.30	67.21	17.29	_	80.0	
00000	But a Man from theory to all	Z	4.11	66.70	16.90	10.00	80.0	1447
10658- AAA	Pulse Waveform (200Hz, 10%)	x	100.00	111.26	26.36	10,00	50.0	± 9.6 %
_		Y	100.00	114.45	28.17		50.0	
		Z	100.00	110.83	26.00		50.0	1
10659- AAA	Pulse Waveform (200Hz, 20%)	×	100.00	108.50	24.19	6.99	60.0	± 9.6 %
	-	Y	100.00	112.09	26.11		60.0	-

Certificate No: ES3-3019_Aug18

Page 38 of 39

August 20, 2018

10660- AAA	Pulse Waveform (200Hz, 40%)	×	100.00	108.32	22.04	3.98	80.0	±9.6 %
		Y	100.00	111.36	24.50		80.0	
		Z	100.00	107.90	22.58		80.0	
10661- AAA	Pulse Waveform (200Hz, 60%)	X	100.00	103.65	19.77	2.22	100.0	±9.6 %
		Y	100.00	112.12	23.59		100.0	
		z	100.00	106.59	20.90		100.0	
10662- AAA	Pulse Waveform (200Hz, 80%)	x	100.00	90.92	13.24	0.97	120.0	± 9.6 %
		Y	100.00	110.88	21.41		120.0	
		Z	100.00	97.17	15.68		120.0	

^c Uncertainty is determined using the max, deviation from linear response applying rectangular distribution and is expressed for the square of the field value.

Certificate No: ES3-3019_Aug18

Page 39 of 39

DIPOLE CALIBRATION CERTIFICATES

Calibration Laboratory of Schmid & Partner Engineering AG zeughausstrasse 43, 8004 Zurich, Switzerland

BACL

Client



Schweizerischer Kalibrierdlenst Service sulsse d'étalonnage Servizlo svizzero di taratura Swiss Calibration Service

Accreditation No.: SCS 0108

Accredited by the Swiss Accreditation Service (SAS) The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

Certificate No: CLA150-4020_Nov16

s

С

S

Object	CLA150 - SN: 4020		
Calibration procedure(s)	QA CAL-15.v8		
		edure for system validation source	ces below 700 MHz
Calibration date:	November 08, 2	016	
This calibration cartificate docum	tents the traceability to na	lional standards, which realize the physical u	nits of measurements (SI)_
		probability are given on the following pages a	
Calibration Equipment used (M&		ary facility: environment temperature $(22 \pm 3)^{\circ}$	'C and humidity < 70%.
Primary Standards	ID #	Cal Date (Certificate No.)	Scheduled Calibration
Power meter NRP	SN: 104778	06-Apr-16 (No. 217-02288/02289)	Apr-17
and a standard of the set of the	SN: 103244		
	PO1.41 1.20201-00-0	00-ADP-16 (No. 217-02288)	
ower sensor NRP-Z91	SN: 103245	06-Apr-16 (No. 217-02288) 06-Apr-16 (No. 217-02289)	Apr-17
ower sensor NRP-Z91		06-Apr-16 (No. 217-02289)	Apr-17
ower sensor NRP-Z91 eference 30 dB Attenuator	SN: 103245 SN: 5129 (30b)	06-Apr-16 (No. 217-02289) 05-Apr-16 (No. 217-02294)	Apr-17 Apr-17
ower sensor NRP-Z91 eference 30 dB Attenuator Ipe-N mismatch combination	SN: 103245	06-Apr-16 (No. 217-02289) 05-Apr-16 (No. 217-02294) 05-Apr-16 (No. 217-02295)	Apr-17 Apr-17 Apr-17
ower sensor NRP-Z91 eferonce 30 dB Attenuator ype-N mismatch combination eference Probe EX3DV4	SN: 103245 SN: 5129 (30b) SN: 5047,2 / 06327	06-Apr-16 (No. 217-02289) 05-Apr-16 (No. 217-02294)	Apr-17 Apr-17
Yower sensor NRP-Z91 leferonce 30 dB Attenuator ype:N mismatch combination leference Probe EX3DV4 WAE4 econdary Standards	SN: 103245 SN: 5129 (30b) SN: 5047.2 / 06327 SN: 3877	06-Apr-16 (No. 217-02289) 05-Apr-16 (No. 217-02294) 05-Apr-16 (No. 217-02295) 31-Dec-15 (No. EX3-3877_Dec15) 12-Aug-16 (No. DAE4-654_Aug16)	Apr-17 Apr-17 Apr-17 Dec-16 Aug-17
ower sensor NRP-291 eferonce 30 dB Attenuator ype-N mismatch combination eference Probe EX3DV4 AE4 econdary Standards ower meter E4419B	SN: 103245 SN: 5129 (30b) SN: 5047.2 / 06327 SN: 3877 SN: 654	06-Apr-16 (No. 217-02289) 05-Apr-16 (No. 217-02294) 05-Apr-16 (No. 217-02295) 31-Dec-15 (No. EX3-3877_Dec15) 12-Aug-16 (No. DAE4-654_Aug16) Check Date (in house)	Apr-17 Apr-17 Apr-17 Dec-16 Aug-17 Scheduled Check
ower sensor NRP-291 leferonce 30 dB Attenuator ype-N mismatch combination leference Probe EX3DV4 wAE4 econdary Standards ower meter E4419B ower sensor E4412A	SN: 103245 SN: 5129 (30b) SN: 5047.2 / 06327 SN: 3877 SN: 654	06-Apr-16 (No. 217-02289) 05-Apr-16 (No. 217-02294) 05-Apr-16 (No. 217-02295) 31-Dec-15 (No. EX3-3877_Dec15) 12-Aug-16 (No. DAE4-654_Aug16)	Apr-17 Apr-17 Apr-17 Dec-16 Aug-17 Scheduled Check In house check: Jun-18
Vower sensor NRP-291 leferonce 30 dB Attenuator ype-N mismatch combination leference Probe EX3DV4 VAE4 econdary Standards ower meter E44198 ower sensor E4412A ower sensor E4412A	SN: 103245 SN: 5129 (30b) SN: 5047,2 / 06327 SN: 654 ID # SN: GB41293674 SN: MY41498087 SN: 000110210	06-Apr-16 (No. 217-02289) 05-Apr-16 (No. 217-02294) 05-Apr-16 (No. 217-02295) 31-Dec-15 (No. EX3-3877_Dec15) 12-Aug-16 (No. DAE4-654_Aug16) Check Date (in house) 08-Apr-16 (No. 217-02285/02284)	Apr-17 Apr-17 Apr-17 Dec-16 Aug-17 Scheduled Check In house check; Jun-18 In house check; Jun-18
Power sensor NRP-291 Reference 30 dB Attenuator Vpe-N mismatch combination Reference Probe EX3DV4 VAE4 Recondary Standards Tower meter E4419B ower sensor E4412A ower sensor E4412A Figurerator HP 8848C	SN: 103245 SN: 5129 (30b) SN: 5047,2 / 06327 SN: 654 SN: 654 SN: 664 SN: 6641293674 SN: MY41498087 SN: 000110210 SN: US3642U01700	06-Apr-16 (No. 217-02289) 05-Apr-16 (No. 217-02294) 05-Apr-16 (No. 217-02295) 31-Dec-15 (No. EX3-3877_Dec15) 12-Aug-16 (No. DAE4-654_Aug16) Check Date (in house) 06-Apr-16 (No. 217-02285/02284) 06-Apr-16 (No. 217-02284)	Apr-17 Apr-17 Apr-17 Dec-16 Aug-17 Scheduled Check In house check; Jun-18 In house check; Jun-18
Vower sensor NRP-291 leferonce 30 dB Attenuator ype-N mismatch combination leference Probe EX3DV4 VAE4 econdary Standards ower meter E4419B ower sensor E4412A ower sensor E4412A F generator HP 8848C	SN: 103245 SN: 5129 (30b) SN: 5047,2 / 06327 SN: 654 ID # SN: GB41293674 SN: MY41498087 SN: 000110210	06-Apr-16 (No. 217-02289) 05-Apr-16 (No. 217-02294) 05-Apr-16 (No. 217-02295) 31-Dec-15 (No. EX3-3877, Dec15) 12-Aug-16 (No. DAE4-654_Aug16) Check Date (in house) 06-Apr-16 (No. 217-02285/02284) 06-Apr-16 (No. 217-02285)	Apr-17 Apr-17 Apr-17 Dec-16 Aug-17 Scheduled Check In house check; Jun-18 In house check; Jun-18
Power sensor NRP-291 Reference 30 dB Attenuator Type: N mismatch combination Reference Probe EX3DV4 VAE4 Secondary Standards Power meter E44198 Power sensor E4412A Power sensor E4412A Power sensor E4412A Figunerator HP 8648C Reference HP 8753E	SN: 103245 SN: 5129 (30b) SN: 5047,2 / 06327 SN: 654 SN: 654 SN: 664 SN: 6641293674 SN: MY41498087 SN: 000110210 SN: US3642U01700	06-Apr-16 (No. 217-02289) 05-Apr-16 (No. 217-02294) 05-Apr-16 (No. 217-02295) 31-Dec-15 (No. EX3-3877_Dec15) 12-Aug-16 (No. DAE4-654_Aug16) Check Date (in hause) 06-Apr-16 (No. 217-02285/02284) 06-Apr-16 (No. 217-02285) 06-Apr-16 (No. 217-02284) 04-Aug-99 (in house check Jun-16)	Apr-17 Apr-17 Apr-17 Dec-16 Aug-17 Scheduled Check In house check: Jun-18 In house check: Jun-18 In house check: Jun-18 In house check: Jun-18 In house check: Jun-18
Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 30 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards Power meter E44198 Power sensor E44198 Power sensor E4412A Te generator HP 8649C Network Analyzer HP 8753E Callibrated by:	SN: 103245 SN: 5129 (30b) SN: 5047,2 / 06327 SN: 654 SN: 654 SN: 6B41293874 SN: MY41498087 SN: 000110210 SN: US3642U01700 SN: US37390585	06-Apr-16 (No. 217-02289) 05-Apr-16 (No. 217-02294) 05-Apr-16 (No. 217-02295) 31-Dec-15 (No. EX3-3877_Dec15) 12-Aug-16 (No. DAE4-654_Aug16) Check Date (in house) 06-Apr-16 (No. 217-02285/02284) 06-Apr-16 (No. 217-02284) 06-Apr-16 (No. 217-02284) 04-Aug-99 (in house check Jun-16) 18-Oct-01 (in house check Oct-16)	Apr-17 Apr-17 Apr-17 Dec-16 Aug-17 Scheduled Check In house check: Jun-18 In house check: Jun-18 In house check: Jun-18 In house check: Jun-18

Certificate No: CLA150-4020_Nov16

Page 1 of 8

Calibration Laboratory of Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland





S Schweizerischer Kallbrierdiumst Service suisse d'étalonnage Servizio svizzero di taratura S swiss Callbration Service

Accreditation No.: SCS 0108

Accredited by the Swiss Accreditation Service (SAS) The Swiss Accreditation Service Is one of the signatories to the EA Multilatoral Agreement for the recognition of calibration cortificates

Glossary:

TSL	tissue simulating liquid
ConvF	sensitivity in TSL / NORM x,y,z
N/A	not applicable or not measured

Calibration is Performed According to the Following Standards:

- a) IEEE Std 1528-2013, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", June 2013
- b) IEC 62209-1, "Procedure to measure the Specific Absorption Rate (SAR) for hand-held devices used in close proximity to the ear (frequency range of 300 MHz to 3 GHz)", February 2005
- c) IEC 62209-2, "Procedure to determine the Specific Absorption Rate (SAR) for wireless communication devices used in close proximity to the human body (frequency range of 30 MHz to 6 GHz)*, March 2010
- d) KDB 865664, "SAR Measurement Requirements for 100 MHz to 6 GHz"

Additional Documentation:

e) DASY4/5 System Handbook

Methods Applied and Interpretation of Parameters:

- Measurement Conditions: Further details are available from the Validation Report at the end
 of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL: The source is mounted in a touch configuration below the center marking of the flat phantom.
- Return Loss: This parameter is measured with the source positioned under the liquid filled phantom (as described in the measurement condition clause). The Return Loss ensures low reflected power. No uncertainty required.
- · SAR measured: SAR measured at the stated antenna input power.
- SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

Gentilicate No: CLA150-4020_Nov16

Page 2 of 8

Measurement Conditions

DASY system configuration, as far as not given on page 1.

DASY Version	DASY5	V52.8.8	
Extrapolation	Advanced Extrapolation		
Phantom	ELI4 Flat Phantom	Shell thickness: 2 ± 0.2 mm	
EUT Positioning	Touch Position		
Zoom Scan Resolution	dx, dy = 4.0 mm, dz = 1.4 mm	Graded Ratio = 1.4 (Z direction)	
Frequency	150 MHz ± 1 MHz		

Head TSL parameters

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	52.3	0.76 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	50.1 ± 6 %	0.75 mho/m ± 6 %
Head TSL temperature change during test	< 0.5 °C		

SAR result with Head TSL

SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition	
SAR measured	1 W input power	3.64 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	3.64 W/kg ± 18.4 % (k=2)

SAR averaged over 10 cm ³ (10 g) of Head TSL	condition	
SAR measured	1 W input power	2.46 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	2.47 W/kg ± 18.0 % (k=2)

Body TSL parameters

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Body TSL parameters	22.0 °C	61.9	0.80 mho/m
Measured Body TSL parameters	(22.0 ± 0.2) °C	61.4 ± 6 %	0.82 mho/m ± 6 %
Body TSL temperature change during test	< 0.5 °C		

SAR result with Body TSL

SAR averaged over 1 cm ³ (1 g) of Body TSL	Condition	
SAR measured	1 W input power	3.81 W/kg
SAR for nominal Body TSL parameters	normalized to 1W	3.73 W/kg ± 18.4 % (k=2)
SAR averaged over 10 cm ³ (10 g) of Body TSL	condition	
SAR measured	1 W input power	2.57 W/kg
		9

Certificate No: CLA150-4020_Nov16

Page 3 of 8

Appendix (Additional assessments outside the scope of SCS 0108)

Antenna Parameters with Head TSL

Impedance, transformed to feed point	44.4 Ω - 1.8 jΩ
Return Loss	- 24.1 dB

Antenna Parameters with Body TSL

Impedance, transformed to feed point	45.9 Ω - 6.0 jΩ
Return Loss	- 22.5 dB

Additional EUT Data

Manufactured by	SPEAG		
Manufactured on	November 25, 2015		

Certificate No: CLA150-4020_Nov16

Page 4 of 8

DASY5 Validation Report for Head TSL

Date: 07.11.2016

Test Laboratory: SPEAG, Zurich, Switzerland

DUT: CLA-150; Type: CLA-150; Serial: 4020


Communication System: UID 0 - CW; Frequency: 150 MHz. Medium parameters used: f = 150 MHz; $\sigma = 0.75 \text{ S/m}$; $\varepsilon_r = 50.1$; $\rho = 1000 \text{ kg/m}^3$ Phantom section: Flat Section Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2011)

DASY52 Configuration:

- Probe: EX3DV4 SN3877; ConvF(12.02, 12.02, 12.02); Calibrated: 31,12.2015;
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn654; Calibrated: 12.08.2016
- Phantom: ELI v4.0; Type: QDOVA001BB; Serial: TP:1003
- DASY52 52.8.8(1258); SEMCAD X 14.6.10(7372)

CLA Calibration for HSL-LF Tissue/CLA150, touch configuration, Pin=1W/Area Scan (81x81x1): Interpolated grid: dx=1.500 mm, dy=1.500 mm Maximum value of SAR (interpolated) = 5.06 W/kg


CLA Calibration for HSL-LF Tissue/CLA150, touch configuration, Pin=1W/Zoom Scan, dist=1.4mm (8x10x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 81.89 V/m; Power Drift = 0.02 dB Pcak SAR (extrapolated) = 6.81 W/kg SAR(1 g) = 3.64 W/kg; SAR(10 g) = 2.46 W/kg Maximum value of SAR (measured) = 5.06 W/kg



Certificate No: CLA150-4020_Nov16

Page 5 of 8

Impedance Measurement Plot for Head TSL



Certificate No: CLA150-4020_Nov16

Page 6 of 8

DASY5 Validation Report for Body TSL

Date: 08.11,2016

Test Laboratory: SPEAG, Zurich, Switzerland

DUT: CLA-150; Type: CLA-150; Serial: 4020


Communication System: UID 0 - CW; Frequency: 150 MHz Medium parameters used: f = 150 MHz; $\sigma = 0.82$ S/m; $\epsilon_r = 61.4$; $\rho = 1000$ kg/m³ Phantom section: Flat Section Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2011)

DASY52 Configuration:

- Probe: EX3DV4 SN3877; ConvF(11.44, 11.44, 11.44); Calibrated: 31.12.2015;
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn654; Calibrated: 12.08.2016
- Phantom: ELI v4.0; Type: QDOVA001BB; Serial: TP:1003
- DASY52 52.8.8(1258); SEMCAD X 14.6.10(7372)


CLA Calibration for MSL-LF Tissue/CLA150, touch configuration, Pin=1W/Area Scan (81x81x1): Interpolated grid: dx=1,500 mm, dy=1,500 mm Maximum value of SAR (interpolated) = 5.32 W/kg


CLA Calibration for MSL-LF Tissue/CLA150, touch configuration, Pin=1W/Zoom Scan, dist=1.4mm (8x10x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 80.95 V/m; Power Drift = -0.07 dB Peak SAR (extrapolated) = 7.02 W/kg SAR(1 g) = 3.81 W/kg; SAR(10 g) = 2.57 W/kg Maximum value of SAR (measured) = 5.27 W/kg



Certificate No: CLA150-4020_Nov16

Page 7 of 8





Certificate No: CLA150-4020_Nov16

Page 8 of 8