

Appendix C for KSCR230200021401

Calibration Certificate

Object	Apply	No	Model	SN	Calibration Date	Due date of calibration
Dipole	<input type="checkbox"/>	1	CLA150	4025	2021/04/26	2024/04/25
	<input type="checkbox"/>	2	D450V3	1103	2021/04/21	2024/04/20
	<input type="checkbox"/>	3	D750V3	1188	2022/03/29	2025/03/28
	<input type="checkbox"/>	4	D835V2	4d114	2022/03/31	2025/03/30
	<input type="checkbox"/>	5	D900V2	1d079	2022/06/07	2025/06/06
	<input type="checkbox"/>	6	D1800V2	2d170	2022/03/31	2025/03/30
	<input type="checkbox"/>	7	D1900V2	5d136	2022/06/07	2025/06/06
	<input type="checkbox"/>	8	D2000V2	1041	2022/06/06	2025/06/05
	<input type="checkbox"/>	9	D2300V2	1096	2022/03/31	2025/03/30
	<input checked="" type="checkbox"/>	10	D2450V2	817	2022/04/01	2025/03/31
	<input type="checkbox"/>	11	D2600V2	1158	2022/03/31	2025/03/30
	<input checked="" type="checkbox"/>	12	D5GHzV2	1095	2022/06/01	2025/05/31
DAE	<input checked="" type="checkbox"/>	13	DAE4	1245	2022/05/30	2023/05/29
Probe	<input checked="" type="checkbox"/>	14	EX3DV4	7767	2022/10/28	2023/10/27



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN_Doccheck@sgs.com
 No.10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300 ☎(86-512)57355888 ☎(86-512)57370818 www.sgs.com.cn
 中国·江苏·昆山市留学院创业园伟业路10号 邮编 215300 ☎(86-512)57355888 ☎(86-512)57370818 sgs.china@sgs.com

1 Dipole

1.1 CLA150 - SN 4025

<p>Calibration Laboratory of Schmid & Partner Engineering AG Zehrgassestrasse 43, 8004 Zurich, Switzerland</p> <p>Accredited by the Swiss Accreditation Service (SAS) The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates</p> <p>Client: SGS-CN (Auden) Certificate No.: CLA150-4025_Apr21</p> <p style="text-align: right;">Accreditation No.: SCS 0108</p> <hr/> <p style="text-align: center;">CALIBRATION CERTIFICATE</p> <p>Object: CLA150 - SN: 4025</p> <p>Calibration procedure(s): QA CAL-15.09 Calibration Procedure for SAR Validation Sources below 700 MHz</p> <p>Calibration date: April 26, 2021</p> <p>The calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.</p> <p>All calibrations have been conducted in the closed laboratory facility: environment temperature (22 ± 3)°C and humidity < 70%.</p> <p>Calibration Equipment used (MPE critical for calibration)</p> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Primary Standards</th> <th>ID #</th> <th>Cal Date (Certificate No.)</th> <th>Scheduled Calibration</th> </tr> </thead> <tbody> <tr> <td>Power meter NRP</td> <td>SN: 10475</td> <td>09-Apr-21 (No. 217-0320102920)</td> <td>Apr-22</td> </tr> <tr> <td>Power sensor NRP Z01</td> <td>SN: 10364</td> <td>09-Apr-21 (No. 217-03201)</td> <td>Apr-22</td> </tr> <tr> <td>Power sensor NRP Z01</td> <td>SN: 10365</td> <td>09-Apr-21 (No. 217-03202)</td> <td>Apr-22</td> </tr> <tr> <td>Reference 20 dB Attenuator</td> <td>SN: C22662 (20)</td> <td>09-Apr-21 (No. 217-03343)</td> <td>Apr-22</td> </tr> <tr> <td>Type-N mismatch combination</td> <td>SN: 31085 / 0037</td> <td>09-Apr-21 (No. 217-03344)</td> <td>Apr-22</td> </tr> <tr> <td>Reference Probe EX30N4 (DIE4)</td> <td>SN: 3877</td> <td>30-Dec-20 (No. EX30377_Dec20)</td> <td>Dec-21</td> </tr> <tr> <td></td> <td>SN: 664</td> <td>26-Jun-20 (No. DMS4-656_Jun20)</td> <td>Jun-21</td> </tr> </tbody> </table> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Secondary Standards</th> <th>ID #</th> <th>Check Date (in house)</th> <th>Scheduled Check</th> </tr> </thead> <tbody> <tr> <td>Power meter E4118B</td> <td>SN: G81282874</td> <td>06-Apr-18 (in house check Jun-20)</td> <td>In house check Jun-22</td> </tr> <tr> <td>Power sensor E4112A</td> <td>SN: MY41498087</td> <td>06-Apr-18 (in house check Jun-20)</td> <td>In house check Jun-22</td> </tr> <tr> <td>Power sensor E4112A</td> <td>SN: 00010210</td> <td>06-Apr-18 (in house check Jun-20)</td> <td>In house check Jun-22</td> </tr> <tr> <td>RF generator HP 8594D</td> <td>SN: US240031070</td> <td>04-Aug-09 (in house check Jun-20)</td> <td>In house check Jun-22</td> </tr> <tr> <td>Network Analyzer Agilent E8363A</td> <td>SN: US41003477</td> <td>31-Mar-14 (in house check Oct-20)</td> <td>In house check Oct-21</td> </tr> </tbody> </table> <p>Calibrated by: Jeffrey Katzman Function: Laboratory Technician Signature: <i>[Signature]</i></p> <p>Approved by: Kate Polovic Technical Manager Signature: <i>[Signature]</i></p> <p>This calibration certificate shall not be reproduced except in full without written approval of the laboratory. Issued: April 26, 2021</p> <p>Certificate No: CLA150-4025_Apr21 Page 1 of 8</p>	Primary Standards	ID #	Cal Date (Certificate No.)	Scheduled Calibration	Power meter NRP	SN: 10475	09-Apr-21 (No. 217-0320102920)	Apr-22	Power sensor NRP Z01	SN: 10364	09-Apr-21 (No. 217-03201)	Apr-22	Power sensor NRP Z01	SN: 10365	09-Apr-21 (No. 217-03202)	Apr-22	Reference 20 dB Attenuator	SN: C22662 (20)	09-Apr-21 (No. 217-03343)	Apr-22	Type-N mismatch combination	SN: 31085 / 0037	09-Apr-21 (No. 217-03344)	Apr-22	Reference Probe EX30N4 (DIE4)	SN: 3877	30-Dec-20 (No. EX30377_Dec20)	Dec-21		SN: 664	26-Jun-20 (No. DMS4-656_Jun20)	Jun-21	Secondary Standards	ID #	Check Date (in house)	Scheduled Check	Power meter E4118B	SN: G81282874	06-Apr-18 (in house check Jun-20)	In house check Jun-22	Power sensor E4112A	SN: MY41498087	06-Apr-18 (in house check Jun-20)	In house check Jun-22	Power sensor E4112A	SN: 00010210	06-Apr-18 (in house check Jun-20)	In house check Jun-22	RF generator HP 8594D	SN: US240031070	04-Aug-09 (in house check Jun-20)	In house check Jun-22	Network Analyzer Agilent E8363A	SN: US41003477	31-Mar-14 (in house check Oct-20)	In house check Oct-21	<p>Calibration Laboratory of Schmid & Partner Engineering AG Zehrgassestrasse 43, 8004 Zurich, Switzerland</p> <p>Accredited by the Swiss Accreditation Service (SAS) The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates</p> <p>Accreditation No.: SCS 0108</p> <hr/> <p>Glossary:</p> <p>TSL: Issue simulating liquid sensitivity in TSL / NORM x,y,z</p> <p>ConvF: not applicable or not measured</p> <p>N/A: not applicable or not measured</p> <p>Calibration is Performed According to the Following Standards:</p> <ol style="list-style-type: none"> IEEE Std 1528-2013, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", June 2013 IEC 62209-1, "Measurement procedure for the assessment of Specific Absorption Rate (SAR) from hand-held and body-mounted devices used next to the ear (frequency range of 300 MHz to 6 GHz)", July 2016 IEC 62209-2, "Procedure to determine the Specific Absorption Rate (SAR) for wireless communication devices used in close proximity to the human body (frequency range of 30 MHz to 6 GHz)", March 2010 KDB 865664, "SAR Measurement Requirements for 100 MHz to 6 GHz" <p>Additional Documentation:</p> <ol style="list-style-type: none"> DASY4/5 System Handbook <p>Methods Applied and Interpretation of Parameters:</p> <ul style="list-style-type: none"> Measurement Conditions: Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated. Antenna Parameters with TSL: The source is mounted in a touch configuration below the center marking of the flat phantom. Return Loss: This parameter is measured with the source positioned under the liquid filled phantom (as described in the measurement condition clause). The Return Loss ensures low reflected power. No uncertainty required. SAR measured: SAR measured at the stated antenna input power. SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna connector. SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result. <p>The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.</p> <hr/> <p>Certificate No: CLA150-4025_Apr21 Page 2 of 8</p>		
Primary Standards	ID #	Cal Date (Certificate No.)	Scheduled Calibration																																																								
Power meter NRP	SN: 10475	09-Apr-21 (No. 217-0320102920)	Apr-22																																																								
Power sensor NRP Z01	SN: 10364	09-Apr-21 (No. 217-03201)	Apr-22																																																								
Power sensor NRP Z01	SN: 10365	09-Apr-21 (No. 217-03202)	Apr-22																																																								
Reference 20 dB Attenuator	SN: C22662 (20)	09-Apr-21 (No. 217-03343)	Apr-22																																																								
Type-N mismatch combination	SN: 31085 / 0037	09-Apr-21 (No. 217-03344)	Apr-22																																																								
Reference Probe EX30N4 (DIE4)	SN: 3877	30-Dec-20 (No. EX30377_Dec20)	Dec-21																																																								
	SN: 664	26-Jun-20 (No. DMS4-656_Jun20)	Jun-21																																																								
Secondary Standards	ID #	Check Date (in house)	Scheduled Check																																																								
Power meter E4118B	SN: G81282874	06-Apr-18 (in house check Jun-20)	In house check Jun-22																																																								
Power sensor E4112A	SN: MY41498087	06-Apr-18 (in house check Jun-20)	In house check Jun-22																																																								
Power sensor E4112A	SN: 00010210	06-Apr-18 (in house check Jun-20)	In house check Jun-22																																																								
RF generator HP 8594D	SN: US240031070	04-Aug-09 (in house check Jun-20)	In house check Jun-22																																																								
Network Analyzer Agilent E8363A	SN: US41003477	31-Mar-14 (in house check Oct-20)	In house check Oct-21																																																								
<p>Measurement Conditions DASY system configuration, as far as not given on page 1.</p> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>DASY Version</th> <th>DASY5</th> <th>V32.10.4</th> </tr> </thead> <tbody> <tr> <td>Extrapolation</td> <td>Advanced Extrapolation</td> <td></td> </tr> <tr> <td>Phantom</td> <td>ELN Flat Phantom</td> <td>Shell thickness: 2 ± 0.2 mm</td> </tr> <tr> <td>EUT Positioning</td> <td>Touch Position</td> <td></td> </tr> <tr> <td>Zoom Scan Resolution</td> <td>dx, dy = 4.0 mm, dz = 1.4 mm</td> <td>Graded Ratio = 1.4 (Z direction)</td> </tr> <tr> <td>Frequency</td> <td>150 MHz ± 1 MHz</td> <td></td> </tr> </tbody> </table> <p>Head TSL parameters The following parameters and calculations were applied.</p> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Parameter</th> <th>Temperature</th> <th>Permittivity</th> <th>Conductivity</th> </tr> </thead> <tbody> <tr> <td>Nominal Head TSL parameters</td> <td>22.0 °C</td> <td>52.3</td> <td>0.75 mho/m</td> </tr> <tr> <td>Measured Head TSL parameters</td> <td>(22.0 ± 0.2) °C</td> <td>51.1 ± 6 %</td> <td>0.75 mho/m ± 6 %</td> </tr> <tr> <td>Head TSL temperature change during test</td> <td>< 0.5 °C</td> <td>---</td> <td>---</td> </tr> </tbody> </table> <p>SAR result with Head TSL</p> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>SAR averaged over 1 cm³ (1 g) of Head TSL</th> <th>Condition</th> <th></th> </tr> </thead> <tbody> <tr> <td>SAR measured</td> <td>1 W input power</td> <td>3.90 W/kg</td> </tr> <tr> <td>SAR for nominal Head TSL parameters</td> <td>normalized to 1W</td> <td>3.88 W/kg ± 18.4 % (k=2)</td> </tr> </tbody> </table> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>SAR averaged over 10 cm³ (10 g) of Head TSL</th> <th>condition</th> <th></th> </tr> </thead> <tbody> <tr> <td>SAR measured</td> <td>1 W input power</td> <td>2.60 W/kg</td> </tr> <tr> <td>SAR for nominal Head TSL parameters</td> <td>normalized to 1W</td> <td>2.59 W/kg ± 18.0 % (k=2)</td> </tr> </tbody> </table> <hr/> <p>Certificate No: CLA150-4025_Apr21 Page 3 of 8</p>	DASY Version	DASY5	V32.10.4	Extrapolation	Advanced Extrapolation		Phantom	ELN Flat Phantom	Shell thickness: 2 ± 0.2 mm	EUT Positioning	Touch Position		Zoom Scan Resolution	dx, dy = 4.0 mm, dz = 1.4 mm	Graded Ratio = 1.4 (Z direction)	Frequency	150 MHz ± 1 MHz		Parameter	Temperature	Permittivity	Conductivity	Nominal Head TSL parameters	22.0 °C	52.3	0.75 mho/m	Measured Head TSL parameters	(22.0 ± 0.2) °C	51.1 ± 6 %	0.75 mho/m ± 6 %	Head TSL temperature change during test	< 0.5 °C	---	---	SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition		SAR measured	1 W input power	3.90 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	3.88 W/kg ± 18.4 % (k=2)	SAR averaged over 10 cm ³ (10 g) of Head TSL	condition		SAR measured	1 W input power	2.60 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	2.59 W/kg ± 18.0 % (k=2)	<p>Appendix (Additional assessments outside the scope of SCS 0108)</p> <p>Antenna Parameters with Head TSL</p> <table border="1" style="width: 100%; border-collapse: collapse;"> <tbody> <tr> <td>Impedance, transformed to feed point</td> <td>47.9 Ω ± 1.5 Ω</td> </tr> <tr> <td>Return Loss</td> <td>-31.4 dB</td> </tr> </tbody> </table> <p>Additional EUT Data</p> <table border="1" style="width: 100%; border-collapse: collapse;"> <tbody> <tr> <td>Manufactured by</td> <td>SPEAG</td> </tr> </tbody> </table> <hr/> <p>Certificate No: CLA150-4025_Apr21 Page 4 of 8</p>	Impedance, transformed to feed point	47.9 Ω ± 1.5 Ω	Return Loss	-31.4 dB	Manufactured by	SPEAG
DASY Version	DASY5	V32.10.4																																																									
Extrapolation	Advanced Extrapolation																																																										
Phantom	ELN Flat Phantom	Shell thickness: 2 ± 0.2 mm																																																									
EUT Positioning	Touch Position																																																										
Zoom Scan Resolution	dx, dy = 4.0 mm, dz = 1.4 mm	Graded Ratio = 1.4 (Z direction)																																																									
Frequency	150 MHz ± 1 MHz																																																										
Parameter	Temperature	Permittivity	Conductivity																																																								
Nominal Head TSL parameters	22.0 °C	52.3	0.75 mho/m																																																								
Measured Head TSL parameters	(22.0 ± 0.2) °C	51.1 ± 6 %	0.75 mho/m ± 6 %																																																								
Head TSL temperature change during test	< 0.5 °C	---	---																																																								
SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition																																																										
SAR measured	1 W input power	3.90 W/kg																																																									
SAR for nominal Head TSL parameters	normalized to 1W	3.88 W/kg ± 18.4 % (k=2)																																																									
SAR averaged over 10 cm ³ (10 g) of Head TSL	condition																																																										
SAR measured	1 W input power	2.60 W/kg																																																									
SAR for nominal Head TSL parameters	normalized to 1W	2.59 W/kg ± 18.0 % (k=2)																																																									
Impedance, transformed to feed point	47.9 Ω ± 1.5 Ω																																																										
Return Loss	-31.4 dB																																																										
Manufactured by	SPEAG																																																										



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

No.10, Weyue Road, Innovation Park, Kunshan, Jiangsu, China 215300 t(86-512)57355888 f(86-512)57370818 www.sgsgroup.com.cn
 中国·江苏·昆山市留学生创业园伟业路10号 邮编 215300 t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com

DASY5 Validation Report for Head TSL

Date: 26.04.2021

Test Laboratory: SPEAG, Zurich, Switzerland

DUT: CLA150; Type: CLA150; Serial: CLA150 - SN: 4025

Communication System: UID 0 - CW; Frequency: 150 MHz
Medium parameters used: $f = 150 \text{ MHz}$; $\sigma = 0.76 \text{ S/m}$; $\epsilon = 51.1$; $\rho = 1000 \text{ kg/m}^3$
Phantom section: Flat Section
Measurement Standard: DASY5 (IEE/IEC/ANSI C63.19-2011)

DASY52 Configuration:

- Probe: EX3DV4 - SN3877; ConvF(12.51, 12.51, 12.51) @ 150 MHz; Calibrated: 30.12.2020
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DA54 Sn654; Calibrated: 26.06.2020
- Phantom: ELI v4.0; Type: QDOVA001BB; Serial: TP.1003
- DASY52 52.10.4(1527); SEMCAD X 14.6.14(7483)

CLA Calibration for HSL-LF Tissue/CLA150, touch configuration, Pin=1W/Zoom Scan, dist=1.4mm (8x10x8)/Cube D: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm
Reference Value = 85.93 W/m; Power Drift = -0.02 dB
Peak SAR (extrapolated) = 7.36 W/kg
SAR(1 g) = 3.90 W/kg; SAR(10 g) = 2.60 W/kg
Smallest distance from peaks to all points 3 dB below: Larger than measurement grid (> 30mm)
Ratio of SAR at M2 to SAR at M1 = 80.4%
Maximum value of SAR (measured) = 5.48 W/kg

0 dB = 5.48 W/kg = 7.39 dBW/kg

Certificate No: CLA150-4025_Apr21 Page 6 of 6

Certificate No: CLA150-4025_Apr21 Page 6 of 6

1.2 D450V3 - SN 1103

Calibration Laboratory of Schmid & Partner Engineering AG
Ziegelhausstrasse 43, 8004 Zurich, Switzerland

Accredited by the Swiss Accreditation Service (SAS)
The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

Client: **SGS-CN (Audein)** Certificate No: **D450V3-1103_Apr21**

CALIBRATION CERTIFICATE

Object: **D450V3 - SN: 1103**

Calibration procedure(s): **QA CAL-15_v9**
Calibration Procedure for SAR Validation Sources below 700 MHz

Calibration date: **April 21, 2021**

This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility; environment temperature (22 ± 0.1) °C and humidity < 70%.

Calibration Equipment used (MTE critical for calibration)

Primary Standards	ID #	Cal Date (Certificate No.)	Scheduled Calibration
Power meter NRP	SN: 104778	09-Apr-21 (No. 217-03021/03030)	Apr-22
Power sensor NRP-291	SN: 102344	09-Apr-21 (No. 217-03021)	Apr-22
Reference 20 dB Attenuator	SN: C2582 (200)	09-Apr-21 (No. 217-03343)	Apr-22
Type-N immersion calibration	SN: 310982 / 06327	09-Apr-21 (No. 217-03344)	Apr-22
Reference Probe E3039A	SN: 3077	30-Dec-20 (No. E30-3077_De2020)	Dec-21
DAEA	SN: 654	05-Jan-20 (No. 0464-664_Jan20)	Jan-21

Secondary Standards	ID #	Check Date (in house)	Scheduled Check
Power meter E4418B	SN: GB41200274	06-Apr-16 (in house check Jun-20)	In house check Jun-22
Power sensor E4412A	SN: MY41496027	06-Apr-16 (in house check Jun-20)	In house check Jun-22
Power sensor E4413A	SN: 00010010	06-Apr-16 (in house check Jun-20)	In house check Jun-22
RF generator HP 8648C	SN: US3496101700	06-Apr-19 (in house check Jun-20)	In house check Jun-22
Network Analyzer Agilent E8358A	SN: US41980477	31-Mar-14 (in house check Oct-20)	In house check Oct-21

Calibrated by: **Christoph Leuber** (Function: Laboratory Technician)

Approved by: **Katja Polovic** (Function: Technical Manager)

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

Certificate No: D450V3-1103_Apr21 Page 1 of 6

Calibration Laboratory of Schmid & Partner Engineering AG
Ziegelhausstrasse 43, 8004 Zurich, Switzerland

Accredited by the Swiss Accreditation Service (SAS)
The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

Client: **SGS-CN (Audein)** Certificate No: **D450V3-1103_Apr21**

Glossary:

TSL: Issue simulating liquid sensitivity in TSL / NORM x,y,z

ConvF: not applicable or not measured

N/A: not applicable or not measured

Calibration is Performed According to the Following Standards:

- IEE Std 1528-2013, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices; Measurement Techniques", June 2013
- IEC 62209-1, "Measurement procedure for the assessment of Specific Absorption Rate (SAR) from hand-held and body-mounted devices used next to the ear (frequency range of 300 MHz to 6 GHz)", July 2016
- IEC 62209-2, "Procedure to determine the Specific Absorption Rate (SAR) for wireless communication devices used in close proximity to the human body (frequency range of 30 MHz to 6 GHz)", March 2010
- KDB 665664, "SAR Measurement Requirements for 100 MHz to 6 GHz"

Additional Documentation:

- DASY4/5 System Handbook

Methods Applied and Interpretation of Parameters:

- Measurement Conditions:** Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL:** The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.
- Feed Point Impedance and Return Loss:** These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.
- Electrical Delay:** One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
- SAR measured:** SAR measured at the stated antenna input power.
- SAR normalized:** SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters:** The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor $k=2$, which for a normal distribution corresponds to a coverage probability of approximately 95%.

Certificate No: D450V3-1103_Apr21 Page 2 of 6



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed on leaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

No.10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300
中国·江苏·昆山市留学生创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgsgroup.com.cn
t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com

Measurement Conditions
DASY system configuration, as far as not given on page 1.

DASY Version	DASY5	V82.10.4
Extrapolation	Advanced Extrapolation	
Phantom	ELJ Flat Phantom	Shell thickness: 2 ± 0.2 mm
Distance Dipole Center - TSL	15 mm	with Spacer
Zoom Scan Resolution	dx, dy, dz = 5 mm	
Frequency	450 MHz ± 1 MHz	

Head TSL parameters
The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	43.5	0.57 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	43.1 ± 0.6 %	0.57 mho/m ± 0 %
Head TSL temperature change during test	< 0.5 °C	---	---

SAR result with Head TSL

SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition	
SAR measured	250 mW input power	1.14 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	4.55 W/kg ± 18.1 % (k=2)

SAR averaged over 10 cm ³ (10 g) of Head TSL	Condition	
SAR measured	250 mW input power	0.757 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	3.06 W/kg ± 17.6 % (k=2)

Certificate No: D450V3-1103_Apr21 Page 3 of 6

Appendix (Additional assessments outside the scope of SCS 0106)

Antenna Parameters with Head TSL

Impedance, transformed to feed point	57.1 Ω - 2.6 jΩ
Return Loss	-23.0 dB

General Antenna Parameters and Design

Electrical Delay (one direction)	1.346 ns
----------------------------------	----------

After long term use with 100W radiated power, only a slight warming of the dipole near the feedpoint can be measured.
The dipole is made of standard straight coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is set according to the Standard.
No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feedpoint may be damaged.

Additional EUT Data

Manufactured by	SPEAG
-----------------	-------

Certificate No: D450V3-1103_Apr21 Page 4 of 6

DASY5 Validation Report for Head TSL

Test Laboratory: SPEAG, Zurich, Switzerland Date: 21.04.2021

DUT: Dipole 450 MHz; Type: D450V3; Serial: D450V3 - SN:1103

Communication System: UID 0 - CW; Frequency: 450 MHz
Medium parameters used: f = 450 MHz, α = 0.87 S/m; ε_r = 43.1; ρ = 1000 kg/m³
Phantom section: Flat Section
Measurement Standard: DASY5 (IEE/IEC/ANSI C63.19-2011)


DASY52 Configuration:

- Probe: EX3DV4 - SN3877; ConvF(10.64, 10.64, 10.64) @ 450 MHz; Calibrated: 30.12.2020
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn654; Calibrated: 26.06.2020
- Phantom: ELJ v4.0; Type: QDOVA001BB; Serial: TP:1003
- DASY52 52.10.4(1527); SEMCAD X 14.6.14(7483)

Dipole Calibration for Head Tissue/d=15mm, Pin=250mW/Zoom Scan (7x7x7)/Cube 0:
Measurement grid: dx=5mm, dy=5mm, dz=5mm
Reference Value = 39.18 V/m; Power Drift = -0.08 dB
Peak SAR (extrapolated) = 1.76 W/kg
SAR(1 g) = 1.14 W/kg; SAR(10 g) = 0.767 W/kg
Smallest distance from peaks to all points 3 dB below: Larger than measurement grid
Ratio of SAR at M2 to SAR at M1 = 64.9%
Maximum value of SAR (measured) = 1.53 W/kg

0 dB = 1.53 W/kg = 1.85 dBW/kg

Certificate No: D450V3-1103_Apr21 Page 5 of 6



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.
Attention: To check the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

No.10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300
中国·江苏·昆山市留学院创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgsgroup.com.cn
t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com

1.3 D750V3 - SN 1188

<div style="display: flex; justify-content: space-between;"> </div> <p> Add: No.52 HuaYuanbei Road, Haidian District, Beijing, 100191, China Tel: +86-10-62306633-2112 Fax: +86-10-62306633-2504 E-mail: cti@chinaetl.com http://www.chinaetl.cn </p> <p> Client: SGS-CN Certificate No: Z22-60103 </p> <h3>CALIBRATION CERTIFICATE</h3> <p> Object: D750V3 - SN: 1188 </p> <p> Calibration Procedure(s): FF-Z11-003-01 Calibration Procedures for dipole validation kits </p> <p> Calibration date: March 29, 2022 </p> <p> This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate. </p> <p> All calibrations have been conducted in the closed laboratory facility: environment temperature (22±3)°C and humidity<70%. </p> <p> Calibration Equipment used (M&TE critical for calibration) </p> <table border="1"> <thead> <tr> <th>Primary Standards</th> <th>ID #</th> <th>Cal Date (Calibrated by, Certificate No.)</th> <th>Scheduled Calibration</th> </tr> </thead> <tbody> <tr> <td>Power Meter NRP2</td> <td>106277</td> <td>24-Sep-21 (CTTL No.J21X08326)</td> <td>Sep-22</td> </tr> <tr> <td>Power sensor NRP88</td> <td>104291</td> <td>24-Sep-21 (CTTL No.J21X08326)</td> <td>Sep-22</td> </tr> <tr> <td>Reference Probe EX30V4</td> <td>SN 7307</td> <td>26-May-21(SPEAG.No.EX3-7307_May21)</td> <td>May-22</td> </tr> <tr> <td>D4E4</td> <td>SN 1556</td> <td>12-Jan-22(CTTL-SPEAG.No.Z22-60007)</td> <td>Jan-23</td> </tr> </tbody> </table> <table border="1"> <thead> <tr> <th>Secondary Standards</th> <th>ID #</th> <th>Cal Date (Calibrated by, Certificate No.)</th> <th>Scheduled Calibration</th> </tr> </thead> <tbody> <tr> <td>Signal Generator E4439C</td> <td>MY49671430</td> <td>13-Jan-22 (CTTL No.J22X00409)</td> <td>Jan-23</td> </tr> <tr> <td>Network Analyzer E5071C</td> <td>MY46110973</td> <td>14-Jan-22 (CTTL No.J22X00409)</td> <td>Jan-23</td> </tr> </tbody> </table> <p> Calibrated by: Zhao Jing SAR Test Engineer (Signature) </p> <p> Reviewed by: Lin Hao SAR Test Engineer (Signature) </p> <p> Approved by: Qi Dianyuan SAR Project Leader (Signature) </p> <p> Issued: April 3, 2022 </p> <p> This calibration certificate shall not be reproduced except in full without written approval of the laboratory. </p> <p> Certificate No: Z22-60103 Page 1 of 6 </p>	Primary Standards	ID #	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration	Power Meter NRP2	106277	24-Sep-21 (CTTL No.J21X08326)	Sep-22	Power sensor NRP88	104291	24-Sep-21 (CTTL No.J21X08326)	Sep-22	Reference Probe EX30V4	SN 7307	26-May-21(SPEAG.No.EX3-7307_May21)	May-22	D4E4	SN 1556	12-Jan-22(CTTL-SPEAG.No.Z22-60007)	Jan-23	Secondary Standards	ID #	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration	Signal Generator E4439C	MY49671430	13-Jan-22 (CTTL No.J22X00409)	Jan-23	Network Analyzer E5071C	MY46110973	14-Jan-22 (CTTL No.J22X00409)	Jan-23	<div style="display: flex; justify-content: space-between;"> </div> <p> Add: No.52 HuaYuanbei Road, Haidian District, Beijing, 100191, China Tel: +86-10-62306633-2079 Fax: +86-10-62306633-2504 E-mail: cti@chinaetl.com http://www.chinaetl.cn </p> <h3>Glossary:</h3> <p> TSL: tissue simulating liquid ConvF: sensitivity in TSL / NORMx.yz N/A: not applicable or not measured </p> <p> Calibration is Performed According to the Following Standards: </p> <ul style="list-style-type: none"> IEC/IEEE 62209-1528, "Measurement Procedure for The Assessment of Specific Absorption Rate of Human Exposure to Radio Frequency Fields from Hand-held and Body-mounted Wireless Communication Devices-Part 1528: Human Models, Instrumentation and Procedures (Frequency range of 4 MHz to 10 GHz)", October 2020 KDB 865684, "SAR Measurement Requirements for 100 MHz to 6 GHz" <p> Additional Documentation: c) DASY4/5 System Handbook </p> <p> Methods Applied and Interpretation of Parameters: </p> <ul style="list-style-type: none"> Measurement Conditions: Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated. Antenna Parameters with TSL: The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis. Feed Point Impedance and Return Loss: These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required. Electrical Delay: One-way delay between the SMA connector and the antenna feed point. No uncertainty required. SAR measured: SAR measured at the stated antenna input power. SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna connector. SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result. <div style="border: 1px solid black; padding: 5px; margin-top: 10px;"> <p>The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.</p> </div> <p> Certificate No: Z22-60103 Page 2 of 6 </p>																												
Primary Standards	ID #	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration																																																										
Power Meter NRP2	106277	24-Sep-21 (CTTL No.J21X08326)	Sep-22																																																										
Power sensor NRP88	104291	24-Sep-21 (CTTL No.J21X08326)	Sep-22																																																										
Reference Probe EX30V4	SN 7307	26-May-21(SPEAG.No.EX3-7307_May21)	May-22																																																										
D4E4	SN 1556	12-Jan-22(CTTL-SPEAG.No.Z22-60007)	Jan-23																																																										
Secondary Standards	ID #	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration																																																										
Signal Generator E4439C	MY49671430	13-Jan-22 (CTTL No.J22X00409)	Jan-23																																																										
Network Analyzer E5071C	MY46110973	14-Jan-22 (CTTL No.J22X00409)	Jan-23																																																										
<div style="display: flex; justify-content: space-between;"> </div> <p> Add: No.52 HuaYuanbei Road, Haidian District, Beijing, 100191, China Tel: +86-10-62306633-2079 Fax: +86-10-62306633-2504 E-mail: cti@chinaetl.com http://www.chinaetl.cn </p> <h3>Measurement Conditions</h3> <p>DASY system configuration, as far as not given on page 1.</p> <table border="1"> <thead> <tr> <th>DASY Version</th> <th>DASY52</th> <th>V52.10.4</th> </tr> </thead> <tbody> <tr> <td>Extrapolation</td> <td>Advanced Extrapolation</td> <td></td> </tr> <tr> <td>Phantom</td> <td>Triple Flat Phantom 5.1C</td> <td></td> </tr> <tr> <td>Distance Dipole Center - TSL</td> <td>15 mm</td> <td>with Spacer</td> </tr> <tr> <td>Zoom Scan Resolution</td> <td>dx, dy, dz = 5 mm</td> <td></td> </tr> <tr> <td>Frequency</td> <td>750 MHz ± 1 MHz</td> <td></td> </tr> </tbody> </table> <h3>Head TSL parameters</h3> <p>The following parameters and calculations were applied:</p> <table border="1"> <thead> <tr> <th></th> <th>Temperature</th> <th>Permittivity</th> <th>Conductivity</th> </tr> </thead> <tbody> <tr> <td>Nominal Head TSL parameters</td> <td>22.0 °C</td> <td>42.0</td> <td>0.90 mho/m</td> </tr> <tr> <td>Measured Head TSL parameters</td> <td>(22.0 ± 0.2) °C</td> <td>41.4 ± 0.6 %</td> <td>0.89 mho/m ± 0.6 %</td> </tr> <tr> <td>Head TSL temperature change during test</td> <td><1.0 °C</td> <td>---</td> <td>---</td> </tr> </tbody> </table> <h3>SAR result with Head TSL</h3> <table border="1"> <thead> <tr> <th>SAR averaged over 1 cm² (1 g) of Head TSL</th> <th>Condition</th> <th></th> </tr> </thead> <tbody> <tr> <td>SAR measured</td> <td>250 mW input power</td> <td>2.07 W/kg</td> </tr> <tr> <td>SAR for nominal Head TSL parameters</td> <td>normalized to 1W</td> <td>8.27 W/kg ± 18.8 % (k=2)</td> </tr> <tr> <th>SAR averaged over 10 cm² (10 g) of Head TSL</th> <th>Condition</th> <th></th> </tr> <tr> <td>SAR measured</td> <td>250 mW input power</td> <td>1.37 W/kg</td> </tr> <tr> <td>SAR for nominal Head TSL parameters</td> <td>normalized to 1W</td> <td>5.48 W/kg ± 18.7 % (k=2)</td> </tr> </tbody> </table> <p> Certificate No: Z22-60103 Page 3 of 6 </p>	DASY Version	DASY52	V52.10.4	Extrapolation	Advanced Extrapolation		Phantom	Triple Flat Phantom 5.1C		Distance Dipole Center - TSL	15 mm	with Spacer	Zoom Scan Resolution	dx, dy, dz = 5 mm		Frequency	750 MHz ± 1 MHz			Temperature	Permittivity	Conductivity	Nominal Head TSL parameters	22.0 °C	42.0	0.90 mho/m	Measured Head TSL parameters	(22.0 ± 0.2) °C	41.4 ± 0.6 %	0.89 mho/m ± 0.6 %	Head TSL temperature change during test	<1.0 °C	---	---	SAR averaged over 1 cm ² (1 g) of Head TSL	Condition		SAR measured	250 mW input power	2.07 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	8.27 W/kg ± 18.8 % (k=2)	SAR averaged over 10 cm ² (10 g) of Head TSL	Condition		SAR measured	250 mW input power	1.37 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	5.48 W/kg ± 18.7 % (k=2)	<div style="display: flex; justify-content: space-between;"> </div> <p> Add: No.52 HuaYuanbei Road, Haidian District, Beijing, 100191, China Tel: +86-10-62306633-2079 Fax: +86-10-62306633-2504 E-mail: cti@chinaetl.com http://www.chinaetl.cn </p> <h3>Appendix (Additional assessments outside the scope of CNAS L0570)</h3> <h4>Antenna Parameters with Head TSL</h4> <table border="1"> <thead> <tr> <th>Impedance, transformed to feed point</th> <th>53.60- 1.13jΩ</th> </tr> </thead> <tbody> <tr> <td>Return Loss</td> <td>-28.7dB</td> </tr> </tbody> </table> <h4>General Antenna Parameters and Design</h4> <table border="1"> <tbody> <tr> <td>Electrical Delay (one direction)</td> <td>0.947 ns</td> </tr> </tbody> </table> <p> After long term use with 100W radiated power, only a slight warming of the dipole near the feed-point can be measured. </p> <p> The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard. No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feed-point may be damaged. </p> <h4>Additional EUT Data</h4> <table border="1"> <tbody> <tr> <td>Manufactured by</td> <td>SPEAG</td> </tr> </tbody> </table> <p> Certificate No: Z22-60103 Page 4 of 6 </p>	Impedance, transformed to feed point	53.60- 1.13jΩ	Return Loss	-28.7dB	Electrical Delay (one direction)	0.947 ns	Manufactured by	SPEAG
DASY Version	DASY52	V52.10.4																																																											
Extrapolation	Advanced Extrapolation																																																												
Phantom	Triple Flat Phantom 5.1C																																																												
Distance Dipole Center - TSL	15 mm	with Spacer																																																											
Zoom Scan Resolution	dx, dy, dz = 5 mm																																																												
Frequency	750 MHz ± 1 MHz																																																												
	Temperature	Permittivity	Conductivity																																																										
Nominal Head TSL parameters	22.0 °C	42.0	0.90 mho/m																																																										
Measured Head TSL parameters	(22.0 ± 0.2) °C	41.4 ± 0.6 %	0.89 mho/m ± 0.6 %																																																										
Head TSL temperature change during test	<1.0 °C	---	---																																																										
SAR averaged over 1 cm ² (1 g) of Head TSL	Condition																																																												
SAR measured	250 mW input power	2.07 W/kg																																																											
SAR for nominal Head TSL parameters	normalized to 1W	8.27 W/kg ± 18.8 % (k=2)																																																											
SAR averaged over 10 cm ² (10 g) of Head TSL	Condition																																																												
SAR measured	250 mW input power	1.37 W/kg																																																											
SAR for nominal Head TSL parameters	normalized to 1W	5.48 W/kg ± 18.7 % (k=2)																																																											
Impedance, transformed to feed point	53.60- 1.13jΩ																																																												
Return Loss	-28.7dB																																																												
Electrical Delay (one direction)	0.947 ns																																																												
Manufactured by	SPEAG																																																												



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

No.10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300 t(86-512)57355888 f(86-512)57370818 www.sgs.com.cn
 中国·江苏·昆山市留学生创业园伟业路10号 邮编 215300 t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com


TTL Speaq Calibration Laboratory
 In Collaboration with **CAICT**

Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
 Tel: +86-10-62506633-2079 Fax: +86-10-62506633-2504
 E-mail: cti@china.ttl.com http://www.china.ttl.com

DASY5 Validation Report for Head TSL Date: 2022-03-29
 Test Laboratory: CCTL, Beijing, China
 DUT: Dipole 750 MHz; Type: D750V3; Serial: D750V3 - SN: 1188
 Communication System: UID 0, CW; Frequency: 750 MHz; Duty Cycle: 1:1
 Medium parameters used: $f = 750 \text{ MHz}$; $\sigma = 0.888 \text{ S/m}$; $\epsilon_r = 41.36$; $\rho = 1000 \text{ kg/m}^3$
 Phantom section: Right Section
 Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007)
 DASY5 Configuration:

- Probe: EX3DV4 - SN7307; ConvF(10.31, 10.31) @ 750 MHz; Calibrated: 2021-05-26
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn1556; Calibrated: 2022-01-12
- Phantom: MFP-V5.1C (20kg probe tilt); Type: QD 000 P51 Cx; Serial: 1062
- DASY52.52.10.4(1535); SEMCADX 14.6.14(7501)

Dipole Calibration/Zoom Scan (7x7x7) (7x7x7) Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm
 Reference Value = 55.06 V/m; Power Drift = 0.00 dB
 Peak SAR (extrapolated) = 2.07 W/kg
 SAR(1 g) = 2.07 W/kg; SAR(10 g) = 1.37 W/kg
 Smallest distance from peaks to all points 3 dB below = 18.9 mm
 Ratio of SAR at M2 to SAR at M1 = 67.1%
 Maximum value of SAR (measured) = 2.74 W/kg




0 dB = 2.74 W/kg = 4.38 dBW/kg

Certificate No: Z22-60103 Page 5 of 6

TTL Speaq Calibration Laboratory
 In Collaboration with **CAICT**

Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
 Tel: +86-10-62506633-2079 Fax: +86-10-62506633-2504
 E-mail: cti@china.ttl.com http://www.china.ttl.com

Impedance Measurement Plot for Head TSL



Certificate No: Z22-60103 Page 6 of 6

1.4 D835V2 - SN 4d114

TTL Speaq Calibration Laboratory
 In Collaboration with **CAICT** and **CNAS**

Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
 Tel: +86-10-62506633-2079 Fax: +86-10-62506633-2504
 E-mail: cti@china.ttl.com http://www.china.ttl.com

Client: **SGS-CN** Certificate No: **Z22-60104**

CALIBRATION CERTIFICATE

Object: D835V2 - SN: 4d114

Calibration Procedure(s): FF-Z11-003-01
 Calibration Procedures for dipole validation kits

Calibration date: March 31, 2022

This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature (22±3)°C and humidity <70%.

Calibration Equipment used (M&TE critical for calibration)

Primary Standards	ID #	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration
Power Meter NRP2	106277	24-Sep-21 (CCTL, No.J21X08326)	Sep-22
Power sensor NRPBS	104291	24-Sep-21 (CCTL, No.J21X08326)	Sep-22
Reference Probe EX3DV4	SN 7307	26-May-21(SPEAG.No.EX3-7307_May21)	May-22
DAE4	SN 1556	12-Jan-22(CCTL-SPEAG.No.Z22-60007)	Jan-23

Secondary Standards	ID #	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration
Signal Generator E4438C	MY49071430	13-Jan-22 (CCTL, No.J22X00409)	Jan-23
Network Analyzer E5071C	MY49110673	14-Jan-22 (CCTL, No.J22X00406)	Jan-23

Calibrated by: Zhao Jing, SAR Test Engineer

Reviewed by: Lin Hao, SAR Test Engineer

Approved by: Qi Dianyuan, SAR Project Leader

Signature: [Signatures]

Issued: April 6, 2022

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

Certificate No: Z22-60104 Page 1 of 6

TTL Speaq Calibration Laboratory
 In Collaboration with **CAICT**

Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
 Tel: +86-10-62506633-2079 Fax: +86-10-62506633-2504
 E-mail: cti@china.ttl.com http://www.china.ttl.com

Glossary:

- TSL: tissue simulating liquid
- ConvF: sensitivity in TSL / NORMx.yz
- N/A: not applicable or not measured

Calibration is Performed According to the Following Standards:

- IEC/IEEE 62209-1528, "Measurement Procedure for The Assessment of Specific Absorption Rate of Human Exposure to Radio Frequency Fields from Hand-held and Body-mounted Wireless Communication Devices- Part 1528: Human Models, Instrumentation and Procedures (Frequency range of 4 MHz to 10 GHz)", October 2020
- KDB 685864, "SAR Measurement Requirements for 100 MHz to 6 GHz"

Additional Documentation:

- DASY4/5 System Handbook

Methods Applied and Interpretation of Parameters:

- Measurement Conditions:** Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL:** The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.
- Feed Point Impedance and Return Loss:** These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.
- Electrical Delay:** One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
- SAR measured:** SAR measured at the stated antenna input power.
- SAR normalized:** SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters:** The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor k=2, which for a normal distribution Corresponds to a coverage probability of approximately 95%.

Certificate No: Z22-60104 Page 2 of 6



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

No. 10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300
 中国·江苏·昆山市留学生创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgsgroup.com.cn
 t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com

In Collaboration with **TTL Speag** CALIBRATION LABORATORY and **CAICT**

Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
 Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504
 E-mail: cti@china.ttl.com http://www.china.ttl.com

Measurement Conditions
 DASY system configuration, as far as not given on page 1

DASY Version	DASY52	VS2 10.4
Extrapolation	Advanced Extrapolation	
Phantom	Triple Flat Phantom 5.1C	
Distance Dipole Center - TSL	15 mm	with Spacer
Zoom Scan Resolution	dx, dy, dz = 5 mm	
Frequency	835 MHz ± 1 MHz	

Head TSL parameters
 The following parameters and calculations were applied:

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	41.5	0.90 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	41.0 ± 8 %	0.91 mho/m ± 8 %
Head TSL temperature change during test	<1.0 °C	---	---

SAR result with Head TSL

SAR averaged over 1 cm ³ (1g) of Head TSL	Condition	
SAR measured	250 mW input power	2.37 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	6.40 W/kg ± 18.6 % (k=2)
SAR averaged over 10 cm ³ (10g) of Head TSL	Condition	
SAR measured	250 mW input power	1.54 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	6.12 W/kg ± 18.7 % (k=2)

Certificate No: Z22-60104 Page 3 of 6

In Collaboration with **TTL Speag** CALIBRATION LABORATORY and **CAICT**

Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
 Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504
 E-mail: cti@china.ttl.com http://www.china.ttl.com

Appendix (Additional assessments outside the scope of CNAS L0570)

Antenna Parameters with Head TSL

Impedance, transformed to feed point	48.7Ω - j22Ω
Return Loss	-25.3dB

General Antenna Parameters and Design

Electrical Delay (one direction)	1.307 ns
----------------------------------	----------

After long term use with 100W radiated power, only a slight warming of the dipole near the feed-point can be measured.

The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard. No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feed-point may be damaged.

Additional EUT Data

Manufactured by	SPEAG
-----------------	-------

Certificate No: Z22-60104 Page 4 of 6

In Collaboration with **TTL Speag** CALIBRATION LABORATORY and **CAICT**

Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
 Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504
 E-mail: cti@china.ttl.com http://www.china.ttl.com

DASY5 Validation Report for Head TSL
 Test Laboratory: CTTL, Beijing, China Date: 2022-03-31
 DUT: Dipole 835 MHz; Type: D835V2; Serial: D835V2 - SN: 4d114
 Communication System: UID 0, CW; Frequency: 835 MHz; Duty Cycle: 1:1
 Medium parameters used: f = 835 MHz; σ = 0.907 S/m; ε_r = 40.98; ρ = 1000 kg/m³
 Phantom section: Right Section
 Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007)
 DASY5 Configuration:

- Probe: EX3DV4 - SN7307; ConvF(10.13, 10.13, 10.13) @ 835 MHz; Calibrated: 2021-05-26
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DA-E4 Sx1556; Calibrated: 2022-01-12
- Phantom: MFP_V5.1C (20deg probe tilt); Type: QD 000 P51 Cx; Serial: 1062
- DASY52 52.10.4(1535); SEMCAD X 14.6.14(7501)

Dipole Calibration/Zoom Scan (7x7x7) (7x7x7) Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm
 Reference Value = 57.88 V/m; Power Drift = 0.04 dB
 Peak SAR (extrapolated) = 3.56 W/kg
 SAR(1g) = 2.37 W/kg; SAR(10g) = 1.54 W/kg
 Smallest distance from peaks to all points 3 dB below = 15.8 mm
 Ratio of SAR at M2 to SAR at M1 = 66.2%
 Maximum value of SAR (measured) = 3.17 W/kg

0 dB = 3.17 W/kg = 5.01 dBW/kg

Certificate No: Z22-60104 Page 5 of 6

In Collaboration with **TTL Speag** CALIBRATION LABORATORY and **CAICT**

Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
 Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504
 E-mail: cti@china.ttl.com http://www.china.ttl.com

Impedance Measurement Plot for Head TSL

Certificate No: Z22-60104 Page 6 of 6



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx>; and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.
Attention: To check the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

No.10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300
 中国·江苏·昆山市留学院创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgs.com.cn
 t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com

1.5 D900V2 - SN 1d079

<div style="display: flex; justify-content: space-between; align-items: center;"> <div style="text-align: center;"> <p>Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-4239633-2117 E-mail: cti@china.ttl.com</p> </div> <div style="text-align: center;"> <p>Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-4239633-2117 E-mail: cti@china.ttl.com</p> </div> <div style="text-align: center;"> <p>Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-4239633-2117 E-mail: cti@china.ttl.com</p> </div> </div> <p style="text-align: center;">Client: SGS-CN Certificate No: Z22-60184</p> <p style="text-align: center;">CALIBRATION CERTIFICATE</p> <p>Object: D900V2 - SN: 1d079</p> <p>Calibration Procedure(s): FF-Z11-003-01 Calibration Procedures for dipole validation kits</p> <p>Calibration date: June 7, 2022</p> <p>The calibration Certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.</p> <p>All calibrations have been conducted in the closed laboratory facility: environment temperature (23±3)°C and humidity <70%.</p> <p>Calibration Equipment used (M&TE critical for calibration)</p> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Primary Standards</th> <th>ID #</th> <th>Cal Date (Calibrated by, Certificate No.)</th> <th>Scheduled Calibration</th> </tr> </thead> <tbody> <tr> <td>Power Meter NRP2</td> <td>106277</td> <td>24-Sep-21 (CTTL No. J21X08326)</td> <td>Sep-22</td> </tr> <tr> <td>Power sensor NRP8S</td> <td>104291</td> <td>24-Sep-21 (CTTL No. J21X08326)</td> <td>Sep-22</td> </tr> <tr> <td>Reference Probe EX3DV4</td> <td>SN 7464</td> <td>26-Jan-22 (SPEAG No. EX3-7464_Jan22)</td> <td>Jan-23</td> </tr> <tr> <td>DAE4</td> <td>SN 1956</td> <td>12-Jan-22 (CTTL-SPEAG No. Z22-60007)</td> <td>Jan-23</td> </tr> </tbody> </table> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Secondary Standards</th> <th>ID #</th> <th>Cal Date (Calibrated by, Certificate No.)</th> <th>Scheduled Calibration</th> </tr> </thead> <tbody> <tr> <td>Signal Generator E4438C</td> <td>MV42071430</td> <td>13-Jan-22 (CTTL No. J22X00409)</td> <td>Jan-23</td> </tr> <tr> <td>Network Analyzer E5071C</td> <td>MV48110673</td> <td>14-Jan-22 (CTTL No. J22X00409)</td> <td>Jan-23</td> </tr> </tbody> </table> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Calibrated by:</th> <th>Name</th> <th>Function</th> <th>Signature</th> </tr> </thead> <tbody> <tr> <td>Reviewed by:</td> <td>Lin Hao</td> <td>SAR Test Engineer</td> <td></td> </tr> <tr> <td>Approved by:</td> <td>Qi Diqiyuan</td> <td>SAR Project Leader</td> <td></td> </tr> </tbody> </table> <p style="text-align: center;">Issued: June 13, 2022</p> <p style="text-align: center;">The calibration certificate shall not be reproduced except in full without written approval of the laboratory.</p> <p style="text-align: center;">Certificate No: Z22-60184 Page 1 of 6</p>	Primary Standards	ID #	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration	Power Meter NRP2	106277	24-Sep-21 (CTTL No. J21X08326)	Sep-22	Power sensor NRP8S	104291	24-Sep-21 (CTTL No. J21X08326)	Sep-22	Reference Probe EX3DV4	SN 7464	26-Jan-22 (SPEAG No. EX3-7464_Jan22)	Jan-23	DAE4	SN 1956	12-Jan-22 (CTTL-SPEAG No. Z22-60007)	Jan-23	Secondary Standards	ID #	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration	Signal Generator E4438C	MV42071430	13-Jan-22 (CTTL No. J22X00409)	Jan-23	Network Analyzer E5071C	MV48110673	14-Jan-22 (CTTL No. J22X00409)	Jan-23	Calibrated by:	Name	Function	Signature	Reviewed by:	Lin Hao	SAR Test Engineer		Approved by:	Qi Diqiyuan	SAR Project Leader		<div style="display: flex; justify-content: space-between; align-items: center;"> <div style="text-align: center;"> <p>Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-4239633-2117 E-mail: cti@china.ttl.com</p> </div> <div style="text-align: center;"> <p>Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-4239633-2117 E-mail: cti@china.ttl.com</p> </div> </div> <p style="text-align: center;">Glossary:</p> <p>TSL: tissue simulating liquid ConvF: sensitivity in TSL / NORM_{x,y,z} N/A: not applicable or not measured</p> <p>Calibration is Performed According to the Following Standards:</p> <p>a) IEC/IEEE 62209-1528, "Measurement Procedure for The Assessment of Specific Absorption Rate of Human Exposure to Radio Frequency Fields from Hand-held and Body-mounted Wireless Communication Devices- Part 1528: Human Models, Instrumentation and Procedures (Frequency range of 4 MHz to 10 GHz)", October 2020 b) KDB 865664, "SAR Measurement Requirements for 100 MHz to 6 GHz" c) DASY4/S System Handbook</p> <p>Additional Documentation:</p> <p>Methods Applied and Interpretation of Parameters:</p> <ul style="list-style-type: none"> Measurement Conditions: Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated. Antenna Parameters with TSL: The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis. Feed Point Impedance and Return Loss: These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required. Electrical Delay: One-way delay between the SMA connector and the antenna feed point. No uncertainty required. SAR measured: SAR measured at the stated antenna input power. SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna connector. SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result. <div style="border: 1px solid black; padding: 5px; margin-top: 10px;"> <p>The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.</p> </div> <p style="text-align: center;">Certificate No: Z22-60184 Page 2 of 6</p>																		
Primary Standards	ID #	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration																																																												
Power Meter NRP2	106277	24-Sep-21 (CTTL No. J21X08326)	Sep-22																																																												
Power sensor NRP8S	104291	24-Sep-21 (CTTL No. J21X08326)	Sep-22																																																												
Reference Probe EX3DV4	SN 7464	26-Jan-22 (SPEAG No. EX3-7464_Jan22)	Jan-23																																																												
DAE4	SN 1956	12-Jan-22 (CTTL-SPEAG No. Z22-60007)	Jan-23																																																												
Secondary Standards	ID #	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration																																																												
Signal Generator E4438C	MV42071430	13-Jan-22 (CTTL No. J22X00409)	Jan-23																																																												
Network Analyzer E5071C	MV48110673	14-Jan-22 (CTTL No. J22X00409)	Jan-23																																																												
Calibrated by:	Name	Function	Signature																																																												
Reviewed by:	Lin Hao	SAR Test Engineer																																																													
Approved by:	Qi Diqiyuan	SAR Project Leader																																																													
<div style="display: flex; justify-content: space-between; align-items: center;"> <div style="text-align: center;"> <p>Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-4239633-2117 E-mail: cti@china.ttl.com</p> </div> <div style="text-align: center;"> <p>Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-4239633-2117 E-mail: cti@china.ttl.com</p> </div> </div> <p>Measurement Conditions</p> <p>DASY system configuration, as far as not given on page 1.</p> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>DASY Version</th> <th>Configuration</th> <th>Value</th> </tr> </thead> <tbody> <tr> <td>Extrapolation</td> <td>Advanced Extrapolation</td> <td>52.10.4</td> </tr> <tr> <td>Phantom</td> <td>Triple Flat Phantom 5.1C</td> <td></td> </tr> <tr> <td>Distance Dipole Center - TSL</td> <td>15 mm</td> <td>with Spacer</td> </tr> <tr> <td>Zoom Scan Resolution</td> <td>dk, dy, dz = 5 mm</td> <td></td> </tr> <tr> <td>Frequency</td> <td>900 MHz ± 1 MHz</td> <td></td> </tr> </tbody> </table> <p>Head TSL parameters</p> <p>The following parameters and calculations were applied.</p> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Nominal Head TSL parameters</th> <th>Temperature</th> <th>Permittivity</th> <th>Conductivity</th> </tr> </thead> <tbody> <tr> <td>Measured Head TSL parameters</td> <td>(22.0 ± 0.2) °C</td> <td>41.5</td> <td>0.07 mho/m</td> </tr> <tr> <td>Head TSL temperature change during test</td> <td><1.0 °C</td> <td>---</td> <td>---</td> </tr> </tbody> </table> <p>SAR result with Head TSL</p> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>SAR averaged over 1 cm³ (1 g) of Head TSL</th> <th>Condition</th> <th>Value</th> </tr> </thead> <tbody> <tr> <td>SAR measured</td> <td>250 mW input power</td> <td>2.70 W/kg</td> </tr> <tr> <td>SAR for nominal Head TSL parameters</td> <td>normalized to 1W</td> <td>11.0 W/kg ± 18.8 % (k=2)</td> </tr> <tr> <td>SAR averaged over 10 cm³ (10 g) of Head TSL</td> <td>Condition</td> <td>Value</td> </tr> <tr> <td>SAR measured</td> <td>250 mW input power</td> <td>1.73 W/kg</td> </tr> <tr> <td>SAR for nominal Head TSL parameters</td> <td>normalized to 1W</td> <td>7.09 W/kg ± 18.7 % (k=2)</td> </tr> </tbody> </table> <p style="text-align: center;">Certificate No: Z22-60184 Page 3 of 6</p>	DASY Version	Configuration	Value	Extrapolation	Advanced Extrapolation	52.10.4	Phantom	Triple Flat Phantom 5.1C		Distance Dipole Center - TSL	15 mm	with Spacer	Zoom Scan Resolution	dk, dy, dz = 5 mm		Frequency	900 MHz ± 1 MHz		Nominal Head TSL parameters	Temperature	Permittivity	Conductivity	Measured Head TSL parameters	(22.0 ± 0.2) °C	41.5	0.07 mho/m	Head TSL temperature change during test	<1.0 °C	---	---	SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition	Value	SAR measured	250 mW input power	2.70 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	11.0 W/kg ± 18.8 % (k=2)	SAR averaged over 10 cm ³ (10 g) of Head TSL	Condition	Value	SAR measured	250 mW input power	1.73 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	7.09 W/kg ± 18.7 % (k=2)	<div style="display: flex; justify-content: space-between; align-items: center;"> <div style="text-align: center;"> <p>Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-4239633-2117 E-mail: cti@china.ttl.com</p> </div> <div style="text-align: center;"> <p>Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-4239633-2117 E-mail: cti@china.ttl.com</p> </div> </div> <p>Appendix (Additional assessments outside the scope of CNAS L0570)</p> <p>Antenna Parameters with Head TSL</p> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Parameter</th> <th>Value</th> </tr> </thead> <tbody> <tr> <td>Impedance, transformed to feed point</td> <td>48.10 - 8.48jΩ</td> </tr> <tr> <td>Return Loss</td> <td>-23.3 dB</td> </tr> </tbody> </table> <p>General Antenna Parameters and Design</p> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Parameter</th> <th>Value</th> </tr> </thead> <tbody> <tr> <td>Electrical Delay (one direction)</td> <td>1.312 ns</td> </tr> </tbody> </table> <p>After long term use with 100W radiated power, only a slight warming of the dipole near the feed-point can be measured.</p> <p>The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard. No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feed-point may be damaged.</p> <p>Additional EUT Data</p> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Manufacturer by</th> <th>Value</th> </tr> </thead> <tbody> <tr> <td>Manufactured by</td> <td>SPEAG</td> </tr> </tbody> </table> <p style="text-align: center;">Certificate No: Z22-60184 Page 4 of 6</p>	Parameter	Value	Impedance, transformed to feed point	48.10 - 8.48jΩ	Return Loss	-23.3 dB	Parameter	Value	Electrical Delay (one direction)	1.312 ns	Manufacturer by	Value	Manufactured by	SPEAG
DASY Version	Configuration	Value																																																													
Extrapolation	Advanced Extrapolation	52.10.4																																																													
Phantom	Triple Flat Phantom 5.1C																																																														
Distance Dipole Center - TSL	15 mm	with Spacer																																																													
Zoom Scan Resolution	dk, dy, dz = 5 mm																																																														
Frequency	900 MHz ± 1 MHz																																																														
Nominal Head TSL parameters	Temperature	Permittivity	Conductivity																																																												
Measured Head TSL parameters	(22.0 ± 0.2) °C	41.5	0.07 mho/m																																																												
Head TSL temperature change during test	<1.0 °C	---	---																																																												
SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition	Value																																																													
SAR measured	250 mW input power	2.70 W/kg																																																													
SAR for nominal Head TSL parameters	normalized to 1W	11.0 W/kg ± 18.8 % (k=2)																																																													
SAR averaged over 10 cm ³ (10 g) of Head TSL	Condition	Value																																																													
SAR measured	250 mW input power	1.73 W/kg																																																													
SAR for nominal Head TSL parameters	normalized to 1W	7.09 W/kg ± 18.7 % (k=2)																																																													
Parameter	Value																																																														
Impedance, transformed to feed point	48.10 - 8.48jΩ																																																														
Return Loss	-23.3 dB																																																														
Parameter	Value																																																														
Electrical Delay (one direction)	1.312 ns																																																														
Manufacturer by	Value																																																														
Manufactured by	SPEAG																																																														



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

No.10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300
中国·江苏·昆山市留学生创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgsgroup.com.cn
t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com

In Collaboration with
CAICT

Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
Tel: +86-10-4230603-2117
E-mail: cti@china.ttl.com

中国认证
国家互认
校准
校准
CAICT
CNAS 15078

Date: 2022-06-07

DASY5 Validation Report for Head TSL

Test Laboratory: CTTL, Beijing, China

DUT: Dipole 900 MHz; Type: D900V2; Serial: D900V2 - SN: 14079

Communication System: UTD 0, CW; Frequency: 900 MHz; Duty Cycle: 1:1

Medium parameters used: $f = 900 \text{ MHz}$; $\sigma = 0.98 \text{ S/m}$; $\epsilon = 42.05$; $\rho = 1000 \text{ kg/m}^3$

Phantom section: Right Section

Measurement Standard: DASY5 (IEE/IEC/ANSI C63.19-2007)

DASY5 Configuration:

- Probe: EX3DV4 - SN7464; ConvF(9.72, 9.72) @ 900 MHz; Calibrated: 2022-01-26
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronic: DA44 SN1556; Calibrated: 2022-01-12
- Phantom: MFP_V5.1C (2ddeg probe tilt); Type: QD 000 P51 Cx; Serial: 1062
- DASY52.10.4(1535); SEMCAD X 14.6.14(7501)

Dipole Calibration/Zoom Scan (7x7) (7x7) Cube 0; Measurement grid: $d_x=5\text{mm}$, $d_y=5\text{mm}$, $d_z=5\text{mm}$

Reference Value = 59.81 V/m; Power Drift = -0.01 dB

Peak SAR (extrapolated) = 4.20 W/kg

SAR(1g) = 2.78 W/kg; SAR(10g) = 1.78 W/kg

Smallest distance from peaks to all points 3 dB below = 16 mm

Ratio of SAR at M2 to SAR at M1 = 65.8%

Maximum value of SAR (measured) = 3.71 W/kg

0 dB = 3.71 W/kg = 5.69 dBW/kg

Certificate No: Z22-60184 Page 3 of 6

In Collaboration with
CAICT

Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
Tel: +86-10-4230603-2117
E-mail: cti@china.ttl.com

中国认证
国家互认
校准
校准
CAICT
CNAS 15078

Date: 2022-06-07

Impedance Measurement Plot for Head TSL

Certificate No: Z22-60184 Page 6 of 6

1.6 D1800V2 - SN 2d170

In Collaboration with
CAICT

Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
Tel: +86-10-4230603-2117
E-mail: cti@china.ttl.com

中国认证
国家互认
校准
校准
CAICT
CNAS 15078

Date: 2022-06-07

Client: SGS-CN Certificate No: Z22-60105

CALIBRATION CERTIFICATE

Object: D1800V2 - SN: 2d170

Calibration Procedure(s): FF-Z11-003-01
Calibration Procedures for dipole validation kits

Calibration date: March 31, 2022

This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility, environment temperature (22±3)°C and humidity <70%.

Calibration Equipment used (M&TE critical for calibration)

Primary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration
Power Meter NRP2	106277	24-Sep-21 (CTTL No.J21X08326)	Sep-22
Power sensor NRP85	104291	24-Sep-21 (CTTL No.J21X08326)	Sep-22
Reference Probe EX3DV4	SN 7307	26-May-21 (SPEAG No.EX3-7307_May21)	May-22
DAE4	SN 1556	12-Jan-22 (CTTL-SPEAG No.Z22-60007)	Jan-23

Secondary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration
Signal Generator E4438C	MY49071430	13-Jan-22 (CTTL No.J22X00406)	Jan-23
Network Analyzer E5071C	MY46110973	14-Jan-22 (CTTL No.J22X00406)	Jan-23

Calibrated by: Zhao Jing, SAR Test Engineer, Signature: [Signature]

Reviewed by: Lin Hao, SAR Test Engineer, Signature: [Signature]

Approved by: Qi Diqiyuan, SAR Project Leader, Signature: [Signature]

Issued: April 6, 2022

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

Certificate No: Z22-60105 Page 1 of 6

In Collaboration with
CAICT

Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
Tel: +86-10-4230603-2079
E-mail: cti@china.ttl.com

中国认证
国家互认
校准
校准
CAICT
CNAS 15078

Date: 2022-06-07

Glossary:

TSL: tissue simulating liquid

ConvF: sensitivity in TSL / NORMx,y,z

N/A: not applicable or not measured

Calibration is Performed According to the Following Standards:

- IEC/IEEE 62209-1528, "Measurement Procedure for The Assessment of Specific Absorption Rate of human Exposure to Radio Frequency Fields from Hand-held and Body-mounted Wireless Communication Devices- Part 1528: Human Models, Instrumentation and Procedures (Frequency range of 4 MHz to 10 GHz)", October 2020
- KDB 855664, "SAR Measurement Requirements for 100 MHz to 6 GHz"

Additional Documentation:

- DASY4/5 System Handbook

Methods Applied and Interpretation of Parameters:

- Measurement Conditions:** Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL:** The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.
- Feed Point Impedance and Return Loss:** These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.
- Electrical Delay:** One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
- SAR measured:** SAR measured at the stated antenna input power.
- SAR normalized:** SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters:** The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor k=2, which for a normal distribution Corresponds to a coverage probability of approximately 95%.

Certificate No: Z22-60105 Page 2 of 6



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/terms-and-conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/terms-and-conditions/terms-e-document.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

No.10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300
中国·江苏·昆山市留学生创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgsgroup.com.cn
t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com

<p>TTL Speaq In Collaboration with CALIBRATION LABORATORY</p> <p>CAICT</p> <p>Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@china.ttl.com http://www.chinat.tl.com</p> <p>Measurement Conditions DASY system configuration, as far as not given on page 1.</p> <table border="1"> <tr><td>DASY Version</td><td>DASY52</td><td>52.10.4</td></tr> <tr><td>Extrapolation</td><td>Advanced Extrapolation</td><td></td></tr> <tr><td>Phantom</td><td>Triple Flat Phantom 5.1C</td><td></td></tr> <tr><td>Distance Dipole Center - TSL</td><td>10 mm</td><td>with Spacer</td></tr> <tr><td>Zoom Scan Resolution</td><td>dx, dy, dz = 5 mm</td><td></td></tr> <tr><td>Frequency</td><td>1800 MHz ± 1 MHz</td><td></td></tr> </table> <p>Head TSL parameters The following parameters and calculations were applied.</p> <table border="1"> <tr><th></th><th>Temperature</th><th>Permittivity</th><th>Conductivity</th></tr> <tr><td>Nominal Head TSL parameters</td><td>22.0 °C</td><td>40.0</td><td>1.40 mho/m</td></tr> <tr><td>Measured Head TSL parameters</td><td>(22.0 ± 0.2) °C</td><td>40.8 ± 8 %</td><td>1.41 mho/m ± 8 %</td></tr> <tr><td>Head TSL temperature change during test</td><td><1.0 °C</td><td>---</td><td>---</td></tr> </table> <p>SAR result with Head TSL</p> <table border="1"> <tr><td>SAR averaged over 1 cm³ (1 g) of Head TSL</td><td>Condition</td><td></td></tr> <tr><td>SAR measured</td><td>250 mW input power</td><td>9.73 W/kg</td></tr> <tr><td>SAR for nominal Head TSL parameters</td><td>normalized to 1W</td><td>38.9 W/kg ± 18.8 % (k=2)</td></tr> <tr><td>SAR averaged over 10 cm³ (10 g) of Head TSL</td><td>Condition</td><td></td></tr> <tr><td>SAR measured</td><td>250 mW input power</td><td>5.11 W/kg</td></tr> <tr><td>SAR for nominal Head TSL parameters</td><td>normalized to 1W</td><td>20.4 W/kg ± 18.7 % (k=2)</td></tr> </table> <p>Certificate No: Z22-60105 Page 3 of 6</p>	DASY Version	DASY52	52.10.4	Extrapolation	Advanced Extrapolation		Phantom	Triple Flat Phantom 5.1C		Distance Dipole Center - TSL	10 mm	with Spacer	Zoom Scan Resolution	dx, dy, dz = 5 mm		Frequency	1800 MHz ± 1 MHz			Temperature	Permittivity	Conductivity	Nominal Head TSL parameters	22.0 °C	40.0	1.40 mho/m	Measured Head TSL parameters	(22.0 ± 0.2) °C	40.8 ± 8 %	1.41 mho/m ± 8 %	Head TSL temperature change during test	<1.0 °C	---	---	SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition		SAR measured	250 mW input power	9.73 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	38.9 W/kg ± 18.8 % (k=2)	SAR averaged over 10 cm ³ (10 g) of Head TSL	Condition		SAR measured	250 mW input power	5.11 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	20.4 W/kg ± 18.7 % (k=2)	<p>TTL Speaq In Collaboration with CALIBRATION LABORATORY</p> <p>CAICT</p> <p>Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@china.ttl.com http://www.chinat.tl.com</p> <p>Appendix (Additional assessments outside the scope of CNAS L0570)</p> <p>Antenna Parameters with Head TSL</p> <table border="1"> <tr><td>Impedance, transformed to feed point</td><td>47.90-2.54jΩ</td></tr> <tr><td>Return Loss</td><td>-29.4dB</td></tr> </table> <p>General Antenna Parameters and Design</p> <table border="1"> <tr><td>Electrical Delay (one direction)</td><td>1.116 ns</td></tr> </table> <p>After long term use with 100W radiated power, only a slight warming of the dipole near the feed-point can be measured.</p> <p>The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard. No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feed-point may be damaged.</p> <p>Additional EUT Data</p> <table border="1"> <tr><td>Manufactured by</td><td>SPEAG</td></tr> </table> <p>Certificate No: Z22-60105 Page 4 of 6</p>	Impedance, transformed to feed point	47.90-2.54jΩ	Return Loss	-29.4dB	Electrical Delay (one direction)	1.116 ns	Manufactured by	SPEAG
DASY Version	DASY52	52.10.4																																																											
Extrapolation	Advanced Extrapolation																																																												
Phantom	Triple Flat Phantom 5.1C																																																												
Distance Dipole Center - TSL	10 mm	with Spacer																																																											
Zoom Scan Resolution	dx, dy, dz = 5 mm																																																												
Frequency	1800 MHz ± 1 MHz																																																												
	Temperature	Permittivity	Conductivity																																																										
Nominal Head TSL parameters	22.0 °C	40.0	1.40 mho/m																																																										
Measured Head TSL parameters	(22.0 ± 0.2) °C	40.8 ± 8 %	1.41 mho/m ± 8 %																																																										
Head TSL temperature change during test	<1.0 °C	---	---																																																										
SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition																																																												
SAR measured	250 mW input power	9.73 W/kg																																																											
SAR for nominal Head TSL parameters	normalized to 1W	38.9 W/kg ± 18.8 % (k=2)																																																											
SAR averaged over 10 cm ³ (10 g) of Head TSL	Condition																																																												
SAR measured	250 mW input power	5.11 W/kg																																																											
SAR for nominal Head TSL parameters	normalized to 1W	20.4 W/kg ± 18.7 % (k=2)																																																											
Impedance, transformed to feed point	47.90-2.54jΩ																																																												
Return Loss	-29.4dB																																																												
Electrical Delay (one direction)	1.116 ns																																																												
Manufactured by	SPEAG																																																												
<p>TTL Speaq In Collaboration with CALIBRATION LABORATORY</p> <p>CAICT</p> <p>Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@china.ttl.com http://www.chinat.tl.com</p> <p>DASY5 Validation Report for Head TSL Date: 2022-03-31 Test Laboratory: CTTL, Beijing, China DUT: Dipole 1800 MHz; Type: D1800V2; Serial: D1800V2 - SN: 2d170 Communication System: UID 0, CW; Frequency: 1800 MHz; Duty Cycle: 1:1 Medium parameters used: f = 1800 MHz; σ = 1.411 S/m; ε = 40.62; ρ = 1000 kg/m³ Phantom section: Right Section Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007) DASY5 Configuration:</p> <ul style="list-style-type: none"> Probe: EX3DV4 - SN7307; ConvF(8.34, 8.34, 8.34) @ 1800 MHz; Calibrated: 2021-05-26 Sensor-Surface: 1.4mm (Mechanical Surface Detection) Electronics: DAE4 Sn1556; Calibrated: 2022-01-12 Phantom: MFP_V5.1C (20deg probe tilt); Type: QD 000 P51 Cx; Serial: 1062 DASY52 52.10.4(1535); SEMCAD X 14.6.14(7501) <p>Dipole Calibration/Zoom Scan (7x7x7) (7x7x7) Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm Reference Value = 98.14 V/m; Power Drift = 0.03 dB Peak SAR (extrapolated) = 18.2 W/kg SAR(1 g) = 9.73 W/kg; SAR(10 g) = 5.11 W/kg Smallest distance from peaks to all points 3 dB below = 10 mm Ratio of SAR at M2 to SAR at M1 = 54% Maximum value of SAR (measured) = 15.2 W/kg</p> <p>0 dB = 15.2 W/kg = 11.82 dBW/kg</p> <p>Certificate No: Z22-60105 Page 5 of 6</p>	<p>TTL Speaq In Collaboration with CALIBRATION LABORATORY</p> <p>CAICT</p> <p>Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@china.ttl.com http://www.chinat.tl.com</p> <p>Impedance Measurement Plot for Head TSL</p> <p>Certificate No: Z22-60105 Page 6 of 6</p>																																																												



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

No.10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300
中国·江苏·昆山市留学院创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgsgroup.com.cn
t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com

1.7 D1900V2 - SN 5d136

<p>Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42204633-2117 E-mail: vt@ttslab.com</p> <p>Client: SGS-CN Certificate No: Z22-60185</p> <h3>CALIBRATION CERTIFICATE</h3> <p>Object: D1900V2 - SN: 5d136</p> <p>Calibration Procedure(s): FF-Z11-003-01 Calibration Procedures for dipole validation kits</p> <p>Calibration date: June 7, 2022</p> <p>This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.</p> <p>All calibrations have been conducted in the closed laboratory facility; environment temperature (23±)°C and humidity <70%.</p> <p>Calibration Equipment used (M&TE critical for calibration)</p> <table border="1"> <thead> <tr> <th>Primary Standards</th> <th>ID #</th> <th>Cal Date (Calibrated by, Certificate No.)</th> <th>Scheduled Calibration</th> </tr> </thead> <tbody> <tr> <td>Power Meter NRP2</td> <td>106277</td> <td>24-Sep-21 (CTTL No. J21X08326)</td> <td>Sep-22</td> </tr> <tr> <td>Power Meter NRP6S</td> <td>104291</td> <td>24-Sep-21 (CTTL No. J21X08326)</td> <td>Sep-22</td> </tr> <tr> <td>Reference Probe EKSDV4</td> <td>SN 7484</td> <td>28-Jan-22 (SPEAG No. EX3-7484_Jan22)</td> <td>Jan-23</td> </tr> <tr> <td>DAE4</td> <td>SN 1656</td> <td>12-Jan-22 (CTTL-SPEAG No. Z22-60007)</td> <td>Jan-23</td> </tr> </tbody> </table> <table border="1"> <thead> <tr> <th>Secondary Standards</th> <th>ID #</th> <th>Cal Date (Calibrated by, Certificate No.)</th> <th>Scheduled Calibration</th> </tr> </thead> <tbody> <tr> <td>Signal Generator E4438C</td> <td>MY48671430</td> <td>13-Jan-22 (CTTL No. J22X00409)</td> <td>Jan-23</td> </tr> <tr> <td>Network Analyser E5071C</td> <td>MY48110073</td> <td>14-Jan-22 (CTTL No. J22X00406)</td> <td>Jan-23</td> </tr> </tbody> </table> <p>Calibrated by: Zhao Jing, SAR Test Engineer</p> <p>Reviewed by: Lin Hao, SAR Test Engineer</p> <p>Approved by: Qi Diaryuan, SAR Project Leader</p> <p>Issued: June 13, 2022</p> <p>This calibration certificate shall not be reproduced except in full without written approval of the laboratory.</p> <p>Certificate No: Z22-60185 Page 1 of 6</p>	Primary Standards	ID #	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration	Power Meter NRP2	106277	24-Sep-21 (CTTL No. J21X08326)	Sep-22	Power Meter NRP6S	104291	24-Sep-21 (CTTL No. J21X08326)	Sep-22	Reference Probe EKSDV4	SN 7484	28-Jan-22 (SPEAG No. EX3-7484_Jan22)	Jan-23	DAE4	SN 1656	12-Jan-22 (CTTL-SPEAG No. Z22-60007)	Jan-23	Secondary Standards	ID #	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration	Signal Generator E4438C	MY48671430	13-Jan-22 (CTTL No. J22X00409)	Jan-23	Network Analyser E5071C	MY48110073	14-Jan-22 (CTTL No. J22X00406)	Jan-23	<p>Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42204633-2117 E-mail: vt@ttslab.com</p> <p>Glossary: TSL: tissue simulating liquid ComF: sensitivity in TSL / NORMx.y.z NA: not applicable or not measured</p> <p>Calibration is Performed According to the Following Standards: a) IEC/IEEE 62209-1528, "Measurement Procedure for The Assessment of Specific Absorption Rate of Human Exposure to Radio Frequency Fields from Hand-held and Body-mounted Wireless Communication Devices- Part 1528: Human Models, Instrumentation and Procedures (Frequency range of 4 MHz to 10 GHz)", October 2020 b) KDB 865984, "SAR Measurement Requirements for 100 MHz to 6 GHz"</p> <p>Additional Documentation: c) DASY4/S System Handbook</p> <p>Methods Applied and Interpretation of Parameters: • Measurement Conditions: Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated. • Antenna Parameters with TSL: The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis. • Feed Point Impedance and Return Loss: These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required. • Electrical Delay: One-way delay between the SMA connector and the antenna feed point. No uncertainty required. • SAR measured: SAR measured at the stated antenna input power. • SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna connector. • SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result.</p> <p>The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.</p> <p>Certificate No: Z22-60185 Page 2 of 6</p>																												
Primary Standards	ID #	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration																																																										
Power Meter NRP2	106277	24-Sep-21 (CTTL No. J21X08326)	Sep-22																																																										
Power Meter NRP6S	104291	24-Sep-21 (CTTL No. J21X08326)	Sep-22																																																										
Reference Probe EKSDV4	SN 7484	28-Jan-22 (SPEAG No. EX3-7484_Jan22)	Jan-23																																																										
DAE4	SN 1656	12-Jan-22 (CTTL-SPEAG No. Z22-60007)	Jan-23																																																										
Secondary Standards	ID #	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration																																																										
Signal Generator E4438C	MY48671430	13-Jan-22 (CTTL No. J22X00409)	Jan-23																																																										
Network Analyser E5071C	MY48110073	14-Jan-22 (CTTL No. J22X00406)	Jan-23																																																										
<p>Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42204633-2117 E-mail: vt@ttslab.com</p> <p>Measurement Conditions DASY system configuration, as far as not given on page 1.</p> <table border="1"> <thead> <tr> <th>DASY Version</th> <th>DASY52</th> <th>52.10.4</th> </tr> </thead> <tbody> <tr> <td>Extrapolation</td> <td>Advanced Extrapolation</td> <td></td> </tr> <tr> <td>Phantom</td> <td>Triple Flat Phantom 5.1C</td> <td></td> </tr> <tr> <td>Distance Dipole Center - TSL</td> <td>10 mm</td> <td>with Spacer</td> </tr> <tr> <td>Zoom Scan Resolution</td> <td>dx, dy, dz = 5 mm</td> <td></td> </tr> <tr> <td>Frequency</td> <td>1900 MHz ± 1 MHz</td> <td></td> </tr> </tbody> </table> <p>Head TSL parameters The following parameters and calculations were applied:</p> <table border="1"> <thead> <tr> <th>Nominal Head TSL parameters</th> <th>Temperature</th> <th>Permittivity</th> <th>Conductivity</th> </tr> </thead> <tbody> <tr> <td></td> <td>22.0 °C</td> <td>40.0</td> <td>1.40 mS/m</td> </tr> <tr> <td>Measured Head TSL parameters</td> <td>(22.0 ± 0.2) °C</td> <td>39.9 ± 0.6 %</td> <td>1.39 mS/m ± 0.6 %</td> </tr> <tr> <td>Head TSL temperature change during test:</td> <td>-0.1 °C</td> <td></td> <td></td> </tr> </tbody> </table> <p>SAR result with Head TSL</p> <table border="1"> <thead> <tr> <th>SAR averaged over 1 cm² (1 g) of Head TSL</th> <th>Condition</th> <th></th> </tr> </thead> <tbody> <tr> <td>SAR measured</td> <td>250 mW input power</td> <td>9.66 W/kg</td> </tr> <tr> <td>SAR for nominal Head TSL parameters</td> <td>normalized to 1W</td> <td>40.0 W/kg ± 18.8 % (k=2)</td> </tr> </tbody> </table> <table border="1"> <thead> <tr> <th>SAR averaged over 10 cm² (10 g) of Head TSL</th> <th>Condition</th> <th></th> </tr> </thead> <tbody> <tr> <td>SAR measured</td> <td>250 mW input power</td> <td>5.18 W/kg</td> </tr> <tr> <td>SAR for nominal Head TSL parameters</td> <td>normalized to 1W</td> <td>20.8 W/kg ± 18.7 % (k=2)</td> </tr> </tbody> </table> <p>Certificate No: Z22-60185 Page 3 of 6</p>	DASY Version	DASY52	52.10.4	Extrapolation	Advanced Extrapolation		Phantom	Triple Flat Phantom 5.1C		Distance Dipole Center - TSL	10 mm	with Spacer	Zoom Scan Resolution	dx, dy, dz = 5 mm		Frequency	1900 MHz ± 1 MHz		Nominal Head TSL parameters	Temperature	Permittivity	Conductivity		22.0 °C	40.0	1.40 mS/m	Measured Head TSL parameters	(22.0 ± 0.2) °C	39.9 ± 0.6 %	1.39 mS/m ± 0.6 %	Head TSL temperature change during test:	-0.1 °C			SAR averaged over 1 cm ² (1 g) of Head TSL	Condition		SAR measured	250 mW input power	9.66 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	40.0 W/kg ± 18.8 % (k=2)	SAR averaged over 10 cm ² (10 g) of Head TSL	Condition		SAR measured	250 mW input power	5.18 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	20.8 W/kg ± 18.7 % (k=2)	<p>Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42204633-2117 E-mail: vt@ttslab.com</p> <p>Appendix (Additional assessments outside the scope of CNAS L0570)</p> <p>Antenna Parameters with Head TSL</p> <table border="1"> <thead> <tr> <th>Impedance, transformed to feed point</th> <th>51.20 ± 7.56(j)</th> </tr> </thead> <tbody> <tr> <td>Return Loss</td> <td>-22.4dB</td> </tr> </tbody> </table> <p>General Antenna Parameters and Design</p> <table border="1"> <thead> <tr> <th>Electrical Delay (one direction)</th> <th>1.109 ns</th> </tr> </thead> </table> <p>After long term use with 100W radiated power, only a slight warming of the dipole near the feed-point can be measured.</p> <p>The dipole is made of standard semi-rigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as specified in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard. No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feed-point may be damaged.</p> <p>Additional EUT Data</p> <table border="1"> <thead> <tr> <th>Manufactured by</th> <th>SPEAG</th> </tr> </thead> </table> <p>Certificate No: Z22-60185 Page 4 of 6</p>	Impedance, transformed to feed point	51.20 ± 7.56(j)	Return Loss	-22.4dB	Electrical Delay (one direction)	1.109 ns	Manufactured by	SPEAG
DASY Version	DASY52	52.10.4																																																											
Extrapolation	Advanced Extrapolation																																																												
Phantom	Triple Flat Phantom 5.1C																																																												
Distance Dipole Center - TSL	10 mm	with Spacer																																																											
Zoom Scan Resolution	dx, dy, dz = 5 mm																																																												
Frequency	1900 MHz ± 1 MHz																																																												
Nominal Head TSL parameters	Temperature	Permittivity	Conductivity																																																										
	22.0 °C	40.0	1.40 mS/m																																																										
Measured Head TSL parameters	(22.0 ± 0.2) °C	39.9 ± 0.6 %	1.39 mS/m ± 0.6 %																																																										
Head TSL temperature change during test:	-0.1 °C																																																												
SAR averaged over 1 cm ² (1 g) of Head TSL	Condition																																																												
SAR measured	250 mW input power	9.66 W/kg																																																											
SAR for nominal Head TSL parameters	normalized to 1W	40.0 W/kg ± 18.8 % (k=2)																																																											
SAR averaged over 10 cm ² (10 g) of Head TSL	Condition																																																												
SAR measured	250 mW input power	5.18 W/kg																																																											
SAR for nominal Head TSL parameters	normalized to 1W	20.8 W/kg ± 18.7 % (k=2)																																																											
Impedance, transformed to feed point	51.20 ± 7.56(j)																																																												
Return Loss	-22.4dB																																																												
Electrical Delay (one direction)	1.109 ns																																																												
Manufactured by	SPEAG																																																												



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

No.10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300
中国·江苏·昆山市留学生创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgs.com.cn
t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com

DASY5 Validation Report for Head TSL

Test Laboratory: CTTL, Beijing, China
 DUT: Dipole 1900 MHz; Type: D1900V2; Serial: D1900V2; SN: 54136
 Communication System: UTD 0, CW; Frequency: 1900 MHz; Duty Cycle: 1:1
 Medium parameters used: $f = 1900 \text{ MHz}$; $\alpha = 1.385 \text{ S/m}$; $\epsilon_r = 39.85$; $\rho = 1000 \text{ kg/m}^3$
 Phantom section: Right Section
 Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007)
 DASY5 Configuration:

- Probe: EX3DV4 - SN7464; ConvF(8, 18, 8, 18) @ 1900 MHz; Calibrated: 2022-01-26
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn1556; Calibrated: 2022-01-12
- Phantom: MFP_V5_1C (20kg probe kit); Type: QD 000 P51 Cx; Serial: 1062
- DASY52.52.10.4(555); SEMCAD X.14.6.14(7501)

Dipole Calibration/Zoom Scan (7x7x7) (7x7x7) Cube D: Measurement grid: dx=5mm, dy=5mm, dz=5mm
 Reference Value = 99.99 V/m; Power Drift = 0.04 dB
 Peak SAR (extrapolated) = 18.6 W/kg
 SAR(1g) = 9.95 W/kg; SAR(10g) = 5.18 W/kg
 Smallest distance from peaks to all points 3 dB below = 9.2 mm
 Ratio of SAR at M2 to SAR at M1 = 54.1%

Maximum value of SAR (measured) = 15.6 W/kg

0 dB = 15.6 W/kg = 11.93 dBW/kg

Certificate No: Z22-60185 Page 6 of 6

Impedance Measurement Plot for Head TSL

Certificate No: Z22-60185 Page 6 of 6

1.8 D2000V2 - SN 1041

Client: SGS-CN Certificate No: Z22-60186

CALIBRATION CERTIFICATE

Object: D2000V2 - SN: 1041

Calibration Procedure(s): FF-Z11-003-01
 Calibration Procedures for dipole validation kits

Calibration date: June 6, 2022

This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility; environment temperature (22±)°C and humidity <70%.

Calibration Equipment used (M&TE critical for calibration)

Primary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration
Power Meter NRP2	106277	24-Sep-21 (CTTL No. J21X06326)	Sep-22
Power sensor NRP35	104291	24-Sep-21 (CTTL No. J21X06326)	Sep-22
Reference Probe EX3DV4	SN 7464	26-Jan-22 (SPEAG No. EX3-7464-Jan22)	Jan-23
DAE4	SN 1556	12-Jan-22 (CTTL-SPEAG No. Z22-60007)	Jan-23

Secondary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration
Signal Generator E4438C	MY49071430	13-Jan-22 (CTTL No. J22X00409)	Jan-23
Network Analyzer E5071C	MY48110673	14-Jan-22 (CTTL No. J22X00406)	Jan-23

Calibrated by: Zhao Jing SAR Test Engineer

Reviewed by: Lin Hao SAR Test Engineer

Approved by: Qi Dianyuan SAR Project Leader

Issued: June 13, 2022

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

Certificate No: Z22-60186 Page 1 of 6

Glossary:

TSL: Issue simulating liquid

ConvF: sensitivity in TSL / NDRMx.y.z

N/A: not applicable or not measured

Calibration is Performed According to the Following Standards:

- IEC/IEEE 62209-1528, "Measurement Procedure for The Assessment of Specific Absorption Rate of Human Exposure to Radio Frequency Fields from Hand-held and Body-mounted Wireless Communication Devices- Part 1528: Human Models, Instrumentation and Procedures (Frequency range of 4 MHz to 10 GHz)", October 2020
- KDB 865664, "SAR Measurement Requirements for 100 MHz to 6 GHz"

Additional Documentation:

- DASY4/5 System Handbook

Methods Applied and Interpretation of Parameters:

- Measurement Conditions:** Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL:** The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.
- Feed Point Impedance and Return Loss:** These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.
- Electrical Delay:** One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
- SAR measured:** SAR measured at the stated antenna input power.
- SAR normalized:** SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters:** The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor k=2, which for a normal distribution Corresponds to a coverage probability of approximately 95%.

Certificate No: Z22-60186 Page 2 of 6





Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

No.10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300
 中国·江苏·昆山市留学生创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgsgroup.com.cn
 t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com

<p>In Collaboration with TTL S P E A G CALIBRATION LABORATORY</p> <p>CAICT</p> <p>Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-4239683-2117 E-mail: ott@china.ttl.com http://www.caict.ac.cn</p> <p>Measurement Conditions DASY system configuration, as far as not given on page 1.</p> <table border="1"> <tr> <td>DASY Version</td> <td>DASY52</td> <td>52.10.4</td> </tr> <tr> <td>Extrapolation</td> <td>Advanced Extrapolation</td> <td></td> </tr> <tr> <td>Phantom</td> <td>Triple Flat Phantom 5.1C</td> <td></td> </tr> <tr> <td>Distance Dipole Center - TSL</td> <td>10 mm</td> <td>with Spacer</td> </tr> <tr> <td>Zoom Scan Resolution</td> <td>dx, dy, dz = 5 mm</td> <td></td> </tr> <tr> <td>Frequency</td> <td>2000 MHz ± 1 MHz</td> <td></td> </tr> </table> <p>Head TSL parameters The following parameters and calculations were applied.</p> <table border="1"> <tr> <td></td> <td>Temperature</td> <td>Permittivity</td> <td>Conductivity</td> </tr> <tr> <td>Nominal Head TSL parameters</td> <td>22.0 °C</td> <td>40.0</td> <td>1.40 mho/m</td> </tr> <tr> <td>Measured Head TSL parameters</td> <td>(22.0 ± 0.2) °C</td> <td>40.2 ± 0.5 %</td> <td>1.39 mho/m ± 0.5 %</td> </tr> <tr> <td>Head TSL temperature change during test</td> <td><1.0 °C</td> <td>---</td> <td>---</td> </tr> </table> <p>SAR result with Head TSL</p> <table border="1"> <tr> <td>SAR averaged over 1 cm³ (1 g) of Head TSL</td> <td>Condition</td> <td></td> </tr> <tr> <td>SAR measured</td> <td>250 mW input power</td> <td>10.4 W/kg</td> </tr> <tr> <td>SAR for nominal Head TSL parameters</td> <td>normalized to 1W</td> <td>41.8 W/kg ± 18.8 % (k=2)</td> </tr> <tr> <td>SAR averaged over 10 cm³ (10 g) of Head TSL</td> <td>Condition</td> <td></td> </tr> <tr> <td>SAR measured</td> <td>250 mW input power</td> <td>5.30 W/kg</td> </tr> <tr> <td>SAR for nominal Head TSL parameters</td> <td>normalized to 1W</td> <td>21.3 W/kg ± 18.7 % (k=2)</td> </tr> </table> <p>Certificate No: Z22-60186 Page 3 of 6</p>	DASY Version	DASY52	52.10.4	Extrapolation	Advanced Extrapolation		Phantom	Triple Flat Phantom 5.1C		Distance Dipole Center - TSL	10 mm	with Spacer	Zoom Scan Resolution	dx, dy, dz = 5 mm		Frequency	2000 MHz ± 1 MHz			Temperature	Permittivity	Conductivity	Nominal Head TSL parameters	22.0 °C	40.0	1.40 mho/m	Measured Head TSL parameters	(22.0 ± 0.2) °C	40.2 ± 0.5 %	1.39 mho/m ± 0.5 %	Head TSL temperature change during test	<1.0 °C	---	---	SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition		SAR measured	250 mW input power	10.4 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	41.8 W/kg ± 18.8 % (k=2)	SAR averaged over 10 cm ³ (10 g) of Head TSL	Condition		SAR measured	250 mW input power	5.30 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	21.3 W/kg ± 18.7 % (k=2)	<p>In Collaboration with TTL S P E A G CALIBRATION LABORATORY</p> <p>CAICT</p> <p>Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-4239683-2117 E-mail: ott@china.ttl.com http://www.caict.ac.cn</p> <p>Appendix (Additional assessments outside the scope of CNAS L6570)</p> <p>Antenna Parameters with Head TSL</p> <table border="1"> <tr> <td>Impedance, transformed to feed point</td> <td>48.4Ω ± 0.74jΩ</td> </tr> <tr> <td>Return Loss</td> <td>-34.9dB</td> </tr> </table> <p>General Antenna Parameters and Design</p> <table border="1"> <tr> <td>Electrical Delay (one direction)</td> <td>1.088 ns</td> </tr> </table> <p>After long term use with 100W radiated power, only a slight warming of the dipole near the feed-point can be measured.</p> <p>The dipole is made of standard semi-rigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, spiral and caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard.</p> <p>No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feed-point may be damaged.</p> <p>Additional EUT Data</p> <table border="1"> <tr> <td>Manufactured by</td> <td>SPEAG</td> </tr> </table> <p>Certificate No: Z22-60186 Page 4 of 6</p>	Impedance, transformed to feed point	48.4Ω ± 0.74jΩ	Return Loss	-34.9dB	Electrical Delay (one direction)	1.088 ns	Manufactured by	SPEAG
DASY Version	DASY52	52.10.4																																																											
Extrapolation	Advanced Extrapolation																																																												
Phantom	Triple Flat Phantom 5.1C																																																												
Distance Dipole Center - TSL	10 mm	with Spacer																																																											
Zoom Scan Resolution	dx, dy, dz = 5 mm																																																												
Frequency	2000 MHz ± 1 MHz																																																												
	Temperature	Permittivity	Conductivity																																																										
Nominal Head TSL parameters	22.0 °C	40.0	1.40 mho/m																																																										
Measured Head TSL parameters	(22.0 ± 0.2) °C	40.2 ± 0.5 %	1.39 mho/m ± 0.5 %																																																										
Head TSL temperature change during test	<1.0 °C	---	---																																																										
SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition																																																												
SAR measured	250 mW input power	10.4 W/kg																																																											
SAR for nominal Head TSL parameters	normalized to 1W	41.8 W/kg ± 18.8 % (k=2)																																																											
SAR averaged over 10 cm ³ (10 g) of Head TSL	Condition																																																												
SAR measured	250 mW input power	5.30 W/kg																																																											
SAR for nominal Head TSL parameters	normalized to 1W	21.3 W/kg ± 18.7 % (k=2)																																																											
Impedance, transformed to feed point	48.4Ω ± 0.74jΩ																																																												
Return Loss	-34.9dB																																																												
Electrical Delay (one direction)	1.088 ns																																																												
Manufactured by	SPEAG																																																												
<p>In Collaboration with TTL S P E A G CALIBRATION LABORATORY</p> <p>CAICT</p> <p>Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-4239683-2117 E-mail: ott@china.ttl.com http://www.caict.ac.cn</p> <p>DASY5 Validation Report for Head TSL Date: 2022-06-06 Test Laboratory: CCTL, Beijing, China DUT: Dipole 2000 MHz; Type: D2000V2; Serial: D2000V2 - SN: 1041 Communication System: UFD 0; CW; Frequency: 2000 MHz; Duty Cycle: 1:1 Medium parameters used: f = 2000 MHz; σ = 1.392 S/m; ε_r = 40.21; ρ = 1000 kg/m³ Phantom section: Right Section Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007) DASY5 Configuration:</p> <ul style="list-style-type: none"> Probe: EX3DV4 - SN7464; ConvF(R,2, 8.2, 8.2) @ 2000 MHz; Calibrated: 2022-01-26 Sensor-Surface: 1.4mm (Mechanical Surface Detection) Electronics: DA64 Sn1556; Calibrated: 2022-01-12 Phantom: MFP_V5.1C (20dkg probe tilt); Type: QD 000 P51 Cx; Serial: 1062 DASY52: S2.10.4(1555); SEMCAD X 14.6.14(7501) <p>Dipole Calibration/Zoom Scan (7x7x7) (7x7x7) Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm Reference Value = 103.4 V/m; Power Drift = 0.03 dB Peak SAR (extrapolated) = 19.6 W/kg SAR(1 g) = 10.4 W/kg; SAR(10 g) = 5.3 W/kg Smallest distance from peaks to all points 3 dB below = 9.1 mm Ratio of SAR at M2 to SAR at M1 = 53.6% Maximum value of SAR (measured) = 16.3 W/kg</p>  <p>Certificate No: Z22-60186 Page 5 of 6</p>	<p>In Collaboration with TTL S P E A G CALIBRATION LABORATORY</p> <p>CAICT</p> <p>Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-4239683-2117 E-mail: ott@china.ttl.com http://www.caict.ac.cn</p> <p>Impedance Measurement Plot for Head TSL</p>  <p>Certificate No: Z22-60186 Page 6 of 6</p>																																																												



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

No.10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300
中国·江苏·昆山市留学院创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgs.com.cn
t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com

1.9 D2300V2 - SN 1096

<p>Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2512 Fax: +86-10-42304633-2504 E-mail: cti@china.ttl.com http://www.china.ttl.com</p>		<p>Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-3079 Fax: +86-10-42304633-2504 E-mail: cti@china.ttl.com http://www.china.ttl.com</p>	
<p>Client: SGS-CN Certificate No: Z22-60106</p>			
<p>CALIBRATION CERTIFICATE</p>			
Object	D2300V2 - SN 1096		
Calibration Procedure(s)	FF-Z11-003-01 Calibration Procedures for dipole validation kits		
Calibration date:	March 31, 2022		
<p>This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.</p>			
<p>All calibrations have been conducted in the closed laboratory facility: environment temperature (22±3)°C and humidity<70%.</p>			
<p>Calibration Equipment used (M&TE critical for calibration)</p>			
Primary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration
Power Meter NRP2	108277	24-Sep-21 (CTTL No.J21X08328)	Sep-22
Power sensor NRP8S	104291	24-Sep-21 (CTTL No.J21X08328)	Sep-22
Reference Probe EX30V4	SN 7307	26-May-21(SPEAG.No.EK3-7307_May21)	May-22
DAE4	SN 1556	12-Jan-22(CTTL-SPEAG.No.Z22-60007)	Jan-23
Secondary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration
Signal Generator E4438C	MY49071430	13-Jan-22 (CTTL No.J22X04006)	Jan-23
Network Analyzer E5071C	MY46110673	14-Jan-22 (CTTL No.J22X04006)	Jan-23
Calibrated by:	Name	Function	Signature
	Zhao Jing	SAR Test Engineer	
Reviewed by:	Lin Hao	SAR Test Engineer	
Approved by:	Qi Diaryuan	SAR Project Leader	
<p>This calibration certificate shall not be reproduced except in full without written approval of the laboratory.</p>			
Certificate No: Z22-60106		Page 1 of 6	

<p>Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-3079 Fax: +86-10-42304633-2504 E-mail: cti@china.ttl.com http://www.china.ttl.com</p>		<p>Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-3079 Fax: +86-10-42304633-2504 E-mail: cti@china.ttl.com http://www.china.ttl.com</p>	
<p>Measurement Conditions DASY system configuration, as far as not given on page 1</p>			
DASY Version	DASY52	52.10.4	
Extrapolation	Advanced Extrapolation		
Phantom	Triple Flat Phantom 5.1C		
Distance Dipole Center - TSL	10 mm	with Spacer	
Zoom Scan Resolution	dx, dy, dz = 5 mm		
Frequency	2300 MHz ± 1 MHz		
<p>Head TSL parameters The following parameters and calculations were applied:</p>			
	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	39.5	1.67 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	39.8 ± 0.6 %	1.70 mho/m ± 6 %
Head TSL temperature change during test	<1.0 °C	---	
<p>SAR result with Head TSL</p>			
SAR averaged over 1 cm ² (1 g) of Head TSL	Condition		
SAR measured	250 mW input power	12.4 W/kg	
SAR for nominal Head TSL parameters	normalized to 1W	49.2 W/kg ± 18.8 % (k=2)	
SAR averaged over 10 cm ² (10 g) of Head TSL	Condition		
SAR measured	250 mW input power	5.88 W/kg	
SAR for nominal Head TSL parameters	normalized to 1W	23.4 W/kg ± 18.7 % (k=2)	
<p>Appendix (Additional assessments outside the scope of CNAS L0570)</p>			
<p>Antenna Parameters with Head TSL</p>			
Impedance, transformed to feed point	49.2Ω -j 4.56(j)		
Return Loss	-26.6dB		
<p>General Antenna Parameters and Design</p>			
Electrical Delay (one direction)	1.083 ns		
<p>After long term use with 100W radiated power, only a slight warming of the dipole near the feed-point can be measured.</p>			
<p>The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small and caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard. No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feed-point may be damaged.</p>			
<p>Additional EUT Data</p>			
Manufactured by	SPEAG		
Certificate No: Z22-60106		Page 4 of 6	





Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

No.10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300
中国·江苏·昆山市留学生创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgs.com.cn
t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com



1.10 D2450V2 - SN 817

Primary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration
Power Meter NRP2	108277	24-Sep-21 (CTTL No.J21X08320)	Sep-22
Power sensor NRP8S	104291	24-Sep-21 (CTTL No.J21X08320)	Sep-22
Reference Probe EX3DV4	SN 7307	26-May-21(SPEAG.No.EX3-7307_May21)	May-22
DAE4	SN 1556	12-Jan-22(CTTL-SPEAG.No.Z22-60007)	Jan-23
Secondary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration
Signal Generator E4438C	MY49071430	13-Jan-22 (CTTL No. J22X00406)	Jan-23
Network Analyzer E5071C	MY46110873	14-Jan-22 (CTTL No. J22X00406)	Jan-23

Client: SGS-CN **Certificate No.:** Z22-60107

Object: D2450V2 - SN 817

Calibration Procedure(s): FF-Z11-003-01
 Calibration Procedures for dipole validation kits

Calibration date: April 1, 2022

Calibration is Performed According to the Following Standards:
 a) IEC/IEEE 62209-1528, "Measurement Procedure for The Assessment of Specific Absorption Rate of Human Exposure to Radio Frequency Fields from Hand-held and Body-mounted Wireless Communication Devices- Part 1528: Human Models, Instrumentation and Procedures (Frequency range of 4 MHz to 10 GHz)", October 2020
 b) KDB 855664, "SAR Measurement Requirements for 100 MHz to 6 GHz"

Additional Documentation:
 c) DASY4/5 System Handbook

Methods Applied and Interpretation of Parameters:
 • **Measurement Conditions:** Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
 • **Antenna Parameters with TSL:** The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.
 • **Feed Point Impedance and Return Loss:** These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.
 • **Electrical Delay:** One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
 • **SAR measured:** SAR measured at the stated antenna input power.
 • **SAR normalized:** SAR as measured, normalized to an input power of 1 W at the antenna connector.
 • **SAR for nominal TSL parameters:** The measured TSL parameters are used to calculate the nominal SAR result.



The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor $k=2$, which for a normal distribution Corresponds to a coverage probability of approximately 95%.



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.
Attention: To check the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

No.10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300
 中国·江苏·昆山市留学生创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgsgroup.com.cn
 t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com

<p>TTL Speaq CALIBRATION LABORATORY In Collaboration with CAICT</p> <p>Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@chinaeui.com http://www.chinaeui.cn</p> <p>Measurement Conditions DASYS system configuration, as far as not given on page 1</p> <table border="1"> <tr><td>DASY Version</td><td>DASY52</td><td>52.10.4</td></tr> <tr><td>Extrapolation</td><td>Advanced Extrapolation</td><td></td></tr> <tr><td>Phantom</td><td>Triple Flat Phantom 5.1C</td><td></td></tr> <tr><td>Distance Dipole Center - TSL</td><td>10 mm</td><td>with Spacer</td></tr> <tr><td>Zoom Scan Resolution</td><td>dx, dy, dz = 5 mm</td><td></td></tr> <tr><td>Frequency</td><td>2450 MHz ± 1 MHz</td><td></td></tr> </table> <p>Head TSL parameters The following parameters and calculations were applied.</p> <table border="1"> <tr><th></th><th>Temperature</th><th>Permittivity</th><th>Conductivity</th></tr> <tr><td>Nominal Head TSL parameters</td><td>22.0 °C</td><td>39.2</td><td>1.80 mho/m</td></tr> <tr><td>Measured Head TSL parameters</td><td>(22.0 ± 0.2) °C</td><td>39.5 ± 6 %</td><td>1.79 mho/m ± 6 %</td></tr> <tr><td>Head TSL temperature change during test</td><td><1.0 °C</td><td>---</td><td>---</td></tr> </table> <p>SAR result with Head TSL</p> <table border="1"> <tr><td>SAR averaged over 1 cm³ (1 g) of Head TSL</td><td>Condition</td><td></td></tr> <tr><td>SAR measured</td><td>250 mW input power</td><td>13.2 W/kg</td></tr> <tr><td>SAR for nominal Head TSL parameters</td><td>normalized to 1W</td><td>53.0 W/kg ± 18.8 % (k=2)</td></tr> <tr><td>SAR averaged over 10 cm³ (10 g) of Head TSL</td><td>Condition</td><td></td></tr> <tr><td>SAR measured</td><td>250 mW input power</td><td>6.15 W/kg</td></tr> <tr><td>SAR for nominal Head TSL parameters</td><td>normalized to 1W</td><td>24.7 W/kg ± 18.7 % (k=2)</td></tr> </table> <p>Certificate No: Z22-60107 Page 3 of 6</p>	DASY Version	DASY52	52.10.4	Extrapolation	Advanced Extrapolation		Phantom	Triple Flat Phantom 5.1C		Distance Dipole Center - TSL	10 mm	with Spacer	Zoom Scan Resolution	dx, dy, dz = 5 mm		Frequency	2450 MHz ± 1 MHz			Temperature	Permittivity	Conductivity	Nominal Head TSL parameters	22.0 °C	39.2	1.80 mho/m	Measured Head TSL parameters	(22.0 ± 0.2) °C	39.5 ± 6 %	1.79 mho/m ± 6 %	Head TSL temperature change during test	<1.0 °C	---	---	SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition		SAR measured	250 mW input power	13.2 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	53.0 W/kg ± 18.8 % (k=2)	SAR averaged over 10 cm ³ (10 g) of Head TSL	Condition		SAR measured	250 mW input power	6.15 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	24.7 W/kg ± 18.7 % (k=2)	<p>TTL Speaq CALIBRATION LABORATORY In Collaboration with CAICT</p> <p>Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@chinaeui.com http://www.chinaeui.cn</p> <p>Appendix (Additional assessments outside the scope of CNAS L0570)</p> <p>Antenna Parameters with Head TSL</p> <table border="1"> <tr><td>Impedance, transformed to feed point</td><td>52.10 ± 3.20Ω</td></tr> <tr><td>Return Loss</td><td>-28.5dB</td></tr> </table> <p>General Antenna Parameters and Design</p> <table border="1"> <tr><td>Electrical Delay (one direction)</td><td>1.066 ns</td></tr> </table> <p>After long term use with 100W radiated power, only a slight warming of the dipole near the feed-point can be measured.</p> <p>The dipole is made of standard serringid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard. No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feed-point may be damaged.</p> <p>Additional EUT Data</p> <table border="1"> <tr><td>Manufactured by</td><td>SPEAG</td></tr> </table> <p>Certificate No: Z22-60107 Page 4 of 6</p>	Impedance, transformed to feed point	52.10 ± 3.20Ω	Return Loss	-28.5dB	Electrical Delay (one direction)	1.066 ns	Manufactured by	SPEAG
DASY Version	DASY52	52.10.4																																																											
Extrapolation	Advanced Extrapolation																																																												
Phantom	Triple Flat Phantom 5.1C																																																												
Distance Dipole Center - TSL	10 mm	with Spacer																																																											
Zoom Scan Resolution	dx, dy, dz = 5 mm																																																												
Frequency	2450 MHz ± 1 MHz																																																												
	Temperature	Permittivity	Conductivity																																																										
Nominal Head TSL parameters	22.0 °C	39.2	1.80 mho/m																																																										
Measured Head TSL parameters	(22.0 ± 0.2) °C	39.5 ± 6 %	1.79 mho/m ± 6 %																																																										
Head TSL temperature change during test	<1.0 °C	---	---																																																										
SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition																																																												
SAR measured	250 mW input power	13.2 W/kg																																																											
SAR for nominal Head TSL parameters	normalized to 1W	53.0 W/kg ± 18.8 % (k=2)																																																											
SAR averaged over 10 cm ³ (10 g) of Head TSL	Condition																																																												
SAR measured	250 mW input power	6.15 W/kg																																																											
SAR for nominal Head TSL parameters	normalized to 1W	24.7 W/kg ± 18.7 % (k=2)																																																											
Impedance, transformed to feed point	52.10 ± 3.20Ω																																																												
Return Loss	-28.5dB																																																												
Electrical Delay (one direction)	1.066 ns																																																												
Manufactured by	SPEAG																																																												
<p>TTL Speaq CALIBRATION LABORATORY In Collaboration with CAICT</p> <p>Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@chinaeui.com http://www.chinaeui.cn</p> <p>DASY5 Validation Report for Head TSL Date: 2022-04-01 Test Laboratory: CTTL, Beijing, China DUT: Dipole 2450 MHz; Type: D2450V2; Serial: D2450V2 - SN: 817 Communication System: UID 0, CW; Frequency: 2450 MHz; Duty Cycle: 1:1 Medium parameters used: f = 2450 MHz; σ = 1.79 S/m; ε_r = 39.52; ρ = 1000 kg/m³ Phantom section: Right Section Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007) DASY5 Configuration:</p> <ul style="list-style-type: none"> Probe: EX3DV4 - SN7307; ConvF(7.75, 7.75) @ 2450 MHz; Calibrated: 2021-05-26 Sensor-Surface: 1.4mm (Mechanical Surface Detection) Electronics: DA14 Sn1556; Calibrated: 2022-01-12 Phantom: MFP_V5.1C (20deg probe tilt); Type: QD 000 P51 Cx; Serial: 1062 DASY5 52.10.4(1535); SEMCAD X 14.6.14(7501) <p>Dipole Calibration Zoom Scan (7x7x7) (Cube 0; Measurement grid: dx=5mm, dy=5mm, dz=5mm Reference Value = 104.6 V/m; Power Drift = -0.03 dB Peak SAR (extrapolated) = 27.0 W/kg SAR(1 g) = 13.2 W/kg; SAR(10 g) = 6.15 W/kg Smallest distance from peaks to all points 3 dB below = 8.9 mm Ratio of SAR at M2 to SAR at M1 = -49.2% Maximum value of SAR (measured) = 22.1 W/kg</p>  <p>0 dB = 22.1 W/kg = 13.44 dBW/kg</p> <p>Certificate No: Z22-60107 Page 1 of 6</p>	<p>TTL Speaq CALIBRATION LABORATORY In Collaboration with CAICT</p> <p>Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@chinaeui.com http://www.chinaeui.cn</p> <p>Impedance Measurement Plot for Head TSL</p>  <p>Certificate No: Z22-60107 Page 6 of 6</p>																																																												



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx>; and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

No. 10, Weiyue Road, Innovation Park, Kunshan, Jiangsu, China 215300
 中国·江苏·昆山市留学院创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgs.com.cn
 t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com

1.11 D2600V2 - SN 1158

<p>TTL S p e a g CALIBRATION LABORATORY</p> <p>Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2512 Fax: +86-10-42304633-2504 E-mail: cti@china.ttl.com http://www.chinatit.com</p> <p>Client: SGS-CN Certificate No: Z22-60108</p> <p>CALIBRATION CERTIFICATE</p> <p>Object: D2600V2 - SN: 1158</p> <p>Calibration Procedure(s): FF-Z11-003-01 Calibration Procedures for dipole validation kits</p> <p>Calibration date: March 31, 2022</p> <p>This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.</p> <p>All calibrations have been conducted in the closed laboratory facility: environment temperature (23±3)°C and humidity<70%.</p> <p>Calibration Equipment used (M&TE critical for calibration)</p> <table border="1"> <thead> <tr> <th>Primary Standards</th> <th>ID #</th> <th>Cal Date (Calibrated by Certificate No.)</th> <th>Scheduled Calibration</th> </tr> </thead> <tbody> <tr> <td>Power Meter NRP2</td> <td>102577</td> <td>24-Sep-21 (CTTL No.J21X08326)</td> <td>Sep-22</td> </tr> <tr> <td>Power sensor NRP8S</td> <td>104291</td> <td>24-Sep-21 (CTTL No.J21X08326)</td> <td>Sep-22</td> </tr> <tr> <td>Reference Probe EX3/DVA</td> <td>SN 7307</td> <td>26-May-21(SPEAG.No.EX3-7307_May21)</td> <td>May-22</td> </tr> <tr> <td>DAE4</td> <td>SN 1556</td> <td>12-Jan-22(CTTL-SPEAG.No.Z22-60007)</td> <td>Jan-23</td> </tr> </tbody> </table> <table border="1"> <thead> <tr> <th>Secondary Standards</th> <th>ID #</th> <th>Cal Date (Calibrated by Certificate No.)</th> <th>Scheduled Calibration</th> </tr> </thead> <tbody> <tr> <td>Signal Generator E4438C</td> <td>MY49071430</td> <td>13-Jan-22 (CTTL No.J22X00406)</td> <td>Jan-23</td> </tr> <tr> <td>Network Analyzer E5071C</td> <td>MY46110673</td> <td>14-Jan-22 (CTTL No.J22X00406)</td> <td>Jan-23</td> </tr> </tbody> </table> <p>Calibrated by: Zhao Jing SAR Test Engineer</p> <p>Reviewed by: Lin Hao SAR Test Engineer</p> <p>Approved by: Qi Dianyuan SAR Project Leader</p> <p>Issued: April 6, 2022</p> <p>This calibration certificate shall not be reproduced except in full without written approval of the laboratory.</p> <p>Certificate No: Z22-60108 Page 1 of 6</p>	Primary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration	Power Meter NRP2	102577	24-Sep-21 (CTTL No.J21X08326)	Sep-22	Power sensor NRP8S	104291	24-Sep-21 (CTTL No.J21X08326)	Sep-22	Reference Probe EX3/DVA	SN 7307	26-May-21(SPEAG.No.EX3-7307_May21)	May-22	DAE4	SN 1556	12-Jan-22(CTTL-SPEAG.No.Z22-60007)	Jan-23	Secondary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration	Signal Generator E4438C	MY49071430	13-Jan-22 (CTTL No.J22X00406)	Jan-23	Network Analyzer E5071C	MY46110673	14-Jan-22 (CTTL No.J22X00406)	Jan-23	<p>TTL S p e a g CALIBRATION LABORATORY</p> <p>Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@china.ttl.com http://www.chinatit.com</p> <p>CAICT</p> <p>Glossary: TSL: tissue simulating liquid ConvF: sensitivity in TSL / NORMx.y.z N/A: not applicable or not measured</p> <p>Calibration is Performed According to the Following Standards: a) IEC/IEEE 62209-1528, "Measurement Procedure for The Assessment of Specific Absorption Rate of Human Exposure to Radio Frequency Fields from Hand-held and Body-mounted Wireless Communication Devices- Part 1528: Human Models, Instrumentation and Procedures (Frequency range of 4 MHz to 10 GHz)", October 2020 b) KDB 865864, "SAR Measurement Requirements for 100 MHz to 6 GHz"</p> <p>Additional Documentation: c) DASY4/S System Handbook</p> <p>Methods Applied and Interpretation of Parameters:</p> <ul style="list-style-type: none"> Measurement Conditions: Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated. Antenna Parameters with TSL: The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis. Feed Point Impedance and Return Loss: These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required. Electrical Delay: One-way delay between the SMA connector and the antenna feed point. No uncertainty required. SAR measured: SAR measured at the stated antenna input power. SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna connector. SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result. <p>The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor k=2, which for a normal distribution Corresponds to a coverage probability of approximately 95%.</p> <p>Certificate No: Z22-60108 Page 2 of 6</p>																																	
Primary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration																																																															
Power Meter NRP2	102577	24-Sep-21 (CTTL No.J21X08326)	Sep-22																																																															
Power sensor NRP8S	104291	24-Sep-21 (CTTL No.J21X08326)	Sep-22																																																															
Reference Probe EX3/DVA	SN 7307	26-May-21(SPEAG.No.EX3-7307_May21)	May-22																																																															
DAE4	SN 1556	12-Jan-22(CTTL-SPEAG.No.Z22-60007)	Jan-23																																																															
Secondary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration																																																															
Signal Generator E4438C	MY49071430	13-Jan-22 (CTTL No.J22X00406)	Jan-23																																																															
Network Analyzer E5071C	MY46110673	14-Jan-22 (CTTL No.J22X00406)	Jan-23																																																															
<p>TTL S p e a g CALIBRATION LABORATORY</p> <p>Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-3079 Fax: +86-10-42304633-2504 E-mail: cti@china.ttl.com http://www.chinatit.com</p> <p>CAICT</p> <p>Measurement Conditions DASY system configuration, as far as not given on page 1.</p> <table border="1"> <thead> <tr> <th>Parameter</th> <th>Value</th> </tr> </thead> <tbody> <tr> <td>DASY Version</td> <td>DASY52</td> </tr> <tr> <td>Extrapolation</td> <td>Advanced Extrapolation</td> </tr> <tr> <td>Phantom</td> <td>Triple Flat Phantom 5.1C</td> </tr> <tr> <td>Distance Dipole Center - TSL</td> <td>10 mm with Spacer</td> </tr> <tr> <td>Zoom Scan Resolution</td> <td>dx, dy, dz = 5 mm</td> </tr> <tr> <td>Frequency</td> <td>2600 MHz ± 1 MHz</td> </tr> </tbody> </table> <p>Head TSL parameters The following parameters and calculations were applied:</p> <table border="1"> <thead> <tr> <th>Parameter</th> <th>Temperature</th> <th>Permittivity</th> <th>Conductivity</th> </tr> </thead> <tbody> <tr> <td>Nominal Head TSL parameters</td> <td>22.0 °C</td> <td>39.0</td> <td>1.96 mho/m</td> </tr> <tr> <td>Measured Head TSL parameters</td> <td>(22.0 ± 0.2) °C</td> <td>38.7 ± 0.5 %</td> <td>1.96 mho/m ± 0.5 %</td> </tr> <tr> <td>Head TSL temperature change during test</td> <td><1.0 °C</td> <td>---</td> <td>---</td> </tr> </tbody> </table> <p>SAR result with Head TSL</p> <table border="1"> <thead> <tr> <th>Parameter</th> <th>Condition</th> <th>Value</th> </tr> </thead> <tbody> <tr> <td>SAR averaged over 1 cm² (1 g) of Head TSL</td> <td>Condition</td> <td>13.7 W/kg</td> </tr> <tr> <td>SAR measured</td> <td>250 mW input power</td> <td>13.7 W/kg</td> </tr> <tr> <td>SAR for nominal Head TSL parameters</td> <td>normalized to 1W</td> <td>54.8 W/kg ± 18.8 % (k=2)</td> </tr> <tr> <td>SAR averaged over 10 cm² (10 g) of Head TSL</td> <td>Condition</td> <td>6.12 W/kg</td> </tr> <tr> <td>SAR measured</td> <td>250 mW input power</td> <td>6.12 W/kg</td> </tr> <tr> <td>SAR for nominal Head TSL parameters</td> <td>normalized to 1W</td> <td>24.6 W/kg ± 18.7 % (k=2)</td> </tr> </tbody> </table> <p>Certificate No: Z22-60108 Page 3 of 6</p>	Parameter	Value	DASY Version	DASY52	Extrapolation	Advanced Extrapolation	Phantom	Triple Flat Phantom 5.1C	Distance Dipole Center - TSL	10 mm with Spacer	Zoom Scan Resolution	dx, dy, dz = 5 mm	Frequency	2600 MHz ± 1 MHz	Parameter	Temperature	Permittivity	Conductivity	Nominal Head TSL parameters	22.0 °C	39.0	1.96 mho/m	Measured Head TSL parameters	(22.0 ± 0.2) °C	38.7 ± 0.5 %	1.96 mho/m ± 0.5 %	Head TSL temperature change during test	<1.0 °C	---	---	Parameter	Condition	Value	SAR averaged over 1 cm ² (1 g) of Head TSL	Condition	13.7 W/kg	SAR measured	250 mW input power	13.7 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	54.8 W/kg ± 18.8 % (k=2)	SAR averaged over 10 cm ² (10 g) of Head TSL	Condition	6.12 W/kg	SAR measured	250 mW input power	6.12 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	24.6 W/kg ± 18.7 % (k=2)	<p>TTL S p e a g CALIBRATION LABORATORY</p> <p>Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-3079 Fax: +86-10-42304633-2504 E-mail: cti@china.ttl.com http://www.chinatit.com</p> <p>CAICT</p> <p>Appendix (Additional assessments outside the scope of CNAS L0570)</p> <p>Antenna Parameters with Head TSL</p> <table border="1"> <thead> <tr> <th>Parameter</th> <th>Value</th> </tr> </thead> <tbody> <tr> <td>Impedance, transformed to feed point</td> <td>49.90- 6.49jΩ</td> </tr> <tr> <td>Return Loss</td> <td>-23.8dB</td> </tr> </tbody> </table> <p>General Antenna Parameters and Design</p> <table border="1"> <thead> <tr> <th>Parameter</th> <th>Value</th> </tr> </thead> <tbody> <tr> <td>Electrical Delay (one direction)</td> <td>1.053 ns</td> </tr> </tbody> </table> <p>After long term use with 100W radiated power, only a slight warming of the dipole near the feed-point can be measured.</p> <p>The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard. No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feed-point may be damaged.</p> <p>Additional EUT Data</p> <table border="1"> <thead> <tr> <th>Parameter</th> <th>Value</th> </tr> </thead> <tbody> <tr> <td>Manufactured by</td> <td>SPEAG</td> </tr> </tbody> </table> <p>Certificate No: Z22-60108 Page 4 of 6</p>	Parameter	Value	Impedance, transformed to feed point	49.90- 6.49jΩ	Return Loss	-23.8dB	Parameter	Value	Electrical Delay (one direction)	1.053 ns	Parameter	Value	Manufactured by	SPEAG
Parameter	Value																																																																	
DASY Version	DASY52																																																																	
Extrapolation	Advanced Extrapolation																																																																	
Phantom	Triple Flat Phantom 5.1C																																																																	
Distance Dipole Center - TSL	10 mm with Spacer																																																																	
Zoom Scan Resolution	dx, dy, dz = 5 mm																																																																	
Frequency	2600 MHz ± 1 MHz																																																																	
Parameter	Temperature	Permittivity	Conductivity																																																															
Nominal Head TSL parameters	22.0 °C	39.0	1.96 mho/m																																																															
Measured Head TSL parameters	(22.0 ± 0.2) °C	38.7 ± 0.5 %	1.96 mho/m ± 0.5 %																																																															
Head TSL temperature change during test	<1.0 °C	---	---																																																															
Parameter	Condition	Value																																																																
SAR averaged over 1 cm ² (1 g) of Head TSL	Condition	13.7 W/kg																																																																
SAR measured	250 mW input power	13.7 W/kg																																																																
SAR for nominal Head TSL parameters	normalized to 1W	54.8 W/kg ± 18.8 % (k=2)																																																																
SAR averaged over 10 cm ² (10 g) of Head TSL	Condition	6.12 W/kg																																																																
SAR measured	250 mW input power	6.12 W/kg																																																																
SAR for nominal Head TSL parameters	normalized to 1W	24.6 W/kg ± 18.7 % (k=2)																																																																
Parameter	Value																																																																	
Impedance, transformed to feed point	49.90- 6.49jΩ																																																																	
Return Loss	-23.8dB																																																																	
Parameter	Value																																																																	
Electrical Delay (one direction)	1.053 ns																																																																	
Parameter	Value																																																																	
Manufactured by	SPEAG																																																																	



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

No.10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300
中国·江苏·昆山市留学生创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgsgroup.com.cn
t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com

TTL Speaq CALIBRATION LABORATORY
 In Collaboration with **CAICT**

Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
 Tel: +86-10-62104633-2079 Fax: +86-10-62104633-2504
 E-mail: cti@china.ttl.com http://www.china.ttl.com

Date: 2022-03-31

DASY5 Validation Report for Head TSL
 Test Laboratory: CTTL, Beijing, China
 DUT: Dipole 2600 MHz; Type: D2600V2; Serial: D2600V2 - SN: 1158
 Communication System: UID 0, CW; Frequency: 2600 MHz; Duty Cycle: 1:1
 Phantom parameters used: $f = 2600$ MHz; $\sigma = 1.955$ S/m; $\epsilon_r = 38.68$; $\rho = 1000$ kg/m³
 Phantom section: Right Section
 Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007)
 DASY5 Configuration:

- Probe: EX3DV4 - SN7307; ConvF(7.5, 7.5, 7.5) @ 2600 MHz; Calibrated: 2021-05-26
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn1556; Calibrated: 2022-01-12
- Phantom: MFP-V5.1C (2ldag probe fill); Type: QD 000 P51 Cx; Serial: 1062
- DASY52.52.10.4(1535); SEMCAD X 14.6.14(7501)

Dipole Calibration/Zoom Scan (7x7x7) (7x7x7)/Cube 0; Measurement grid: dx=5mm, dy=5mm, dz=5mm
 Reference Value = 103.3 V/m; Power Drift = 0.04 dB
 Peak SAR (extrapolated) = 29.9 W/kg
 SAR(1 g) = 13.7 W/kg; SAR(10 g) = 6.12 W/kg
 Smallest distance from peaks to all points 3 dB below = 8.9 mm
 Ratio of SAR at M2 to SAR at M1 = 47.5%
 Maximum value of SAR (measured) = 23.4 W/kg

0 dB = 23.4 W/kg = 13.69 dBW/kg

Certificate No: Z22-60108 Page 5 of 6

TTL Speaq CALIBRATION LABORATORY
 In Collaboration with **CAICT**

Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
 Tel: +86-10-62104633-2079 Fax: +86-10-62104633-2504
 E-mail: cti@china.ttl.com http://www.china.ttl.com

Impedance Measurement Plot for Head TSL

Certificate No: Z22-60108 Page 6 of 6

1.12 D5GHzV2 - SN 1095

TTL Speaq CALIBRATION LABORATORY
 In Collaboration with **CAICT**

Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
 Tel: +86-10-62104633-2079 Fax: +86-10-62104633-2504
 E-mail: cti@china.ttl.com http://www.china.ttl.com

Client: **SGS-CN** Certificate No: **Z22-60187**

CALIBRATION CERTIFICATE

Object: D5GHzV2 - SN: 1095

Calibration Procedure(s): FF-Z11-003-01
 Calibration Procedures for dipole validation kits

Calibration date: June 1, 2022

This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature (23±1)°C and humidity < 70%.

Calibration Equipment used (MTE critical for calibration)

Primary Standards	ID #	Cal Date (Calibrated by: Certificate No.)	Scheduled Calibration
Power Meter NRP2	106277	24-Sep-21 (CTTL No. J21008326)	Sep-22
Power sensor NRP8S	104291	24-Sep-21 (CTTL No. J21008326)	Sep-22
Reference Probe EX3DV4	SN 7464	26-Jan-22(SPEAG No. EX3-7464_Jan22)	Jan-23
DAE4	SN 1556	12-Jan-22(CTTL-SPEAG No. Z22-60007)	Jan-23

Secondary Standards	ID #	Cal Date (Calibrated by: Certificate No.)	Scheduled Calibration
Signal Generator E4438C	MY46071430	13-Jan-22 (CTTL No. J2200406)	Jan-23
Network Analyzer E5071C	MY46110673	14-Jan-22 (CTTL No. J2200406)	Jan-23

Calibrated by: Zhao Jing, SAR Test Engineer

Reviewed by: Lin Hao, SAR Test Engineer

Approved by: Qi Dianyan, SAR Project Leader

Issued: June 6, 2022

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

Certificate No: Z22-60187 Page 1 of 10

TTL Speaq CALIBRATION LABORATORY
 In Collaboration with **CAICT**

Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
 Tel: +86-10-62104633-2079 Fax: +86-10-62104633-2504
 E-mail: cti@china.ttl.com http://www.china.ttl.com

Glossary:

TSL: Issue simulating liquid
 ConvF: sensitivity in TSL / NORMx,y,z
 N/A: not applicable or not measured

Calibration is Performed According to the Following Standards:

- IEC/IEEE 62209-1528, "Measurement Procedure for The Assessment of Specific Absorption Rate of Human Exposure to Radio Frequency Fields from Hand-held and Body-mounted Wireless Communication Devices- Part 1528: Human Models, Instrumentation and Procedures (Frequency range of 4 MHz to 10 GHz)", October 2020
- KDB 665664, "SAR Measurement Requirements for 100 MHz to 6 GHz"

Additional Documentation:

- DASY4/G System Handbook

Methods Applied and Interpretation of Parameters:

- Measurement Conditions: Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL: The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.
- Feed Point Impedance and Return Loss: These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.
- Electrical Delay: One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
- SAR measured: SAR measured at the stated antenna input power.
- SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor k=2, which for a normal distribution Corresponds to a coverage probability of approximately 95%.

Certificate No: Z22-60187 Page 2 of 10



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx>; and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN_Doccheck@sgs.com

No. 10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300
 中国·江苏·昆山市留学生创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgsgroup.com.cn
 t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com

In Collaboration with
TTL
CALIBRATION LABORATORY

Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
Tel: +86-10-62382117
E-mail: ott@ttsnarl.com <http://www.ttsnarl.com>

Measurement Conditions
DAEY system configuration, as far as not given on page 1.

DAEY Version	Temperature	Permittivity	Conductivity
DAEY Version	DAEYS2		52.10.4
Extrapolation	Advanced Extrapolation		
Phantom	Triple Flat Phantom 5.1C		
Distance Dipole Center - TSL	10 mm		with Spacer
Zoom Scan Resolution	dx, dy = 4 mm, dz = 1.4 mm	Graded Ratio = 1.4 (Z direction)	
Frequency	5000 MHz ± 1 MHz 5300 MHz ± 1 MHz 5500 MHz ± 1 MHz 5600 MHz ± 1 MHz 5800 MHz ± 1 MHz		

Head TSL parameters at 5200MHz
The following parameters and calculations were applied.

Nominal Head TSL parameters	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	35.0	4.66 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	35.4 ± 6 %	4.82 mho/m ± 6 %
Head TSL temperature change during test	<1.0 °C	---	---

SAR result with Head TSL at 5200MHz

SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition	SAR measured
SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition	7.79 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	77.8 W/kg ± 24.4 % (k=2)
SAR averaged over 10 cm ³ (10 g) of Head TSL	Condition	2.22 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	22.1 W/kg ± 24.2 % (k=2)

Certificate No: Z22-60187 Page 3 of 10

In Collaboration with
TTL
CALIBRATION LABORATORY

Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
Tel: +86-10-62382117
E-mail: ott@ttsnarl.com <http://www.ttsnarl.com>

Head TSL parameters at 5300MHz
The following parameters and calculations were applied.

Nominal Head TSL parameters	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	35.9	4.76 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	35.2 ± 6 %	4.73 mho/m ± 6 %
Head TSL temperature change during test	<1.0 °C	---	---

SAR result with Head TSL at 5300MHz

SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition	SAR measured
SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition	7.94 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	79.1 W/kg ± 24.4 % (k=2)
SAR averaged over 10 cm ³ (10 g) of Head TSL	Condition	2.27 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	22.6 W/kg ± 24.2 % (k=2)

Head TSL parameters at 5500MHz
The following parameters and calculations were applied.

Nominal Head TSL parameters	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	35.6	4.96 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	34.8 ± 6 %	4.94 mho/m ± 6 %
Head TSL temperature change during test	<1.0 °C	---	---

SAR result with Head TSL at 5500MHz

SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition	SAR measured
SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition	8.29 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	82.6 W/kg ± 24.4 % (k=2)
SAR averaged over 10 cm ³ (10 g) of Head TSL	Condition	2.34 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	23.3 W/kg ± 24.2 % (k=2)

Certificate No: Z22-60187 Page 4 of 10

In Collaboration with
TTL
CALIBRATION LABORATORY

Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
Tel: +86-10-62382117
E-mail: ott@ttsnarl.com <http://www.ttsnarl.com>

Head TSL parameters at 5600MHz
The following parameters and calculations were applied.

Nominal Head TSL parameters	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	35.5	5.07 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	34.7 ± 6 %	5.05 mho/m ± 6 %
Head TSL temperature change during test	<1.0 °C	---	---

SAR result with Head TSL at 5600MHz

SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition	SAR measured
SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition	8.12 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	80.8 W/kg ± 24.4 % (k=2)
SAR averaged over 10 cm ³ (10 g) of Head TSL	Condition	2.30 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	22.9 W/kg ± 24.2 % (k=2)

Head TSL parameters at 5800MHz
The following parameters and calculations were applied.

Nominal Head TSL parameters	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	35.3	5.27 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	34.4 ± 6 %	5.25 mho/m ± 6 %
Head TSL temperature change during test	<1.0 °C	---	---

SAR result with Head TSL at 5800MHz

SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition	SAR measured
SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition	7.71 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	76.7 W/kg ± 24.4 % (k=2)
SAR averaged over 10 cm ³ (10 g) of Head TSL	Condition	2.16 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	21.8 W/kg ± 24.2 % (k=2)

Certificate No: Z22-60187 Page 5 of 10

In Collaboration with
TTL
CALIBRATION LABORATORY

Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
Tel: +86-10-62382117
E-mail: ott@ttsnarl.com <http://www.ttsnarl.com>

Appendix (Additional assessments outside the scope of CNAS L0570)

Antenna Parameters with Head TSL at 5200MHz

Impedance, transformed to feed point	Return Loss
46.1D-8.03jD	-23.6dB

Antenna Parameters with Head TSL at 5300MHz

Impedance, transformed to feed point	Return Loss
47.8D-2.42jD	-28.5dB

Antenna Parameters with Head TSL at 5500MHz

Impedance, transformed to feed point	Return Loss
50.3D-4.26jD	-27.4dB

Antenna Parameters with Head TSL at 5600MHz

Impedance, transformed to feed point	Return Loss
54.5D-4.80jD	-24.0dB

Antenna Parameters with Head TSL at 5800MHz

Impedance, transformed to feed point	Return Loss
51.5D-5.61jD	-24.9dB

Certificate No: Z22-60187 Page 6 of 10



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx>; and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

No. 10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300
 中国 · 江苏 · 昆山市留学生创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgs.com.cn
 t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com

<p>In Collaboration with speaq CALIBRATION LABORATORY</p> <p>CAICT</p> <p>Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-62021117 E-mail: ott@ttsnet.com http://www.caict.ac.cn</p> <p>General Antenna Parameters and Design</p> <p>Electrical Delay (one direction): 1.101 ns</p> <p>After long term use with 100W radiated power, only a slight warming of the dipole near the feed-point can be measured.</p> <p>The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard. No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feed-point may be damaged.</p> <p>Additional EUT Data</p> <p>Manufactured by: SPEAG</p> <p>Certificate No: Z22-60187 Page 7 of 10</p>	<p>In Collaboration with speaq CALIBRATION LABORATORY</p> <p>CAICT</p> <p>Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-62021117 E-mail: ott@ttsnet.com http://www.caict.ac.cn</p> <p>DASY5 Validation Report for Head TSL</p> <p>Test Laboratory: CTTL, Beijing, China Date: 2022-06-01</p> <p>DUT: Dipole 5GHz; Type: D5GHzV2; Serial: D5GHzV2 - SN: 1095</p> <p>Communication System: CW; Frequency: 5200 MHz; Frequency: 5300 MHz; Frequency: 5500 MHz; Frequency: 5600 MHz; Frequency: 5800 MHz; Duty Cycle: 1:1 Medium parameters used: f = 5200 MHz; $\sigma = 4.62$ S/m; $\epsilon_r = 35.38$; $\rho = 1000$ kg/m³ Medium parameters used: f = 5300 MHz; $\sigma = 4.73$ S/m; $\epsilon_r = 35.19$; $\rho = 1000$ kg/m³ Medium parameters used: f = 5500 MHz; $\sigma = 4.939$ S/m; $\epsilon_r = 34.83$; $\rho = 1000$ kg/m³ Medium parameters used: f = 5600 MHz; $\sigma = 5.051$ S/m; $\epsilon_r = 34.68$; $\rho = 1000$ kg/m³ Medium parameters used: f = 5800 MHz; $\sigma = 5.247$ S/m; $\epsilon_r = 34.42$; $\rho = 1000$ kg/m³ Phantom section: Right Section Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007) DASY5 Configuration:</p> <ul style="list-style-type: none"> Probe: EX3DV4 - SN7484; ConvF(5.6, 5.6, 5.6) @ 5200 MHz; ConvF(5.32, 5.32, 5.32) @ 5300 MHz; ConvF(5.11, 5.11, 5.11) @ 5500 MHz; ConvF(4.91, 4.91, 4.91) @ 5600 MHz; ConvF(5, 5, 5) @ 5800 MHz; Calibrated: 2022-01-26 Sensor-Surface: 1.4mm (Mechanical Surface Detection) Electronics: DAE4 Sn1556; Calibrated: 2022-01-12 Phantom: MFP_V5.1C (20deg probe tilt); Type: QD 000 P1 Cx; Serial: 1062 DASY52 52.10.4(1535); SEMCAD X 14.6.14(7501) <p>Dipole Calibration /Pin=100mW, d=10mm, f=5200 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0; Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 60.80 V/m; Power Drift = -0.06 dB Peak SAR (extrapolated) = 29.8 W/kg SAR(1 g) = 7.73 W/kg; SAR(10 g) = 2.22 W/kg Smallest distance from peaks to all points 3 dB below = 7.2 mm Ratio of SAR at M2 to SAR at M1 = 66.8% Maximum value of SAR (measured) = 18.3 W/kg</p> <p>Dipole Calibration /Pin=100mW, d=10mm, f=5300 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0; Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 61.08 V/m; Power Drift = -0.07 dB Peak SAR (extrapolated) = 31.5 W/kg SAR(1 g) = 7.94 W/kg; SAR(10 g) = 2.27 W/kg Smallest distance from peaks to all points 3 dB below = 7.2 mm Ratio of SAR at M2 to SAR at M1 = 65.5% Maximum value of SAR (measured) = 19.0 W/kg</p> <p>Certificate No: Z22-60187 Page 8 of 10</p>
<p>In Collaboration with speaq CALIBRATION LABORATORY</p> <p>CAICT</p> <p>Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-62021117 E-mail: ott@ttsnet.com http://www.caict.ac.cn</p> <p>Dipole Calibration /Pin=100mW, d=10mm, f=5500 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0; Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 61.92 V/m; Power Drift = -0.08 dB Peak SAR (extrapolated) = 34.7 W/kg SAR(1 g) = 8.28 W/kg; SAR(10 g) = 2.34 W/kg Smallest distance from peaks to all points 3 dB below = 7.2 mm Ratio of SAR at M2 to SAR at M1 = 63.9% Maximum value of SAR (measured) = 20.2 W/kg</p> <p>Dipole Calibration /Pin=100mW, d=10mm, f=5600 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0; Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 65.08 V/m; Power Drift = -0.07 dB Peak SAR (extrapolated) = 35.2 W/kg SAR(1 g) = 8.12 W/kg; SAR(10 g) = 2.3 W/kg Smallest distance from peaks to all points 3 dB below = 7.2 mm Ratio of SAR at M2 to SAR at M1 = 62.5% Maximum value of SAR (measured) = 19.1 W/kg</p> <p>Dipole Calibration /Pin=100mW, d=10mm, f=5800 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0; Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 62.13 V/m; Power Drift = -0.06 dB Peak SAR (extrapolated) = 34.8 W/kg SAR(1 g) = 7.71 W/kg; SAR(10 g) = 2.16 W/kg Smallest distance from peaks to all points 3 dB below = 7.2 mm Ratio of SAR at M2 to SAR at M1 = 61.6% Maximum value of SAR (measured) = 18.7 W/kg</p> <p>0 dB = 18.7 W/kg = 12.72 dBW/kg</p> <p>Certificate No: Z22-60187 Page 9 of 10</p>	<p>In Collaboration with speaq CALIBRATION LABORATORY</p> <p>CAICT</p> <p>Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-62021117 E-mail: ott@ttsnet.com http://www.caict.ac.cn</p> <p>Impedance Measurement Plot for Head TSL</p> <p>Certificate No: Z22-60187 Page 10 of 10</p>



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

No.10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300
中国·江苏·昆山市留学生创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgs.com.cn
t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com

2 DAE4 - SN 1245

<div style="display: flex; justify-content: space-between; align-items: center;"> <div style="font-size: 8px;"> <p>Schmid & Partner Engineering AG Zugstrasse 10, 8001 Zurich, Switzerland Phone +41 44 248 9700, Fax +41 44 248 9770 www.sgs.com, info@sgs.com</p> </div> <div style="text-align: center; font-weight: bold; font-size: 12px;">s p e a g</div> <div style="font-size: 8px;"> <p>S Schweizerischer Kalibrierdienst C Service suisse d'étalonnage S Servizio svizzero di taratura S Swiss Calibration Service</p> </div> </div> <p style="text-align: center; color: red; font-weight: bold; margin-top: 10px;">IMPORTANT NOTICE</p> <p>USAGE OF THE DAE4</p> <p>The DAE unit is a delicate, high precision instrument and requires careful treatment by the user. There are no serviceable parts inside the DAE. Special attention shall be given to the following points:</p> <p>Battery Exchange: The battery cover of the DAE4 unit is fixed using a screw, over tightening the screw may cause the threads inside the DAE to wear out.</p> <p>Shipping of the DAE: Before shipping the DAE to SPEAG for calibration, remove the batteries and pack the DAE in an anti-static bag. This anti-static bag shall then be packed into a larger box or container which protects the DAE from impacts during transportation. The package shall be marked to indicate that a fragile instrument is inside.</p> <p>E-Stop Failures: Touch detection may be malfunctioning due to broken magnets in the E-stop. Rough handling of the E-stop may lead to damage of these magnets. Touch and collision errors are often caused by dust and dirt accumulated in the E-stop. To prevent E-stop failures, the customer shall always mount the probe to the DAE carefully and keep the DAE unit in a non-dusty environment if not used for measurements.</p> <p>Repair: Minor repairs are performed at no extra cost during the annual calibration. However, SPEAG reserves the right to charge for any repair especially if rough/unprofessional handling caused the defect.</p> <p>DASY Configuration Files: Since the exact values of the DAE input resistances, as measured during the calibration procedure of a DAE unit, are not used by the DASY software, a nominal value of 200 MOhm is given in the corresponding configuration file.</p> <div style="border: 1px solid black; padding: 2px; margin-top: 5px;"> <p>Important Note: Warranty and calibration is void if the DAE unit is disassembled partly or fully by the Customer.</p> </div> <div style="border: 1px solid black; padding: 2px; margin-top: 5px;"> <p>Important Note: Never attempt to grease or oil the E-stop assembly. Cleaning and readjusting of the E-stop assembly is allowed by certified SPEAG personnel only and is part of the annual calibration procedure.</p> </div> <div style="border: 1px solid black; padding: 2px; margin-top: 5px;"> <p>Important Note: To prevent damage of the DAE probe connector pins, use great care when installing the probe to the DAE. Carefully connect the probe with the connector notch oriented in the mating position. Avoid any rotational movement of the probe body versus the DAE while turning the locking nut of the connector. The same care shall be used when disconnecting the probe from the DAE.</p> </div> <p style="font-size: 8px; margin-top: 10px;">TN_EH160306AE_DAE4.docx 07.03.2019</p>	<div style="display: flex; justify-content: space-between; align-items: center;"> <div style="font-size: 8px;"> <p>Schmid & Partner Engineering AG Zugstrasse 10, 8001 Zurich, Switzerland Phone +41 44 248 9700, Fax +41 44 248 9770 www.sgs.com, info@sgs.com</p> </div> <div style="text-align: center; font-weight: bold; font-size: 12px;">s p e a g</div> <div style="font-size: 8px;"> <p>S Schweizerischer Kalibrierdienst C Service suisse d'étalonnage S Servizio svizzero di taratura S Swiss Calibration Service</p> </div> </div> <p style="font-size: 8px; margin-top: 5px;">Accredited by the Swiss Accreditation Service (SAS) The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates</p> <p style="text-align: right; font-size: 8px;">Accreditation No.: SCS 0108</p> <p style="font-size: 8px;">Client: SGS-CN (Auden) Certificate No.: DAE4-1245_May22</p> <div style="border: 1px solid black; padding: 5px; margin-top: 5px;"> <p style="text-align: center; font-weight: bold; font-size: 10px;">CALIBRATION CERTIFICATE</p> <p>Object: DAE4 - SD 000 D04 BM - SN: 1245</p> <p>Calibration procedure(s): QA CAL-06 v30 Calibration procedure for the data acquisition electronics (DAE)</p> <p>Calibration date: May 30, 2022</p> <p style="font-size: 8px; margin-top: 5px;">This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate. All calibrations have been conducted in the closed laboratory facility, environment temperature (22 ± 3)°C and humidity < 70%.</p> <p style="font-size: 8px;">Calibration Equipment used (M&E critical for calibration)</p> <table border="1" style="width: 100%; font-size: 8px; border-collapse: collapse;"> <thead> <tr> <th>Primary Standards</th> <th>ID #</th> <th>Exp. Date (Certificate No.)</th> <th>Scheduled Calibration</th> </tr> </thead> <tbody> <tr> <td>Kelway Multimeter Type 2001</td> <td>SN: 0810276</td> <td>31-Aug-21 (No:31368)</td> <td>Aug-22</td> </tr> </tbody> </table> <table border="1" style="width: 100%; font-size: 8px; border-collapse: collapse; margin-top: 5px;"> <thead> <tr> <th>Secondary Standards</th> <th>ID #</th> <th>Check Date (in house)</th> <th>Scheduled Check</th> </tr> </thead> <tbody> <tr> <td>Auto DAE Calibration Unit</td> <td>SE LWS 003 AA 1001</td> <td>24-Jan-22 (in house check)</td> <td>In house check Jan-23</td> </tr> <tr> <td>Calibrator class V0.1</td> <td>SE LWS 100 AA 1002</td> <td>24-Jan-22 (in house check)</td> <td>In house check Jan-23</td> </tr> </tbody> </table> <div style="display: flex; justify-content: space-between; margin-top: 10px;"> <div style="font-size: 8px;"> <p>Calibrated by: Dominique Shelten Function: Laboratory Technician</p> </div> <div style="font-size: 8px;"> <p>Approved by: Steven Kohn Technical Manager</p> </div> </div> <p style="font-size: 8px; margin-top: 5px;">This calibration certificate shall not be reproduced except in full without written approval of the laboratory.</p> <p style="font-size: 8px; text-align: right;">Issued: May 30, 2022</p> </div>	Primary Standards	ID #	Exp. Date (Certificate No.)	Scheduled Calibration	Kelway Multimeter Type 2001	SN: 0810276	31-Aug-21 (No:31368)	Aug-22	Secondary Standards	ID #	Check Date (in house)	Scheduled Check	Auto DAE Calibration Unit	SE LWS 003 AA 1001	24-Jan-22 (in house check)	In house check Jan-23	Calibrator class V0.1	SE LWS 100 AA 1002	24-Jan-22 (in house check)	In house check Jan-23
Primary Standards	ID #	Exp. Date (Certificate No.)	Scheduled Calibration																		
Kelway Multimeter Type 2001	SN: 0810276	31-Aug-21 (No:31368)	Aug-22																		
Secondary Standards	ID #	Check Date (in house)	Scheduled Check																		
Auto DAE Calibration Unit	SE LWS 003 AA 1001	24-Jan-22 (in house check)	In house check Jan-23																		
Calibrator class V0.1	SE LWS 100 AA 1002	24-Jan-22 (in house check)	In house check Jan-23																		
<div style="display: flex; justify-content: space-between; align-items: center;"> <div style="font-size: 8px;"> <p>Schmid & Partner Engineering AG Zugstrasse 10, 8001 Zurich, Switzerland Phone +41 44 248 9700, Fax +41 44 248 9770 www.sgs.com, info@sgs.com</p> </div> <div style="text-align: center; font-weight: bold; font-size: 12px;">s p e a g</div> <div style="font-size: 8px;"> <p>S Schweizerischer Kalibrierdienst C Service suisse d'étalonnage S Servizio svizzero di taratura S Swiss Calibration Service</p> </div> </div> <p style="font-size: 8px; margin-top: 5px;">Accredited by the Swiss Accreditation Service (SAS) The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates</p> <p style="text-align: right; font-size: 8px;">Accreditation No.: SCS 0108</p> <p>Glossary</p> <p>DAE: data acquisition electronics Connector angle: information used in DASY system to align probe sensor X to the robot coordinate system.</p> <p>Methods Applied and Interpretation of Parameters</p> <ul style="list-style-type: none"> DC Voltage Measurement: Calibration Factor assessed for use in DASY system by comparison with a calibrated instrument traceable to national standards. The figure given corresponds to the full scale range of the voltmeter in the respective range. Connector angle: The angle of the connector is assessed measuring the angle mechanically by a tool inserted. Uncertainty is not required. The following parameters as documented in the Appendix contain technical information as a result from the performance test and require no uncertainty. <ul style="list-style-type: none"> DC Voltage Measurement Linearity: Verification of the Linearity at +10% and -10% of the nominal calibration voltage. Influence of offset voltage is included in this measurement. Common mode sensitivity: Influence of a positive or negative common mode voltage on the differential measurement. Channel separation: Influence of a voltage on the neighbor channels not subject to an input voltage. AD Converter Values with inputs shorted: Values on the internal AD converter corresponding to zero input voltage Input Offset Measurement: Output voltage and statistical results over a large number of zero voltage measurements. Input Offset Current: Typical value for information; Maximum channel input offset current, not considering the input resistance. Input resistance: Typical value for information; DAE input resistance at the connector, during internal auto-zeroing and during measurement. Low Battery Alarm Voltage: Typical value for information. Below this voltage, a battery alarm signal is generated. Power consumption: Typical value for information. Supply currents in various operating modes. <p style="font-size: 8px; margin-top: 10px;">Certificate No: DAE4-1245_May22 Page 2 of 5</p>	<p>DC Voltage Measurement</p> <p>AD - Converter Resolution nominal High Range: 1LSB = 6 mV, full range = -100...+320 mV Low Range: 1LSB = 61 mV, full range = -1...+3mV DASY measurement parameters: Auto Zero-Time: 3 sec; Measuring time: 3 sec</p> <table border="1" style="width: 100%; font-size: 8px; border-collapse: collapse; margin-top: 5px;"> <thead> <tr> <th>Calibration Factors</th> <th>X</th> <th>Y</th> <th>Z</th> </tr> </thead> <tbody> <tr> <td>High Range</td> <td>405.265 ± 0.02% (k=2)</td> <td>403.974 ± 0.02% (k=2)</td> <td>406.092 ± 0.02% (k=2)</td> </tr> <tr> <td>Low Range</td> <td>3.99534 ± 1.50% (k=2)</td> <td>3.99508 ± 1.50% (k=2)</td> <td>4.01015 ± 1.50% (k=2)</td> </tr> </tbody> </table> <p>Connector Angle</p> <table border="1" style="width: 100%; font-size: 8px; border-collapse: collapse; margin-top: 5px;"> <tr> <td>Connector Angle to be used in DASY system</td> <td>30.0° ± 1°</td> </tr> </table> <p style="font-size: 8px; margin-top: 10px;">Certificate No: DAE4-1245_May22 Page 3 of 5</p>	Calibration Factors	X	Y	Z	High Range	405.265 ± 0.02% (k=2)	403.974 ± 0.02% (k=2)	406.092 ± 0.02% (k=2)	Low Range	3.99534 ± 1.50% (k=2)	3.99508 ± 1.50% (k=2)	4.01015 ± 1.50% (k=2)	Connector Angle to be used in DASY system	30.0° ± 1°						
Calibration Factors	X	Y	Z																		
High Range	405.265 ± 0.02% (k=2)	403.974 ± 0.02% (k=2)	406.092 ± 0.02% (k=2)																		
Low Range	3.99534 ± 1.50% (k=2)	3.99508 ± 1.50% (k=2)	4.01015 ± 1.50% (k=2)																		
Connector Angle to be used in DASY system	30.0° ± 1°																				



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing /inspection report & certificate, please contact us by telephone: (86-755) 8307 1443, or email: CN_Doccheck@sgs.com

No.10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300 t(86-512)57355888 f(86-512)57370818 www.sgsgroup.com.cn
 中国·江苏·昆山市留学生创业园伟业路10号 邮编 215300 t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com

Appendix (Additional assessments outside the scope of SCS0108)			
1. DC Voltage Linearity			
High Range			
Channel	Reading (µV)	Difference (µV)	Error (%)
Channel X + Input	19994.45	1.52	0.00
Channel X - Input	20004.58	2.22	0.01
Channel Y + Input	-20000.14	1.12	-0.01
Channel Y - Input	19994.72	1.58	0.00
Channel Z + Input	20001.22	-1.00	-0.00
Channel Z - Input	-20003.05	-1.57	0.01
Channel X + Input	19992.84	0.19	0.00
Channel X - Input	20003.09	0.58	0.00
Channel Y + Input	-20001.73	-0.27	0.00
Low Range			
Channel	Reading (µV)	Difference (µV)	Error (%)
Channel X + Input	2001.91	0.41	0.02
Channel X - Input	2002.54	0.65	0.32
Channel Y + Input	-197.86	0.07	-0.04
Channel Y - Input	2002.05	0.58	0.03
Channel Z + Input	201.27	-0.57	-0.28
Channel Z - Input	-199.23	-0.06	0.03
Channel X + Input	2001.98	0.08	0.00
Channel X - Input	2002.09	-1.53	-0.76
Channel Y + Input	-199.85	-1.57	0.79
2. Common mode sensitivity			
DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec			
Channel	Common mode Input Voltage (mV)	High Range Average Reading (µV)	Low Range Average Reading (µV)
Channel X	200	-8.87	-7.69
	-200	9.12	7.79
Channel Y	200	-8.68	-8.28
	-200	8.52	8.36
Channel Z	200	-5.36	-5.80
	-200	3.58	3.08
3. Channel separation			
DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec			
Channel X (µV)	Channel Y (µV)	Channel Z (µV)	
200	-	4.07	
200	9.36	-	
200	10.11	7.14	

4. AD-Converter Values with inputs shorted			
DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec			
Channel	High Range (LSB)	Low Range (LSB)	
Channel X	15984	17040	
Channel Y	16562	16768	
Channel Z	16035	15668	

5. Input Offset Measurement				
DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec				
Input 10MΩ				
Channel	Average (µV)	min. Offset (µV)	max. Offset (µV)	Std. Deviation (µV)
Channel X	1.00	-0.15	1.93	0.45
Channel Y	-0.18	-1.28	0.94	0.45
Channel Z	-0.58	-2.61	0.58	0.60

6. Input Offset Current		
Nominal input circuitry offset current on all channels: <25nA		
Channel X	200	200
Channel Y	200	200
Channel Z	200	200

7. Input Resistance (Typical values for information)		
Channel	Zeroing (ΩOhm)	Measuring (MΩhm)
Channel X	200	200
Channel Y	200	200
Channel Z	200	200

8. Low Battery Alarm Voltage (Typical values for information)	
Typical values	Alarm Level (VDC)
Supply (+ Vcc)	+7.9
Supply (- Vcc)	-7.6

9. Power Consumption (Typical values for information)			
Typical values	Switched off (mA)	Stand by (mA)	Transmitting (mA)
Supply (+ Vcc)	+0.01	+8	+14
Supply (- Vcc)	-0.01	-8	-9

3 EX3DV4 - SN 7767

<p>Calibration Laboratory of Schmid & Partner Engineering AG Zürcherstrasse 43, 8004 Zürich, Switzerland</p> <p>Accredited by the Swiss Accreditation Service (SAS) The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates</p> <p>Client: SGS-CN (Audien) Certificate No: EX-7767_Oct22</p> <p>CALIBRATION CERTIFICATE</p> <p>Object: EX3DV4 - SN:7767</p> <p>Calibration procedure(s): QA CAL-01-v9, QA CAL-12-v9, QA CAL-14-v6, QA CAL-23-v5, QA CAL-25-v7 Calibration procedure for dosimetric E-field probes</p> <p>Calibration date: October 28, 2022</p> <p>This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.</p> <p>All calibrations have been conducted in the closed laboratory facility: environment temperature (22 ± 3) °C and humidity < 70%.</p> <p>Calibration Equipment used (MATE critical for calibration)</p> <table border="1"> <thead> <tr> <th>Primary Standards</th> <th>ID</th> <th>Cal. Date (Certificate No.)</th> <th>Scheduled Calibration</th> </tr> </thead> <tbody> <tr> <td>Power meter NRP</td> <td>SN: 104778</td> <td>04-Apr-22 (No. 21 F20228-03564)</td> <td>Apr-23</td> </tr> <tr> <td>Power sensor NRP201</td> <td>SN: 103844</td> <td>04-Apr-22 (No. 21 F20224)</td> <td>Apr-23</td> </tr> <tr> <td>DCP DM-3-3 (Impedance)</td> <td>SN: 1048</td> <td>20-Sep-22 (DCP DM-3-3 (Imp. Oct22))</td> <td>Oct-23</td> </tr> <tr> <td>DCP DM-12</td> <td>SN: 1018</td> <td>20-Sep-22 (DCP DM-12 (Imp. Oct22))</td> <td>Oct-23</td> </tr> <tr> <td>Reference 30 dB Attenuator</td> <td>SN: 102582 (20)</td> <td>04-Apr-22 (No. 21 F20227)</td> <td>Apr-23</td> </tr> <tr> <td>DAE4</td> <td>SN: 981</td> <td>19-Oct-22 (No. DAE4-980 Oct22)</td> <td>Oct-23</td> </tr> <tr> <td>Reference Probe E53502</td> <td>SN: 3013</td> <td>27-Dec-21 (No. E53502-3013 Oct21)</td> <td>Dec-22</td> </tr> </tbody> </table> <p>Secondary Standards</p> <table border="1"> <thead> <tr> <th>Power meter E4418B</th> <th>SN: GB41936874</th> <th>05-Apr-18 (In house check Jun-20)</th> <th>Scheduled Check</th> </tr> </thead> <tbody> <tr> <td>Power sensor E4418A <td>SN: 10449889</td> <td>05-Apr-18 (In house check Jun-20)</td> <td>In house check Jun-24</td> </td></tr> <tr> <td>Power sensor E4418A <td>SN: 10017120</td> <td>05-Apr-18 (In house check Jun-20)</td> <td>In house check Jun-24</td> </td></tr> <tr> <td>RF generator HP 8548C <td>SN: US8548017193</td> <td>04-Aug-09 (In house check Jun-20)</td> <td>In house check Jun-24</td> </td></tr> <tr> <td>Network Analyzer E6656A <td>SN: GB41936877</td> <td>31-May-14 (In house check Oct-20)</td> <td>In house check Oct-24</td> </td></tr> </tbody> </table> <p>Calibrated by: Aldona Georgiadiu Laboratory Technician</p> <p>Approved by: Gven Kilm Technical Manager</p> <p>This calibration certificate shall not be reproduced except in full without written approval of the laboratory.</p> <p>Certificate No: EX-7767_Oct22 Page 1 of 9</p>	Primary Standards	ID	Cal. Date (Certificate No.)	Scheduled Calibration	Power meter NRP	SN: 104778	04-Apr-22 (No. 21 F20228-03564)	Apr-23	Power sensor NRP201	SN: 103844	04-Apr-22 (No. 21 F20224)	Apr-23	DCP DM-3-3 (Impedance)	SN: 1048	20-Sep-22 (DCP DM-3-3 (Imp. Oct22))	Oct-23	DCP DM-12	SN: 1018	20-Sep-22 (DCP DM-12 (Imp. Oct22))	Oct-23	Reference 30 dB Attenuator	SN: 102582 (20)	04-Apr-22 (No. 21 F20227)	Apr-23	DAE4	SN: 981	19-Oct-22 (No. DAE4-980 Oct22)	Oct-23	Reference Probe E53502	SN: 3013	27-Dec-21 (No. E53502-3013 Oct21)	Dec-22	Power meter E4418B	SN: GB41936874	05-Apr-18 (In house check Jun-20)	Scheduled Check	Power sensor E4418A <td>SN: 10449889</td> <td>05-Apr-18 (In house check Jun-20)</td> <td>In house check Jun-24</td>	SN: 10449889	05-Apr-18 (In house check Jun-20)	In house check Jun-24	Power sensor E4418A <td>SN: 10017120</td> <td>05-Apr-18 (In house check Jun-20)</td> <td>In house check Jun-24</td>	SN: 10017120	05-Apr-18 (In house check Jun-20)	In house check Jun-24	RF generator HP 8548C <td>SN: US8548017193</td> <td>04-Aug-09 (In house check Jun-20)</td> <td>In house check Jun-24</td>	SN: US8548017193	04-Aug-09 (In house check Jun-20)	In house check Jun-24	Network Analyzer E6656A <td>SN: GB41936877</td> <td>31-May-14 (In house check Oct-20)</td> <td>In house check Oct-24</td>	SN: GB41936877	31-May-14 (In house check Oct-20)	In house check Oct-24	<p>Calibration Laboratory of Schmid & Partner Engineering AG Zürcherstrasse 43, 8004 Zürich, Switzerland</p> <p>Accredited by the Swiss Accreditation Service (SAS) The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates</p> <p>Accreditation No: SCS 0108</p> <p>Glossary</p> <p>TSL: Issue simulating liquid NORM_{μ,z}: sensitivity in TSL; NORM_{μ,z}: diode compression point DCP: crest factor (10log₁₀ cycle) of the RF signal A, B, C, D: modulation dependent linearization parameters Polarization: or rotation around probe axis Polarization: if rotation around an axis that is in the plane normal to probe axis (at measurement center), i.e., θ = 0 is normal to probe axis Connector Angle: information used in DASY system to align probe sensor X to the robot coordinate system</p> <p>Calibration is Performed According to the Following Standards:</p> <p>a) IEC61028-1528: Measurement Procedure For The Assessment Of Specific Absorption Rate Of Human Exposure To Radio Frequency Fields From Hand-Held And Body-Worn Wireless Communication Devices - Part 1528: Human Models, Instrumentation And Procedures (Frequency Range of 4 MHz to 10 GHz), October 2020. b) IEC 60888-4: SAR Measurement Requirements for 100 MHz to 6 GHz</p> <p>Methods Applied and Interpretation of Parameters:</p> <ul style="list-style-type: none"> NORM_{μ,z}: Assessed for E-field polarization θ = 0 (f = 900 MHz in TEM-cell; f = 1800 MHz: R22 waveguide). NORM_{μ,z} are only intermediate values, i.e., the uncertainties of NORM_{μ,z} does not affect the E-field uncertainty inside TSL (see below CompF). NORM_{μ,z} = NORM_{μ,z} * frequency response (see Frequency Response Chart). This linearization is implemented in DASY4 software versions later than 4.2. The uncertainty of the frequency response is included in the stated uncertainty of CompF. DCP_{μ,z}: DCP are numerical linearization parameters assessed based on the data of power sweep with CW signal. DCP does not depend on frequency nor media. PAR: PAR is the Peak to Average Ratio that is not calibrated but determined based on the signal characteristics. A_{μ,z}, B_{μ,z}, C_{μ,z}, D_{μ,z}, VR_{μ,z}, A, B, C, D are numerical linearization parameters assessed based on the data of power sweep for specific modulation signals. The parameters do not depend on frequency nor media, VR is the maximum calibration range expressed in RMS voltage across the diode. CompF and Boundary Effect Parameters: Assessed in far field phantom using E-field (or Temperature Transfer Standard for f < 800 MHz) and media waveguide using analytical field distributions based on power measurements for f > 800 MHz. The same setups are used for assessment of the parameters applied for boundary compensation (alpha, depth) of which typical uncertainty values are given. These parameters are used in DASY4 software to improve probe accuracy close to the boundary. The sensitivity in TSL corresponds to NORM_{μ,z} * CompF whereby the uncertainty corresponds to that given for CompF. A frequency dependent CompF is used in DASY system 4.4 and higher which allows extending the validity from <math>±50 MHz</math> to <math>±100 MHz</math>. Spherical Antenna (SD) deviation from isotropy: In a field of low gradients realized using a flat phantom exposed by a patch antenna. Sensor Offset: The sensor offset corresponds to the offset of virtual measurement center from the probe tip (on probe axis). No tolerance required. Connector Angle: The angle is assessed using the information gained by determining the NORM_{μ,z} (no uncertainty required). <p>Certificate No: EX-7767_Oct22 Page 2 of 9</p>
Primary Standards	ID	Cal. Date (Certificate No.)	Scheduled Calibration																																																		
Power meter NRP	SN: 104778	04-Apr-22 (No. 21 F20228-03564)	Apr-23																																																		
Power sensor NRP201	SN: 103844	04-Apr-22 (No. 21 F20224)	Apr-23																																																		
DCP DM-3-3 (Impedance)	SN: 1048	20-Sep-22 (DCP DM-3-3 (Imp. Oct22))	Oct-23																																																		
DCP DM-12	SN: 1018	20-Sep-22 (DCP DM-12 (Imp. Oct22))	Oct-23																																																		
Reference 30 dB Attenuator	SN: 102582 (20)	04-Apr-22 (No. 21 F20227)	Apr-23																																																		
DAE4	SN: 981	19-Oct-22 (No. DAE4-980 Oct22)	Oct-23																																																		
Reference Probe E53502	SN: 3013	27-Dec-21 (No. E53502-3013 Oct21)	Dec-22																																																		
Power meter E4418B	SN: GB41936874	05-Apr-18 (In house check Jun-20)	Scheduled Check																																																		
Power sensor E4418A <td>SN: 10449889</td> <td>05-Apr-18 (In house check Jun-20)</td> <td>In house check Jun-24</td>	SN: 10449889	05-Apr-18 (In house check Jun-20)	In house check Jun-24																																																		
Power sensor E4418A <td>SN: 10017120</td> <td>05-Apr-18 (In house check Jun-20)</td> <td>In house check Jun-24</td>	SN: 10017120	05-Apr-18 (In house check Jun-20)	In house check Jun-24																																																		
RF generator HP 8548C <td>SN: US8548017193</td> <td>04-Aug-09 (In house check Jun-20)</td> <td>In house check Jun-24</td>	SN: US8548017193	04-Aug-09 (In house check Jun-20)	In house check Jun-24																																																		
Network Analyzer E6656A <td>SN: GB41936877</td> <td>31-May-14 (In house check Oct-20)</td> <td>In house check Oct-24</td>	SN: GB41936877	31-May-14 (In house check Oct-20)	In house check Oct-24																																																		



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic formal documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com


No.10, Weyue Road, Innovation Park, Kunshan, Jiangsu, China 215300
中国·江苏·昆山市留学生创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgsgroup.com.cn
t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com

<p>EX3DV4 - SN:7767 October 28, 2022</p> <p>Parameters of Probe: EX3DV4 - SN:7767</p> <p>Basic Calibration Parameters</p> <table border="1"><thead><tr><th></th><th>Sensor X</th><th>Sensor Y</th><th>Sensor Z</th><th>Unc (k = 2)</th></tr></thead><tbody><tr><td>Norm. (μV/(mm²)^{1/2})^A</td><td>0.67</td><td>0.89</td><td>0.56</td><td>±10.1%</td></tr><tr><td>DCP (mV)^B</td><td>103.4</td><td>107.3</td><td>105.7</td><td>±4.7%</td></tr></tbody></table> <p>Calibration Results for Modulation Response</p> <table border="1"><thead><tr><th>URD</th><th>Communication System Name</th><th>A dB</th><th>B dB/μV</th><th>C</th><th>D dB</th><th>VR mV</th><th>Max dev.</th><th>Max Unc^C k=2</th></tr></thead><tbody><tr><td>0</td><td>CW</td><td>X: 0.00 Y: 0.00 Z: 0.00</td><td>0.00 0.00 0.00</td><td>1.00 1.00 1.00</td><td>0.00 184.7 179.3</td><td></td><td>±3.5% ±4.7%</td><td></td></tr></tbody></table> <p>The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.</p> <p>^A The uncertainties of Norm X,Y,Z do not affect the E₁-field uncertainty inside TSS. (see Page 5). ^B Uncertainty parameter uncertainty for maximum specified lead strength. ^C Uncertainty is determined using the max. deviation from three response coupling rectangular distribution and is expressed for the square of the field value.</p> <p>Certificate No: EX-7767_Oct22 Page 5 of 9</p>		Sensor X	Sensor Y	Sensor Z	Unc (k = 2)	Norm. (μV/(mm ²) ^{1/2}) ^A	0.67	0.89	0.56	±10.1%	DCP (mV) ^B	103.4	107.3	105.7	±4.7%	URD	Communication System Name	A dB	B dB/μV	C	D dB	VR mV	Max dev.	Max Unc ^C k=2	0	CW	X: 0.00 Y: 0.00 Z: 0.00	0.00 0.00 0.00	1.00 1.00 1.00	0.00 184.7 179.3		±3.5% ±4.7%		<p>EX3DV4 - SN:7767 October 28, 2022</p> <p>Parameters of Probe: EX3DV4 - SN:7767</p> <p>Other Probe Parameters</p> <table border="1"><thead><tr><th>Parameter</th><th>Value</th></tr></thead><tbody><tr><td>Sensor Arrangement</td><td>Triangular</td></tr><tr><td>Connector Angle</td><td>144.9°</td></tr><tr><td>Mechanical Surface Detection Mode</td><td>enabled</td></tr><tr><td>Optical Surface Detection Mode</td><td>disabled</td></tr><tr><td>Probe Overall Length</td><td>337mm</td></tr><tr><td>Probe Body Diameter</td><td>10mm</td></tr><tr><td>Tip Length</td><td>2mm</td></tr><tr><td>Tip Diameter</td><td>2.5mm</td></tr><tr><td>Probe Tip to Sensor X Calibration Point</td><td>1mm</td></tr><tr><td>Probe Tip to Sensor Y Calibration Point</td><td>1mm</td></tr><tr><td>Probe Tip to Sensor Z Calibration Point</td><td>1mm</td></tr><tr><td>Recommended Measurement Distance from Surface</td><td>1.4mm</td></tr></tbody></table> <p>Note: Measurement distance from surface can be increased to 3-4mm for an Area Scan job.</p> <p>Certificate No: EX-7767_Oct22 Page 4 of 9</p>	Parameter	Value	Sensor Arrangement	Triangular	Connector Angle	144.9°	Mechanical Surface Detection Mode	enabled	Optical Surface Detection Mode	disabled	Probe Overall Length	337mm	Probe Body Diameter	10mm	Tip Length	2mm	Tip Diameter	2.5mm	Probe Tip to Sensor X Calibration Point	1mm	Probe Tip to Sensor Y Calibration Point	1mm	Probe Tip to Sensor Z Calibration Point	1mm	Recommended Measurement Distance from Surface	1.4mm																																																																																																																																																																															
	Sensor X	Sensor Y	Sensor Z	Unc (k = 2)																																																																																																																																																																																																																																							
Norm. (μV/(mm ²) ^{1/2}) ^A	0.67	0.89	0.56	±10.1%																																																																																																																																																																																																																																							
DCP (mV) ^B	103.4	107.3	105.7	±4.7%																																																																																																																																																																																																																																							
URD	Communication System Name	A dB	B dB/μV	C	D dB	VR mV	Max dev.	Max Unc ^C k=2																																																																																																																																																																																																																																			
0	CW	X: 0.00 Y: 0.00 Z: 0.00	0.00 0.00 0.00	1.00 1.00 1.00	0.00 184.7 179.3		±3.5% ±4.7%																																																																																																																																																																																																																																				
Parameter	Value																																																																																																																																																																																																																																										
Sensor Arrangement	Triangular																																																																																																																																																																																																																																										
Connector Angle	144.9°																																																																																																																																																																																																																																										
Mechanical Surface Detection Mode	enabled																																																																																																																																																																																																																																										
Optical Surface Detection Mode	disabled																																																																																																																																																																																																																																										
Probe Overall Length	337mm																																																																																																																																																																																																																																										
Probe Body Diameter	10mm																																																																																																																																																																																																																																										
Tip Length	2mm																																																																																																																																																																																																																																										
Tip Diameter	2.5mm																																																																																																																																																																																																																																										
Probe Tip to Sensor X Calibration Point	1mm																																																																																																																																																																																																																																										
Probe Tip to Sensor Y Calibration Point	1mm																																																																																																																																																																																																																																										
Probe Tip to Sensor Z Calibration Point	1mm																																																																																																																																																																																																																																										
Recommended Measurement Distance from Surface	1.4mm																																																																																																																																																																																																																																										
<p>EX3DV4 - SN:7767 October 28, 2022</p> <p>Parameters of Probe: EX3DV4 - SN:7767</p> <p>Calibration Parameter Determined in Head Tissue Simulating Media</p> <table border="1"><thead><tr><th>f (MHz)^C</th><th>Relative Permittivity^D</th><th>Conductivity^E (S/m)</th><th>CompF X</th><th>CompF Y</th><th>CompF Z</th><th>Alpha^F</th><th>Depth^G (mm)</th><th>Unc (k = 2)</th></tr></thead><tbody><tr><td>150</td><td>50.3</td><td>0.79</td><td>14.08</td><td>14.08</td><td>14.08</td><td>0.00</td><td>1.00</td><td>±13.2%</td></tr><tr><td>450</td><td>42.5</td><td>0.87</td><td>11.50</td><td>11.50</td><td>11.50</td><td>0.16</td><td>1.20</td><td>±13.3%</td></tr><tr><td>750</td><td>41.8</td><td>0.88</td><td>10.26</td><td>10.26</td><td>10.26</td><td>0.30</td><td>0.80</td><td>±12.0%</td></tr><tr><td>835</td><td>41.5</td><td>0.90</td><td>10.00</td><td>10.00</td><td>10.00</td><td>0.43</td><td>0.80</td><td>±12.0%</td></tr><tr><td>1750</td><td>40.1</td><td>1.37</td><td>9.32</td><td>9.32</td><td>9.32</td><td>0.36</td><td>0.86</td><td>±12.0%</td></tr><tr><td>1900</td><td>40.0</td><td>1.40</td><td>8.91</td><td>8.91</td><td>8.91</td><td>0.33</td><td>0.86</td><td>±12.0%</td></tr><tr><td>2100</td><td>39.8</td><td>1.49</td><td>8.60</td><td>8.60</td><td>8.60</td><td>0.30</td><td>0.86</td><td>±12.0%</td></tr><tr><td>2200</td><td>39.5</td><td>1.67</td><td>8.44</td><td>8.44</td><td>8.44</td><td>0.33</td><td>0.80</td><td>±12.0%</td></tr><tr><td>2450</td><td>39.2</td><td>1.80</td><td>8.24</td><td>8.24</td><td>8.24</td><td>0.32</td><td>0.90</td><td>±12.0%</td></tr><tr><td>2600</td><td>39.0</td><td>1.96</td><td>7.99</td><td>7.99</td><td>7.99</td><td>0.27</td><td>0.90</td><td>±12.0%</td></tr><tr><td>3300</td><td>38.2</td><td>2.71</td><td>7.55</td><td>7.55</td><td>7.55</td><td>0.30</td><td>1.25</td><td>±13.1%</td></tr><tr><td>3600</td><td>37.9</td><td>2.91</td><td>7.45</td><td>7.45</td><td>7.45</td><td>0.30</td><td>1.35</td><td>±13.1%</td></tr><tr><td>3700</td><td>37.7</td><td>3.12</td><td>7.20</td><td>7.20</td><td>7.20</td><td>0.30</td><td>1.05</td><td>±13.1%</td></tr><tr><td>3900</td><td>37.5</td><td>3.32</td><td>6.84</td><td>6.84</td><td>6.84</td><td>0.40</td><td>1.06</td><td>±13.1%</td></tr><tr><td>4100</td><td>37.2</td><td>3.53</td><td>6.63</td><td>6.63</td><td>6.63</td><td>0.40</td><td>1.00</td><td>±13.1%</td></tr><tr><td>4300</td><td>37.1</td><td>3.60</td><td>6.30</td><td>6.30</td><td>6.30</td><td>0.40</td><td>1.70</td><td>±13.1%</td></tr><tr><td>4400</td><td>36.9</td><td>3.84</td><td>6.17</td><td>6.17</td><td>6.17</td><td>0.40</td><td>1.70</td><td>±13.1%</td></tr><tr><td>4600</td><td>36.7</td><td>4.04</td><td>6.15</td><td>6.15</td><td>6.15</td><td>0.40</td><td>1.70</td><td>±13.1%</td></tr><tr><td>4800</td><td>36.4</td><td>4.25</td><td>6.13</td><td>6.13</td><td>6.13</td><td>0.40</td><td>1.90</td><td>±13.1%</td></tr><tr><td>4900</td><td>36.3</td><td>4.40</td><td>6.07</td><td>6.07</td><td>6.07</td><td>0.40</td><td>1.80</td><td>±13.1%</td></tr><tr><td>5000</td><td>36.0</td><td>4.66</td><td>5.65</td><td>5.65</td><td>5.65</td><td>0.40</td><td>1.80</td><td>±13.1%</td></tr><tr><td>5300</td><td>35.9</td><td>4.75</td><td>5.48</td><td>5.48</td><td>5.48</td><td>0.40</td><td>1.80</td><td>±13.1%</td></tr><tr><td>5900</td><td>35.6</td><td>4.99</td><td>5.30</td><td>5.30</td><td>5.30</td><td>0.40</td><td>1.80</td><td>±13.1%</td></tr><tr><td>5900</td><td>35.5</td><td>5.07</td><td>5.14</td><td>5.14</td><td>5.14</td><td>0.40</td><td>1.80</td><td>±13.1%</td></tr><tr><td>5900</td><td>35.3</td><td>5.27</td><td>5.10</td><td>5.10</td><td>5.10</td><td>0.40</td><td>1.80</td><td>±13.1%</td></tr></tbody></table> <p>^C Frequency validity above 300 MHz at ±100 MHz only applies for DCP v1.4 and higher (see Page 5), else it is limited to ±50 MHz. The uncertainty is the RMS of the CompF uncertainty of calibration frequency and the uncertainty for the relative permittivity. Frequency validity below 300 MHz is ±1%, 2%, 4%, 6%, and 7% for CompF assessments at 30, 60, 120, 150, and 300 MHz, respectively. Validity of CompF assessed at 60 MHz is ±4.66%, and CompF assessed at 150 MHz is ±0.16%. Asses at 60 MHz validity can be extended to ±100 MHz. ^D At frequencies below 3 GHz, the validity of tissue parameters (α and ρ) can be related to a 10% liquid composition formula is applied to measured S10 values. At frequencies above 3 GHz, the validity of tissue parameters (α and ρ) is not related to 10%. The uncertainty is the RMS of the CompF uncertainty for individual target tissue parameters. ^E Alpha values are determined using calibration. SRRAD warns that the remaining residual due to the uncertainty when after composition is applied has less than ±1% for frequencies below 3 GHz and below ±2% for frequencies between 3-6 GHz at any distance larger than half the probe tip diameter from the targets. ^F Alpha values are determined using calibration. SRRAD warns that the remaining residual due to the uncertainty when after composition is applied has less than ±1% for frequencies below 3 GHz and below ±2% for frequencies between 3-6 GHz at any distance larger than half the probe tip diameter from the targets. ^G Alpha values are determined using calibration. SRRAD warns that the remaining residual due to the uncertainty when after composition is applied has less than ±1% for frequencies below 3 GHz and below ±2% for frequencies between 3-6 GHz at any distance larger than half the probe tip diameter from the targets.</p> <p>Certificate No: EX-7767_Oct22 Page 5 of 9</p>	f (MHz) ^C	Relative Permittivity ^D	Conductivity ^E (S/m)	CompF X	CompF Y	CompF Z	Alpha ^F	Depth ^G (mm)	Unc (k = 2)	150	50.3	0.79	14.08	14.08	14.08	0.00	1.00	±13.2%	450	42.5	0.87	11.50	11.50	11.50	0.16	1.20	±13.3%	750	41.8	0.88	10.26	10.26	10.26	0.30	0.80	±12.0%	835	41.5	0.90	10.00	10.00	10.00	0.43	0.80	±12.0%	1750	40.1	1.37	9.32	9.32	9.32	0.36	0.86	±12.0%	1900	40.0	1.40	8.91	8.91	8.91	0.33	0.86	±12.0%	2100	39.8	1.49	8.60	8.60	8.60	0.30	0.86	±12.0%	2200	39.5	1.67	8.44	8.44	8.44	0.33	0.80	±12.0%	2450	39.2	1.80	8.24	8.24	8.24	0.32	0.90	±12.0%	2600	39.0	1.96	7.99	7.99	7.99	0.27	0.90	±12.0%	3300	38.2	2.71	7.55	7.55	7.55	0.30	1.25	±13.1%	3600	37.9	2.91	7.45	7.45	7.45	0.30	1.35	±13.1%	3700	37.7	3.12	7.20	7.20	7.20	0.30	1.05	±13.1%	3900	37.5	3.32	6.84	6.84	6.84	0.40	1.06	±13.1%	4100	37.2	3.53	6.63	6.63	6.63	0.40	1.00	±13.1%	4300	37.1	3.60	6.30	6.30	6.30	0.40	1.70	±13.1%	4400	36.9	3.84	6.17	6.17	6.17	0.40	1.70	±13.1%	4600	36.7	4.04	6.15	6.15	6.15	0.40	1.70	±13.1%	4800	36.4	4.25	6.13	6.13	6.13	0.40	1.90	±13.1%	4900	36.3	4.40	6.07	6.07	6.07	0.40	1.80	±13.1%	5000	36.0	4.66	5.65	5.65	5.65	0.40	1.80	±13.1%	5300	35.9	4.75	5.48	5.48	5.48	0.40	1.80	±13.1%	5900	35.6	4.99	5.30	5.30	5.30	0.40	1.80	±13.1%	5900	35.5	5.07	5.14	5.14	5.14	0.40	1.80	±13.1%	5900	35.3	5.27	5.10	5.10	5.10	0.40	1.80	±13.1%	<p>EX3DV4 - SN:7767 October 28, 2022</p> <p>Frequency Response of E-Field (TEM-Cell#119 EXL, Waveguide#22)</p> <p>Certificate No: EX-7767_Oct22 Page 6 of 9</p>
f (MHz) ^C	Relative Permittivity ^D	Conductivity ^E (S/m)	CompF X	CompF Y	CompF Z	Alpha ^F	Depth ^G (mm)	Unc (k = 2)																																																																																																																																																																																																																																			
150	50.3	0.79	14.08	14.08	14.08	0.00	1.00	±13.2%																																																																																																																																																																																																																																			
450	42.5	0.87	11.50	11.50	11.50	0.16	1.20	±13.3%																																																																																																																																																																																																																																			
750	41.8	0.88	10.26	10.26	10.26	0.30	0.80	±12.0%																																																																																																																																																																																																																																			
835	41.5	0.90	10.00	10.00	10.00	0.43	0.80	±12.0%																																																																																																																																																																																																																																			
1750	40.1	1.37	9.32	9.32	9.32	0.36	0.86	±12.0%																																																																																																																																																																																																																																			
1900	40.0	1.40	8.91	8.91	8.91	0.33	0.86	±12.0%																																																																																																																																																																																																																																			
2100	39.8	1.49	8.60	8.60	8.60	0.30	0.86	±12.0%																																																																																																																																																																																																																																			
2200	39.5	1.67	8.44	8.44	8.44	0.33	0.80	±12.0%																																																																																																																																																																																																																																			
2450	39.2	1.80	8.24	8.24	8.24	0.32	0.90	±12.0%																																																																																																																																																																																																																																			
2600	39.0	1.96	7.99	7.99	7.99	0.27	0.90	±12.0%																																																																																																																																																																																																																																			
3300	38.2	2.71	7.55	7.55	7.55	0.30	1.25	±13.1%																																																																																																																																																																																																																																			
3600	37.9	2.91	7.45	7.45	7.45	0.30	1.35	±13.1%																																																																																																																																																																																																																																			
3700	37.7	3.12	7.20	7.20	7.20	0.30	1.05	±13.1%																																																																																																																																																																																																																																			
3900	37.5	3.32	6.84	6.84	6.84	0.40	1.06	±13.1%																																																																																																																																																																																																																																			
4100	37.2	3.53	6.63	6.63	6.63	0.40	1.00	±13.1%																																																																																																																																																																																																																																			
4300	37.1	3.60	6.30	6.30	6.30	0.40	1.70	±13.1%																																																																																																																																																																																																																																			
4400	36.9	3.84	6.17	6.17	6.17	0.40	1.70	±13.1%																																																																																																																																																																																																																																			
4600	36.7	4.04	6.15	6.15	6.15	0.40	1.70	±13.1%																																																																																																																																																																																																																																			
4800	36.4	4.25	6.13	6.13	6.13	0.40	1.90	±13.1%																																																																																																																																																																																																																																			
4900	36.3	4.40	6.07	6.07	6.07	0.40	1.80	±13.1%																																																																																																																																																																																																																																			
5000	36.0	4.66	5.65	5.65	5.65	0.40	1.80	±13.1%																																																																																																																																																																																																																																			
5300	35.9	4.75	5.48	5.48	5.48	0.40	1.80	±13.1%																																																																																																																																																																																																																																			
5900	35.6	4.99	5.30	5.30	5.30	0.40	1.80	±13.1%																																																																																																																																																																																																																																			
5900	35.5	5.07	5.14	5.14	5.14	0.40	1.80	±13.1%																																																																																																																																																																																																																																			
5900	35.3	5.27	5.10	5.10	5.10	0.40	1.80	±13.1%																																																																																																																																																																																																																																			



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx>; and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.
Attention: To check the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com
No.10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300
中国·江苏·昆山市留学院创业园伟业路10号 邮编 215300
t(86-512)57355888 f(86-512)57370818 www.sgsgroup.com.cn
t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.
Attention: To check the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN_Doccheck@sgs.com
 No.10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300 t(86-512)57355888 f(86-512)57370818 www.sgs.com.cn
 中国·江苏·昆山市留学院创业园伟业路10号 邮编 215300 t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com

4 Impedance and return loss

Dipole CLA150 SN 4025				
Head Liquid				
Date of Measurement	Return Loss(dB)	Δ %	Impedance (Ω)	$\Delta\Omega$
2021/4/26	-31.4	/	47.8	/
2022/4/26	-32.5	-3.5%	47.1	-0.7
Dipole D450V3 SN 1103				
Head Liquid				
Date of Measurement	Return Loss(dB)	Δ %	Impedance (Ω)	$\Delta\Omega$
2021/4/21	-23	/	57.1	/
2022/4/26	-23.4	-1.74%	56.6	-0.5



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Document.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN_Doccheck@sgs.com
 No.10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300 t(86-512)57355888 f(86-512)57370818 www.sgs.com.cn
 中国·江苏·昆山市留学生创业园伟业路10号 邮编 215300 t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com