

Vscan Air Wireless Charger User Manual

Overview

The Vscan Air Wireless Charger is specifically designed to charge the Vscan Air probe.

Signal Indicator

| Indicator | Status |
|-------------------------------------|--------------------------|
| Solid BLUE for 3 seconds, then off. | Power on, not charging: |
| Blink green | Charging |
| Orange | Foreign metals detected: |
| Solid BLUE | Completed charging |

Cautions

1. Do not place any foreign materials, such as metal objects, magnets, and magnetic stripe cards between the Vscan Air probe and the charger.
2. Use this charger in an ambient temperature between 0°C ~ 40°C.
3. Do not disassemble the charger.
4. Use original or certified cable and charger for safe charging.
5. Keep a distance of at least 20 cm (7.87 inches) between implanted medical devices (pacemakers, cochlear implants, etc.) and this charger, to avoid potential interference with the medical devices.

Product Information

Product name: Vscan Air Wireless Charger

Model: GP310000 / GP320000

Authorized by: GE HealthCare India

Address: Odyssey Building 3rd Floor Wing A, EPIP 122 (Phase II), Whitefield Road John F. Welch Technology Center, Hoodi Village, Bengaluru, State-Karnataka, India. Pincode-560066

Manufacturer: BizLink (Kunshan) Co., Ltd.

Specifications

| | |
|--------|--|
| Input | 5V = 2.1A |
| Output | 5W |
| Size | 179.44 x 83.56 x 65.4 mm / 7.06 x 3.29 x 2.57 inch |
| Weight | 180 g / 6.35 oz |

Service life

5 years

FCC Statement

This equipment has been tested and found to comply with the limits for a Class B digital device, pursuant to Part 15 of the FCC Rules. These limits are designed to provide reasonable protection against harmful interference in a residential installation. This equipment generates, uses and can radiate radio frequency energy and, if not installed and used in accordance with the instructions, may cause harmful interference to radio communications. However, there is no guarantee that interference will not occur in a particular installation. If this equipment does cause harmful interference to radio or television reception, which can be determined by turning the equipment off and on, the user is encouraged to try to correct the interference by one or more of the following measures:

- Reorient or relocate the receiving antenna.
- Increase the separation between the equipment and receiver.
- Connect the equipment into an outlet on a circuit different from that to which the receiver is connected.
- Consult the dealer or an experienced radio/TV technician for help.

FCC Caution: Any changes or modifications not expressly approved by the party responsible for compliance could void the user's authority to operate this equipment.

This device complies with Part 15 of the FCC Rules. Operation is subject to the following two conditions: (1) This device may not cause harmful interference, and (2) this device must accept any interference received, including interference that may cause undesired operation.

This device and its antenna(s) must not be co-located or operating in conjunction with any other antenna or transmitter.

FCC Radiation Exposure Statement

This equipment complies with FCC radiation exposure limits set forth for an uncontrolled environment. This equipment should be installed and operated with minimum distance 20cm between the radiator & your body.

For Canada

This device contains licence-exempt transmitter(s)/receiver(s) that comply with Innovation, Science and Economic Development Canada's licence-exempt RSS(s). Operation is subject to the following two conditions:

1. This device may not cause interference.
2. This device must accept any interference, including interference that may cause undesired operation of the device.

L'émetteur/récepteur exempt de licence contenu dans le présent appareil est conforme aux CNR d'Innovation, Sciences et Développement économique Canada applicables aux appareils radio exempts de licence. L'exploitation est autorisée aux deux conditions suivantes :

1. L'appareil ne doit pas produire de brouillage;
2. L'appareil doit accepter tout brouillage radioélectrique subi, même si le brouillage est susceptible d'en compromettre le fonctionnement.

IC Radiation Exposure Statement:

This equipment complies with IC RSS-102 radiation exposure limits set forth for an uncontrolled environment. This equipment should be installed and operated with minimum distance 20cm between the radiator & your body.

Cet équipement est conforme aux limites d'exposition aux rayonnements IC établies pour un environnement non contrôlé. Cet équipement doit être installé et utilisé avec un minimum de 20cm de distance entre la source de rayonnement et votre corps.

For EU/UK

Safety statement: The minimum distance between the user and/or any bystander and the radiating structure of the transmitter is 20cm.

The following table provides information on the frequency bands used and the maximum RF transmit power of the product for sale in the EU and UK:

| Frequency range | Max. Transmit Power |
|-----------------|---------------------|
| 112-148.5 kHz | 5.44 dBuA/m |



Letter of Authorization

Date: March 30, 2023

Dear Sir:

I hereby have entrusted the following person to be a proxy regarding application for EU-type examination certificate and UKCA approval. I am therefore responsible for the contents of the application.

Product Description: Vscan Air Wireless Charger

Model type: GP310000, GP320000

Brand Name: GE HealthCare

Company name : Sporton International Inc.

Company Address : No.52, Huaya 1st Rd., Guishan Dist., Taoyuan City 333, Taiwan
(R.O.C.)

Name : Leo Huang

Job Title : VP

Sincerely,

Applicant's company name : BizLink (Kunshan) Co., Ltd.

Applicant's company address : No.168, Nanhe Rd., Kunshan Economic & Technology
Development Zone, Kunshan City, Jiangsu 215300, China

Signature : **Tony Duan**
数字签名者 : Tony Duan
日期 : 2023.03.30 11:39:17
+08'00'

Name and Job Title : Tony Duan, Manager

Tel : +86 512 57630855

E-mail : tony_duan@bizlinktech.com

EU Declaration of Conformity

We,

Name of Manufacturer: BizLink (Kunshan) Co., Ltd.

Address: No.168, Nanhe Rd., Kunshan Economic & Technology Development Zone,
Kunshan City, Jiangsu 215300, China

Telephone number: +86 512 57630855

hereby, declare under our sole responsibility that the requirements set out in the **Directive 2014/53/EU** has been fully fulfilled on our product with indication below:


Product Name: Vscan Air Wireless Charger

Model Number: GP310000, GP320000

Object of the Declaration:



GP310000



GP320000

The object of the declaration described above is in conformity with the relevant Union harmonization legislation:

Radio Equipment Directive (RED) 2014/53/EU

Restriction of Hazardous Substances Directive (RoHS) 2011/65/EU

Waste Electrical and Electronic Equipment Directive (WEEE) 2012/19/EU

The following standards and technical specifications have been applied:

Article 3.2 & 3.3 : EN 303 417 V1.1.1

**Article 3.1(b) : EN 301 489-1 V2.2.3, EN 301 489-3 V2.2.0, EN 55032:2015/A1:2020, EN 55032:2015/A11:2020,
EN 55035:2017/A11:2020, EN IEC 61000-3-2:2019, EN 61000-3-3:2013+A1:2019**

Article 3.1(a) : EN IEC 62311:2020, EN 50665:2017

Article 3.1(a) : IEC 62368-1:2018, EN IEC 62368-1:2020+A11:2020

Notified Body: Timco Engineering, Inc. 1177

EU-type examination certificate: with Module B+C

Additional Information:

Software: v1.04

Accessories: Molex Cable (USB Type-C to Molex), USB Cable (USB Type-C to USB Type-A)

Signed for and on behalf of:

Signature : **Tony Duan** 数字签名者 : Tony Duan
日期 : 2023.03.30
11:39:38 +08'00'

March 30, 2023

Name, Function(Title) : Tony Duan, Manager



E-mail : tony_duan@bizlinktech.com

Declaration of Conformity

Declaration of conformity

No.: (optional)

Name of Manufacturer: BizLink (Kunshan) Co., Ltd.

Address: No.168, Nanhe Rd., Kunshan Economic & Technology Development Zone,
Kunshan City, Jiangsu 215300, China


Product Name: Vscan Air Wireless Charger

Model Number: GP310000, GP320000

Object of the Declaration:



GP310000



GP320000

We, BizLink (Kunshan) Co., Ltd., declare under our sole responsibility that the above product is in conformity with the relevant statutory requirements: **RER 2017 (SI 2017/1206)**

The following designated standards or technical specifications were used in relation to which conformity is declared:

Regulation 6.2 : EN 303 417 V1.1.1

Regulation 6.1b : EN 301 489-1 V2.2.3, EN 301 489-3 V2.2.0, EN 55032:2015/A1:2020, EN 55032:2015/A11:2020, EN 55035:2017/A11:2020, EN IEC 61000-3-2:2019, EN 61000-3-3:2013+A1:2019

Regulation 6.1a : EN IEC 62311:2020, EN 50665:2017

Regulation 6.1a : IEC 62368-1:2018, EN IEC 62368-1:2020+A11:2020

UK Approved Body:

TIMCO Engineering, Inc. with number AB1177 has performed a conformity assessment in accordance with the procedures in Schedule 3 Module against the essential requirements Regulation 6.1a, 6.1b, and 6.2 of the RER 2017 (SI 2017/1206), and issued the Type examination certificate:

Description of accessories and components, including software, which allow the radio equipment to operate as intended and covered by the Declaration of conformity

Software: v1.04

Accessories: Molex Cable (USB Type-C to Molex), USB Cable (USB Type-C to USB Type-A)

Signed for and on behalf of:

Signature : **Tony Duan**
数字签名者 : Tony Duan
日期 : 2023.03.30 11:40:05
+08'00'



March 30, 2023

Name, Function(Title) : Tony Duan, Manager

E-mail : tony_duan@bizlinktech.com