



Wipro GE Healthcare Pvt. Ltd.

John F Welch Technology Centre
Odyssey Building, 122, EPIP, Phase II
Whitefield, Bangalore 560 066
India
T 91 80 4088 1000
F 91 80 2841 1645
www.wiproge.com

GE Healthcare
John F Welch Technology Center
Odyssey, #122, EPIP Phase II, Whitefield,
Bangalore, India

Date: 25th Nov 2020

Federal Communications Commission
Authorization and Evaluation division

Subject: Frequency plan in UNII Radio Band for the product **Vscan Air CL** with **FCC ID: YOM-VSCANAIR**

We declare the frequency plan of the product **Vscan Air CL** is as below:

Mode of Operation	Low Frequency (MHz)	High Frequency (MHz)
Non-DFS	5150	5250
Non-DFS	5725	5825
DFS (Client without radar detection)	Not Supported	Not Supported

Wireless module(WL1837MOD from Texas Instruments) used in the product support DFS bands. However, the DFS bands are disabled in module firmware and the end user will not have access to enable the transmission on DFS bands.

Sincerely,

Gunaseelan Murugan
Sr. Systems Engineer
GE Healthcare
John F Welch Technology Center
Odyssey, #122, EPIP Phase II, Whitefield, Bangalore, India
Phone: +91 80 4088 1405
E-Mail ID: guna.m@ge.com

Thank you for partnering with us on our mission of being at work for a healthier India

Wipro GE Healthcare Pvt. Ltd. Regd. Off : No. 4, Kadugodi Industrial Area, Bangalore 560 067, Karnataka India
CIN : U33111KA1990PTC016063 T 91 80 4180 1000 F 91 80 4180 1290