

Test report No. : 30DE0169-HO-01-A-R1  
Page : 44 of 82  
FCC ID : YR7AERODRP1  
Issued date : September 16, 2010  
Revised date : October 26, 2010

---

### APPENDIX 3 : Test instruments

---

**UL Japan, Inc.**  
**Head Office EMC Lab.**  
4383-326 Asama-cho, Ise-shi, Mie-ken 516-0021 JAPAN  
Telephone : +81 596 24 8116  
Facsimile : +81 596 24 8124

## 1. Equipment used

| Control No.    | Instrument                   | Manufacturer                  | Model No            | Serial No  | Test Item         | Calibration Date Interval(month) * |
|----------------|------------------------------|-------------------------------|---------------------|------------|-------------------|------------------------------------|
| MAT-22         | Attenuator(10dB) DC-18GHz    | Orient Microwave              | BX10-0476-00        | -          | Power Measurement | 2010/03/01 * 12                    |
| MPM-09         | Power Meter                  | Anritsu                       | ML2495A             | 6K00003348 | Power Measurement | 2009/09/09 * 12                    |
| MPSE-12        | Power sensor                 | Anritsu                       | MA2411B             | 011598     | Power Measurement | 2009/09/09 * 12                    |
| MPM-01         | Power Meter                  | Agilent                       | E4417A              | GB41290639 | SAR               | 2010/02/16 * 12                    |
| MPSE-01        | Power Sensor                 | Agilent                       | E9300B              | US40010300 | SAR               | 2010/02/11 * 12                    |
| MPSE-03        | Power sensor                 | Agilent                       | E9327A              | US40440576 | SAR               | 2010/02/19 * 12                    |
| MAT-15         | Attenuator(30dB)             | Agilent                       | 8498A               | US40010300 | SAR               | 2010/02/24 * 12                    |
| MSG-10         | Signal Generator             | Agilent                       | N5181A              | MY47421098 | SAR               | 2009/06/30 * 24                    |
| MRFA-08        | Pre Amplifier                | TSJ                           | TCBP0206            | -          | SAR               | 2010/03/19 * 12                    |
| MHDC-12        | Dual Directional Coupler     | Hewlett Packard               | 772D                | 2839A0016  | SAR               | Pre Check                          |
| MNA-01         | Network Analyzer             | Agilent/HP                    | E8358A              | US41080381 | SAR               | 2009/08/28 * 12                    |
| MDPK-01        | Dielectric probe kit         | Agilent                       | 85070D              | -          | SAR               | Pre Check                          |
| MNCK-01        | Type N Calibration Kit       | Agilent                       | 85032F              | MY41495257 | SAR               | 2009/08/28 * 12                    |
| MPB-03         | Dosimetric E-Field Probe     | Schmid&Partner Engineering AG | EX3DV3              | 3507       | SAR               | 2010/02/19 * 12                    |
| MDAE-01        | Data Acquisition Electronics | Schmid&Partner Engineering AG | DAE3 V1             | 509        | SAR               | 2009/07/08* 12                     |
| COTS-MSTW-16   | DASY4                        | Schmid&Partner Engineering AG | DASY4 V4.7 Build71  | -          | SAR               | -                                  |
| COTS-MSTW-17   | S-Parameter Network Analyzer | Agilent                       | -                   | -          | SAR               | -                                  |
| MDA-08         | Dipole Antenna               | Schmid&Partner Engineering AG | D5GHzV2             | 1020       | SAR               | 2009/08/19 * 24                    |
| MPF-02         | 2mmOval Flat Phantom ERI 4.0 | Schmid&Partner Engineering AG | QD VA 001B (ERI4.0) | 1045       | SAR               | -                                  |
| MOS-05         | Thermo-Hygrometer            | Custom                        | CTH-190             | 810201     | SAR               | 2010/04/21 * 12                    |
| MOS-10         | Digital thermometer          | HANNA                         | Checktemp-2         | MOS-10     | SAR               | 2010/02/05* 12                     |
| MBM-13         | Barometer                    | Sunoh                         | SBR121              | 837        | SAR               | 2008/03/14 * 36                    |
| Muscle 5800MHz |                              |                               |                     |            | Daily check       | Target value $\pm 5$ to 10%        |
| SAR room       |                              |                               |                     |            | Daily check       | Ambient Noise<0.012W/kg            |

The expiration date of the calibration is the end of the expired month.

As for some calibrations performed after the tested dates, those test equipment have been controlled by means of an unbroken chains of calibrations.

**UL Japan, Inc.**

**Head Office EMC Lab.**

4383-326 Asama-cho, Ise-shi, Mie-ken 516-0021 JAPAN

Telephone : +81 596 24 8116

Facsimile : +81 596 24 8124

## 2. Dosimetry assessment setup

These measurements were performed with the automated near-field scanning system DASY4 from Schmid & Partner Engineering AG (SPEAG). The system is based on a high precision robot (working range greater than 0.9 m), which positions the probes with a positional repeatability of better than +/- 0.02 mm. Special E- and H-field probes have been developed for measurements close to material discontinuity, the sensors of which are directly loaded with a Schottky diode and connected via highly resistive lines to the data acquisition unit. The SAR measurements were conducted with the dosimetry probe EX3DV3, SN: 3507(manufactured by SPEAG), designed in the classical triangular configuration and optimized for dosimetric evaluation. The probe has been calibrated according to the procedure described in [2] with accuracy of better than +/-10%. The spherical isotropy was evaluated with the procedure described in [3] and found to be better than +/-0.25 dB. The phantom used was the SAM Twin Phantom as described in FCC supplement C, IEEE P1528 and CENELEC EN50361.

---

**UL Japan, Inc.**


**Head Office EMC Lab.**

4383-326 Asama-cho, Ise-shi, Mie-ken 516-0021 JAPAN

Telephone : +81 596 24 8116

Facsimile : +81 596 24 8124

### 3. Configuration and peripherals



The DASY4 system for performing compliance tests consist of the following items:

1. A standard high precision 6-axis robot (Stäubli RX family) with controller and software.  
An arm extension for accommodating the data acquisition electronics (DAE).
2. A dosimetric probe, i.e., an isotropic E-field probe optimized and calibrated for usage in tissue simulating liquid. The probe is equipped with an optical surface detector system.
3. A data acquisition electronic (DAE), which performs the signal amplification, signal multiplexing, AD-conversion, offset measurements, mechanical surface detection, collision detection, etc. The unit is battery powered with standard or rechargeable batteries. The signal is optically transmitted to the EOC.
4. The Electro-optical converter (EOC) performs the conversion between optical and electrical of the signals for the digital communication to the DAE and for the analog signal from the optical surface detection.  
The EOC is connected to the measurement server.
5. The function of the measurement server is to perform the time critical tasks such as signal filtering, control of the robot operation and fast movement interrupts.
6. A probe alignment unit which improves the (absolute) accuracy of the probe positioning.
7. A computer operating Windows 2000.
8. DASY4 software.
9. Remote control with teaches pendant and additional circuitry for robot safety such as warning lamps, etc.
10. The 2mm Flat phantom ERI4.0
11. The device holder for handheld mobile phones.
12. Tissue simulating liquid mixed according to the given recipes.
13. Validation dipole kits allowing to validate the proper functioning of the system.

**UL Japan, Inc.**

**Head Office EMC Lab.**

4383-326 Asama-cho, Ise-shi, Mie-ken 516-0021 JAPAN

Telephone : +81 596 24 8116

Facsimile : +81 596 24 8124

#### 4. System components

##### EX3DV3 Probe Specification

**Construction:**

Symmetrical design with triangular core  
Built-in shielding against static charges  
PEEK enclosure material (resistant to organic solvents, e.g., glycol ether)

**Calibration:**

Basic Broad Band calibration in air : 10-3000 MHz

Conversion Factors (Head and Body):

**Frequency:**

10 MHz to > 6GHz; Linearity: +/-0.2 dB(30 MHz to 3 GHz)

**Directivity:**

+/-0.3 dB in HSL (rotation around probe axis)

+/-0.5 dB in tissue material (rotation normal probe axis)

**Dynamic Range:**

10uW/g to > 100 mW/g; Linearity: +/-0.2 dB(noise: typically < 1uW/g)

**Dimensions:**

Overall length: 330 mm (Tip: 20 mm)

Tip diameter: 2.5mm (Body: 12 mm)

Typical distance from probe tip to dipole centers: 1 mm

**Application:**

Highprecision dosimetric measurement in any exposure scenario  
(e.g., very strong gradient fields). Only probe which enables compliance  
testing for frequencies up to 6GHz with precision of better 30%.



**EX3DV3 E-field Probe**

---

**UL Japan, Inc.**

**Head Office EMC Lab.**

4383-326 Asama-cho, Ise-shi, Mie-ken 516-0021 JAPAN

Telephone : +81 596 24 8116

Facsimile : +81 596 24 8124

## 2mm Flat phantom ERI4.0

### Description

Phantom for compliance testing of handheld and body-mounted wireless devices in the frequency range of 30 MHz to 6 GHz. ELI4 is fully compatible with the latest draft of the standard IEC 62209 Part II and all known tissue simulating liquids. ELI4 has been optimized regarding its performance and can be integrated into our standard phantom tables. A cover prevents evaporation of the liquid. Reference markings on the phantom allow installation of the complete setup, including all predefined phantom positions and measurement grids, by teaching three points. The phantom is supported by software version DASY4.5 and higher and is compatible with all SPEAG dosimetric probes and dipoles.

### Shell Thickness

2.0 ± 0.2 mm (sagging: <1%)

### Filling Volume

approx. 30 liters

### Dimensions

Major ellipse axis : 600 mm

Minor axis : 400 mm

### Compatibilities

- Standard: IEC 62209 Part II (Draft 0.9 and higher)
- Software release: DASY 4.5 or higher
- SPEAG standard phantom table
- all SPEAG dosimetric probes and dipoles

### Device Holder

For this measurement, the urethane foam was used as device holder.

---

## UL Japan, Inc.

### Head Office EMC Lab.

4383-326 Asama-cho, Ise-shi, Mie-ken 516-0021 JAPAN

Telephone : +81 596 24 8116

Facsimile : +81 596 24 8124

## 5. Test system specifications

### Robot RX60L

|                      |   |   |
|----------------------|---|---|
| Number of Axes       | : | 6   |
| Payload              | : | 1.6 kg                                    |
| Reach                | : | 800mm                                     |
| Repeatability        | : | +/-0.025mm                                |
| Control Unit         | : | CS7M                                      |
| Programming Language | : | V+  |
| Manufacture          | : | Stäubli Unimation Corp. Robot Model: RX60 |

### DASY4 Measurement server

|             |   |  |
|-------------|---|--|
| Features    | : | 166MHz low power Pentium MMX 32MB chipdisk and 64MB RAM<br>Serial link to DAE (with watchdog supervision)<br>16 Bit A/D converter for surface detection system<br>Two serial links to robot (one for real-time communication which is supervised by watchdog)<br>Ethernet link to PC (with watchdog supervision)<br>Emergency stop relay for robot safety chain<br>Two expansion slots for future applications |
| Manufacture | : | Schimid & Partner Engineering AG   |

### Data Acquisition Electronic (DAE)

|                      |   |   |
|----------------------|---|---|
| Features             | : | Signal amplifier, multiplexer, A/D converter and control logic<br>Serial optical link for communication with DASY4 embedded system (fully remote controlled)<br>2 step probe touch detector for mechanical surface detection and emergency robot stop (not in -R version) |
| Measurement Range    | : | 1 $\mu$ V to > 200 mV (16 bit resolution and two range settings: 4mV, 400mV)  |
| Input Offset voltage | : | < 1 $\mu$ V (with auto zero)  |
| Input Resistance     | : | 200 M $\Omega$  |
| Battery Power        | : | > 10 h of operation (with two 9 V battery)  |
| Dimension            | : | 60 x 60 x 68 mm   |
| Manufacture          | : | Schimid & Partner Engineering AG  |

### Software

|                      |   |                                   |
|----------------------|---|-----------------------------------|
| Item                 | : | Dosimetric Assesment System DASY4 |
| Type No.             | : | SD 000 401A, SD 000 402A          |
| Software version No. | : | DASY4 V4.7 Build80                |
| Manufacture / Origin | : | Schimid & Partner Engineering AG  |

### E-Field Probe

|              |   |   |
|--------------|---|---|
| Model        | : | EX3DV3                                  |
| Serial No.   | : | 3507                                    |
| Construction | : | Symmetrical design with triangular core |
| Frequency    | : | 10 MHz to 6 GHz                         |
| Linearity    | : | +/-0.2 dB (30 MHz to 3 GHz)             |
| Manufacture  | : | Schimid & Partner Engineering AG        |

### Phantom

|                 |   |   |
|-----------------|---|---|
| Type            | : | 2mm   |
| Shell Thickness | : | 2.0 $\pm$ 0.2 mm (sagging: <1%)               |
| Filling Volume  | : | approx. 30 liters                             |
| Dimensions      | : | Major ellipse axis: 600 mm Minor axis: 400 mm |
| Manufacture     | : | Schimid & Partner Engineering AG              |

---

**UL Japan, Inc.**

**Head Office EMC Lab.**

4383-326 Asama-cho, Ise-shi, Mie-ken 516-0021 JAPAN

Telephone : +81 596 24 8116

Facsimile : +81 596 24 8124

## 6. Simulated Tissues Composition (5GHz)

| Ingredient         | MiXTURE(%)     |
|--------------------|----------------|
|                    | Muscle 5800MHz |
| Water              | 78.0           |
| Mineral Oil        | 11.0           |
| Emulsifiers        | 9.0            |
| Additives and salt | 2.0            |

## 7. Simulated Tissue Liquid Parameter confirmation (5GHz)

The dielectric parameters were checked prior to assessment using the HP85070D dielectric probe kit. The dielectric parameters measurement are reported in each correspondent section.

### 7-a Muscle 5GHz

Type of liquid : **Muscle 5GHz**  
Ambient temperature (deg.c.) : **24.0(18-Jun), 24.5(21-Jun)**  
Relative Humidity (%) : **51(18-Jun),42 (21-Jun)**  
Liquid depth (cm) : **15.0**

| DIELECTRIC PARAMETERS MEASUREMENT RESULTS |                 |                     |       |                                    |                |          |               |           |
|---|-----------------|---------------------|-------|------------------------------------|----------------|----------|---------------|-----------|
| Date                                      | Frequency [MHz] | Liquid Temp [deg.c] |       | Parameters                         | Target Value*1 | Measured | Deviation [%] | Limit [%] |
|   |                 | Before              | After |                                    |                |          |               |           |
| 18-Jun                                    | 5200            | 23.0                | 23.0  | Relative Permittivity $\epsilon_r$ | 49.0           | 46.6     | -4.9          | +/-5      |
|   |                 |                     |       | Coductivity $\sigma$ [mho/m]       | 5.30           | 5.47     | 3.2           | +/-5      |
| 18-Jun                                    | 5300            | 23.0                | 23.0  | Relative Permittivity $\epsilon_r$ | 48.9           | 46.6     | -4.7          | +/-5      |
|   |                 |                     |       | Coductivity $\sigma$ [mho/m]       | 5.42           | 5.65     | 4.2           | +/-5      |
| 21-Jun                                    | 5800            | 23.5                | 23.5  | Relative Permittivity $\epsilon_r$ | 48.2           | 46.0     | -4.6          | +/-5      |
|   |                 |                     |       | Coductivity $\sigma$ [mho/m]       | 6.00           | 6.13     | 2.2           | +/-5      |

\*1 The target values are a parameter defined in FCC OET 65.

| DIELECTRIC PARAMETERS MEASUREMENT RESULTS |                 |                     |       |                                    |                |          |               |           |
|---|-----------------|---------------------|-------|------------------------------------|----------------|----------|---------------|-----------|
| Date                                      | Frequency [MHz] | Liquid Temp [deg.c] |       | Parameters                         | Target Value*2 | Measured | Deviation [%] | Limit [%] |
|   |                 | Before              | After |                                    |                |          |               |           |
| 18-Jun                                    | 5200            | 23.0                | 23.0  | Relative Permittivity $\epsilon_r$ | 48.2           | 46.6     | -3.3          | +/-10     |
|   |                 |                     |       | Coductivity $\sigma$ [mho/m]       | 5.43           | 5.47     | 0.7           | +/-10     |
| 21-Jun                                    | 5800            | 23.5                | 23.5  | Relative Permittivity $\epsilon_r$ | 46.9           | 46.0     | -1.9          | +/-10     |
|   |                 |                     |       | Coductivity $\sigma$ [mho/m]       | 6.17           | 6.13     | -0.6          | +/-10     |

\*2 The target values are the calibrated dipole MSL parameters. (D5GHzV2 SN:1020/p68)

**UL Japan, Inc.**

**Head Office EMC Lab.**

4383-326 Asama-cho, Ise-shi, Mie-ken 516-0021 JAPAN

Telephone : +81 596 24 8116

Facsimile : +81 596 24 8124

**7-b Decision on Simulated Tissues of 5200MHz and 5600MHz**

In the current standards (e.g., IEEE P1528, OET 65 Supplement C), the dielectric parameters suggested for head and body tissue simulating liquid are given at 3000MHz and 5800MHz. As an intermediate solution, dielectric parameters for the frequencies between 5000to 5800 MHz were obtained using linear interpolation. Therefore the dielectric parameters of 5200MHz and 5300MHz(The frequency for the validation) were decided as following.

| f (MHz) | Head Tissue |         | Body Tissue |         | Reference    |
|---------|-------------|---------|-------------|---------|--------------|
|         | r           | [mho/m] | r           | [mho/m] |              |
| 3000    | 38.5        | 2.40    | 52.0        | 2.73    | Standard     |
| 5800    | 35.3        | 5.27    | 48.2        | 6.00    | Standard     |
| 5000    | 36.2        | 4.45    | 49.3        | 5.07    | Interpolated |
| 5100    | 36.1        | 4.55    | 49.1        | 5.18    | Interpolated |
| 5200    | 36.0        | 4.66    | 49.0        | 5.30    | Interpolated |
| 5300    | 35.9        | 4.76    | 48.9        | 5.42    | Interpolated |
| 5400    | 35.8        | 4.86    | 48.7        | 5.53    | Interpolated |
| 5500    | 35.6        | 4.96    | 48.6        | 5.65    | Interpolated |
| 5600    | 35.5        | 5.07    | 48.5        | 5.77    | Interpolated |
| 5700    | 35.4        | 5.17    | 48.3        | 5.88    | Interpolated |

Standard and interpolated dielectric parameters for head and body tissue simulating liquid in the frequency range 3000 to 5800MHz.

**8. System validation data(5GHz)**


Prior to the assessment, the system validation kit was used to test whether the system was operating within its specifications of +/-10%. The validation results are in the table below.

**8-a BODY System validation of 5GHz**

Type of liquid : Muscle 5GHz  
Ambient temperature (deg.c.) : 24.0(18-Jun), 24.5(21-Jun)  
Relative Humidity (%) : 51(18-Jun),42 (21-Jun)  
Dipole : D5GHzV2 SN:1020  
Power : 100mW

| SYSTEM PERFORMANCE CHECK        |           |                      |       |                                    |          |                               |          |  |          |               |           |
|---------------------------------|-----------|----------------------|-------|------------------------------------|----------|-------------------------------|----------|--|----------|---------------|-----------|
| Liquid ( Muscle 5100-5800 MHz ) |           |                      |       |                                    |          |                               |          | System dipole validation target & measured |          |               |           |
| Date                            | Frequency | Liquid Temp [deg.c.] |       | Relative Permittivity $\epsilon_r$ |          | Conductivity $\sigma$ [mho/m] |          | SAR 1g [W/kg]                              |          | Deviation [%] | Limit [%] |
|                                 |           | Before               | After | Target*1                           | Measured | Target*1                      | Measured | Target*2                                   | Measured |               |           |
| 18-Jun                          | 5200      | 23.0                 | 23.0  | 48.2                               | 46.6     | 5.43                          | 5.47     | 7.75                                       | 8.03     | 3.6           | +/-10     |
| 21-Jun                          | 5800      | 23.5                 | 23.5  | 46.9                               | 46.0     | 6.17                          | 6.13     | 7.13                                       | 7.72     | 8.3           | +/-10     |

\*1 The target values are the calibrated dipole MSL parameters. (D5GHzV2 SN:1020/p68)  
\*2 The target value is a parameter defined in calibration data of Dipole antenna (D5GHzV2 SN:1020)  
Note: Please refer to Attachment for the result representation in plot forma



Test system for the system performance check setup diagram

## 9. Validation uncertainty

The uncertainty budget has been determined for the DASY4 measurement system according to the SPEAG documents[6][7] and is given in the following Table.

| Error Description                    | Uncertainty value $\pm$ % | Probability distribution | divisor    | (ci) 1g | Standard Uncertainty (1g)      | vi or veff |
|--------------------------------------|---------------------------|--------------------------|------------|---------|--------------------------------|------------|
| <b>Measurement System</b>            |                           |                          |            |         |                                |            |
| Probe calibration                    | $\pm 6.8$                 | Normal                   | 1          | 1       | $\pm 6.8$                      | $\infty$   |
| Axial isotropy of the probe          | $\pm 4.7$                 | Rectangular              | $\sqrt{3}$ | 1       | $\pm 2.7$                      | $\infty$   |
| Spherical isotropy of the probe      | $\pm 9.6$                 | Rectangular              | 0          | 0       | 0                              | $\infty$   |
| Boundary effects                     | $\pm 2.0$                 | Rectangular              | $\sqrt{3}$ | 1       | $\pm 1.2$                      | $\infty$   |
| Probe linearity                      | $\pm 4.7$                 | Rectangular              | $\sqrt{3}$ | 1       | $\pm 2.7$                      | $\infty$   |
| Detection limit                      | $\pm 1.0$                 | Rectangular              | $\sqrt{3}$ | 1       | $\pm 0.6$                      | $\infty$   |
| Readout electronics                  | $\pm 0.3$                 | Normal                   | 1          | 1       | $\pm 0.3$                      | $\infty$   |
| Response time                        | 0                         | Rectangular              | $\sqrt{3}$ | 1       | 0                              | $\infty$   |
| Integration time                     | 0                         | Rectangular              | $\sqrt{3}$ | 1       | 0                              | $\infty$   |
| RF ambient Noise                     | $\pm 3.0$                 | Rectangular              | $\sqrt{3}$ | 1       | $\pm 1.7$                      | $\infty$   |
| RF ambient Reflections               | $\pm 3.0$                 | Rectangular              | $\sqrt{3}$ | 1       | $\pm 1.7$                      | $\infty$   |
| Probe Positioner                     | $\pm 0.8$                 | Rectangular              | $\sqrt{3}$ | 1       | $\pm 0.5$                      | $\infty$   |
| Probe positioning                    | $\pm 9.9$                 | Rectangular              | 1          | 1       | $\pm 5.7$                      | $\infty$   |
| Algorithms for Max.SAR Eval.         | $\pm 4.0$                 | Rectangular              | $\sqrt{3}$ | 1       | $\pm 2.3$                      | $\infty$   |
| <b>Dipole</b>                        |                           |                          |            |         |                                |            |
| Dipole Axis to Liquid Distance       | $\pm 2.0$                 | Rectangular              | $\sqrt{3}$ | 1       | $\pm 1.2$                      | $\infty$   |
| Input power and SAR drift meas.      | $\pm 4.7$                 | Rectangular              | $\sqrt{3}$ | 1       | $\pm 2.7$                      | $\infty$   |
| <b>Phantom and Setup</b>             |                           |                          |            |         |                                |            |
| Phantom uncertainty                  | $\pm 4.0$                 | Rectangular              | $\sqrt{3}$ | 1       | $\pm 2.3$                      | $\infty$   |
| Liquid conductivity (target)         | $\pm 5.0$                 | Rectangular              | $\sqrt{3}$ | 0.64    | $\pm 1.8$                      | $\infty$   |
| Liquid conductivity (meas.)          | $\pm 5.0$                 | Rectangular              | 1          | 0.64    | $\pm 3.2$                      | $\infty$   |
| Liquid permittivity (target)         | $\pm 5.0$                 | Rectangular              | $\sqrt{3}$ | 0.6     | $\pm 1.7$                      | $\infty$   |
| Liquid permittivity (meas.)          | $\pm 5.0$                 | Rectangular              | 1          | 0.6     | $\pm 3.0$                      | $\infty$   |
| <b>Combined Standard Uncertainty</b> |                           |                          |            |         | <b><math>\pm 12.079</math></b> |            |
| <b>Expanded Uncertainty (k=2)</b>    |                           |                          |            |         | <b><math>\pm 24.2</math></b>   |            |

**UL Japan, Inc.**

**Head Office EMC Lab.**

4383-326 Asama-cho, Ise-shi, Mie-ken 516-0021 JAPAN

Telephone : +81 596 24 8116

Facsimile : +81 596 24 8124

## 10. Validation Measurement data

### BODY 5200MHz System Validation / Dipole 5GHz / Forward Conducted Power : 100mW

Crest factor:1

Medium: M5200 Medium parameters used:  $f = 5200$  MHz;  $\sigma = 5.47$  mho/m;  $\epsilon_r = 46.6$ ;  $\rho = 1000$  kg/m<sup>3</sup>

Phantom section: Flat Section

DASY4 Configuration:

- Probe: EX3DV3 - SN3507; ConvF(4.41, 4.41, 4.41); Calibrated: 2010/02/19

- Sensor-Surface: 2mm (Mechanical Surface Detection)

- Phantom: Flat Phantom ELI4.0

- Measurement SW: DASY4, V4.7 Build 80; Postprocessing SW: SEMCAD, V1.8 Build 186

**Area Scan (101x101x1):** Measurement grid: dx=10mm, dy=10mm

Maximum value of SAR (interpolated) = 17.9 mW/g

**Zoom Scan (8x8x10)/Cube 0:** Measurement grid: dx=4mm, dy=4mm, dz=2.5mm

Reference Value = 63.9 V/m; Power Drift = -0.025 dB

Peak SAR (extrapolated) = 27.2 W/kg



**SAR(1 g) = 8.03 mW/g; SAR(10 g) = 2.29 mW/g**

Maximum value of SAR (measured) = 15.1 mW/g

Test Date = 06/18/10

Ambient Temperature = 24.0 degree.C.

Liquid Temperature = Before 23.0 degree.C. , After 23.0 degree.C.



**UL Japan, Inc.**

**Head Office EMC Lab.**

4383-326 Asama-cho, Ise-shi, Mie-ken 516-0021 JAPAN

Telephone : +81 596 24 8116

Facsimile : +81 596 24 8124

**BODY 5800MHz System Validation / Dipole 5GHz / Forward Conducted Power : 100mW**

Crest factor: 1

Medium: M5800 Medium parameters used:  $f = 5800$  MHz;  $\sigma = 6.13$  mho/m;  $\epsilon_r = 46$ ;  $\rho = 1000$  kg/m<sup>3</sup>

Phantom section: Flat Section

DASY4 Configuration:

- Probe: EX3DV3 - SN3507; ConvF(3.59, 3.59, 3.59); Calibrated: 2010/02/19

- Sensor-Surface: 2mm (Mechanical Surface Detection)

- Phantom: Flat Phantom ELI4.0

- Measurement SW: DASY4, V4.7 Build 80; Postprocessing SW: SEMCAD, V1.8 Build 186

**Area Scan (101x101x1):** Measurement grid: dx=10mm, dy=10mm

Maximum value of SAR (interpolated) = 17.0 mW/g

**Zoom Scan (8x8x10)/Cube 0:** Measurement grid: dx=4mm, dy=4mm, dz=2.5mm

Reference Value = 58.5 V/m; Power Drift = -0.084 dB

Peak SAR (extrapolated) = 28.0 W/kg



**SAR(1 g) = 7.72 mW/g; SAR(10 g) = 2.11 mW/g**

Maximum value of SAR (measured) = 15.1 mW/g

Test Date = 06/21/10

Ambient Temperature = 24.0 degree.C.

Liquid Temperature = Before 23.0 degree.C. , After 23.0 degree.C.



**UL Japan, Inc.**

**Head Office EMC Lab.**

4383-326 Asama-cho, Ise-shi, Mie-ken 516-0021 JAPAN

Telephone : +81 596 24 8116

Facsimile : +81 596 24 8124

11. System Validation Dipole (D5GHzV2,S/N: 1020)

Calibration Laboratory of  
Schmid & Partner  
Engineering AG  
Zeughausstrasse 43, 8004 Zurich, Switzerland



S Schweizerischer Kalibrierdienst  
C Service suisse d'étalonnage  
S Servizio svizzero di taratura  
S Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS)  
The Swiss Accreditation Service is one of the signatories to the EA  
Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: SCS 108

Client: UL Japan (PTT)

Certificate No: D5GHzV2-1020\_Aug09

**CALIBRATION CERTIFICATE**

|  |  |                                   |                         |
|--|--|-----------------------------------|-------------------------|
| Object   | D5GHzV2 - SN: 1020   |                                   |                         |
| Calibration procedure(s)   | QA CAL-22.v1<br>Calibration procedure for dipole validation kits between 3-6 GHz |                                   |                         |
| Calibration date:  | August 19, 2009  |                                   |                         |
| Condition of the calibrated item   | In Tolerance   |                                   |                         |
| This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate. |  |                                   |                         |
| All calibrations have been conducted in the closed laboratory facility; environment temperature (22 ± 3)°C and humidity < 70%.   |  |                                   |                         |
| Calibration Equipment used (M&TE critical for calibration)   |  |                                   |                         |
| Primary Standards  | ID #   | Cal Date (Certificate No.)        | Scheduled Calibration   |
| Power meter EPM-442A   | GB37480704   | 08-Oct-08 (No. 217-00999)         | Oct-09                  |
| Power sensor HP 8481A  | US37292783   | 08-Oct-08 (No. 217-00898)         | Oct-09                  |
| Reference 20 dB Attenuator   | SN: 5086 (20g)   | 31-Mar-09 (No. 217-01025)         | Mar-10                  |
| Type-N mismatch combination  | SN: 5047.2 / 06327   | 31-Mar-09 (No. 217-01029)         | Mar-10                  |
| Reference Probe EX3DV4   | SN: 3503   | 11-Mar-09 (No. EX3-3503_Mar09)    | Mar-10                  |
| DAE4   | SN: 801  | 07-Mar-09 (No. DAE4-801_Mar09)    | Mar-10                  |
| Secondary Standards  | ID #   | Check Date (In house)             | Scheduled Check         |
| Power sensor HP 8481A  | MY41092317   | 18-Oct-02 (In house check Oct-07) | In house check: Oct-09  |
| RF generator R&S SMT-08  | 100005   | 4-Aug-99 (In house check Oct-07)  | In house check: Oct-09  |
| Network Analyzer HP 8753E  | US37380585 S4206   | 18-Oct-01 (In house check Oct-08) | In house check: Oct-09  |
| Calibrated by:   | Name<br>Jeton Kasrati  | Function<br>Laboratory Technician | Signature<br>           |
| Approved by:   | Name<br>Katja Pokovic  | Function<br>Technical Manager     | Signature<br>           |
|  |  |                                   | Issued: August 19, 2009 |
| This calibration certificate shall not be reproduced except in full without written approval of the laboratory.  |  |                                   |                         |

Calibration Laboratory of  
Schmid & Partner  
Engineering AG  
Zoughausstrasse 43, 8004 Zurich, Switzerland



S Schweizerischer Kalibrierdienst  
C Service suisse d'étalonnage  
S Servizio svizzero di taratura  
S Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS)  
The Swiss Accreditation Service is one of the signatories to the EA  
Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: SCS 108

#### Glossary:

TSL tissue simulating liquid  
ConvF sensitivity in TSL / NORM x,y,z  
N/A not applicable or not measured

#### Calibration is Performed According to the Following Standards:

- a) IEC Std 62209 Part 2, "Evaluation of Human Exposure to Radio Frequency Fields from Handheld and Body-Mounted Wireless Communication Devices in the Frequency Range of 30 MHz to 6 GHz: Human models, Instrumentation, and Procedures"; Part 2: "Procedure to determine the Specific Absorption Rate (SAR) for including accessories and multiple transmitters", Draft Version 0.9, December 2004
- b) Federal Communications Commission Office of Engineering & Technology (FCC OET), "Evaluating Compliance with FCC Guidelines for Human Exposure to Radiofrequency Electromagnetic Fields; Additional Information for Evaluating Compliance of Mobile and Portable Devices with FCC Limits for Human Exposure to Radiofrequency Emissions", Supplement C (Edition 01-01) to Bulletin 65

#### Additional Documentation:

- c) DASY4/5 System Handbook

#### Methods Applied and Interpretation of Parameters:

- *Measurement Conditions:* Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- *Antenna Parameters with TSL:* The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.
- *Feed Point Impedance and Return Loss:* These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.
- *Electrical Delay:* One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
- *SAR measured:* SAR measured at the stated antenna input power.
- *SAR normalized:* SAR as measured, normalized to an input power of 1 W at the antenna connector.
- *SAR for nominal TSL parameters:* The measured TSL parameters are used to calculate the nominal SAR result.

UL Japan, Inc.

Head Office EMC Lab.

4383-326 Asama-cho, Ise-shi, Mie-ken 516-0021 JAPAN

Telephone : +81 596 24 8116

Facsimile : +81 596 24 8124

### Measurement Conditions

DASY system configuration, as far as not given on page 1.

|                              |  |             |
|------------------------------|--|-------------|
| DASY Version                 | DASY5  | V5.0        |
| Extrapolation                | Advanced Extrapolation                                   |             |
| Phantom                      | Modular Flat Phantom V5.0                                |             |
| Distance Dipole Center - TSL | 10 mm  | with Spacer |
| Area Scan resolution         | dx, dy = 10 mm   |             |
| Zoom Scan Resolution         | dx, dy = 4.0 mm, dz = 2.5 mm                             |             |
| Frequency                    | 5200 MHz ± 1 MHz<br>5500 MHz ± 1 MHz<br>5800 MHz ± 1 MHz |             |

### Head TSL parameters at 5200 MHz

The following parameters and calculations were applied.

|                                  | Temperature     | Permittivity | Conductivity     |
|----------------------------------|-----------------|--------------|------------------|
| Nominal Head TSL parameters      | 22.0 °C         | 36.0         | 4.66 mho/m       |
| Measured Head TSL parameters     | (22.0 ± 0.2) °C | 36.2 ± 6 %   | 4.65 mho/m ± 6 % |
| Head TSL temperature during test | (22.0 ± 0.2) °C | ---          | ---              |

### SAR result with Head TSL at 5200 MHz

| SAR averaged over 1 cm <sup>3</sup> (1 g) of Head TSL | condition          |                            |
|---|--------------------|----------------------------|
| SAR measured  | 100 mW input power | 7.70 mW / g                |
| SAR normalized  | normalized to 1W   | 77.0 mW / g                |
| SAR for nominal Head TSL parameters <sup>1</sup>      | normalized to 1W   | 77.1 mW / g ± 19.9 % (k=2) |

| SAR averaged over 10 cm <sup>3</sup> (10 g) of Head TSL | condition          |                            |
|---|--------------------|----------------------------|
| SAR measured  | 100 mW input power | 2.19 mW / g                |
| SAR normalized  | normalized to 1W   | 21.9 mW / g                |
| SAR for nominal Head TSL parameters <sup>1</sup>        | normalized to 1W   | 21.9 mW / g ± 19.5 % (k=2) |

<sup>1</sup> Correction to nominal TSL parameters according to c), chapter "SAR Sensitivities"

**UL Japan, Inc.**

**Head Office EMC Lab.**

4383-326 Asama-cho, Ise-shi, Mie-ken 516-0021 JAPAN

Telephone : +81 596 24 8116

Facsimile : +81 596 24 8124

#### Head TSL parameters at 5500 MHz

The following parameters and calculations were applied.

|                                  | Temperature     | Permittivity | Conductivity     |
|----------------------------------|-----------------|--------------|------------------|
| Nominal Head TSL parameters      | 22.0 °C         | 35.6         | 4.98 mho/m       |
| Measured Head TSL parameters     | (22.0 ± 0.2) °C | 35.6 ± 6 %   | 4.93 mho/m ± 6 % |
| Head TSL temperature during test | (22.0 ± 0.2) °C | ----         | ----             |

#### SAR result with Head TSL at 5500 MHz

| SAR averaged over 1 cm <sup>3</sup> (1 g) of Head TSL | condition          |                            |
|---|--------------------|----------------------------|
| SAR measured  | 100 mW input power | 8.07 mW / g                |
| SAR normalized  | normalized to 1W   | 80.7 mW / g                |
| SAR for nominal Head TSL parameters <sup>1</sup>      | normalized to 1W   | 80.7 mW / g ± 19.9 % (k=2) |

| SAR averaged over 10 cm <sup>3</sup> (10 g) of Head TSL | condition          |                            |
|---|--------------------|----------------------------|
| SAR measured  | 100 mW input power | 2.28 mW / g                |
| SAR normalized  | normalized to 1W   | 22.8 mW / g                |
| SAR for nominal Head TSL parameters <sup>1</sup>        | normalized to 1W   | 22.8 mW / g ± 19.5 % (k=2) |

#### Head TSL parameters at 5800 MHz

The following parameters and calculations were applied.

|                                  | Temperature     | Permittivity | Conductivity     |
|----------------------------------|-----------------|--------------|------------------|
| Nominal Head TSL parameters      | 22.0 °C         | 35.3         | 5.27 mho/m       |
| Measured Head TSL parameters     | (22.0 ± 0.2) °C | 35.1 ± 6 %   | 5.20 mho/m ± 6 % |
| Head TSL temperature during test | (22.0 ± 0.2) °C | ----         | ----             |

#### SAR result with Head TSL at 5800 MHz

| SAR averaged over 1 cm <sup>3</sup> (1 g) of Head TSL | condition          |                            |
|---|--------------------|----------------------------|
| SAR measured  | 100 mW input power | 7.49 mW / g                |
| SAR normalized  | normalized to 1W   | 74.9 mW / g                |
| SAR for nominal Head TSL parameters <sup>1</sup>      | normalized to 1W   | 74.7 mW / g ± 19.9 % (k=2) |

| SAR averaged over 10 cm <sup>3</sup> (10 g) of Head TSL | condition          |                            |
|---|--------------------|----------------------------|
| SAR measured  | 100 mW input power | 2.12 mW / g                |
| SAR normalized  | normalized to 1W   | 21.2 mW / g                |
| SAR for nominal Head TSL parameters <sup>1</sup>        | normalized to 1W   | 21.1 mW / g ± 19.5 % (k=2) |

<sup>1</sup> Correction to nominal TSL parameters according to c), chapter "SAR Sensitivities"

**UL Japan, Inc.**

**Head Office EMC Lab.**

4383-326 Asama-cho, Ise-shi, Mie-ken 516-0021 JAPAN

Telephone : +81 596 24 8116

Facsimile : +81 596 24 8124

**Body TSL parameters at 5200 MHz**

The following parameters and calculations were applied.

|                                  | Temperature     | Permittivity | Conductivity     |
|----------------------------------|-----------------|--------------|------------------|
| Nominal Body TSL parameters      | 22.0 °C         | 49.0         | 5.30 mho/m       |
| Measured Body TSL parameters     | (22.0 ± 0.2) °C | 48.2 ± 6 %   | 5.46 mho/m ± 6 % |
| Body TSL temperature during test | (22.5 ± 0.2) °C | ----         | ----             |

**SAR result with Body TSL at 5200 MHz**

| SAR averaged over 1 cm <sup>3</sup> (1 g) of Body TSL | condition          |                            |
|---|--------------------|----------------------------|
| SAR measured  | 100 mW input power | 7.75 mW / g                |
| SAR normalized  | normalized to 1W   | 77.5 mW / g                |
| SAR for nominal Body TSL parameters <sup>1</sup>      | normalized to 1W   | 77.2 mW / g ± 19.9 % (k=2) |

| SAR averaged over 10 cm <sup>3</sup> (10 g) of Body TSL | condition          |                            |
|---|--------------------|----------------------------|
| SAR measured  | 100 mW input power | 2.16 mW / g                |
| SAR normalized  | normalized to 1W   | 21.6 mW / g                |
| SAR for nominal Body TSL parameters <sup>1</sup>        | normalized to 1W   | 21.5 mW / g ± 19.5 % (k=2) |

**Body TSL parameters at 5500 MHz**

The following parameters and calculations were applied.

|                                  | Temperature     | Permittivity | Conductivity     |
|----------------------------------|-----------------|--------------|------------------|
| Nominal Body TSL parameters      | 22.0 °C         | 48.6         | 5.85 mho/m       |
| Measured Body TSL parameters     | (22.0 ± 0.2) °C | 47.4 ± 6 %   | 5.85 mho/m ± 6 % |
| Body TSL temperature during test | (22.5 ± 0.2) °C | ----         | ----             |

**SAR result with Body TSL at 5500 MHz**

| SAR averaged over 1 cm <sup>3</sup> (1 g) of Body TSL | condition          |                            |
|---|--------------------|----------------------------|
| SAR measured  | 100 mW input power | 8.25 mW / g                |
| SAR normalized  | normalized to 1W   | 82.5 mW / g                |
| SAR for nominal Body TSL parameters <sup>1</sup>      | normalized to 1W   | 82.1 mW / g ± 19.9 % (k=2) |

| SAR averaged over 10 cm <sup>3</sup> (10 g) of Body TSL | condition          |                            |
|---|--------------------|----------------------------|
| SAR measured  | 100 mW input power | 2.27 mW / g                |
| SAR normalized  | normalized to 1W   | 22.7 mW / g                |
| SAR for nominal Body TSL parameters <sup>1</sup>        | normalized to 1W   | 22.6 mW / g ± 19.5 % (k=2) |

<sup>1</sup> Correction to nominal TSL parameters according to c), chapter "SAR Sensitivities"

**Body TSL parameters at 5800 MHz**

The following parameters and calculations were applied.

|                                  | Temperature     | Permittivity | Conductivity     |
|----------------------------------|-----------------|--------------|------------------|
| Nominal Body TSL parameters      | 22.0 °C         | 48.2         | 6.00 mho/m       |
| Measured Body TSL parameters     | (22.0 ± 0.2) °C | 46.9 ± 6 %   | 6.21 mho/m ± 6 % |
| Body TSL temperature during test | (22.5 ± 0.2) °C | ----         | ----             |

**SAR result with Body TSL at 5800 MHz**

| SAR averaged over 1 cm <sup>3</sup> (1 g) of Body TSL | condition          |                            |
|---|--------------------|----------------------------|
| SAR measured  | 100 mW input power | 7.13 mW / g                |
| SAR normalized  | normalized to 1W   | 71.3 mW / g                |
| SAR for nominal Body TSL parameters <sup>1</sup>      | normalized to 1W   | 71.0 mW / g ± 19.9 % (k=2) |

| SAR averaged over 10 cm <sup>3</sup> (10 g) of Body TSL | condition          |                            |
|---|--------------------|----------------------------|
| SAR measured  | 100 mW input power | 1.96 mW / g                |
| SAR normalized  | normalized to 1W   | 19.6 mW / g                |
| SAR for nominal Body TSL parameters <sup>1</sup>        | normalized to 1W   | 19.5 mW / g ± 19.5 % (k=2) |

<sup>1</sup> Correction to nominal TSL parameters according to c), chapter "SAR Sensitivities"

**Appendix**

**Antenna Parameters with Head TSL at 5200 MHz**

|                                      |                                |
|--------------------------------------|--------------------------------|
| Impedance, transformed to feed point | 48.7 $\Omega$ - 9.9 j $\Omega$ |
| Return Loss                          | -19.9 dB                       |

**Antenna Parameters with Head TSL at 5500 MHz**

|                                      |                                |
|--------------------------------------|--------------------------------|
| Impedance, transformed to feed point | 49.5 $\Omega$ - 3.4 j $\Omega$ |
| Return Loss                          | -29.2 dB                       |

**Antenna Parameters with Head TSL at 5800 MHz**

|                                      |                                |
|--------------------------------------|--------------------------------|
| Impedance, transformed to feed point | 56.0 $\Omega$ - 1.3 j $\Omega$ |
| Return Loss                          | -24.8 dB                       |

**Antenna Parameters with Body TSL at 5200 MHz**

|                                      |                                |
|--------------------------------------|--------------------------------|
| Impedance, transformed to feed point | 49.2 $\Omega$ - 9.3 j $\Omega$ |
| Return Loss                          | -20.5 dB                       |

**Antenna Parameters with Body TSL at 5500 MHz**

|                                      |                                |
|--------------------------------------|--------------------------------|
| Impedance, transformed to feed point | 49.5 $\Omega$ - 1.7 j $\Omega$ |
| Return Loss                          | -34.9 dB                       |

**Antenna Parameters with Body TSL at 5800 MHz**

|                                      |                                |
|--------------------------------------|--------------------------------|
| Impedance, transformed to feed point | 55.5 $\Omega$ - 0.2 j $\Omega$ |
| Return Loss                          | -25.6 dB                       |

#### General Antenna Parameters and Design

|                                  |          |
|----------------------------------|----------|
| Electrical Delay (one direction) | 1.199 ns |
|----------------------------------|----------|

After long term use with 40 W radiated power, only a slight warming of the dipole near the feedpoint can be measured.

The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals.

No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feedpoint may be damaged.

#### Additional EUT Data

|                 |                   |
|-----------------|-------------------|
| Manufactured by | SPEAG             |
| Manufactured on | February 05, 2004 |

## DASY5 Validation Report for Head TSL

Date/Time: 13.08.2009 14:59:05

Test Laboratory: SPEAG, Zurich, Switzerland

**DUT: Dipole 5GHz; Type: D5GHz; Serial: D5GHzV2 - SN:1020**

Communication System: CW-5GHz; Frequency: 5200 MHzFrequency: 5500 MHzFrequency: 5800 MHz;  
Duty Cycle: 1:1

Medium: HSL 3-6 GHz

Medium parameters used:  $f = 5200$  MHz;  $\sigma = 4.65$  mho/m;  $\epsilon_r = 36.2$ ;  $\rho = 1000$  kg/m<sup>3</sup> Medium parameters used:  $f = 5500$  MHz;  $\sigma = 4.93$  mho/m;  $\epsilon_r = 35.6$ ;  $\rho = 1000$  kg/m<sup>3</sup> Medium parameters used:  $f = 5800$  MHz;  $\sigma = 5.2$  mho/m;  $\epsilon_r = 35.1$ ;  $\rho = 1000$  kg/m<sup>3</sup>

Phantom section: Flat Section

Measurement Standard: DASY5 (IEEE/IEC)

DASY5 Configuration:

- Probe: EX3DV4 - SN3503; ConvF(5.36, 5.36, 5.36)ConvF(4.85, 4.85, 4.85)ConvF(4.74, 4.74, 4.74); Calibrated: 11.03.2009
- Sensor-Surface: 2mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn601; Calibrated: 07.03.2009
- Phantom: Flat Phantom 5.0 (front); Type: QD000P50AA; Serial: 1001
- Measurement SW: DASY5, V5.0 Build 120; SEMCAD X Version 13.4 Build 45

**d=10mm, Pin=100mW, f=5200 MHz/Area Scan (61x61x1):** Measurement grid: dx=10mm, dy=10mm

Maximum value of SAR (interpolated) = 15.8 mW/g

**d=10mm, Pin=100mW, f=5200 MHz/Zoom Scan (8x8x10), dist=2mm (8x8x10)/Cube 0:**

Measurement grid: dx=4mm, dy=4mm, dz=2.5mm

Reference Value = 61.5 V/m; Power Drift = 0.086 dB

Peak SAR (extrapolated) = 30.1 W/kg

**SAR(1 g) = 7.7 mW/g; SAR(10 g) = 2.19 mW/g**

Maximum value of SAR (measured) = 16 mW/g

**d=10mm, Pin=100mW, f=5500 MHz 2/Zoom Scan (8x8x10), dist=2mm (8x8x10)/Cube 0:**

Measurement grid: dx=4mm, dy=4mm, dz=2.5mm

Reference Value = 62 V/m; Power Drift = 0.094 dB

Peak SAR (extrapolated) = 32.8 W/kg

**SAR(1 g) = 8.07 mW/g; SAR(10 g) = 2.28 mW/g**

Maximum value of SAR (measured) = 16.9 mW/g

**d=10mm, Pin=100mW, f=5800 MHz 2/Zoom Scan (8x8x10), dist=2mm (8x8x10)/Cube 0:**

Measurement grid: dx=4mm, dy=4mm, dz=2.5mm

Reference Value = 59 V/m; Power Drift = 0.090 dB

Peak SAR (extrapolated) = 31.9 W/kg

**SAR(1 g) = 7.49 mW/g; SAR(10 g) = 2.12 mW/g**

---


**UL Japan, Inc.**

**Head Office EMC Lab.**

4383-326 Asama-cho, Ise-shi, Mie-ken 516-0021 JAPAN


Telephone : +81 596 24 8116

Facsimile : +81 596 24 8124



0 dB = 16mW/g

Impedance Measurement Plot for Head TSL



## DASY5 Validation Report for Body TSL

Date/Time: 19.08.2009 13:11:30

Test Laboratory: SPEAG, Zurich, Switzerland

**DUT: Dipole 5GHz; Type: D5GHz; Serial: D5GHzV2 - SN:1020**

Communication System: CW-5GHz; Frequency: 5200 MHzFrequency: 5500 MHzFrequency: 5800 MHz;  
Duty Cycle: 1:1

Medium: MSL 3-6 GHz

Medium parameters used:  $f = 5200$  MHz;  $\sigma = 5.43$  mho/m;  $\epsilon_r = 48.2$ ;  $\rho = 1000$  kg/m<sup>3</sup> Medium parameters used:  $f = 5500$  MHz;  $\sigma = 5.82$  mho/m;  $\epsilon_r = 47.5$ ;  $\rho = 1000$  kg/m<sup>3</sup> Medium parameters used:  $f = 5800$  MHz;  $\sigma = 6.17$  mho/m;  $\epsilon_r = 46.9$ ;  $\rho = 1000$  kg/m<sup>3</sup>

Phantom section: Flat Section

Measurement Standard: DASY5 (IEEE/IEC)

DASY5 Configuration:

- Probe: EX3DV4 - SN3503; ConvF(4.88, 4.88, 4.88)ConvF(4.37, 4.37, 4.37)ConvF(4.57, 4.57, 4.57); Calibrated: 11.03.2009
- Sensor-Surface: 2mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn601; Calibrated: 07.03.2009
- Phantom: Flat Phantom 5.0 (back); Type: QD000P50AA; Serial: 1002
- Measurement SW: DASY5, V5.0 Build 120; SEMCAD X Version 13.4 Build 45

**d=10mm, Pin=100mW, f=5200 MHz/Area Scan (61x61x1):** Measurement grid: dx=10mm, dy=10mm  
Maximum value of SAR (interpolated) = 16.3 mW/g

**d=10mm, Pin=100mW, f=5200 MHz/Zoom Scan (8x8x10), dist=2mm (8x8x10)/Cube 0:**  
Measurement grid: dx=4mm, dy=4mm, dz=2.5mm  
Reference Value = 59.5 V/m; Power Drift = 9.72e-005 dB  
Peak SAR (extrapolated) = 29.4 W/kg  
**SAR(1 g) = 7.75 mW/g; SAR(10 g) = 2.16 mW/g**  
Maximum value of SAR (measured) = 16.3 mW/g

**d=10mm, Pin=100mW, f=5500 MHz/Zoom Scan (8x8x10), dist=2mm (8x8x10)/Cube 0:**  
Measurement grid: dx=4mm, dy=4mm, dz=2.5mm  
Reference Value = 59.4 V/m; Power Drift = 0.00123 dB  
Peak SAR (extrapolated) = 33.5 W/kg  
**SAR(1 g) = 8.25 mW/g; SAR(10 g) = 2.27 mW/g**  
Maximum value of SAR (measured) = 17.7 mW/g

**d=10mm, Pin=100mW, f=5800 MHz/Zoom Scan (8x8x10), dist=2mm (8x8x10)/Cube 0:**  
Measurement grid: dx=4mm, dy=4mm, dz=2.5mm  
Reference Value = 54 V/m; Power Drift = 0.00784 dB  
Peak SAR (extrapolated) = 31 W/kg  
**SAR(1 g) = 7.13 mW/g; SAR(10 g) = 1.96 mW/g**  
Maximum value of SAR (measured) = 15.6 mW/g

---


**UL Japan, Inc.**

**Head Office EMC Lab.**

4383-326 Asama-cho, Ise-shi, Mie-ken 516-0021 JAPAN


Telephone : +81 596 24 8116

Facsimile : +81 596 24 8124



0 dB = 15.6mW/g

**Impedance Measurement Plot for Body TSL**



12. Dosimetric E-Field Probe Calibration (EX3DV3,S/N: 3507)

Calibration Laboratory of  
Schmid & Partner  
Engineering AG  
Zeughausstrasse 43, 8004 Zurich, Switzerland



S Schweizerischer Kalibrierdienst  
S Service suisse d'étalonnage  
C Servizio svizzero di taratura  
S Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS)  
The Swiss Accreditation Service is one of the signatories to the EA  
Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: SCS 108

Client **UL Japan (PTT)**

Certificate No: **EX3-3507\_Feb10**

**CALIBRATION CERTIFICATE**

Object: **EX3DV3 - SN:3507**

Calibration procedure(s): **QA CAL-01.v6, QA CAL-14.v3, QA CAL-23.v3 and QA CAL-25.v2  
Calibration procedure for dosimetric E-field probes**

Calibration date: **February 19, 2010**

This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI).  
The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature (22 ± 3)°C and humidity < 70%.

Calibration Equipment used (M&TE critical for calibration)

| Primary Standards          | ID #            | Cal Date (Certificate No.)        | Scheduled Calibration  |
|----------------------------|-----------------|-----------------------------------|------------------------|
| Power meter E4418B         | GB41293874      | 1-Apr-09 (No. 217-01030)          | Apr-10                 |
| Power sensor E4412A        | MY41495277      | 1-Apr-09 (No. 217-01030)          | Apr-10                 |
| Power sensor E4412A        | MY41498087      | 1-Apr-09 (No. 217-01030)          | Apr-10                 |
| Reference 3 dB Attenuator  | SN: S9054 (3c)  | 31-Mar-09 (No. 217-01026)         | Mar-10                 |
| Reference 20 dB Attenuator | SN: S5086 (20b) | 31-Mar-09 (No. 217-01028)         | Mar-10                 |
| Reference 30 dB Attenuator | SN: S5129 (30b) | 31-Mar-09 (No. 217-01027)         | Mar-10                 |
| Reference Probe ES3DV2     | SN: 3013        | 30-Dec-09 (No. ES3-3013_Dec09)    | Dec-10                 |
| DAE4                       | SN: 660         | 29-Sep-09 (No. DAE4-660_Sep09)    | Sep-10                 |
| Secondary Standards        | ID #            | Check Date (in house)             | Scheduled Check        |
| RF generator HP 8648C      | US3642U01700    | 4-Aug-99 (in house check Oct-09)  | In house check: Oct-11 |
| Network Analyzer HP 8753E  | US37390585      | 18-Oct-01 (in house check Oct-09) | In house check: Oct10  |

Calibrated by: **Katja Pokovic** (Name), **Technical Manager** (Function), *[Signature]* (Signature)

Approved by: **Niels Kuster** (Name), **Quality Manager** (Function), *[Signature]* (Signature)

Issued: February 19, 2010

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

Calibration Laboratory of  
Schmid & Partner  
Engineering AG  
Zeughausstrasse 43, 8004 Zurich, Switzerland



S Schweizerischer Kalibrierdienst  
C Service suisse d'étalonnage  
S Servizio svizzero di taratura  
S Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS)  
The Swiss Accreditation Service is one of the signatories to the EA  
Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: SCS 108

#### Glossary:

|                          |   |
|--------------------------|---|
| TSL                      | tissue simulating liquid  |
| NORM <sub>x,y,z</sub>    | sensitivity in free space   |
| ConvF                    | sensitivity in TSL / NORM <sub>x,y,z</sub>  |
| DCP                      | diode compression point   |
| CF                       | crest factor (1/duty_cycle) of the RF signal  |
| A, B, C                  | modulation dependent linearization parameters   |
| Polarization $\varphi$   | $\varphi$ rotation around probe axis  |
| Polarization $\vartheta$ | $\vartheta$ rotation around an axis that is in the plane normal to probe axis (at measurement center),<br>i.e., $\vartheta = 0$ is normal to probe axis |

#### Calibration is Performed According to the Following Standards:

- IEEE Std 1528-2003, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", December 2003
- IEC 62209-1, "Procedure to measure the Specific Absorption Rate (SAR) for hand-held devices used in close proximity to the ear (frequency range of 300 MHz to 3 GHz)", February 2005

#### Methods Applied and Interpretation of Parameters:

- NORM<sub>x,y,z</sub>: Assessed for E-field polarization  $\vartheta = 0$  ( $f \leq 900$  MHz in TEM-cell;  $f > 1800$  MHz: R22 waveguide). NORM<sub>x,y,z</sub> are only intermediate values, i.e., the uncertainties of NORM<sub>x,y,z</sub> does not effect the E<sup>2</sup>-field uncertainty inside TSL (see below ConvF).
- NORM(f)<sub>x,y,z</sub> = NORM<sub>x,y,z</sub> \* frequency\_response (see Frequency Response Chart). This linearization is implemented in DASY4 software versions later than 4.2. The uncertainty of the frequency response is included in the stated uncertainty of ConvF.
- DCP<sub>x,y,z</sub>: DCP are numerical linearization parameters assessed based on the data of power sweep with CW signal (no uncertainty required). DCP does not depend on frequency nor media.
- A<sub>x,y,z</sub>; B<sub>x,y,z</sub>; C<sub>x,y,z</sub>; VR<sub>x,y,z</sub>: A, B, C are numerical linearization parameters assessed based on the data of power sweep for specific modulation signal. The parameters do not depend on frequency nor media. VR is the maximum calibration range expressed in RMS voltage across the diode.
- ConvF and Boundary Effect Parameters: Assessed in flat phantom using E-field (or Temperature Transfer Standard for  $f < 800$  MHz) and inside waveguide using analytical field distributions based on power measurements for  $f > 800$  MHz. The same setups are used for assessment of the parameters applied for boundary compensation (alpha, depth) of which typical uncertainty values are given. These parameters are used in DASY4 software to improve probe accuracy close to the boundary. The sensitivity in TSL corresponds to NORM<sub>x,y,z</sub> \* ConvF whereby the uncertainty corresponds to that given for ConvF. A frequency dependent ConvF is used in DASY version 4.4 and higher which allows extending the validity from  $\pm 50$  MHz to  $\pm 100$  MHz.
- Spherical isotropy (3D deviation from isotropy): in a field of low gradients realized using a flat phantom exposed by a patch antenna.
- Sensor Offset: The sensor offset corresponds to the offset of virtual measurement center from the probe tip (on probe axis). No tolerance required.

UL Japan, Inc.

Head Office EMC Lab.

4383-326 Asama-cho, Ise-shi, Mie-ken 516-0021 JAPAN

Telephone : +81 596 24 8116

Facsimile : +81 596 24 8124

EX3DV3 SN:3507

February 19, 2010

# Probe EX3DV3

## SN:3507

Manufactured: December 15, 2003  
Last calibrated: February 12, 2009  
Recalibrated: February 19, 2010

Calibrated for DASYS Systems

(Note: non-compatible with DASYS2 system!)

---

**UL Japan, Inc.**

**Head Office EMC Lab.**

4383-326 Asama-cho, Ise-shi, Mie-ken 516-0021 JAPAN

Telephone : +81 596 24 8116

Facsimile : +81 596 24 8124

EX3DV3 SN:3507

February 19, 2010

**DASY - Parameters of Probe: EX3DV3 SN:3507**

**Basic Calibration Parameters**

|   | Sensor X | Sensor Y | Sensor Z | Unc (k=2) |
|---|----------|----------|----------|-----------|
| Norm ( $\mu\text{V}/(\text{V}/\text{m})^2$ ) <sup>A</sup> | 0.68     | 0.73     | 0.69     | ± 10.1%   |
| DCP (mV) <sup>B</sup>                                     | 94.7     | 90.2     | 92.4     |           |

**Modulation Calibration Parameters**

| UID   | Communication System Name | PAR  |   | A<br>dB | B<br>dBuV | C    | VR<br>mV | Unc <sup>E</sup><br>(k=2) |
|-------|---------------------------|------|---|---------|-----------|------|----------|---------------------------|
| 10000 | CW                        | 0.00 | X | 0.00    | 0.00      | 1.00 | 300      | ± 1.5%                    |
|       |                           |      | Y | 0.00    | 0.00      | 1.00 | 300      |                           |
|       |                           |      | Z | 0.00    | 0.00      | 1.00 | 300      |                           |

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

<sup>A</sup> The uncertainties of NormX,Y,Z do not affect the E<sup>2</sup>-field uncertainty inside TSL (see Pages 5 and 8).

<sup>B</sup> Numerical linearization parameter: uncertainty not required.

<sup>E</sup> Uncertainty is determined using the maximum deviation from linear response applying rectangular distribution and is expressed for the square of the field value.

EX3DV3 SN:3507

February 19, 2010

## DASY - Parameters of Probe: EX3DV3 SN:3507

### Calibration Parameter Determined in Head Tissue Simulating Media

| f [MHz] | Validity [MHz] <sup>c</sup> | Permittivity | Conductivity | ConvF X | ConvF Y | ConvF Z | Alpha | Depth Unc (k=2) |
|---------|-----------------------------|--------------|--------------|---------|---------|---------|-------|-----------------|
| 835     | ± 50 / ± 100                | 41.5 ± 5%    | 0.80 ± 5%    | 10.59   | 10.59   | 10.59   | 0.41  | 0.77 ± 11.0%    |
| 900     | ± 50 / ± 100                | 41.5 ± 5%    | 0.97 ± 5%    | 10.25   | 10.25   | 10.25   | 0.60  | 0.63 ± 11.0%    |
| 1640    | ± 50 / ± 100                | 40.3 ± 5%    | 1.29 ± 5%    | 9.45    | 9.45    | 9.45    | 0.59  | 0.62 ± 11.0%    |
| 1750    | ± 50 / ± 100                | 40.1 ± 5%    | 1.37 ± 5%    | 9.16    | 9.16    | 9.16    | 0.83  | 0.50 ± 11.0%    |
| 1810    | ± 50 / ± 100                | 40.0 ± 5%    | 1.40 ± 5%    | 8.90    | 8.90    | 8.90    | 0.98  | 0.46 ± 11.0%    |
| 1900    | ± 50 / ± 100                | 40.0 ± 5%    | 1.40 ± 5%    | 8.73    | 8.73    | 8.73    | 0.83  | 0.47 ± 11.0%    |
| 2000    | ± 50 / ± 100                | 40.0 ± 5%    | 1.40 ± 5%    | 8.66    | 8.66    | 8.66    | 0.66  | 0.54 ± 11.0%    |
| 2450    | ± 50 / ± 100                | 39.2 ± 5%    | 1.80 ± 5%    | 8.01    | 8.01    | 8.01    | 0.27  | 0.85 ± 11.0%    |
| 5200    | ± 50 / ± 100                | 36.0 ± 5%    | 4.66 ± 5%    | 5.14    | 5.14    | 5.14    | 0.30  | 1.90 ± 13.1%    |
| 5300    | ± 50 / ± 100                | 35.9 ± 5%    | 4.76 ± 5%    | 4.78    | 4.76    | 4.76    | 0.35  | 1.90 ± 13.1%    |
| 5500    | ± 50 / ± 100                | 35.6 ± 5%    | 4.96 ± 5%    | 4.36    | 4.36    | 4.36    | 0.40  | 1.90 ± 13.1%    |
| 5600    | ± 50 / ± 100                | 35.5 ± 5%    | 5.07 ± 5%    | 4.12    | 4.12    | 4.12    | 0.50  | 1.90 ± 13.1%    |
| 5800    | ± 50 / ± 100                | 35.3 ± 5%    | 5.27 ± 5%    | 4.10    | 4.10    | 4.10    | 0.50  | 1.90 ± 13.1%    |

<sup>c</sup> The validity of ± 100 MHz only applies for DASY v4.4 and higher (see Page 2). The uncertainty is the RSS of the ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band.

EX3DV3 SN:3507

February 19, 2010

## DASY - Parameters of Probe: EX3DV3 SN:3507

### Calibration Parameter Determined in Body Tissue Simulating Media

| f [MHz] | Validity [MHz] <sup>C</sup> | Permittivity | Conductivity | ConvF X | ConvF Y | ConvF Z | Alpha | Depth Unc (k=2) |
|---------|-----------------------------|--------------|--------------|---------|---------|---------|-------|-----------------|
| 835     | ± 50 / ± 100                | 55.2 ± 5%    | 0.97 ± 5%    | 10.29   | 10.29   | 10.29   | 0.45  | 0.77 ± 11.0%    |
| 900     | ± 50 / ± 100                | 55.0 ± 5%    | 1.05 ± 5%    | 10.13   | 10.13   | 10.13   | 0.58  | 0.68 ± 11.0%    |
| 1640    | ± 50 / ± 100                | 53.8 ± 5%    | 1.40 ± 5%    | 9.49    | 9.49    | 9.49    | 0.99  | 0.47 ± 11.0%    |
| 1750    | ± 50 / ± 100                | 53.4 ± 5%    | 1.49 ± 5%    | 8.79    | 8.79    | 8.79    | 0.59  | 0.80 ± 11.0%    |
| 1810    | ± 50 / ± 100                | 53.3 ± 5%    | 1.52 ± 5%    | 8.66    | 8.66    | 8.66    | 0.49  | 0.70 ± 11.0%    |
| 1900    | ± 50 / ± 100                | 53.3 ± 5%    | 1.52 ± 5%    | 8.47    | 8.47    | 8.47    | 0.28  | 1.00 ± 11.0%    |
| 2000    | ± 50 / ± 100                | 53.3 ± 5%    | 1.52 ± 5%    | 8.61    | 8.61    | 8.61    | 0.39  | 0.80 ± 11.0%    |
| 2450    | ± 50 / ± 100                | 52.7 ± 5%    | 1.95 ± 5%    | 8.11    | 8.11    | 8.11    | 0.28  | 0.98 ± 11.0%    |
| 5200    | ± 50 / ± 100                | 49.0 ± 5%    | 5.30 ± 5%    | 4.41    | 4.41    | 4.41    | 0.55  | 1.95 ± 13.1%    |
| 5300    | ± 50 / ± 100                | 48.5 ± 5%    | 5.42 ± 5%    | 4.22    | 4.22    | 4.22    | 0.60  | 1.95 ± 13.1%    |
| 5500    | ± 50 / ± 100                | 48.6 ± 5%    | 5.65 ± 5%    | 3.78    | 3.78    | 3.78    | 0.60  | 1.95 ± 13.1%    |
| 5600    | ± 50 / ± 100                | 48.5 ± 5%    | 5.77 ± 5%    | 3.59    | 3.59    | 3.59    | 0.63  | 1.95 ± 13.1%    |
| 5800    | ± 50 / ± 100                | 48.2 ± 5%    | 6.00 ± 5%    | 3.59    | 3.59    | 3.59    | 0.65  | 1.95 ± 13.1%    |


<sup>C</sup> The validity of ± 100 MHz only applies for DASY v4.4 and higher (see Page 2). The uncertainty is the RSS of the ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band.

EX3DV3 SN:3507

February 19, 2010

### Frequency Response of E-Field

(TEM-Cell: ifi110 EXX, Waveguide: R22)





Uncertainty of Frequency Response of E-field:  $\pm 6.3\%$  (k=2)

EX3DV3 SN:3507

February 19, 2010

Receiving Pattern ( $\phi$ ),  $\theta = 0^\circ$




Uncertainty of Axial Isotropy Assessment:  $\pm 0.5\%$  ( $k=2$ )

EX3DV3 SN:3507

February 19, 2010

**Dynamic Range f(SAR<sub>head</sub>)**  
(Waveguide R22, f = 1800 MHz)




Uncertainty of Linearity Assessment:  $\pm 0.6\%$  (k=2)

EX3DV3 SN:3507


February 19, 2010

### Conversion Factor Assessment



### Deviation from Isotropy in HSL

Error ( $\phi$ ,  $\theta$ ), f = 900 MHz



Uncertainty of Spherical Isotropy Assessment:  $\pm 2.6\%$  (k=2)

EX3DV3 SN:3507

February 19, 2010

### Other Probe Parameters

|   |                |
|---|----------------|
| Sensor Arrangement                            | Triangular     |
| Connector Angle (°)                           | Not applicable |
| Mechanical Surface Detection Mode             | enabled        |
| Optical Surface Detection Mode                | disabled       |
| Probe Overall Length                          | 337 mm         |
| Probe Body Diameter                           | 10 mm          |
| Tip Length                                    | 9 mm           |
| Tip Diameter                                  | 2.5 mm         |
| Probe Tip to Sensor X Calibration Point       | 1 mm           |
| Probe Tip to Sensor Y Calibration Point       | 1 mm           |
| Probe Tip to Sensor Z Calibration Point       | 1 mm           |
| Recommended Measurement Distance from Surface | 2 mm           |

### 13. References

- [1]ANSI, ANSI/IEEE C95.1-1992: IEEE Standard for Safety Levels with Respect to Human Exposure to Radio Frequency Electromagnetic Fields, 3 kHz to 300 GHz, The Institute of Electrical and Electronics Engineers, Inc., New York, NY 10017, 1992.
- [2] Katja Pokovic, Thomas Schmid, and Niels Kuster, "Robust setup for precise calibration of E-field probes in tissue simulating liquids at mobile communications frequencies", in ICECOM '97, Dubrovnik, October 15-17, 1997, pp. 120-124.
- [3] Katja Pokovic, Thomas Schmid, and Niels Kuster, "E-field probe with improved isotropy in brain simulating liquids", in Proceedings of the ELMAR, Zadar, Croatia, 23-25 June, 1996, pp.172-175.
- [4] W. Gander, Computermathematik, Birkhaeuser, Basel, 1992.
- [5] W. H. Press, S. A. Teukolsky, W. T. Vetterling, and B. P. Flannery, Numerical Recipes in C, The Art of Scientific Computing, Second Edition, Cambridge University Press, 1992.
- [6]SPEAG uncertainty document for DASY 4 System from SPEAG (Shimid & Partner Engineering AG).
- [7]SPEAG uncertainty document for "the 5-6GHz Extension" from SPEAG (Shimid & Partner Engineering AG).

---

**UL Japan, Inc.**

**Head Office EMC Lab.**

4383-326 Asama-cho, Ise-shi, Mie-ken 516-0021 JAPAN

Telephone : +81 596 24 8116

Facsimile : +81 596 24 8124