

Appendix 2-2: SAR measurement data / SAR test data plot of other test conditions (cont'd)

Step 8: W58 band (Head)

Plot 8-3: (Head) Antenna Main; Front (Patient)-main & touch, 11n(40HT) (MCS0), 5755 MHz

EUT: SKR3000; Type: P-61; Serial: A8CE-S002

Mode: n40(MCS0,BPSK/OFDM)(UID 0, Frame Length in ms: 0; PAR: 0; PMF: 1); Frequency: 5755 MHz; Crest Factor: 1.0

Medium: HSL5GHz(1608); Medium parameters used: $f = 5755$ MHz; $\sigma = 5.018$ S/m; $\epsilon_r = 35.37$; $\rho = 1000$ kg/m³

Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007)

DASY Configuration: -Probe: EX3DV4 - SN3907; ConvF(4.3, 4.3, 4.3); Calibrated: 2016/05/12; -DASY52 52.8.8(1222); SEMCAD X 14.6.10(7331)
-Sensor-Surface: 1.4mm (Mechanical Surface Detection), z = 1.0, 25.0 -Electronics: DAE4 Sn626; Calibrated: 2015/09/15
-Phantom: ELI v4.0; Type: QDOVA001BA; Serial: 1059; Phantom section: Flat Section

head,w58,ant-main(chain0)/5g66w58h14,ant0,front(patient)&d0,n40(m0,p12),h5755/


Area Scan:80x120,stp10 (9x13x1): Measurement grid: dx=10mm, dy=10mm; Maximum value of SAR (measured) = 0.0977 W/kg

Area Scan:80x120,stp10 (81x121x1): Interpolated grid: dx=1.000 mm, dy=1.000 mm; Maximum value of SAR (interpolated) = 0.00351 W/kg

Zoom:28x28x24,xy4-z1.4(ratio) (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm;

Reference Value = 4.884 V/m; Power Drift = 0.14 dB; Maximum value of SAR (measured) = 0.0928 W/kg; Peak SAR (extrapolated) = 0.191 W/kg

SAR(1 g) = 0.026 W/kg; SAR(10 g) = 0.00364 W/kg



Remarks: * Date tested: 2016/08/04; Tested by: Hiroshi Naka; Tested place: No.7 shielded room,
* liquid depth: 153 mm; Position: distance of EUT to phantom: 0 mm (2 mm to liquid); ambient: 24 ± 1 deg.C. / 50 ± 10 %RH,
* liquid temperature: 23.0(start)/23.0(end)/22.8(in check) deg.C.; * White cubic: zoom scan area, Red cubic: big-SAR(10g)/small-SAR(1g)

Plot 8-4: (Head) Antenna Sub; Front (Patient)-sub & touch, 11n(40HT) (MCS0), 5755 MHz

EUT: SKR3000; Type: P-61; Serial: A8CE-S002

Mode: n40(MCS0,BPSK/OFDM)(UID 0, Frame Length in ms: 0; PAR: 0; PMF: 1); Frequency: 5755 MHz; Crest Factor: 1.0

Medium: HSL5GHz(1608); Medium parameters used: $f = 5755$ MHz; $\sigma = 5.018$ S/m; $\epsilon_r = 35.37$; $\rho = 1000$ kg/m³

Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007)

DASY Configuration: -Probe: EX3DV4 - SN3907; ConvF(4.3, 4.3, 4.3); Calibrated: 2016/05/12; -DASY52 52.8.8(1222); SEMCAD X 14.6.10(7331)
-Sensor-Surface: 1.4mm (Mechanical Surface Detection), z = 1.0, 25.0 -Electronics: DAE4 Sn626; Calibrated: 2015/09/15
-Phantom: ELI v4.0; Type: QDOVA001BA; Serial: 1059; Phantom section: Flat Section

head,w58,ant-sub(chain1)/5g67w58h15,ant1,front(patient)&d0,n40(m0,p12),h5755/


Area Scan:80x130,stp10 (9x14x1): Measurement grid: dx=10mm, dy=10mm; Maximum value of SAR (measured) = 0.0782 W/kg

Area Scan:80x130,stp10 (81x131x1): Interpolated grid: dx=1.000 mm, dy=1.000 mm; Maximum value of SAR (interpolated) = 0.00801 W/kg

Zoom:28x28x24,xy4-z1.4(ratio) (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm;

Reference Value = 4.275 V/m; Power Drift = -0.10 dB; Maximum value of SAR (measured) = 0.0665 W/kg; Peak SAR (extrapolated) = 0.287 W/kg

SAR(1 g) = 0.012 W/kg; SAR(10 g) = 0.0015 W/kg



Remarks: * Date tested: 2016/08/04; Tested by: Hiroshi Naka; Tested place: No.7 shielded room,
* liquid depth: 153 mm; Position: distance of EUT to phantom: 0 mm (2 mm to liquid); ambient: 24 ± 1 deg.C. / 50 ± 10 %RH,
* liquid temperature: 23.0(start)/23.0(end)/22.8(in check) deg.C.; * White cubic: zoom scan area, Red cubic: big-SAR(10g)/small-SAR(1g)

Appendix 2-2: SAR measurement data / SAR test data plot of other test conditions / Step 8: W58 band (Head) (cont'd)

Plot 8-5: (Head) Antenna Main; Back-main & touch, 11n(40HT) (MCS0), 5755 MHz

EUT: SKR3000; Type: P-61; Serial: A8CE-S002

Mode: n40(MCS0,BPSK/OFDM)(UID 0, Frame Length in ms: 0; PAR: 0; PMF: 1); Frequency: 5755 MHz; Crest Factor: 1.0

Medium: HSL5GHz(1608); Medium parameters used: $f = 5755$ MHz; $\sigma = 5.018$ S/m; $\epsilon_r = 35.37$; $\rho = 1000$ kg/m³

Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007)

DASY Configuration: -Probe: EX3DV4 - SN3907; ConvF(4.3, 4.3, 4.3); Calibrated: 2016/05/12; -DASY52 52.8.8(1222); SEMCAD X 14.6.10(7331)
-Sensor-Surface: 1.4mm (Mechanical Surface Detection), z= 1.0, 25.0 -Electronics: DAE4 Sn626; Calibrated: 2015/09/15
-Phantom: ELI v4.0; Type: QDOVA001BA; Serial: 1059; Phantom section: Flat Section

head,w58,ant-main(chain0)/5g69w58h17,ant0,back&d0,n40(m0,p12),h5755/


Area Scan:80x120,stp10 (9x13x1): Measurement grid: dx=10mm, dy=10mm; Maximum value of SAR (measured) = 0.176 W/kg

Area Scan:80x120,stp10 (81x121x1): Interpolated grid: dx=1.000 mm, dy=1.000 mm; Maximum value of SAR (interpolated) = 0.213 W/kg

Zoom:28x28x24,xy4-z1.4(ratio) (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm;

Reference Value = 6.491 V/m; Power Drift = -0.14 dB; Maximum value of SAR (measured) = 0.181 W/kg; Peak SAR (extrapolated) = 0.328 W/kg

SAR(1 g) = 0.056 W/kg; SAR(10 g) = 0.012 W/kg



Remarks: * Date tested: 2016/08/04; Tested by: Hiroshi Naka; Tested place: No.7 shielded room,
* liquid depth: 153 mm; Position: distance of EUT to phantom: 0 mm (2 mm to liquid); ambient: 24 ± 1 deg.C. / 50 ± 10 %RH,
* liquid temperature: 23.0(start)/23.0(end)/22.8(in check) deg.C.; * White cubic: zoom scan area, Red cubic: big=SAR(10g)/small=SAR(1g)

Plot 8-6: (Head) Antenna Sub; Back-sub & touch, 11n(40HT) (MCS0), 5755 MHz

EUT: SKR3000; Type: P-61; Serial: A8CE-S002

Mode: n40(MCS0,BPSK/OFDM)(UID 0, Frame Length in ms: 0; PAR: 0; PMF: 1); Frequency: 5755 MHz; Crest Factor: 1.0

Medium: HSL5GHz(1608); Medium parameters used: $f = 5755$ MHz; $\sigma = 5.018$ S/m; $\epsilon_r = 35.37$; $\rho = 1000$ kg/m³

Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007)

DASY Configuration: -Probe: EX3DV4 - SN3907; ConvF(4.3, 4.3, 4.3); Calibrated: 2016/05/12; -DASY52 52.8.8(1222); SEMCAD X 14.6.10(7331)
-Sensor-Surface: 1.4mm (Mechanical Surface Detection), z= 1.0, 25.0 -Electronics: DAE4 Sn626; Calibrated: 2015/09/15
-Phantom: ELI v4.0; Type: QDOVA001BA; Serial: 1059; Phantom section: Flat Section

head,w58,ant-sub(chain1)/5g68w58h16,ant1,back&d0,n40(m0,p12),h5755/


Area Scan:80x120,stp10 (9x13x1): Measurement grid: dx=10mm, dy=10mm; Maximum value of SAR (measured) = 0.116 W/kg

Area Scan:80x120,stp10 (81x121x1): Interpolated grid: dx=1.000 mm, dy=1.000 mm; Maximum value of SAR (interpolated) = 0.00890 W/kg

Zoom:28x28x24,xy4-z1.4(ratio) (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm;

Reference Value = 5.432 V/m; Power Drift = -0.20 dB; Maximum value of SAR (measured) = 0.153 W/kg; Peak SAR (extrapolated) = 0.311 W/kg

SAR(1 g) = 0.021 W/kg; SAR(10 g) = 0.00374 W/kg



Remarks: * Date tested: 2016/08/04; Tested by: Hiroshi Naka; Tested place: No.7 shielded room,
* liquid depth: 153 mm; Position: distance of EUT to phantom: 0 mm (2 mm to liquid); ambient: 24 ± 1 deg.C. / 50 ± 10 %RH,
* liquid temperature: 23.0(start)/23.0(end)/22.8(in check) deg.C.; * White cubic: zoom scan area, Red cubic: big=SAR(10g)/small=SAR(1g)

Appendix 2-2: SAR measurement data / SAR test data plot of other test conditions / Step 8: W58 band (Head) (cont'd)

Plot 8-7: (Head) Antenna Main; Long side-main & touch, 11n(40HT) (MCS0), 5755 MHz

EUT: SKR3000; Type: P-61; Serial: A8CE-S002

Mode: n40(MCS0,BPSK/OFDM)(UID 0, Frame Length in ms: 0; PAR: 0; PMF: 1); Frequency: 5755 MHz; Crest Factor: 1.0

Medium: HSL5GHz(1608); Medium parameters used: $f = 5755$ MHz; $\sigma = 5.018$ S/m; $\epsilon_r = 35.37$; $\rho = 1000$ kg/m³

Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007)

DASY Configuration: -Probe: EX3DV4 - SN3907; ConvF(4.3, 4.3, 4.3); Calibrated: 2016/05/12; -DASY52 52.8.8(1222); SEMCAD X 14.6.10(7331)
-Sensor-Surface: 1.4mm (Mechanical Surface Detection), z= 1.0, 25.0 -Electronics: DAE4 Sn626; Calibrated: 2015/09/15
-Phantom: ELI v4.0; Type: QDOVA001BA; Serial: 1059; Phantom section: Flat Section

head,w58,ant-main(chain0)/5g53w58h1,bw40;ant0,side&d0,n40(m0,p12),h5755/


Area scan:120x60,stp10 (13x7x1): Measurement grid: dx=10mm, dy=10mm; Maximum value of SAR (measured) = 1.54 W/kg

Area scan:120x60,stp10 (121x61x1): Interpolated grid: dx=1.000 mm, dy=1.000 mm; Maximum value of SAR (interpolated) = 2.76 W/kg

Zoom:28x28x24,xy4-z1.4(ratio) (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm;

Reference Value = 24.33 V/m; Power Drift = -0.06 dB; Maximum value of SAR (measured) = 2.06 W/kg; Peak SAR (extrapolated) = 4.39 W/kg

SAR(1 g) = 0.702 W/kg; SAR(10 g) = 0.139 W/kg



Remarks: * Date tested: 2016/08/04; Tested by: Hiroshi Naka; Tested place:No.7 shielded room,
* liquid depth: 153 mm; Position: distance of EUT to phantom: 0 mm (2 mm to liquid); ambient: 24 ± 1 deg.C. / 50 ± 10 %RH,
* liquid temperature: 22.8(start)/22.8(end)/22.8(in check) deg.C.; * White cubic: zoom scan area, Red cubic: big=SAR(10g)/small=SAR(1g)

Plot 8-8: (Head) Antenna Main; Long side-main & touch, 11n(40HT) (MCS0), 5795 MHz

EUT: SKR3000; Type: P-61; Serial: A8CE-S002

Mode: n40(MCS0,BPSK/OFDM)(UID 0, Frame Length in ms: 0; PAR: 0; PMF: 1); Frequency: 5795 MHz; Crest Factor: 1.0

Medium: HSL5GHz(1608); Medium parameters used: $f = 5795$ MHz; $\sigma = 5.038$ S/m; $\epsilon_r = 35.32$; $\rho = 1000$ kg/m³

Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007)

DASY Configuration: -Probe: EX3DV4 - SN3907; ConvF(4.3, 4.3, 4.3); Calibrated: 2016/05/12; -DASY52 52.8.8(1222); SEMCAD X 14.6.10(7331)
-Sensor-Surface: 1.4mm (Mechanical Surface Detection), z= 1.0, 25.0 -Electronics: DAE4 Sn626; Calibrated: 2015/09/15
-Phantom: ELI v4.0; Type: QDOVA001BA; Serial: 1059; Phantom section: Flat Section

head,w58,ant-main(chain0)/5g54w58h2,bw40;ch;ant0,side&d0,n40(m0,p12),h5795/


Area Scan:100x40,stp10 (11x5x1): Measurement grid: dx=10mm, dy=10mm; Maximum value of SAR (measured) = 1.33 W/kg

Area Scan:100x40,stp10 (101x41x1): Interpolated grid: dx=1.000 mm, dy=1.000 mm; Maximum value of SAR (interpolated) = 2.27 W/kg

Zoom:28x28x24,xy4-z1.4(ratio) (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm;

Reference Value = 22.68 V/m; Power Drift = 0.03 dB; Maximum value of SAR (measured) = 1.87 W/kg; Peak SAR (extrapolated) = 3.89 W/kg

SAR(1 g) = 0.629 W/kg; SAR(10 g) = 0.124 W/kg



Remarks: * Date tested: 2016/08/04; Tested by: Hiroshi Naka; Tested place:No.7 shielded room,
* liquid depth: 153 mm; Position: distance of EUT to phantom: 0 mm (2 mm to liquid); ambient: 24 ± 1 deg.C. / 50 ± 10 %RH,
* liquid temperature: 22.8(start)/22.8(end)/22.8(in check) deg.C.; * White cubic: zoom scan area, Red cubic: big=SAR(10g)/small=SAR(1g)

UL Japan, Inc.

Shonan EMC Lab.

1-22-3 Megumigaoka, Hiratsuka-shi, Kanagawa-ken, 259-1220 JAPAN
Telephone: +81 463 50 6400 / Facsimile: +81 463 50 6401

Appendix 2-2: SAR measurement data / SAR test data plot of other test conditions / Step 8: W58 band (Head) (cont'd)

Plot 8-9: (Head) Antenna Sub: Short side-sub & touch, 11n(40HT) (MCS0), 5755 MHz

EUT: SKR3000; Type: P-61; Serial: A8CE-S002

Mode: n40(MCS0,BPSK/OFDM)(UID 0, Frame Length in ms: 0; PAR: 0; PMF: 1); Frequency: 5755 MHz; Crest Factor: 1.0

Medium: HSL5GHz(1608); Medium parameters used: $f = 5755$ MHz; $\sigma = 5.018$ S/m; $\epsilon_r = 35.37$; $\rho = 1000$ kg/m³

Measurement Standard: DASYS5 (IEEE/IEC/ANSI C63.19-2007)

DASY Configuration: -Probe: EX3DV4 - SN3907; ConvF(4.3, 4.3, 4.3); Calibrated: 2016/05/12; -DASY52 52.8.8(1222); SEMCAD X 14.6.10(7331)
-Sensor-Surface: 1.4mm (Mechanical Surface Detection), z= 1.0, 25.0 -Electronics: DAE4 Sn626; Calibrated: 2015/09/15
-Phantom: ELL v4.0; Type: QDOVA001BA; Serial: 1059; Phantom section: Flat Section

head,w58,ant-sub(chain1)/5g60w58h8,bw40;ant1,side&d0,n40(m0,p12),h5755/


Area Scan:60x140,stp10 (7x15x1): Measurement grid: dx=10mm, dy=10mm; Maximum value of SAR (measured) = 0.733 W/kg

Area Scan:60x140,stp10 (61x141x1): Interpolated grid: dx=1.000 mm, dy=1.000 mm; Maximum value of SAR (interpolated) = 0.735 W/kg

Zoom:28x28x24,xy4-z1.4(ratio) (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm;

Reference Value = 13.32 V/m; Power Drift = -0.03 dB; Maximum value of SAR (measured) = 0.689 W/kg; Peak SAR (extrapolated) = 1.68 W/kg

SAR(1 g) = 0.238 W/kg; SAR(10 g) = 0.054 W/kg



Remarks: * Date tested: 2016/08/04; Tested by: Hiroshi Naka; Tested place: No.7 shielded room,
* liquid depth: 153 mm; Position: distance of EUT to phantom: 0 mm (2 mm to liquid); ambient: 24 ± 1 deg.C. / 50 ± 10 %RH,
* liquid temperature: 22.9(start)/22.9(end)/22.8(in check) deg.C.; * White cubic: zoom scan area, Red cubic: big=SAR(10g)/small=SAR(1g)

Plot 8-10: (Head) Antenna Sub: Short side-sub & touch, 11n(40HT) (MCS0), 5795 MHz

EUT: SKR3000; Type: P-61; Serial: A8CE-S002

Mode: n40(MCS0,BPSK/OFDM)(UID 0, Frame Length in ms: 0; PAR: 0; PMF: 1); Frequency: 5795 MHz; Crest Factor: 1.0

Medium: HSL5GHz(1608); Medium parameters used: $f = 5795$ MHz; $\sigma = 5.038$ S/m; $\epsilon_r = 35.32$; $\rho = 1000$ kg/m³

Measurement Standard: DASYS5 (IEEE/IEC/ANSI C63.19-2007)

DASY Configuration: -Probe: EX3DV4 - SN3907; ConvF(4.3, 4.3, 4.3); Calibrated: 2016/05/12; -DASY52 52.8.8(1222); SEMCAD X 14.6.10(7331)
-Sensor-Surface: 1.4mm (Mechanical Surface Detection), z= 1.0, 25.0 -Electronics: DAE4 Sn626; Calibrated: 2015/09/15
-Phantom: ELL v4.0; Type: QDOVA001BA; Serial: 1059; Phantom section: Flat Section

head,w58,ant-sub(chain1)/5g61w58h9,bw40;ch;ant1,side&d0,n40(m0,p12),h5795/


Area Scan:40x100,stp10 (5x11x1): Measurement grid: dx=10mm, dy=10mm; Maximum value of SAR (measured) = 0.576 W/kg

Area Scan:40x100,stp10 (41x101x1): Interpolated grid: dx=1.000 mm, dy=1.000 mm; Maximum value of SAR (interpolated) = 0.579 W/kg

Zoom:28x28x24,xy4-z1.4(ratio) (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm;

Reference Value = 11.91 V/m; Power Drift = -0.02 dB; Maximum value of SAR (measured) = 0.544 W/kg; Peak SAR (extrapolated) = 0.913 W/kg

SAR(1 g) = 0.184 W/kg; SAR(10 g) = 0.042 W/kg



Remarks: * Date tested: 2016/08/04; Tested by: Hiroshi Naka; Tested place: No.7 shielded room,
* liquid depth: 153 mm; Position: distance of EUT to phantom: 0 mm (2 mm to liquid); ambient: 24 ± 1 deg.C. / 50 ± 10 %RH,
* liquid temperature: 22.9(start)/22.9(end)/22.8(in check) deg.C.; * White cubic: zoom scan area, Red cubic: big=SAR(10g)/small=SAR(1g)

Appendix 2-2: SAR measurement data / SAR test data plot of other test conditions / Step 8: W58 band (Head) (cont'd)

Plot 8-11: (Head) Antenna Main; Long side-main & touch, 11a(6Mbps), 5785 MHz

EUT: SKR3000; Type: P-61; Serial: A8CE-S002

Mode: 11a(6Mbps,BPSK/OFDM)(UID 0, Frame Length in ms: 0; PAR: 0; PMF: 1); Frequency: 5785 MHz; Crest Factor: 1.0

Medium: HSL5GHz(1608); Medium parameters used: $f = 5785$ MHz; $\sigma = 5.02$ S/m; $\epsilon_r = 35.32$; $\rho = 1000$ kg/m³

Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007)

DASY Configuration: -Probe: EX3DV4 - SN3907; ConvF(4.3, 4.3, 4.3); Calibrated: 2016/05/12; -Electronics: DAE4 Sn626; Calibrated: 2015/09/15
-Sensor-Surface: 1.4mm (Mechanical Surface Detection (Locations From Previous Scan Used)), Sensor-Surface: 1.4mm (Mechanical Surface Detection), z = 1.0, 25.0
-Phantom: ELI v4.0; Type: QDOVA001BA; Serial: 1059; Phantom section: Flat Section -DASY52 52.8.8(1222); SEMCAD X 14.6.10(7331)

head,w58,ant-main(chain0)/5g56&w58h4,mode2/ch;ant0,side&d0,a(6m,p12),h5785/


Area Scan:100x40,stp10 (11x5x1): Measurement grid: dx=10mm, dy=10mm; Maximum value of SAR (measured) = 1.43 W/kg

Area Scan:100x40,stp10 (101x41x1): Interpolated grid: dx=1.000 mm, dy=1.000 mm; Maximum value of SAR (interpolated) = 1.59 W/kg

Zoom:28x28x24,xy4-z1.4(ratio) (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm;

Reference Value = 22.99 V/m; Power Drift = -0.10 dB; Maximum value of SAR (measured) = 2.29 W/kg; Peak SAR (extrapolated) = 4.25 W/kg

SAR(1 g) = 0.692 W/kg; SAR(10 g) = 0.137 W/kg



Remarks: * Date tested: 2016/08/04; Tested by: Hiroshi Naka; Tested place: No.7 shielded room,
* liquid depth: 153 mm; Position: distance of EUT to phantom: 0 mm (2 mm to liquid); ambient: 24 ± 1 deg.C. / 50 ± 10 %RH,
* liquid temperature: 22.8(start)/22.9(end)/22.8(in check) deg.C.; * White cubic: zoom scan area, Red cubic: big=SAR(10g)/small=SAR(1g)

Plot 8-12: (Head) Antenna Main; Long side-main & touch, 11a(6Mbps), 5825 MHz

EUT: SKR3000; Type: P-61; Serial: A8CE-S002

Mode: 11a(6Mbps,BPSK/OFDM)(UID 0, Frame Length in ms: 0; PAR: 0; PMF: 1); Frequency: 5825 MHz; Crest Factor: 1.0

Medium: HSL5GHz(1608); Medium parameters used: $f = 5825$ MHz; $\sigma = 5.084$ S/m; $\epsilon_r = 35.30$; $\rho = 1000$ kg/m³

Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007)

DASY Configuration: -Probe: EX3DV4 - SN3907; ConvF(4.3, 4.3, 4.3); Calibrated: 2016/05/12; -Electronics: DAE4 Sn626; Calibrated: 2015/09/15
-Sensor-Surface: 1.4mm (Mechanical Surface Detection (Locations From Previous Scan Used)), Sensor-Surface: 1.4mm (Mechanical Surface Detection), z = 1.0, 25.0
-Phantom: ELI v4.0; Type: QDOVA001BA; Serial: 1059; Phantom section: Flat Section -DASY52 52.8.8(1222); SEMCAD X 14.6.10(7331)

head,w58,ant-main(chain0)/5g57&w58h5,mode2/ch;ant0,side&d0,a(6m,p12),h5825/


Area Scan:100x40,stp10 (11x5x1): Measurement grid: dx=10mm, dy=10mm; Maximum value of SAR (measured) = 1.12 W/kg

Area Scan:100x40,stp10 (101x41x1): Interpolated grid: dx=1.000 mm, dy=1.000 mm; Maximum value of SAR (interpolated) = 1.26 W/kg

Zoom:28x28x24,xy4-z1.4(ratio) (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm;

Reference Value = 20.37 V/m; Power Drift = -0.09 dB; Maximum value of SAR (measured) = 1.79 W/kg; Peak SAR (extrapolated) = 3.34 W/kg

SAR(1 g) = 0.542 W/kg; SAR(10 g) = 0.107 W/kg



Remarks: * Date tested: 2016/08/04; Tested by: Hiroshi Naka; Tested place: No.7 shielded room,
* liquid depth: 153 mm; Position: distance of EUT to phantom: 0 mm (2 mm to liquid); ambient: 24 ± 1 deg.C. / 50 ± 10 %RH,
* liquid temperature: 22.9(start)/22.9(end)/22.8(in check) deg.C.; * White cubic: zoom scan area, Red cubic: big=SAR(10g)/small=SAR(1g)

Appendix 2-2: SAR measurement data / SAR test data plot of other test conditions / Step 8: W58 band (Head) (cont'd)

Plot 8-13: (Head) Antenna Sub: Short side-sub & touch, 11a(6Mbps), 5785 MHz

EUT: SKR3000; Type: P-61; Serial: A8CE-S002

Mode: 11a(6Mbps,BPSK/OFDM)(UID 0, Frame Length in ms: 0; PAR: 0; PMF: 1); Frequency: 5785 MHz; Crest Factor: 1.0

Medium: HSL5GHz(1608); Medium parameters used: $f = 5785$ MHz; $\sigma = 5.02$ S/m; $\epsilon_r = 35.32$; $\rho = 1000$ kg/m³

Measurement Standard: DASYS5 (IEEE/IEC/ANSI C63.19-2007)

DASY Configuration: -Probe: EX3DV4 - SN3907; ConvF(4.3, 4.3, 4.3); Calibrated: 2016/05/12; -Electronics: DAE4 Sn626; Calibrated: 2015/09/15
-Sensor-Surface: 1.4mm (Mechanical Surface Detection (Locations From Previous Scan Used)), Sensor-Surface: 1.4mm (Mechanical Surface Detection), z = 1.0, 25.0
-Phantom: ELI v4.0; Type: QDOVA001BA; Serial: 1059; Phantom section: Flat Section -DASY52 52.8.8(1222); SEMCAD X 14.6.10(7331)

head,w58,ant-sub(chain1)/5g63w58h11,mode2/ch;ant1,side&d0,a(6m,p12),h5785/


Area Scan:40x100,stp10 (5x11x1): Measurement grid: dx=10mm, dy=10mm; Maximum value of SAR (measured) = 0.658 W/kg

Area Scan:40x100,stp10 (41x101x1): Interpolated grid: dx=1.000 mm, dy=1.000 mm; Maximum value of SAR (interpolated) = 0.721 W/kg

Zoom:28x28x24,xy4-z1.4(ratio) (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm;

Reference Value = 12.80 V/m; Power Drift = -0.07 dB; Maximum value of SAR (measured) = 0.604 W/kg; Peak SAR (extrapolated) = 2.18 W/kg

SAR(1 g) = 0.214 W/kg; SAR(10 g) = 0.048 W/kg



Remarks: * Date tested: 2016/08/04; Tested by: Hiroshi Naka; Tested place: No.7 shielded room,
* liquid depth: 153 mm; Position: distance of EUT to phantom: 0 mm (2 mm to liquid); ambient: 24 ± 1 deg.C. / 50 ± 10 %RH,
* liquid temperature: 22.9(start)/22.9(end)/22.8(in check) deg.C.; * White cubic: zoom scan area, Red cubic: big=SAR(10g)/small=SAR(1g)

Plot 8-14: (Head) Antenna Sub: Short side-sub & touch, 11a(6Mbps), 5825 MHz

EUT: SKR3000; Type: P-61; Serial: A8CE-S002

Mode: 11a(6Mbps,BPSK/OFDM)(UID 0, Frame Length in ms: 0; PAR: 0; PMF: 1); Frequency: 5825 MHz; Crest Factor: 1.0

Medium: HSL5GHz(1608); Medium parameters used: $f = 5825$ MHz; $\sigma = 5.084$ S/m; $\epsilon_r = 35.30$; $\rho = 1000$ kg/m³

Measurement Standard: DASYS5 (IEEE/IEC/ANSI C63.19-2007)

DASY Configuration: -Probe: EX3DV4 - SN3907; ConvF(4.3, 4.3, 4.3); Calibrated: 2016/05/12; -Electronics: DAE4 Sn626; Calibrated: 2015/09/15
-Sensor-Surface: 1.4mm (Mechanical Surface Detection (Locations From Previous Scan Used)), Sensor-Surface: 1.4mm (Mechanical Surface Detection), z = 1.0, 25.0
-Phantom: ELI v4.0; Type: QDOVA001BA; Serial: 1059; Phantom section: Flat Section -DASY52 52.8.8(1222); SEMCAD X 14.6.10(7331)

head,w58,ant-sub(chain1)/5g64w58h12,mode2/ch;ant1,side&d0,a(6m,p12),h5825/


Area Scan:40x100,stp10 (5x11x1): Measurement grid: dx=10mm, dy=10mm; Maximum value of SAR (measured) = 0.695 W/kg

Area Scan:40x100,stp10 (41x101x1): Interpolated grid: dx=1.000 mm, dy=1.000 mm; Maximum value of SAR (interpolated) = 0.758 W/kg

Zoom:28x28x24,xy4-z1.4(ratio) (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm;

Reference Value = 13.05 V/m; Power Drift = -0.08 dB; Maximum value of SAR (measured) = 0.650 W/kg; Peak SAR (extrapolated) = 1.10 W/kg

SAR(1 g) = 0.227 W/kg; SAR(10 g) = 0.053 W/kg



Remarks: * Date tested: 2016/08/04; Tested by: Hiroshi Naka; Tested place: No.7 shielded room,
* liquid depth: 153 mm; Position: distance of EUT to phantom: 0 mm (2 mm to liquid); ambient: 24 ± 1 deg.C. / 50 ± 10 %RH,
* liquid temperature: 22.9(start)/23.0(end)/22.8(in check) deg.C.; * White cubic: zoom scan area, Red cubic: big=SAR(10g)/small=SAR(1g)

Appendix 2-2: SAR measurement data / SAR test data plot of other test conditions (cont'd)

Step 9: SAR Measurement Variability

Plot 9-1: (Body) Antenna Main; Long side-main & touch, 11a(6Mbps), 5260 MHz

EUT: SKR3000; Type: P-61; Serial: A8CE-S002

Mode: 11a(6Mbps,BPSK/OFDM)(UID 0, Frame Length in ms: 0; PAR: 0; PMF: 1); Frequency: 5260 MHz; Crest Factor: 1.0

Medium: MSL5800(1607); Medium parameters used: $f = 5260$ MHz; $\sigma = 5.416$ S/m; $\epsilon_r = 47.28$; $\rho = 1000$ kg/m³

Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007)

DASY Configuration: -Probe: EX3DV4 - SN3907; ConvF(4.37, 4.37, 4.37); Calibrated: 2016/05/12; -DASY52 52.8.8(1222); SEMCAD X 14.6.10(7331)
 -Sensor-Surface: 1.4mm (Mechanical Surface Detection), z = 1.0, 25.0 -Electronics: DAE4 Sn626; Calibrated: 2015/09/15
 -Phantom: ELI v4.0; Type: QDOVA001BA; Serial: 1059; Phantom section: Flat Section

body-touch,w53,ant-main(chain0)/Repeat.max.measure/5g51w53b13/Re;5g44&w53b6,ant0,side&d0,a(6mp12),b5260/


Area Scan:100x40,stp10 (11x5x1): Measurement grid: dx=10mm, dy=10mm; Maximum value of SAR (measured) = 2.36 W/kg

Area Scan:100x40,stp10 (101x41x1): Interpolated grid: dx=1.000 mm, dy=1.000 mm; Maximum value of SAR (interpolated) = 3.18 W/kg

Zoom:28x28x24,xy4-z1.4(ratio) (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm;

Reference Value = 26.76 V/m; Power Drift = -0.13 dB; Maximum value of SAR (measured) = 2.79 W/kg; Peak SAR (extrapolated) = 5.67 W/kg

SAR(1 g) = 0.996 W/kg; SAR(10 g) = 0.213 W/kg



Remarks: * Date tested: 2016/07/29; Tested by: Hiroshi Naka; Tested place: No.7 shielded room,
 * liquid depth: 152 mm; Position: distance of EUT to phantom: 0 mm (2 mm to liquid); ambient: 24~25 deg.C. / 50 ± 10 %RH,
 * liquid temperature: 23.6(start)/23.6(end)/23.7(in check) deg.C.; * White cubic: zoom scan area, Red cubic: big=SAR(10g)/small=SAR(1g)

Plot 9-2: (Head) Antenna Main; Long side-main & touch, 11a(6Mbps), 5260 MHz

EUT: SKR3000; Type: P-61; Serial: A8CE-S002

Mode: 11a(6Mbps,BPSK/OFDM)(UID 0, Frame Length in ms: 0; PAR: 0; PMF: 1); Frequency: 5260 MHz; Crest Factor: 1.0

Medium: HSL5GHz(1608); Medium parameters used: $f = 5260$ MHz; $\sigma = 4.501$ S/m; $\epsilon_r = 36.07$; $\rho = 1000$ kg/m³

Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007)

DASY Configuration: -Probe: EX3DV4 - SN3907; ConvF(4.94, 4.94, 4.94); Calibrated: 2016/05/12; -DASY52 52.8.8(1222); SEMCAD X 14.6.10(7331)
 -Sensor-Surface: 1.4mm (Mechanical Surface Detection), z = 1.0, 25.0 -Electronics: DAE4 Sn626; Calibrated: 2015/09/15
 -Phantom: ELI v4.0; Type: QDOVA001BA; Serial: 1059; Phantom section: Flat Section

head,w53,ant-main(chain0)/Repeat.max.measure/5g11w53h11/Re;5g3w53h3,ant0,side&d0,a(6mp12),h5260/


Area Scan:100x40,stp10 (11x5x1): Measurement grid: dx=10mm, dy=10mm; Maximum value of SAR (measured) = 2.53 W/kg

Area Scan:100x40,stp10 (101x41x1): Interpolated grid: dx=1.000 mm, dy=1.000 mm; Maximum value of SAR (interpolated) = 2.89 W/kg

Zoom:28x28x24,xy4-z1.4(ratio) (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm;

Reference Value = 27.48 V/m; Power Drift = -0.11 dB; Maximum value of SAR (measured) = 3.16 W/kg; Peak SAR (extrapolated) = 5.94 W/kg

SAR(1 g) = 1.04 W/kg; SAR(10 g) = 0.231 W/kg



Remarks: * Date tested: 2016/08/01; Tested by: Hiroshi Naka; Tested place: No.7 shielded room,
 * liquid depth: 153 mm; Position: distance of EUT to phantom: 0 mm (2 mm to liquid); ambient: 24 ± 1 deg.C. / 50 ± 10 %RH,
 * liquid temperature: 23.0(start)/23.0(end)/22.8(in check) deg.C.; * White cubic: zoom scan area, Red cubic: big=SAR(10g)/small=SAR(1g)

APPENDIX 3: Test instruments

Appendix 3-1: Equipment used

| Control No. | Instrument | Manufacturer | Model No | Serial No | Test Item | Calibration Date * Interval(month) |
|--------------|---|-------------------------------|---------------------------------|-------------------------|------------|------------------------------------|
| KPM-08 | Power meter | Anritsu | ML2495A | 6K00003356 | AT | 2015/09/09 * 12 |
| KPSS-04 | Power sensor | Anritsu | MA2411B | 012088 | AT | 2015/09/09 * 12 |
| KAT10-S3 | Attenuator | Agilent | 8490D 010 | 50924 | AT | 2015/12/24 * 12 |
| KTM-G1 | Terminator | Hirose Electric | HRM-TMP-05(40) | - | AT | - |
| COTS-SSAR-02 | DASY52 | Schmid&Partner Engineering AG | DASY52(ver.52.8.8(1222)) | - | SAR | - |
| COTS-SSEP-02 | Dielectric assessment kit | Schmid&Partner Engineering AG | DAK(ver1.10.317.11) | - | SAR(daily) | - |
| SSAR-02 | SAR measurement system | Schmid&Partner Engineering AG | DASY5 | 1324 | SAR | Pre Check |
| SSRBT-02 | SAR robot | Schmid&Partner Engineering AG | TX60 Lspeag | F12/5L2QA1/A/01 | SAR | 2015/09/10 * 12 |
| KDAE-01 | Data Acquisition Electronics | Schmid&Partner Engineering AG | DAE4 | 626 | SAR | 2015/09/15 * 12 |
| SPB-02 | Dosimetric E-Field Probe | Schmid&Partner Engineering AG | EX3DV4 | 3907 | SAR | 2016/05/12 * 12 |
| KSDA-01 | Dipole Antenna | Schmid&Partner Engineering AG | D2450V2 | 822 | SAR(daily) | 2016/01/14 * 12 |
| KSDA-02 | Dipole Antenna | Schmid&Partner Engineering AG | D5GHzV2 | 1070 | SAR(daily) | 2016/03/10 * 12 |
| KPFL-01 | Flat Phantom | Schmid&Partner Engineering AG | Oval flat phantom ELI 4.0 | 1059 | SAR | 2015/08/06 * 12 |
| SSNA-01 | Network Analyzer | Agilent | 8753ES | US39171777 | SAR(daily) | 2015/12/24 * 12 |
| SEPP-R03 | Dielectric probe | Schmid&Partner Engineering AG | DAK3.5 | 1191 | SAR(daily) | 2016/05/10 * 12 |
| KSG-08 | Signal Generator | Rohde & Schwarz | SMT06 | 100763 | SAR(daily) | 2015/07/02 * 12 |
| SSG-01 | Signal Generator | Agilent | E4438C | MY47271584 | SAR(daily) | 2016/03/24 * 12 |
| KPA-12 | RF Power Amplifier | MILMEGA | AS2560-50 | 1018582 | SAR(daily) | Pre Check |
| KCPL-07 | Directional Coupler | Pulsar Microwave Corp. | CCS30-B26 | 0621 | SAR(daily) | Pre Check |
| KPM-06 | Power Meter | Rohde & Schwarz | NRVD | 101599 | SAR(daily) | 2015/09/08 * 12 |
| KIU-08 | Power sensor | Rohde & Schwarz | NRV-24 | 100372 | SAR(daily) | 2015/09/08 * 12 |
| KIU-09 | Power sensor | Rohde & Schwarz | NRV-24 | 100371 | SAR(daily) | 2015/09/08 * 12 |
| KAT10-P1 | Attenuator | Weinschel | 24-10-34 | BY5927 | SAR(daily) | 2015/12/24 * 12 |
| KPM-05 | Power meter | Agilent | E4417A | GB41290718 | SAR(daily) | 2016/04/13 * 12 |
| KPSS-01 | Power sensor | Agilent | E9327A | US40440544 | SAR(daily) | 2016/04/13 * 12 |
| SAT20-SAR1 | Attenuator | TME | SFA-01AXPJ-20 | - | SAR(daily) | 2015/12/24 * 12 |
| SCC-SAR2 | Coaxial Cable | HUBER+SUHNER | SF104A/11PC3542/11N451/4M | MY699/4A | SAR(daily) | Pre Check |
| KRU-01 | Ruler(300mm) | Shinwa | 13134 | - | SAR | 2016/02/24 * 12 |
| KRU-02 | Ruler(150mm,L) | Shinwa | 12103 | - | SAR | 2016/02/24 * 12 |
| KRU-04 | Ruler(300mm) | Shinwa | 13134 | - | SAR | 2016/05/16 * 12 |
| KOS-13 | Digital thermometer | HANNA | Checktemp-2 | KOS-13 | SAR | 2015/12/07 * 12 |
| KOS-14 | Thermo-Hygrometer data logger | SATO KEIRYOKI | SK-L20THII α / SK-LTHII α-2 | 015246/08169 | SAR | 2015/12/07 * 12 |
| SOS-11 | Humidity Indicator | A&D | AD-5681 | 4063424 | SAR(daily) | 2015/12/07 * 12 |
| SOS-12 | Digital thermometer | HANNA | Checktemp-4 | SOS-12 | SAR(daily) | 2016/02/24 * 12 |
| SOS-SAR1 | Digital thermometer | LKMelectronic | DTM3000 | 3171 | SAR(daily) | 2015/10/20 * 12 |
| SSA-04 | Spectrum Analyzer | Advantest | R3272 | 101100994 | SAR(moni.) | Pre Check |
| KSDH-01 | Device holder | Schmid&Partner Engineering AG | Mounting device for transmitter | - | SAR | 2015/09/10 * 12 |
| SSDH-02 | Laptop holder | Schmid&Partner Engineering AG | SM LHI 001 C | - | SAR | Pre Check |
| SWTR-03 | DI water | MonotaRo | 34557433 | - | SAR | Pre Check |
| SALC-01 | Primepure Ethanol | Kanto Chemical Co., Inc. | 14032-79 | - | SAR(daily) | Pre Check |
| KSLH245-01 | Tissue simulation liquid (2450MHz,head) | Schmid&Partner Engineering AG | HSL2450V2 | SL AAH 245 BA | SAR | Pre Check |
| KSLM245-01 | Tissue simulation liquid (2450MHz,body) | Schmid&Partner Engineering AG | MSL2450V2 | SL AAM 245 BA | SAR | Pre Check |
| KSLM580-02 | Tissue simulation liquid (5800MHz,body) | Schmid&Partner Engineering AG | MBSL3500-5800V5 | SL AAM 501 AB(110520-3) | SAR | Pre Check |
| KSLH580-04 | Tissue simulation liquid (5800MHz,head) | Schmid&Partner Engineering AG | HBBL3500-5800V5 | SL AAH 502 AD(140305-1) | SAR | Pre Check |

The expiration date of calibration is the end of the expired month.


As for some calibrations performed after the tested dates, those test equipment have been controlled by means of an unbroken chains of calibrations.

All equipment is calibrated with valid calibrations. Each measurement data is traceable to the national or international standards.

[Test Item] SAR: Specific Absorption Rate, AT: Antenna terminal conducted power

Appendix 3-2: Configuration and peripherals

These measurements were performed with the automated near-field scanning system DASY5 from Schmid & Partner Engineering AG (SPEAG). The system is based on a high precision robot, which positions the probes with a positional repeatability of better than ± 0.02 mm. Special E- and H-field probes have been developed for measurements close to material discontinuity, the sensors of which are directly loaded with a Schottky diode and connected via highly resistive lines to the data acquisition unit. The SAR measurements were conducted with the dosimetry probes EX3DV4 (manufactured by SPEAG), designed in the classical triangular configuration and optimized for dosimetric evaluation.



The DASY5 system for performing compliance tests consist of the following items:

| | |
|----|---|
| 1 | A standard high precision 6-axis robot (Stäubli TX/RX family) with controller, teach pendant and software. An arm extension for accommodating the data acquisition electronics (DAE). |
| 2 | An isotropic field probe optimized and calibrated for the targeted measurement. |
| 3 | A data acquisition electronics (DAE) which performs the signal amplification, signal multiplexing, AD-conversion, offset measurements, mechanical surface detection, collision detection, etc. The unit is battery powered with standard or rechargeable batteries. The signal is optically transmitted to the EOC. |
| 4 | The Electro-optical converter (EOC) performs the conversion from optical to electrical signals for the digital communication to the DAE. To use optical surface detection, a special version of the EOC is required. The EOC signal is transmitted to the measurement server. |
| 5 | The function of the measurement server is to perform the time critical tasks such as signal filtering, control of the robot operation and fast movement interrupts. |
| 6 | The Light Beam used is for probe alignment. This improves the (absolute) accuracy of the probe positioning. |
| 7 | A computer running Win7 professional operating system and the DASY5 software. |
| 8 | R Remote control and teach pendant as well as additional circuitry for robot safety such as warning lamps, etc. |
| 9 | The phantom. |
| 10 | The device holder for EUT. (low-loss dielectric palette) (*. when it was used.) |
| 11 | Tissue simulating liquid mixed according to the given recipes. |
| 12 | Validation dipole kits allowing to validate the proper functioning of the system. |

Appendix 3-3: Test system specification

TX60 Lspeag robot/CS8Cspeag-TX60 robot controller

- Number of Axes : 6
- Repeatability : $\pm 0.02\text{mm}$
- Manufacture : Stäubli Unimation Corp.

DASY5 Measurement server

- Features : The DASY5 measurement server is based on a PC/104 CPU board with a 400MHz intel ULV Celeron, 128MB chip-disk and 128MB RAM. The necessary circuits for communication with the DAE4 electronics box, as well as the 16 bit AD converter system for optical detection and digital I/O interface are contained on the DASY5 I/O board, which is directly connected to the PC/104 bus of the CPU board.
- Calibration : No calibration required.
- Manufacture : Schmid & Partner Engineering AG

Data Acquisition Electronic (DAE)

- Features : Signal amplifier, multiplexer, A/D converter and control logic. Serial optical link for communication with DASY5 embedded system (fully remote controlled). 2 step probe touch detector for mechanical surface detection and emergency robot stop (not in -R version)
- Measurement Range : $1\mu\text{V}$ to $> 200\text{mV}$ (16bit resolution and 2 range settings: 4mV, 400mV)
- Input Offset voltage : $< 1\mu\text{V}$ (with auto zero)
- Input Resistance : $200\text{M}\Omega$
- Battery Power : $> 10\text{hr}$ of operation (with two 9V battery)
- Manufacture : Schmid & Partner Engineering AG

Electro-Optical Converter (EOC61)

- Manufacture : Schmid & Partner Engineering AG

Light Beam Switch (LB5/80)

- Manufacture : Schmid & Partner Engineering AG

SAR measurement software

- Item : Dosimetric Assessment System DASY5
- Software version : DASY52, V8.2 B969
- Manufacture : Schmid & Partner Engineering AG

E-Field Probe


- Model : **EX3DV4 (serial number: 3679)**
- Construction : Symmetrical design with triangular core. Built-in shielding against static charges. PEEK enclosure material (resistant to organic solvents, e.g., DGBE).
- Frequency : 10MHz to 6GHz, Linearity: $\pm 0.2\text{ dB}$ (30MHz to 6GHz)
- Conversion Factors : 2.45, 5.2, 5.25, 5.5, 5.6, 5.75, 5.8 GHz (Head)
2.45, 5.25, 5.6, 5.75 GHz (Body)
- Directivity : $\pm 0.3\text{ dB}$ in HSL (rotation around probe axis)
 $\pm 0.5\text{ dB}$ in tissue material (rotation normal to probe axis)
- Dynamic Range : $10\mu\text{W/g}$ to $> 100\text{ mW/g}$; Linearity: $\pm 0.2\text{ dB}$ (noise: typically $< 1\mu\text{W/g}$)
- Dimension : Overall length: 330mm (Tip: 20mm)
Tip diameter: 2.5mm (Body: 12mm)
Typical distance from probe tip to dipole centers: 1mm
- Application : High precision dosimetric measurement in any exposure scenario (e.g., very strong gradient fields). Only probe which enables compliance testing for frequencies up to 6GHz with precision of better 30%.
- Manufacture : Schmid & Partner Engineering AG

Phantom

- Type : **ELI 4.0 oval flat phantom**
- Shell Material : Fiberglass
- Shell Thickness : Bottom plate: $2 \pm 0.2\text{mm}$
- Dimensions : Bottom elliptical: 600x400mm, Depth: 190mm (Volume: Approx. 30 liters)
- Manufacture : Schmid & Partner Engineering AG

Device Holder

- Urethane foam
- KSDH-01: In combination with the ELI4, the Mounting Device enables the rotation of the mounted transmitter device in spherical coordinates. Transmitter devices can be easily and accurately positioned. The low-loss dielectric urethane foam was used for the mounting section of device holder.
 - Material : POM
 - Manufacture : Schmid & Partner Engineering AG
- SSDH-02: Device holder for the laptop computer.



Appendix 3-4: Simulated tissue composition and parameter confirmation

| Liquid type | Body | Body | Head | Head |
|-------------------------|--|--|--|---|
| Control No. | KSLM245-01 | KSLM580-02 | KSLH245-01 | KSLH580-04 |
| Model No. / Product No. | MSL2450V2 / SL AAM 245 BA | MBBL3500-5800V5 / SL AAM 501 AB | HSL2450V2 / SL AAH 245 BA | HBBL 3500-5800V5 / SL AAH 502 AD |
| Ingredient: Mixture(%) | Water: 52-75%, DGBE: 25-48%, NaCl: <1.0% | Water: 60-80%, Ester/Emulsifiers/Inhibitors: 20-40%, Sodium salt: 0-1.5% | Water: 52-75%, DGBE: 25-48%, NaCl: <1.0% | Water: 50-65%, Mineral oil: 10-30%, Emulsifiers: 8-25%, Sodium salt: 0-1.5% |
| Manufacture | Schmid & Partner Engineering AG | | | |

*. The dielectric parameters were checked prior to assessment using the DAK3.5 dielectric probe kit.

| Measured date | Freq. [MHz] | Liquid type | Ambient [deg.C.] [%RH] | Liquid temp. [deg.C.] | Liquid Depth [mm] | Liquid parameters (*a) | | | | | | | ASAR (1g) [%] (*b) | |
|---------------------|-------------|-------------|------------------------|-----------------------|-------------------|------------------------|----------|------|-------|--------------------|----------|------|--------------------|-------|
| | | | | | | Permittivity (εr) [-] | | | | Conductivity [S/m] | | | | |
| | | | | | | Target | Measured | | Limit | Target | Measured | | | Limit |
| July 25, 2016 | 2450 | Head | 24.4/49 | 23.8 | (153) | 39.2 | 38.20 | -2.6 | ±5% | 1.80 | 1.860 | +3.3 | ±5% | +2.17 |
| July 26, 2016 | 2450 | Body | 23.6/48 | 22.5 | (153) | 52.7 | 50.87 | -3.5 | ±5% | 1.95 | 1.975 | +1.3 | ±5% | +1.39 |
| July 27, 2016 (*1) | 5600 | Body | 24.0/57 | 23.7 | (152) | 48.47 | 46.98 | -3.1 | ±5% | 5.766 | 5.904 | +2.4 | ±5% | +0.51 |
| July 28, 2016 (*1) | 5600 | Body | 24.0/57 | 23.7 | (152) | 48.47 | 46.98 | -3.1 | ±5% | 5.766 | 5.904 | +2.4 | ±5% | +0.51 |
| July 28, 2016 | 5750 | Body | 23.9/59 | 23.7 | (152) | 48.27 | 46.52 | -3.6 | ±5% | 5.942 | 6.081 | +2.4 | ±5% | +0.61 |
| July 29, 2016 | 5250 | Body | 23.9/59 | 23.7 | (152) | 48.95 | 47.33 | -3.3 | ±5% | 5.358 | 5.454 | +1.8 | ±5% | +0.60 |
| August 1, 2016 | 5250 | Head | 23.4/49 | 22.8 | (153) | 35.93 | 36.03 | +0.3 | ±5% | 4.706 | 4.489 | -4.6 | ±5% | +0.09 |
| August 2, 2016 (*2) | 5250 | Head | 23.4/49 | 22.8 | (153) | 35.93 | 36.03 | +0.3 | ±5% | 4.706 | 4.489 | -4.6 | ±5% | +0.09 |
| August 2, 2016 | 5600 | Head | 23.4/49 | 22.8 | (153) | 35.53 | 35.67 | +0.4 | ±5% | 5.065 | 4.832 | -4.6 | ±5% | +0.12 |
| August 3, 2016 (*3) | 5600 | Head | 23.4/49 | 22.8 | (153) | 35.53 | 35.67 | +0.4 | ±5% | 5.065 | 4.832 | -4.6 | ±5% | +0.12 |
| August 4, 2016 | 5800 | Head | 23.4/49 | 22.8 | (153) | 35.3 | 35.26 | -0.1 | ±5% | 5.27 | 5.056 | -4.1 | ±5% | +0.21 |

*1. It was within 24 hours from measurement on July 27, 2016 and same liquid temperature, so measured parameters on June 27 were used on June 28 continuously.

*2. It was within 24 hours from measurement on August 1, 2016 and same liquid temperature, so measured parameters on August 1 were used on August 2 continuously.

*3. It was within 24 hours from measurement on August 2, 2016 and same liquid temperature, so measured parameters on August 2 were used on August 3 continuously.

a. The target value is a parameter defined in Appendix A of KDB865664 D01 (v01r04), the dielectric parameters suggested for head and body tissue simulating liquid are given at 2000, 2450, 3000 and 5800MHz. (. The parameters of the head liquid are the same value as IEC 62209-2.) Parameters for the frequencies between 2000-3000, 3000-5800MHz were obtained using linear interpolation. Above 5800MHz were obtained using linear extrapolation.

| f (MHz) | Standard | | | | | | | | Interpolated & Extrapolated | | | | | | | | | | |
|-------------|-------------|---------|-------------|---------|---------|-------------|---------|-------------|-----------------------------|---------|-------------|---------|-------------|---------|------|---------|---------|-------|-------|
| | Head Tissue | | Body Tissue | | f (MHz) | Head Tissue | | Body Tissue | | f (MHz) | Head Tissue | | Body Tissue | | | | | | |
| | εr | σ [S/m] | εr | σ [S/m] | | εr | σ [S/m] | εr | σ [S/m] | | εr | σ [S/m] | εr | σ [S/m] | | | | | |
| (1800-)2000 | 40.0 | 1.40 | 53.3 | 1.52 | 3000 | 38.5 | 2.40 | 52.0 | 2.73 | 5250 | 35.93 | 4.706 | 48.95 | 5.358 | 5750 | not use | not use | 48.27 | 5.942 |
| 2450 | 39.2 | 1.80 | 52.7 | 1.95 | 5800 | 35.3 | 5.27 | 48.2 | 6.00 | 5600 | 35.53 | 5.065 | 48.47 | 5.766 | | | | | |

*b. The coefficients are parameters defined in IEEE Std 1528-2013.

$$ASAR(1g) = C_{\epsilon r} \times \Delta \epsilon r + C_{\sigma} \times \Delta \sigma, C_{\epsilon r} = -7.854E-4 \times f^3 + 9.402E-3 \times f^2 - 2.742E-2 \times f - 0.2026 / C_{\sigma} = 9.804E-3 \times f^3 - 8.661E-2 \times f^2 + 2.981E-2 \times f + 0.7829$$

Appendix 3-5: Daily check results

Prior to the SAR assessment of EUT, the Daily check was performed to test whether the SRA system was operating within its target of ±10%. The Daily check results are in the table below. (*. Refer to Appendix 3-6 of measurement data.)


| Date | Freq. [MHz] | Liquid Type | Ambient [deg.C.] [%RH] | Liquid Temp. [deg.C.] Check/Before/After | Liquid Depth [mm] | Dielectric parameter | | Power drift [dB] | Daily check target & measured | | | | | Pass? | |
|----------------|-------------|-------------|------------------------|--|-------------------|----------------------|---------|------------------|-------------------------------|----------------|-----------|-----------|---------------|-------|-----------|
| | | | | | | εr [-] | σ [S/m] | | SAR (1g) [W/kg] (°C) | | | | Deviation [%] | | Limit [%] |
| | | | | | | | | | Measured | ASAR-corrected | 1W scaled | Target | | | |
| July 25, 2016 | 2450 | Head | 24.8/52 | 23.8/23.8/23.8 | 153 | 38.20 | 1.86 | 0.03 | 12.7 (250mW) | 12.43 | 49.72 | 52.4 (*c) | -5.1 | ±10 | Pass |
| July 26, 2016 | 2450 | Body | 23.9/48 | 22.5/22.7/22.7 | 153 | 50.87 | 1.975 | -0.02 | 12.6 (250mW) | 12.42 | 49.68 | 51.2 (*d) | -3.0 | ±10 | Pass |
| July 27, 2016 | 5600 | Body | 24.4/57 | 23.7/23.6/23.6 | 152 | 46.98 | 5.904 | 0.10 | 7.99 (100mW) | 7.95 | 79.5 | 77.7 (*d) | +2.3 | ±10 | Pass |
| July 28, 2016 | 5600 | Body | 24.6/59 | 23.7/23.7/23.6 | 152 | 46.98 | 5.904 | -0.01 | 7.94 (100mW) | 7.90 | 79.0 | 77.7 (*d) | +1.7 | ±10 | Pass |
| July 28, 2016 | 5750 | Body | 24.6/57 | 23.7/23.6/23.6 | 152 | 46.52 | 6.081 | 0.09 | 7.16 (100mW) | 7.12 | 71.2 | 74.0 (*d) | -3.8 | ±10 | Pass |
| June 29, 2016 | 5250 | Body | 24.7/44 | 23.7/23.6/23.6 | 152 | 47.33 | 5.454 | -0.04 | 7.41 (100mW) | 7.36 | 73.6 | 72.2 (*d) | +1.9 | ±10 | Pass |
| August 1, 2016 | 5250 | Head | 24.4/41 | 22.8/23.0/23.0 | 153 | 36.03 | 4.489 | 0.05 | 7.78 (100mW) | 7.77 | 77.7 | 75.3 (*d) | +3.2 | ±10 | Pass |
| August 2, 2016 | 5250 | Head | 24.1/50 | 22.8/22.8/22.8 | 153 | 36.03 | 4.489 | 0.03 | 7.57 (100mW) | 7.56 | 75.6 | 75.3 (*d) | +0.4 | ±10 | Pass |
| August 2, 2016 | 5600 | Head | 24.1/50 | 22.8/22.8/22.8 | 153 | 35.67 | 4.832 | -0.01 | 7.96 (100mW) | 7.95 | 79.5 | 78.6 (*d) | +1.1 | ±10 | Pass |
| August 3, 2016 | 5600 | Head | 24.1/52 | 22.8/22.7/22.7 | 153 | 35.67 | 4.832 | 0.02 | 7.91 (100mW) | 7.90 | 79.0 | 78.6 (*d) | +0.5 | ±10 | Pass |
| August 4, 2016 | 5800 | Head | 23.9/50 | 22.8/22.8/22.8 | 153 | 35.26 | 5.056 | 0.03 | 7.6 (100mW) | 7.58 | 75.8 | 75.2 (*d) | +0.8 | ±10 | Pass |

*. Calculating formula: ASAR corrected SAR (1g) (W/kg) = (Observed SAR(1g) (W/kg)) × (100 - (ΔASAR(%))) / 100

*c. The measured SAR value of Daily check was compensated for tissue dielectric deviations (ASAR) and scaled to 1W of output power in order to compare with the manufacturer's calibration target value which was normalized.

*d. The target value is a parameter defined in the calibration data sheet of D2450V2 (sn:822) and D5GHZV2 (sn:1070) dipole calibrated by Schmid & Partner Engineering AG (Certification No. D2450V2-822 Jan16 / D5GHZV2-1070 Mar16, the data sheet was filed in this report). For 2450GHz, the manufacturer's calibration data of dipole for head liquid were within 2.5 % (0.1 dB) of IEEE Std.1528 head liquid target value (=52.4 W/kg, cal.=51.4 W/kg, -1.9% vs. standard). This calibration result is enough, using this dipole as a reference. We decided to use body liquid calibration data of this dipole for the Daily check target.

*e. The target value (normalized to 1W) is defined in IEEE Std.1528.



Test setup for the system performance check


Appendix 3-6: Daily check uncertainty

| | | |
|---|-----------------|-----------------|
| Uncertainty of Daily check (2.4-6GHz) (*ε&σ tolerance: ≤±5%, DAK3.5, CW) (v08) | 1g SAR | 10g SAR |
| Combined measurement uncertainty of the measurement system (k=1) | ± 11.0 % | ± 10.9 % |
| Expanded uncertainty (k=2) | ± 22.1 % | ± 21.8 % |

| | Error Description (v08) | Uncertainty Value | Probability distribution | Divisor | ci (1g) | ci (10g) | ui (1g) | ui (10g) | Vi, veff |
|----------|---|-------------------|--------------------------|---------|---------|----------|--------------------|--------------------|----------|
| A | Measurement System (DASY5) | | | | | | (std. uncertainty) | (std. uncertainty) | |
| 1 | Probe Calibration Error (2.45,5.25,3.5,5.5,6.5,8GHz±100MHz) | ±6.55 % | Normal | 1 | 1 | 1 | ±6.55 % | ±6.55 % | ∞ |
| 2 | Axial isotropy error | ±4.7 % | Rectangular | √3 | √0.5 | √0.5 | ±1.9 % | ±1.9 % | ∞ |
| 3 | Hemispherical isotropy error | ±9.6 % | Rectangular | √3 | 0 | 0 | 0 % | 0 % | ∞ |
| 4 | Probe linearity | ±4.7 % | Rectangular | √3 | 1 | 1 | ±2.7 % | ±2.7 % | ∞ |
| 5 | Probe modulation response (CW) | ±0.0 % | Rectangular | √3 | 1 | 1 | 0 % | 0 % | ∞ |
| 6 | System detection limit | ±1.0 % | Rectangular | √3 | 1 | 1 | ±0.6 % | ±0.6 % | ∞ |
| 7 | Boundary effects | ±4.8 % | Rectangular | √3 | 1 | 1 | ±2.8 % | ±2.8 % | ∞ |
| 8 | System readout electronics (DAE) | ±0.3 % | Normal | 1 | 1 | 1 | ±0.3 % | ±0.3 % | ∞ |
| 9 | Response Time Error (<5ms/100ms wait) | ±0.0 % | Rectangular | √3 | 1 | 1 | 0 % | 0 % | ∞ |
| 10 | Integration Time Error (CW) | ±0.0 % | Rectangular | √3 | 1 | 1 | 0 % | 0 % | ∞ |
| 11 | RF ambient conditions-noise | ±3.0 % | Rectangular | √3 | 1 | 1 | ±1.7 % | ±1.7 % | ∞ |
| 12 | RF ambient conditions-reflections | ±3.0 % | Rectangular | √3 | 1 | 1 | ±1.7 % | ±1.7 % | ∞ |
| 13 | Probe positioner mechanical tolerance | ±3.3 % | Rectangular | √3 | 1 | 1 | ±1.9 % | ±1.9 % | ∞ |
| 14 | Probe positioning with respect to phantom shell | ±6.7 % | Rectangular | √3 | 1 | 1 | ±3.9 % | ±3.9 % | ∞ |
| 15 | Max. SAR evaluation (Post-processing) | ±4.0 % | Rectangular | √3 | 1 | 1 | ±2.3 % | ±2.3 % | ∞ |
| B | Test Sample Related | | | | | | | | |
| 16 | Deviation of the experimental source | ±3.5 % | Normal | 1 | 1 | 1 | ±3.5 % | ±3.5 % | ∞ |
| 17 | Dipole to liquid distance (10mm±0.2mm, <2deg.) | ±2.0 % | Rectangular | √3 | 1 | 1 | ±1.2 % | ±1.2 % | ∞ |
| 18 | Drift of output power (measured, <0.2dB) | ±2.3 % | Rectangular | √3 | 1 | 1 | ±1.3 % | ±1.3 % | ∞ |
| C | Phantom and Setup | | | | | | | | |
| 19 | Phantom uncertainty | ±2.0 % | Rectangular | √3 | 1 | 1 | ±1.2 % | ±1.2 % | ∞ |
| 20 | Algorithm for correcting SAR (ε',σ: ≤5%) | ±1.2 % | Normal | 1 | 1 | 0.84 | ±1.2 % | ±0.97 % | ∞ |
| 21 | Liquid conductivity (meas.) (DAK3.5) | ±3.0 % | Normal | 1 | 0.78 | 0.71 | ±2.3 % | ±2.1 % | ∞ |
| 22 | Liquid permittivity (meas.) (DAK3.5) | ±3.1 % | Normal | 1 | 0.23 | 0.26 | ±0.7 % | ±0.8 % | ∞ |
| 23 | Liquid Conductivity-temp.uncertainty (≤2deg.C.) | ±5.3 % | Rectangular | √3 | 0.78 | 0.71 | ±2.4 % | ±2.2 % | ∞ |
| 24 | Liquid Permittivity-temp.uncertainty (≤2deg.C.) | ±0.9 % | Rectangular | √3 | 0.23 | 0.26 | ±0.1 % | ±0.1 % | ∞ |
| | Combined Standard Uncertainty | | | | | | ±11.0 % | ±10.9 % | |
| | Expanded Uncertainty (k=2) | | | | | | ±22.1 % | ±21.8 % | |

* This measurement uncertainty budget is suggested by IEEE Std. 1528(2013) and determined by Schmid & Partner Engineering AG (DASY5 Uncertainty Budget).

Appendix 3-7: Daily check measurement data



(July 25, 2016) EUT: Dipole(2.45GHz)(sn822); Type: D2450V2; Serial: 822; Forward conducted power: 250mW

Communication System: CW (*. Frame Length in ms: 0; PAR: 0; PMF: 1); **Frequency: 2450 MHz; Crest Factor: 1.0**
Medium: HSL2450(1607); Medium parameters used: f = 2450 MHz; $\sigma = 1.86$ S/m; $\epsilon_r = 38.20$; $\rho = 1000$ kg/m³
Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007)

DASY Configuration: -Probe: EX3DV4 - SN3907; ConvF(7.22, 7.22, 7.22); Calibrated: 2016/05/12; -DASY52 52.8.8(1222); SEMCAD X 14.6.10(7331)
-Sensor-Surface: 2mm (Mechanical Surface Detection), z = 1.0, 31.0, 161.0 -Electronics: DAE4 Sn626; Calibrated: 2015/09/15
-Phantom: ELI v4.0; Type: QDOVA001BA; Serial: 1059; Phantom section: Flat Section

Area Scan:60x60,stp15 (5x5x1): Measurement grid: dx=15mm, dy=15mm; Maximum value of SAR (measured) = 19.6 W/kg


Area Scan:60x60,stp15 (41x41x1): Interpolated grid: dx=1.500 mm, dy=1.500 mm; Maximum value of SAR (interpolated) = 19.6 W/kg

Z Scan:160,stp5(mm) (1x1x33): Measurement grid: dx=20mm, dy=20mm, dz=5mm; Maximum value of SAR (measured) = 19.8 W/kg

Zoom Scan:30x30x30,stp5,5,5 (7x7x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm;
Reference Value = 104.6 V/m; Power Drift = 0.03 dB; Maximum value of SAR (measured) = 19.7 W/kg

Peak SAR (extrapolated) = 26.8 W/kg (+3.1% vs. IEC 62209-2) = 26 W/kg)

SAR(1 g) = 12.7 W/kg; SAR(10 g) = 5.86 W/kg



Remarks: * Date tested: 2016/07/25; Tested by: Hiroshi Naka; Tested place: No.7 shielded room,
* liquid depth: 153 mm; Position: distance of dipole to phantom: 8mm (10mm to liquid); ambient: 24.8 deg.C. / 52 %RH,
* liquid temperature: 23.8(start)/23.8(end)/23.8(in check) deg.C.; * White cubic: zoom scan area, Red cubic: big=SAR(10g)/small=SAR(1g)

(July 26, 2016) EUT: Dipole(2.45GHz)(sn822); Type: D2450V2; Serial: 822; Forward conducted power: 250mW

Communication System: CW (*. Frame Length in ms: 0; PAR: 0; PMF: 1); **Frequency: 2450 MHz; Crest Factor: 1.0**
Medium: M2450(1607); Medium parameters used: f = 2450 MHz; $\sigma = 1.975$ S/m; $\epsilon_r = 50.87$; $\rho = 1000$ kg/m³
Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007)

DASY Configuration: -Probe: EX3DV4 - SN3907; ConvF(7.16, 7.16, 7.16); Calibrated: 2016/05/12; -DASY52 52.8.8(1222); SEMCAD X 14.6.10(7331)
-Sensor-Surface: 2mm (Mechanical Surface Detection), z = 1.0, 31.0, 161.0 -Electronics: DAE4 Sn626; Calibrated: 2015/09/15
-Phantom: ELI v4.0; Type: QDOVA001BA; Serial: 1059; Phantom section: Flat Section

Area Scan:60x60,stp15 (5x5x1): Measurement grid: dx=15mm, dy=15mm; Maximum value of SAR (measured) = 18.9 W/kg


Area Scan:60x60,stp15 (41x41x1): Interpolated grid: dx=1.500 mm, dy=1.500 mm; Maximum value of SAR (interpolated) = 19.1 W/kg

Z Scan:160,stp5(mm) (1x1x33): Measurement grid: dx=20mm, dy=20mm, dz=5mm; Maximum value of SAR (measured) = 19.1 W/kg

Zoom Scan:30x30x30,stp5,5,5 (7x7x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm;
Reference Value = 99.40 V/m; Power Drift = -0.02 dB; Maximum value of SAR (measured) = 19.2 W/kg

Peak SAR (extrapolated) = 25.5 W/kg (-3.0% vs. speag-cal) = 26.3 W/kg)

SAR(1 g) = 12.6 W/kg; SAR(10 g) = 5.88 W/kg




Remarks: * Date tested: 2016/07/26; Tested by: Hiroshi Naka; Tested place: No.7 shielded room,
* liquid depth: 153 mm; Position: distance of dipole to phantom: 8mm (10mm to liquid); ambient: 23.9 deg.C. / 48 %RH,
* liquid temperature: 22.7(start)/22.7(end)/22.5(in check) deg.C.; * White cubic: zoom scan area, Red cubic: big=SAR(10g)/small=SAR(1g)

UL Japan, Inc.

Shonan EMC Lab.

1-22-3 Megumigaoka, Hiratsuka-shi, Kanagawa-ken, 259-1220 JAPAN
Telephone: +81 463 50 6400 / Facsimile: +81 463 50 6401

Appendix 3-7: Daily check measurement data (cont'd)




(July 27, 2016) EUT: Dipole(5GHz)(1070); Type: D5GHzV2; Serial: 1070; Forward conducted power: 100mW

Communication System: CW (*. Frame Length in ms: 0; PAR: 0; PMF: 1); **Frequency: 5600 MHz; Crest Factor: 1.0**
Medium: MSL5800(1607); Medium parameters used: f = 5600 MHz; $\sigma = 5.904$ S/m; $\epsilon_r = 46.98$; $\rho = 1000$ kg/m³
Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007)

DASY Configuration: -Probe: EX3DV4 - SN3907; ConvF(3.65, 3.65, 3.65); Calibrated: 2016/05/12; -DASY52 52.8.8(1222); SEMCAD X 14.6.10(7331)
-Sensor-Surface: 1.4mm (Mechanical Surface Detection), z = 1.0, 25.0, 156.0 -Electronics: DAE4 Sn626; Calibrated: 2015/09/15
-Phantom: ELI v4.0; Type: QDOVA001BA; Serial: 1059; Phantom section: Flat Section

Area:60x60,stp10 (7x7x1): Measurement grid: dx=10mm, dy=10mm; Maximum value of SAR (measured) = 20.2 W/kg
Area:60x60,stp10 (61x61x1): Interpolated grid: dx=1.000 mm, dy=1.000 mm; Maximum value of SAR (interpolated) = 20.6 W/kg
Z Scan:155,5 (1x1x32): Measurement grid: dx=20mm, dy=20mm, dz=5mm; Maximum value of SAR (measured) = 18.9 W/kg
Zoom:28x28x24,xy4-z1.4(ratio) (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm;
Reference Value = 68.44 V/m; Power Drift = 0.10 dB; Maximum value of SAR (measured) = 19.0 W/kg
Peak SAR (extrapolated) = 32.6 W/kg (+1.6% vs speag-cal.=32.1 W/kg)
SAR(1 g) = 7.99 W/kg; SAR(10 g) = 2.23 W/kg




Remarks: * Date tested: 2016/07/27; Tested by: Hiroshi Naka; Tested place: No.7 shielded room,
* liquid depth: 152 mm; Position: distance of dipole to phantom: 8mm (10mm to liquid); ambient: 24.4 deg.C. / 57%RH,
* liquid temperature: 23.6(start)/23.6(end)/23.7(in check) deg.C.; * White cubic: zoom scan area, Red cubic: big=SAR(10g)/small=SAR(1g)

(July 28, 2016) EUT: Dipole(5GHz)(1070); Type: D5GHzV2; Serial: 1070; Forward conducted power: 100mW

Communication System: CW (*. Frame Length in ms: 0; PAR: 0; PMF: 1); **Frequency: 5600 MHz; Crest Factor: 1.0**
Medium: MSL5800(1607); Medium parameters used: f = 5600 MHz; $\sigma = 5.904$ S/m; $\epsilon_r = 46.98$; $\rho = 1000$ kg/m³
Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007)


DASY Configuration: -Probe: EX3DV4 - SN3907; ConvF(3.65, 3.65, 3.65); Calibrated: 2016/05/12; -DASY52 52.8.8(1222); SEMCAD X 14.6.10(7331)
-Sensor-Surface: 1.4mm (Mechanical Surface Detection), z = 1.0, 25.0, 156.0 -Electronics: DAE4 Sn626; Calibrated: 2015/09/15
-Phantom: ELI v4.0; Type: QDOVA001BA; Serial: 1059; Phantom section: Flat Section

Area:60x60,stp10 (7x7x1): Measurement grid: dx=10mm, dy=10mm; Maximum value of SAR (measured) = 20.4 W/kg
Area:60x60,stp10 (61x61x1): Interpolated grid: dx=1.000 mm, dy=1.000 mm; Maximum value of SAR (interpolated) = 20.8 W/kg
Z Scan:155,5 (1x1x32): Measurement grid: dx=20mm, dy=20mm, dz=5mm; Maximum value of SAR (measured) = 18.8 W/kg
Zoom:28x28x24,xy4-z1.4(ratio) (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm;
Reference Value = 68.51 V/m; Power Drift = -0.01 dB; Maximum value of SAR (measured) = 18.7 W/kg
Peak SAR (extrapolated) = 32.1 W/kg (0% vs speag-cal.=32.1 W/kg)
SAR(1 g) = 7.94 W/kg; SAR(10 g) = 2.21 W/kg



Remarks: * Date tested: 2016/07/28; Tested by: Hiroshi Naka; Tested place: No.7 shielded room,
* liquid depth: 152 mm; Position: distance of dipole to phantom: 8mm (10mm to liquid); ambient: 24.6 deg.C. / 59%RH,
* liquid temperature: 23.7(start)/23.6(end)/23.7(in check) deg.C.; * White cubic: zoom scan area, Red cubic: big=SAR(10g)/small=SAR(1g)


Appendix 3-7: Daily check measurement data (cont'd)



(July 28, 2016) EUT: Dipole(5GHz)(1070); Type: D5GHzV2; Serial: 1070; Forward conducted power: 100mW
Communication System: CW (*. Frame Length in ms: 0; PAR: 0; PMF: 1); **Frequency: 5750 MHz; Crest Factor: 1.0**
Medium: MSL5800(1607); Medium parameters used: f = 5750 MHz; $\sigma = 6.081$ S/m; $\epsilon_r = 46.52$; $\rho = 1000$ kg/m³
Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007)

DASY Configuration: -Probe: EX3DV4 - SN3907; ConvF(3.96, 3.96, 3.96); Calibrated: 2016/05/12; -DASY52 52.8.8(1222); SEMCAD X 14.6.10(7331)
-Sensor-Surface: 1.4mm (Mechanical Surface Detection), z = 1.0, 25.0, 156.0 -Electronics: DAE4 Sn626; Calibrated: 2015/09/15
-Phantom: ELI v4.0; Type: QDOVA001BA; Serial: 1059; Phantom section: Flat Section

Area:60x60,stp10 (7x7x1): Measurement grid: dx=10mm, dy=10mm; Maximum value of SAR (measured) = 17.7 W/kg
Area:60x60,stp10 (61x61x1): Interpolated grid: dx=1.000 mm, dy=1.000 mm; Maximum value of SAR (interpolated) = 18.1 W/kg
Z Scan:155,5 (1x1x32): Measurement grid: dx=20mm, dy=20mm, dz=5mm; Maximum value of SAR (measured) = 17.9 W/kg
Zoom:28x28x24,xy4-z1.4(ratio) (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm;
Reference Value = 63.55 V/m; Power Drift = 0.09 dB; Maximum value of SAR (measured) = 17.9 W/kg
Peak SAR (extrapolated) = 30.8 W/kg (-3.8 % vs.speag-cal=32 W/kg)
SAR(1 g) = 7.16 W/kg; SAR(10 g) = 1.99 W/kg




Remarks: * Date tested: 2016/07/28; Tested by: Hiroshi Naka; Tested place: No.7 shielded room,
* liquid depth: 152 mm; Position: distance of dipole to phantom: 8mm (10mm to liquid); ambient: 24.6 deg.C. / 57%RH,
* liquid temperature: 23.6(start)/23.6(end)/23.7(in check) deg.C.; * White cubic: zoom scan area, Red cubic: big=SAR(10g)/small=SAR(1g)

(July 29, 2016) EUT: Dipole(5GHz)(1070); Type: D5GHzV2; Serial: 1070; Forward conducted power: 100mW
Communication System: CW (*. Frame Length in ms: 0; PAR: 0; PMF: 1); **Frequency: 5250 MHz; Crest Factor: 1.0**
Medium: MSL5800(1607); Medium parameters used: f = 5250 MHz; $\sigma = 5.454$ S/m; $\epsilon_r = 47.33$; $\rho = 1000$ kg/m³
Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007)


DASY Configuration: -Probe: EX3DV4 - SN3907; ConvF(4.37, 4.37, 4.37); Calibrated: 2016/05/12; -DASY52 52.8.8(1222); SEMCAD X 14.6.10(7331)
-Sensor-Surface: 1.4mm (Mechanical Surface Detection), z = 1.0, 25.0, 156.0 -Electronics: DAE4 Sn626; Calibrated: 2015/09/15
-Phantom: ELI v4.0; Type: QDOVA001BA; Serial: 1059; Phantom section: Flat Section

Area:60x60,stp10 (7x7x1): Measurement grid: dx=10mm, dy=10mm; Maximum value of SAR (measured) = 18.6 W/kg
Area:60x60,stp10 (61x61x1): Interpolated grid: dx=1.000 mm, dy=1.000 mm; Maximum value of SAR (interpolated) = 18.8 W/kg
Z Scan:155,5 (1x1x32): Measurement grid: dx=20mm, dy=20mm, dz=5mm; Maximum value of SAR (measured) = 16.9 W/kg
Zoom:28x28x24,xy4-z1.4(ratio) (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm;
Reference Value = 66.11 V/m; Power Drift = -0.04 dB; Maximum value of SAR (measured) = 17.0 W/kg
Peak SAR (extrapolated) = 29.8 W/kg (-8.8 % vs.speag-cal=27.4 W/kg)
SAR(1 g) = 7.41 W/kg; SAR(10 g) = 2.08 W/kg



Remarks: * Date tested: 2016/07/29; Tested by: Hiroshi Naka; Tested place: No.7 shielded room,
* liquid depth: 152 mm; Position: distance of dipole to phantom: 8mm (10mm to liquid); ambient: 24.7 deg.C. / 44%RH,
* liquid temperature: 23.6(start)/23.6(end)/23.7(in check) deg.C.; * White cubic: zoom scan area, Red cubic: big=SAR(10g)/small=SAR(1g)

Appendix 3-7: Daily check measurement data (cont'd)




(August 1, 2016) EUT: Dipole(5GHz)(1070); Type: D5GHzV2; Serial: 1070; Forward conducted power: 100mW
Communication System: CW (*. Frame Length in ms: 0; PAR: 0; PMF: 1); **Frequency: 5250 MHz; Crest Factor: 1.0**
Medium: HSL5GHz(1608); Medium parameters used: f = 5250 MHz; $\sigma = 4.489$ S/m; $\epsilon_r = 36.03$; $\rho = 1000$ kg/m³
Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007)

DASY Configuration: -Probe: EX3DV4 - SN3907; ConvF(4.94, 4.94, 4.94); Calibrated: 2016/05/12; -DASY52 52.8.8(1222); SEMCAD X 14.6.10(7331)
-Sensor-Surface: 1.4mm (Mechanical Surface Detection), z = 1.0, 25.0, 156.0 -Electronics: DAE4 Sn626; Calibrated: 2015/09/15
-Phantom: ELI v4.0; Type: QDOVA001BA; Serial: 1059; Phantom section: Flat Section

Area:60x60,stp10 (7x7x1): Measurement grid: dx=10mm, dy=10mm; Maximum value of SAR (measured) = 18.8 W/kg
Area:60x60,stp10 (61x61x1): Interpolated grid: dx=1.000 mm, dy=1.000 mm; Maximum value of SAR (interpolated) = 19.0 W/kg
Z Scan:155.5 (1x1x32): Measurement grid: dx=20mm, dy=20mm, dz=5mm; Maximum value of SAR (measured) = 18.2 W/kg

Zoom:28x28x24,xy4-z1.4(ratio) (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm;
Reference Value = 70.42 V/m; Power Drift = 0.05 dB; Maximum value of SAR (measured) = 18.2 W/kg
Peak SAR (extrapolated) = 31.2 W/kg (+13.0 %, vs.speag-cal.=27.6 W/kg)
SAR(1 g) = 7.78 W/kg; SAR(10 g) = 2.22 W/kg




Remarks: * Date tested: 2016/08/01; Tested by: Hiroshi Naka; Tested place:No.7 shielded room,
* liquid depth: 153 mm; Position: distance of dipole to phantom: 8mm (10mm to liquid); ambient: 24.4 deg.C. / 41 %RH,
* liquid temperature: 23.0(start)/23.0(end)/22.8(in check) deg.C.; * White cubic: zoom scan area, Red cubic: big=SAR(10g)/small=SAR(1g)

(August 2, 2016) EUT: Dipole(5GHz)(1070); Type: D5GHzV2; Serial: 1070; Forward conducted power: 100mW
Communication System: CW (*. Frame Length in ms: 0; PAR: 0; PMF: 1); **Frequency: 5250 MHz; Crest Factor: 1.0**
Medium: HSL5GHz(1608); Medium parameters used: f = 5250 MHz; $\sigma = 4.489$ S/m; $\epsilon_r = 36.03$; $\rho = 1000$ kg/m³
Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007)

DASY Configuration: -Probe: EX3DV4 - SN3907; ConvF(4.94, 4.94, 4.94); Calibrated: 2016/05/12; -DASY52 52.8.8(1222); SEMCAD X 14.6.10(7331)
-Sensor-Surface: 1.4mm (Mechanical Surface Detection), z = 1.0, 25.0, 156.0 -Electronics: DAE4 Sn626; Calibrated: 2015/09/15
-Phantom: ELI v4.0; Type: QDOVA001BA; Serial: 1059; Phantom section: Flat Section

Area:60x60,stp10 (7x7x1): Measurement grid: dx=10mm, dy=10mm; Maximum value of SAR (measured) = 18.4 W/kg
Area:60x60,stp10 (61x61x1): Interpolated grid: dx=1.000 mm, dy=1.000 mm; Maximum value of SAR (interpolated) = 18.7 W/kg
Z Scan:155.5 (1x1x32): Measurement grid: dx=20mm, dy=20mm, dz=5mm; Maximum value of SAR (measured) = 17.8 W/kg

Zoom:28x28x24,xy4-z1.4(ratio) (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm;
Reference Value = 69.90 V/m; Power Drift = 0.03 dB; Maximum value of SAR (measured) = 17.8 W/kg
Peak SAR (extrapolated) = 30.6 W/kg (+10.9 %, vs.speag-cal.=27.6 W/kg)
SAR(1 g) = 7.57 W/kg; SAR(10 g) = 2.16 W/kg




Remarks: * Date tested: 2016/08/02; Tested by: Hiroshi Naka; Tested place:No.7 shielded room,
* liquid depth: 153 mm; Position: distance of dipole to phantom: 8mm (10mm to liquid); ambient: 24.1 deg.C. / 50 %RH,
* liquid temperature: 22.8(start)/22.8(end)/22.8(in check) deg.C.; * White cubic: zoom scan area, Red cubic: big=SAR(10g)/small=SAR(1g)

UL Japan, Inc.

Shonan EMC Lab.

1-22-3 Megumigaoka, Hiratsuka-shi, Kanagawa-ken, 259-1220 JAPAN
Telephone: +81 463 50 6400 / Facsimile: +81 463 50 6401

Appendix 3-7: Daily check measurement data (cont'd)



(August 2, 2016) EUT: Dipole(5GHz)(1070); Type: D5GHzV2; Serial: 1070; Forward conducted power: 100mW
Communication System: CW (*. Frame Length in ms: 0; PAR: 0; PMF: 1); **Frequency: 5600 MHz; Crest Factor: 1.0**
Medium: HSL5GHz(1608); Medium parameters used: f = 5600 MHz; $\sigma = 4.832$ S/m; $\epsilon_r = 35.67$; $\rho = 1000$ kg/m³
Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007)


DASY Configuration: -Probe: EX3DV4 - SN3907; ConvF(4.33, 4.33, 4.33); Calibrated: 2016/05/12; -DASY52 52.8.8(1222); SEMCAD X 14.6.10(7331)
-Sensor-Surface: 1.4mm (Mechanical Surface Detection), z = 1.0, 25.0, 156.0 -Electronics: DAE4 Sn626; Calibrated: 2015/09/15
-Phantom: ELI v4.0; Type: QDOVA001BA; Serial: 1059; Phantom section: Flat Section

Area:60x60,stp10 (7x7x1): Measurement grid: dx=10mm, dy=10mm; Maximum value of SAR (measured) = 19.7 W/kg
Area:60x60,stp10 (61x61x1): Interpolated grid: dx=1.000 mm, dy=1.000 mm; Maximum value of SAR (interpolated) = 20.1 W/kg
Z Scan:155.5 (1x1x32): Measurement grid: dx=20mm, dy=20mm, dz=5mm; Maximum value of SAR (measured) = 19.5 W/kg

Zoom:28x28x24,xy4-z1.4(ratio) (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm;
Reference Value = 71.93 V/m; Power Drift = -0.01 dB; Maximum value of SAR (measured) = 19.2 W/kg;

Peak SAR (extrapolated) = 32.3 W/kg (+4.2% vs.speag-cal.=31 W/kg)

SAR(1 g) = 7.96 W/kg; SAR(10 g) = 2.26 W/kg



Remarks: * Date tested: 2016/08/02; Tested by: Hiroshi Naka; Tested place: No.7 shielded room,
* liquid depth: 153 mm; Position: distance of dipole to phantom: 8mm (10mm to liquid); ambient: 24.1 deg.C. / 50%RH,
* liquid temperature: 22.8(start)/22.8(end)/22.8(in check) deg.C.; * White cubic: zoom scan area, Red cubic: big=SAR(10g)/small=SAR(1g)

(August 3, 2016) EUT: Dipole(5GHz)(1070); Type: D5GHzV2; Serial: 1070; Forward conducted power: 100mW
Communication System: CW (*. Frame Length in ms: 0; PAR: 0; PMF: 1); **Frequency: 5600 MHz; Crest Factor: 1.0**
Medium: HSL5GHz(1608); Medium parameters used: f = 5600 MHz; $\sigma = 4.832$ S/m; $\epsilon_r = 35.67$; $\rho = 1000$ kg/m³
Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007)


DASY Configuration: -Probe: EX3DV4 - SN3907; ConvF(4.33, 4.33, 4.33); Calibrated: 2016/05/12; -DASY52 52.8.8(1222); SEMCAD X 14.6.10(7331)
-Sensor-Surface: 1.4mm (Mechanical Surface Detection), z = 1.0, 25.0, 156.0 -Electronics: DAE4 Sn626; Calibrated: 2015/09/15
-Phantom: ELI v4.0; Type: QDOVA001BA; Serial: 1059; Phantom section: Flat Section

Area:60x60,stp10 (7x7x1): Measurement grid: dx=10mm, dy=10mm; Maximum value of SAR (measured) = 19.9 W/kg
Area:60x60,stp10 (61x61x1): Interpolated grid: dx=1.000 mm, dy=1.000 mm; Maximum value of SAR (interpolated) = 20.2 W/kg
Z Scan:155.5 (1x1x32): Measurement grid: dx=20mm, dy=20mm, dz=5mm; Maximum value of SAR (measured) = 18.7 W/kg

Zoom:28x28x24,xy4-z1.4(ratio) (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm;
Reference Value = 71.63 V/m; Power Drift = 0.02 dB; Maximum value of SAR (measured) = 18.9 W/kg;

Peak SAR (extrapolated) = 32.6 W/kg (+5.2% vs.speag-cal.=31 W/kg)

SAR(1 g) = 7.91 W/kg; SAR(10 g) = 2.24 W/kg



Remarks: * Date tested: 2016/08/03; Tested by: Hiroshi Naka; Tested place: No.7 shielded room,
* liquid depth: 153 mm; Position: distance of dipole to phantom: 8mm (10mm to liquid); ambient: 24.1 deg.C. / 52%RH,
* liquid temperature: 22.7(start)/22.7(end)/22.8(in check) deg.C.; * White cubic: zoom scan area, Red cubic: big=SAR(10g)/small=SAR(1g)

UL Japan, Inc.

Shonan EMC Lab.

1-22-3 Megumigaoka, Hiratsuka-shi, Kanagawa-ken, 259-1220 JAPAN
Telephone: +81 463 50 6400 / Facsimile: +81 463 50 6401

Appendix 3-7: Daily check measurement data (cont'd)



(August 4, 2016) EUT: Dipole(5GHz)(1070); Type: D5GHzV2; Serial: 1070; Forward conducted power: 100mW
Communication System: CW (*. Frame Length in ms: 0; PAR: 0; PMF: 1); **Frequency: 5800 MHz; Crest Factor: 1.0**
Medium: HSL5GHz(1608); Medium parameters used: f = 5800 MHz; $\sigma = 5.056$ S/m; $\epsilon_r = 35.26$; $\rho = 1000$ kg/m³
 Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007)


DASY Configuration: -Probe: EX3DV4 - SN3907; ConvF(4.3, 4.3, 4.3); Calibrated: 2016/05/12; -DASY52 52.8.8(1222); SEMCAD X 14.6.10(7331)
 -Sensor-Surface: 1.4mm (Mechanical Surface Detection), z= 1.0, 25.0, 156.0 -Electronics: DAE4 Sn626; Calibrated: 2015/09/15
 -Phantom: ELI v4.0; Type: QDOVA001BA; Serial: 1059; Phantom section: Flat Section

Area:60x60,stp10 (7x7x1): Measurement grid: dx=10mm, dy=10mm; Maximum value of SAR (measured) = 19.5 W/kg
Area:60x60,stp10 (61x61x1): Interpolated grid: dx=1.000 mm, dy=1.000 mm; Maximum value of SAR (interpolated) = 19.6 W/kg
Z Scan;155,5 (1x1x32): Measurement grid: dx=20mm, dy=20mm, dz=5mm; Maximum value of SAR (measured) = 18.2 W/kg

Zoom:28x28x24,xy4-z1.4(ratio) (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm;
 Reference Value = 70.19 V/m; Power Drift = 0.03 dB; Maximum value of SAR (measured) = 18.1 W/kg

Peak SAR (extrapolated) = 32.4 W/kg (+3.8 %, vs.speag-cal.=31.2 W/kg)

SAR(1 g) = 7.6 W/kg; SAR(10 g) = 2.15 W/kg



Remarks: * Date tested: 2016/08/04; Tested by: Hiroshi Naka; Tested place: No.7 shielded room,
 * liquid depth: 153 mm; Position: distance of dipole to phantom: 8mm (10mm to liquid); ambient: 23.9 deg.C. / 50%RH,
 * liquid temperature: 22.8(start)22.8(end)22.8(in check) deg.C.; * White cubic: zoom scan area, Red cubic: big=SAR(10g)/small=SAR(1g)

Appendix 3-8: Calibration certificate: E-Field Probe (EX3DV4)

**Calibration Laboratory of
 Schmid & Partner
 Engineering AG**
 Zeughausstrasse 43, 8004 Zurich, Switzerland



S Schweizerischer Kalibrierdienst
C Service suisse d'étalonnage
S Servizio svizzero di taratura
 Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS)
 The Swiss Accreditation Service is one of the signatories to the EA
 Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 0108**

Client **UL Japan Shonan (Vitec)**

Certificate No: **EX3-3907_May16**

CALIBRATION CERTIFICATE

Object: **EX3DV4 - SN:3907**

Calibration procedure(s): **QA CAL-01.v9, QA CAL-14.v4, QA CAL-23.v5, QA CAL-25.v6
 Calibration procedure for dosimetric E-field probes**

Calibration date: **May 12, 2016**

This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI).
 The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature (22 ± 3)°C and humidity < 70%.

Calibration Equipment used (M&TE critical for calibration)

| Primary Standards | ID | Cal Date (Certificate No.) | Scheduled Calibration |
|----------------------------|------------------|-----------------------------------|------------------------|
| Power meter NRP | SN: 104778 | 06-Apr-16 (No. 217-02288/02289) | Apr-17 |
| Power sensor NRP-Z91 | SN: 103244 | 06-Apr-16 (No. 217-02288) | Apr-17 |
| Power sensor NRP-Z91 | SN: 103245 | 06-Apr-16 (No. 217-02289) | Apr-17 |
| Reference 20 dB Attenuator | SN: S5277 (20x) | 05-Apr-16 (No. 217-02293) | Apr-17 |
| Reference Probe ES3DV2 | SN: 3013 | 31-Dec-15 (No. ES3-3013_Dec15) | Dec-16 |
| DAE4 | SN: 660 | 23-Dec-15 (No. DAE4-660_Dec15) | Dec-16 |
| Secondary Standards | ID | Check Date (in house) | Scheduled Check |
| Power meter E4419B | SN: GB41293874 | 06-Apr-16 (No. 217-02285/02284) | In house check: Jun-16 |
| Power sensor E4412A | SN: MY41498087 | 06-Apr-16 (No. 217-02285) | In house check: Jun-16 |
| Power sensor E4412A | SN: 000110210 | 06-Apr-16 (No. 217-02284) | In house check: Jun-16 |
| RF generator HP 8648C | SN: US3642U01700 | 04-Aug-99 (in house check Apr-13) | In house check: Jun-16 |
| Network Analyzer HP 8753E | SN: US37390585 | 18-Oct-01 (in house check Oct-15) | In house check: Oct-16 |

Calibrated by: **Name: Claudio Leubler, Function: Laboratory Technician, Signature: [Handwritten Signature]**

Approved by: **Name: Katja Pokovic, Function: Technical Manager, Signature: [Handwritten Signature]**

Issued: May 14, 2016

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

Appendix 3-8: Calibration certificate: E-Field Probe (EX3DV4) (cont'd)

Calibration Laboratory of
Schmid & Partner
Engineering AG
Zeughausstrasse 43, 8004 Zurich, Switzerland



S Schweizerischer Kalibrierdienst
C Service suisse d'étalonnage
S Servizio svizzero di taratura
S Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS)
The Swiss Accreditation Service is one of the signatories to the EA
Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: SCS 0108

Glossary:

| | |
|--------------------------|---|
| TSL | tissue simulating liquid |
| NORM _{x,y,z} | sensitivity in free space |
| ConvF | sensitivity in TSL / NORM _{x,y,z} |
| DCP | diode compression point |
| CF | crest factor (1/duty_cycle) of the RF signal |
| A, B, C, D | modulation dependent linearization parameters |
| Polarization φ | φ rotation around probe axis |
| Polarization ϑ | ϑ rotation around an axis that is in the plane normal to probe axis (at measurement center), i.e., $\vartheta = 0$ is normal to probe axis |
| Connector Angle | information used in DASY system to align probe sensor X to the robot coordinate system |

Calibration is Performed According to the Following Standards:

- IEEE Std 1528-2013, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", June 2013
- IEC 62209-1, "Procedure to measure the Specific Absorption Rate (SAR) for hand-held devices used in close proximity to the ear (frequency range of 300 MHz to 3 GHz)", February 2005
- IEC 62209-2, "Procedure to determine the Specific Absorption Rate (SAR) for wireless communication devices used in close proximity to the human body (frequency range of 30 MHz to 6 GHz)", March 2010
- KDB 865664, "SAR Measurement Requirements for 100 MHz to 6 GHz"

Methods Applied and Interpretation of Parameters:

- NORM_{x,y,z}: Assessed for E-field polarization $\vartheta = 0$ ($f \leq 900$ MHz in TEM-cell; $f > 1800$ MHz: R22 waveguide). NORM_{x,y,z} are only intermediate values, i.e., the uncertainties of NORM_{x,y,z} does not affect the E²-field uncertainty inside TSL (see below ConvF).
- NORM(f)_{x,y,z} = NORM_{x,y,z} * frequency_response (see Frequency Response Chart). This linearization is implemented in DASY4 software versions later than 4.2. The uncertainty of the frequency response is included in the stated uncertainty of ConvF.
- DCP_{x,y,z}: DCP are numerical linearization parameters assessed based on the data of power sweep with CW signal (no uncertainty required). DCP does not depend on frequency nor media.
- PAR: PAR is the Peak to Average Ratio that is not calibrated but determined based on the signal characteristics
- A_{x,y,z}; B_{x,y,z}; C_{x,y,z}; D_{x,y,z}; VR_{x,y,z}: A, B, C, D are numerical linearization parameters assessed based on the data of power sweep for specific modulation signal. The parameters do not depend on frequency nor media. VR is the maximum calibration range expressed in RMS voltage across the diode.
- ConvF and Boundary Effect Parameters: Assessed in flat phantom using E-field (or Temperature Transfer Standard for $f \leq 800$ MHz) and inside waveguide using analytical field distributions based on power measurements for $f > 800$ MHz. The same setups are used for assessment of the parameters applied for boundary compensation (alpha, depth) of which typical uncertainty values are given. These parameters are used in DASY4 software to improve probe accuracy close to the boundary. The sensitivity in TSL corresponds to NORM_{x,y,z} * ConvF whereby the uncertainty corresponds to that given for ConvF. A frequency dependent ConvF is used in DASY version 4.4 and higher which allows extending the validity from ± 50 MHz to ± 100 MHz.
- Spherical isotropy (3D deviation from isotropy): in a field of low gradients realized using a flat phantom exposed by a patch antenna.
- Sensor Offset: The sensor offset corresponds to the offset of virtual measurement center from the probe tip (on probe axis). No tolerance required.
- Connector Angle: The angle is assessed using the information gained by determining the NORM_x (no uncertainty required).

Appendix 3-8: Calibration certificate: E-Field Probe (EX3DV4) (cont'd)

EX3DV4 – SN:3907

May 12, 2016

Probe EX3DV4

SN:3907

Manufactured: September 4, 2012
Calibrated: May 12, 2016

Calibrated for DASY/EASY Systems
(Note: non-compatible with DASY2 system!)

Appendix 3-8: Calibration certificate: E-Field Probe (EX3DV4) (cont'd)

EX3DV4- SN:3907

May 12, 2016

DASY/EASY - Parameters of Probe: EX3DV4 - SN:3907

Basic Calibration Parameters

| | Sensor X | Sensor Y | Sensor Z | Unc (k=2) |
|---|----------|----------|----------|-----------|
| Norm ($\mu\text{V}/(\text{V}/\text{m})^2$) ^A | 0.44 | 0.45 | 0.46 | ± 10.1 % |
| DCP (mV) ^B | 101.0 | 99.2 | 101.3 | |

Modulation Calibration Parameters

| UID | Communication System Name | | A dB | B dB√μV | C | D dB | VR mV | Unc ^E (k=2) |
|-----|---------------------------|---|---------|------------|-----|---------|----------|---------------------------|
| 0 | CW | X | 0.0 | 0.0 | 1.0 | 0.00 | 150.6 | ±3.0 % |
| | | Y | 0.0 | 0.0 | 1.0 | | 152.4 | |
| | | Z | 0.0 | 0.0 | 1.0 | | 157.7 | |

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

^A The uncertainties of Norm X,Y,Z do not affect the E²-field uncertainty inside TSL (see Pages 5 and 6).
^B Numerical linearization parameter: uncertainty not required.
^E Uncertainty is determined using the max. deviation from linear response applying rectangular distribution and is expressed for the square of the field value.

Appendix 3-8: Calibration certificate: E-Field Probe (EX3DV4) (cont'd)

EX3DV4- SN:3907

May 12, 2016

DASY/EASY - Parameters of Probe: EX3DV4 - SN:3907

Calibration Parameter Determined in Head Tissue Simulating Media

| f (MHz) ^C | Relative Permittivity ^F | Conductivity (S/m) ^F | ConvF X | ConvF Y | ConvF Z | Alpha ^G | Depth (mm) ^G | Unc (k=2) |
|----------------------|------------------------------------|---------------------------------|---------|---------|---------|--------------------|-------------------------|-----------|
| 2450 | 39.2 | 1.80 | 7.22 | 7.22 | 7.22 | 0.27 | 0.90 | ± 12.0 % |
| 5200 | 36.0 | 4.66 | 5.21 | 5.21 | 5.21 | 0.30 | 1.80 | ± 13.1 % |
| 5250 | 35.9 | 4.71 | 4.94 | 4.94 | 4.94 | 0.35 | 1.80 | ± 13.1 % |
| 5500 | 35.6 | 4.96 | 4.60 | 4.60 | 4.60 | 0.40 | 1.80 | ± 13.1 % |
| 5600 | 35.5 | 5.07 | 4.33 | 4.33 | 4.33 | 0.45 | 1.80 | ± 13.1 % |
| 5750 | 35.4 | 5.22 | 4.47 | 4.47 | 4.47 | 0.45 | 1.80 | ± 13.1 % |
| 5800 | 35.3 | 5.27 | 4.30 | 4.30 | 4.30 | 0.45 | 1.80 | ± 13.1 % |

^C Frequency validity above 300 MHz of ± 100 MHz only applies for DASY v4.4 and higher (see Page 2), else it is restricted to ± 50 MHz. The uncertainty is the RSS of the ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band. Frequency validity below 300 MHz is ± 10, 25, 40, 50 and 70 MHz for ConvF assessments at 30, 64, 128, 150 and 220 MHz respectively. Above 5 GHz frequency validity can be extended to ± 110 MHz.

^F At frequencies below 3 GHz, the validity of tissue parameters (ϵ and σ) can be relaxed to ± 10% if liquid compensation formula is applied to measured SAR values. At frequencies above 3 GHz, the validity of tissue parameters (ϵ and σ) is restricted to ± 5%. The uncertainty is the RSS of the ConvF uncertainty for indicated target tissue parameters.

^G Alpha/Depth are determined during calibration. SPEAG warrants that the remaining deviation due to the boundary effect after compensation is always less than ± 1% for frequencies below 3 GHz and below ± 2% for frequencies between 3-6 GHz at any distance larger than half the probe tip diameter from the boundary.

Appendix 3-8: Calibration certificate: E-Field Probe (EX3DV4) (cont'd)

EX3DV4- SN:3907

May 12, 2016

DASY/EASY - Parameters of Probe: EX3DV4 - SN:3907

Calibration Parameter Determined in Body Tissue Simulating Media

| f (MHz) ^C | Relative Permittivity ^F | Conductivity (S/m) ^F | ConvF X | ConvF Y | ConvF Z | Alpha ^G | Depth ^G (mm) | Unc (k=2) |
|----------------------|------------------------------------|---------------------------------|---------|---------|---------|--------------------|-------------------------|-----------|
| 2450 | 52.7 | 1.95 | 7.16 | 7.16 | 7.16 | 0.37 | 0.90 | ± 12.0 % |
| 5250 | 48.9 | 5.36 | 4.37 | 4.37 | 4.37 | 0.45 | 1.90 | ± 13.1 % |
| 5600 | 48.5 | 5.77 | 3.65 | 3.65 | 3.65 | 0.55 | 1.90 | ± 13.1 % |
| 5750 | 48.3 | 5.94 | 3.96 | 3.96 | 3.96 | 0.55 | 1.90 | ± 13.1 % |

^C Frequency validity above 300 MHz of ± 100 MHz only applies for DASY v4.4 and higher (see Page 2), else it is restricted to ± 50 MHz. The uncertainty is the RSS of the ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band. Frequency validity below 300 MHz is ± 10, 25, 40, 50 and 70 MHz for ConvF assessments at 30, 64, 128, 150 and 220 MHz respectively. Above 5 GHz frequency validity can be extended to ± 110 MHz.

^F At frequencies below 3 GHz, the validity of tissue parameters (ϵ and σ) can be relaxed to ± 10% if liquid compensation formula is applied to measured SAR values. At frequencies above 3 GHz, the validity of tissue parameters (ϵ and σ) is restricted to ± 5%. The uncertainty is the RSS of the ConvF uncertainty for indicated target tissue parameters.


^G Alpha/Depth are determined during calibration. SPEAG warrants that the remaining deviation due to the boundary effect after compensation is always less than ± 1% for frequencies below 3 GHz and below ± 2% for frequencies between 3-6 GHz at any distance larger than half the probe tip diameter from the boundary.

Appendix 3-8: Calibration certificate: E-Field Probe (EX3DV4) (cont'd)

EX3DV4- SN:3907

May 12, 2016

Frequency Response of E-Field
(TEM-Cell:ifi110 EXX, Waveguide: R22)



Uncertainty of Frequency Response of E-field: $\pm 6.3\%$ (k=2)


Appendix 3-8: Calibration certificate: E-Field Probe (EX3DV4) (cont'd)

EX3DV4- SN:3907



May 12, 2016

Receiving Pattern (ϕ), $\theta = 0^\circ$

f=600 MHz,TEM



f=1800 MHz,R22





Uncertainty of Axial Isotropy Assessment: $\pm 0.5\%$ (k=2)

Appendix 3-8: Calibration certificate: E-Field Probe (EX3DV4) (cont'd)

EX3DV4- SN:3907

May 12, 2016

Dynamic Range f(SAR_{head})
 (TEM cell , f_{eval}= 1900 MHz)




Uncertainty of Linearity Assessment: ± 0.6% (k=2)

Appendix 3-8: Calibration certificate: E-Field Probe (EX3DV4) (cont'd)


EX3DV4- SN:3907

May 12, 2016

Conversion Factor Assessment



Deviation from Isotropy in Liquid
 Error (ϕ , θ), $f = 900$ MHz



Appendix 3-8: Calibration certificate: E-Field Probe (EX3DV4) (cont'd)

EX3DV4- SN:3907

May 12, 2016

DASY/EASY - Parameters of Probe: EX3DV4 - SN:3907

Other Probe Parameters

| | |
|---|------------|
| Sensor Arrangement | Triangular |
| Connector Angle (°) | 112.4 |
| Mechanical Surface Detection Mode | enabled |
| Optical Surface Detection Mode | disabled |
| Probe Overall Length | 337 mm |
| Probe Body Diameter | 10 mm |
| Tip Length | 9 mm |
| Tip Diameter | 2.5 mm |
| Probe Tip to Sensor X Calibration Point | 1 mm |
| Probe Tip to Sensor Y Calibration Point | 1 mm |
| Probe Tip to Sensor Z Calibration Point | 1 mm |
| Recommended Measurement Distance from Surface | 1.4 mm |

Appendix 3-9: Calibration certificate: Dipole (D2450V2)

**Calibration Laboratory of
 Schmid & Partner
 Engineering AG**
 Zeughausstrasse 43, 8004 Zurich, Switzerland



S Schweizerischer Kalibrierdienst
C Service suisse d'étalonnage
S Servizio svizzero di taratura
S Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS)
 The Swiss Accreditation Service is one of the signatories to the EA
 Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 0108**

Client **UL Japan Shonan (Vitec)**

Certificate No: **D2450V2-822_Jan16**

CALIBRATION CERTIFICATE

Object **D2450V2 - SN:822**

Calibration procedure(s) **QA CAL-05.v9
 Calibration procedure for dipole validation kits above 700 MHz**

Calibration date: **January 14, 2016**

This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI).
 The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature (22 ± 3)°C and humidity < 70%.

Calibration Equipment used (M&TE critical for calibration)

| Primary Standards | ID # | Cal Date (Certificate No.) | Scheduled Calibration |
|-----------------------------|--------------------|-----------------------------------|------------------------|
| Power meter EPM-442A | GB37480704 | 07-Oct-15 (No. 217-02222) | Oct-16 |
| Power sensor HP 8481A | US37292783 | 07-Oct-15 (No. 217-02222) | Oct-16 |
| Power sensor HP 8481A | MY41092317 | 07-Oct-15 (No. 217-02223) | Oct-16 |
| Reference 20 dB Attenuator | SN: 5058 (20k) | 01-Apr-'5 (No. 217-02131) | Mar-16 |
| Type-N mismatch combination | SN: 5047.2 / 06327 | 01-Apr-'5 (No. 217-02134) | Mar-16 |
| Reference Probe EX3DV4 | SN: 7349 | 31-Dec-15 (No. EX3-7349_Dec15) | Dec-16 |
| DAE4 | SN: 601 | 30-Dec-15 (No. DAE4-601_Dec15) | Dec-16 |
| Secondary Standards | ID # | Check Date (in house) | Scheduled Check |
| RF generator R&S SMT-06 | 100972 | 15-Jun-15 (in house check Jun-15) | In house check: Jun-18 |
| Network Analyzer HP 8753E | US37390585 S4206 | 18-Oct-01 (in house check Oct-15) | In house check: Oct-16 |

Calibrated by: **Name: Michael Weber, Function: Laboratory Technician, Signature: [Handwritten Signature]**

Approved by: **Name: Katja Pokovic, Function: Technical Manager, Signature: [Handwritten Signature]**

Issued: January 15, 2016

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

Appendix 3-9: Calibration certificate: Dipole (D2450V2) (cont'd)

**Calibration Laboratory of
Schmid & Partner
Engineering AG**
Zeughausstrasse 43, 8004 Zurich, Switzerland



S Schweizerischer Kalibrierdienst
C Service suisse d'étalonnage
S Servizio svizzero di taratura
S Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS)
The Swiss Accreditation Service is one of the signatories to the EA
Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 0108**

Glossary:

TSL tissue simulating liquid
ConvF sensitivity in TSL / NORM x,y,z
N/A not applicable or not measured

Calibration is Performed According to the Following Standards:

- a) IEEE Std 1528-2013, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", June 2013
- b) IEC 62209-1, "Procedure to measure the Specific Absorption Rate (SAR) for hand-held devices used in close proximity to the ear (frequency range of 300 MHz to 3 GHz)", February 2005
- c) IEC 62209-2, "Procedure to determine the Specific Absorption Rate (SAR) for wireless communication devices used in close proximity to the human body (frequency range of 30 MHz to 6 GHz)", March 2010
- d) KDB 865664, "SAR Measurement Requirements for 100 MHz to 6 GHz"

Additional Documentation:

- e) DASY4/5 System Handbook

Methods Applied and Interpretation of Parameters:

- *Measurement Conditions:* Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- *Antenna Parameters with TSL:* The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.
- *Feed Point Impedance and Return Loss:* These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.
- *Electrical Delay:* One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
- *SAR measured:* SAR measured at the stated antenna input power.
- *SAR normalized:* SAR as measured, normalized to an input power of 1 W at the antenna connector.
- *SAR for nominal TSL parameters:* The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor $k=2$, which for a normal distribution corresponds to a coverage probability of approximately 95%.

Appendix 3-9: Calibration certificate: Dipole (D2450V2) (cont'd)

Measurement Conditions

DASY system configuration, as far as not given on page 1.

| | | |
|------------------------------|------------------------|-------------|
| DASY Version | DASY5 | V52.8.8 |
| Extrapolation | Advanced Extrapolation | |
| Phantom | Modular Flat Phantom | |
| Distance Dipole Center - TSL | 10 mm | with Spacer |
| Zoom Scan Resolution | dx, dy, dz = 5 mm | |
| Frequency | 2450 MHz ± 1 MHz | |

Head TSL parameters

The following parameters and calculations were applied.

| | Temperature | Permittivity | Conductivity |
|---|-----------------|--------------|------------------|
| Nominal Head TSL parameters | 22.0 °C | 39.2 | 1.80 mho/m |
| Measured Head TSL parameters | (22.0 ± 0.2) °C | 37.8 ± 6 % | 1.87 mho/m ± 6 % |
| Head TSL temperature change during test | < 0.5 °C | ---- | ---- |

SAR result with Head TSL

| SAR averaged over 1 cm ³ (1 g) of Head TSL | Condition | |
|---|--------------------|--------------------------|
| SAR measured | 250 mW input power | 13.2 W/kg |
| SAR for nominal Head TSL parameters | normalized to 1W | 51.4 W/kg ± 17.0 % (k=2) |

| SAR averaged over 10 cm ³ (10 g) of Head TSL | condition | |
|---|--------------------|--------------------------|
| SAR measured | 250 mW input power | 6.09 W/kg |
| SAR for nominal Head TSL parameters | normalized to 1W | 24.0 W/kg ± 16.5 % (k=2) |

Body TSL parameters

The following parameters and calculations were applied

| | Temperature | Permittivity | Conductivity |
|---|-----------------|--------------|------------------|
| Nominal Body TSL parameters | 22.0 °C | 52.7 | 1.95 mho/m |
| Measured Body TSL parameters | (22.0 ± 0.2) °C | 52.1 ± 6 % | 2.04 mho/m ± 6 % |
| Body TSL temperature change during test | < 0.5 °C | ---- | ---- |

SAR result with Body TSL

| SAR averaged over 1 cm ³ (1 g) of Body TSL | Condition | |
|---|--------------------|--------------------------|
| SAR measured | 250 mW input power | 13.1 W/kg |
| SAR for nominal Body TSL parameters | normalized to 1W | 51.2 W/kg ± 17.0 % (k=2) |

| SAR averaged over 10 cm ³ (10 g) of Body TSL | condition | |
|---|--------------------|--------------------------|
| SAR measured | 250 mW input power | 6.12 W/kg |
| SAR for nominal Body TSL parameters | normalized to 1W | 24.2 W/kg ± 16.5 % (k=2) |

Appendix 3-9: Calibration certificate: Dipole (D2450V2) (cont'd)

Appendix (Additional assessments outside the scope of SCS 0108)

Antenna Parameters with Head TSL

| | |
|--------------------------------------|--------------------------------|
| Impedance, transformed to feed point | 54.9 Ω + 4.1 j Ω |
| Return Loss | - 24.3 dB |

Antenna Parameters with Body TSL

| | |
|--------------------------------------|--------------------------------|
| Impedance, transformed to feed point | 50.6 Ω + 6.3 j Ω |
| Return Loss | - 24.1 dB |

General Antenna Parameters and Design

| | |
|----------------------------------|----------|
| Electrical Delay (one direction) | 1.158 ns |
|----------------------------------|----------|

After long term use with 100W radiated power, only a slight warming of the dipole near the feedpoint can be measured.

The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard.

No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feedpoint may be damaged.

Additional EUT Data

| | |
|-----------------|-------------------|
| Manufactured by | SPEAG |
| Manufactured on | December 11, 2008 |

Appendix 3-9: Calibration certificate: Dipole (D2450V2) (cont'd)

DASY5 Validation Report for Head TSL

Date: 14.01.2016

Test Laboratory: SPEAG, Zurich, Switzerland

DUT: Dipole 2450 MHz D2450V2; Type: D2450V2; Serial: D2450V2 - SN: 822


Communication System: UID 0 - CW; Frequency: 2450 MHz
Medium parameters used: $f = 2450$ MHz; $\sigma = 1.87$ S/m; $\epsilon_r = 37.8$; $\rho = 1000$ kg/m³
Phantom section: Flat Section
Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2011)

DASY52 Configuration:

- Probe: EX3DV4 - SN7349; ConvF(7.76, 7.76, 7.76); Calibrated: 31.12.2015;
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn601; Calibrated: 30.12.2015
- Phantom: Flat Phantom 5.0 (front); Type: QD000P50AA; Serial: 1001
- DASY52 52.8.8(1258); SEMCAD X 14.6.10(7372)


Dipole Calibration for Head Tissue/Pin=250 mW, d=10mm/Zoom Scan (7x7x7)/Cube 0:

Measurement grid: dx=5mm, dy=5mm, dz=5mm
Reference Value = 114.0 V/m; Power Drift = 0.03 dB
Peak SAR (extrapolated) = 27.0 W/kg
SAR(1 g) = 13.2 W/kg; SAR(10 g) = 6.09 W/kg
Maximum value of SAR (measured) = 21.9 W/kg



Appendix 3-9: Calibration certificate: Dipole (D2450V2) (cont'd)

Impedance Measurement Plot for Head TSL



Appendix 3-9: Calibration certificate: Dipole (D2450V2) (cont'd)

DASY5 Validation Report for Body TSL

Date: 14.01.2016

Test Laboratory: SPEAG, Zurich, Switzerland

DUT: Dipole 2450 MHz D2450V2; Type: D2450V2; Serial: D2450V2 - SN: 822

Communication System: UID 0 - CW; Frequency: 2450 MHz

Medium parameters used: $f = 2450$ MHz; $\sigma = 2.04$ S/m; $\epsilon_r = 52.1$; $\rho = 1000$ kg/m³

Phantom section: Flat Section

Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2011)

DASY52 Configuration:

- Probe: EX3DV4 - SN7349; ConvF(7.79, 7.79, 7.79); Calibrated: 31.12.2015;
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn601; Calibrated: 30.12.2015
- Phantom: Flat Phantom 5.0 (back); Type: QD000P50AA; Serial: 1002
- DASY52 52.8.8(1258); SEMCAD X 14.6.10(7372)

Dipole Calibration for Body Tissue/Pin=250 mW, d=10mm/Zoom Scan (7x7x7)/Cube 0:


Measurement grid: dx=5mm, dy=5mm, dz=5mm

Reference Value = 106.0 V/m; Power Drift = 0.01 dB

Peak SAR (extrapolated) = 26.3 W/kg


SAR(1 g) = 13.1 W/kg; SAR(10 g) = 6.12 W/kg

Maximum value of SAR (measured) = 21.2 W/kg



Appendix 3-9: Calibration certificate: Dipole (D2450V2) (cont'd)

Impedance Measurement Plot for Body TSL



Appendix 3-10: Calibration certificate: Dipole (D5GHzV2)

**Calibration Laboratory of
 Schmid & Partner
 Engineering AG**
 Zeughausstrasse 43, 8004 Zurich, Switzerland



S Schweizerischer Kalibrierdienst
C Service suisse d'étalonnage
S Servizio svizzero di taratura
S Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS)
 The Swiss Accreditation Service is one of the signatories to the EA
 Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 0108**

Client **UL Japan Shonan (Vitec)**

Certificate No: **D5GHzV2-1070_Mar16**

| CALIBRATION CERTIFICATE | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
|---|--|-----------------------------------|------------------------|-------------------|------|----------------------------|-----------------------|----------------------|------------|---------------------------|--------|-----------------------|------------|---------------------------|--------|-----------------------|------------|---------------------------|--------|----------------------------|----------------|---------------------------|--------|-----------------------------|--------------------|---------------------------|--------|------------------------|----------|--------------------------------|--------|------|---------|--------------------------------|--------|---------------------|------|-----------------------|-----------------|-------------------------|--------|-----------------------------------|------------------------|---------------------------|------------------|-----------------------------------|------------------------|
| Object | D5GHzV2 - SN:1070 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Calibration procedure(s) | QA CAL-22.v2 Calibration procedure for dipole validation kits between 3-6 GHz | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Calibration date: | March 10, 2016 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| <p>This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.</p> <p>All calibrations have been conducted in the closed laboratory facility: environment temperature (22 ± 3)°C and humidity < 70%.</p> <p>Calibration Equipment used (M&TE critical for calibration)</p> <table border="1"> <thead> <tr> <th>Primary Standards</th> <th>ID #</th> <th>Cal Date (Certificate No.)</th> <th>Scheduled Calibration</th> </tr> </thead> <tbody> <tr> <td>Power meter EPM-442A</td> <td>GB37480704</td> <td>07-Oct-15 (No. 217-02222)</td> <td>Oct-16</td> </tr> <tr> <td>Power sensor HP 8481A</td> <td>US37292783</td> <td>07-Oct-15 (No. 217-02222)</td> <td>Oct-16</td> </tr> <tr> <td>Power sensor HP 8481A</td> <td>MY41092317</td> <td>07-Oct-15 (No. 217-02223)</td> <td>Oct-16</td> </tr> <tr> <td>Reference 20 dB Attenuator</td> <td>SN: 5058 (20k)</td> <td>01-Apr-15 (No. 217-02131)</td> <td>Mar-16</td> </tr> <tr> <td>Type-N mismatch combination</td> <td>SN: 5047.2 / 06327</td> <td>01-Apr-15 (No. 217-02134)</td> <td>Mar-16</td> </tr> <tr> <td>Reference Probe EX3DV4</td> <td>SN: 3503</td> <td>31-Dec-15 (No. EX3-3503_Dec15)</td> <td>Dec-16</td> </tr> <tr> <td>DAE4</td> <td>SN: 601</td> <td>30-Dec-15 (No. DAE4-601_Dec15)</td> <td>Dec-16</td> </tr> </tbody> </table> <table border="1"> <thead> <tr> <th>Secondary Standards</th> <th>ID #</th> <th>Check Date (in house)</th> <th>Scheduled Check</th> </tr> </thead> <tbody> <tr> <td>RF generator R&S SMT-06</td> <td>100972</td> <td>15-Jun-15 (in house check Jun-15)</td> <td>In house check: Jun-16</td> </tr> <tr> <td>Network Analyzer HP 8753E</td> <td>US37390565 S4206</td> <td>18-Oct-01 (in house check Oct-15)</td> <td>In house check: Oct-16</td> </tr> </tbody> </table> | | | | Primary Standards | ID # | Cal Date (Certificate No.) | Scheduled Calibration | Power meter EPM-442A | GB37480704 | 07-Oct-15 (No. 217-02222) | Oct-16 | Power sensor HP 8481A | US37292783 | 07-Oct-15 (No. 217-02222) | Oct-16 | Power sensor HP 8481A | MY41092317 | 07-Oct-15 (No. 217-02223) | Oct-16 | Reference 20 dB Attenuator | SN: 5058 (20k) | 01-Apr-15 (No. 217-02131) | Mar-16 | Type-N mismatch combination | SN: 5047.2 / 06327 | 01-Apr-15 (No. 217-02134) | Mar-16 | Reference Probe EX3DV4 | SN: 3503 | 31-Dec-15 (No. EX3-3503_Dec15) | Dec-16 | DAE4 | SN: 601 | 30-Dec-15 (No. DAE4-601_Dec15) | Dec-16 | Secondary Standards | ID # | Check Date (in house) | Scheduled Check | RF generator R&S SMT-06 | 100972 | 15-Jun-15 (in house check Jun-15) | In house check: Jun-16 | Network Analyzer HP 8753E | US37390565 S4206 | 18-Oct-01 (in house check Oct-15) | In house check: Oct-16 |
| Primary Standards | ID # | Cal Date (Certificate No.) | Scheduled Calibration | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Power meter EPM-442A | GB37480704 | 07-Oct-15 (No. 217-02222) | Oct-16 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Power sensor HP 8481A | US37292783 | 07-Oct-15 (No. 217-02222) | Oct-16 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Power sensor HP 8481A | MY41092317 | 07-Oct-15 (No. 217-02223) | Oct-16 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Reference 20 dB Attenuator | SN: 5058 (20k) | 01-Apr-15 (No. 217-02131) | Mar-16 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Type-N mismatch combination | SN: 5047.2 / 06327 | 01-Apr-15 (No. 217-02134) | Mar-16 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Reference Probe EX3DV4 | SN: 3503 | 31-Dec-15 (No. EX3-3503_Dec15) | Dec-16 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| DAE4 | SN: 601 | 30-Dec-15 (No. DAE4-601_Dec15) | Dec-16 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Secondary Standards | ID # | Check Date (in house) | Scheduled Check | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| RF generator R&S SMT-06 | 100972 | 15-Jun-15 (in house check Jun-15) | In house check: Jun-16 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Network Analyzer HP 8753E | US37390565 S4206 | 18-Oct-01 (in house check Oct-15) | In house check: Oct-16 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Calibrated by: | Name Jeton Kastrati | Function Laboratory Technician | Signature | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Approved by: | Name Katja Pokovic | Function Technical Manager | Signature | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Issued: March 11, 2016 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| This calibration certificate shall not be reproduced except in full without written approval of the laboratory. | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |

Appendix 3-10: Calibration certificate: Dipole (D5GHzV2) (cont'd)

Calibration Laboratory of
Schmid & Partner
Engineering AG
Zeughausstrasse 43, 8004 Zurich, Switzerland



S Schweizerischer Kalibrierdienst
C Service suisse d'étalonnage
S Servizio svizzero di taratura
S Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS)
The Swiss Accreditation Service is one of the signatories to the EA
Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: SCS 0108

Glossary:

| | |
|-------|---------------------------------|
| TSL | tissue simulating liquid |
| ConvF | sensitivity in TSL / NORM x,y,z |
| N/A | not applicable or not measured |

Calibration is Performed According to the Following Standards:

- IEEE Std 1528-2013, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", June 2013
- IEC 62209-2, "Procedure to determine the Specific Absorption Rate (SAR) for wireless communication devices used in close proximity to the human body (frequency range of 30 MHz to 6 GHz)", March 2010
- KDB 865664, "SAR Measurement Requirements for 100 MHz to 6 GHz"

Additional Documentation:

- DASY4/5 System Handbook

Methods Applied and Interpretation of Parameters:

- Measurement Conditions:** Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL:** The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.
- Feed Point Impedance and Return Loss:** These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.
- Electrical Delay:** One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
- SAR measured:** SAR measured at the stated antenna input power.
- SAR normalized:** SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters:** The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor $k=2$, which for a normal distribution corresponds to a coverage probability of approximately 95%.

Appendix 3-10: Calibration certificate: Dipole (D5GHzV2) (cont'd)

Measurement Conditions

DASY system configuration, as far as not given on page 1.

| | | |
|------------------------------|--|----------------------------------|
| DASY Version | DASY5 | V52.8.8 |
| Extrapolation | Advanced Extrapolation | |
| Phantom | Modular Flat Phantom V5.0 | |
| Distance Dipole Center - TSL | 10 mm | with Spacer |
| Zoom Scan Resolution | dx, dy = 4.0 mm, dz = 1.4 mm | Graded Ratio = 1.4 (Z direction) |
| Frequency | 5200 MHz ± 1 MHz 5250 MHz ± 1 MHz 5500 MHz ± 1 MHz 5600 MHz ± 1 MHz 5750 MHz ± 1 MHz 5800 MHz ± 1 MHz | |

Head TSL parameters at 5200 MHz

The following parameters and calculations were applied.

| | Temperature | Permittivity | Conductivity |
|---|-----------------|--------------|------------------|
| Nominal Head TSL parameters | 22.0 °C | 36.0 | 4.66 mho/m |
| Measured Head TSL parameters | (22.0 ± 0.2) °C | 34.8 ± 6 % | 4.49 mho/m ± 6 % |
| Head TSL temperature change during test | < 0.5 °C | ---- | ---- |

SAR result with Head TSL at 5200 MHz

| SAR averaged over 1 cm ³ (1 g) of Head TSL | Condition | |
|---|--------------------|---------------------------------|
| SAR measured | 100 mW input power | 7.44 W/kg |
| SAR for nominal Head TSL parameters | normalized to 1W | 73.8 W/kg ± 19.9 % (k=2) |

| SAR averaged over 10 cm ³ (10 g) of Head TSL | condition | |
|---|--------------------|---------------------------------|
| SAR measured | 100 mW input power | 2.15 W/kg |
| SAR for nominal Head TSL parameters | normalized to 1W | 21.3 W/kg ± 19.5 % (k=2) |

Appendix 3-10: Calibration certificate: Dipole (D5GHzV2) (cont'd)

Head TSL parameters at 5250 MHz

The following parameters and calculations were applied.

| | Temperature | Permittivity | Conductivity |
|---|-----------------|--------------|------------------|
| Nominal Head TSL parameters | 22.0 °C | 35.9 | 4.71 mho/m |
| Measured Head TSL parameters | (22.0 ± 0.2) °C | 34.7 ± 6 % | 4.53 mho/m ± 6 % |
| Head TSL temperature change during test | < 0.5 °C | ---- | ---- |

SAR result with Head TSL at 5250 MHz

| SAR averaged over 1 cm ³ (1 g) of Head TSL | Condition | |
|---|--------------------|----------------------------|
| SAR measured | 100 mW input power | 7.59 W/kg |
| SAR for nominal Head TSL parameters | normalized to 1W | 75.3 W / kg ± 19.9 % (k=2) |

| SAR averaged over 10 cm ³ (10 g) of Head TSL | condition | |
|---|--------------------|--------------------------|
| SAR measured | 100 mW input power | 2.19 W/kg |
| SAR for nominal Head TSL parameters | normalized to 1W | 21.7 W/kg ± 19.5 % (k=2) |

Head TSL parameters at 5500 MHz

The following parameters and calculations were applied.

| | Temperature | Permittivity | Conductivity |
|---|-----------------|--------------|------------------|
| Nominal Head TSL parameters | 22.0 °C | 35.6 | 4.96 mho/m |
| Measured Head TSL parameters | (22.0 ± 0.2) °C | 34.4 ± 6 % | 4.77 mho/m ± 6 % |
| Head TSL temperature change during test | < 0.5 °C | ---- | ---- |

SAR result with Head TSL at 5500 MHz

| SAR averaged over 1 cm ³ (1 g) of Head TSL | Condition | |
|---|--------------------|--------------------------|
| SAR measured | 100 mW input power | 7.71 W/kg |
| SAR for nominal Head TSL parameters | normalized to 1W | 76.4 W/kg ± 19.9 % (k=2) |

| SAR averaged over 10 cm ³ (10 g) of Head TSL | condition | |
|---|--------------------|--------------------------|
| SAR measured | 100 mW input power | 2.21 W/kg |
| SAR for nominal Head TSL parameters | normalized to 1W | 21.9 W/kg ± 19.5 % (k=2) |

Appendix 3-10: Calibration certificate: Dipole (D5GHzV2) (cont'd)

Head TSL parameters at 5600 MHz

The following parameters and calculations were applied.

| | Temperature | Permittivity | Conductivity |
|---|-----------------|--------------|------------------|
| Nominal Head TSL parameters | 22.0 °C | 35.5 | 5.07 mho/m |
| Measured Head TSL parameters | (22.0 ± 0.2) °C | 34.2 ± 6 % | 4.87 mho/m ± 6 % |
| Head TSL temperature change during test | < 0.5 °C | ---- | ---- |

SAR result with Head TSL at 5600 MHz

| SAR averaged over 1 cm ³ (1 g) of Head TSL | Condition | |
|---|--------------------|--------------------------|
| SAR measured | 100 mW input power | 7.93 W/kg |
| SAR for nominal Head TSL parameters | normalized to 1W | 78.6 W/kg ± 19.9 % (k=2) |

| SAR averaged over 10 cm ³ (10 g) of Head TSL | condition | |
|---|--------------------|--------------------------|
| SAR measured | 100 mW input power | 2.28 W/kg |
| SAR for nominal Head TSL parameters | normalized to 1W | 22.5 W/kg ± 19.5 % (k=2) |

Head TSL parameters at 5750 MHz

The following parameters and calculations were applied.

| | Temperature | Permittivity | Conductivity |
|---|-----------------|--------------|------------------|
| Nominal Head TSL parameters | 22.0 °C | 35.4 | 5.22 mho/m |
| Measured Head TSL parameters | (22.0 ± 0.2) °C | 34.0 ± 6 % | 5.03 mho/m ± 6 % |
| Head TSL temperature change during test | < 0.5 °C | ---- | ---- |

SAR result with Head TSL at 5750 MHz

| SAR averaged over 1 cm ³ (1 g) of Head TSL | Condition | |
|---|--------------------|--------------------------|
| SAR measured | 100 mW input power | 7.61 W/kg |
| SAR for nominal Head TSL parameters | normalized to 1W | 75.3 W/kg ± 19.9 % (k=2) |

| SAR averaged over 10 cm ³ (10 g) of Head TSL | condition | |
|---|--------------------|--------------------------|
| SAR measured | 100 mW input power | 2.18 W/kg |
| SAR for nominal Head TSL parameters | normalized to 1W | 21.5 W/kg ± 19.5 % (k=2) |

Appendix 3-10: Calibration certificate: Dipole (D5GHzV2) (cont'd)

Head TSL parameters at 5800 MHz

The following parameters and calculations were applied.

| | Temperature | Permittivity | Conductivity |
|---|-----------------|--------------|------------------|
| Nominal Head TSL parameters | 22.0 °C | 35.3 | 5.27 mho/m |
| Measured Head TSL parameters | (22.0 ± 0.2) °C | 34.0 ± 6 % | 5.08 mho/m ± 6 % |
| Head TSL temperature change during test | < 0.5 °C | ---- | ---- |

SAR result with Head TSL at 5800 MHz

| SAR averaged over 1 cm ³ (1 g) of Head TSL | Condition | |
|---|--------------------|--------------------------|
| SAR measured | 100 mW input power | 7.59 W/kg |
| SAR for nominal Head TSL parameters | normalized to 1W | 75.2 W/kg ± 19.9 % (k=2) |

| SAR averaged over 10 cm ³ (10 g) of Head TSL | condition | |
|---|--------------------|--------------------------|
| SAR measured | 100 mW input power | 2.17 W/kg |
| SAR for nominal Head TSL parameters | normalized to 1W | 21.5 W/kg ± 19.5 % (k=2) |

Appendix 3-10: Calibration certificate: Dipole (D5GHzV2) (cont'd)

Body TSL parameters at 5250 MHz

The following parameters and calculations were applied.

| | Temperature | Permittivity | Conductivity |
|---|-----------------|--------------|------------------|
| Nominal Body TSL parameters | 22.0 °C | 48.9 | 5.36 mho/m |
| Measured Body TSL parameters | (22.0 ± 0.2) °C | 47.3 ± 6 % | 5.43 mho/m ± 6 % |
| Body TSL temperature change during test | < 0.5 °C | ---- | ---- |

SAR result with Body TSL at 5250 MHz

| SAR averaged over 1 cm ³ (1 g) of Body TSL | Condition | |
|---|--------------------|--------------------------|
| SAR measured | 100 mW input power | 7.27 W/kg |
| SAR for nominal Body TSL parameters | normalized to 1W | 72.2 W/kg ± 19.9 % (k=2) |

| SAR averaged over 10 cm ³ (10 g) of Body TSL | condition | |
|---|--------------------|--------------------------|
| SAR measured | 100 mW input power | 2.06 W/kg |
| SAR for nominal Body TSL parameters | normalized to 1W | 20.4 W/kg ± 19.5 % (k=2) |

Body TSL parameters at 5600 MHz

The following parameters and calculations were applied.

| | Temperature | Permittivity | Conductivity |
|---|-----------------|--------------|------------------|
| Nominal Body TSL parameters | 22.0 °C | 48.5 | 5.77 mho/m |
| Measured Body TSL parameters | (22.0 ± 0.2) °C | 46.7 ± 6 % | 5.90 mho/m ± 6 % |
| Body TSL temperature change during test | < 0.5 °C | ---- | ---- |

SAR result with Body TSL at 5600 MHz

| SAR averaged over 1 cm ³ (1 g) of Body TSL | Condition | |
|---|--------------------|--------------------------|
| SAR measured | 100 mW input power | 7.82 W/kg |
| SAR for nominal Body TSL parameters | normalized to 1W | 77.7 W/kg ± 19.9 % (k=2) |

| SAR averaged over 10 cm ³ (10 g) of Body TSL | condition | |
|---|--------------------|--------------------------|
| SAR measured | 100 mW input power | 2.20 W/kg |
| SAR for nominal Body TSL parameters | normalized to 1W | 21.8 W/kg ± 19.5 % (k=2) |

Appendix 3-10: Calibration certificate: Dipole (D5GHzV2) (cont'd)

Body TSL parameters at 5750 MHz

The following parameters and calculations were applied.

| | Temperature | Permittivity | Conductivity |
|---|-----------------|--------------|------------------|
| Nominal Body TSL parameters | 22.0 °C | 48.3 | 5.94 mho/m |
| Measured Body TSL parameters | (22.0 ± 0.2) °C | 46.4 ± 6 % | 6.12 mho/m ± 6 % |
| Body TSL temperature change during test | < 0.5 °C | ---- | ---- |

SAR result with Body TSL at 5750 MHz

| SAR averaged over 1 cm ³ (1 g) of Body TSL | Condition | |
|---|--------------------|---------------------------------|
| SAR measured | 100 mW input power | 7.45 W/kg |
| SAR for nominal Body TSL parameters | normalized to 1W | 74.0 W/kg ± 19.9 % (k=2) |

| SAR averaged over 10 cm ³ (10 g) of Body TSL | condition | |
|---|--------------------|---------------------------------|
| SAR measured | 100 mW input power | 2.09 W/kg |
| SAR for nominal Body TSL parameters | normalized to 1W | 20.7 W/kg ± 19.5 % (k=2) |

Appendix 3-10: Calibration certificate: Dipole (D5GHzV2) (cont'd)

Appendix (Additional assessments outside the scope of SCS 0108)

Antenna Parameters with Head TSL at 5200 MHz

| | |
|--------------------------------------|---------------------------------|
| Impedance, transformed to feed point | 51.1 Ω - 13.1 j Ω |
| Return Loss | - 17.8 dB |

Antenna Parameters with Head TSL at 5250 MHz

| | |
|--------------------------------------|--------------------------------|
| Impedance, transformed to feed point | 52.0 Ω - 8.8 j Ω |
| Return Loss | - 21.1 dB |

Antenna Parameters with Head TSL at 5500 MHz

| | |
|--------------------------------------|--------------------------------|
| Impedance, transformed to feed point | 48.0 Ω - 8.7 j Ω |
| Return Loss | - 20.8 dB |

Antenna Parameters with Head TSL at 5600 MHz

| | |
|--------------------------------------|--------------------------------|
| Impedance, transformed to feed point | 55.5 Ω - 8.5 j Ω |
| Return Loss | - 20.4 dB |

Antenna Parameters with Head TSL at 5750 MHz

| | |
|--------------------------------------|--------------------------------|
| Impedance, transformed to feed point | 55.3 Ω - 2.7 j Ω |
| Return Loss | - 25.0 dB |

Antenna Parameters with Head TSL at 5800 MHz

| | |
|--------------------------------------|--------------------------------|
| Impedance, transformed to feed point | 53.4 Ω - 3.2 j Ω |
| Return Loss | - 26.9 dB |

Antenna Parameters with Body TSL at 5250 MHz

| | |
|--------------------------------------|--------------------------------|
| Impedance, transformed to feed point | 52.5 Ω - 6.6 j Ω |
| Return Loss | - 23.3 dB |

Antenna Parameters with Body TSL at 5600 MHz

| | |
|--------------------------------------|--------------------------------|
| Impedance, transformed to feed point | 56.8 Ω - 7.0 j Ω |
| Return Loss | - 20.8 dB |

Antenna Parameters with Body TSL at 5750 MHz

| | |
|--------------------------------------|--------------------------------|
| Impedance, transformed to feed point | 57.6 Ω - 1.0 j Ω |
| Return Loss | - 23.0 dB |

Appendix 3-10: Calibration certificate: Dipole (D5GHzV2) (cont'd)

General Antenna Parameters and Design

| | |
|----------------------------------|----------|
| Electrical Delay (one direction) | 1.202 ns |
|----------------------------------|----------|

After long term use with 100W radiated power, only a slight warming of the dipole near the feedpoint can be measured.

The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard.

No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feedpoint may be damaged.

Additional EUT Data

| | |
|-----------------|--------------------|
| Manufactured by | SPEAG |
| Manufactured on | September 26, 2008 |

Appendix 3-10: Calibration certificate: Dipole (D5GHzV2) (cont'd)

DASY5 Validation Report for Head TSL

Date: 07.03.2016

Test Laboratory: SPEAG, Zurich, Switzerland

DUT: Dipole D5GHzV2; Type: D5GHzV2; Serial: D5GHzV2 - SN:1070

Communication System: UID 0 - CW; Frequency: 5200 MHz, Frequency: 5250 MHz, Frequency: 5500 MHz, Frequency: 5600 MHz, Frequency: 5750 MHz, Frequency: 5800 MHz
Medium parameters used: $f = 5200$ MHz; $\sigma = 4.49$ S/m; $\epsilon_r = 34.8$; $\rho = 1000$ kg/m³, Medium parameters used: $f = 5250$ MHz; $\sigma = 4.53$ S/m; $\epsilon_r = 34.7$; $\rho = 1000$ kg/m³, Medium parameters used: $f = 5500$ MHz; $\sigma = 4.77$ S/m; $\epsilon_r = 34.4$; $\rho = 1000$ kg/m³, Medium parameters used: $f = 5600$ MHz; $\sigma = 4.87$ S/m; $\epsilon_r = 34.2$; $\rho = 1000$ kg/m³, Medium parameters used: $f = 5750$ MHz; $\sigma = 5.03$ S/m; $\epsilon_r = 34$; $\rho = 1000$ kg/m³, Medium parameters used: $f = 5800$ MHz; $\sigma = 5.08$ S/m; $\epsilon_r = 34$; $\rho = 1000$ kg/m³
Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2011)

DASY52 Configuration:

- Probe: EX3DV4 - SN3503; ConvF(5.59, 5.59, 5.59); Calibrated: 31.12.2015, ConvF(5.53, 5.53, 5.53); Calibrated: 31.12.2015, ConvF(5.18, 5.18, 5.18); Calibrated: 31.12.2015, ConvF(4.99, 4.99, 4.99); Calibrated: 31.12.2015, ConvF(4.95, 4.95, 4.95); Calibrated: 31.12.2015, ConvF(4.95, 4.95, 4.95); Calibrated: 31.12.2015;
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn601; Calibrated: 30.12.2015
- Phantom Type: QD000P50AA
- DASY52 52.8.8(1258); SEMCAD X 14.6.10(7372)

Dipole Calibration for Head Tissue/Pin=100mW, dist=10mm, f=5200 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm
Reference Value = 70.46 V/m; Power Drift = -0.02 dB
Peak SAR (extrapolated) = 26.9 W/kg
SAR(1 g) = 7.44 W/kg; SAR(10 g) = 2.15 W/kg
Maximum value of SAR (measured) = 17.5 W/kg

Dipole Calibration for Head Tissue/Pin=100mW, dist=10mm, f=5250 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm
Reference Value = 71.40 V/m; Power Drift = 0.03 dB
Peak SAR (extrapolated) = 27.6 W/kg
SAR(1 g) = 7.59 W/kg; SAR(10 g) = 2.19 W/kg
Maximum value of SAR (measured) = 17.9 W/kg


Appendix 3-10: Calibration certificate: Dipole (D5GHzV2) (cont'd)

Dipole Calibration for Head Tissue/Pin=100mW, dist=10mm, f=5500 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm
Reference Value = 70.66 V/m; Power Drift = 0.05 dB
Peak SAR (extrapolated) = 30.1 W/kg
SAR(1 g) = 7.71 W/kg; SAR(10 g) = 2.21 W/kg
Maximum value of SAR (measured) = 18.7 W/kg

Dipole Calibration for Head Tissue/Pin=100mW, dist=10mm, f=5600 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm
Reference Value = 71.61 V/m; Power Drift = 0.05 dB
Peak SAR (extrapolated) = 31.0 W/kg
SAR(1 g) = 7.93 W/kg; SAR(10 g) = 2.28 W/kg
Maximum value of SAR (measured) = 19.3 W/kg


Dipole Calibration for Head Tissue/Pin=100mW, dist=10mm, f=5750 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm
Reference Value = 69.14 V/m; Power Drift = 0.06 dB
Peak SAR (extrapolated) = 31.0 W/kg
SAR(1 g) = 7.61 W/kg; SAR(10 g) = 2.18 W/kg
Maximum value of SAR (measured) = 18.8 W/kg

Dipole Calibration for Head Tissue/Pin=100mW, dist=10mm, f=5800 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm
Reference Value = 69.07 V/m; Power Drift = 0.06 dB
Peak SAR (extrapolated) = 31.2 W/kg
SAR(1 g) = 7.59 W/kg; SAR(10 g) = 2.17 W/kg
Maximum value of SAR (measured) = 18.9 W/kg




Appendix 3-10: Calibration certificate: Dipole (D5GHzV2) (cont'd)

Impedance Measurement Plot for Head TSL (5200, 5500, 5750)



Appendix 3-10: Calibration certificate: Dipole (D5GHzV2) (cont'd)

Impedance Measurement Plot for Head TSL (5250, 5600, 5800)



Appendix 3-10: Calibration certificate: Dipole (D5GHzV2) (cont'd)

DASY5 Validation Report for Body TSL

Date: 10.03.2016

Test Laboratory: SPEAG, Zurich, Switzerland

DUT: Dipole D5GHzV2; Type: D5GHzV2; Serial: D5GHzV2 - SN:1070

Communication System: UID 0 - CW; Frequency: 5250 MHz, Frequency: 5600 MHz, Frequency: 5750 MHz
Medium parameters used: $f = 5250$ MHz; $\sigma = 5.43$ S/m; $\epsilon_r = 47.3$; $\rho = 1000$ kg/m³, Medium parameters used: $f = 5600$ MHz; $\sigma = 5.9$ S/m; $\epsilon_r = 46.7$; $\rho = 1000$ kg/m³, Medium parameters used: $f = 5750$ MHz; $\sigma = 6.12$ S/m; $\epsilon_r = 46.4$; $\rho = 1000$ kg/m³
Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2011)

DASY52 Configuration:


- Probe: EX3DV4 - SN3503; ConvF(4.85, 4.85, 4.85); Calibrated: 31.12.2015, ConvF(4.35, 4.35, 4.35); Calibrated: 31.12.2015, ConvF(4.3, 4.3, 4.3); Calibrated: 31.12.2015;
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn601; Calibrated: 30.12.2015
- Phantom Type: QD000P50AA
- DASY52 52.8.8(1258); SEMCAD X 14.6.10(7372)

Dipole Calibration for Body Tissue/Pin=100mW, dist=10mm, f=5250 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm
Reference Value = 65.77 V/m; Power Drift = -0.01 dB
Peak SAR (extrapolated) = 27.4 W/kg
SAR(1 g) = 7.27 W/kg; SAR(10 g) = 2.06 W/kg
Maximum value of SAR (measured) = 16.7 W/kg

Dipole Calibration for Body Tissue/Pin=100mW, dist=10mm, f=5600 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm
Reference Value = 66.43 V/m; Power Drift = 0.04 dB
Peak SAR (extrapolated) = 32.1 W/kg
SAR(1 g) = 7.82 W/kg; SAR(10 g) = 2.2 W/kg
Maximum value of SAR (measured) = 18.5 W/kg

Dipole Calibration for Body Tissue/Pin=100mW, dist=10mm, f=5750 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm
Reference Value = 64.35 V/m; Power Drift = 0.00 dB
Peak SAR (extrapolated) = 32.0 W/kg
SAR(1 g) = 7.45 W/kg; SAR(10 g) = 2.09 W/kg
Maximum value of SAR (measured) = 17.9 W/kg

Appendix 3-10: Calibration certificate: Dipole (D5GHzV2) (cont'd)



Appendix 3-10: Calibration certificate: Dipole (D5GHzV2) (cont'd)

Impedance Measurement Plot for Body TSL

