



166 South Carter, Genoa City, WI 53128

Company: GE Medical Systems, LLC  
Model Tested: 5436008-2 rev1  
Report Number: 17001 (RF Exposure)  
Project Number: 4679

## FCC Code of Federal Regulations 47 Part 1.1307(b) (1)

### RF Exposure Statement of Compliance

THE FOLLOWING **MEETS** THE ABOVE TEST SPECIFICATION

Formal Name: URP Radio

Kind of Equipment: UWB Radio

FCC ID Number: YYJ-5406102

Frequency Range: 3168 - 4752 MHz, 6336 – 7920 MHz

Test Configuration: Hand-held transceiver tested table-top in worst case configuration of three orthogonal planes.

Model Number(s): 5436008 rev3, 5436008-2 rev1

Model(s) Tested: 5436008-2 rev1

Serial Number(s): 11100098

Date of Tests: May 23 – 27, 2011

Test Conducted For: GE Medical Systems, LLC  
3000 N. Grandview Blvd  
Mailstop W622  
Waukesha, WI 53188



166 South Carter, Genoa City, WI 53128

Company: GE Medical Systems, LLC  
Model Tested: 5436008-2 rev1  
Report Number: 17001 (RF Exposure)  
Project Number: 4679

**Maximum EIRP (Peak):**

BG1: -32.57 dBm 4.15 GHz  
BG3: -32.61 dBm 7.51 GHz

**Exposure Limit:**

Maximum Permissible Exposure (MPE) limit for General Population / Uncontrolled Exposure in the frequency range 1500 – 100,000 MHz:

$$S = 1 \text{ mW/cm}^2$$

**MPE Calculation:**

Power Density (mW/cm<sup>2</sup>):

$$S = \frac{PG}{4\pi R^2} \quad \text{or} \quad S = \frac{EIRP}{4\pi R^2}$$

- S = Power Density (mW/cm<sup>2</sup>)
- P = Power Input to the antenna (mW)
- G = Numeric Power Gain of the antenna
- R = Distance to the center of the radiation of the antenna (cm)
- EIRP = Equivalent (effective) isotropically radiated power

**Results:**

**BG1:**  $S = \frac{0.000553 \text{ mW}}{4\pi(20)^2} = 0.000000110 \frac{\text{mW}}{\text{cm}^2}$

**BG3:**  $S = \frac{0.000548 \text{ mW}}{4\pi(20)^2} = 0.000000109 \frac{\text{mW}}{\text{cm}^2}$

**Conclusion:**

The GE Medical Systems, LLC, URP Radio operating under FCC Part 15.519 complies with the requirements of FCC Part 1.1307(b) (1) for RF Exposure Evaluation by a large margin.