

FCC Class II Permissive Change Request Letter

Date: December 9, 2021

To Federal Communications Commission:

RE: FCC Class II Permissive Change Request for Company: Stryker Medical FCC ID: Z7A-6507

We are submitting an application for a class II permissive change to the FCC approval of the Company name: Stryker Medical, product description: Near Field Magnetic Induction Communication Module (FCC ID: Z7A-6507, Original Grant Date: 9/17/2021). The transmitter module itself has not changed. The antenna and rated power have also not changed. Here are the changes:

Updated the User Manual to include installation instructions so that the module can be installed in Stryker devices by third parties.

KDB 996369 D01 allows modules to be end-user plug-and-play replaceable as long as any control that will cause operation of the device in violation of the regulations is not provided. The module is designed to interface with Stryker products using a proprietary connector and the firmware drivers to enable the module are embedded into the host such that the installer and the end user cannot change the operation of the module.

Sincerely,



Divya Murali

Director, Global Regulatory Affairs

Divya.Murali@stryker.com

Medical

3800 E. Centre Avenue, Portage, MI 49002 USA
P +1 269 389 6998