



Excellence in Compliance Testing

5015 B. U. Bowman Dr.  
Buford, GA 30518

ACS TCB  
5015 B.U. Bowman Drive  
Buford, GA 30518

Re: FCC ID: Z9O-FAS1510

To Whom It May Concern:

The following application is submitted on behalf of our client, UltraClenz LLC, for evaluation of their model FAS1510-02 for certification under FCC Part 15.209, 15.231 and 15.249.

The FCC ID: Z9O-FAS1510 PSS Dispenser Beacon will be used in the healthcare industry as part of the UltraClenz's patient safeguard system. The product includes three radio transceivers which transmit at 125 kHz, 433.9 MHz and 2405 MHz, respectively. The FCC ID: Z9O-FAS1510 includes multiple model variants which differ only by the color of their enclosure. The model variants are identified as FAS1510-00, FAS1510-01, FAS1510-02, FAS1510-03, and FAS1510-04.

The radiated emissions measurements were performed with the product set in the orientation of typical installation. The FCC ID: Z9O-FAS1510 was tested in full to the requirements of the aforementioned rules as well for inter-modulation products of the co-located 433.9 MHz and 2405 MHz radios transmitting at the same time and was found to be in compliance.

Sincerely,

Sam Wismer  
Vice President, Technology  
Advanced Compliance Solutions, Inc.