



Excellence in Compliance Testing

5015 B. U. Bowman Dr.
Buford, GA 30518

ACS TCB
5015 B.U. Bowman Drive
Buford, GA 30518

Re: FCC ID: Z9O-FAS1521

To Whom It May Concern:

The following application is submitted on behalf of our client, UltraClenz LLC, for evaluation of their model FAS1521-01 for certification under FCC Part 15.231.

The UltraClenz LLC PSS door monitor, model FAS1521-01, is a 433 MHz battery operated transceiver. The device is designed to track healthcare worker hand hygiene compliance with respect to a specific hospital room, by recording whether or not a soap or sanitizer dispenser was used during patient room entries and exits.

The FCC ID: Z9O-FAS1521 was tested in full to the aforementioned test requirements and was found to be in compliance.

Sincerely,

A handwritten signature in black ink, appearing to read 'R. Sam Wismer'.

Sam Wismer
Vice President, Technology
Advanced Compliance Solutions, Inc.