



Appendix (Additional assessments outside the scope of SCS 0108)

Antenna Parameters with Head TSL at 3900 MHz

Impedance, transformed to feed point	46.3 Ω - 6.6 jΩ	
Return Loss	- 22.1 dB	

Antenna Parameters with Head TSL at 4000 MHz

Impedance, transformed to feed point	52.1 Ω - 2.7 jΩ	
Return Loss	- 29.5 dB	

Antenna Parameters with Head TSL at 4100 MHz

Impedance, transformed to feed point	59.8 Ω - 1.9 jΩ	
Return Loss	- 20.8 dB	

General Antenna Parameters and Design

Electrical Delay (one direction)	1.107 ns	

After long term use with 100W radiated power, only a slight warming of the dipole near the feedpoint can be measured.

The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard.

No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feedpoint may be damaged.

Additional EUT Data

	Non-the Control of the Control of th
Manufactured by	SPEAG

Certificate No: D3900V2-1024_Jul22

Page 5 of 8





DASY5 Validation Report for Head TSL

Date: 01.07.2022

Test Laboratory: SPEAG, Zurich, Switzerland

DUT: Dipole 3900 MHz; Type: D3900V2; Serial: D3900V2 - SN:1024

Communication System: UID 0 - CW; Frequency: 3900 MHz, Frequency: 4000 MHz, Frequency: 4100 MHz

Medium parameters used: f = 3900 MHz; σ = 3.24 S/m; ϵ_r = 36.7; ρ = 1000 kg/m³, Medium parameters used: f = 4000 MHz; σ = 3.33 S/m; ϵ_r = 36.6; ρ = 1000 kg/m³, Medium parameters used: f = 4100 MHz; σ = 3.41 S/m; ϵ_r = 36.5; ρ = 1000 kg/m³ Phantom section: Flat Section Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2011)

DASY52 Configuration:

- Probe: EX3DV4 SN3503; ConvF(7.39, 7.39, 7.39) @ 3900 MHz, ConvF(7.39, 7.39, 7.39) @ 4000 MHz, ConvF(7.26, 7.26, 7.26) @ 4100 MHz; Calibrated: 08.03.2022
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn601; Calibrated: 02.05.2022
- Phantom: Flat Phantom 5.0 (front); Type: QD 000 P50 AA; Serial: 1001
- DASY52 52.10.4(1535); SEMCAD X 14.6.14(7501)

Dipole Calibration for Head Tissue/Pin=100 mW, d=10mm, f=3900MHz/Zoom Scan,

dist=1.4mm (8x8x8)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 71.06 V/m; Power Drift = 0.00 dB Peak SAR (extrapolated) = 20.0 W/kg SAR(1 g) = 6.96 W/kg; SAR(10 g) = 2.42 W/kg Smallest distance from peaks to all points 3 dB below = 8 mm Ratio of SAR at M2 to SAR at M1 = 74.2% Maximum value of SAR (measured) = 13.9 W/kg

```
Dipole Calibration for Head Tissue/Pin=100 mW, d=10mm, f=4000MHz/Zoom Scan,
```

dist=1.4mm (8x8x8)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 71.32 V/m; Power Drift = 0.02 dB Peak SAR (extrapolated) = 20.0 W/kg SAR(1 g) = 6.82 W/kg; SAR(10 g) = 2.38 W/kg Smallest distance from peaks to all points 3 dB below = 8 mm Ratio of SAR at M2 to SAR at M1 = 73.6% Maximum value of SAR (measured) = 13.8 W/kg

Certificate No: D3900V2-1024_Jul22

Page 6 of 8





Dipole Calibration for Head Tissue/Pin=100 mW, d=10mm, f=4100MHz/Zoom Scan,

dist=1.4mm (8x8x8)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm
Reference Value = 69.19 V/m; Power Drift = 0.04 dB
Peak SAR (extrapolated) = 19.7 W/kg
SAR(1 g) = 6.82 W/kg; SAR(10 g) = 2.37 W/kg
Smallest distance from peaks to all points 3 dB below = 8 mm
Ratio of SAR at M2 to SAR at M1 = 74.2%
Maximum value of SAR (measured) = 13.7 W/kg



0 dB = 13.9 W/kg = 11.43 dBW/kg

Certificate No: D3900V2-1024_Jul22

Page 7 of 8





Scale Channel Sweep Marker System Window Help Calibration Trace File ⊻iew 46.299 Ω 3.900000 GHz 6.1975 pF -6.5852 Ω 4.000000 GHz 52.062 Ω -2.7128 Ω 14.667 pF 100000 GHz 59.839 Ω 20.087 pF -1.9338 Ω 3.900000 GHz 78.261 mU -115.43 ° Ch 1 Avg = 20 Ch1: Start 3.70000 GHz Stop 4.30000 GHz 10.00 900000 Hz -29.531 dB 4.000000 GHz 0.00 4.100000 GHz -20.793 dB -5.00 -10.00 -15.00 -20.00 -25.00 30.00 -35.00 40.00 Ch 1 Avg = 20 Ch 1: Start 3.70000 GHz Stop 4.30000 GHz Avg=20 Delay LCL C* 1-Port Status CH 1: 511

Impedance Measurement Plot for Head TSL

Certificate No: D3900V2-1024_Jul22

Page 8 of 8





5GHz Dipole Calibration Certificate

Zeughausstrasse 43, 8004 Zurich,	, Switzerland		Service suisse d'étalonnage Servizio svizzero di taratura Swiss Calibration Service
Accredited by the Swiss Accreditation The Swiss Accreditation Service	on Service (SAS) is one of the signatorie	es to the EA	Accreditation No.: SCS 0108
Client CTTL (Auden)	sognition of canoration	Certificate M	No: D5GHzV2-1262_Jan22
CALIBRATION C	ERTIFICAT	Ξ	
Object	D5GHzV2 - SN:1	262	
Calibration procedure(s)	QA CAL-22.v6		
	Calibration Proce	edure for SAR Validation Source	s between 3-10 GHz
Calibration date:	January 27, 2022	2	
This calibration certificate documer The measurements and the uncert All calibrations have been conduct	nts the traceability to nati ainties with confidence p ed in the closed laborator	onal standards, which realize the physical un robability are given on the following pages a ry facility: environment temperature $(22 \pm 3)^{4}$	nits of measurements (SI). Ind are part of the certificate. °C and humidity < 70%.
This calibration certificate documer The measurements and the uncert All calibrations have been conduct Calibration Equipment used (M&TE	the traceability to nati ainties with confidence p ed in the closed laborator E critical for calibration)	onal standards, which realize the physical ur robability are given on the following pages a ry facility: environment temperature (22 ± 3)*	nits of measurements (SI). Ind are part of the certificate. °C and humidity < 70%.
This calibration certificate documer The measurements and the uncert All calibrations have been conducto Calibration Equipment used (M&TE Primary Standards Bourse meter NBB	to the traceability to nati ainties with confidence p ad in the closed laborator critical for calibration)	Cal Date (Certificate No.)	nits of measurements (SI). Ind are part of the certificate. °C and humidity < 70%. Scheduled Calibration
This calibration certificate documer The measurements and the uncert All calibrations have been conducto Calibration Equipment used (M&TE Primary Standards Power meter NRP Power sensor NRP-791	ts the traceability to nati ainties with confidence p ed in the closed laborator critical for calibration) ID # SN: 104778 SN: 104244	Cal Date (Certificate No.) 09-Apr-21 (No. 217-03291/03292) 09-Apr-21 (No. 217-03291/03292)	hits of measurements (SI). Ind are part of the certificate. "C and humidity < 70%. Scheduled Calibration Apr-22 Apr-22
This calibration certificate documer The measurements and the uncert All calibrations have been conducto Calibration Equipment used (M&TE Primary Standards Power meter NRP Power sensor NRP-Z91 Power sensor NRP-Z91	ts the traceability to nati ainties with confidence p ed in the closed laborator critical for calibration) ID # SN: 104778 SN: 103244 SN: 103245	onal standards, which realize the physical ur robability are given on the following pages a ry facility: environment temperature (22 ± 3)* Cal Date (Certificate No.) 09-Apr-21 (No. 217-03291/03292) 09-Apr-21 (No. 217-03291) 09-Apr-21 (No. 217-03291)	hits of measurements (SI). Ind are part of the certificate. "C and humidity < 70%. Scheduled Calibration Apr-22 Apr-22 Apr-22
This calibration certificate documer The measurements and the uncert All calibrations have been conduct Calibration Equipment used (M&TE <u>Primary Standards</u> Power meter NRP Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator	ts the traceability to nati ainties with confidence p ad in the closed laborator critical for calibration) ID # SN: 104778 SN: 103244 SN: 103245 SN: BH9394 (20k)	onal standards, which realize the physical ur robability are given on the following pages a ry facility: environment temperature (22 ± 3)* Cal Date (Certificate No.) 09-Apr-21 (No. 217-03291/03292) 09-Apr-21 (No. 217-03291) 09-Apr-21 (No. 217-03292) 09-Apr-21 (No. 217-03343)	hits of measurements (SI). Ind are part of the certificate. C and humidity < 70%. Scheduled Calibration Apr-22 Apr-22 Apr-22 Apr-22 Apr-22
This calibration certificate documer The measurements and the uncert All calibrations have been conduct Calibration Equipment used (M&TE Primary Standards Power meter NRP Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Type-N mismatch combination	ts the traceability to nati ainties with confidence p ad in the closed laborator critical for calibration) ID # SN: 104778 SN: 103244 SN: 103245 SN: 81H9394 (20k) SN: 310982 / 06327	onal standards, which realize the physical ur robability are given on the following pages a ry facility: environment temperature (22 ± 3)* Cal Date (Certificate No.) 09-Apr-21 (No. 217-03291/03292) 09-Apr-21 (No. 217-03291) 09-Apr-21 (No. 217-03343) 09-Apr-21 (No. 217-03344)	Scheduled Calibration Apr-22 Apr-32
This calibration certificate documer The measurements and the uncert All calibrations have been conduct Calibration Equipment used (M&TE Primary Standards Power meter NRP Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4	ts the traceability to nati ainties with confidence p ad in the closed laborator critical for calibration) ID # SN: 104778 SN: 103244 SN: 103245 SN: BH9394 (20k) SN: 310982 / 06327 SN: 3503	onal standards, which realize the physical ur robability are given on the following pages a ry facility: environment temperature (22 ± 3)* <u>Cal Date (Certificate No.)</u> 09-Apr-21 (No. 217-03291/03292) 09-Apr-21 (No. 217-03292) 09-Apr-21 (No. 217-03343) 09-Apr-21 (No. 217-03344) 31-Dec-21 (No. EX3-3503_Dec21)	Scheduled Calibration Apr-22 Apr-22
This calibration certificate documer The measurements and the uncert All calibrations have been conduct Calibration Equipment used (M&TE Primary Standards Power meter NRP Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4	tis the traceability to nati ainties with confidence p ad in the closed laborator critical for calibration) ID # SN: 104778 SN: 103244 SN: 103244 SN: 103245 SN: 103245 SN: 8H9394 (20k) SN: 310982 / 06327 SN: 3503 SN: 601	Cal Date (Certificate No.) 09-Apr-21 (No. 217-03291/03292) 09-Apr-21 (No. 217-03291) 09-Apr-21 (No. 217-03291) 09-Apr-21 (No. 217-03291) 09-Apr-21 (No. 217-03291) 09-Apr-21 (No. 217-03343) 09-Apr-21 (No. 217-03344) 31-Dec-21 (No. EX3-3503_Dec21) 01-Nov-21 (No. DAE4-601_Nov21)	nits of measurements (SI). ind are part of the certificate. °C and humidity < 70%. <u>Scheduled Calibration</u> Apr-22 Apr-22 Apr-22 Apr-22 Apr-22 Dec-22 Nov-22
This calibration certificate documer The measurements and the uncert All calibrations have been conduct Calibration Equipment used (M&TE <u>Primary Standards</u> Power meter NRP Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards	ts the traceability to nati ainties with confidence p ed in the closed laborator critical for calibration) ID # SN: 104778 SN: 103244 SN: 103245 SN: 103245 SN: 8H9394 (20k) SN: 310982 / 06327 SN: 3503 SN: 601 ID #	onal standards, which realize the physical ur robability are given on the following pages a ry facility: environment temperature (22 ± 3)* <u>Cal Date (Certificate No.)</u> 09-Apr-21 (No. 217-03291/03292) 09-Apr-21 (No. 217-03291) 09-Apr-21 (No. 217-03291) 09-Apr-21 (No. 217-03343) 09-Apr-21 (No. 217-03344) 31-Dec-21 (No. EX3-3503_Dec21) 01-Nov-21 (No. DAE4-601_Nov21) Check Date (in house)	nits of measurements (SI). Ind are part of the certificate. C and humidity < 70%. Scheduled Calibration Apr-22 Apr-22 Apr-22 Apr-22 Apr-22 Dec-22 Nov-22 Scheduled Check
This calibration certificate documer The measurements and the uncert All calibrations have been conduct Calibration Equipment used (M&TE <u>Primary Standards</u> Power meter NRP Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 <u>Secondary Standards</u> Power meter E4419B	ts the traceability to nati ainties with confidence p ed in the closed laborator critical for calibration) ID # SN: 104778 SN: 103244 SN: 103245 SN: 103245 SN: 8H9394 (20k) SN: 310982 / 06327 SN: 3503 SN: 601 ID # SN: GB39512475	onal standards, which realize the physical ur robability are given on the following pages a ry facility: environment temperature (22 ± 3)* <u>Cal Date (Certificate No.)</u> 09-Apr-21 (No. 217-03291/03292) 09-Apr-21 (No. 217-03291) 09-Apr-21 (No. 217-03292) 09-Apr-21 (No. 217-03343) 09-Apr-21 (No. 217-03344) 31-Dec-21 (No. EX3-3503_Dec21) 01-Nov-21 (No. DAE4-601_Nov21) <u>Check Date (in house)</u> 30-Oct-14 (in house check Oct-20)	hits of measurements (SI). Ind are part of the certificate. C and humidity < 70%. Scheduled Calibration Apr-22 Apr-22 Apr-22 Apr-22 Apr-22 Dec-22 Nov-22 Scheduled Check In house check: Oct-22
This calibration certificate documer The measurements and the uncert All calibrations have been conduct Calibration Equipment used (M&TE <u>Primary Standards</u> Power meter NRP Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 <u>Secondary Standards</u> Power meter E4419B Power sensor HP 8481A	ts the traceability to nati ainties with confidence p ed in the closed laborator critical for calibration) ID # SN: 104778 SN: 103244 SN: 103245 SN: 103245 SN: 103245 SN: 310982 / 06327 SN: 3503 SN: 601 ID # SN: GB39512475 SN: US37292783	onal standards, which realize the physical ur robability are given on the following pages a ry facility: environment temperature (22 ± 3)* <u>Cal Date (Certificate No.)</u> 09-Apr-21 (No. 217-03291/03292) 09-Apr-21 (No. 217-03291) 09-Apr-21 (No. 217-03292) 09-Apr-21 (No. 217-03343) 09-Apr-21 (No. 217-03344) 31-Dec-21 (No. EX3-3503_Dec21) 01-Nov-21 (No. DAE4-601_Nov21) <u>Check Date (in house)</u> 30-Oct-14 (in house check Oct-20) 07-Oct-15 (in house check Oct-20)	hits of measurements (SI). Ind are part of the certificate. C and humidity < 70%. Scheduled Calibration Apr-22 Apr-22 Apr-22 Apr-22 Apr-22 Dec-22 Nov-22 Scheduled Check In house check: Oct-22 In house check: Oct-22
This calibration certificate documer The measurements and the uncert All calibrations have been conduct Calibration Equipment used (M&TE <u>Primary Standards</u> Power meter NRP Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 <u>Secondary Standards</u> Power meter E4419B Power sensor HP 8481A Power sensor HP 8481A	ts the traceability to nati ainties with confidence p ed in the closed laborator critical for calibration) ID # SN: 104778 SN: 103244 SN: 103245 SN: 103245 SN: 103245 SN: 103245 SN: 310982 / 06327 SN: 3503 SN: 601 ID # SN: GB39512475 SN: US37292783 SN: MY41093315	Cal Date (Certificate No.) 09-Apr-21 (No. 217-03291/03292) 09-Apr-21 (No. 217-03291/03292) 09-Apr-21 (No. 217-03291) 09-Apr-21 (No. 217-03291) 09-Apr-21 (No. 217-03292) 09-Apr-21 (No. 217-03292) 09-Apr-21 (No. 217-03343) 09-Apr-21 (No. 217-03344) 31-Dec-21 (No. EX3-3503_Dec21) 01-Nov-21 (No. DAE4-601_Nov21) Check Date (in house) 30-Oct-14 (in house check Oct-20) 07-Oct-15 (in house check Oct-20)	Scheduled Calibration PC and humidity < 70%.
This calibration certificate documer The measurements and the uncert All calibrations have been conducts Calibration Equipment used (M&TE <u>Primary Standards</u> Power meter NRP Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 <u>Secondary Standards</u> Power meter E4419B Power sensor HP 8481A Power sensor HP 8481A Power sensor HP 8481A RF generator R&S SMT-06 Network Apalyzer Acident E62504	tis the traceability to nati ainties with confidence p ad in the closed laborator critical for calibration) ID # SN: 104778 SN: 103244 SN: 103245 SN: 103245 SN: 310982 / 06327 SN: 3503 SN: 601 ID # SN: GB39512475 SN: US37292783 SN: MY41093315 SN: 100972 SN: 109420777	Cal Date (Certificate No.) 09-Apr-21 (No. 217-03291/03292) 09-Apr-21 (No. 217-03291/03292) 09-Apr-21 (No. 217-03291) 09-Apr-21 (No. 217-03291) 09-Apr-21 (No. 217-03292) 09-Apr-21 (No. 217-03292) 09-Apr-21 (No. 217-03343) 09-Apr-21 (No. 217-03344) 31-Dec-21 (No. DAE4-601_Nov21) Check Date (in house) 30-Oct-14 (in house check Oct-20) 07-Oct-15 (in house check Oct-20) 07-Oct-15 (in house check Oct-20) 07-Oct-14 (in house check Oct-20) 07-Oct-15 (in house check Oct-20) 07-Oct-14 (in house check Oct-20) 07-Oct-15 (in house check Oct-20)	Scheduled Calibration PC and humidity < 70%.
This calibration certificate documer The measurements and the uncert All calibrations have been conduct Calibration Equipment used (M&TE Primary Standards Power meter NRP Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards Power sensor HP 8481A Power sensor HP 8481A RF generator R&S SMT-06 Network Analyzer Agilent E8358A	Its the traceability to natiainties with confidence p ainties with confidence p ad in the closed laborator E critical for calibration) ID # SN: 104778 SN: 103244 SN: 103245 SN: 103245 SN: 103245 SN: 103245 SN: 310982 / 06327 SN: 310982 / 06327 SN: 3503 SN: 601 ID # SN: GB39512475 SN: US37292783 SN: 100972 SN: US41080477	Cal Date (Certificate No.) 09-Apr-21 (No. 217-03291/03292) 09-Apr-21 (No. 217-03291) 09-Apr-21 (No. 217-03291) 09-Apr-21 (No. 217-03292) 09-Apr-21 (No. 217-03292) 09-Apr-21 (No. 217-03292) 09-Apr-21 (No. 217-03292) 09-Apr-21 (No. 217-03343) 09-Apr-21 (No. 217-03344) 31-Dec-21 (No. DAE4-601_Nov21) Check Date (in house) 30-Oct-14 (in house check Oct-20) 07-Oct-15 (in house check Oct-20) 15-Jun-15 (in house check Oct-20) 31-Mar-14 (in house check Oct-20)	Scheduled Calibration °C and humidity < 70%.
This calibration certificate documer The measurements and the uncert All calibrations have been conduct Calibration Equipment used (M&TE <u>Primary Standards</u> Power meter NRP Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 <u>Secondary Standards</u> Power meter E4419B Power sensor HP 8481A Power sensor HP 8481A RF generator R&S SMT-06 Network Analyzer Agilent E8358A	Its the traceability to natiainties with confidence p ainties with confidence p ad in the closed laborator E critical for calibration) ID # SN: 104778 SN: 103244 SN: 103245 SN: 103245 SN: 103245 SN: 310982 / 06327 SN: 3503 SN: 601 ID # SN: GB39512475 SN: US37292783 SN: 100972 SN: US41080477 Name	Cal Date (Certificate No.) 09-Apr-21 (No. 217-03291/03292) 09-Apr-21 (No. 217-03291) 09-Apr-21 (No. 217-03292) 09-Apr-21 (No. 217-03343) 09-Apr-21 (No. 217-03344) 31-Dec-21 (No. EX3-3503_Dec21) 01-Nov-21 (No. DAE4-601_Nov21) Check Date (in house) 30-Oct-14 (in house check Oct-20) 07-Oct-15 (in house check Oct-20) 07-Oct-15 (in house check Oct-20) 15-Jun-15 (in house check Oct-20) 31-Mar-14 (in house check Oct-20)	nits of measurements (SI). ind are part of the certificate. "C and humidity < 70%. "C and humidity < 70%. Scheduled Calibration Apr-22 Apr-22 Apr-22 Apr-22 Apr-22 Dec-22 Nov-22 Scheduled Check In house check: Oct-22 In house check:
This calibration certificate documer The measurements and the uncert All calibrations have been conduct Calibration Equipment used (M&TE <u>Primary Standards</u> Power meter NRP Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 <u>Secondary Standards</u> Power meter E4419B Power sensor HP 8481A Power sensor HP 8481A RF generator R&S SMT-06 Network Analyzer Agilent E8358A	tis the traceability to nati ainties with confidence p ed in the closed laborator critical for calibration) ID # SN: 104778 SN: 103244 SN: 103244 SN: 103245 SN: 103245 SN: 103245 SN: 310982 / 06327 SN: 3503 SN: 601 ID # SN: GB39512475 SN: US37292783 SN: WY41093315 SN: 100972 SN: US41080477 Name Aldonia Georgiadou	onal standards, which realize the physical unrobability are given on the following pages a robability are given on the following pages a ry facility: environment temperature (22 ± 3)* Q9-Apr-21 (No. 217-03291/03292) 09-Apr-21 (No. 217-03291) 09-Apr-21 (No. 217-03291) 09-Apr-21 (No. 217-03292) 09-Apr-21 (No. 217-03343) 09-Apr-21 (No. 217-03344) 31-Dec-21 (No. 217-03344) 31-Dec-21 (No. EX3-3503_Dec21) 01-Nov-21 (No. DAE4-601_Nov21) Check Date (in house) 30-Oct-14 (in house check Oct-20) 07-Oct-15 (in house check Oct-20) 07-Oct-15 (in house check Oct-20) 15-Jun-15 (in house check Oct-20) 31-Mar-14 (in house check Oct-20) St-Mar-14 (in house check Oct-20)	Scheduled Calibration PC and humidity < 70%.
This calibration certificate documer The measurements and the uncert All calibrations have been conducts Calibration Equipment used (M&TE Primary Standards Power meter NRP Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards Power meter E44198 Power sensor HP 8481A RF generator R&S SMT-06 Network Analyzer Agilent E8358A Calibrated by:	ID # ID # SN: 104778 SN: 104778 SN: 103244 SN: 103244 SN: 103245 SN: 103244 SN: 103245 SN: 8H9394 (20k) SN: 310982 / 06327 SN: 3503 SN: 601 ID # SN: GB39512475 SN: US37292783 SN: WY41093315 SN: US37292783 SN: WY41093315 SN: US41080477 Name Aldonia Georgiadou Sven Kühn	onal standards, which realize the physical un robability are given on the following pages a ry facility: environment temperature (22 ± 3)* <u>Cal Date (Certificate No.)</u> 09-Apr-21 (No. 217-03291/03292) 09-Apr-21 (No. 217-03291) 09-Apr-21 (No. 217-03292) 09-Apr-21 (No. 217-03343) 09-Apr-21 (No. 217-03344) 31-Dec-21 (No. EX3-3503_Dec21) 01-Nov-21 (No. DAE4-601_Nov21) <u>Check Date (in house)</u> 30-Oct-14 (in house check Oct-20) 07-Oct-15 (in house check Oct-20) 07-Oct-15 (in house check Oct-20) 31-Mar-14 (in house check Oct-20) Function Laboratory Technician	nits of measurements (SI). Ind are part of the certificate. PC and humidity < 70%. PC and humidity





Calibration Laboratory of Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland



Schweizerischer Kalibrierdienst Service suisse d'étalonnage Servizio svizzero di taratura Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS)

The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

Glossary:

TSL	tissue simulating liquid
ConvF	sensitivity in TSL / NORM x,y,z
N/A	not applicable or not measured

Calibration is Performed According to the Following Standards:

- a) IEC/IEEE 62209-1528, "Measurement Procedure For The Assessment Of Specific Absorption Rate Of Human Exposure To Radio Frequency Fields From Hand-Held And Body-Worn Wireless Communication Devices - Part 1528: Human Models, Instrumentation And Procedures (Frequency Range of 4 MHz to 10 GHz)", October 2020.
- b) KDB 865664, "SAR Measurement Requirements for 100 MHz to 6 GHz"

Additional Documentation:

c) DASY System Handbook

Methods Applied and Interpretation of Parameters:

- · Measurement Conditions: Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL: The source is mounted in a touch configuration below the center marking of the flat phantom.
- Return Loss: This parameter is measured with the source positioned under the liquid filled . phantom (as described in the measurement condition clause). The Return Loss ensures low reflected power. No uncertainty required.
- SAR measured: SAR measured at the stated antenna input power.
- SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

Certificate No: D5GHzV2-1262 Jan22

Page 2 of 8

Page 460 of 475

С S

Accreditation No.: SCS 0108

S





Measurement Conditions

DASY system configuration, as far as not given on page 1.

DASY Version	DASY52	V52.10.4
Extrapolation	Advanced Extrapolation	
Phantom	Modular Flat Phantom V5.0	
Distance Dipole Center - TSL	10 mm	with Spacer
Zoom Scan Resolution	dx, dy = 4.0 mm, dz = 1.4 mm	Graded Ratio = 1.4 (Z direction)
	5250 MHz ± 1 MHz	
Frequency	5600 MHz ± 1 MHz 5750 MHz ± 1 MHz	

Head TSL parameters at 5250 MHz

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	35.9	4.71 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	34.9 ± 6 %	4.52 mho/m ± 6 %
Head TSL temperature change during test	< 0.5 °C		

SAR result with Head TSL at 5250 MHz

SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition	
SAR measured	100 mW input power	8.15 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	80.9 W/kg ± 19.9 % (k=2)
SAR averaged over 10 cm ³ (10 g) of Head TSL	condition	
SAR averaged over 10 cm ³ (10 g) of Head TSL SAR measured	condition 100 mW input power	2.33 W/kg

Head TSL parameters at 5600 MHz

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	35.5	5.07 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	34.4 ± 6 %	4.87 mho/m ± 6 %
Head TSL temperature change during test	< 0.5 °C		

SAR result with Head TSL at 5600 MHz

SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition	
SAR measured	100 mW input power	8.51 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	84.4 W/kg ± 19.9 % (k=2)

SAR averaged over 10 cm ³ (10 g) of Head TSL	condition	
SAR measured	100 mW input power	2.41 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	23.9 W/kg ± 19.5 % (k=2)

Certificate No: D5GHzV2-1262_Jan22

Page 3 of 8





Head TSL parameters at 5750 MHz The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	35.4	5.22 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	34.2 ± 6 %	5.02 mho/m ± 6 %
Head TSL temperature change during test	< 0.5 °C		

SAR result with Head TSL at 5750 MHz

SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition	
SAR measured	100 mW input power	8.19 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	81.2 W/kg ± 19.9 % (k=2)
SAR averaged over 10 cm ³ (10 g) of Head TSL	condition	
SAR averaged over 10 cm ³ (10 g) of Head TSL SAR measured	condition 100 mW input power	2.31 W/kg

Certificate No: D5GHzV2-1262_Jan22

Page 4 of 8





Appendix (Additional assessments outside the scope of SCS 0108)

Antenna Parameters with Head TSL at 5250 MHz

Impedance, transformed to feed point	49.9 Ω - 4.0 jΩ			
Return Loss	- 27.8 dB			

Antenna Parameters with Head TSL at 5600 MHz

Impedance, transformed to feed point	51.5 Ω + 1.0 jΩ	
Return Loss	- 34.9 dB	

Antenna Parameters with Head TSL at 5750 MHz

Impedance, transformed to feed point	53.3 Ω + 1.4 jΩ				
Return Loss	- 29.2 dB				

General Antenna Parameters and Design

Electrical Delay (one direction)	1.193 ns

After long term use with 100W radiated power, only a slight warming of the dipole near the feedpoint can be measured.

The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard.

No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feedpoint may be damaged.

Additional EUT Data

Manufactured by	SPEAG

Certificate No: D5GHzV2-1262_Jan22

Page 5 of 8





DASY5 Validation Report for Head TSL

Date: 27.01.2022

Test Laboratory: SPEAG, Zurich, Switzerland

DUT: Dipole D5GHzV2; Type: D5GHzV2; Serial: D5GHzV2 - SN:1262

Communication System: UID 0 - CW; Frequency: 5250 MHz, Frequency: 5600 MHz, Frequency: 5750 MHz Medium parameters used: f = 5250 MHz; $\sigma = 4.52$ S/m; $\varepsilon_r = 34.9$; $\rho = 1000$ kg/m³, Medium parameters used: f = 5600 MHz; $\sigma = 4.87$ S/m; $\varepsilon_r = 34.4$; $\rho = 1000$ kg/m³, Medium parameters used: f = 5750 MHz; $\sigma = 5.02$ S/m; $\varepsilon_r = 34.2$; $\rho = 1000$ kg/m³ Phantom section: Flat Section Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2011)

DASY52 Configuration:

- Probe: EX3DV4 SN3503; ConvF(5.5, 5.5, 5.5) @ 5250 MHz, ConvF(5.1, 5.1, 5.1) @ 5600 MHz, ConvF(5.08, 5.08, 5.08) @ 5750 MHz; Calibrated: 31.12.2021
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn601; Calibrated: 01.11.2021
- Phantom: Flat Phantom 5.0 (front); Type: QD 000 P50 AA; Serial: 1001
- DASY52 52.10.4(1535); SEMCAD X 14.6.14(7501)

Dipole Calibration for Head Tissue/Pin=100mW, dist=10mm, f=5250 MHz/Zoom Scan,

dist=1.4mm (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 79.04 V/m; Power Drift = 0.04 dB Peak SAR (extrapolated) = 28.0 W/kg SAR(1 g) = 8.15 W/kg; SAR(10 g) = 2.33 W/kg Smallest distance from peaks to all points 3 dB below = 7.2 mm Ratio of SAR at M2 to SAR at M1 = 69.9% Maximum value of SAR (measured) = 18.9 W/kg

Dipole Calibration for Head Tissue/Pin=100mW, dist=10mm, f=5600 MHz/Zoom Scan,

dist=1.4mm (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 78.74 V/m; Power Drift = 0.06 dB Peak SAR (extrapolated) = 31.4 W/kg SAR(1 g) = 8.51 W/kg; SAR(10 g) = 2.41 W/kg Smallest distance from peaks to all points 3 dB below = 7.2 mm Ratio of SAR at M2 to SAR at M1 = 67.6% Maximum value of SAR (measured) = 20.3 W/kg

Certificate No: D5GHzV2-1262_Jan22

Page 6 of 8





Dipole Calibration for Head Tissue/Pin=100mW, dist=10mm, f=5750 MHz/Zoom Scan,

dist=1.4mm (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm
Reference Value = 75.87 V/m; Power Drift = 0.04 dB
Peak SAR (extrapolated) = 31.9 W/kg
SAR(1 g) = 8.19 W/kg; SAR(10 g) = 2.31 W/kg
Smallest distance from peaks to all points 3 dB below = 7.4 mm
Ratio of SAR at M2 to SAR at M1 = 65.8%
Maximum value of SAR (measured) = 19.9 W/kg



0 dB = 20.3 W/kg = 13.07 dBW/kg

Certificate No: D5GHzV2-1262_Jan22

Page 7 of 8





Impedance Measurement Plot for Head TSL



Certificate No: D5GHzV2-1262_Jan22

Page 8 of 8

1987

Page 466 of 475





W2-NA	ANTO	ANT1	ANT4	ANT5	ANT6
Front Side	18	16	18	24	24
Back Side	21	20	24	28	28
Left Side	/	15	18	/	/
Top Side	/	/	23	28	28
Bottom Side	22	16	/	/	/

ANNEX I SAR Sensor Triggering Data Summary

Per FCC KDB Publication 616217 D04v01r02, this device was tested by the manufacturer to determine the proximity sensor triggering distances for the rear and bottom edge of the device. The measured output power within \pm 5mm of the triggering points (or until touching the phantom) is included for rear and each applicable edge.

To ensure all production units are compliant it is necessary to test SAR at a distance 1mm less than the smallest distance from the device and SAR phantom (determined from these triggering tests according to the KDB 616217 D04v01r02) with the device at maximum output power without power reduction. These SAR tests are included in addition to the SAR tests for the device touching the SAR phantom, with reduced power.

We tested the power and got the different proximity sensor triggering distances for front/rear/bottom edge for ANT0. The manufacturer has declared 18mm/21mm/22mm is the most conservative triggering distance for ANT0 with front/rear/bottom edge. So base on the most conservative triggering distance of 18mm/21mm/22mm, additional SAR measurements were required at 17mm/20mm/21mm from the highest SAR position between front/rear/bottom edge of ANT0.

We tested the power and got the different proximity sensor triggering distances for front/rear /bottom/left edge for ANT1. The manufacturer has declared 16mm/20mm/16mm/15mm is the most conservative triggering distance for ANT1 with front/rear/bottom/left edge. So base on the most conservative triggering distance of 16mm/20mm/16mm/15mm, additional SAR measurements were required at 15mm/19mm/15mm/14mm from the highest SAR position between front/rear/bottom/left edge of ANT1.

We tested the power and got the different proximity sensor triggering distances for front/rear /top/left edge for ANT4. The manufacturer has declared 18mm/24mm/23mm/18mm is the most conservative triggering distance for ANT4 with front/rear/top/left edge. So base on the most conservative triggering distance of 18mm/24mm/23mm/18mm, additional SAR measurements were required at 17mm/23mm/22mm/17mm from the highest SAR position between front/rear/top/left edge of ANT4.





We tested the power and got the different proximity sensor triggering distances for front/rear /top edge for ANT5. The manufacturer has declared 24mm/28mm/28mm is the most conservative triggering distance for ANT5 with front/rear/top edge. So base on the most conservative triggering distance of 24mm/28mm/28mm, additional SAR measurements were required at 23mm/27mm/27mm from the highest SAR position between front/rear/top edge of ANT5.

We tested the power and got the different proximity sensor triggering distances for front/rear /top edge for ANT6. The manufacturer has declared 24mm/28mm/28mm is the most conservative triggering distance for ANT6 with front/rear/top edge. So base on the most conservative triggering distance of 24mm/28mm/28mm, additional SAR measurements were required at 23mm/27mm/27mm from the highest SAR position between front/rear/top edge of ANT6.





ANT0:

Front Edge

Moving device toward the phantom:

sensor near or far(KDB 616217 6.2.6)											
Distance [mm]	23	22	21	20	19	18	17	16	15	14	13
Main antenna	Far	Far	Far	Far	Far	Near	Near	Near	Near	Near	Near

Moving device away from the phantom:

sensor near or far(KDB 616217 6.2.6)											
Distance [mm]	13	14	15	16	17	18	19	20	21	22	23
Main antenna	Near	Near	Near	Near	Near	Near	Far	Far	Far	Far	Far

Rear Edge

Moving device toward the phantom:

sensor near or far(KDB 616217 6.2.6)											
Distance [mm]	26	25	24	23	22	21	20	19	18	17	16
Main antenna	Far	Far	Far	Far	Far	Near	Near	Near	Near	Near	Near

Moving device away from the phantom:

	sensor near or far(KDB 616217 6.2.6)													
Distance [mm]	Distance [mm] 16 17 18 19 20 21 22 23 24 25 26													
Main antenna	Near	Near	Near	Near	Near	Near	Far	Far	Far	Far	Far			

Bottom Edge

Moving device toward the phantom:

	sensor near or far(KDB 616217 6.2.6)													
Distance [mm]	27	26	25	24	23	22	21	20	19	18	17			
Main antenna	Far	Far	Far	Far	Far	Near	Near	Near	Near	Near	Near			

Moving device away from the phantom:

	sensor near or far(KDB 616217 6.2.6)													
Distance [mm] 17 18 19 20 21 22 23 24 25 26 27														
Main antenna	Near	Near	Near	Near	Near	Near	Far	Far	Far	Far	Far			





ANT1:

Front Edge

Moving device toward the phantom:

	sensor near or far(KDB 616217 6.2.6)													
Distance [mm]	Distance [mm] 21 20 19 18 17 16 15 14 13 12 11													
Main antenna	Far	Far	Far	Far	Far	Near	Near	Near	Near	Near	Near			

Moving device away from the phantom:

	sensor near or far(KDB 616217 6.2.6)													
Distance [mm] 11 12 13 14 15 16 17 18 19 20 21														
Main antenna	Near	Near	Near	Near	Near	Near	Far	Far	Far	Far	Far			

Rear Edge

Moving device toward the phantom:

	sensor near or far(KDB 616217 6.2.6)													
Distance [mm]	25	24	23	22	21	20	19	18	17	16	15			
Main antenna	Far	Far	Far	Far	Far	Near	Near	Near	Near	Near	Near			

Moving device away from the phantom:

	sensor near or far(KDB 616217 6.2.6)													
sensor near or far(KDB 616217 6.2.6) Distance [mm] 15 16 17 18 19 20 21 22 23 24 25 Main antenna Near Near Near Near Far Far														
Main antenna	Near	Near	Near	Near	Near	Near	Far	Far	Far	Far	Far			

Left Edge

Moving device toward the phantom:

	sensor near or far(KDB 616217 6.2.6)													
Distance [mm]	20	19	18	17	16	15	14	13	12	11	10			
Main antenna	Far	Far	Far	Far	Far	Near	Near	Near	Near	Near	Near			
NA 1 1														

Moving device away from the phantom:

	sensor near or far(KDB 616217 6.2.6)													
Distance [mm]	10	11	12	13	14	15	16	17	18	19	20			
Main antenna	Near	Near	Near	Near	Near	Near	Far	Far	Far	Far	Far			

Bottom Edge

Moving device toward the phantom:

	sensor near or far(KDB 616217 6.2.6)													
Distance [mm]	21	20	19	18	17	16	15	14	13	12	11			
Main antenna	Far	Far	Far	Far	Far	Near	Near	Near	Near	Near	Near			
Maxim a star														

Moving device away from the phantom:

	sensor near or far(KDB 616217 6.2.6)													
Distance [mm]	Distance [mm] 11 12 13 14 15 16 17 18 19 20 21													
Main antenna	Near	Near	Near	Near	Near	Near	Far	Far	Far	Far	Far			





ANT4:

Front Edge

Moving device toward the phantom:

	sensor near or far(KDB 616217 6.2.6)													
Distance [mm]	Distance [mm] 23 22 21 20 19 18 17 16 15 14 13													
Main antenna	Far	Far	Far	Far	Far	Near	Near	Near	Near	Near	Near			

Moving device away from the phantom:

			senso	r near or	far(KDB	616217 6.2	2.6)				
Distance [mm]	13	14	15	16	17	18	19	20	21	22	23
Main antenna	Near	Near	Near	Near	Near	Near	Far	Far	Far	Far	Far

Rear Edge

Moving device toward the phantom:

sensor near or far(KDB 616217 6.2.6)													
Distance [mm]	29	28	27	26	25	24	23	22	21	20	19		
Main antenna	Far	Far	Far	Far	Far	Near	Near	Near	Near	Near	Near		

Moving device away from the phantom:

			senso	r near or	far(KDB 6	16217 6.2	2.6)				
Distance [mm]	19	20	21	22	23	24	25	26	27	28	29
Main antenna	Near	Near	Near	Near	Near	Near	Far	Far	Far	Far	Far

Top Edge

Moving device toward the phantom:

	sensor near or far(KDB 616217 6.2.6)													
Distance [mm]	28	27	26	25	24	23	22	21	20	19	18			
Main antenna	Far	Far	Far	Far	Far	Near	Near	Near	Near	Near	Near			

Moving device away from the phantom:

	sensor near or far(KDB 616217 6.2.6)											
Distance [mm]	18	19	20	21	22	23	24	25	26	27	28	
Main antenna	Near	Near	Near	Near	Near	Near	Far	Far	Far	Far	Far	

Left Edge

Moving device toward the phantom:

	sensor near or far(KDB 616217 6.2.6)													
Distance [mm]	23	22	21	20	19	18	17	16	15	14	13			
Main antenna	Far	Far	Far	Far	Far	Near	Near	Near	Near	Near	Near			
Maxima a alar d														

Moving device away from the phantom:

			senso	r near or	far(KDB	616217 6.2	2.6)				
Distance [mm]	13	14	15	16	17	18	19	20	21	22	23
Main antenna	Near	Near	Near	Near	Near	Near	Far	Far	Far	Far	Far





ANT4/ANT5:

Front Edge

Moving device toward the phantom:

	sensor near or far(KDB 616217 6.2.6)												
Distance [mm]	29	28	27	26	25	24	23	22	21	20	19		
Main antenna	Far	Far	Far	Far	Far	Near	Near	Near	Near	Near	Near		

Moving device away from the phantom:

	sensor near or far(KDB 616217 6.2.6)												
Distance [mm]	nce [mm] 19 20 21 22 23 24 25 26 27 28 29												
Main antenna	Near	Near	Near	Near	Near	Near	Far	Far	Far	Far	Far		

Rear Edge

Moving device toward the phantom:

sensor near or far(KDB 616217 6.2.6)													
Distance [mm]	33	32	31	30	29	28	27	26	25	24	23		
Main antenna	Far	Far	Far	Far	Far	Near	Near	Near	Near	Near	Near		
Maxim a day													

Moving device away from the phantom:

	sensor near or far(KDB 616217 6.2.6)												
Distance [mm]	23	24	25	26	27	28	29	30	31	32	33		
Main antenna	Near	Near	Near	Near	Near	Near	Far	Far	Far	Far	Far		

Top Edge

Moving device toward the phantom:

			senso	or near or	far(KDB 6	516217 6.2	2.6)				
Distance [mm]	33	32	31	30	29	28	27	26	25	24	23
Main antenna	Far	Far	Far	Far	Far	Near	Near	Near	Near	Near	Near
		6 (1)									

Moving device away from the phantom:

sensor near or far(KDB 616217 6.2.6)											
Distance [mm]	23	24	25	26	27	28	29	30	31	32	33
Main antenna	Near	Near	Near	Near	Near	Near	Far	Far	Far	Far	Far





Per FCC KDB Publication 616217 D04v01r02, the influence of table tilt angles to proximity sensor triggering is determined by positioning each edge that contains a transmitting antenna, perpendicular to the flat phantom, at the smallest sensor triggering test distanceby rotating the device around the edge next to the phantom in $\leq 10^{\circ}$ increments until the tablet is ±45° or more from the vertical position at 0°.



The bottom/top edge evaluation

Page 473 of 475







The left edge evaluation

Based on the above evaluation, we come to the conclusion that the sensor triggering is not released and normal maximum output power is not restored within the $\pm 45^{\circ}$ range at the smallest sensor triggering test distance declared by manufacturer.



Per FCC KDB Publication 616217 D04v01r02, conditions where the antenna is next to the user but the sensor is laterally further away to ensure sensor coverage is sufficient for reducing the power to maintain compliance. The device has been verified that the sensor coverage includes the device itself. Depending on the overlap between antenna and sensor, it can be proved that different media have no effect on sensor firing and a single medium type can be used. ©Copyright. All rights reserved by CTTL. Page 474 of 475





ANNEX J Accreditation Certificate

