

Dipole Intermediate Check

Justification for Extended SAR Dipole Calibrations

Usage of SAR dipoles calibrated less than 3 years ago but more than 1 year ago were confirmed in maintaining return loss (< -20 dB, within 20% of prior calibration) and impedance (within 5 ohm from prior calibration) requirements per extended calibrations in KDB 865664 D01:

Test Date: 2024/01/30

835 MHz

| Dipole 835 V2 SN: 445 | | | | |
|------------------------------|------------------------|-----------|-----------------------------|-----------|
| Head Liquid | | | | |
| Test By: Jason Lu | | | Checked By: Bard Liu | |
| Date of Measurement | Return Loss(dB) | Δ% | Impedance (Ω) | ΔΩ |
| 2023/02/10 | -21.8 | N/A | 51.9 | N/A |
| 2024/01/30 | -20.4 | 6.422 | 51.3 | 0.6 |
| | | | | |