10607- AAB	IEEE 802.11ac WiFi (20MHz, MCS0, 90pc duty cycle)	X	4.38	66.13	16.09	0.46	130.0	± 9.6 %
		Y	4.36	65.74	15.89		130.0	
		Z	4.38	66.10	16.07		130.0	
10608- AAB	IEEE 802.11ac WiFi (20MHz, MCS1, 90pc duty cycle)	Х	4.51	66.43	16.22	0.46	130.0	± 9.6 %
		Y	4.49	66.06	16.03		130.0	
		Z	4.50	66.39	16.20		130.0	
10609- AAB	IEEE 802.11ac WiFi (20MHz, MCS2, 90pc duty cycle)	X	4.41	66.26	16.04	0.46	130.0	± 9.6 %
		Y	4.39	65.88	15.84		130.0	
		Z	4.40	66.23	16.02		130.0	
10610- AAB	IEEE 802.11ac WiFi (20MHz, MCS3, 90pc duty cycle)	X	4.46	66.44	16.22	0.46	130.0	± 9.6 %
		Y	4.44	66.06	16.02		130.0	
		Z	4.45	66.40	16.20		130.0	
10611- AAB	IEEE 802.11ac WiFi (20MHz, MCS4, 90pc duty cycle)	X	4.37	66.22	16.05	0.46	130.0	± 9.6 %
		Y	4.35	65.85	15.85		130.0	
		Z	4.37	66.19	16.03		130.0	
10612- AAB	IEEE 802.11ac WiFi (20MHz, MCS5, 90pc duty cycle)	X	4.35	66.33	16.08	0.46	130.0	± 9.6 %
	A Committee of the comm	Y	4.34	65.96	15.88		130.0	
	Lackson and the second second	Z	4.35	66.29	16.06		130.0	
10613- AAB	IEEE 802.11ac WiFi (20MHz, MCS6, 90pc duty cycle)	X	4.35	66.14	15.92	0.46	130.0	± 9.6 %
		Y	4.33	65.78	15.73		130.0	
		Z	4.34	66.10	15.90		130.0	
10614- IEEE 802.11ac WiFi (20MHz, MC AAB 90pc duty cycle)	IEEE 802.11ac WiFi (20MHz, MCS7, 90pc duty cycle)	X	4.33	66.39	16.19	0.46	130.0	± 9.6 %
		Υ	4.31	66.02	15.99		130.0	
		Z	4.32	66.35	16.17		130.0	
10615- AAB	IEEE 802.11ac WiFi (20MHz, MCS8, 90pc duty cycle)	X	4.36	66.05	15.81	0.46	130.0	± 9.6 %
		Y	4.34	65.67	15.61		130.0	
		Z	4.36	66.02	15.80		130.0	
10616- AAB	IEEE 802.11ac WiFi (40MHz, MCS0, 90pc duty cycle)	X	5.02	66.37	16.26	0.46	130.0	± 9.6 %
		Y	5.01	66.11	16.12		130.0	
		Z	5.02	66.34	16.25		130.0	
10617- AAB	IEEE 802.11ac WiFi (40MHz, MCS1, 90pc duty cycle)	X	5.05	66.46	16.29	0.46	130.0	± 9.6 %
		Y	5.06	66.24	16.16		130.0	
	The state of the s	Z	5.05	66.44	16.27		130.0	
10618- AAB	IEEE 802.11ac WiFi (40MHz, MCS2, 90pc duty cycle)	X	4.96	66.54	16.34	0.46	130.0	± 9.6 %
		Y	4.96	66.29	16.20		130.0	1
		Z	4.96	66.50	16.32		130.0	
10619- AAB	IEEE 802.11ac WiFi (40MHz, MCS3, 90pc duty cycle)	X	5.00	66.42	16.21	0.46	130.0	± 9.6 %
		Y	4.99	66.14	16.06		130.0	
72.27		Z	5.00	66.39	16.20		130.0	
10620- AAB	IEEE 802.11ac WiFi (40MHz, MCS4, 90pc duty cycle)	×	5.05	66.36	16.23	0.46	130.0	± 9.6 %
		Y	5.06	66.13	16.10		130.0	
10001		Z	5.05	66.34	16.22		130.0	
10621- AAB	IEEE 802.11ac WiFi (40MHz, MCS5, 90pc duty cycle)	X	5.07	66.49	16.42	0.46	130.0	± 9.6 %
		Y	5.06	66.23	16.28		130.0	
10000		Z	5.07	66.46	16.40		130.0	
10622- AAB	IEEE 802.11ac WiFi (40MHz, MCS6, 90pc duty cycle)	X	5.05	66.57	16.46	0.46	130.0	± 9.6 %
		Y	5.05	66.32	16.32		130.0	
		Z	5.05	66.54	16.44		130.0	

10623- AAB	IEEE 802.11ac WiFi (40MHz, MCS7, 90pc duty cycle)	X	4.95	66.15	16.10	0.46	130.0	± 9.6 %
		Y	4.94	65.85	15.94		130.0	
		Z	4.95	66.12	16.09		130.0	
10624- IEEE 802.11ac WiFi (40MHz, MCS8, 90pc duty cycle)	Х	5.14	66.40	16.29	0.46	130.0	± 9.6 %	
		Y	5.14	66.14	16.16		130.0	
		Z	5.14	66.37	16.28	L 1	130.0	V
10625- AAB	IEEE 802.11ac WiFi (40MHz, MCS9, 90pc duty cycle)	X	5.23	66.56	16.44	0.46	130.0	± 9.6 %
		Y	5.23	66.29	16.30		130.0	
40000	IEEE 000 44 MEET (001 H) MOOO	Z	5.23	66.54	16.43		130.0	
10626- AAB	IEEE 802.11ac WiFi (80MHz, MCS0, 90pc duty cycle)	X	5.36	66.38	16.21	0.46	130.0	± 9.6 %
		Y	5.35	66.13	16.08		130.0	N. Comment
10627-	JEEE 902 44 co WIE: (DOM II - MOO4	Z	5.36	66.36	16.20		130.0	
AAB	IEEE 802.11ac WiFi (80MHz, MCS1, 90pc duty cycle)	X	5.58	66.98	16.48	0.46	130.0	± 9.6 %
		Y	5.60	66.82	16.40		130.0	
10628-	IEEE 802 1100 M/IE: /90M/II- M000	Z	5.58	66.96	16.48	0.1-	130.0	
AAB	IEEE 802.11ac WiFi (80MHz, MCS2, 90pc duty cycle)	X	5.35	66.35	16.10	0.46	130.0	± 9.6 %
		Y	5.34	66.11	15.97		130.0	
10629-	IEEE 802.11ac WiFi (80MHz, MCS3,		5.35	66.32	16.09	0.10	130.0	
AAB	90pc duty cycle)	X	5.47	66.59	16.22	0.46	130.0	± 9.6 %
		Z	5.46	66.35	16.08 16.22		130.0	
	IEEE 802.11ac WiFi (80MHz, MCS4, 90pc duty cycle)	X	5.47 5.64	66.58 67.33	16.22	0.46	130.0 130.0	± 9.6 %
	, , , , , , ,	Y	5.71	67.34	16.58		130.0	
La Company		Z	5.64	67.31	16.59		130.0	
10631- AAB	IEEE 802.11ac WiFi (80MHz, MCS5, 90pc duty cycle)	X	5.64	67.43	16.83	0.46	130.0	± 9.6 %
		Y	5.66	67.30	16.76		130.0	
		Z	5.64	67.40	16.82		130.0	
10632- AAB	IEEE 802.11ac WiFi (80MHz, MCS6, 90pc duty cycle)	X	5.60	67.22	16.75	0.46	130.0	± 9.6 %
		Y	5.60	67.01	16.64		130.0	
		Z	5.60	67.21	16.75		130.0	
10633- AAB	IEEE 802.11ac WiFi (80MHz, MCS7, 90pc duty cycle)	X	5.37	66.42	16.17	0.46	130.0	± 9.6 %
		Y	5.37	66.21	16.05		130.0	
10001	1000	Z	5.37	66.39	16.16		130.0	
10634- AAB	IEEE 802.11ac WiFi (80MHz, MCS8, 90pc duty cycle)	Х	5.41	66.64	16.33	0.46	130.0	± 9.6 %
		Y	5.39	66.38	16.20		130.0	
10635- AAB	IEEE 802.11ac WiFi (80MHz, MCS9, 90pc duty cycle)	Z	5.40 5.25	66.61 65.86	16.32 15.66	0.46	130.0 130.0	± 9.6 %
7010	cope daty cycle)	Y	5.25	65.60	15.50		400.0	
		Z	5.25	65.62 65.84	15.53 15.66		130.0	
10636-	IEEE 802.11ac WiFi (160MHz, MCS0,	X	5.80	66.73	16.30	0.46	130.0	+000
AAC	90pc duty cycle)	Y	5.79			0.46	130.0	± 9.6 %
		Z	5.79	66.51	16.18		130.0	
10637- AAC	IEEE 802.11ac WiFi (160MHz, MCS1, 90pc duty cycle)	X	5.90	66.71 67.00	16.29 16.42	0.46	130.0 130.0	± 9.6 %
		Y	5.91	66.84	16.34		130.0	
200		Z	5.90	66.98	16.42		130.0	
10638- AAC	IEEE 802.11ac WiFi (160MHz, MCS2, 90pc duty cycle)	X	5.95	67.12	16.46	0.46	130.0	± 9.6 %
2		Y	5.93	66.88	16.34		130.0	
		Z	5.95		16.45			

10639- AAC	IEEE 802.11ac WiFi (160MHz, MCS3, 90pc duty cycle)	X	5.89	66.96	16.42	0.46	130.0	± 9.6 %
		Y	5.88	66.75	16.31		130.0	
	2.76	Z	5.89	66.93	16.41		130.0	
10640- AAC	IEEE 802.11ac WiFi (160MHz, MCS4, 90pc duty cycle)	Х	5.83	66.79	16.28	0.46	130.0	± 9.6 %
	Y	5.84	66.63	16.19		130.0		
		Z	5.83	66.77	16.27		130.0	
10641- AAC	IEEE 802.11ac WiFi (160MHz, MCS5, 90pc duty cycle)	Х	5.95	66.93	16.37	0.46	130.0	± 9.6 %
		Υ	5.96	66.76	16.28		130.0	
10010	1555 000 44 1115 4400 11	Z	5.95	66.91	16.37		130.0	
10642- AAC	IEEE 802.11ac WiFi (160MHz, MCS6, 90pc duty cycle)	X	5.96	67.11	16.63	0.46	130.0	± 9.6 %
		Y	5.97	66.93	16.54		130.0	
10643-	IEEE 902 1100 W/IE: (100MH - 11007	Z	5.96	67.08	16.62		130.0	
AAC	IEEE 802.11ac WiFi (160MHz, MCS7, 90pc duty cycle)	X	5.81	66.80	16.36	0.46	130.0	± 9.6 %
		Y	5.82	66.62	16.27		130.0	
10644-	IEEE 902 11co WiE: /460MI - MOCC	Z	5.81	66.77	16.35	6.1-	130.0	
AAC	IEEE 802.11ac WiFi (160MHz, MCS8, 90pc duty cycle)	X	5.86	66.95	16.46	0.46	130.0	± 9.6 %
		Y	5.87	66.79	16.37		130.0	
10645-	IEEE 802.11ac WiFi (160MHz, MCS9,	Z	5.86	66.93	16.45	0.10	130.0	
AAC	90pc duty cycle)	X	5.99	67.02	16.46	0.46	130.0	± 9.6 %
		Y	6.05	67.00	16.45		130.0	
10646-	LTE-TDD (SC-FDMA, 1 RB, 5 MHz,	X	5.99	67.00	16.45	0.00	130.0	
AAD	AAD QPSK, UL Subframe=2,7)		6.44	91.69	32.34	9.30	60.0	± 9.6 %
		Y	6.42	89.28	30.79		60.0	
10647- AAC	LTE-TDD (SC-FDMA, 1 RB, 20 MHz, QPSK, UL Subframe=2,7)	Z X	6.55 5.54	92.39 88.53	32.93 31.29	9.30	60.0 60.0	± 9.6 %
7010	Q1 ON, OL Subilanie-2,1)	Y	5.68	86.96	30.05		00.0	
		Z	5.63	89.14	31.84		60.0 60.0	-
10648- AAA	CDMA2000 (1x Advanced)	X	0.45	61.10	7.60	0.00	150.0	± 9.6 %
		Y	0.39	60.00	6.45		150.0	
		Z	0.43	60.76	7.29		150.0	
10652- AAB	LTE-TDD (OFDMA, 5 MHz, E-TM 3.1, Clipping 44%)	Х	3.12	66.25	15.82	2.23	80.0	± 9.6 %
		Y	3.02	65.42	15.43		80.0	
Marie Land		Z	3.15	66.39	15.92		80.0	
10653- AAB	LTE-TDD (OFDMA, 10 MHz, E-TM 3.1, Clipping 44%)	X	3.68	65.59	16.18	2.23	80.0	± 9.6 %
		Y	3.61	65.02	15.90		80.0	
		Z	3.70	65.67	16.25		80.0	
10654- AAB	LTE-TDD (OFDMA, 15 MHz, E-TM 3.1, Clipping 44%)	Х	3.71	65.19	16.23	2.23	80.0	± 9.6 %
		Y	3.64	64.69	15.97		80.0	
40055	LTE TOO (OFFILM	Z	3.73	65.26	16.29		80.0	
10655- AAB	LTE-TDD (OFDMA, 20 MHz, E-TM 3.1, Clipping 44%)	X	3.79	65.10	16.27	2.23	80.0	± 9.6 %
		Y	3.72	64.62	16.02		80.0	
10658- AAA	Pulse Waveform (200Hz, 10%)	Z	3.81 3.27	65.16 68.60	16.33 11.27	10.00	80.0 50.0	± 9.6 %
, , , ,		Υ	3.23	68.00	14 22		FOO	
					11.33		50.0	
10659- AAA	Pulse Waveform (200Hz, 20%)	X	6.24 2.86	75.59 70.12	14.25 10.90	6.99	50.0 60.0	± 9.6 %
AAA		Y	1.79	65.86	9.30		60.0	
		V 1	1 70	66 96	0.20		I GOO	

10660- AAA	Pulse Waveform (200Hz, 40%)	×	100.00	96.01	16.35	3.98	80.0	± 9.6 %
	BART I. TO THE RESERVE TO THE	Y	0.53	60.90	5.66		80.0	1
		Z	100.00	100.94	18.50	14.5	80.0	
10661- Pulse Waveform (200Hz, 60%) AAA	Х	100.00	93.46	14.46	2.22	100.0	± 9.6 %	
	0.000	Y	0.26	60.00	3.84		100.0	
		Z	100.00	99.57	16.97		100.0	
10662- AAA Pulse Waveform (200Hz, 80%)	Х	0.12	60.00	3.65	0.97	120.0	± 9.6 %	
	Y	3.12	205.83	11.04	n Litter in	120.0	-10 (State 1) H1	
		Z	100.00	84.98	10.36		120.0	

<sup>&</sup>lt;sup>E</sup> Uncertainty is determined using the max. deviation from linear response applying rectangular distribution and is expressed for the square of the field value.

REPORT NO: UL-SAR-RP12291733-116A V3.0 Issue Date: 09 November 2018

### 12.5. Calibration Certificate for Dipoles

This sub-section contains Cal Certificates for Dipoles, and is not included in the total number of pages for this report.

UL VS Ltd. Report. No.: 3.0

PRE0178324

### Calibration Laboratory of Schmid & Partner Engineering AG

Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland





Schweizerischer Kalibrierdiens

Service suisse d'étalonnage

Servizio svizzero di taratura

Accreditation No.: SCS 0108

S Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS)

The Swiss Accreditation Service is one of the signatories to the EA

Multilateral Agreement for the recognition of calibration certificates

Client

**UL RFI UK** 

Certificate No: D900V2-1d199\_Mar18

# **CALIBRATION CERTIFICATE**

Object

D900V2 - SN:1d199

Calibration procedure(s)

QA CAL-05.v10

Calibration procedure for dipole validation kits above 700 MHz

Calibration date:

March 07, 2018

This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature (22  $\pm$  3)°C and humidity < 70%.

Calibration Equipment used (M&TE critical for calibration)

Primary Standards	ID#	Cal Date (Certificate No.)	Scheduled Calibration
Power meter NRP	SN: 104778	04-Apr-17 (No. 217-02521/02522)	Apr-18
Power sensor NRP-Z91	SN: 103244	04-Apr-17 (No. 217-02521)	Apr-18
Power sensor NRP-Z91	SN: 103245	04-Apr-17 (No. 217-02522)	Apr-18
Reference 20 dB Attenuator	SN: 5058 (20k)	07-Apr-17 (No. 217-02528)	Apr-18
Type-N mismatch combination	SN: 5047.2 / 06327	07-Apr-17 (No. 217-02529)	Apr-18
Reference Probe EX3DV4	SN: 7349	30-Dec-17 (No. EX3-7349_Dec17)	Dec-18
DAE4	SN: 601	26-Oct-17 (No. DAE4-601_Oct17)	Oct-18
Secondary Standards	ID#	Check Date (in house)	Scheduled Check
Power meter EPM-442A	SN: GB37480704	07-Oct-15 (in house check Oct-16)	In house check: Oct-18
Power sensor HP 8481A	SN: US37292783	07-Oct-15 (in house check Oct-16)	In house check: Oct-18
Power sensor HP 8481A	SN: MY41092317	07-Oct-15 (in house check Oct-16)	In house check: Oct-18
RF generator R&S SMT-06	SN: 100972	15-Jun-15 (in house check Oct-16)	In house check: Oct-18
Network Analyzer HP 8753E	SN: US37390585	18-Oct-01 (in house check Oct-17)	In house check: Oct-18
	Name	Function	Signature
Calibrated by:	Jeton Kastrati	Laboratory Technician	Official Control
Approved by:	Katja Pokovic	Technical Manager	00 m

Issued: March 12, 2018

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

### **Calibration Laboratory of**

Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland





Schweizerischer Kalibrierdienst Service suisse d'étalonnage Servizio svizzero di taratura Swiss Calibration Service

Accreditation No.: SCS 0108

Accredited by the Swiss Accreditation Service (SAS)

The Swiss Accreditation Service is one of the signatories to the EA

Multilateral Agreement for the recognition of calibration certificates

Glossary:

TSL

tissue simulating liquid

ConvF N/A

sensitivity in TSL / NORM x,y,z not applicable or not measured

Calibration is Performed According to the Following Standards:

 a) IEEE Std 1528-2013, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", June 2013

b) IEC 62209-1, "Measurement procedure for the assessment of Specific Absorption Rate (SAR) from hand-held and body-mounted devices used next to the ear (frequency range of 300 MHz to 6 GHz)", July 2016

c) IEC 62209-2, "Procedure to determine the Specific Absorption Rate (SAR) for wireless communication devices used in close proximity to the human body (frequency range of 30 MHz to 6 GHz)", March 2010

d) KDB 865664, "SAR Measurement Requirements for 100 MHz to 6 GHz"

### **Additional Documentation:**

e) DASY4/5 System Handbook

### **Methods Applied and Interpretation of Parameters:**

- Measurement Conditions: Further details are available from the Validation Report at the end
  of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL: The dipole is mounted with the spacer to position its feed
  point exactly below the center marking of the flat phantom section, with the arms oriented
  parallel to the body axis.
- Feed Point Impedance and Return Loss: These parameters are measured with the dipole
  positioned under the liquid filled phantom. The impedance stated is transformed from the
  measurement at the SMA connector to the feed point. The Return Loss ensures low
  reflected power. No uncertainty required.
- Electrical Delay: One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
- SAR measured: SAR measured at the stated antenna input power.
- SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

Certificate No: D900V2-1d199 Mar18

## **Measurement Conditions**

DASY system configuration, as far as not given on page 1.

DASY Version	DASY5	V52.10.0
Extrapolation	Advanced Extrapolation	
Phantom	Modular Flat Phantom	
Distance Dipole Center - TSL	15 mm	with Spacer
Zoom Scan Resolution	dx, $dy$ , $dz = 5 mm$	313274
Frequency	900 MHz ± 1 MHz	

## **Head TSL parameters**

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	41.5	0.97 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	41.0 ± 6 %	0.94 mho/m ± 6 %
Head TSL temperature change during test	< 0.5 °C		

### SAR result with Head TSL

SAR averaged over 1 cm <sup>3</sup> (1 g) of Head TSL	Condition	
SAR measured	250 mW input power	2.63 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	10.7 W/kg ± 17.0 % (k=2)

SAR averaged over 10 cm <sup>3</sup> (10 g) of Head TSL	condition	
SAR measured	250 mW input power	1.69 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	6.87 W/kg ± 16.5 % (k=2)

# **Body TSL parameters**

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Body TSL parameters	22.0 °C	55.0	1.05 mho/m
Measured Body TSL parameters	(22.0 ± 0.2) °C	54.8 ± 6 %	1.02 mho/m ± 6 %
Body TSL temperature change during test	< 0.5 °C		

## SAR result with Body TSL

SAR averaged over 1 cm <sup>3</sup> (1 g) of Body TSL	Condition	
SAR measured	250 mW input power	2.68 W/kg
SAR for nominal Body TSL parameters	normalized to 1W	10.9 W/kg ± 17.0 % (k=2)

SAR averaged over 10 cm <sup>3</sup> (10 g) of Body TSL	condition	
SAR measured	250 mW input power	1.75 W/kg
SAR for nominal Body TSL parameters	normalized to 1W	7.12 W/kg ± 16.5 % (k=2)

# Appendix (Additional assessments outside the scope of SCS 0108)

## **Antenna Parameters with Head TSL**

Impedance, transformed to feed point	50.6 Ω - 1.2 jΩ
Return Loss	- 37.5 dB

## **Antenna Parameters with Body TSL**

Impedance, transformed to feed point	46.4 Ω - 3.0 jΩ		
Return Loss	- 26.3 dB		

## **General Antenna Parameters and Design**

Electrical Delay (one direction)	1.405 ns
	1.400 113

After long term use with 100W radiated power, only a slight warming of the dipole near the feedpoint can be measured.

The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard.

No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feedpoint may be damaged.

#### **Additional EUT Data**

Manufactured by	SPEAG
Manufactured on	June 10, 2015

### **DASY5 Validation Report for Head TSL**

Date: 07.03.2018

Test Laboratory: SPEAG, Zurich, Switzerland

DUT: Dipole 900 MHz; Type: D900V2; Serial: D900V2 - SN:1d199

Communication System: UID 0 - CW; Frequency: 900 MHz

Medium parameters used: f = 900 MHz;  $\sigma = 0.94 \text{ S/m}$ ;  $\varepsilon_r = 41$ ;  $\rho = 1000 \text{ kg/m}^3$ 

Phantom section: Flat Section

Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2011)

#### DASY52 Configuration:

- Probe: EX3DV4 SN7349; ConvF(9.71, 9.71, 9.71); Calibrated: 30.12.2017;
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn601; Calibrated: 26.10.2017
- Phantom: Flat Phantom 4.9 (front); Type: QD 00L P49 AA; Serial: 1001
- DASY52 52.10.0(1446); SEMCAD X 14.6.10(7417)

# Dipole Calibration for Head Tissue/Pin=250 mW, d=15mm/Zoom Scan (7x7x7)/Cube 0:


Measurement grid: dx=5mm, dy=5mm, dz=5mm

Reference Value = 64.85 V/m; Power Drift = -0.01 dB

Peak SAR (extrapolated) = 3.97 W/kg


SAR(1 g) = 2.63 W/kg; SAR(10 g) = 1.69 W/kg

Maximum value of SAR (measured) = 3.53 W/kg



0 dB = 3.53 W/kg = 5.48 dBW/kg

## Impedance Measurement Plot for Head TSL



### **DASY5 Validation Report for Body TSL**

Date: 07.03.2018

Test Laboratory: SPEAG, Zurich, Switzerland

## DUT: Dipole 900 MHz; Type: D900V2; Serial: D900V2 - SN:1d199

Communication System: UID 0 - CW; Frequency: 900 MHz

Medium parameters used: f = 900 MHz;  $\sigma = 1.02$  S/m;  $\epsilon_r = 54.8$ ;  $\rho = 1000$  kg/m<sup>3</sup>

Phantom section: Flat Section

Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2011)

#### DASY52 Configuration:

- Probe: EX3DV4 SN7349; ConvF(9.83, 9.83, 9.83); Calibrated: 30.12.2017;
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn601; Calibrated: 26.10.2017
- Phantom: Flat Phantom 4.9 (Back); Type: QD 00R P49 AA; Serial: 1005
- DASY52 52.10.0(1446); SEMCAD X 14.6.10(7417)

# Dipole Calibration for Body Tissue/Pin=250 mW, d=15mm/Zoom Scan (7x7x7)/Cube 0:


Measurement grid: dx=5mm, dy=5mm, dz=5mm

Reference Value = 62.90 V/m; Power Drift = -0.03 dB


Peak SAR (extrapolated) = 3.87 W/kg

SAR(1 g) = 2.68 W/kg; SAR(10 g) = 1.75 W/kg

Maximum value of SAR (measured) = 3.51 W/kg



## Impedance Measurement Plot for Body TSL



A2200

### Calibration Laboratory of Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland





Schweizerischer Kalibrierdienst

Service suisse d'étalonnage

Checked

Servizio svizzero di taratura
Swiss Calibration Service

Accreditation No.: SCS 0108

Issued: February 7, 2018

Accredited by the Swiss Accreditation Service (SAS)

The Swiss Accreditation Service is one of the signatories to the EA

Multilateral Agreement for the recognition of calibration certificates

Client

**UL RFI UK** 

Certificate No: D1900V2-537\_Feb18

## **CALIBRATION CERTIFICATE**

Object

D1900V2 - SN:537

Calibration procedure(s)

QA CAL-05.v9

Calibration procedure for dipole validation kits above 700 MHz

Calibration date:

February 07, 2018

This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature (22  $\pm$  3)°C and humidity < 70%.

Calibration Equipment used (M&TE critical for calibration)

Primary Standards	ID#	Cal Date (Certificate No.)	Scheduled Calibration
Power meter NRP	SN: 104778	04-Apr-17 (No. 217-02521/02522)	Apr-18
Power sensor NRP-Z91	SN: 103244	04-Apr-17 (No. 217-02521)	Apr-18
Power sensor NRP-Z91	SN: 103245	04-Apr-17 (No. 217-02522)	Apr-18
Reference 20 dB Attenuator	SN: 5058 (20k)	07-Apr-17 (No. 217-02528)	Apr-18
Type-N mismatch combination	SN: 5047.2 / 06327	07-Apr-17 (No. 217-02529)	Apr-18
Reference Probe EX3DV4	SN: 7349	30-Dec-17 (No. EX3-7349_Dec17)	Dec-18
DAE4	SN: 601	26-Oct-17 (No. DAE4-601_Oct17)	Oct-18
Secondary Standards	ID#	Check Date (in house)	Scheduled Check
Power meter EPM-442A	SN: GB37480704	07-Oct-15 (in house check Oct-16)	In house check: Oct-18
Power sensor HP 8481A	SN: US37292783	07-Oct-15 (in house check Oct-16)	In house check: Oct-18
Power sensor HP 8481A	SN: MY41092317	07-Oct-15 (in house check Oct-16)	In house check: Oct-18
RF generator R&S SMT-06	SN: 100972	15-Jun-15 (in house check Oct-16)	In house check: Oct-18
Network Analyzer HP 8753E	SN: US37390585	18-Oct-01 (in house check Oct-17)	In house check: Oct-18
	Name	Function	Signature 1
Calibrated by:	Claudio Leubler	Laboratory Technician	
Approved by:	Katja Pokovic	Technical Manager	RUS

Certificate No: D1900V2-537\_Feb18

Page 1 of 8

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

### Calibration Laboratory of

Schmid & Partner
Engineering AG
Zeughausstrasse 43, 8004 Zurich, Switzerland





S Schweizerischer Kalibrierdienst
C Service suisse d'étalonnage
Servizio svizzero di taratura
Swiss Calibration Service

Accreditation No.: SCS 0108

Accredited by the Swiss Accreditation Service (SAS)

The Swiss Accreditation Service is one of the signatories to the EA

Multilateral Agreement for the recognition of calibration certificates

Glossary:

TSL

N/A

tissue simulating liquid

ConvF

sensitivity in TSL / NORM x,y,z not applicable or not measured

Calibration is Performed According to the Following Standards:

a) IEEE Std 1528-2013, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", June 2013

b) IEC 62209-1, "Measurement procedure for the assessment of Specific Absorption Rate (SAR) from hand-held and body-mounted devices used next to the ear (frequency range of 300 MHz to 6 GHz)", July 2016

c) IEC 62209-2, "Procedure to determine the Specific Absorption Rate (SAR) for wireless communication devices used in close proximity to the human body (frequency range of 30 MHz to 6 GHz)", March 2010

d) KDB 865664, "SAR Measurement Requirements for 100 MHz to 6 GHz"

#### **Additional Documentation:**

e) DASY4/5 System Handbook

### **Methods Applied and Interpretation of Parameters:**

- Measurement Conditions: Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL: The dipole is mounted with the spacer to position its feed
  point exactly below the center marking of the flat phantom section, with the arms oriented
  parallel to the body axis.
- Feed Point Impedance and Return Loss: These parameters are measured with the dipole
  positioned under the liquid filled phantom. The impedance stated is transformed from the
  measurement at the SMA connector to the feed point. The Return Loss ensures low
  reflected power. No uncertainty required.
- Electrical Delay: One-way delay between the SMA connector and the antenna feed point.
   No uncertainty required.
- SAR measured: SAR measured at the stated antenna input power.
- SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

#### **Measurement Conditions**

DASY system configuration, as far as not given on page 1.

DASY Version	DASY5	V52.10.0
Extrapolation	Advanced Extrapolation	
Phantom	Modular Flat Phantom	
Distance Dipole Center - TSL	10 mm	with Spacer
Zoom Scan Resolution	dx, $dy$ , $dz = 5 mm$	
Frequency	1900 MHz ± 1 MHz	

### **Head TSL parameters**

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	40.0	1.40 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	40.7 ± 6 %	1.39 mho/m ± 6 %
Head TSL temperature change during test	< 0.5 °C		

### **SAR** result with Head TSL

SAR averaged over 1 cm <sup>3</sup> (1 g) of Head TSL	Condition	
SAR measured	250 mW input power	9.98 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	40.2 W/kg ± 17.0 % (k=2)

SAR averaged over 10 cm <sup>3</sup> (10 g) of Head TSL	condition	
SAR measured	250 mW input power	5.24 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	21.1 W/kg ± 16.5 % (k=2)

### **Body TSL parameters**

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Body TSL parameters	22.0 °C	53.3	1.52 mho/m
Measured Body TSL parameters	(22.0 ± 0.2) °C	55.2 ± 6 %	1.48 mho/m ± 6 %
Body TSL temperature change during test	< 0.5 °C		

### SAR result with Body TSL

SAR averaged over 1 cm <sup>3</sup> (1 g) of Body TSL	Condition	
SAR measured	250 mW input power	10.0 W/kg
SAR for nominal Body TSL parameters	normalized to 1W	41.0 W/kg ± 17.0 % (k=2)

SAR averaged over 10 cm <sup>3</sup> (10 g) of Body TSL	condition	
SAR measured	250 mW input power	5.29 W/kg
SAR for nominal Body TSL parameters	normalized to 1W	21.5 W/kg ± 16.5 % (k=2)

## Appendix (Additional assessments outside the scope of SCS 0108)

#### **Antenna Parameters with Head TSL**

Impedance, transformed to feed point	47.5 Ω - 6.9 jΩ	
Return Loss	- 22.5 dB	

### **Antenna Parameters with Body TSL**

Impedance, transformed to feed point	43.6 Ω - 6.4 jΩ	
Return Loss	- 20.3 dB	

### **General Antenna Parameters and Design**

Electrical Delay (one direction)	1.183 ns
Lieuthour Boldy (one direction)	1.103118

After long term use with 100W radiated power, only a slight warming of the dipole near the feedpoint can be measured.

The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard.

No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feedpoint may be damaged.

#### **Additional EUT Data**

Manufactured by	SPEAG
Manufactured on	March 22, 2001

### **DASY5 Validation Report for Head TSL**

Date: 07.02.2018

Test Laboratory: SPEAG, Zurich, Switzerland

### DUT: Dipole 1900 MHz; Type: D1900V2; Serial: D1900V2 - SN:537

Communication System: UID 0 - CW; Frequency: 1900 MHz

Medium parameters used; f = 1900 MHz;  $\sigma$  = 1.39 S/m;  $\varepsilon_r$  = 40.7;  $\rho$  = 1000 kg/m<sup>3</sup>

Phantom section: Flat Section

Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2011)

#### DASY52 Configuration:

- Probe: EX3DV4 SN7349; ConvF(8.18, 8.18, 8.18); Calibrated: 30.12.2017;
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn601; Calibrated: 26.10.2017
- Phantom: Flat Phantom 5.0 (front); Type: QD 000 P50 AA; Serial: 1001
- DASY52 52.10.0(1446); SEMCAD X 14.6.10(7417)

## Dipole Calibration for Head Tissue/Pin=250 mW, d=10mm/Zoom Scan (7x7x7)/Cube 0:


Measurement grid: dx=5mm, dy=5mm, dz=5mm

Reference Value = 109.7 V/m; Power Drift = -0.06 dB

Peak SAR (extrapolated) = 18.5 W/kg

SAR(1 g) = 9.98 W/kg; SAR(10 g) = 5.24 W/kg

Maximum value of SAR (measured) = 15.3 W/kg



0 dB = 15.3 W/kg = 11.85 dBW/kg