



TZ Medical
Sparked by your ideas

Subject: Validity of Test Results from Testing / Certification of the Integrated Module

To Whom It May Concern:

We, TZ Medical, Inc., 17750 SW Upper Boones Ferry RD STE 150, Portland, OR 97224, do hereby declare that the test results from conducted tests applied to the radio module:

Manufacturer: Thales Group

Model: PLS63-W

FCC ID: QIPPLS63-W

IC: 7830A-PLS63W

which is incorporated in our host equipment:

Manufacturer: TZ Medical, Inc.

Model: H41L

FCC ID: ZIMH41L

IC: 9647A-H41L

and as reported in module test report(s):

Report # 200722016RFM-2, Issue Date: 2021-02-03

- FCC 47 CFR Part 22
- FCC 47 CFR Part 24
- FCC 47 CFR Part 27
- FCC 47 CFR Part 90
- RSS-130 Issue 2
- RSS-132 Issue 3
- RSS-133 Issue 6
- RSS-139 Issue 3
- RSS-199 Issue 3
- RSS-Gen Issue 5

Report # 210811001RFM-1, Issue Date: 2021-08-26

- FCC 47 CFR Part 27 Subpart P

Report # 200722016EMC-1, Issue Date: 2021-02-03

- FCC 47 CFR Part 15 Subpart B

accurately represent the test results under the new conditions when the identified radio module is integrated in the identified host equipment.



By integration of the said radio module into the identified host equipment no changes were made to it or by external adaptation circuitry that impacts the validity of the results of the original module's test reports; hence conducted signal measurements in the above listed reports also apply to Product/Model: Trident 40L (Model: H41L)

Sincerely,

A handwritten signature in black ink that reads "John D. Moore". The signature is written in a cursive style with a large, stylized 'J' and 'M'.

John Moore
Engineering Manager

