



269 Mill Road
Chelmsford, Massachusetts 01824-4105
978-421-9655 (main)
978-421-0025 (fax)
www.zoll.com

Declaration for DFS client devices

March 5, 2012

We, ZOLL Medical Corporation, declare that the device does not support any non-US channels in all the operational mode(s) for the following product. Also, the client software and associated drivers will not initiate any transmission on DFS frequencies without initiation by a master. This includes restriction on transmissions for beacons. This device does not support ad-hoc peer-to-peer operation.

FCC ID: ZKP-RDC001

Signature of FCC Authorized submitter

Name: Carl Mamuszka

Title: Principal compliance Engineer

Tel: 978-421-9330

Fax: 978 421-0049

E-mail: CMamuszka@zoll.com