



an Asahi Kasei company

269 Mill Road  
Chelmsford, MA 01824  
978.421.9655 (main)  
978.421.0025 (fax)  
[zoll.com](http://zoll.com)

Date: February 2, 2022

### **Justification for Re-use of Radio Test Results**

We, ZOLL Medical Corporation, believe that the conducted signal test data on original reports provided by the Quectel Wireless Solutions Company Limited (Single Module, FCC ID: XMR201808EC25AF, approved on August 2, 2018) can be re-used for the following reasons:

1. Applicant has built in the radio module according to the specific radio module integration guides.
2. No changes have been made on radio module except for some functions disabled by software setting.
3. Re-test the conducted power for spot check verification.

Conducted signal tests in following reports re-used.

FCC Part	Reference ID	Report No.	Approval date
Applicable FCC Parts. 1. Part 22 2. Part 24 3. Part 27 4. Part 90	FCC ID: XMR201808EC25AF	1. R1806A0301-R1V1 2. R1806A0301-R2V1 3. R1806A0301-R3V1 4. R1806A0301-R4V1	Approved on August 2, 2018

Sincerely,

\_\_\_\_\_  
Signature

Name: Carl Mamuszka  
Title: Principal Compliance Engineer  
On behalf of ZOLL Medical Corporation  
Phone: +1 978-421-9330