

Modular Approval Declaration Letter

Shenzhen Mindray BIO-Medical electronics Co.,LTD.

Jun. 10, 2020

(Bluetooth module) FCC ID : ZLZ-PANLINK3

is seeking FCC Authorization as a **Single Modular transmitter** / **Single Limited Modular Approval**.

The EUT meets the requirements for **Single Modular approval** / **Single Limited Modular Approval**

as detailed in FCC public Notice DA00-1407. Compliance to each of the requirements is described below:

Item	Modular requirement	Yes	No
1	The radio elements must have the radio frequency circuitry shielded. Physical components and tuning capacitor(s) may be located external to the shield, but must be on the module assembly;	<input checked="" type="checkbox"/>	<input type="checkbox"/>
2	The module must have buffered modulation/data inputs to ensure that the device will comply with Part 15 requirements with any type of input signal;	<input checked="" type="checkbox"/>	<input type="checkbox"/>
3	The module must contain power supply regulation on the module;	<input checked="" type="checkbox"/>	<input type="checkbox"/>
4	The module must contain a permanently attached antenna, or contain a unique antenna connector, and be marketed and operated only with specific antenna(s), per Sections 15.203, 15.204(b), 15.204(c), 15.212(a), 2.929(b);	<input checked="" type="checkbox"/>	<input type="checkbox"/>
5	The module must demonstrate compliance in a stand-alone configuration;	<input checked="" type="checkbox"/>	<input type="checkbox"/>
6	The module must be labelled with its permanently affixed FCC ID label, or use an electronic display (See KDB Publication 784748 about labelling requirements);	<input checked="" type="checkbox"/>	<input type="checkbox"/>
7	The module must comply with all specific rules applicable to the transmitter including all the conditions provided in the integration instructions by the grantee;	<input checked="" type="checkbox"/>	<input type="checkbox"/>
8	The module must comply with RF exposure requirements	<input checked="" type="checkbox"/>	<input type="checkbox"/>

Leo Shao

Client's signature

Leo Shao / Manager

Client's name / title

liuyan_d@mindray.com / Mindray Building, Keji 12th Road South, High-tech Industrial Park, Nanshan, 518057 Shenzhen, PEOPLE'S
REPUBLIC OF CHINA

Contact information / address

(Shenzhen Mindray Bio-Medical Electronics Co.,Ltd.)