

Shenzhen Mindray BIO-Medical electronics Co.,LTD.

Jun. 16, 2020

Equipment Authorization Division  
Federal Communications Commission  
7435 Oakland Mills Road  
Columbia, MD 21046

FCC ID: ZLZ-PMACS

To Whom It May Concern:

Shenzhen Mindray BIO-Medical electronics Co.,LTD., has changed Antenna gain value and Use distance. The hardware and software remain the same, the contents of the certification are the same with the original one.

New antenna information:


Antenna installation: ANT1 FPC dipole  
ANT2 FPC dipole

Antenna Gain: ANT1: 2.6dBi  
ANT2: 0.88dBi

Contact Information:

Shenzhen Mindray BIO-Medical electronics Co.,LTD.  
Mindray Building, Keji 12th Road South, Hi-tech Ind, Shenzhen, China  
Contact Name: Leo Shao  
Email Address: liuyan\_d@mindray.com  
Phone Number: +86-755-81885333

Sincerely,



.....

Signature

Leo Shao / Manager

.....

Typed name and Title