

Shenzhen Mindray BIO-Medical electronics Co.,LTD.

March 27, 2023

To Whom It May Concern:

We, Shenzhen Mindray BIO-Medical electronics Co.,LTD., hereby apply for Class II permissive change to the approved product under FCC ID: ZLZ-WLINK, date of original grant: 05/19/2021 .

Three optional antennas have been added to the original.

Radiated emission is tested again and demonstrates to compliant to the limit per FCC part15.247 and part15.209.

And, according to KDB178919 D01 Permissive Change Policy v06 section II, for the antenna with lowest gain, DFS validation test is performed.

Sincerely,

Shenzhen Mindray BIO-Medical electronics Co.,LTD.

.....  
Signature

Leo Shao / Manager 

.....  
Typed name and Title