### 20150423\_SystemPerformanceCheck-D835V2 SN 4d002

Frequency: 835 MHz; Duty Cycle: 1:1; Room Ambient Temperature: 24.0°C; Liquid Temperature: 23.0°C Medium parameters used: f = 835 MHz;  $\sigma = 0.981$  S/m;  $\epsilon_r = 53.973$ ;  $\rho = 1000$  kg/m<sup>3</sup> DASY5 Configuration:

- Area Scan Setting: Find Secondary Maximum Within: 2.0 dB and with a peak SAR value greater than 0.0012W/kg

Date/Time: 4/23/2015 8:13:33 PM

- Electronics: DAE4 Sn1257: Calibrated: 9/29/2014
- Probe: EX3DV4 SN3772; ConvF(8.46, 8.46, 8.46); Calibrated: 2/23/2015;
- Sensor-Surface: 2.5mm (Mechanical Surface Detection)
- Phantom: ELI v5.0 A; Type: QD OVA 002 AA; Serial: 1180

### Body/Pin=100 mW/Area Scan (7x7x1): Measurement grid: dx=15mm, dy=15mm


Maximum value of SAR (measured) = 1.19 W/kg

#### Body/Pin=100 mW/Zoom Scan (7x7x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm

Reference Value = 36.146 V/m; Power Drift = 0.01 dB

Peak SAR (extrapolated) = 1.46 W/kg

SAR(1 g) = 0.990 W/kg; SAR(10 g) = 0.656 W/kg Maximum value of SAR (measured) = 1.20 W/kg




0 dB = 1.20 W/kg = 0.79 dBW/kg

## 20150423\_SystemPerformanceCheck-D835V2 SN 4d002

Frequency: 835 MHz; Duty Cycle: 1:1

**Body/Pin=100 mW/Z Scan (1x1x21):** Measurement grid: dx=20mm, dy=20mm, dz=5mm Maximum value of SAR (measured) = 1.19 W/kg



### 20150421\_SystemPerformanceCheck-D1900V2 SN 5d043

Frequency: 1900 MHz; Duty Cycle: 1:1; Room Ambient Temperature: 24.0°C; Liquid Temperature: 23.0°C Medium parameters used: f = 1900 MHz;  $\sigma = 1.568$  S/m;  $\epsilon_r = 51.351$ ;  $\rho = 1000$  kg/m<sup>3</sup> DASY5 Configuration:

- Area Scan Setting: Find Secondary Maximum Within: 2.0 dB and with a peak SAR value greater than 0.0012W/kg

Date/Time: 4/21/2015 11:48:20 AM

- Electronics: DAE4 Sn1258: Calibrated: 5/15/2014
- Probe: EX3DV4 SN7335; ConvF(7.9, 7.9, 7.9); Calibrated: 3/13/2015;
- Sensor-Surface: 2.5mm (Mechanical Surface Detection)
- Phantom: ELI v5.0 B; Type: QDOVA002AA; Serial: 1248

#### Body/Pin=100 mW/Area Scan (7x7x1): Measurement grid: dx=15mm, dy=15mm


Maximum value of SAR (measured) = 5.46 W/kg

#### Body/Pin=100 mW/Zoom Scan (7x7x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm

Reference Value = 61.922 V/m; Power Drift = 0.03 dB

Peak SAR (extrapolated) = 7.36 W/kg

SAR(1 g) = 4.05 W/kg; SAR(10 g) = 2.1 W/kg Maximum value of SAR (measured) = 5.45 W/kg



0 dB = 5.45 W/kg = 7.36 dBW/kg

Test Laboratory: UL Verification Services Inc. SAR Lab H Date/Time: 4/21/2015 12:06:34 PM

# 20150421\_SystemPerformanceCheck-D1900V2 SN 5d043

Frequency: 1900 MHz; Duty Cycle: 1:1

**Body/Pin=100 mW/Z Scan (1x1x21):** Measurement grid: dx=20mm, dy=20mm, dz=5mm Maximum value of SAR (measured) = 5.45 W/kg

