

HAC-RF Emission

Communication System: UID 0, CW; Frequency: 835 MHz; Duty Cycle: 1:1
 Phantom section: RF Section
 DASY5 Configuration:

- Probe: ER3DV6 - SN2339; ConvF(1, 1, 1); Calibrated: 2/26/2015;
- Sensor-Surface: (Fix Surface)
- Electronics: DAE3 Sn427; Calibrated: 1/14/2015
- Phantom: HAC Test Arch with AMCC; Type: SD HAC P01 BB
- Measurement SW: DASY52, Version 52.8 (7);SEMCAD X Version 14.6.10 (7164)

Dipole E-Field measurement/835 MHz/Hearing Aid Compatibility Test at 15mm distance (41x361x1): Interpolated grid:

dx=0.5000 mm, dy=0.5000 mm
 Device Reference Point: 0, 0, -6.3 mm
 Reference Value = 118.0 V/m; Power Drift = -0.01 dB
 PMR not calibrated. PMF = 1.000 is applied.
 E-field emissions = 106.2 V/m


Near-field category: M4 (AWF 0 dB)

PMF scaled E-field

Grid 1 M4 104.4 V/m	Grid 2 M4 106.2 V/m	Grid 3 M4 104.5 V/m
Grid 4 M4 60.39 V/m	Grid 5 M4 61.30 V/m	Grid 6 M4 60.56 V/m
Grid 7 M4 103.9 V/m	Grid 8 M4 105.1 V/m	Grid 9 M4 102.7 V/m

Cursor:

Total = 106.2 V/m
 E Category: M4
 Location: 0, -78, 9.7 mm



0 dB = 106.2 V/m = 40.52 dBV/m

HAC-RF Emission

Communication System: UID 0, CW; Frequency: 1880 MHz; Duty Cycle: 1:1
 Phantom section: RF Section
 DASY5 Configuration:

- Probe: ER3DV6 - SN2339; ConvF(1, 1, 1); Calibrated: 2/26/2015;
- Sensor-Surface: (Fix Surface)
- Electronics: DAE3 Sn427; Calibrated: 1/14/2015
- Phantom: HAC Test Arch with AMCC; Type: SD HAC P01 BB
- Measurement SW: DASY52, Version 52.8 (7);SEMCAD X Version 14.6.10 (7164)

Dipole E-Field measurement/1880 MHz/Hearing Aid Compatibility Test at 15mm distance (41x181x1): Interpolated grid:

dx=0.5000 mm, dy=0.5000 mm
 Device Reference Point: 0, 0, -6.3 mm
 Reference Value = 137.9 V/m; Power Drift = 0.02 dB
 PMR not calibrated. PMF = 1.000 is applied.
 E-field emissions = 92.32 V/m


Near-field category: M3 (AWF 0 dB)

PMF scaled E-field

Grid 1 M3 79.66 V/m	Grid 2 M3 80.85 V/m	Grid 3 M3 80.22 V/m
Grid 4 M3 72.46 V/m	Grid 5 M3 73.08 V/m	Grid 6 M3 71.91 V/m
Grid 7 M3 91.95 V/m	Grid 8 M3 92.32 V/m	Grid 9 M3 89.39 V/m

Cursor:

Total = 92.32 V/m
 E Category: M3
 Location: 1.5, 31, 9.7 mm



0 dB = 92.32 V/m = 39.31 dBV/m

HAC-RF Emission

Communication System: UID 0, CW; Frequency: 1880 MHz; Duty Cycle: 1:1

Phantom section: RF Section

DASY5 Configuration:

- Probe: ER3DV6 - SN2339; ConvF(1, 1, 1); Calibrated: 2/26/2015;
- Sensor-Surface: (Fix Surface)
- Electronics: DAE3 Sn427; Calibrated: 1/14/2015
- Phantom: HAC Test Arch with AMCC; Type: SD HAC P01 BB
- Measurement SW: DASY52, Version 52.8 (7);SEMCAD X Version 14.6.10 (7164)

Dipole E-Field measurement/1880 MHz/Hearing Aid Compatibility Test at 15mm distance (41x181x1): Interpolated grid:

dx=0.5000 mm, dy=0.5000 mm

Device Reference Point: 0, 0, -6.3 mm

Reference Value = 133.7 V/m; Power Drift = 0.02 dB

PMR not calibrated. PMF = 1.000 is applied.

E-field emissions = 89.83 V/m

Near-field category: M3 (AWF 0 dB)

PMF scaled E-field


Grid 1 M3 84.09 V/m	Grid 2 M3 84.96 V/m	Grid 3 M3 83.66 V/m
Grid 4 M3 71.36 V/m	Grid 5 M3 72.62 V/m	Grid 6 M3 72.15 V/m
Grid 7 M3 88.36 V/m	Grid 8 M3 89.83 V/m	Grid 9 M3 88.43 V/m

Cursor:

Total = 89.83 V/m

E Category: M3

Location: 0, 30.5, 9.7 mm



0 dB = 89.83 V/m = 39.07 dBV/m

HAC-RF Emission

Communication System: UID 0, CW (0); Frequency: 2600 MHz; Duty Cycle: 1:1
 Phantom section: RF Section
 DASY5 Configuration:
 - Probe: ER3DV6 - SN2339; ConvF(1, 1, 1); Calibrated: 2/26/2015;
 - Sensor-Surface: (Fix Surface)
 - Electronics: DAE3 Sn427; Calibrated: 1/14/2015
 - Phantom: HAC Test Arch with AMCC; Type: SD HAC P01 BB
 - Measurement SW: DASY52, Version 52.8 (7);SEMCAD X Version 14.6.10 (7164)

Dipole E-Field measurement/2600 MHz/Hearing Aid Compatibility Test at 15mm distance (41x181x1): Interpolated grid:

dx=0.5000 mm, dy=0.5000 mm
 Device Reference Point: 0, 0, -6.3 mm
 Reference Value = 68.28 V/m; Power Drift = -0.01 dB
 PMR not calibrated. PMF = 1.000 is applied.
 E-field emissions = 90.99 V/m


Near-field category: M3 (AWF 0 dB)

PMF scaled E-field

Grid 1 M3 84.08 V/m	Grid 2 M3 84.62 V/m	Grid 3 M3 83.17 V/m
Grid 4 M3 87.27 V/m	Grid 5 M3 88.25 V/m	Grid 6 M3 86.62 V/m
Grid 7 M3 90.18 V/m	Grid 8 M3 90.99 V/m	Grid 9 M3 88.79 V/m

Cursor:

Total = 90.99 V/m
 E Category: M3
 Location: 1, 20, 9.7 mm



0 dB = 90.99 V/m = 39.18 dBV/m